

Cover Page

Order ID : Q2263

Project ID : NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258

Client : AECOM Technical Services, Inc.

Lab Sample Number

Q2263-01
Q2263-02

Client Sample Number

RW9-MW01D3-20250606
RW9-MW01S-20250606

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature : _____

Date: 6/20/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

AECOM Technical Services, Inc.

Project Name: NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258

Project # N/A

Order ID # Q2263

Test Name: VOCMS Group1

A. Number of Samples and Date of Receipt:

2 Water samples were received on 06/06/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: SVOC-SIMGroup1 and VOCMS Group1. This data package contains results for VOCMS Group1.

C. Analytical Techniques:

The analysis performed on instrument MSVOA_N were done using GC column Rxi-624SIL MS 30m, 0.25mm, 1.4 um, Cat. #13868. The analysis of VOCMS Group1 was based on method 8260D.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria.

The Internal Standards Areas met the acceptable requirements.

The Retention Times were acceptable for all samples.

The RPD met criteria.

The Blank Spike met requirements for all samples.

The Blank Spike Duplicate met requirements for all samples.

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements.

The Continuous Calibration File ID VN086940.D met the requirements except for Bromoform is failing high but no positive hit in associate sample therefore no corrective action taken.

The Tuning criteria met requirements.



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

E. Additional Comments:

Samples for MS/MSD for VOC analysis were not provided with this set of samples. The Blank Spike Duplicate is reported with the data.

Trip Blank was not provided with this set of samples.

The not QT review data is reported in the Miscellaneous.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <20% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 20% for the Initial Calibration curve for SW-846 analysis.

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following “ Results Qualifiers” are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. “10 U”. This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as “12 B”.
E	Indicates the analyte ‘s concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a “P”.
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

ALLIANCE 284 Sheffield Street, Mountainside New Jersey 07092

NEW JERSEY LAB ID#: 20012; NEW YORK LAB ID#: 11376

GC/MS VOA CONFORMANCE/NON-CONFORMANCE SUMMARY

ORDER ID: Q2263

MATRIX: Water

METHOD: 8260D

		NA	NO	YES
1.	Chromatograms Labeled/Compounds Identified. (Field samples and Method Blanks)			✓
2.	GC/MS Tuning Specifications BFB Meet Criteria (NOTE THAT THERE ARE DIFFERENT CRITERIA FOR NY ASP CLP, CLP AND NJ)			✓
3.	GC/MS Tuning Frequency - Performed every 24 hours for 600 series and 12 hours for 8000 Series.			✓
4.	GC/MS Calibration - Initial Calibration performed before sample analysis and continuing calibration performed within 24 hours of sample analysis for 600 series and 12 hours for 8000 series.			✓
5.	GC/MS Calibration Requirements. The Initial Calibration met the requirements. The Continuous Calibration File ID VN086940.D met the requirements except for Bromoform is failing high but no positive hit in associate sample therefore no corrective action taken.			✓
6.	Blank Contamination - If yes, list compounds and concentrations in each blank:		✓	
7.	Surrogate Recoveries Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable ranges.			✓
8.	Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range.			✓
9.	Internal Standard Area/Retention Time Shift Meet Criteria Comments:			✓
10.	Analysis Holding Time Met If not met, list number of days exceeded for each sample:			✓

ALLIANCE 284 Sheffield Street, Mountainside New Jersey 07092

NEW JERSEY LAB ID#: 20012; NEW YORK LAB ID#: 11376

GC/MS VOA CONFORMANCE/NON-CONFORMANCE SUMMARY (CONTINUED)

NA NO YES

ADDITIONAL COMMENTS:

Samples for MS/MSD for VOC analysis were not provided with this set of samples. The Blank Spike Duplicate is reported with the data.

Trip Blank was not provided with this set of samples.

The not QT review data is reported in the Miscellaneous.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <20% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is >20% for the Initial Calibration curve for SW-846 analysis.

QA REVIEW

Date

APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: Q2263

Completed

For thorough review, the report must have the following:

GENERAL:

Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)

✓

Check chain-of-custody for proper relinquish/return of samples

✓

Is the chain of custody signed and complete

✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts

✓

Collect information for each project id from server. Were all requirements followed

✓

COVER PAGE:

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page

✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody

✓

CHAIN OF CUSTODY:

Do requested analyses on Chain of Custody agree with form I results

✓

Do requested analyses on Chain of Custody agree with the log-in page

✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Custody

✓

Were the samples received within hold time

✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle

✓

ANALYTICAL:

Was method requirement followed?

✓

Was client requirement followed?

✓

Does the case narrative summarize all QC failure?

✓

All runlogs and manual integration are reviewed for requirements

✓

All manual calculations and /or hand notations verified

✓

QA Review Signature: SOHIL JODHANI

Date: 06/20/2025

LAB CHRONICLE

OrderID:	Q2263	OrderDate:	6/6/2025 12:42:00 PM
Client:	AECOM Technical Services, Inc.	Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258
Contact:	Eleanor Vivadou	Location:	D21,VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2263-01	RW9-MW01D3-20250606	Water	VOCMS Group1	8260-Low	06/06/25		06/11/25	06/06/25
Q2263-02	RW9-MW01S-20250606	Water	VOCMS Group1	8260-Low	06/06/25		06/11/25	06/06/25



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: Q2263
Client: AECOM Technical Services, Inc.

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	LOD	RDL	Units
Client ID:	RW9-MW01D3-20250606								
Q2263-01	RW9-MW01D3-20250606	Water	Acetone	1.90	J	1.50	3.80	5.00	ug/L
			Total Voc :	1.90					
			Total Concentration:	1.90					



QC SUMMARY

Surrogate Summary

SDG No.: Q2263

Client: AECOM Technical Services, Inc.

Analytical Method: SW8260-Low

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q2263-01	RW9-MW01D3-20250606	1,2-Dichloroethane-d4	50	50.6	101	81	118
		Dibromofluoromethane	50	50.8	101	80	119
		Toluene-d8	50	51.9	104	89	112
		4-Bromofluorobenzene	50	50.0	100	85	114
Q2263-02	RW9-MW01S-20250606	1,2-Dichloroethane-d4	50	53.8	108	81	118
		Dibromofluoromethane	50	51.3	103	80	119
		Toluene-d8	50	51.7	103	89	112
		4-Bromofluorobenzene	50	50.0	100	85	114
VN0611WBL01	VN0611WBL01	1,2-Dichloroethane-d4	50	48.3	97	81	118
		Dibromofluoromethane	50	49.3	99	80	119
		Toluene-d8	50	51.7	103	89	112
		4-Bromofluorobenzene	50	49.4	99	85	114
VN0611WBS02	VN0611WBS02	1,2-Dichloroethane-d4	50	47.1	94	81	118
		Dibromofluoromethane	50	51.1	102	80	119
		Toluene-d8	50	48.7	97	89	112
		4-Bromofluorobenzene	50	50.0	100	85	114
VN0611WBSD0	VN0611WBSD02	1,2-Dichloroethane-d4	50	47.9	96	81	118
		Dibromofluoromethane	50	52.3	105	80	119
		Toluene-d8	50	49.5	99	89	112
		4-Bromofluorobenzene	50	50.1	100	85	114

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2263

Client: AECOM Technical Services, Inc.

Analytical Method: SW8260-Low

Datafile : VN086944.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN0611WBS02	Dichlorodifluoromethane	20	19.7	ug/L	99			32	152	
	Chloromethane	20	16.1	ug/L	81			50	139	
	Vinyl chloride	20	18.9	ug/L	95			58	137	
	Bromomethane	20	16.6	ug/L	83			53	141	
	Chloroethane	20	19.4	ug/L	97			60	138	
	Trichlorofluoromethane	20	19.2	ug/L	96			65	141	
	1,1,2-Trichlorotrifluoroethane	20	19.6	ug/L	98			70	136	
	1,1-Dichloroethene	20	19.2	ug/L	96			71	131	
	Acetone	100	99.0	ug/L	99			39	160	
	Carbon disulfide	20	17.8	ug/L	89			64	133	
	Methyl tert-butyl Ether	20	19.6	ug/L	98			71	124	
	Methyl Acetate	20	19.0	ug/L	95			56	136	
	Methylene Chloride	20	19.0	ug/L	95			74	124	
	trans-1,2-Dichloroethene	20	18.5	ug/L	93			75	124	
	1,1-Dichloroethane	20	19.4	ug/L	97			77	125	
	Cyclohexane	20	17.6	ug/L	88			71	130	
	2-Butanone	100	90.4	ug/L	90			56	143	
	Carbon Tetrachloride	20	19.2	ug/L	96			72	136	
	cis-1,2-Dichloroethene	20	19.5	ug/L	98			78	123	
	Bromochloromethane	20	19.9	ug/L	100			78	123	
	Chloroform	20	19.5	ug/L	98			79	124	
	1,1,1-Trichloroethane	20	19.0	ug/L	95			74	131	
	Methylcyclohexane	20	16.9	ug/L	85			72	132	
	Benzene	20	19.3	ug/L	97			79	120	
	1,2-Dichloroethane	20	19.5	ug/L	98			73	128	
	Trichloroethene	20	19.9	ug/L	100			79	123	
	1,2-Dichloropropane	20	19.6	ug/L	98			78	122	
	Bromodichloromethane	20	19.6	ug/L	98			79	125	
	4-Methyl-2-Pentanone	100	98.0	ug/L	98			67	130	
	Toluene	20	19.5	ug/L	98			80	121	
	t-1,3-Dichloropropene	20	19.9	ug/L	100			73	127	
	cis-1,3-Dichloropropene	20	20.1	ug/L	101			75	124	
	1,1,2-Trichloroethane	20	20.4	ug/L	102			80	119	
	2-Hexanone	100	89.2	ug/L	89			57	139	
	Dibromochloromethane	20	20.5	ug/L	103			74	126	
	1,2-Dibromoethane	20	19.7	ug/L	99			77	121	
	Tetrachloroethene	20	19.2	ug/L	96			74	129	
	Chlorobenzene	20	20.0	ug/L	100			82	118	
	Ethyl Benzene	20	19.6	ug/L	98			79	121	
	m/p-Xylenes	40	39.6	ug/L	99			80	121	
	o-Xylene	20	20.2	ug/L	101			78	122	
	Styrene	20	20.1	ug/L	101			78	123	
	Bromoform	20	21.0	ug/L	105			66	130	
	Isopropylbenzene	20	19.3	ug/L	97			72	131	
	1,1,2,2-Tetrachloroethane	20	21.0	ug/L	105			71	121	
	1,3-Dichlorobenzene	20	20.3	ug/L	102			80	119	
	1,4-Dichlorobenzene	20	20.4	ug/L	102			79	118	
	1,2-Dichlorobenzene	20	20.0	ug/L	100			80	119	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2263

Client: AECOM Technical Services, Inc.

Analytical Method: SW8260-Low

Datafile : VN086944.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	RPD
									High	
VN0611WBS02	1,2-Dibromo-3-Chloropropane	20	19.2	ug/L	96			62	128	
	1,2,4-Trichlorobenzene	20	18.4	ug/L	92			69	130	
	1,2,3-Trichlorobenzene	20	17.8	ug/L	89			69	129	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2263

Client: AECOM Technical Services, Inc.

Analytical Method: SW8260-Low

Datafile : VN086945.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN0611WBSD02	Dichlorodifluoromethane	20	19.1	ug/L	96	3		32	152	20
	Chloromethane	20	15.5	ug/L	78	4		50	139	20
	Vinyl chloride	20	18.7	ug/L	94	1		58	137	20
	Bromomethane	20	16.2	ug/L	81	2		53	141	20
	Chloroethane	20	18.9	ug/L	95	2		60	138	20
	Trichlorofluoromethane	20	19.0	ug/L	95	1		65	141	20
	1,1,2-Trichlorotrifluoroethane	20	18.7	ug/L	94	4		70	136	20
	1,1-Dichloroethene	20	19.1	ug/L	96	0		71	131	20
	Acetone	100	92.4	ug/L	92	7		39	160	20
	Carbon disulfide	20	17.2	ug/L	86	3		64	133	20
	Methyl tert-butyl Ether	20	20.2	ug/L	101	3		71	124	20
	Methyl Acetate	20	20.1	ug/L	101	6		56	136	20
	Methylene Chloride	20	19.1	ug/L	96	1		74	124	20
	trans-1,2-Dichloroethene	20	18.6	ug/L	93	0		75	124	20
	1,1-Dichloroethane	20	19.3	ug/L	97	0		77	125	20
	Cyclohexane	20	17.2	ug/L	86	2		71	130	20
	2-Butanone	100	94.9	ug/L	95	5		56	143	20
	Carbon Tetrachloride	20	19.5	ug/L	98	2		72	136	20
	cis-1,2-Dichloroethene	20	19.6	ug/L	98	0		78	123	20
	Bromochloromethane	20	21.1	ug/L	106	6		78	123	20
	Chloroform	20	19.4	ug/L	97	1		79	124	20
	1,1,1-Trichloroethane	20	18.8	ug/L	94	1		74	131	20
	Methylcyclohexane	20	16.6	ug/L	83	2		72	132	20
	Benzene	20	19.7	ug/L	99	2		79	120	20
	1,2-Dichloroethane	20	20.3	ug/L	102	4		73	128	20
	Trichloroethene	20	20.1	ug/L	101	1		79	123	20
	1,2-Dichloropropane	20	20.0	ug/L	100	2		78	122	20
	Bromodichloromethane	20	20.7	ug/L	104	6		79	125	20
	4-Methyl-2-Pentanone	100	100	ug/L	100	2		67	130	20
	Toluene	20	19.9	ug/L	100	2		80	121	20
	t-1,3-Dichloropropene	20	20.7	ug/L	104	4		73	127	20
	cis-1,3-Dichloropropene	20	20.7	ug/L	104	3		75	124	20
	1,1,2-Trichloroethane	20	20.9	ug/L	104	2		80	119	20
	2-Hexanone	100	93.8	ug/L	94	5		57	139	20
	Dibromochloromethane	20	21.2	ug/L	106	3		74	126	20
	1,2-Dibromoethane	20	20.7	ug/L	104	5		77	121	20
	Tetrachloroethene	20	18.5	ug/L	93	3		74	129	20
	Chlorobenzene	20	20.2	ug/L	101	1		82	118	20
	Ethyl Benzene	20	19.3	ug/L	97	1		79	121	20
	m/p-Xylenes	40	39.5	ug/L	99	0		80	121	20
	o-Xylene	20	20.0	ug/L	100	1		78	122	20
	Styrene	20	20.6	ug/L	103	2		78	123	20
	Bromoform	20	22.4	ug/L	112	6		66	130	20
	Isopropylbenzene	20	19.3	ug/L	97	0		72	131	20
	1,1,2,2-Tetrachloroethane	20	21.7	ug/L	109	4		71	121	20
	1,3-Dichlorobenzene	20	20.5	ug/L	103	1		80	119	20
	1,4-Dichlorobenzene	20	20.4	ug/L	102	0		79	118	20
	1,2-Dichlorobenzene	20	20.6	ug/L	103	3		80	119	20

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2263

Client: AECOM Technical Services, Inc.

Analytical Method: SW8260-Low

Datafile : VN086945.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN0611WBSD02	1,2-Dibromo-3-Chloropropane	20	20.8	ug/L	104	8		62	128	20
	1,2,4-Trichlorobenzene	20	18.7	ug/L	94	2		69	130	20
	1,2,3-Trichlorobenzene	20	18.1	ug/L	91	2		69	129	20

VOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

VN0611WBL01

Lab Name: CHEMTECH

Contract: AECO15

Lab Code: CHEM Case No.: Q2263

SAS No.: Q2263 SDG NO.: Q2263

Lab File ID: VN086942.D

Lab Sample ID: VN0611WBL01

Date Analyzed: 06/11/2025

Time Analyzed: 12:28

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: (Y/N) N

Instrument ID: MSVOA_N

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
VN0611WBS02	VN0611WBS02	VN086944.D	06/11/2025
VN0611WBSD02	VN0611WBSD02	VN086945.D	06/11/2025
RW9-MW01D3-20250606	Q2263-01	VN086963.D	06/11/2025
RW9-MW01S-20250606	Q2263-02	VN086964.D	06/11/2025

COMMENTS: _____



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: AECO15
Lab Code: CHEM Case No.: Q2263 SAS No.: Q2263 SDG NO.: Q2263
Lab File ID: VN086861.D BFB Injection Date: 06/06/2025
Instrument ID: MSVOA_N BFB Injection Time: 07:59
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	17.3
75	30.0 - 60.0% of mass 95	48.1
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.4
173	Less than 2.0% of mass 174	0.7 (1) 1
174	50.0 - 100.0% of mass 95	66.6
175	5.0 - 9.0% of mass 174	4.7 (7.1) 1
176	95.0 - 101.0% of mass 174	65.3 (98.1) 1
177	5.0 - 9.0% of mass 176	4.4 (6.8) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDIC001	VSTDIC001	VN086862.D	06/06/2025	12:44
VSTDIC005	VSTDIC005	VN086863.D	06/06/2025	13:17
VSTDIC020	VSTDIC020	VN086864.D	06/06/2025	13:40
VSTDIC050	VSTDIC050	VN086865.D	06/06/2025	14:03
VSTDIC100	VSTDIC100	VN086866.D	06/06/2025	14:26
VSTDIC150	VSTDIC150	VN086867.D	06/06/2025	14:49



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: AECO15
Lab Code: CHEM Case No.: Q2263 SAS No.: Q2263 SDG NO.: Q2263
Lab File ID: VN086939.D BFB Injection Date: 06/11/2025
Instrument ID: MSVOA_N BFB Injection Time: 10:22
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	15.7
75	30.0 - 60.0% of mass 95	46.5
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	7.1
173	Less than 2.0% of mass 174	0.7 (1) 1
174	50.0 - 100.0% of mass 95	71.1
175	5.0 - 9.0% of mass 174	5.4 (7.5) 1
176	95.0 - 101.0% of mass 174	68.8 (96.8) 1
177	5.0 - 9.0% of mass 176	4.8 (7) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VN086940.D	06/11/2025	11:32
VN0611WBL01	VN0611WBL01	VN086942.D	06/11/2025	12:28
VN0611WBS02	VN0611WBS02	VN086944.D	06/11/2025	13:27
VN0611WBSD02	VN0611WBSD02	VN086945.D	06/11/2025	14:01
RW9-MW01D3-20250606	Q2263-01	VN086963.D	06/11/2025	20:39
RW9-MW01S-20250606	Q2263-02	VN086964.D	06/11/2025	21:01
VSTDCCC050EC	VSTDCCC050	VN086965.D	06/11/2025	21:23

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: AECO15
Lab Code: CHEM Case No.: Q2263 SAS No.: Q2263 SDG NO.: Q2263
Lab File ID: VN086940.D Date Analyzed: 06/11/2025
Instrument ID: MSVOA_N Time Analyzed: 11:32
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	205098	8.23	357643	9.11	312173	11.87
UPPER LIMIT	410196	8.73	715286	9.606	624346	12.365
LOWER LIMIT	102549	7.73	178822	8.606	156087	11.365
EPA SAMPLE NO.						
RW9-MW01D3-20250606	255804	8.24	486937	9.11	433933	11.87
RW9-MW01S-20250606	212082	8.24	417259	9.11	370384	11.87
VN0611WBL01	328120	8.23	618651	9.11	543433	11.87
VN0611WBS02	206799	8.24	370524	9.11	321188	11.87
VN0611WBSD02	199904	8.23	354069	9.11	310598	11.87

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: AECO15
 Lab Code: CHEM Case No.: Q2263 SAS No.: Q2263 SDG NO.: Q2263
 Lab File ID: VN086940.D Date Analyzed: 06/11/2025
 Instrument ID: MSVOA_N Time Analyzed: 11:32
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	152136	13.788				
UPPER LIMIT	304272	14.288				
LOWER LIMIT	76068	13.288				
EPA SAMPLE NO.						
RW9-MW01D3-20250606	207409	13.79				
RW9-MW01S-20250606	174250	13.79				
VN0611WBL01	257257	13.79				
VN0611WBS02	157403	13.79				
VN0611WBSD02	151691	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.



SAMPLE DATA

Report of Analysis

Client:	AECOM Technical Services, Inc.			Date Collected:	06/06/25		
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258			Date Received:	06/06/25		
Client Sample ID:	RW9-MW01D3-20250606			SDG No.:	Q2263		
Lab Sample ID:	Q2263-01			Matrix:	Water		
Analytical Method:	8260D			% Solid:	0		
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000	uL	
Soil Aliquot Vol:			uL	Test:	VOCMS Group1		
GC Column:	RXI-624	ID :	0.25	Level :	LOW		
Prep Method :							

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086963.D	1		06/11/25 20:39	VN061125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	1.90	J	1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.22	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.			Date Collected:	06/06/25		
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258			Date Received:	06/06/25		
Client Sample ID:	RW9-MW01D3-20250606			SDG No.:	Q2263		
Lab Sample ID:	Q2263-01			Matrix:	Water		
Analytical Method:	8260D			% Solid:	0		
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000	uL	
Soil Aliquot Vol:			uL	Test:	VOCMS Group1		
GC Column:	RXI-624	ID :	0.25	Level :	LOW		
Prep Method :							

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086963.D	1		06/11/25 20:39	VN061125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	50.6		81 - 118		101%	SPK: 50
1868-53-7	Dibromofluoromethane	50.7		80 - 119		101%	SPK: 50
2037-26-5	Toluene-d8	51.9		89 - 112		104%	SPK: 50
460-00-4	4-Bromofluorobenzene	50.0		85 - 114		100%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	256000	8.235				
540-36-3	1,4-Difluorobenzene	487000	9.106				
3114-55-4	Chlorobenzene-d5	434000	11.865				
3855-82-1	1,4-Dichlorobenzene-d4	207000	13.794				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	06/06/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	06/06/25	
Client Sample ID:	RW9-MW01D3-20250606		SDG No.:	Q2263	
Lab Sample ID:	Q2263-01		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086963.D	1		06/11/25 20:39	VN061125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061125\
 Data File : VN086963.D
 Acq On : 11 Jun 2025 20:39
 Operator : JC\MD
 Sample : Q2263-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RW9-MW01D3-20250606

Quant Time: Jun 12 01:47:14 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	8.235	168	255804	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	486937	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	433933	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	207409	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	173262	50.588	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 101.180%			
35) Dibromofluoromethane	8.177	113	146437	50.746	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 101.500%			
50) Toluene-d8	10.571	98	592829	51.894	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 103.780%			
62) 4-Bromofluorobenzene	12.847	95	212200	49.997	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 100.000%			
Target Compounds						
16) Acetone	4.453	43	3409	1.877	ug/l #	82
95) Naphthalene	15.641	128	10758	0.691	ug/l	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061125\
Data File : VN086963.D
Acq On : 11 Jun 2025 20:39
Operator : JC\MD
Sample : Q2263-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RW9-MW01D3-20250606

Quant Time: Jun 12 01:47:14 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 02:12:50 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.235 min Scan# 1068
Delta R.T. -0.001 min
Lab File: VN086963.D
Acq: 11 Jun 2025 20:39




Instrument :
MSVOA_N
ClientSampleId :
RW9-MW01D3-20250606


Tgt Ion:168 Resp: 255804
Ion Ratio Lower Upper
168 100
99 58.0 49.1 73.7



#16
Acetone
Concen: 1.877 ug/l
RT: 4.453 min Scan# 425
Delta R.T. 0.006 min
Lab File: VN086963.D
Acq: 11 Jun 2025 20:39

Tgt Ion: 43 Resp: 3409
Ion Ratio Lower Upper
43 100
58 24.6 28.0 42.0#





#33

1,2-Dichloroethane-d4

Concen: 50.588 ug/l

RT: 8.588 min Scan# 1128

Delta R.T. -0.000 min

Lab File: VN086963.D

Acq: 11 Jun 2025 20:39

Instrument :

MSVOA_N

ClientSampleId :





RW9-MW01D3-20250606

Tgt Ion: 65 Resp: 173262

Ion Ratio Lower Upper

65 100

67 54.5 0.0 105.6



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.106 min Scan# 1216

Delta R.T. -0.000 min

Lab File: VN086963.D

Acq: 11 Jun 2025 20:39




Tgt Ion: 114 Resp: 486937


Ion Ratio Lower Upper

114 100

63 20.1 0.0 39.6

88 15.7 0.0 30.2





#35

Dibromofluoromethane

Concen: 50.746 ug/l

RT: 8.177 min Scan# 1057

Delta R.T. 0.006 min

Lab File: VN086963.D

Acq: 11 Jun 2025 20:39

Instrument :

MSVOA_N

ClientSampleId :

RW9-MW01D3-20250606





Tgt Ion:113 Resp: 146437

Ion Ratio Lower Upper

113 100

111 100.8 84.2 126.2

192 17.0 14.2 21.4



#50

Toluene-d8

Concen: 51.894 ug/l

RT: 10.571 min Scan# 1465

Delta R.T. -0.000 min

Lab File: VN086963.D




Acq: 11 Jun 2025 20:39


Tgt Ion: 98 Resp: 592829

Ion Ratio Lower Upper

98 100

100 66.1 53.4 80.0









#62
4-Bromofluorobenzene
Concen: 49.997 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. -0.000 min
Lab File: VN086963.D
Acq: 11 Jun 2025 20:39




Instrument :
MSVOA_N
ClientSampleId :
RW9-MW01D3-20250606


Tgt Ion: 95 Resp: 212200
Ion Ratio Lower Upper
95 100
174 70.4 0.0 141.8
176 68.6 0.0 132.6



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 1685
Delta R.T. -0.000 min
Lab File: VN086963.D
Acq: 11 Jun 2025 20:39

Tgt Ion: 117 Resp: 433933
Ion Ratio Lower Upper
117 100
82 53.9 44.6 67.0
119 32.3 25.5 38.3









#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.794 min Scan# 2013
Delta R.T. -0.000 min
Lab File: VN086963.D
Acq: 11 Jun 2025 20:39




Instrument :
MSVOA_N
ClientSampleId :
RW9-MW01D3-20250606

Tgt Ion:152 Resp: 207409
Ion Ratio Lower Upper
152 100
115 59.9 30.1 90.5
150 158.7 0.0 345.0



#95
Naphthalene
Concen: 0.691 ug/l
RT: 15.641 min Scan# 2327
Delta R.T. -0.000 min
Lab File: VN086963.D
Acq: 11 Jun 2025 20:39

Tgt Ion:128 Resp: 10758
Ion Ratio Lower Upper
128 100
127 12.0 10.2 15.2
129 10.7 8.8 13.2



Report of Analysis

Client:	AECOM Technical Services, Inc.			Date Collected:	06/06/25		
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258			Date Received:	06/06/25		
Client Sample ID:	RW9-MW01S-20250606			SDG No.:	Q2263		
Lab Sample ID:	Q2263-02			Matrix:	Water		
Analytical Method:	8260D			% Solid:	0		
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000	uL	
Soil Aliquot Vol:			uL	Test:	VOCMS Group1		
GC Column:	RXI-624	ID :	0.25	Level :	LOW		
Prep Method :							

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086964.D	1		06/11/25 21:01	VN061125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	3.80	U	1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.22	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	06/06/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	06/06/25	
Client Sample ID:	RW9-MW01S-20250606		SDG No.:	Q2263	
Lab Sample ID:	Q2263-02		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086964.D	1		06/11/25 21:01	VN061125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	53.8		81 - 118		108%	SPK: 50
1868-53-7	Dibromofluoromethane	51.3		80 - 119		103%	SPK: 50
2037-26-5	Toluene-d8	51.7		89 - 112		103%	SPK: 50
460-00-4	4-Bromofluorobenzene	50.0		85 - 114		100%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	212000	8.235				
540-36-3	1,4-Difluorobenzene	417000	9.106				
3114-55-4	Chlorobenzene-d5	370000	11.865				
3855-82-1	1,4-Dichlorobenzene-d4	174000	13.794				

Report of Analysis

Client:	AECOM Technical Services, Inc.	Date Collected:	06/06/25
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258	Date Received:	06/06/25
Client Sample ID:	RW9-MW01S-20250606	SDG No.:	Q2263
Lab Sample ID:	Q2263-02	Matrix:	Water
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086964.D	1		06/11/25 21:01	VN061125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061125\
 Data File : VN086964.D
 Acq On : 11 Jun 2025 21:01
 Operator : JC\MD
 Sample : Q2263-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RW9-MW01S-20250606

Quant Time: Jun 12 01:47:39 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	8.235	168	212082	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	417259	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	370384	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	174250	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	152843	53.827	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 107.660%		
35) Dibromofluoromethane	8.177	113	126870	51.307	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 102.620%		
50) Toluene-d8	10.571	98	506520	51.743	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 103.480%		
62) 4-Bromofluorobenzene	12.847	95	181778	49.981	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 99.960%		
Target Compounds						
95) Naphthalene	15.641	128	8833	0.675	ug/l	Qvalue 99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061125\
Data File : VN086964.D
Acq On : 11 Jun 2025 21:01
Operator : JC\MD
Sample : Q2263-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RW9-MW01S-20250606

Quant Time: Jun 12 01:47:39 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 02:12:50 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.235 min Scan# 1068
Delta R.T. -0.001 min
Lab File: VN086964.D
Acq: 11 Jun 2025 21:01




Instrument :
MSVOA_N
ClientSampleId :
RW9-MW01S-20250606


Tgt Ion:168 Resp: 212082
Ion Ratio Lower Upper
168 100
99 57.6 49.1 73.7



#33
1,2-Dichloroethane-d4
Concen: 53.827 ug/l
RT: 8.588 min Scan# 1128
Delta R.T. -0.000 min
Lab File: VN086964.D
Acq: 11 Jun 2025 21:01

Tgt Ion: 65 Resp: 152843
Ion Ratio Lower Upper
65 100
67 54.6 0.0 105.6





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.106 min Scan# 11

Delta R.T. -0.000 min

Lab File: VN086964.D

Acq: 11 Jun 2025 21:01

Instrument :

MSVOA_N

ClientSampleId :

RW9-MW01S-20250606





Tgt Ion:114 Resp: 417259

Ion Ratio Lower Upper

114 100

63 20.3 0.0 39.6

88 15.8 0.0 30.2



#35

Dibromofluoromethane

Concen: 51.307 ug/l

RT: 8.177 min Scan# 1058

Delta R.T. 0.006 min

Lab File: VN086964.D

Acq: 11 Jun 2025 21:01




Tgt Ion:113 Resp: 126870


Ion Ratio Lower Upper

113 100

111 102.8 84.2 126.2

192 17.4 14.2 21.4









#50
Toluene-d8
Concen: 51.743 ug/l
RT: 10.571 min Scan# 1465
Delta R.T. -0.000 min
Lab File: VN086964.D
Acq: 11 Jun 2025 21:01




Instrument :
MSVOA_N
ClientSampleId :
RW9-MW01S-20250606


Tgt Ion: 98 Resp: 506520
Ion Ratio Lower Upper
98 100
100 66.6 53.4 80.0



#62
4-Bromofluorobenzene
Concen: 49.981 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. -0.000 min
Lab File: VN086964.D
Acq: 11 Jun 2025 21:01

Tgt Ion: 95 Resp: 181778
Ion Ratio Lower Upper
95 100
174 70.3 0.0 141.8
176 67.5 0.0 132.6





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.865 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086964.D

Acq: 11 Jun 2025 21:01

Instrument :

MSVOA_N

ClientSampleId :

RW9-MW01S-20250606





Tgt Ion:117 Resp: 370384

Ion Ratio Lower Upper

117 100

82 55.4 44.6 67.0

119 33.0 25.5 38.3



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.794 min Scan# 2013

Delta R.T. -0.000 min

Lab File: VN086964.D

Acq: 11 Jun 2025 21:01




Tgt Ion:152 Resp: 174250


Ion Ratio Lower Upper

152 100

115 61.2 30.1 90.5

150 159.1 0.0 345.0





#95

Naphthalene

Concen: 0.675 ug/l

RT: 15.641 min Scan# 21

Delta R.T. -0.000 min

Lab File: VN086964.D


Acq: 11 Jun 2025 21:01

Instrument :

MSVOA_N

ClientSampleId :

RW9-MW01S-20250606





Tgt Ion:128 Resp: 8833

Ion	Ratio	Lower	Upper

128	100		

127	12.1	10.2	15.2

129	10.7	8.8	13.2





CALIBRATION SUMMARY



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: AECO15
 Lab Code: CHEM Case No.: Q2263 SAS No.: Q2263 SDG No.: Q2263
 Instrument ID: MSVOA_N Calibration Date(s): 06/06/2025 06/06/2025
 Heated Purge: (Y/N) N Calibration Time(s): 12:44 14:49
 GC Column: RXI-624 ID: 0.25 (mm)

LAB FILE ID:		RRF001 = VN086862.D		RRF005 = VN086863.D		RRF020 = VN086864.D			
		RRF050 = VN086865.D		RRF100 = VN086866.D		RRF150 = VN086867.D			
COMPOUND	RRF001	RRF005	RRF020	RRF050	RRF100	RRF150	RRF	% RSD	
Dichlorodifluoromethane	0.467	0.444	0.539	0.501	0.535	0.506	0.499	7.5	
Chloromethane	0.762	0.654	0.645	0.597	0.617	0.587	0.644	9.9	
Vinyl Chloride	0.670	0.670	0.684	0.640	0.673	0.648	0.664	2.5	
Bromomethane		0.375	0.380	0.357	0.379	0.368	0.372	2.6	
Chloroethane	0.460	0.444	0.442	0.408	0.418	0.402	0.429	5.4	
Trichlorofluoromethane	0.882	0.903	0.904	0.834	0.858	0.825	0.868	3.9	
1,1,2-Trichlorotrifluoroethane	0.554	0.567	0.563	0.519	0.546	0.520	0.545	3.8	
1,1-Dichloroethene	0.573	0.593	0.563	0.533	0.550	0.527	0.557	4.4	
Acetone	0.426	0.366	0.366	0.322	0.334	0.316	0.355	11.5	
Carbon Disulfide	1.718	1.622	1.542	1.426	1.496	1.433	1.539	7.4	
Methyl tert-butyl Ether	2.120	2.038	2.051	1.933	2.021	1.926	2.015	3.7	
Methyl Acetate	1.035	1.049	1.078	0.986	1.049	1.011	1.035	3.1	
Methylene Chloride	0.822	0.688	0.643	0.605	0.629	0.601	0.665	12.5	
trans-1,2-Dichloroethene	0.700	0.674	0.621	0.567	0.591	0.561	0.619	9.3	
1,1-Dichloroethane	1.192	1.153	1.156	1.063	1.110	1.043	1.120	5.2	
Cyclohexane		1.303	1.116	1.004	1.030	0.976	1.086	12.2	
2-Butanone	0.604	0.598	0.604	0.551	0.573	0.533	0.577	5.2	
Carbon Tetrachloride	0.453	0.449	0.434	0.409	0.435	0.421	0.433	3.9	
cis-1,2-Dichloroethene	0.786	0.766	0.762	0.699	0.729	0.701	0.740	4.9	
Bromochloromethane	0.579	0.564	0.616	0.466	0.517	0.560	0.550	9.5	
Chloroform	1.235	1.152	1.145	1.061	1.085	1.030	1.118	6.7	
1,1,1-Trichloroethane	1.029	0.995	0.969	0.895	0.925	0.893	0.951	5.9	
Methylcyclohexane	0.633	0.645	0.588	0.570	0.603	0.589	0.605	4.8	
Benzene	1.588	1.501	1.444	1.345	1.414	1.371	1.444	6.2	
1,2-Dichloroethane	0.473	0.456	0.444	0.411	0.430	0.413	0.438	5.6	
Trichloroethene	0.359	0.360	0.341	0.327	0.340	0.328	0.342	4.2	
1,2-Dichloropropane	0.366	0.369	0.354	0.332	0.352	0.335	0.351	4.4	
Bromodichloromethane	0.510	0.484	0.480	0.457	0.483	0.465	0.480	3.8	
4-Methyl-2-Pentanone	0.505	0.549	0.576	0.538	0.562	0.528	0.543	4.6	
Toluene	0.918	0.914	0.885	0.835	0.883	0.859	0.882	3.6	

* Compounds with required minimum RRF and maximum %RSD values.
 All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: AECO15
 Lab Code: CHEM Case No.: Q2263 SAS No.: Q2263 SDG No.: Q2263
 Instrument ID: MSVOA_N Calibration Date(s): 06/06/2025 06/06/2025
 Heated Purge: (Y/N) N Calibration Time(s): 12:44 14:49
 GC Column: RXI-624 ID: 0.25 (mm)

LAB FILE ID:		RRF001 = VN086862.D		RRF005 = VN086863.D		RRF020 = VN086864.D		
		RRF050 = VN086865.D		RRF100 = VN086866.D		RRF150 = VN086867.D		
COMPOUND	RRF001	RRF005	RRF020	RRF050	RRF100	RRF150	RRF	% RSD
t-1,3-Dichloropropene	0.571	0.526	0.524	0.522	0.548	0.530	0.537	3.6
cis-1,3-Dichloropropene	0.609	0.577	0.564	0.551	0.584	0.561	0.574	3.6
1,1,2-Trichloroethane	0.359	0.355	0.342	0.322	0.335	0.323	0.340	4.7
2-Hexanone	0.312	0.282	0.363	0.368	0.397	0.376	0.350	12.4
Dibromochloromethane	0.361	0.351	0.358	0.340	0.363	0.349	0.354	2.5
1,2-Dibromoethane	0.353	0.358	0.354	0.329	0.354	0.341	0.348	3.2
Tetrachloroethene	0.355	0.331	0.312	0.294	0.313	0.293	0.316	7.5
Chlorobenzene	1.233	1.135	1.107	1.023	1.089	1.030	1.103	7
Ethyl Benzene	1.989	1.975	1.907	1.796	1.913	1.816	1.900	4.2
m/p-Xylenes	0.730	0.751	0.741	0.701	0.736	0.703	0.727	2.8
o-Xylene	0.714	0.699	0.702	0.674	0.711	0.678	0.696	2.4
Styrene	1.177	1.193	1.226	1.162	1.229	1.164	1.192	2.5
Bromoform	0.217	0.266	0.276	0.265	0.286	0.267	0.263	9.1
Isopropylbenzene	3.864	3.749	3.649	3.426	3.621	3.546	3.643	4.2
1,1,2,2-Tetrachloroethane	1.292	1.299	1.273	1.178	1.205	1.157	1.234	5
1,3-Dichlorobenzene	1.763	1.709	1.657	1.554	1.612	1.566	1.644	5
1,4-Dichlorobenzene	1.820	1.786	1.657	1.572	1.642	1.576	1.676	6.3
1,2-Dichlorobenzene	1.675	1.651	1.596	1.500	1.557	1.496	1.579	4.8
1,2-Dibromo-3-Chloropropane	0.339	0.317	0.290	0.272	0.283	0.270	0.295	9.3
1,2,4-Trichlorobenzene	1.042	1.037	0.991	0.969	1.016	0.994	1.008	2.8
1,2,3-Trichlorobenzene	1.073	1.009	0.993	0.950	1.000	0.987	1.002	4
1,2-Dichloroethane-d4		0.732	0.707	0.500	0.656	0.751	0.669	15.1
Dibromofluoromethane		0.303	0.310	0.219	0.298	0.351	0.296	16.2
Toluene-d8		1.245	1.203	0.861	1.178	1.377	1.173	16.2
4-Bromofluorobenzene		0.441	0.446	0.325	0.446	0.521	0.436	16.2

* Compounds with required minimum RRF and maximum %RSD values.
 All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.

Method Path : Z:\voasrv\HPCHEM1\MSVOA_N\methods\
 Method File : 82N060625W.M
 Title : SW846 8260
 Last Update : Sat Jun 07 02:12:50 2025
 Response Via : Initial Calibration

Calibration Files

1 =VN086862.D 5 =VN086863.D 20 =VN086864.D 50 =VN086865.D 100 =VN086866.D 150 =VN086867.D

	Compound	1	5	20	50	100	150	Avg	%RSD
1) I	Pentafluorobenzene	-----ISTD-----							
2) T	Dichlorodifluo...	0.467	0.444	0.539	0.501	0.535	0.506	0.499	7.51
3) P	Chloromethane	0.762	0.654	0.645	0.597	0.617	0.587	0.644	9.85
4) C	Vinyl Chloride	0.670	0.670	0.684	0.640	0.673	0.648	0.664	2.49#
5) T	Bromomethane		0.375	0.380	0.357	0.379	0.368	0.372	2.57
6) T	Chloroethane	0.460	0.444	0.442	0.408	0.418	0.402	0.429	5.37
7) T	Trichlorofluor...	0.882	0.903	0.904	0.834	0.858	0.825	0.868	3.93
8) T	Diethyl Ether	0.372	0.397	0.383	0.363	0.384	0.369	0.378	3.28
9) T	1,1,2-Trichlor...	0.554	0.567	0.563	0.519	0.546	0.520	0.545	3.83
10) T	Methyl Iodide		0.720	0.729	0.683	0.722	0.679	0.707	3.35
11) T	Tert butyl alc...		0.192	0.194	0.176	0.180	0.166	0.182	6.37
12) CM	1,1-Dichloroet...	0.573	0.593	0.563	0.533	0.550	0.527	0.557	4.44#
13) T	Acrolein		0.073	0.051	0.045	0.052	0.066	0.057	20.26
14) T	Allyl chloride	0.985	0.911	0.925	0.865	0.905	0.947	0.923	4.40
15) T	Acrylonitrile	0.434	0.434	0.445	0.407	0.423	0.404	0.425	3.82
16) T	Acetone	0.426	0.366	0.366	0.322	0.334	0.316	0.355	11.53
17) T	Carbon Disulfide	1.718	1.621	1.542	1.426	1.496	1.433	1.539	7.38
18) T	Methyl Acetate	1.035	1.049	1.078	0.986	1.049	1.011	1.035	3.10
19) T	Methyl tert-bu...	2.120	2.038	2.051	1.933	2.021	1.926	2.015	3.69
20) T	Methylene Chlo...	0.822	0.688	0.643	0.605	0.629	0.601	0.665	12.52
21) T	trans-1,2-Dich...	0.700	0.674	0.621	0.567	0.591	0.561	0.619	9.25
22) T	Diisopropyl ether	2.018	2.020	2.036	1.856	1.915	1.828	1.945	4.70
23) T	Vinyl Acetate	1.743	1.692	1.715	1.591	1.604	1.517	1.644	5.29
24) P	1,1-Dichloroet...	1.192	1.152	1.156	1.063	1.110	1.043	1.120	5.17
25) T	2-Butanone	0.604	0.598	0.604	0.551	0.573	0.533	0.577	5.18
26) T	2,2-Dichloropr...	0.967	0.796	0.778	0.905	0.912	0.868	0.871	8.32
27) T	cis-1,2-Dichlo...	0.786	0.766	0.762	0.699	0.729	0.701	0.740	4.93
28) T	Bromochloromet...	0.579	0.564	0.616	0.466	0.517	0.560	0.550	9.48
29) T	Tetrahydrofuran	0.390	0.390	0.399	0.358	0.372	0.346	0.376	5.46
30) C	Chloroform	1.235	1.151	1.145	1.061	1.085	1.030	1.118	6.66#
31) T	Cyclohexane		1.303	1.116	1.004	1.030	0.976	1.086	12.21
32) T	1,1,1-Trichlor...	1.029	0.995	0.969	0.895	0.925	0.893	0.951	5.88
33) S	1,2-Dichloroet...		0.732	0.707	0.500	0.656	0.751	0.669	15.08
34) I	1,4-Difluorobenzene	-----ISTD-----							
35) S	Dibromofluorom...		0.303	0.310	0.219	0.298	0.351	0.296	16.15
36) T	1,1-Dichloropr...	0.467	0.458	0.438	0.418	0.442	0.426	0.442	4.23
37) T	Ethyl Acetate	0.544	0.615	0.577	0.536	0.565	0.535	0.562	5.48
38) T	Carbon Tetrach...	0.453	0.449	0.434	0.409	0.435	0.421	0.433	3.88
39) T	Methylcyclohexane	0.633	0.645	0.588	0.570	0.603	0.589	0.605	4.75
40) TM	Benzene	1.588	1.501	1.444	1.345	1.414	1.371	1.444	6.20
41) T	Methacrylonitrile	0.356	0.319	0.318	0.303	0.299	0.299	0.316	6.84
42) TM	1,2-Dichloroet...	0.473	0.456	0.444	0.411	0.430	0.413	0.438	5.60
43) T	Isopropyl Acetate	0.975	0.950	0.906	0.852	0.884	0.845	0.902	5.79
44) TM	Trichloroethene	0.359	0.360	0.341	0.327	0.340	0.327	0.342	4.17
45) C	1,2-Dichloropr...	0.366	0.369	0.354	0.332	0.352	0.335	0.351	4.40#
46) T	Dibromomethane	0.233	0.242	0.241	0.221	0.237	0.226	0.233	3.54
47) T	Bromodichlorom...	0.510	0.484	0.480	0.457	0.483	0.465	0.480	3.80
48) T	Methyl methacr...	0.444	0.415	0.421	0.394	0.421	0.397	0.415	4.40
49) T	1,4-Dioxane	0.006	0.007	0.008	0.008	0.008	0.007	0.008	9.40
50) S	Toluene-d8		1.245	1.203	0.861	1.178	1.377	1.173	16.23
51) T	4-Methyl-2-Pen...	0.505	0.549	0.576	0.538	0.562	0.528	0.543	4.64
52) CM	Toluene	0.918	0.914	0.885	0.835	0.883	0.859	0.882	3.61#
53) T	t-1,3-Dichloro...	0.571	0.526	0.524	0.522	0.548	0.530	0.537	3.59
54) T	cis-1,3-Dichlo...	0.609	0.577	0.564	0.551	0.584	0.561	0.574	3.61
55) T	1,1,2-Trichlor...	0.359	0.355	0.342	0.322	0.335	0.323	0.340	4.66
56) T	Ethyl methacry...	0.433	0.516	0.567	0.556	0.595	0.578	0.541	10.92

Method Path : Z:\voasrv\HPCHEM1\MSVOA_N\methods\
 Method File : 82N060625W.M

57)	T	1,3-Dichloropr...	0.643	0.600	0.587	0.561	0.583	0.559	0.589	5.27
58)	T	2-Chloroethyl ...	0.278	0.305	0.334	0.274	0.340	0.405	0.323	15.07
59)	T	2-Hexanone	0.312	0.282	0.363	0.368	0.397	0.376	0.350	12.36
60)	T	Dibromochlorom...	0.361	0.351	0.358	0.340	0.363	0.349	0.354	2.49
61)	T	1,2-Dibromoethane	0.353	0.358	0.354	0.329	0.354	0.341	0.348	3.18
62)	S	4-Bromofluorob...	0.441	0.446	0.325	0.446	0.521	0.436		16.16
63)	I	Chlorobenzene-d5	-----ISTD-----							
64)	T	Tetrachloroethene	0.355	0.331	0.312	0.294	0.313	0.293	0.316	7.51
65)	PM	Chlorobenzene	1.233	1.135	1.107	1.023	1.089	1.030	1.103	7.01
66)	T	1,1,1,2-Tetrac...	0.362	0.374	0.358	0.338	0.356	0.338	0.354	3.99
67)	C	Ethyl Benzene	1.989	1.975	1.907	1.796	1.913	1.816	1.899	4.19#
68)	T	m/p-Xylenes	0.730	0.751	0.741	0.701	0.736	0.703	0.727	2.82
69)	T	o-Xylene	0.714	0.699	0.702	0.674	0.711	0.678	0.696	2.42
70)	T	Styrene	1.177	1.193	1.226	1.162	1.229	1.164	1.192	2.51
71)	P	Bromoform	0.217	0.266	0.276	0.265	0.286	0.267	0.263	9.09
72)	I	1,4-Dichlorobenzen...	-----ISTD-----							
73)	T	Isopropylbenzene	3.864	3.749	3.649	3.426	3.621	3.546	3.643	4.21
74)	T	N-amyl acetate	1.376	1.110	1.187	1.266	1.372	1.327	1.273	8.42
75)	P	1,1,2,2-Tetrac...	1.292	1.299	1.273	1.178	1.205	1.157	1.234	4.99
76)	T	1,2,3-Trichlor...	1.297	1.189	1.129	1.173	1.201	1.142	1.189	5.03
77)	T	Bromobenzene	0.870	0.855	0.840	0.790	0.838	0.819	0.835	3.36
78)	T	n-propylbenzene	4.530	4.627	4.449	4.211	4.424	4.317	4.427	3.35
79)	T	2-Chlorotoluene	2.852	2.723	2.674	2.507	2.618	2.554	2.655	4.69
80)	T	1,3,5-Trimethy...	3.004	3.091	3.035	2.898	3.064	2.953	3.007	2.39
81)	T	trans-1,4-Dich...	0.510	0.522	0.476	0.546	0.528	0.516		5.06
82)	T	4-Chlorotoluene	2.899	2.757	2.671	2.531	2.664	2.595	2.686	4.81
83)	T	tert-Butylbenzene	2.942	2.810	2.736	2.617	2.745	2.668	2.753	4.14
84)	T	1,2,4-Trimethy...	2.952	3.122	3.063	2.914	3.075	2.968	3.016	2.73
85)	T	sec-Butylbenzene	4.114	4.126	4.037	3.829	4.018	3.861	3.998	3.15
86)	T	p-Isopropyltol...	3.305	3.425	3.321	3.184	3.366	3.227	3.305	2.68
87)	T	1,3-Dichlorobe...	1.763	1.709	1.657	1.554	1.612	1.566	1.644	5.01
88)	T	1,4-Dichlorobe...	1.820	1.786	1.657	1.572	1.642	1.576	1.676	6.27
89)	T	n-Butylbenzene	3.473	3.292	3.140	3.059	3.188	3.054	3.201	4.99
90)	T	Hexachloroethane	0.580	0.537	0.569	0.537	0.577	0.562	0.560	3.43
91)	T	1,2-Dichlorobe...	1.675	1.651	1.596	1.500	1.557	1.496	1.579	4.77
92)	T	1,2-Dibromo-3-...	0.339	0.317	0.290	0.272	0.283	0.270	0.295	9.29
93)	T	1,2,4-Trichlor...	1.042	1.037	0.991	0.969	1.016	0.994	1.008	2.82
94)	T	Hexachlorobuta...	0.364	0.391	0.385	0.363	0.382	0.369	0.376	3.11
95)	T	Naphthalene	3.820	3.727	3.761	3.600	3.839	3.772	3.753	2.27
96)	T	1,2,3-Trichlor...	1.073	1.009	0.993	0.950	1.000	0.987	1.002	4.05

(#) = Out of Range

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060625\
 Data File : VN086862.D
 Acq On : 06 Jun 2025 12:44
 Operator : JC\MD
 Sample : VSTDICC001
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC001

Manual Integrations
 APPROVED

Reviewed By : John Carlone 06/09/2025
 Supervised By : Mahesh Dadoda 06/09/2025

Quant Time: Jun 07 01:54:26 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 01:51:02 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	212682	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	393457	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	348010	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	167399	50.000	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	0.000	65	0d	0.000	ug/l	
Spiked Amount	50.000	Range	74 - 125	Recovery	=	0.000%#
35) Dibromofluoromethane	0.000	113	0d	0.000	ug/l	
Spiked Amount	50.000	Range	75 - 124	Recovery	=	0.000%#
50) Toluene-d8	0.000	98	0d	0.000	ug/l	
Spiked Amount	50.000	Range	86 - 113	Recovery	=	0.000%#
62) 4-Bromofluorobenzene	0.000	95	0d	0.000	ug/l	
Spiked Amount	50.000	Range	77 - 121	Recovery	=	0.000%#

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	2.153	85	1985	0.936	ug/l	72
3) Chloromethane	2.400	50	3241	1.184	ug/l	98
4) Vinyl Chloride	2.559	62	2848	1.008	ug/l #	88
6) Chloroethane	3.159	64	1957	1.073	ug/l #	79
7) Trichlorofluoromethane	3.530	101	3752	1.017	ug/l	97
8) Diethyl Ether	3.994	74	1583	0.984	ug/l	76
9) 1,1,2-Trichlorotrifluo...	4.377	101	2358m	1.017	ug/l	
12) 1,1-Dichloroethene	4.365	96	2436	1.029	ug/l	85
14) Allyl chloride	5.041	41	4189	1.067	ug/l #	75
15) Acrylonitrile	5.735	53	9222	5.106	ug/l	96
16) Acetone	4.459	43	9069	6.006	ug/l	95
17) Carbon Disulfide	4.730	76	7306	1.116	ug/l	97
18) Methyl Acetate	5.053	43	4402	1.000	ug/l #	66
19) Methyl tert-butyl Ether	5.824	73	9019	1.052	ug/l	92
20) Methylene Chloride	5.294	84	3496	1.237	ug/l	93
21) trans-1,2-Dichloroethene	5.800	96	2979	1.131	ug/l	87
22) Diisopropyl ether	6.682	45	8584	1.037	ug/l #	96
23) Vinyl Acetate	6.618	43	37075	5.303	ug/l	95
24) 1,1-Dichloroethane	6.588	63	5071	1.065	ug/l #	81
25) 2-Butanone	7.500	43	12841	5.231	ug/l	93
26) 2,2-Dichloropropane	7.500	77	4112	1.110	ug/l	94
27) cis-1,2-Dichloroethene	7.494	96	3345	1.062	ug/l	95
28) Bromochloromethane	7.824	49	2464	1.052	ug/l #	88
29) Tetrahydrofuran	7.853	42	8301	5.191	ug/l	100
30) Chloroform	7.971	83	5255	1.105	ug/l	97
32) 1,1,1-Trichloroethane	8.171	97	4378	1.082	ug/l #	50
36) 1,1-Dichloropropene	8.376	75	3678	1.059	ug/l	96
37) Ethyl Acetate	7.577	43	4282	0.968	ug/l #	78
38) Carbon Tetrachloride	8.376	117	3563	1.045	ug/l #	93
39) Methylcyclohexane	9.606	83	4981	1.046	ug/l #	90
40) Benzene	8.612	78	12498	1.100	ug/l	100
41) Methacrylonitrile	7.782	41	2799	1.127	ug/l	89
42) 1,2-Dichloroethane	8.682	62	3725	1.081	ug/l	82
43) Isopropyl Acetate	8.694	43	7673	1.081	ug/l #	96
44) Trichloroethene	9.365	130	2822	1.047	ug/l	71
45) 1,2-Dichloropropane	9.629	63	2883	1.043	ug/l	95

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060625\
 Data File : VN086862.D
 Acq On : 06 Jun 2025 12:44
 Operator : JC\MD
 Sample : VSTDICC001
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC001

Manual Integrations
 APPROVED

Reviewed By : John Carlone 06/09/2025
 Supervised By : Mahesh Dadoda 06/09/2025

Quant Time: Jun 07 01:54:26 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 01:51:02 2025
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) Dibromomethane	9.712	93	1835	0.999	ug/l	94
47) Bromodichloromethane	9.888	83	4017	1.063	ug/l #	96
48) Methyl methacrylate	9.688	41	3492	1.069	ug/l	88
49) 1,4-Dioxane	9.706	88	1004	16.891	ug/l #	74
51) 4-Methyl-2-Pentanone	10.453	43	19886	4.653	ug/l	99
52) Toluene	10.635	92	7227	1.041	ug/l	95
53) t-1,3-Dichloropropene	10.841	75	4494	1.064	ug/l	97
54) cis-1,3-Dichloropropene	10.312	75	4792	1.060	ug/l	97
55) 1,1,2-Trichloroethane	11.023	97	2828	1.058	ug/l	89
56) Ethyl methacrylate	10.882	69	3406	0.800	ug/l #	85
57) 1,3-Dichloropropane	11.170	76	5062	1.092	ug/l	98
58) 2-Chloroethyl Vinyl ether	10.165	63	10924	4.304	ug/l	97
59) 2-Hexanone	11.235	43	12295m	4.466	ug/l	
60) Dibromochloromethane	11.359	129	2843	1.021	ug/l	96
61) 1,2-Dibromoethane	11.470	107	2777	1.014	ug/l	96
64) Tetrachloroethene	11.106	164	2473	1.123	ug/l	94
65) Chlorobenzene	11.894	112	8581	1.118	ug/l #	87
66) 1,1,1,2-Tetrachloroethane	11.959	131	2523	1.023	ug/l #	63
67) Ethyl Benzene	11.970	91	13845	1.047	ug/l	94
68) m/p-Xylenes	12.070	106	10163	2.008	ug/l	95
69) o-Xylene	12.400	106	4970	1.025	ug/l	93
70) Styrene	12.417	104	8189	0.987	ug/l	98
71) Bromoform	12.582	173	1508	0.825	ug/l #	78
73) Isopropylbenzene	12.694	105	12938	1.061	ug/l	98
74) N-amyl acetate	12.647	43	4608m	1.138	ug/l	
75) 1,1,2,2-Tetrachloroethane	12.935	83	4326	1.047	ug/l #	93
76) 1,2,3-Trichloropropane	13.000	75	4343m	1.088	ug/l	
77) Bromobenzene	12.982	156	2913	1.042	ug/l	96
78) n-propylbenzene	13.035	91	15168	1.023	ug/l	98
79) 2-Chlorotoluene	13.123	91	9550	1.074	ug/l	97
80) 1,3,5-Trimethylbenzene	13.170	105	10058	0.999	ug/l	96
82) 4-Chlorotoluene	13.223	91	9706	1.079	ug/l	96
83) tert-Butylbenzene	13.435	119	9849	1.069	ug/l	98
84) 1,2,4-Trimethylbenzene	13.482	105	9882	0.979	ug/l	99
85) sec-Butylbenzene	13.617	105	13775	1.029	ug/l	100
86) p-Isopropyltoluene	13.729	119	11065	1.000	ug/l	97
87) 1,3-Dichlorobenzene	13.735	146	5903	1.073	ug/l	94
88) 1,4-Dichlorobenzene	13.811	146	6094m	1.086	ug/l	
89) n-Butylbenzene	14.058	91	11627	1.085	ug/l	94
90) Hexachloroethane	14.329	117	1942	1.036	ug/l	93
91) 1,2-Dichlorobenzene	14.106	146	5607	1.061	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.717	75	1136	1.150	ug/l	78
93) 1,2,4-Trichlorobenzene	15.388	180	3489	1.034	ug/l	98
94) Hexachlorobutadiene	15.500	225	1219	0.969	ug/l	88
95) Naphthalene	15.635	128	12789	1.018	ug/l	98
96) 1,2,3-Trichlorobenzene	15.835	180	3594	1.072	ug/l	96


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001

Manual Integrations


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.235 min Scan# 1068
Delta R.T. -0.001 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001



Tgt Ion:168 Resp: 21268
Ion Ratio Lower Upper
168 100
99 58.1 49.1 73.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#2
Dichlorodifluoromethane
Concen: 0.936 ug/l
RT: 2.153 min Scan# 34
Delta R.T. -0.000 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44


Tgt Ion: 85 Resp: 1985
Ion Ratio Lower Upper
85 100
87 48.3 16.2 48.6





#3
Chloromethane
Concen: 1.184 ug/l
RT: 2.400 min Scan# 70
Delta R.T. 0.006 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001



Tgt Ion: 50 Resp: 324
Ion Ratio Lower Upper
50 100
52 30.7 25.7 38.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#4
Vinyl Chloride
Concen: 1.008 ug/l
RT: 2.559 min Scan# 103
Delta R.T. 0.006 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44


Tgt Ion: 62 Resp: 2848
Ion Ratio Lower Upper
62 100
64 25.9 26.0 39.0#





#6
Chloroethane
Concen: 1.073 ug/l
RT: 3.159 min Scan# 205
Delta R.T. 0.006 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001



Tgt Ion: 64 Resp: 195
Ion Ratio Lower Upper
64 100
66 42.3 24.6 36.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#7
Trichlorofluoromethane
Concen: 1.017 ug/l
RT: 3.530 min Scan# 268
Delta R.T. 0.006 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44


Tgt Ion:101 Resp: 3752
Ion Ratio Lower Upper
101 100
103 67.2 52.2 78.2





#8
Diethyl Ether
Concen: 0.984 ug/l
RT: 3.994 min Scan# 345
Delta R.T. 0.011 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001



Tgt Ion: 74 Resp: 158
Ion Ratio Lower Upper
74 100
45 113.8 45.5 136.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#9
1,1,2-Trichlorotrifluoroethane
Concen: 1.017 ug/l m
RT: 4.377 min Scan# 412
Delta R.T. -0.018 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44

Tgt Ion:101 Resp: 2358
Ion Ratio Lower Upper
101 100
85 27.8 33.8 50.8#
151 72.1 58.7 88.1





#12

1,1-Dichloroethene

Concen: 1.029 ug/l

RT: 4.365 min Scan# 410

Delta R.T. 0.006 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001

Tgt Ion: 96 Resp: 2430

Ion Ratio Lower Upper

96 100

61 137.1 125.8 188.6





98 55.8 52.5 78.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#14

Allyl chloride

Concen: 1.067 ug/l

RT: 5.041 min Scan# 525

Delta R.T. -0.000 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44




Tgt Ion: 41 Resp: 4189


Ion Ratio Lower Upper

41 100

39 35.3 51.3 76.9#


76 36.4 31.4 47.0





#15
Acrylonitrile
Concen: 5.106 ug/l
RT: 5.735 min Scan# 644
Delta R.T. -0.006 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001



Tgt Ion: 53 Resp: 922
Ion Ratio Lower Upper
53 100
52 85.4 65.0 97.6
51 32.7 27.2 40.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#16
Acetone
Concen: 6.006 ug/l
RT: 4.459 min Scan# 426
Delta R.T. 0.011 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44

Tgt Ion: 43 Resp: 9069
Ion Ratio Lower Upper
43 100
58 38.0 28.0 42.0





#17

Carbon Disulfide

Concen: 1.116 ug/l

RT: 4.730 min Scan# 41

Delta R.T. -0.000 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC001

Tgt Ion: 76 Resp: 7300

Ion Ratio Lower Upper

76 100





78 10.3 7.3 10.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#18

Methyl Acetate

Concen: 1.000 ug/l

RT: 5.053 min Scan# 527

Delta R.T. 0.011 min

Lab File: VN086862.D




Acq: 06 Jun 2025 12:44


Tgt Ion: 43 Resp: 4402

Ion Ratio Lower Upper

43 100

74 8.7 20.7 31.1#





#19

Methyl tert-butyl Ether

Concen: 1.052 ug/l

RT: 5.824 min Scan# 656

Delta R.T. 0.011 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001

Tgt Ion: 73 Resp: 9019

Ion Ratio Lower Upper

73 100





57 18.0 17.4 26.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#20

Methylene Chloride

Concen: 1.237 ug/l

RT: 5.294 min Scan# 568

Delta R.T. -0.000 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44

Tgt Ion: 84 Resp: 3496




Ion Ratio Lower Upper


84 100

49 125.7 90.5 135.7

51 35.9 28.5 42.7

86 60.5 48.1 72.1





#21

trans-1,2-Dichloroethene

Concen: 1.131 ug/l

RT: 5.800 min Scan# 61

Delta R.T. -0.006 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001

Tgt Ion: 96 Resp: 2979

Ion Ratio Lower Upper

96 100

61 119.4 109.2 163.8





98 54.3 49.8 74.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#22

Diisopropyl ether

Concen: 1.037 ug/l

RT: 6.682 min Scan# 804

Delta R.T. -0.006 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44

Tgt Ion: 45 Resp: 8584




Ion Ratio Lower Upper


45 100

43 51.5 41.9 62.9

87 24.5 22.4 33.6


59 8.4 9.8 14.6#





#23
Vinyl Acetate
Concen: 5.303 ug/l
RT: 6.618 min Scan# 793
Delta R.T. -0.000 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001



Tgt Ion: 43 Resp: 3707
Ion Ratio Lower Upper
43 100
86 10.0 9.5 14.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#24
1,1-Dichloroethane
Concen: 1.065 ug/l
RT: 6.588 min Scan# 788
Delta R.T. 0.006 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44

Tgt Ion: 63 Resp: 5071
Ion Ratio Lower Upper
63 100
98 0.0 3.6 10.8#
100 0.0 2.6 7.8#









#25
2-Butanone
Concen: 5.231 ug/l
RT: 7.500 min Scan# 942
Delta R.T. 0.006 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001

Tgt Ion: 43 Resp: 1284
Ion Ratio Lower Upper
43 100
72 32.7 23.0 34.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#26
2,2-Dichloropropane
Concen: 1.110 ug/l
RT: 7.500 min Scan# 943
Delta R.T. -0.000 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44

Tgt Ion: 77 Resp: 4112
Ion Ratio Lower Upper
77 100
97 20.1 11.6 34.7





#27

cis-1,2-Dichloroethene

Concen: 1.062 ug/l

RT: 7.494 min Scan# 942

Delta R.T. -0.006 min

Lab File: VN086862.D


Acq: 06 Jun 2025 12:44

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC001






Tgt Ion:	96	Resp:	334
Ion	Ratio	Lower	Upper
96	100		
61	134.6	0.0	278.0
98	58.7	0.0	128.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#28

Bromochloromethane


Concen: 1.052 ug/l

RT: 7.824 min Scan# 998



Delta R.T. -0.000 min


Lab File: VN086862.D

Acq: 06 Jun 2025 12:44



Tgt Ion:	49	Resp:	2464
Ion	Ratio	Lower	Upper
49	100		
129	0.0	0.0	3.8
130	67.5	62.2	93.2





#29

Tetrahydrofuran

Concen: 5.191 ug/l

RT: 7.853 min Scan# 1003

Delta R.T. -0.000 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001





Tgt Ion:	42	Resp:	830
Ion	Ratio	Lower	Upper
42	100		
72	47.2	37.8	56.6
71	44.4	35.6	53.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#30

Chloroform

Concen: 1.105 ug/l




RT: 7.971 min Scan# 1023


Delta R.T. -0.006 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44

Tgt Ion:	83	Resp:	5255
Ion	Ratio	Lower	Upper
83	100		
85	68.6	52.9	79.3





#32

1,1,1-Trichloroethane

Concen: 1.082 ug/l

RT: 8.171 min Scan# 1058

Delta R.T. -0.006 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC001

Tgt Ion: 97 Resp: 4378

Ion Ratio Lower Upper

97 100

99 0.0 52.0 78.0





61 42.5 35.9 53.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#33

1,2-Dichloroethane-d4

Concen: 0.000 ug/l

Expected RT: 8.59 min

Lab File: VN086862.D


Acq: 06 Jun 2025 12:44


Tgt Ion: 65

Sig Exp Ratio

65 100

67 52.8





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.106 min Scan# 11

Delta R.T. -0.000 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC001

Tgt Ion:114 Resp: 39345

Ion Ratio Lower Upper

114 100

63 20.2 0.0 39.6





88 16.0 0.0 30.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#35

Dibromofluoromethane

Concen: 0.000 ug/l

Expected RT: 8.17 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44


Tgt Ion: 113


Sig Exp Ratio

113 100

111 105.2

192 17.8





#36

1,1-Dichloropropene

Concen: 1.059 ug/l

RT: 8.376 min Scan# 1092

Delta R.T. -0.000 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001





Tgt Ion:	75	Resp:	3678
Ion	Ratio	Lower	Upper
75	100		
110	31.8	17.6	52.9
77	30.2	25.3	37.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#37

Ethyl Acetate

Concen: 0.968 ug/l




RT: 7.577 min Scan# 956


Delta R.T. -0.000 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44


Tgt Ion:	43	Resp:	4282
Ion	Ratio	Lower	Upper
43	100		
61	0.0	12.1	18.1#
70	14.1	9.8	14.6





#38
Carbon Tetrachloride
Concen: 1.045 ug/l
RT: 8.376 min Scan# 1091
Delta R.T. 0.006 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001



Tgt Ion: 117 Resp: 356
Ion Ratio Lower Upper
117 100
119 89.0 75.0 112.6
121 23.8 24.9 37.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#39
Methylcyclohexane
Concen: 1.046 ug/l
RT: 9.606 min Scan# 1301
Delta R.T. -0.000 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44

Tgt Ion: 83 Resp: 4981
Ion Ratio Lower Upper
83 100
55 79.7 61.0 91.6
98 61.6 38.9 58.3





#40

Benzene

Concen: 1.100 ug/l

RT: 8.612 min Scan# 1132

Delta R.T. -0.000 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001

Tgt Ion: 78 Resp: 12498

Ion Ratio Lower Upper

78 100





77 23.2 18.7 28.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#41

Methacrylonitrile

Concen: 1.127 ug/l

RT: 7.782 min Scan# 991

Delta R.T. -0.006 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44

Tgt Ion: 41 Resp: 2799




Ion Ratio Lower Upper


41 100

39 50.3 41.2 61.8

67 62.4 61.4 92.2

52 22.9 22.7 34.1





#42

1,2-Dichloroethane

Concen: 1.081 ug/l

RT: 8.682 min Scan# 1144

Delta R.T. 0.006 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001

Tgt Ion: 62 Resp: 3725

Ion Ratio Lower Upper

62 100





98 3.9 0.0 21.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#43

Isopropyl Acetate

Concen: 1.081 ug/l

RT: 8.694 min Scan# 1146

Delta R.T. -0.000 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44




Tgt Ion: 43 Resp: 7673


Ion Ratio Lower Upper

43 100

61 26.2 22.1 33.1


87 11.7 11.8 17.6#





#44
Trichloroethene
Concen: 1.047 ug/l
RT: 9.365 min Scan# 1260
Delta R.T. 0.006 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001



Tgt Ion:130 Resp: 282
Ion Ratio Lower Upper
130 100
95 69.1 0.0 196.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#45
1,2-Dichloropropane
Concen: 1.043 ug/l
RT: 9.629 min Scan# 1305
Delta R.T. -0.000 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44

Tgt Ion: 63 Resp: 2883
Ion Ratio Lower Upper
63 100
65 29.0 25.5 38.3





#46

Dibromomethane

Concen: 0.999 ug/l

RT: 9.712 min Scan# 1319

Delta R.T. -0.006 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC001

Tgt Ion: 93 Resp: 183

Ion Ratio Lower Upper

93 100

95 87.6 67.1 100.7





174 100.5 74.2 111.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#47

Bromodichloromethane

Concen: 1.063 ug/l

RT: 9.888 min Scan# 1349

Delta R.T. -0.006 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44




Tgt Ion: 83 Resp: 4017


Ion Ratio Lower Upper

83 100

85 62.7 50.9 76.3

127 0.0 6.6 10.0#





#48

Methyl methacrylate

Concen: 1.069 ug/l

RT: 9.688 min Scan# 1315

Delta R.T. 0.006 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001





Tgt Ion:	41	Resp:	349
Ion Ratio	Lower	Upper	
41	100		
69	79.4	74.6	112.0
39	44.9	40.2	60.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#49

1,4-Dioxane

Concen: 16.891 ug/l




RT: 9.706 min Scan# 1318


Delta R.T. 0.006 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44

Tgt Ion:	88	Resp:	1004
Ion Ratio	Lower	Upper	
88	100		
43	0.0	0.0	0.0
58	94.0	57.7	86.5





#50
Toluene-d8
Concen: 0.000 ug/l
Expected RT: 10.57 min



Lab File: VN086862.D
Acq: 06 Jun 2025 12:44

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001

Tgt Ion: 98
Sig Exp Ratio
98 100
100 66.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#51
4-Methyl-2-Pentanone
Concen: 4.653 ug/l
RT: 10.453 min Scan# 1445
Delta R.T. 0.006 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44

Tgt Ion: 43 Resp: 19886
Ion Ratio Lower Upper
43 100
58 42.3 34.2 51.4









#52
Toluene
Concen: 1.041 ug/l
RT: 10.635 min Scan# 1476
Delta R.T. -0.000 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001

Tgt Ion: 92 Resp: 722
Ion Ratio Lower Upper
92 100
91 178.1 136.5 204.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#53
t-1,3-Dichloropropene
Concen: 1.064 ug/l
RT: 10.841 min Scan# 1511
Delta R.T. -0.000 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44


Tgt Ion: 75 Resp: 4494
Ion Ratio Lower Upper
75 100
77 31.4 26.4 39.6





#54
cis-1,3-Dichloropropene
Concen: 1.060 ug/l
RT: 10.312 min Scan# 1421
Delta R.T. -0.006 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001



Tgt Ion: 75 Resp: 4792
Ion Ratio Lower Upper
75 100
77 35.9 26.2 39.4
39 41.1 32.3 48.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#55
1,1,2-Trichloroethane
Concen: 1.058 ug/l
RT: 11.023 min Scan# 1542
Delta R.T. 0.006 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44


Tgt Ion: 97 Resp: 2828
Ion Ratio Lower Upper
97 100
83 80.6 69.7 104.5
85 46.5 45.1 67.7
99 52.5 50.6 75.8





#56
Ethyl methacrylate
Concen: 0.800 ug/l
RT: 10.882 min Scan# 1518
Delta R.T. -0.000 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001



Tgt Ion: 69 Resp: 3400
Ion Ratio Lower Upper
69 100
41 73.5 47.2 70.8
39 31.6 23.0 34.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#57
1,3-Dichloropropane
Concen: 1.092 ug/l
RT: 11.170 min Scan# 1567
Delta R.T. 0.006 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44

Tgt Ion: 76 Resp: 5062
Ion Ratio Lower Upper
76 100
78 32.8 25.3 37.9









#58
2-Chloroethyl Vinyl ether
Concen: 4.304 ug/l
RT: 10.165 min Scan# 1396
Delta R.T. -0.000 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001

Tgt Ion: 63 Resp: 10924
Ion Ratio Lower Upper
63 100
106 30.4 23.0 34.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#59
2-Hexanone
Concen: 4.466 ug/l m
RT: 11.235 min Scan# 1578
Delta R.T. 0.035 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44


Tgt Ion: 43 Resp: 12295
Ion Ratio Lower Upper
43 100
58 4.7 30.4 91.3#





#60
Dibromochloromethane
Concen: 1.021 ug/l
RT: 11.359 min Scan# 1599
Delta R.T. -0.000 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001




Tgt Ion:129 Resp: 284
Ion Ratio Lower Upper
129 100
127 73.8 38.6 115.8

Manual Integrations
APPROVED



Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025




#61
1,2-Dibromoethane
Concen: 1.014 ug/l
RT: 11.470 min Scan# 1618
Delta R.T. -0.000 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44



Tgt Ion:107 Resp: 2777
Ion Ratio Lower Upper
107 100
109 91.2 76.3 114.5





#62
4-Bromofluorobenzene
Concen: 0.000 ug/l
Expected RT: 12.85 min



Lab File: VN086862.D
Acq: 06 Jun 2025 12:44

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001

Tgt Ion: 95
Sig Exp Ratio
95 100
174 70.9
176 66.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 1685
Delta R.T. -0.000 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44


Tgt Ion:117 Resp: 348010
Ion Ratio Lower Upper
117 100
82 57.5 44.6 67.0
119 32.5 25.5 38.3





#64
Tetrachloroethene
Concen: 1.123 ug/l
RT: 11.106 min Scan# 1556
Delta R.T. -0.000 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001



Tgt Ion:164 Resp: 247
Ion Ratio Lower Upper
164 100
166 121.4 101.8 152.8
129 104.4 83.0 124.4
131 84.3 78.2 117.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#65
Chlorobenzene
Concen: 1.118 ug/l
RT: 11.894 min Scan# 1690
Delta R.T. -0.000 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44


Tgt Ion:112 Resp: 8581
Ion Ratio Lower Upper
112 100
114 25.7 26.2 39.2#





#66
1,1,1,2-Tetrachloroethane
Concen: 1.023 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. -0.006 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001



Tgt Ion:131 Resp: 252
Ion Ratio Lower Upper
131 100
133 100.9 48.3 144.8
119 0.0 34.0 101.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#67
Ethyl Benzene
Concen: 1.047 ug/l
RT: 11.970 min Scan# 1703
Delta R.T. 0.006 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44


Tgt Ion: 91 Resp: 13845
Ion Ratio Lower Upper
91 100
106 28.7 25.4 38.2





#68
m/p-Xylenes
Concen: 2.008 ug/l
RT: 12.070 min Scan# 1720
Delta R.T. -0.000 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001




Tgt Ion:106 Resp: 1016
Ion Ratio Lower Upper
106 100
91 207.3 159.4 239.0

Manual Integrations
APPROVED



Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025




#69
o-Xylene
Concen: 1.025 ug/l
RT: 12.400 min Scan# 1776
Delta R.T. -0.000 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44



Tgt Ion:106 Resp: 4970
Ion Ratio Lower Upper
106 100
91 223.6 106.1 318.1









#70
Styrene
Concen: 0.987 ug/l
RT: 12.417 min Scan# 1779
Delta R.T. 0.006 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001

Tgt Ion:104 Resp: 818
Ion Ratio Lower Upper
104 100
78 49.6 38.5 57.7
103 52.5 43.1 64.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#71
Bromoform
Concen: 0.825 ug/l
RT: 12.582 min Scan# 1807
Delta R.T. -0.000 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44


Tgt Ion:173 Resp: 1508
Ion Ratio Lower Upper
173 100
175 63.1 24.1 72.4
254 0.0 0.0 0.0





#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.794 min Scan# 2013
Delta R.T. -0.000 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001



Tgt Ion:152 Resp: 167399
Ion Ratio Lower Upper
152 100
115 61.1 30.1 90.5
150 158.0 0.0 345.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#73
Isopropylbenzene
Concen: 1.061 ug/l
RT: 12.694 min Scan# 1826
Delta R.T. -0.000 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44


Tgt Ion:105 Resp: 12938
Ion Ratio Lower Upper
105 100
120 28.1 13.5 40.4





#74
N-amy1 acetate
Concen: 1.138 ug/l m
RT: 12.647 min Scan# 1818
Delta R.T. 0.135 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001



Tgt Ion: 43 Resp: 4608
Ion Ratio Lower Upper
43 100
70 0.0 37.6 56.4
55 0.0 19.0 28.6
61 0.0 20.8 31.2#




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#75
1,1,2,2-Tetrachloroethane
Concen: 1.047 ug/l
RT: 12.935 min Scan# 1867
Delta R.T. -0.006 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44

Tgt Ion: 83 Resp: 4326
Ion Ratio Lower Upper
83 100
131 4.1 4.9 14.7#
85 68.6 32.0 96.0









#76
1,2,3-Trichloropropane
Concen: 1.088 ug/l m
RT: 13.000 min Scan# 1878
Delta R.T. 0.006 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44

Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001

Tgt Ion: 75 Resp: 434
Ion Ratio Lower Upper
75 100
77 154.4 79.3 237.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#77
Bromobenzene
Concen: 1.042 ug/l
RT: 12.982 min Scan# 1875
Delta R.T. -0.000 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44

Tgt Ion:156 Resp: 2913
Ion Ratio Lower Upper
156 100
77 230.2 117.7 353.1
158 102.5 48.1 144.4









#78
n-propylbenzene
Concen: 1.023 ug/l
RT: 13.035 min Scan# 1884
Delta R.T. -0.000 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001

Tgt Ion: 91 Resp: 1516
Ion Ratio Lower Upper
91 100
120 21.9 11.4 34.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#79
2-Chlorotoluene
Concen: 1.074 ug/l
RT: 13.123 min Scan# 1899
Delta R.T. -0.000 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44

Tgt Ion: 91 Resp: 9550
Ion Ratio Lower Upper
91 100
126 30.4 16.2 48.5









#80
1,3,5-Trimethylbenzene
Concen: 0.999 ug/l
RT: 13.170 min Scan# 1907
Delta R.T. -0.000 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001

Tgt Ion:105 Resp: 10058
Ion Ratio Lower Upper
105 100
120 52.7 24.9 74.6




Manual Integrations
APPROVED


Reviewed By :John Carbone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#82
4-Chlorotoluene
Concen: 1.079 ug/l
RT: 13.223 min Scan# 1916
Delta R.T. -0.000 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44


Tgt Ion: 91 Resp: 9706
Ion Ratio Lower Upper
91 100
126 31.1 16.7 50.1





#83
tert-Butylbenzene
Concen: 1.069 ug/l
RT: 13.435 min Scan# 1952
Delta R.T. -0.000 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001



Tgt Ion:119 Resp: 9849
Ion Ratio Lower Upper
119 100
91 65.3 32.1 96.5
134 23.9 13.4 40.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#84
1,2,4-Trimethylbenzene
Concen: 0.979 ug/l
RT: 13.482 min Scan# 1960
Delta R.T. -0.000 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44


Tgt Ion:105 Resp: 9882
Ion Ratio Lower Upper
105 100
120 47.2 23.2 69.6





#85
sec-Butylbenzene
Concen: 1.029 ug/l
RT: 13.617 min Scan# 1983
Delta R.T. -0.000 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001




Tgt Ion:105 Resp: 13775
Ion Ratio Lower Upper
105 100
134 20.2 10.0 30.0

Manual Integrations
APPROVED



Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025




#86
p-Isopropyltoluene
Concen: 1.000 ug/l
RT: 13.729 min Scan# 2002
Delta R.T. -0.000 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44



Tgt Ion:119 Resp: 11065
Ion Ratio Lower Upper
119 100
134 25.4 13.5 40.4
91 24.1 11.5 34.4





#87

1,3-Dichlorobenzene

Concen: 1.073 ug/l

RT: 13.735 min Scan# 2003

Delta R.T. -0.000 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC001





Tgt Ion:	146	Resp:	590
Ion Ratio	Lower	Upper	
146	100		
111	45.8	20.9	62.8
148	59.7	32.0	96.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#88

1,4-Dichlorobenzene

Concen: 1.086 ug/l m




RT: 13.811 min Scan# 2016


Delta R.T. -0.000 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44

Tgt Ion:	146	Resp:	6094
Ion Ratio	Lower	Upper	
146	100		
111	44.4	19.9	59.7
148	57.8	31.8	95.4





#89

n-Butylbenzene

Concen: 1.085 ug/l

RT: 14.058 min Scan# 2057

Delta R.T. 0.006 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC001

Tgt Ion:	91	Resp:	1162
Ion	Ratio	Lower	Upper
91	100		
92	50.8	27.5	82.4
134	22.6	13.0	39.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#90

Hexachloroethane

Concen: 1.036 ug/l




RT: 14.329 min Scan# 2104


Delta R.T. -0.006 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44

Tgt Ion:	117	Resp:	1942
Ion	Ratio	Lower	Upper
117	100		
201	60.6	33.2	99.6





#91

1,2-Dichlorobenzene

Concen: 1.061 ug/l

RT: 14.106 min Scan# 2066

Delta R.T. -0.000 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC001





Tgt Ion:	146	Resp:	560
Ion	Ratio	Lower	Upper
146	100		
111	43.1	21.6	64.6
148	64.7	31.7	95.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#92

1,2-Dibromo-3-Chloropropane

Concen: 1.150 ug/l




RT: 14.717 min Scan# 2170


Delta R.T. -0.000 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44

Tgt Ion:	75	Resp:	1136
Ion	Ratio	Lower	Upper
75	100		
155	63.6	41.1	123.4
157	81.0	52.7	158.1





#93

1,2,4-Trichlorobenzene

Concen: 1.034 ug/l

RT: 15.388 min Scan# 21

Delta R.T. -0.000 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001





Tgt Ion:	180	Resp:	348
Ion Ratio	Lower	Upper	
180	100		
182	92.1	46.9	140.7
145	30.4	16.0	47.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#94

Hexachlorobutadiene

Concen: 0.969 ug/l




RT: 15.500 min Scan# 2303


Delta R.T. -0.000 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44


Tgt Ion:	225	Resp:	1219
Ion Ratio	Lower	Upper	
225	100		
223	53.6	33.5	100.4
227	57.5	31.1	93.5





#95
Naphthalene
Concen: 1.018 ug/l
RT: 15.635 min Scan# 2327
Delta R.T. -0.006 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001




Tgt Ion:128 Resp: 12789
Ion Ratio Lower Upper
128 100
127 12.4 10.2 15.2
129 9.5 8.8 13.2

Manual Integrations
APPROVED



Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#96
1,2,3-Trichlorobenzene
Concen: 1.072 ug/l
RT: 15.835 min Scan# 2360
Delta R.T. -0.000 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44



Tgt Ion:180 Resp: 3594
Ion Ratio Lower Upper
180 100
182 91.6 47.0 141.0
145 29.0 16.6 49.8



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060625\
 Data File : VN086863.D
 Acq On : 06 Jun 2025 13:17
 Operator : JC\MD
 Sample : VSTDICC005
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC005

Manual Integrations
 APPROVED

Reviewed By : John Carlone 06/09/2025
 Supervised By : Mahesh Dadoda 06/09/2025

Quant Time: Jun 07 01:55:22 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 01:51:02 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	229701	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	423980	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	371851	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	181065	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	16818	5.468	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	10.940%#		
35) Dibromofluoromethane	8.177	113	12867	5.121	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	10.240%#		
50) Toluene-d8	10.571	98	52794	5.308	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	10.620%#		
62) 4-Bromofluorobenzene	12.847	95	18697	5.059	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	10.120%#		
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	2.154	85	10196	4.452	ug/l	93
3) Chloromethane	2.395	50	15025	5.080	ug/l	98
4) Vinyl Chloride	2.554	62	15384	5.043	ug/l	97
5) Bromomethane	2.989	94	8606	5.041	ug/l	97
6) Chloroethane	3.154	64	10190	5.171	ug/l	96
7) Trichlorofluoromethane	3.518	101	20740	5.203	ug/l	99
8) Diethyl Ether	3.971	74	9121	5.252	ug/l	95
9) 1,1,2-Trichlorotrifluo...	4.401	101	13034	5.205	ug/l	97
10) Methyl Iodide	4.606	142	16530	5.092	ug/l	98
11) Tert butyl alcohol	5.542	59	22000	26.363	ug/l	99
12) 1,1-Dichloroethene	4.359	96	13610	5.323	ug/l	94
13) Acrolein	4.195	56	8401	31.804	ug/l	99
14) Allyl chloride	5.042	41	20925	4.935	ug/l	99
15) Acrylonitrile	5.736	53	49826	25.545	ug/l	99
16) Acetone	4.448	43	41981	25.741	ug/l	99
17) Carbon Disulfide	4.736	76	37246	5.267	ug/l	97
18) Methyl Acetate	5.048	43	24096	5.070	ug/l	97
19) Methyl tert-butyl Ether	5.812	73	46823	5.058	ug/l	95
20) Methylene Chloride	5.300	84	15809	5.178	ug/l	96
21) trans-1,2-Dichloroethene	5.806	96	15480	5.442	ug/l #	81
22) Diisopropyl ether	6.689	45	46396	5.191	ug/l	99
23) Vinyl Acetate	6.618	43	194334	25.735	ug/l	100
24) 1,1-Dichloroethane	6.583	63	26473	5.147	ug/l	99
25) 2-Butanone	7.494	43	68629	25.888	ug/l	99
26) 2,2-Dichloropropane	7.500	77	18288	4.571	ug/l	100
27) cis-1,2-Dichloroethene	7.500	96	17594	5.172	ug/l	98
28) Bromochloromethane	7.830	49	12958	5.124	ug/l #	97
29) Tetrahydrofuran	7.853	42	44749	25.912	ug/l	98
30) Chloroform	7.977	83	26450	5.149	ug/l	95
31) Cyclohexane	8.271	56	29938	6.002	ug/l	95
32) 1,1,1-Trichloroethane	8.177	97	22857	5.232	ug/l #	50
36) 1,1-Dichloropropene	8.377	75	19410	5.184	ug/l	98
37) Ethyl Acetate	7.571	43	26073	5.470	ug/l	97
38) Carbon Tetrachloride	8.371	117	19056	5.184	ug/l	99
39) Methylcyclohexane	9.606	83	27353	5.333	ug/l	95
40) Benzene	8.612	78	63646	5.199	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060625\
 Data File : VN086863.D
 Acq On : 06 Jun 2025 13:17
 Operator : JC\MD
 Sample : VSTDICC005
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC005

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025

Quant Time: Jun 07 01:55:22 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M

Quant Title : SW846 8260

QLast Update : Sat Jun 07 01:51:02 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.783	41	13518	5.050	ug/l	99
42) 1,2-Dichloroethane	8.683	62	19329	5.205	ug/l	98
43) Isopropyl Acetate	8.700	43	40268	5.264	ug/l	98
44) Trichloroethene	9.359	130	15245	5.251	ug/l	96
45) 1,2-Dichloropropane	9.624	63	15632	5.248	ug/l	99
46) Dibromomethane	9.718	93	10241	5.176	ug/l	99
47) Bromodichloromethane	9.888	83	20512	5.039	ug/l	97
48) Methyl methacrylate	9.688	41	17591	4.996	ug/l	94
49) 1,4-Dioxane	9.700	88	6141	95.874	ug/l #	93
51) 4-Methyl-2-Pentanone	10.447	43	116434	25.283	ug/l	97
52) Toluene	10.630	92	38744	5.178	ug/l	99
53) t-1,3-Dichloropropene	10.841	75	22285	4.896	ug/l	98
54) cis-1,3-Dichloropropene	10.318	75	24477	5.026	ug/l	93
55) 1,1,2-Trichloroethane	11.018	97	15068	5.234	ug/l	97
56) Ethyl methacrylate	10.888	69	21893	4.774	ug/l	95
57) 1,3-Dichloropropane	11.165	76	25452	5.096	ug/l	98
58) 2-Chloroethyl Vinyl ether	10.165	63	64722	23.664	ug/l	98
59) 2-Hexanone	11.212	43	59884	20.188	ug/l	99
60) Dibromochloromethane	11.359	129	14884	4.962	ug/l	98
61) 1,2-Dibromoethane	11.471	107	15172	5.141	ug/l	99
64) Tetrachloroethene	11.106	164	12322	5.236	ug/l	92
65) Chlorobenzene	11.894	112	42195	5.145	ug/l	98
66) 1,1,1,2-Tetrachloroethane	11.965	131	13897	5.271	ug/l	97
67) Ethyl Benzene	11.965	91	73443	5.199	ug/l	97
68) m/p-Xylenes	12.071	106	55859	10.330	ug/l	100
69) o-Xylene	12.400	106	26004	5.021	ug/l	96
70) Styrene	12.412	104	44380	5.008	ug/l	98
71) Bromoform	12.582	173	9890	5.063	ug/l #	96
73) Isopropylbenzene	12.694	105	67878	5.146	ug/l	100
74) N-amyl acetate	12.559	43	20096m	4.590	ug/l	
75) 1,1,2,2-Tetrachloroethane	12.935	83	23516	5.262	ug/l	98
76) 1,2,3-Trichloropropane	12.994	75	21527m	4.986	ug/l	
77) Bromobenzene	12.976	156	15485	5.119	ug/l	92
78) n-propylbenzene	13.035	91	83781	5.226	ug/l	99
79) 2-Chlorotoluene	13.124	91	49297	5.128	ug/l	100
80) 1,3,5-Trimethylbenzene	13.171	105	55961	5.139	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.741	75	9234	4.939	ug/l	94
82) 4-Chlorotoluene	13.223	91	49924	5.132	ug/l	100
83) tert-Butylbenzene	13.435	119	50874	5.103	ug/l	98
84) 1,2,4-Trimethylbenzene	13.482	105	56525	5.176	ug/l	100
85) sec-Butylbenzene	13.618	105	74711	5.161	ug/l	99
86) p-Isopropyltoluene	13.729	119	62016	5.182	ug/l	99
87) 1,3-Dichlorobenzene	13.735	146	30949	5.200	ug/l	98
88) 1,4-Dichlorobenzene	13.812	146	32342	5.330	ug/l	96
89) n-Butylbenzene	14.053	91	59601	5.142	ug/l	99
90) Hexachloroethane	14.335	117	9725	4.794	ug/l	95
91) 1,2-Dichlorobenzene	14.106	146	29897	5.228	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.717	75	5731	5.363	ug/l	93
93) 1,2,4-Trichlorobenzene	15.394	180	18778	5.143	ug/l	99
94) Hexachlorobutadiene	15.500	225	7072	5.199	ug/l	99
95) Naphthalene	15.641	128	67481	4.965	ug/l	99
96) 1,2,3-Trichlorobenzene	15.841	180	18262	5.034	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060625\
Data File : VN086863.D
Acq On : 06 Jun 2025 13:17
Operator : JC\MD
Sample : VSTDICC005
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED

Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025

Quant Time: Jun 07 01:55:22 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 01:51:02 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.230 min Scan# 1067
Delta R.T. -0.006 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005



Tgt Ion:168 Resp: 22970
Ion Ratio Lower Upper
168 100
99 59.2 49.1 73.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#2
Dichlorodifluoromethane
Concen: 4.452 ug/l
RT: 2.154 min Scan# 34
Delta R.T. -0.000 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17


Tgt Ion: 85 Resp: 10196
Ion Ratio Lower Upper
85 100
87 28.6 16.2 48.6





#3
Chloromethane
Concen: 5.080 ug/l
RT: 2.395 min Scan# 75
Delta R.T. 0.000 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005



Tgt Ion: 50 Resp: 1502
Ion Ratio Lower Upper
50 100
52 30.9 25.7 38.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#4
Vinyl Chloride
Concen: 5.043 ug/l
RT: 2.554 min Scan# 102
Delta R.T. 0.000 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17


Tgt Ion: 62 Resp: 15384
Ion Ratio Lower Upper
62 100
64 34.3 26.0 39.0





#5
Bromomethane
Concen: 5.041 ug/l
RT: 2.989 min Scan# 11
Delta R.T. -0.006 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005



Tgt Ion: 94 Resp: 860
Ion Ratio Lower Upper
94 100
96 100.6 77.9 116.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#6
Chloroethane
Concen: 5.171 ug/l
RT: 3.154 min Scan# 204
Delta R.T. -0.000 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17


Tgt Ion: 64 Resp: 10190
Ion Ratio Lower Upper
64 100
66 28.8 24.6 36.8





#7
Trichlorofluoromethane
Concen: 5.203 ug/l
RT: 3.518 min Scan# 20740
Delta R.T. -0.006 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005



Tgt Ion:101 Resp: 20740
Ion Ratio Lower Upper
101 100
103 64.8 52.2 78.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#8
Diethyl Ether
Concen: 5.252 ug/l
RT: 3.971 min Scan# 343
Delta R.T. -0.012 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17

Tgt Ion: 74 Resp: 9121
Ion Ratio Lower Upper
74 100
45 95.8 45.5 136.5





#9

1,1,2-Trichlorotrifluoroethane

Concen: 5.205 ug/l

RT: 4.401 min Scan# 415

Delta R.T. 0.006 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005





Tgt Ion:	101	Resp:	13034
Ion	Ratio	Lower	Upper
101	100		
85	43.3	33.8	50.8
151	70.8	58.7	88.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#10

Methyl Iodide

Concen: 5.092 ug/l




RT: 4.606 min Scan# 451


Delta R.T. -0.000 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17

Tgt Ion:	142	Resp:	16530
Ion	Ratio	Lower	Upper
142	100		
127	39.8	30.9	46.3
141	14.5	10.4	15.6





#11

Tert butyl alcohol

Concen: 26.363 ug/l

RT: 5.542 min Scan# 611

Delta R.T. -0.006 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC005

Tgt Ion: 59 Resp: 22000

Ion Ratio Lower Upper

59 100





57 9.6 8.1 12.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#12

1,1-Dichloroethene

Concen: 5.323 ug/l

RT: 4.359 min Scan# 409

Delta R.T. -0.000 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17




Tgt Ion: 96 Resp: 13610


Ion Ratio Lower Upper

96 100

61 148.1 125.8 188.6

98 67.5 52.5 78.7





#13

Acrolein

Concen: 31.804 ug/l

RT: 4.195 min Scan# 381

Delta R.T. -0.006 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC005

Tgt Ion: 56 Resp: 840

Ion Ratio Lower Upper

56 100





55 70.2 55.6 83.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#14

Allyl chloride

Concen: 4.935 ug/l

RT: 5.042 min Scan# 525

Delta R.T. -0.000 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17




Tgt Ion: 41 Resp: 20925


Ion Ratio Lower Upper

41 100

39 63.2 51.3 76.9


76 39.0 31.4 47.0





#15
Acrylonitrile
Concen: 25.545 ug/l
RT: 5.736 min Scan# 644
Delta R.T. -0.006 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005



Tgt Ion: 53 Resp: 49820
Ion Ratio Lower Upper
53 100
52 82.4 65.0 97.6
51 32.9 27.2 40.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#16
Acetone
Concen: 25.741 ug/l
RT: 4.448 min Scan# 424
Delta R.T. -0.000 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17

Tgt Ion: 43 Resp: 41981
Ion Ratio Lower Upper
43 100
58 34.1 28.0 42.0





#17

Carbon Disulfide

Concen: 5.267 ug/l

RT: 4.736 min Scan# 472

Delta R.T. 0.006 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC005

Tgt Ion: 76 Resp: 3724

Ion Ratio Lower Upper

76 100





78 8.0 7.3 10.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#18

Methyl Acetate

Concen: 5.070 ug/l

RT: 5.048 min Scan# 526

Delta R.T. 0.006 min

Lab File: VN086863.D




Acq: 06 Jun 2025 13:17


Tgt Ion: 43 Resp: 24096

Ion Ratio Lower Upper

43 100

74 24.4 20.7 31.1





#19

Methyl tert-butyl Ether

Concen: 5.058 ug/l

RT: 5.812 min Scan# 61

Delta R.T. -0.000 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005

Tgt Ion: 73 Resp: 4682

Ion Ratio Lower Upper

73 100





57 24.3 17.4 26.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#20

Methylene Chloride

Concen: 5.178 ug/l

RT: 5.300 min Scan# 569

Delta R.T. 0.006 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17

Tgt Ion: 84 Resp: 15809




Ion Ratio Lower Upper


84 100

49 106.8 90.5 135.7

51 36.6 28.5 42.7


86 60.4 48.1 72.1





#21
trans-1,2-Dichloroethene
Concen: 5.442 ug/l
RT: 5.806 min Scan# 61
Delta R.T. -0.000 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005



Tgt Ion: 96 Resp: 15480
Ion Ratio Lower Upper
96 100
61 105.6 109.2 163.8
98 58.4 49.8 74.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#22
Diisopropyl ether
Concen: 5.191 ug/l
RT: 6.689 min Scan# 805
Delta R.T. -0.000 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17

Tgt Ion: 45 Resp: 46396
Ion Ratio Lower Upper
45 100
43 53.2 41.9 62.9
87 27.9 22.4 33.6
59 12.6 9.8 14.6





#23
Vinyl Acetate
Concen: 25.735 ug/l
RT: 6.618 min Scan# 793
Delta R.T. -0.000 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC005

Tgt Ion: 43 Resp: 194334
Ion Ratio Lower Upper
43 100
86 11.8 9.5 14.3

Manual Integrations




APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#24
1,1-Dichloroethane
Concen: 5.147 ug/l
RT: 6.583 min Scan# 787
Delta R.T. 0.000 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17

Tgt Ion: 63 Resp: 26473
Ion Ratio Lower Upper
63 100
98 6.4 3.6 10.8
100 5.1 2.6 7.8









#25
2-Butanone
Concen: 25.888 ug/l
RT: 7.494 min Scan# 942
Delta R.T. 0.000 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17

Instrument :
MSVOA_N
Client Sample Id :
VSTDIC005

Tgt Ion: 43 Resp: 68629
Ion Ratio Lower Upper
43 100
72 28.1 23.0 34.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#26
2,2-Dichloropropane
Concen: 4.571 ug/l
RT: 7.500 min Scan# 943
Delta R.T. 0.000 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17

Tgt Ion: 77 Resp: 18288
Ion Ratio Lower Upper
77 100
97 23.3 11.6 34.7





#27

cis-1,2-Dichloroethene

Concen: 5.172 ug/l

RT: 7.500 min Scan# 943

Delta R.T. 0.000 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005





Tgt Ion:	96	Resp:	17594
Ion	Ratio	Lower	Upper
96	100		
61	136.0	0.0	278.0
98	64.0	0.0	128.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#28

Bromochloromethane

Concen: 5.124 ug/l




RT: 7.830 min Scan# 999


Delta R.T. 0.006 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17

Tgt Ion:	49	Resp:	12958
Ion	Ratio	Lower	Upper
49	100		
129	0.0	0.0	3.8
130	75.2	62.2	93.2





#29

Tetrahydrofuran

Concen: 25.912 ug/l

RT: 7.853 min Scan# 1003

Delta R.T. 0.000 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005





Tgt Ion:	42	Resp:	44749
Ion	Ratio	Lower	Upper
42	100		
72	48.6	37.8	56.6
71	43.5	35.6	53.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#30

Chloroform

Concen: 5.149 ug/l




RT: 7.977 min Scan# 1024


Delta R.T. -0.000 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17

Tgt Ion:	83	Resp:	26450
Ion	Ratio	Lower	Upper
83	100		
85	69.9	52.9	79.3





#31

Cyclohexane

Concen: 6.002 ug/l

RT: 8.271 min Scan# 1074

Delta R.T. -0.000 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005

Tgt Ion: 56 Resp: 29938

Ion Ratio Lower Upper

56 100

69 33.2 25.4 38.0





84 81.1 68.8 103.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#32

1,1,1-Trichloroethane

Concen: 5.232 ug/l

RT: 8.177 min Scan# 1058

Delta R.T. -0.000 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17




Tgt Ion: 97 Resp: 22857


Ion Ratio Lower Upper

97 100

99 0.0 52.0 78.0

61 42.9 35.9 53.9





#33

1,2-Dichloroethane-d4

Concen: 5.468 ug/l

RT: 8.588 min Scan# 1128

Delta R.T. 0.000 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005

Tgt Ion: 65 Resp: 16818

Ion Ratio Lower Upper

65 100





67 54.4 0.0 105.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.106 min Scan# 1216

Delta R.T. -0.000 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17




Tgt Ion:114 Resp: 423980


Ion Ratio Lower Upper

114 100

63 20.8 0.0 39.6


88 15.9 0.0 30.2





#35
Dibromofluoromethane
Concen: 5.121 ug/l
RT: 8.177 min Scan# 1057
Delta R.T. 0.006 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005



Tgt Ion: 113 Resp: 1286
Ion Ratio Lower Upper
113 100
111 106.5 84.2 126.2
192 18.0 14.2 21.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#36
1,1-Dichloropropene
Concen: 5.184 ug/l
RT: 8.377 min Scan# 1092
Delta R.T. 0.000 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17

Tgt Ion: 75 Resp: 19410
Ion Ratio Lower Upper
75 100
110 37.4 17.6 52.9
77 31.7 25.3 37.9





#37

Ethyl Acetate

Concen: 5.470 ug/l

RT: 7.571 min Scan# 91

Delta R.T. -0.006 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005





Tgt Ion:	43	Resp:	2607
Ion	Ratio	Lower	Upper
43	100		
61	14.1	12.1	18.1
70	10.9	9.8	14.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#38

Carbon Tetrachloride

Concen: 5.184 ug/l




RT: 8.371 min Scan# 1091


Delta R.T. 0.000 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17

Tgt Ion:	117	Resp:	19056
Ion	Ratio	Lower	Upper
117	100		
119	93.8	75.0	112.6
121	28.5	24.9	37.3





#39

Methylcyclohexane

Concen: 5.333 ug/l

RT: 9.606 min Scan# 1130

Delta R.T. -0.000 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005





Tgt Ion:	83	Resp:	2735
Ion	Ratio	Lower	Upper
83	100		
55	73.6	61.0	91.6
98	42.6	38.9	58.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#40

Benzene

Concen: 5.199 ug/l




RT: 8.612 min Scan# 1132


Delta R.T. -0.000 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17


Tgt Ion:	78	Resp:	63646
Ion	Ratio	Lower	Upper
78	100		
77	22.6	18.7	28.1





#41
Methacrylonitrile
Concen: 5.050 ug/l
RT: 7.783 min Scan# 991
Delta R.T. -0.006 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005



Tgt Ion: 41 Resp: 13518
Ion Ratio Lower Upper
41 100
39 53.1 41.2 61.8
67 77.0 61.4 92.2
52 29.8 22.7 34.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#42
1,2-Dichloroethane
Concen: 5.205 ug/l
RT: 8.683 min Scan# 1144
Delta R.T. 0.006 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17

Tgt Ion: 62 Resp: 19329
Ion Ratio Lower Upper
62 100
98 9.9 0.0 21.2





#43

Isopropyl Acetate

Concen: 5.264 ug/l

RT: 8.700 min Scan# 1146

Delta R.T. 0.006 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005





Tgt Ion: 43	Resp: 4026
Ion Ratio	Lower Upper
43 100	
61 26.2	22.1 33.1
87 14.1	11.8 17.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#44

Trichloroethene

Concen: 5.251 ug/l




RT: 9.359 min Scan# 1259


Delta R.T. 0.000 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17

Tgt Ion:130	Resp: 15245
Ion Ratio	Lower Upper
130 100	
95 101.6	0.0 196.2





#45

1,2-Dichloropropane

Concen: 5.248 ug/l

RT: 9.624 min Scan# 1304

Delta R.T. -0.006 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005

Tgt Ion: 63 Resp: 1563

Ion Ratio Lower Upper

63 100





65 32.7 25.5 38.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#46

Dibromomethane

Concen: 5.176 ug/l

RT: 9.718 min Scan# 1320

Delta R.T. -0.000 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17




Tgt Ion: 93 Resp: 10241


Ion Ratio Lower Upper

93 100

95 84.9 67.1 100.7

174 93.4 74.2 111.4









#47
Bromodichloromethane
Concen: 5.039 ug/l
RT: 9.888 min Scan# 1349
Delta R.T. -0.006 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Tgt Ion: 83 Resp: 2051
Ion Ratio Lower Upper
83 100
85 61.3 50.9 76.3
127 7.0 6.6 10.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#48
Methyl methacrylate
Concen: 4.996 ug/l
RT: 9.688 min Scan# 1315
Delta R.T. 0.006 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17

Tgt Ion: 41 Resp: 17591
Ion Ratio Lower Upper
41 100
69 85.2 74.6 112.0
39 52.3 40.2 60.4





#49

1,4-Dioxane

Concen: 95.874 ug/l

RT: 9.700 min Scan# 1317

Delta R.T. -0.000 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005

Tgt Ion: 88 Resp: 614

Ion Ratio Lower Upper

88 100

43 0.0 0.0 0.0





58 77.9 57.7 86.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#50

Toluene-d8

Concen: 5.308 ug/l

RT: 10.571 min Scan# 1465

Delta R.T. -0.000 min

Lab File: VN086863.D




Acq: 06 Jun 2025 13:17


Tgt Ion: 98 Resp: 52794

Ion Ratio Lower Upper

98 100


100 68.1 53.4 80.0





#51
4-Methyl-2-Pentanone
Concen: 25.283 ug/l
RT: 10.447 min Scan# 1444
Delta R.T. 0.000 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005



Tgt Ion: 43 Resp: 116434
Ion Ratio Lower Upper
43 100
58 40.9 34.2 51.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#52
Toluene
Concen: 5.178 ug/l
RT: 10.630 min Scan# 1475
Delta R.T. -0.006 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17


Tgt Ion: 92 Resp: 38744
Ion Ratio Lower Upper
92 100
91 169.5 136.5 204.7





#53
t-1,3-Dichloropropene
Concen: 4.896 ug/l
RT: 10.841 min Scan# 1511
Delta R.T. -0.000 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005



Tgt Ion: 75 Resp: 2228
Ion Ratio Lower Upper
75 100
77 31.9 26.4 39.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#54
cis-1,3-Dichloropropene
Concen: 5.026 ug/l
RT: 10.318 min Scan# 1422
Delta R.T. 0.000 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17


Tgt Ion: 75 Resp: 24477
Ion Ratio Lower Upper
75 100
77 34.5 26.2 39.4
39 46.5 32.3 48.5





#55
1,1,2-Trichloroethane
Concen: 5.234 ug/l
RT: 11.018 min Scan# 1541
Delta R.T. 0.000 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005



Tgt Ion: 97 Resp: 1506
Ion Ratio Lower Upper
97 100
83 82.5 69.7 104.5
85 53.5 45.1 67.7
99 62.6 50.6 75.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#56
Ethyl methacrylate
Concen: 4.774 ug/l
RT: 10.888 min Scan# 1519
Delta R.T. 0.006 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17

Tgt Ion: 69 Resp: 21893
Ion Ratio Lower Upper
69 100
41 63.5 47.2 70.8
39 29.9 23.0 34.6





#57

1,3-Dichloropropane

Concen: 5.096 ug/l

RT: 11.165 min Scan# 1566

Delta R.T. 0.000 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005

Tgt Ion: 76 Resp: 25452

Ion Ratio Lower Upper

76 100





78 32.8 25.3 37.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#58

2-Chloroethyl Vinyl ether

Concen: 23.664 ug/l

RT: 10.165 min Scan# 1396

Delta R.T. 0.000 min

Lab File: VN086863.D




Acq: 06 Jun 2025 13:17


Tgt Ion: 63 Resp: 64722

Ion Ratio Lower Upper

63 100


106 27.9 23.0 34.6





#59
2-Hexanone
Concen: 20.188 ug/l
RT: 11.212 min Scan# 1574
Delta R.T. 0.012 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005



Tgt Ion: 43 Resp: 59884
Ion Ratio Lower Upper
43 100
58 60.5 30.4 91.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#60
Dibromochloromethane
Concen: 4.962 ug/l
RT: 11.359 min Scan# 1599
Delta R.T. -0.000 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17


Tgt Ion:129 Resp: 14884
Ion Ratio Lower Upper
129 100
127 75.9 38.6 115.8





#61
1,2-Dibromoethane
Concen: 5.141 ug/l
RT: 11.471 min Scan# 1618
Delta R.T. 0.000 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005



Tgt Ion:107 Resp: 15172
Ion Ratio Lower Upper
107 100
109 94.2 76.3 114.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#62
4-Bromofluorobenzene
Concen: 5.059 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. -0.000 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17

Tgt Ion: 95 Resp: 18697
Ion Ratio Lower Upper
95 100
174 75.5 0.0 141.8
176 70.2 0.0 132.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 1685
Delta R.T. -0.000 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Tgt Ion:117 Resp: 37185
Ion Ratio Lower Upper
117 100
82 57.3 44.6 67.0
119 32.3 25.5 38.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#64
Tetrachloroethene
Concen: 5.236 ug/l
RT: 11.106 min Scan# 1556
Delta R.T. 0.000 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17


Tgt Ion:164 Resp: 12322
Ion Ratio Lower Upper
164 100
166 117.9 101.8 152.8
129 96.6 83.0 124.4
131 88.0 78.2 117.2





#65
Chlorobenzene
Concen: 5.145 ug/l
RT: 11.894 min Scan# 1690
Delta R.T. 0.000 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005



Tgt Ion:112 Resp: 42198
Ion Ratio Lower Upper
112 100
114 31.5 26.2 39.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#66
1,1,1,2-Tetrachloroethane
Concen: 5.271 ug/l
RT: 11.965 min Scan# 1702
Delta R.T. 0.000 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17

Tgt Ion:131 Resp: 13897
Ion Ratio Lower Upper
131 100
133 91.7 48.3 144.8
119 69.1 34.0 101.8





#67

Ethyl Benzene

Concen: 5.199 ug/l

RT: 11.965 min Scan# 1702

Delta R.T. 0.000 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005

Tgt Ion: 91 Resp: 7344

Ion Ratio Lower Upper

91 100





106 30.3 25.4 38.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#68

m/p-Xylenes

Concen: 10.330 ug/l

RT: 12.071 min Scan# 1720

Delta R.T. -0.000 min

Lab File: VN086863.D




Acq: 06 Jun 2025 13:17


Tgt Ion:106 Resp: 55859

Ion Ratio Lower Upper

106 100

91 199.0 159.4 239.0





#69

o-Xylene

Concen: 5.021 ug/l

RT: 12.400 min Scan# 1776

Delta R.T. 0.000 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005

Tgt Ion:106 Resp: 26004

Ion Ratio Lower Upper

106 100





91 218.1 106.1 318.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#70

Styrene

Concen: 5.008 ug/l

RT: 12.412 min Scan# 1778

Delta R.T. -0.000 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17




Tgt Ion:104 Resp: 44380


Ion Ratio Lower Upper

104 100

78 50.4 38.5 57.7


103 54.4 43.1 64.7





#71
Bromoform
Concen: 5.063 ug/l
RT: 12.582 min Scan# 1807
Delta R.T. -0.000 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005




Tgt Ion:173 Resp: 9890
Ion Ratio Lower Upper
173 100
175 45.8 24.1 72.4
254 0.0 0.0 0.0

Manual Integrations
APPROVED



Reviewed By :John Carbone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.794 min Scan# 2013
Delta R.T. -0.000 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17



Tgt Ion:152 Resp: 181065
Ion Ratio Lower Upper
152 100
115 61.2 30.1 90.5
150 157.8 0.0 345.0









#73
Isopropylbenzene
Concen: 5.146 ug/l
RT: 12.694 min Scan# 1803
Delta R.T. 0.000 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Tgt Ion:105 Resp: 67878
Ion Ratio Lower Upper
105 100
120 26.8 13.5 40.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#74
N-ethyl acetate
Concen: 4.590 ug/l m
RT: 12.559 min Scan# 1803
Delta R.T. 0.047 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17


Tgt Ion: 43 Resp: 20096
Ion Ratio Lower Upper
43 100
70 0.0 37.6 56.4#
55 0.0 19.0 28.6#
61 10.0 20.8 31.2#





#75
1,1,2-Tetrachloroethane
Concen: 5.262 ug/l
RT: 12.935 min Scan# 1867
Delta R.T. -0.006 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005



Tgt Ion: 83 Resp: 23510
Ion Ratio Lower Upper
83 100
131 8.4 4.9 14.7
85 65.1 32.0 96.0




Manual Integrations
APPROVED


Reviewed By :John Carbone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#76
1,2,3-Trichloropropane
Concen: 4.986 ug/l m
RT: 12.994 min Scan# 1877
Delta R.T. -0.000 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17

Tgt Ion: 75 Resp: 21527
Ion Ratio Lower Upper
75 100
77 179.7 79.3 237.9









#77
Bromobenzene
Concen: 5.119 ug/l
RT: 12.976 min Scan# 1874
Delta R.T. -0.006 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17

Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005

Tgt Ion:156 Resp: 1548
Ion Ratio Lower Upper
156 100
77 249.9 117.7 353.1
158 102.9 48.1 144.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#78
n-propylbenzene
Concen: 5.226 ug/l
RT: 13.035 min Scan# 1884
Delta R.T. 0.000 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17

Tgt Ion: 91 Resp: 83781
Ion Ratio Lower Upper
91 100
120 22.1 11.4 34.2





#79

2-Chlorotoluene

Concen: 5.128 ug/l

RT: 13.124 min Scan# 1899

Delta R.T. 0.000 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005

Tgt Ion: 91 Resp: 4929

Ion Ratio Lower Upper

91 100





126 32.5 16.2 48.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#80

1,3,5-Trimethylbenzene

Concen: 5.139 ug/l

RT: 13.171 min Scan# 1907

Delta R.T. -0.000 min

Lab File: VN086863.D




Acq: 06 Jun 2025 13:17


Tgt Ion:105 Resp: 55961

Ion Ratio Lower Upper

105 100


120 49.8 24.9 74.6





#81
trans-1,4-Dichloro-2-butene
Concen: 4.939 ug/l
RT: 12.741 min Scan# 1834
Delta R.T. 0.006 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005



Tgt Ion: 75 Resp: 9234
Ion Ratio Lower Upper
75 100
53 88.6 66.6 99.8
89 41.4 36.7 55.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#82
4-Chlorotoluene
Concen: 5.132 ug/l
RT: 13.223 min Scan# 1916
Delta R.T. -0.000 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17

Tgt Ion: 91 Resp: 49924
Ion Ratio Lower Upper
91 100
126 33.3 16.7 50.1





#83

tert-Butylbenzene

Concen: 5.103 ug/l

RT: 13.435 min Scan# 1952

Delta R.T. -0.000 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005





Tgt Ion:	119	Resp:	50874
Ion	Ratio	Lower	Upper
119	100		
91	65.3	32.1	96.5
134	24.5	13.4	40.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#84

1,2,4-Trimethylbenzene

Concen: 5.176 ug/l




RT: 13.482 min Scan# 1960


Delta R.T. 0.000 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17

Tgt Ion:	105	Resp:	56525
Ion	Ratio	Lower	Upper
105	100		
120	46.6	23.2	69.6









#85
sec-Butylbenzene
Concen: 5.161 ug/l
RT: 13.618 min Scan# 1983
Delta R.T. 0.000 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Tgt Ion:105 Resp: 7471
Ion Ratio Lower Upper
105 100
134 19.5 10.0 30.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#86
p-Isopropyltoluene
Concen: 5.182 ug/l
RT: 13.729 min Scan# 2002
Delta R.T. -0.000 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17

Tgt Ion:119 Resp: 62016
Ion Ratio Lower Upper
119 100
134 26.4 13.5 40.4
91 22.9 11.5 34.4





#87

1,3-Dichlorobenzene

Concen: 5.200 ug/l

RT: 13.735 min Scan# 2003

Delta R.T. -0.000 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005





Tgt Ion:	146	Resp:	30949
Ion Ratio	Lower	Upper	
146	100		
111	41.6	20.9	62.8
148	61.1	32.0	96.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#88

1,4-Dichlorobenzene

Concen: 5.330 ug/l




RT: 13.812 min Scan# 2016


Delta R.T. -0.000 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17

Tgt Ion:	146	Resp:	32342
Ion Ratio	Lower	Upper	
146	100		
111	46.1	19.9	59.7
148	63.3	31.8	95.4





#89

n-Butylbenzene

Concen: 5.142 ug/l

RT: 14.053 min Scan# 2057

Delta R.T. 0.000 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005





Tgt Ion:	91	Resp:	5960
Ion Ratio	Lower	Upper	
91	100		
92	54.1	27.5	82.4
134	24.9	13.0	39.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#90

Hexachloroethane

Concen: 4.794 ug/l




RT: 14.335 min Scan# 2105


Delta R.T. 0.000 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17

Tgt Ion:	117	Resp:	9725
Ion Ratio	Lower	Upper	
117	100		
201	70.6	33.2	99.6









#91
1,2-Dichlorobenzene
Concen: 5.228 ug/l
RT: 14.106 min Scan# 2066
Delta R.T. -0.000 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Tgt Ion:	146	Resp:	2989
Ion Ratio	Lower	Upper	
146	100		
111	44.0	21.6	64.6
148	64.3	31.7	95.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#92
1,2-Dibromo-3-Chloropropane
Concen: 5.363 ug/l
RT: 14.717 min Scan# 2170
Delta R.T. -0.000 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17

Tgt Ion:	75	Resp:	5731
Ion Ratio	Lower	Upper	
75	100		
155	76.0	41.1	123.4
157	98.8	52.7	158.1





#93

1,2,4-Trichlorobenzene

Concen: 5.143 ug/l

RT: 15.394 min Scan# 2284

Delta R.T. 0.006 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005

Tgt Ion:180 Resp: 18778

Ion Ratio Lower Upper

180 100

182 92.6 46.9 140.7





145 31.7 16.0 47.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#94

Hexachlorobutadiene

Concen: 5.199 ug/l

RT: 15.500 min Scan# 2303

Delta R.T. 0.000 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17




Tgt Ion:225 Resp: 7072


Ion Ratio Lower Upper

225 100

223 66.4 33.5 100.4

227 61.7 31.1 93.5









#95
Naphthalene
Concen: 4.965 ug/l
RT: 15.641 min Scan# 2327
Delta R.T. 0.000 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17

Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005

Tgt Ion	Ratio	Lower	Upper
128	100		
127	13.0	10.2	15.2
129	10.3	8.8	13.2




Manual Integrations
APPROVED

Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#96
1,2,3-Trichlorobenzene
Concen: 5.034 ug/l
RT: 15.841 min Scan# 2361
Delta R.T. 0.006 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17

Tgt Ion	Ratio	Lower	Upper
180	100		
182	90.9	47.0	141.0
145	35.3	16.6	49.8



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060625\
 Data File : VN086864.D
 Acq On : 06 Jun 2025 13:40
 Operator : JC\MD
 Sample : VSTDICC020
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC020

Manual Integrations
 APPROVED

Reviewed By : John Carlone 06/09/2025
 Supervised By : Mahesh Dadoda 06/09/2025

Quant Time: Jun 07 01:56:20 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 01:51:02 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	224006	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	418154	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	363172	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	181178	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	63390	21.136	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	42.280%#		
35) Dibromofluoromethane	8.177	113	51797	20.902	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	41.800%#		
50) Toluene-d8	10.571	98	201268	20.516	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	41.040%#		
62) 4-Bromofluorobenzene	12.853	95	74622	20.474	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	40.940%#		
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	2.159	85	48254	21.605	ug/l	92
3) Chloromethane	2.401	50	57778	20.033	ug/l	93
4) Vinyl Chloride	2.559	62	61254	20.589	ug/l	99
5) Bromomethane	3.006	94	34022	20.435	ug/l	95
6) Chloroethane	3.159	64	39625	20.619	ug/l	99
7) Trichlorofluoromethane	3.524	101	80978	20.832	ug/l	94
8) Diethyl Ether	3.983	74	34288	20.245	ug/l	98
9) 1,1,2-Trichlorotrifluo...	4.395	101	50458	20.661	ug/l	98
10) Methyl Iodide	4.612	142	65360	20.646	ug/l	98
11) Tert butyl alcohol	5.547	59	87102	107.031	ug/l	99
12) 1,1-Dichloroethene	4.359	96	50485	20.249	ug/l	96
13) Acrolein	4.200	56	22866	88.767	ug/l	97
14) Allyl chloride	5.047	41	82924	20.055	ug/l	98
15) Acrylonitrile	5.742	53	199498	104.881	ug/l	99
16) Acetone	4.447	43	163979	103.099	ug/l	98
17) Carbon Disulfide	4.736	76	138195	20.038	ug/l	97
18) Methyl Acetate	5.047	43	96547	20.830	ug/l	99
19) Methyl tert-butyl Ether	5.818	73	183780	20.358	ug/l	98
20) Methylene Chloride	5.294	84	57611	19.348	ug/l	98
21) trans-1,2-Dichloroethene	5.806	96	55643	20.059	ug/l	94
22) Diisopropyl ether	6.689	45	182472	20.935	ug/l	98
23) Vinyl Acetate	6.624	43	768556	104.365	ug/l	99
24) 1,1-Dichloroethane	6.583	63	103578	20.650	ug/l	99
25) 2-Butanone	7.494	43	270477	104.621	ug/l	100
26) 2,2-Dichloropropane	7.506	77	69692	17.862	ug/l	99
27) cis-1,2-Dichloroethene	7.500	96	68292	20.585	ug/l	98
28) Bromochloromethane	7.824	49	55196	22.381	ug/l	99
29) Tetrahydrofuran	7.853	42	178571	106.031	ug/l	99
30) Chloroform	7.977	83	102615	20.485	ug/l	93
31) Cyclohexane	8.271	56	100026	20.563	ug/l	98
32) 1,1,1-Trichloroethane	8.177	97	86810	20.376	ug/l	98
36) 1,1-Dichloropropene	8.383	75	73252	19.838	ug/l	100
37) Ethyl Acetate	7.571	43	96467	20.522	ug/l	98
38) Carbon Tetrachloride	8.377	117	72628	20.034	ug/l	97
39) Methylcyclohexane	9.606	83	98425	19.455	ug/l	99
40) Benzene	8.612	78	241444	19.996	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060625\
 Data File : VN086864.D
 Acq On : 06 Jun 2025 13:40
 Operator : JC\MD
 Sample : VSTDICC020
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC020

Manual Integrations
 APPROVED

Reviewed By : John Carlone 06/09/2025
 Supervised By : Mahesh Dadoda 06/09/2025

Quant Time: Jun 07 01:56:20 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 01:51:02 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.788	41	53244	20.169	ug/l	98
42) 1,2-Dichloroethane	8.677	62	74238	20.270	ug/l	100
43) Isopropyl Acetate	8.694	43	151601	20.094	ug/l	99
44) Trichloroethene	9.359	130	57110	19.946	ug/l	96
45) 1,2-Dichloropropane	9.630	63	59293	20.184	ug/l	97
46) Dibromomethane	9.712	93	40290	20.648	ug/l	98
47) Bromodichloromethane	9.894	83	80353	20.015	ug/l	98
48) Methyl methacrylate	9.682	41	70457	20.290	ug/l	97
49) 1,4-Dioxane	9.706	88	28231	446.886	ug/l #	98
51) 4-Methyl-2-Pentanone	10.447	43	481741	106.066	ug/l	99
52) Toluene	10.635	92	148000	20.056	ug/l	98
53) t-1,3-Dichloropropene	10.841	75	87683	19.532	ug/l	99
54) cis-1,3-Dichloropropene	10.318	75	94315	19.636	ug/l	99
55) 1,1,2-Trichloroethane	11.018	97	57195	20.143	ug/l	99
56) Ethyl methacrylate	10.882	69	94829	20.968	ug/l	98
57) 1,3-Dichloropropane	11.165	76	98205	19.938	ug/l	98
58) 2-Chloroethyl Vinyl ether	10.165	63	278986	103.426	ug/l	99
59) 2-Hexanone	11.206	43	303686	103.803	ug/l	95
60) Dibromochloromethane	11.359	129	59961	20.268	ug/l	99
61) 1,2-Dibromoethane	11.471	107	59139	20.320	ug/l	97
64) Tetrachloroethene	11.106	164	45374	19.742	ug/l	96
65) Chlorobenzene	11.894	112	160803	20.076	ug/l	97
66) 1,1,1,2-Tetrachloroethane	11.959	131	52069	20.223	ug/l	99
67) Ethyl Benzene	11.965	91	277087	20.083	ug/l	98
68) m/p-Xylenes	12.071	106	215378	40.780	ug/l	100
69) o-Xylene	12.400	106	102050	20.175	ug/l	98
70) Styrene	12.412	104	178061	20.572	ug/l	99
71) Bromoform	12.576	173	40070	21.004	ug/l #	99
73) Isopropylbenzene	12.694	105	264442	20.035	ug/l	100
74) N-amyl acetate	12.523	43	86003	19.633	ug/l	98
75) 1,1,2,2-Tetrachloroethane	12.941	83	92264	20.634	ug/l	99
76) 1,2,3-Trichloropropane	12.994	75	81846m	18.945	ug/l	
77) Bromobenzene	12.982	156	60877	20.111	ug/l	98
78) n-propylbenzene	13.035	91	322436	20.102	ug/l	100
79) 2-Chlorotoluene	13.123	91	193774	20.144	ug/l	100
80) 1,3,5-Trimethylbenzene	13.170	105	219937	20.183	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.735	75	37808	20.209	ug/l	92
82) 4-Chlorotoluene	13.223	91	193544	19.885	ug/l	99
83) tert-Butylbenzene	13.435	119	198303	19.880	ug/l	99
84) 1,2,4-Trimethylbenzene	13.482	105	221990	20.315	ug/l	100
85) sec-Butylbenzene	13.617	105	292563	20.196	ug/l	100
86) p-Isopropyltoluene	13.729	119	240682	20.099	ug/l	100
87) 1,3-Dichlorobenzene	13.735	146	120099	20.166	ug/l	99
88) 1,4-Dichlorobenzene	13.812	146	120102	19.780	ug/l	97
89) n-Butylbenzene	14.053	91	227536	19.617	ug/l	100
90) Hexachloroethane	14.335	117	41202	20.299	ug/l	100
91) 1,2-Dichlorobenzene	14.106	146	115645	20.212	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.717	75	21025	19.661	ug/l	98
93) 1,2,4-Trichlorobenzene	15.388	180	71826	19.659	ug/l	98
94) Hexachlorobutadiene	15.500	225	27887	20.486	ug/l	97
95) Naphthalene	15.641	128	272584	20.043	ug/l	100
96) 1,2,3-Trichlorobenzene	15.841	180	71929	19.815	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060625\
Data File : VN086864.D
Acq On : 06 Jun 2025 13:40
Operator : JC\MD
Sample : VSTDICC020
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Manual Integrations
APPROVED

Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025

Quant Time: Jun 07 01:56:20 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 01:51:02 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060625\
 Data File : VN086864.D
 Acq On : 06 Jun 2025 13:40
 Operator : JC\MD
 Sample : VSTDIC020
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1


Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC020

Manual Integrations APPROVED

Reviewed By : John Carlone 06/09/2025
 Supervised By : Mahesh Dadoda 06/09/2025


Quant Time: Jun 07 01:56:20 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 01:51:02 2025
 Response via : Initial Calibration





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.235 min Scan# 1068
Delta R.T. -0.001 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020



Tgt Ion:168 Resp: 224000
Ion Ratio Lower Upper
168 100
99 61.1 49.1 73.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#2
Dichlorodifluoromethane
Concen: 21.605 ug/l
RT: 2.159 min Scan# 35
Delta R.T. 0.006 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40

Tgt Ion: 85 Resp: 48254
Ion Ratio Lower Upper
85 100
87 37.1 16.2 48.6









#3
Chloromethane
Concen: 20.033 ug/l
RT: 2.401 min Scan# 70
Delta R.T. 0.006 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Tgt Ion: 50 Resp: 57778
Ion Ratio Lower Upper
50 100
52 35.8 25.7 38.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#4
Vinyl Chloride
Concen: 20.589 ug/l
RT: 2.559 min Scan# 103
Delta R.T. 0.006 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40

Tgt Ion: 62 Resp: 61254
Ion Ratio Lower Upper
62 100
64 31.7 26.0 39.0









#5
Bromomethane
Concen: 20.435 ug/l
RT: 3.006 min Scan# 11
Delta R.T. 0.012 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40

Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020

Tgt Ion: 94 Resp: 3402
Ion Ratio Lower Upper
94 100
96 92.9 77.9 116.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#6
Chloroethane
Concen: 20.619 ug/l
RT: 3.159 min Scan# 205
Delta R.T. 0.006 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40

Tgt Ion: 64 Resp: 39625
Ion Ratio Lower Upper
64 100
66 30.4 24.6 36.8









#7
Trichlorofluoromethane
Concen: 20.832 ug/l
RT: 3.524 min Scan# 267
Delta R.T. -0.000 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Tgt Ion:101 Resp: 80978
Ion Ratio Lower Upper
101 100
103 60.4 52.2 78.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#8
Diethyl Ether
Concen: 20.245 ug/l
RT: 3.983 min Scan# 345
Delta R.T. -0.000 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40

Tgt Ion: 74 Resp: 34288
Ion Ratio Lower Upper
74 100
45 92.9 45.5 136.5





#9

1,1,2-Trichlorotrifluoroethane

Concen: 20.661 ug/l

RT: 4.395 min Scan# 415

Delta R.T. -0.000 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC020

Tgt Ion:	101	Resp:	5045
Ion	Ratio	Lower	Upper
101	100		
85	41.7	33.8	50.8
151	71.3	58.7	88.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#10

Methyl Iodide

Concen: 20.646 ug/l




RT: 4.612 min Scan# 452


Delta R.T. 0.006 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40

Tgt Ion:	142	Resp:	65360
Ion	Ratio	Lower	Upper
142	100		
127	39.2	30.9	46.3
141	14.3	10.4	15.6





#11

Tert butyl alcohol

Concen: 107.031 ug/l

RT: 5.547 min Scan# 611

Delta R.T. -0.000 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC020

Tgt Ion: 59 Resp: 87102

Ion Ratio Lower Upper

59 100





57 9.9 8.1 12.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#12

1,1-Dichloroethene

Concen: 20.249 ug/l

RT: 4.359 min Scan# 409

Delta R.T. -0.000 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40




Tgt Ion: 96 Resp: 50485


Ion Ratio Lower Upper

96 100

61 153.1 125.8 188.6

98 61.0 52.5 78.7





#13

Acrolein

Concen: 88.767 ug/l

RT: 4.200 min Scan# 382

Delta R.T. -0.000 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC020

Tgt Ion: 56 Resp: 2286

Ion Ratio Lower Upper

56 100





55 71.6 55.6 83.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#14

Allyl chloride

Concen: 20.055 ug/l

RT: 5.047 min Scan# 526

Delta R.T. 0.006 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40




Tgt Ion: 41 Resp: 82924


Ion Ratio Lower Upper

41 100

39 62.6 51.3 76.9


76 38.1 31.4 47.0





#15
Acrylonitrile
Concen: 104.881 ug/l
RT: 5.742 min Scan# 644
Delta R.T. -0.000 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020



Tgt Ion: 53 Resp: 199498
Ion Ratio Lower Upper
53 100
52 80.3 65.0 97.6
51 33.7 27.2 40.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#16
Acetone
Concen: 103.099 ug/l
RT: 4.447 min Scan# 424
Delta R.T. -0.000 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40


Tgt Ion: 43 Resp: 163979
Ion Ratio Lower Upper
43 100
58 35.9 28.0 42.0





#17
Carbon Disulfide
Concen: 20.038 ug/l
RT: 4.736 min Scan# 473
Delta R.T. 0.006 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020



Tgt Ion: 76 Resp: 13819
Ion Ratio Lower Upper
76 100
78 7.8 7.3 10.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#18
Methyl Acetate
Concen: 20.830 ug/l
RT: 5.047 min Scan# 526
Delta R.T. 0.006 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40


Tgt Ion: 43 Resp: 96547
Ion Ratio Lower Upper
43 100
74 26.2 20.7 31.1





#19
Methyl tert-butyl Ether
Concen: 20.358 ug/l
RT: 5.818 min Scan# 656
Delta R.T. 0.006 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020



Tgt Ion: 73 Resp: 183780
Ion Ratio Lower Upper
73 100
57 22.9 17.4 26.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#20
Methylene Chloride
Concen: 19.348 ug/l
RT: 5.294 min Scan# 568
Delta R.T. -0.000 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40


Tgt Ion: 84 Resp: 57611
Ion Ratio Lower Upper
84 100
49 114.5 90.5 135.7
51 35.9 28.5 42.7
86 63.0 48.1 72.1





#21
trans-1,2-Dichloroethene
Concen: 20.059 ug/l
RT: 5.806 min Scan# 611
Delta R.T. -0.000 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020



Tgt Ion: 96 Resp: 5564
Ion Ratio Lower Upper
96 100
61 126.9 109.2 163.8
98 61.8 49.8 74.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#22
Diisopropyl ether
Concen: 20.935 ug/l
RT: 6.689 min Scan# 805
Delta R.T. -0.000 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40


Tgt Ion: 45 Resp: 182472
Ion Ratio Lower Upper
45 100
43 50.7 41.9 62.9
87 27.0 22.4 33.6
59 13.1 9.8 14.6





#23
Vinyl Acetate
Concen: 104.365 ug/l
RT: 6.624 min Scan# 794
Delta R.T. 0.006 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020



Tgt Ion: 43 Resp: 768550
Ion Ratio Lower Upper
43 100
86 11.4 9.5 14.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#24
1,1-Dichloroethane
Concen: 20.650 ug/l
RT: 6.583 min Scan# 787
Delta R.T. -0.000 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40


Tgt Ion: 63 Resp: 103578
Ion Ratio Lower Upper
63 100
98 7.5 3.6 10.8
100 4.7 2.6 7.8





#25
2-Butanone
Concen: 104.621 ug/l
RT: 7.494 min Scan# 942
Delta R.T. -0.000 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020



Tgt Ion: 43 Resp: 27047
Ion Ratio Lower Upper
43 100
72 28.6 23.0 34.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#26
2,2-Dichloropropane
Concen: 17.862 ug/l
RT: 7.506 min Scan# 944
Delta R.T. 0.006 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40

Tgt Ion: 77 Resp: 69692
Ion Ratio Lower Upper
77 100
97 23.7 11.6 34.7





#27

cis-1,2-Dichloroethene

Concen: 20.585 ug/l

RT: 7.500 min Scan# 943

Delta R.T. -0.000 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40

Instrument :

MSVOA_N

ClientSampleId :





VSTDICC020

Tgt Ion:	96	Resp:	6829
Ion	Ratio	Lower	Upper
96	100		
61	135.5	0.0	278.0
98	63.0	0.0	128.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#28

Bromochloromethane

Concen: 22.381 ug/l




RT: 7.824 min Scan# 998


Delta R.T. -0.000 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40

Tgt Ion:	49	Resp:	55196
Ion	Ratio	Lower	Upper
49	100		
129	1.9	0.0	3.8
130	77.1	62.2	93.2





#29

Tetrahydrofuran

Concen: 106.031 ug/l

RT: 7.853 min Scan# 1003

Delta R.T. -0.000 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020

Tgt Ion: 42 Resp: 17857

Ion Ratio Lower Upper

42 100

72 46.6 37.8 56.6





71 43.5 35.6 53.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#30

Chloroform

Concen: 20.485 ug/l

RT: 7.977 min Scan# 1024

Delta R.T. -0.000 min

Lab File: VN086864.D




Acq: 06 Jun 2025 13:40


Tgt Ion: 83 Resp: 102615

Ion Ratio Lower Upper

83 100


85 60.7 52.9 79.3





#31
Cyclohexane
Concen: 20.563 ug/l
RT: 8.271 min Scan# 1074
Delta R.T. -0.000 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020



Tgt Ion: 56 Resp: 100020
Ion Ratio Lower Upper
56 100
69 29.0 25.4 38.0
84 85.6 68.8 103.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#32
1,1,1-Trichloroethane
Concen: 20.376 ug/l
RT: 8.177 min Scan# 1058
Delta R.T. -0.000 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40

Tgt Ion: 97 Resp: 86810
Ion Ratio Lower Upper
97 100
99 63.8 52.0 78.0
61 43.3 35.9 53.9





#33

1,2-Dichloroethane-d4

Concen: 21.136 ug/l

RT: 8.588 min Scan# 1128

Delta R.T. -0.000 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020

Tgt Ion: 65 Resp: 63390

Ion Ratio Lower Upper

65 100





67 55.3 0.0 105.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.106 min Scan# 1216

Delta R.T. -0.000 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40




Tgt Ion:114 Resp: 418154


Ion Ratio Lower Upper

114 100

63 20.6 0.0 39.6

88 15.8 0.0 30.2





#35

Dibromofluoromethane

Concen: 20.902 ug/l

RT: 8.177 min Scan# 1057

Delta R.T. 0.006 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020





Tgt Ion:	113	Resp:	5179
Ion Ratio	Lower	Upper	
113	100		
111	105.1	84.2	126.2
192	17.7	14.2	21.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#36

1,1-Dichloropropene

Concen: 19.838 ug/l




RT: 8.383 min Scan# 1093


Delta R.T. 0.006 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40

Tgt Ion:	75	Resp:	73252
Ion Ratio	Lower	Upper	
75	100		
110	35.6	17.6	52.9
77	31.8	25.3	37.9





#37

Ethyl Acetate

Concen: 20.522 ug/l

RT: 7.571 min Scan# 91

Delta R.T. -0.006 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020

Tgt Ion: 43 Resp: 9646

Ion Ratio Lower Upper

43 100

61 13.9 12.1 18.1





70 12.6 9.8 14.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#38

Carbon Tetrachloride

Concen: 20.034 ug/l

RT: 8.377 min Scan# 1092

Delta R.T. 0.006 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40




Tgt Ion:117 Resp: 72628


Ion Ratio Lower Upper

117 100

119 91.3 75.0 112.6

121 28.1 24.9 37.3





#39

Methylcyclohexane

Concen: 19.455 ug/l

RT: 9.606 min Scan# 1130

Delta R.T. -0.000 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020





Tgt Ion:	83	Resp:	9842
Ion	Ratio	Lower	Upper
83	100		
55	75.7	61.0	91.6
98	47.4	38.9	58.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#40

Benzene

Concen: 19.996 ug/l




RT: 8.612 min Scan# 1132


Delta R.T. -0.000 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40


Tgt Ion:	78	Resp:	241444
Ion	Ratio	Lower	Upper
78	100		
77	23.8	18.7	28.1





#41
Methacrylonitrile
Concen: 20.169 ug/l
RT: 7.788 min Scan# 992
Delta R.T. -0.000 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020



Tgt Ion: 41 Resp: 53244
Ion Ratio Lower Upper
41 100
39 49.5 41.2 61.8
67 77.3 61.4 92.2
52 29.6 22.7 34.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#42
1,2-Dichloroethane
Concen: 20.270 ug/l
RT: 8.677 min Scan# 1143
Delta R.T. -0.000 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40

Tgt Ion: 62 Resp: 74238
Ion Ratio Lower Upper
62 100
98 10.6 0.0 21.2





#43

Isopropyl Acetate

Concen: 20.094 ug/l

RT: 8.694 min Scan# 1146

Delta R.T. -0.000 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020

Tgt Ion: 43 Resp: 15160

Ion Ratio Lower Upper

43 100

61 27.9 22.1 33.1





87 14.4 11.8 17.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#44

Trichloroethene

Concen: 19.946 ug/l

RT: 9.359 min Scan# 1259

Delta R.T. -0.000 min

Lab File: VN086864.D




Acq: 06 Jun 2025 13:40


Tgt Ion:130 Resp: 57110

Ion Ratio Lower Upper

130 100

95 102.1 0.0 196.2





#45

1,2-Dichloropropane

Concen: 20.184 ug/l

RT: 9.630 min Scan# 1305

Delta R.T. -0.000 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC020

Tgt Ion: 63 Resp: 5929

Ion Ratio Lower Upper

63 100





65 30.4 25.5 38.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#46

Dibromomethane

Concen: 20.648 ug/l

RT: 9.712 min Scan# 1319

Delta R.T. -0.006 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40




Tgt Ion: 93 Resp: 40290


Ion Ratio Lower Upper

93 100

95 84.1 67.1 100.7

174 90.1 74.2 111.4





#47

Bromodichloromethane

Concen: 20.015 ug/l

RT: 9.894 min Scan# 1350

Delta R.T. -0.000 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020





Tgt Ion:	83	Resp:	8035
Ion	Ratio	Lower	Upper
83	100		
85	62.3	50.9	76.3
127	7.6	6.6	10.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#48

Methyl methacrylate

Concen: 20.290 ug/l




RT: 9.682 min Scan# 1314


Delta R.T. -0.000 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40

Tgt Ion:	41	Resp:	70457
Ion	Ratio	Lower	Upper
41	100		
69	89.4	74.6	112.0
39	50.3	40.2	60.4





#49

1,4-Dioxane

Concen: 446.886 ug/l

RT: 9.706 min Scan# 1317

Delta R.T. 0.006 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC020

Tgt Ion: 88 Resp: 2823

Ion Ratio Lower Upper

88 100

43 19.7 0.0 0.0





58 70.1 57.7 86.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#50

Toluene-d8

Concen: 20.516 ug/l

RT: 10.571 min Scan# 1465

Delta R.T. -0.000 min

Lab File: VN086864.D




Acq: 06 Jun 2025 13:40


Tgt Ion: 98 Resp: 201268

Ion Ratio Lower Upper

98 100

100 67.8 53.4 80.0









#51
4-Methyl-2-Pentanone
Concen: 106.066 ug/l
RT: 10.447 min Scan# 1444
Delta R.T. -0.000 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Tgt Ion: 43 Resp: 48174
Ion Ratio Lower Upper
43 100
58 42.0 34.2 51.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#52
Toluene
Concen: 20.056 ug/l
RT: 10.635 min Scan# 1476
Delta R.T. -0.000 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40

Tgt Ion: 92 Resp: 148000
Ion Ratio Lower Upper
92 100
91 172.9 136.5 204.7









#53
t-1,3-Dichloropropene
Concen: 19.532 ug/l
RT: 10.841 min Scan# 1511
Delta R.T. -0.000 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40

Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020

Tgt Ion: 75 Resp: 8768
Ion Ratio Lower Upper
75 100
77 32.7 26.4 39.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#54
cis-1,3-Dichloropropene
Concen: 19.636 ug/l
RT: 10.318 min Scan# 1422
Delta R.T. -0.000 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40


Tgt Ion: 75 Resp: 94315
Ion Ratio Lower Upper
75 100
77 32.5 26.2 39.4
39 41.3 32.3 48.5





#55
1,1,2-Trichloroethane
Concen: 20.143 ug/l
RT: 11.018 min Scan# 1541
Delta R.T. -0.000 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020



Tgt Ion: 97 Resp: 5719
Ion Ratio Lower Upper
97 100
83 87.1 69.7 104.5
85 54.0 45.1 67.7
99 62.2 50.6 75.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#56
Ethyl methacrylate
Concen: 20.968 ug/l
RT: 10.882 min Scan# 1518
Delta R.T. -0.000 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40


Tgt Ion: 69 Resp: 94829
Ion Ratio Lower Upper
69 100
41 58.9 47.2 70.8
39 25.9 23.0 34.6





#57
1,3-Dichloropropane
Concen: 19.938 ug/l
RT: 11.165 min Scan# 1566
Delta R.T. -0.000 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020



Tgt Ion: 76 Resp: 9820
Ion Ratio Lower Upper
76 100
78 32.4 25.3 37.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#58
2-Chloroethyl Vinyl ether
Concen: 103.426 ug/l
RT: 10.165 min Scan# 1396
Delta R.T. -0.000 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40

Tgt Ion: 63 Resp: 278986
Ion Ratio Lower Upper
63 100
106 28.1 23.0 34.6





#59

2-Hexanone

Concen: 103.803 ug/l

RT: 11.206 min Scan# 1572

Delta R.T. 0.006 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020

Tgt Ion: 43 Resp: 30368

Ion Ratio Lower Upper

43 100





58 64.8 30.4 91.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#60

Dibromochloromethane

Concen: 20.268 ug/l

RT: 11.359 min Scan# 1599

Delta R.T. -0.000 min

Lab File: VN086864.D




Acq: 06 Jun 2025 13:40


Tgt Ion:129 Resp: 59961

Ion Ratio Lower Upper

129 100


127 76.3 38.6 115.8





#61
1,2-Dibromoethane
Concen: 20.320 ug/l
RT: 11.471 min Scan# 1618
Delta R.T. -0.000 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020




Tgt Ion:107 Resp: 59139
Ion Ratio Lower Upper
107 100
109 93.0 76.3 114.5

Manual Integrations
APPROVED



Reviewed By :John Carbone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025




#62
4-Bromofluorobenzene
Concen: 20.474 ug/l
RT: 12.853 min Scan# 1853
Delta R.T. 0.006 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40




Tgt Ion: 95 Resp: 74622
Ion Ratio Lower Upper
95 100
174 71.7 0.0 141.8
176 68.0 0.0 132.6





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 1685
Delta R.T. -0.000 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020



Tgt Ion:117 Resp: 363172
Ion Ratio Lower Upper
117 100
82 58.0 44.6 67.0
119 32.6 25.5 38.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#64
Tetrachloroethene
Concen: 19.742 ug/l
RT: 11.106 min Scan# 1556
Delta R.T. -0.000 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40

Tgt Ion:164 Resp: 45374
Ion Ratio Lower Upper
164 100
166 124.4 101.8 152.8
129 100.4 83.0 124.4
131 92.1 78.2 117.2





#65

Chlorobenzene

Concen: 20.076 ug/l

RT: 11.894 min Scan# 11

Delta R.T. -0.000 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020

Tgt Ion:112 Resp: 16080

Ion Ratio Lower Upper

112 100





114 30.8 26.2 39.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#66

1,1,1,2-Tetrachloroethane

Concen: 20.223 ug/l

RT: 11.959 min Scan# 1701

Delta R.T. -0.006 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40




Tgt Ion:131 Resp: 52069


Ion Ratio Lower Upper

131 100

133 96.3 48.3 144.8

119 69.6 34.0 101.8





#67

Ethyl Benzene

Concen: 20.083 ug/l

RT: 11.965 min Scan# 1702

Delta R.T. -0.000 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020

Tgt Ion: 91 Resp: 27708

Ion Ratio Lower Upper

91 100





106 30.5 25.4 38.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#68

m/p-Xylenes

Concen: 40.780 ug/l

RT: 12.071 min Scan# 1720

Delta R.T. -0.000 min

Lab File: VN086864.D




Acq: 06 Jun 2025 13:40


Tgt Ion:106 Resp: 215378

Ion Ratio Lower Upper

106 100

91 199.8 159.4 239.0





#69

o-Xylene

Concen: 20.175 ug/l

RT: 12.400 min Scan# 1776

Delta R.T. -0.000 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020

Tgt Ion:106 Resp: 102050

Ion Ratio Lower Upper

106 100





91 215.9 106.1 318.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#70

Styrene

Concen: 20.572 ug/l

RT: 12.412 min Scan# 1778

Delta R.T. -0.000 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40




Tgt Ion:104 Resp: 178061


Ion Ratio Lower Upper

104 100

78 48.3 38.5 57.7


103 54.4 43.1 64.7





#71
Bromoform
Concen: 21.004 ug/l
RT: 12.576 min Scan# 11
Delta R.T. -0.006 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020



Tgt Ion:173 Resp: 40070
Ion Ratio Lower Upper
173 100
175 47.6 24.1 72.4
254 0.0 0.0 0.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.788 min Scan# 2012
Delta R.T. -0.006 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40


Tgt Ion:152 Resp: 181178
Ion Ratio Lower Upper
152 100
115 61.4 30.1 90.5
150 164.5 0.0 345.0





#73
Isopropylbenzene
Concen: 20.035 ug/l
RT: 12.694 min Scan# 1826
Delta R.T. -0.000 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020



Tgt Ion:105 Resp: 26444
Ion Ratio Lower Upper
105 100
120 27.0 13.5 40.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#74
N-ethyl acetate
Concen: 19.633 ug/l
RT: 12.523 min Scan# 1797
Delta R.T. 0.012 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40


Tgt Ion: 43 Resp: 86003
Ion Ratio Lower Upper
43 100
70 47.7 37.6 56.4
55 26.6 19.0 28.6
61 25.9 20.8 31.2





#75
1,1,2-Tetrachloroethane
Concen: 20.634 ug/l
RT: 12.941 min Scan# 1868
Delta R.T. -0.000 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020



Tgt Ion: 83 Resp: 92264
Ion Ratio Lower Upper
83 100
131 9.8 4.9 14.7
85 64.5 32.0 96.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#76
1,2,3-Trichloropropane
Concen: 18.945 ug/l m
RT: 12.994 min Scan# 1877
Delta R.T. -0.000 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40

Tgt Ion: 75 Resp: 81846
Ion Ratio Lower Upper
75 100
77 176.5 79.3 237.9









#77
Bromobenzene
Concen: 20.111 ug/l
RT: 12.982 min Scan# 1875
Delta R.T. -0.000 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Tgt Ion:156 Resp: 6087
Ion Ratio Lower Upper
156 100
77 237.3 117.7 353.1
158 99.0 48.1 144.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#78
n-propylbenzene
Concen: 20.102 ug/l
RT: 13.035 min Scan# 1884
Delta R.T. -0.000 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40

Tgt Ion: 91 Resp: 322436
Ion Ratio Lower Upper
91 100
120 22.9 11.4 34.2





#79

2-Chlorotoluene

Concen: 20.144 ug/l

RT: 13.123 min Scan# 1899

Delta R.T. -0.000 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020

Tgt Ion: 91 Resp: 193774

Ion Ratio Lower Upper

91 100





126 32.3 16.2 48.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#80

1,3,5-Trimethylbenzene

Concen: 20.183 ug/l

RT: 13.170 min Scan# 1907

Delta R.T. -0.000 min

Lab File: VN086864.D




Acq: 06 Jun 2025 13:40


Tgt Ion:105 Resp: 219937

Ion Ratio Lower Upper

105 100


120 49.6 24.9 74.6





#81
trans-1,4-Dichloro-2-butene
Concen: 20.209 ug/l
RT: 12.735 min Scan# 1833
Delta R.T. -0.000 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020



Tgt Ion: 75 Resp: 37808
Ion Ratio Lower Upper
75 100
53 77.3 66.6 99.8
89 39.6 36.7 55.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#82
4-Chlorotoluene
Concen: 19.885 ug/l
RT: 13.223 min Scan# 1916
Delta R.T. -0.000 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40


Tgt Ion: 91 Resp: 193544
Ion Ratio Lower Upper
91 100
126 33.7 16.7 50.1





#83
tert-Butylbenzene
Concen: 19.880 ug/l
RT: 13.435 min Scan# 1952
Delta R.T. -0.000 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020



Tgt Ion:119 Resp: 19830
Ion Ratio Lower Upper
119 100
91 65.2 32.1 96.5
134 26.5 13.4 40.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#84
1,2,4-Trimethylbenzene
Concen: 20.315 ug/l
RT: 13.482 min Scan# 1960
Delta R.T. -0.000 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40


Tgt Ion:105 Resp: 221990
Ion Ratio Lower Upper
105 100
120 46.1 23.2 69.6





#85
sec-Butylbenzene
Concen: 20.196 ug/l
RT: 13.617 min Scan# 1983
Delta R.T. -0.000 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020



Tgt Ion:105 Resp: 29256
Ion Ratio Lower Upper
105 100
134 20.0 10.0 30.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#86
p-Isopropyltoluene
Concen: 20.099 ug/l
RT: 13.729 min Scan# 2002
Delta R.T. -0.000 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40

Tgt Ion:119 Resp: 240682
Ion Ratio Lower Upper
119 100
134 27.1 13.5 40.4
91 22.7 11.5 34.4





#87

1,3-Dichlorobenzene

Concen: 20.166 ug/l

RT: 13.735 min Scan# 2003

Delta R.T. -0.000 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020

Tgt Ion:146 Resp: 120099

Ion Ratio Lower Upper

146 100

111 41.5 20.9 62.8





148 62.5 32.0 96.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#88

1,4-Dichlorobenzene

Concen: 19.780 ug/l

RT: 13.812 min Scan# 2016

Delta R.T. -0.000 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40




Tgt Ion:146 Resp: 120102


Ion Ratio Lower Upper

146 100

111 42.5 19.9 59.7

148 64.7 31.8 95.4





#89

n-Butylbenzene

Concen: 19.617 ug/l

RT: 14.053 min Scan# 2057

Delta R.T. -0.000 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020





Tgt Ion:	91	Resp:	227530
Ion	Ratio	Lower	Upper
91	100		
92	54.5	27.5	82.4
134	26.2	13.0	39.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#90

Hexachloroethane

Concen: 20.299 ug/l




RT: 14.335 min Scan# 2105


Delta R.T. -0.000 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40

Tgt Ion:	117	Resp:	41202
Ion	Ratio	Lower	Upper
117	100		
201	66.2	33.2	99.6





#91

1,2-Dichlorobenzene

Concen: 20.212 ug/l

RT: 14.106 min Scan# 2066

Delta R.T. -0.000 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020

Tgt Ion:146 Resp: 11564

Ion Ratio Lower Upper

146 100

111 42.9 21.6 64.6





148 63.9 31.7 95.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#92

1,2-Dibromo-3-Chloropropane

Concen: 19.661 ug/l

RT: 14.717 min Scan# 2170

Delta R.T. -0.000 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40




Tgt Ion: 75 Resp: 21025


Ion Ratio Lower Upper

75 100

155 81.8 41.1 123.4

157 109.0 52.7 158.1









#93
1,2,4-Trichlorobenzene
Concen: 19.659 ug/l
RT: 15.388 min Scan# 2182
Delta R.T. -0.000 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Tgt Ion:180 Resp: 71820
Ion Ratio Lower Upper
180 100
182 95.4 46.9 140.7
145 33.3 16.0 47.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#94
Hexachlorobutadiene
Concen: 20.486 ug/l
RT: 15.500 min Scan# 2303
Delta R.T. -0.000 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40


Tgt Ion:225 Resp: 27887
Ion Ratio Lower Upper
225 100
223 63.0 33.5 100.4
227 63.3 31.1 93.5





#95
Naphthalene
Concen: 20.043 ug/l
RT: 15.641 min Scan# 2327
Delta R.T. -0.000 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020




Tgt Ion:128 Resp: 272584
Ion Ratio Lower Upper
128 100
127 12.7 10.2 15.2
129 11.0 8.8 13.2

Manual Integrations
APPROVED



Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#96
1,2,3-Trichlorobenzene
Concen: 19.815 ug/l
RT: 15.841 min Scan# 2361
Delta R.T. 0.006 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40



Tgt Ion:180 Resp: 71929
Ion Ratio Lower Upper
180 100
182 95.3 47.0 141.0
145 33.1 16.6 49.8



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060625\
 Data File : VN086865.D
 Acq On : 06 Jun 2025 14:03
 Operator : JC\MD
 Sample : VSTDICCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICCC050

Manual Integrations
 APPROVED

Reviewed By : John Carlone 06/09/2025
 Supervised By : Mahesh Dadoda 06/09/2025

Quant Time: Jun 07 01:57:17 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 01:51:02 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	207354	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	378587	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	332766	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	167532	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	103736	37.366	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	74.740%		
35) Dibromofluoromethane	8.171	113	83058	37.020	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	74.040%#		
50) Toluene-d8	10.571	98	326105	36.716	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	73.440%#		
62) 4-Bromofluorobenzene	12.847	95	122965	37.264	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	74.520%#		
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	2.154	85	103958	50.283	ug/l	100
3) Chloromethane	2.395	50	123866	46.397	ug/l	100
4) Vinyl Chloride	2.554	62	132783	48.215	ug/l	100
5) Bromomethane	2.995	94	73974	47.999	ug/l	100
6) Chloroethane	3.154	64	84505	47.504	ug/l	100
7) Trichlorofluoromethane	3.524	101	172860	48.041	ug/l	100
8) Diethyl Ether	3.983	74	75193	47.963	ug/l	100
9) 1,1,2-Trichlorotrifluo...	4.395	101	107695	47.640	ug/l	100
10) Methyl Iodide	4.606	142	141678	48.346	ug/l	100
11) Tert butyl alcohol	5.548	59	182315	242.020	ug/l	100
12) 1,1-Dichloroethene	4.359	96	110549	47.900	ug/l	100
13) Acrolein	4.201	56	47026	197.217	ug/l	100
14) Allyl chloride	5.042	41	179313	46.848	ug/l	100
15) Acrylonitrile	5.742	53	421941	239.638	ug/l	100
16) Acetone	4.448	43	333857	226.765	ug/l	100
17) Carbon Disulfide	4.730	76	295753	46.328	ug/l	100
18) Methyl Acetate	5.042	43	204538	47.673	ug/l	100
19) Methyl tert-butyl Ether	5.812	73	400775	47.960	ug/l	100
20) Methylene Chloride	5.295	84	125498	45.531	ug/l	100
21) trans-1,2-Dichloroethene	5.806	96	117660	45.822	ug/l	100
22) Diisopropyl ether	6.689	45	384939	47.711	ug/l	100
23) Vinyl Acetate	6.618	43	1649023	241.910	ug/l	100
24) 1,1-Dichloroethane	6.583	63	220477	47.485	ug/l	100
25) 2-Butanone	7.494	43	571221	238.693	ug/l	100
26) 2,2-Dichloropropane	7.500	77	187567	51.933	ug/l	100
27) cis-1,2-Dichloroethene	7.500	96	144911	47.189	ug/l	100
28) Bromochloromethane	7.824	49	96707	42.362	ug/l	100
29) Tetrahydrofuran	7.853	42	371357	238.210	ug/l	100
30) Chloroform	7.977	83	220093	47.467	ug/l	100
31) Cyclohexane	8.271	56	208105	46.217	ug/l	100
32) 1,1,1-Trichloroethane	8.177	97	185502	47.037	ug/l	100
36) 1,1-Dichloropropene	8.377	75	158341	47.362	ug/l	100
37) Ethyl Acetate	7.577	43	203091	47.720	ug/l	100
38) Carbon Tetrachloride	8.371	117	154677	47.126	ug/l	100
39) Methylcyclohexane	9.606	83	215849	47.126	ug/l	100
40) Benzene	8.612	78	509334	46.590	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060625\
 Data File : VN086865.D
 Acq On : 06 Jun 2025 14:03
 Operator : JC\MD
 Sample : VSTDICCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICCC050

Manual Integrations
 APPROVED

Reviewed By :John Carlone 06/09/2025
 Supervised By :Mahesh Dadoda 06/09/2025

Quant Time: Jun 07 01:57:17 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 01:51:02 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.789	41	114727	48.000	ug/l	100
42) 1,2-Dichloroethane	8.677	62	155713	46.960	ug/l	100
43) Isopropyl Acetate	8.694	43	322741	47.248	ug/l	100
44) Trichloroethene	9.359	130	123847	47.776	ug/l	100
45) 1,2-Dichloropropane	9.630	63	125673	47.251	ug/l	100
46) Dibromomethane	9.718	93	83811	47.442	ug/l	100
47) Bromodichloromethane	9.894	83	173200	47.651	ug/l	100
48) Methyl methacrylate	9.683	41	149019	47.399	ug/l	100
49) 1,4-Dioxane	9.700	88	59060	1032.607	ug/l #	100
51) 4-Methyl-2-Pentanone	10.447	43	1017688	247.484	ug/l	100
52) Toluene	10.635	92	316219	47.331	ug/l	100
53) t-1,3-Dichloropropene	10.841	75	197475	48.587	ug/l	100
54) cis-1,3-Dichloropropene	10.318	75	208523	47.950	ug/l	100
55) 1,1,2-Trichloroethane	11.018	97	121969	47.445	ug/l	100
56) Ethyl methacrylate	10.882	69	210467	51.401	ug/l	100
57) 1,3-Dichloropropane	11.165	76	212243	47.593	ug/l	100
58) 2-Chloroethyl Vinyl ether	10.165	63	519139	212.570	ug/l	100
59) 2-Hexanone	11.200	43	696567	262.977	ug/l	100
60) Dibromochloromethane	11.359	129	128715	48.056	ug/l	100
61) 1,2-Dibromoethane	11.471	107	124522	47.258	ug/l	100
64) Tetrachloroethene	11.106	164	97679	46.382	ug/l	100
65) Chlorobenzene	11.894	112	340439	46.386	ug/l	100
66) 1,1,1,2-Tetrachloroethane	11.965	131	112333	47.615	ug/l	100
67) Ethyl Benzene	11.965	91	597593	47.271	ug/l	100
68) m/p-Xylenes	12.071	106	466620	96.424	ug/l	100
69) o-Xylene	12.400	106	224299	48.395	ug/l	100
70) Styrene	12.412	104	386596	48.745	ug/l	100
71) Bromoform	12.582	173	88127	50.415	ug/l #	100
73) Isopropylbenzene	12.694	105	573905	47.023	ug/l	100
74) N-amyl acetate	12.512	43	212064	52.353	ug/l	100
75) 1,1,2,2-Tetrachloroethane	12.941	83	197360	47.732	ug/l	100
76) 1,2,3-Trichloropropane	12.994	75	196473m	49.183	ug/l	
77) Bromobenzene	12.982	156	132381	47.296	ug/l	100
78) n-propylbenzene	13.035	91	705524	47.567	ug/l	100
79) 2-Chlorotoluene	13.124	91	419986	47.216	ug/l	100
80) 1,3,5-Trimethylbenzene	13.171	105	485499	48.181	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.735	75	79732	46.089	ug/l	100
82) 4-Chlorotoluene	13.224	91	424066	47.117	ug/l	100
83) tert-Butylbenzene	13.435	119	438417	47.531	ug/l	100
84) 1,2,4-Trimethylbenzene	13.482	105	488189	48.314	ug/l	100
85) sec-Butylbenzene	13.618	105	641523	47.893	ug/l	100
86) p-Isopropyltoluene	13.729	119	533366	48.169	ug/l	100
87) 1,3-Dichlorobenzene	13.735	146	260310	47.269	ug/l	100
88) 1,4-Dichlorobenzene	13.812	146	263358	46.907	ug/l	100
89) n-Butylbenzene	14.053	91	512519	47.787	ug/l	100
90) Hexachloroethane	14.335	117	89896	47.896	ug/l	100
91) 1,2-Dichlorobenzene	14.106	146	251235	47.486	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	14.718	75	45535	46.049	ug/l	100
93) 1,2,4-Trichlorobenzene	15.388	180	162374	48.061	ug/l	100
94) Hexachlorobutadiene	15.500	225	60866	48.356	ug/l	100
95) Naphthalene	15.641	128	603088	47.958	ug/l	100
96) 1,2,3-Trichlorobenzene	15.835	180	159113	47.402	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060625\
Data File : VN086865.D
Acq On : 06 Jun 2025 14:03
Operator : JC\MD
Sample : VSTDICCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Manual Integrations
APPROVED

Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025

Quant Time: Jun 07 01:57:17 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 01:51:02 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


```
Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060625\  
Data File : VN086865.D  
Acq On    : 06 Jun 2025  14:03  
Operator   : JC\MD  
Sample     : VSTDICCC050  
Misc      : 5.0mL/MSVOA_N/WATER  
ALS Vial   : 5      Sample Multiplier: 1
```


Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED

Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025


Quant Time: Jun 07 01:57:17 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 01:51:02 2025
Response via : Initial Calibration





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.236 min Scan# 1068
Delta R.T. -0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050



Tgt Ion:168 Resp: 207354
Ion Ratio Lower Upper
168 100
99 61.4 49.1 73.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#2
Dichlorodifluoromethane
Concen: 50.283 ug/l
RT: 2.154 min Scan# 34
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03


Tgt Ion: 85 Resp: 103958
Ion Ratio Lower Upper
85 100
87 32.4 16.2 48.6





#3
Chloromethane
Concen: 46.397 ug/l
RT: 2.395 min Scan# 75
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050



Tgt Ion: 50 Resp: 123860
Ion Ratio Lower Upper
50 100
52 32.1 25.7 38.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#4
Vinyl Chloride
Concen: 48.215 ug/l
RT: 2.554 min Scan# 102
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03


Tgt Ion: 62 Resp: 132783
Ion Ratio Lower Upper
62 100
64 32.5 26.0 39.0





#5
Bromomethane
Concen: 47.999 ug/l
RT: 2.995 min Scan# 177
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050



Tgt Ion: 94 Resp: 73974
Ion Ratio Lower Upper
94 100
96 97.4 77.9 116.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#6
Chloroethane
Concen: 47.504 ug/l
RT: 3.154 min Scan# 204
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03


Tgt Ion: 64 Resp: 84505
Ion Ratio Lower Upper
64 100
66 30.7 24.6 36.8





#7
Trichlorofluoromethane
Concen: 48.041 ug/l
RT: 3.524 min Scan# 267
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050



Tgt Ion:101 Resp: 172860
Ion Ratio Lower Upper
101 100
103 65.2 52.2 78.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#8
Diethyl Ether
Concen: 47.963 ug/l
RT: 3.983 min Scan# 345
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03

Tgt Ion: 74 Resp: 75193
Ion Ratio Lower Upper
74 100
45 91.0 45.5 136.5





#9

1,1,2-Trichlorotrifluoroethane

Concen: 47.640 ug/l

RT: 4.395 min Scan# 415

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050





Tgt Ion:	101	Resp:	10769
Ion	Ratio	Lower	Upper
101	100		
85	42.3	33.8	50.8
151	73.4	58.7	88.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#10

Methyl Iodide

Concen: 48.346 ug/l




RT: 4.606 min Scan# 451


Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03

Tgt Ion:	142	Resp:	141678
Ion	Ratio	Lower	Upper
142	100		
127	38.6	30.9	46.3
141	13.0	10.4	15.6





#11

Tert butyl alcohol

Concen: 242.020 ug/l

RT: 5.548 min Scan# 611

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion: 59 Resp: 18231

Ion Ratio Lower Upper

59 100





57 10.1 8.1 12.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#12

1,1-Dichloroethene

Concen: 47.900 ug/l

RT: 4.359 min Scan# 409

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03




Tgt Ion: 96 Resp: 110549


Ion Ratio Lower Upper

96 100

61 157.2 125.8 188.6

98 65.6 52.5 78.7





#13

Acrolein

Concen: 197.217 ug/l

RT: 4.201 min Scan# 382

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion: 56 Resp: 47020

Ion Ratio Lower Upper

56 100





55 69.5 55.6 83.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#14

Allyl chloride

Concen: 46.848 ug/l

RT: 5.042 min Scan# 525

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03




Tgt Ion: 41 Resp: 179313


Ion Ratio Lower Upper

41 100

39 64.1 51.3 76.9

76 39.2 31.4 47.0





#15

Acrylonitrile

Concen: 239.638 ug/l

RT: 5.742 min Scan# 644

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion: 53 Resp: 42194

Ion Ratio Lower Upper

53 100

52 81.3 65.0 97.6





51 34.0 27.2 40.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#16

Acetone

Concen: 226.765 ug/l

RT: 4.448 min Scan# 424

Delta R.T. 0.000 min

Lab File: VN086865.D




Acq: 06 Jun 2025 14:03


Tgt Ion: 43 Resp: 333857

Ion Ratio Lower Upper

43 100


58 35.0 28.0 42.0





#17
Carbon Disulfide
Concen: 46.328 ug/l
RT: 4.730 min Scan# 472
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050



Tgt Ion: 76 Resp: 29575
Ion Ratio Lower Upper
76 100
78 9.1 7.3 10.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#18
Methyl Acetate
Concen: 47.673 ug/l
RT: 5.042 min Scan# 525
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03

Tgt Ion: 43 Resp: 204538
Ion Ratio Lower Upper
43 100
74 25.9 20.7 31.1









#19
Methyl tert-butyl Ether
Concen: 47.960 ug/l
RT: 5.812 min Scan# 656
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Tgt Ion: 73 Resp: 400775
Ion Ratio Lower Upper
73 100
57 21.8 17.4 26.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#20
Methylene Chloride
Concen: 45.531 ug/l
RT: 5.295 min Scan# 568
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03


Tgt Ion: 84 Resp: 125498
Ion Ratio Lower Upper
84 100
49 113.1 90.5 135.7
51 35.6 28.5 42.7
86 60.1 48.1 72.1





#21
trans-1,2-Dichloroethene
Concen: 45.822 ug/l
RT: 5.806 min Scan# 611
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050



Tgt Ion: 96 Resp: 117660
Ion Ratio Lower Upper
96 100
61 136.5 109.2 163.8
98 62.3 49.8 74.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#22
Diisopropyl ether
Concen: 47.711 ug/l
RT: 6.689 min Scan# 805
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03


Tgt Ion: 45 Resp: 384939
Ion Ratio Lower Upper
45 100
43 52.4 41.9 62.9
87 28.0 22.4 33.6
59 12.2 9.8 14.6





#23
Vinyl Acetate
Concen: 241.910 ug/l
RT: 6.618 min Scan# 793
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050



Tgt Ion: 43 Resp: 164902
Ion Ratio Lower Upper
43 100
86 11.9 9.5 14.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#24
1,1-Dichloroethane
Concen: 47.485 ug/l
RT: 6.583 min Scan# 787
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03

Tgt Ion: 63 Resp: 220477
Ion Ratio Lower Upper
63 100
98 7.2 3.6 10.8
100 5.2 2.6 7.8









#25
2-Butanone
Concen: 238.693 ug/l
RT: 7.494 min Scan# 942
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Tgt Ion: 43 Resp: 57122
Ion Ratio Lower Upper
43 100
72 28.8 23.0 34.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#26
2,2-Dichloropropane
Concen: 51.933 ug/l
RT: 7.500 min Scan# 943
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03

Tgt Ion: 77 Resp: 187567
Ion Ratio Lower Upper
77 100
97 23.1 11.6 34.7





#27

cis-1,2-Dichloroethene

Concen: 47.189 ug/l

RT: 7.500 min Scan# 943

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03

Instrument :

MSVOA_N

ClientSampleId :





VSTDICCC050

Tgt Ion:	96	Resp:	14491
Ion	Ratio	Lower	Upper
96	100		
61	139.0	0.0	278.0
98	64.1	0.0	128.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#28

Bromochloromethane

Concen: 42.362 ug/l




RT: 7.824 min Scan# 998


Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03

Tgt Ion:	49	Resp:	96707
Ion	Ratio	Lower	Upper
49	100		
129	1.9	0.0	3.8
130	77.7	62.2	93.2





#29

Tetrahydrofuran

Concen: 238.210 ug/l

RT: 7.853 min Scan# 1003

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03

Instrument :

MSVOA_N

ClientSampleId :





VSTDICCC050

Tgt Ion:	42	Resp:	37135
Ion	Ratio	Lower	Upper
42	100		
72	47.2	37.8	56.6
71	44.5	35.6	53.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#30

Chloroform

Concen: 47.467 ug/l




RT: 7.977 min Scan# 1024


Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03

Tgt Ion:	83	Resp:	220093
Ion	Ratio	Lower	Upper
83	100		
85	66.1	52.9	79.3





#31

Cyclohexane

Concen: 46.217 ug/l

RT: 8.271 min Scan# 1074

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion: 56 Resp: 20810

Ion Ratio Lower Upper

56 100

69 31.7 25.4 38.0





84 86.0 68.8 103.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#32

1,1,1-Trichloroethane

Concen: 47.037 ug/l

RT: 8.177 min Scan# 1058

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03




Tgt Ion: 97 Resp: 185502


Ion Ratio Lower Upper

97 100

99 65.0 52.0 78.0

61 44.9 35.9 53.9





#33

1,2-Dichloroethane-d4

Concen: 37.366 ug/l

RT: 8.588 min Scan# 1128

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion: 65 Resp: 103730

Ion Ratio Lower Upper

65 100





67 52.8 0.0 105.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.106 min Scan# 1216

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03




Tgt Ion:114 Resp: 378587


Ion Ratio Lower Upper

114 100

63 19.8 0.0 39.6

88 15.1 0.0 30.2





#35

Dibromofluoromethane

Concen: 37.020 ug/l

RT: 8.171 min Scan# 1057

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion:113 Resp: 83058

Ion Ratio Lower Upper

113 100

111 105.2 84.2 126.2





192 17.8 14.2 21.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#36

1,1-Dichloropropene

Concen: 47.362 ug/l

RT: 8.377 min Scan# 1092

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03




Tgt Ion: 75 Resp: 158341


Ion Ratio Lower Upper

75 100

110 35.3 17.6 52.9

77 31.6 25.3 37.9





#37

Ethyl Acetate

Concen: 47.720 ug/l

RT: 7.577 min Scan# 956

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion: 43 Resp: 20309

Ion Ratio Lower Upper

43 100

61 15.1 12.1 18.1





70 12.2 9.8 14.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#38

Carbon Tetrachloride

Concen: 47.126 ug/l

RT: 8.371 min Scan# 1091

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03




Tgt Ion:117 Resp: 154677


Ion Ratio Lower Upper

117 100

119 93.8 75.0 112.6

121 31.1 24.9 37.3





#39

Methylcyclohexane

Concen: 47.126 ug/l

RT: 9.606 min Scan# 1130

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion: 83 Resp: 215849

Ion Ratio Lower Upper

83 100

55 76.3 61.0 91.6





98 48.6 38.9 58.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#40

Benzene

Concen: 46.590 ug/l

RT: 8.612 min Scan# 1132

Delta R.T. 0.000 min

Lab File: VN086865.D




Acq: 06 Jun 2025 14:03


Tgt Ion: 78 Resp: 509334

Ion Ratio Lower Upper

78 100


77 23.4 18.7 28.1





#41
Methacrylonitrile
Concen: 48.000 ug/l
RT: 7.789 min Scan# 992
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050



Tgt Ion: 41 Resp: 11472
Ion Ratio Lower Upper
41 100
39 51.5 41.2 61.8
67 76.8 61.4 92.2
52 28.4 22.7 34.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#42
1,2-Dichloroethane
Concen: 46.960 ug/l
RT: 8.677 min Scan# 1143
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03

Tgt Ion: 62 Resp: 155713
Ion Ratio Lower Upper
62 100
98 10.6 0.0 21.2





#43

Isopropyl Acetate

Concen: 47.248 ug/l

RT: 8.694 min Scan# 1146

Delta R.T. 0.000 min

Lab File: VN086865.D


Acq: 06 Jun 2025 14:03

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050





Tgt Ion:	43	Resp:	32274
Ion	Ratio	Lower	Upper
43	100		
61	27.6	22.1	33.1
87	14.7	11.8	17.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#44

Trichloroethene


Concen: 47.776 ug/l

RT: 9.359 min Scan# 1259



Delta R.T. 0.000 min


Lab File: VN086865.D

Acq: 06 Jun 2025 14:03



Tgt Ion:	130	Resp:	123847
Ion	Ratio	Lower	Upper
130	100		
95	98.1	0.0	196.2





#45

1,2-Dichloropropane

Concen: 47.251 ug/l

RT: 9.630 min Scan# 1305

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion: 63 Resp: 12567

Ion Ratio Lower Upper

63 100





65 31.9 25.5 38.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#46

Dibromomethane

Concen: 47.442 ug/l

RT: 9.718 min Scan# 1320

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03




Tgt Ion: 93 Resp: 83811


Ion Ratio Lower Upper

93 100

95 83.9 67.1 100.7

174 92.8 74.2 111.4





#47

Bromodichloromethane

Concen: 47.651 ug/l

RT: 9.894 min Scan# 1350

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050





Tgt Ion:	83	Resp:	173200
Ion Ratio	Lower	Upper	
83	100		
85	63.6	50.9	76.3
127	8.3	6.6	10.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#48

Methyl methacrylate

Concen: 47.399 ug/l




RT: 9.683 min Scan# 1314


Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03

Tgt Ion:	41	Resp:	149019
Ion Ratio	Lower	Upper	
41	100		
69	93.3	74.6	112.0
39	50.3	40.2	60.4





#49

1,4-Dioxane

Concen: 1032.607 ug/l

RT: 9.700 min Scan# 1317

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion: 88 Resp: 59060

Ion Ratio Lower Upper

88 100

43 0.0 0.0 0.0





58 72.1 57.7 86.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#50

Toluene-d8

Concen: 36.716 ug/l

RT: 10.571 min Scan# 1465

Delta R.T. 0.000 min

Lab File: VN086865.D




Acq: 06 Jun 2025 14:03


Tgt Ion: 98 Resp: 326105

Ion Ratio Lower Upper

98 100

100 66.7 53.4 80.0









#51
4-Methyl-2-Pentanone
Concen: 247.484 ug/l
RT: 10.447 min Scan# 1444
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Tgt Ion: 43 Resp: 101768
Ion Ratio Lower Upper
43 100
58 42.8 34.2 51.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#52
Toluene
Concen: 47.331 ug/l
RT: 10.635 min Scan# 1476
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03


Tgt Ion: 92 Resp: 316219
Ion Ratio Lower Upper
92 100
91 170.6 136.5 204.7





#53
t-1,3-Dichloropropene
Concen: 48.587 ug/l
RT: 10.841 min Scan# 1511
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050



Tgt Ion: 75 Resp: 197475
Ion Ratio Lower Upper
75 100
77 33.0 26.4 39.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#54
cis-1,3-Dichloropropene
Concen: 47.950 ug/l
RT: 10.318 min Scan# 1422
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03


Tgt Ion: 75 Resp: 208523
Ion Ratio Lower Upper
75 100
77 32.8 26.2 39.4
39 40.4 32.3 48.5





#55
1,1,2-Trichloroethane
Concen: 47.445 ug/l
RT: 11.018 min Scan# 1541
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050



Tgt Ion: 97 Resp: 121969
Ion Ratio Lower Upper
97 100
83 87.1 69.7 104.5
85 56.4 45.1 67.7
99 63.2 50.6 75.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#56
Ethyl methacrylate
Concen: 51.401 ug/l
RT: 10.882 min Scan# 1518
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03

Tgt Ion: 69 Resp: 210467
Ion Ratio Lower Upper
69 100
41 59.0 47.2 70.8
39 28.8 23.0 34.6









#57
1,3-Dichloropropane
Concen: 47.593 ug/l
RT: 11.165 min Scan# 1566
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Tgt Ion: 76 Resp: 21224
Ion Ratio Lower Upper
76 100
78 31.6 25.3 37.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#58
2-Chloroethyl Vinyl ether
Concen: 212.570 ug/l
RT: 10.165 min Scan# 1396
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03


Tgt Ion: 63 Resp: 519139
Ion Ratio Lower Upper
63 100
106 28.8 23.0 34.6





#59
2-Hexanone
Concen: 262.977 ug/l
RT: 11.200 min Scan# 1572
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050



Tgt Ion: 43 Resp: 69656
Ion Ratio Lower Upper
43 100
58 60.9 30.4 91.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#60
Dibromochloromethane
Concen: 48.056 ug/l
RT: 11.359 min Scan# 1599
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03


Tgt Ion:129 Resp: 128715
Ion Ratio Lower Upper
129 100
127 77.2 38.6 115.8





#61
1,2-Dibromoethane
Concen: 47.258 ug/l
RT: 11.471 min Scan# 1618
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050



Tgt Ion: 107 Resp: 12452
Ion Ratio Lower Upper
107 100
109 95.4 76.3 114.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#62
4-Bromofluorobenzene
Concen: 37.264 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03


Tgt Ion: 95 Resp: 122965
Ion Ratio Lower Upper
95 100
174 70.9 0.0 141.8
176 66.3 0.0 132.6





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 1685
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050



Tgt Ion:117 Resp: 332760
Ion Ratio Lower Upper
117 100
82 55.8 44.6 67.0
119 31.9 25.5 38.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#64
Tetrachloroethene
Concen: 46.382 ug/l
RT: 11.106 min Scan# 1556
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03

Tgt Ion:164 Resp: 97679
Ion Ratio Lower Upper
164 100
166 127.3 101.8 152.8
129 103.7 83.0 124.4
131 97.7 78.2 117.2









#65
Chlorobenzene
Concen: 46.386 ug/l
RT: 11.894 min Scan# 1690
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Tgt Ion:112 Resp: 340439
Ion Ratio Lower Upper
112 100
114 32.7 26.2 39.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#66
1,1,1,2-Tetrachloroethane
Concen: 47.615 ug/l
RT: 11.965 min Scan# 1702
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03


Tgt Ion:131 Resp: 112333
Ion Ratio Lower Upper
131 100
133 96.5 48.3 144.8
119 67.9 34.0 101.8





#67
Ethyl Benzene
Concen: 47.271 ug/l
RT: 11.965 min Scan# 1702
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050



Tgt Ion: 91 Resp: 597593
Ion Ratio Lower Upper
91 100
106 31.8 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#68
m/p-Xylenes
Concen: 96.424 ug/l
RT: 12.071 min Scan# 1720
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03

Tgt Ion:106 Resp: 466620
Ion Ratio Lower Upper
106 100
91 199.2 159.4 239.0





#69

o-Xylene

Concen: 48.395 ug/l

RT: 12.400 min Scan# 1776

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion:106 Resp: 224299

Ion Ratio Lower Upper

106 100





91 212.1 106.1 318.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#70

Styrene

Concen: 48.745 ug/l

RT: 12.412 min Scan# 1778

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03




Tgt Ion:104 Resp: 386596


Ion Ratio Lower Upper

104 100

78 48.1 38.5 57.7


103 53.9 43.1 64.7





#71
Bromoform
Concen: 50.415 ug/l
RT: 12.582 min Scan# 1807
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050



Tgt Ion:173 Resp: 8812
Ion Ratio Lower Upper
173 100
175 48.3 24.1 72.4
254 0.0 0.0 0.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.794 min Scan# 2013
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03


Tgt Ion:152 Resp: 167532
Ion Ratio Lower Upper
152 100
115 60.3 30.1 90.5
150 172.5 0.0 345.0





#73
Isopropylbenzene
Concen: 47.023 ug/l
RT: 12.694 min Scan# 1826
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050



Tgt Ion: 105 Resp: 57390
Ion Ratio Lower Upper
105 100
120 26.9 13.5 40.4




Manual Integrations
APPROVED


Reviewed By : John Carlone 06/09/2025
Supervised By : Mahesh Dadoda 06/09/2025



#74
N-ethyl acetate
Concen: 52.353 ug/l
RT: 12.512 min Scan# 1795
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03


Tgt Ion: 43 Resp: 212064
Ion Ratio Lower Upper
43 100
70 47.0 37.6 56.4
55 23.8 19.0 28.6
61 26.0 20.8 31.2





#75
1,1,2-Tetrachloroethane
Concen: 47.732 ug/l
RT: 12.941 min Scan# 1868
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050



Tgt Ion: 83 Resp: 197360
Ion Ratio Lower Upper
83 100
131 9.8 4.9 14.7
85 64.0 32.0 96.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#76
1,2,3-Trichloropropane
Concen: 49.183 ug/l m
RT: 12.994 min Scan# 1877
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03


Tgt Ion: 75 Resp: 196473
Ion Ratio Lower Upper
75 100
77 158.6 79.3 237.9





#77
Bromobenzene
Concen: 47.296 ug/l
RT: 12.982 min Scan# 1875
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050




Tgt Ion:156 Resp: 13238
Ion Ratio Lower Upper
156 100
77 235.4 117.7 353.1
158 96.3 48.1 144.4

Manual Integrations
APPROVED



Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025




#78
n-propylbenzene
Concen: 47.567 ug/l
RT: 13.035 min Scan# 1884
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03



Tgt Ion: 91 Resp: 705524
Ion Ratio Lower Upper
91 100
120 22.8 11.4 34.2









#79
2-Chlorotoluene
Concen: 47.216 ug/l
RT: 13.124 min Scan# 1899
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Tgt Ion: 91 Resp: 419980
Ion Ratio Lower Upper
91 100
126 32.3 16.2 48.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#80
1,3,5-Trimethylbenzene
Concen: 48.181 ug/l
RT: 13.171 min Scan# 1907
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03


Tgt Ion:105 Resp: 485499
Ion Ratio Lower Upper
105 100
120 49.7 24.9 74.6





#81
trans-1,4-Dichloro-2-butene
Concen: 46.089 ug/l
RT: 12.735 min Scan# 1833
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050



Tgt Ion: 75 Resp: 79732
Ion Ratio Lower Upper
75 100
53 83.2 66.6 99.8
89 45.9 36.7 55.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#82
4-Chlorotoluene
Concen: 47.117 ug/l
RT: 13.224 min Scan# 1916
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03

Tgt Ion: 91 Resp: 424066
Ion Ratio Lower Upper
91 100
126 33.4 16.7 50.1









#83
tert-Butylbenzene
Concen: 47.531 ug/l
RT: 13.435 min Scan# 1952
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Tgt Ion:119 Resp: 43841
Ion Ratio Lower Upper
119 100
91 64.3 32.1 96.5
134 26.7 13.4 40.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#84
1,2,4-Trimethylbenzene
Concen: 48.314 ug/l
RT: 13.482 min Scan# 1960
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03


Tgt Ion:105 Resp: 488189
Ion Ratio Lower Upper
105 100
120 46.4 23.2 69.6





#85
sec-Butylbenzene
Concen: 47.893 ug/l
RT: 13.618 min Scan# 1983
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050



Tgt Ion:105 Resp: 641523
Ion Ratio Lower Upper
105 100
134 20.0 10.0 30.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#86
p-Isopropyltoluene
Concen: 48.169 ug/l
RT: 13.729 min Scan# 2002
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03

Tgt Ion:119 Resp: 533366
Ion Ratio Lower Upper
119 100
134 26.9 13.5 40.4
91 22.9 11.5 34.4





#87

1,3-Dichlorobenzene

Concen: 47.269 ug/l

RT: 13.735 min Scan# 2003

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion:146 Resp: 260310

Ion Ratio Lower Upper

146 100

111 41.9 20.9 62.8





148 64.1 32.0 96.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#88

1,4-Dichlorobenzene

Concen: 46.907 ug/l

RT: 13.812 min Scan# 2016

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03




Tgt Ion:146 Resp: 263358


Ion Ratio Lower Upper

146 100

111 39.8 19.9 59.7

148 63.6 31.8 95.4





#89

n-Butylbenzene

Concen: 47.787 ug/l

RT: 14.053 min Scan# 2057

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion: 91 Resp: 512519

Ion Ratio Lower Upper

91 100

92 54.9 27.5 82.4





134 26.0 13.0 39.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#90

Hexachloroethane

Concen: 47.896 ug/l

RT: 14.335 min Scan# 2105

Delta R.T. 0.000 min

Lab File: VN086865.D




Acq: 06 Jun 2025 14:03


Tgt Ion:117 Resp: 89896

Ion Ratio Lower Upper

117 100

201 66.4 33.2 99.6









#91
1,2-Dichlorobenzene
Concen: 47.486 ug/l
RT: 14.106 min Scan# 2066
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Tgt Ion:146 Resp: 251235
Ion Ratio Lower Upper
146 100
111 43.1 21.6 64.6
148 63.4 31.7 95.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#92
1,2-Dibromo-3-Chloropropane
Concen: 46.049 ug/l
RT: 14.718 min Scan# 2170
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03

Tgt Ion: 75 Resp: 45535
Ion Ratio Lower Upper
75 100
155 82.3 41.1 123.4
157 105.4 52.7 158.1





#93

1,2,4-Trichlorobenzene

Concen: 48.061 ug/l

RT: 15.388 min Scan# 2284

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion:180 Resp: 162374

Ion Ratio Lower Upper

180 100

182 93.8 46.9 140.7





145 31.9 16.0 47.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#94

Hexachlorobutadiene

Concen: 48.356 ug/l

RT: 15.500 min Scan# 2303

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03




Tgt Ion:225 Resp: 60866


Ion Ratio Lower Upper

225 100

223 66.9 33.5 100.4


227 62.3 31.1 93.5





#95
Naphthalene
Concen: 47.958 ug/l
RT: 15.641 min Scan# 2327
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050



Tgt Ion:128 Resp: 603088
Ion Ratio Lower Upper
128 100
127 12.7 10.2 15.2
129 11.0 8.8 13.2




Manual Integrations
APPROVED

Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#96
1,2,3-Trichlorobenzene
Concen: 47.402 ug/l
RT: 15.835 min Scan# 2360
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03

Tgt Ion:180 Resp: 159113
Ion Ratio Lower Upper
180 100
182 94.0 47.0 141.0
145 33.2 16.6 49.8



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060625\
 Data File : VN086866.D
 Acq On : 06 Jun 2025 14:26
 Operator : JC\MD
 Sample : VSTDICC100
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC100

Manual Integrations
 APPROVED

Reviewed By : John Carlone 06/09/2025
 Supervised By : Mahesh Dadoda 06/09/2025

Quant Time: Jun 07 01:58:16 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 01:51:02 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	181996	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	327782	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	290313	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	147098	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	238957	98.065	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 196.120%#			
35) Dibromofluoromethane	8.177	113	195212	100.495	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 200.980%#			
50) Toluene-d8	10.570	98	772210	100.419	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 200.840%#			
62) 4-Bromofluorobenzene	12.847	95	292385	102.339	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 204.680%#			
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	2.153	85	194799	107.350	ug/l	100
3) Chloromethane	2.395	50	224558	95.833	ug/l	98
4) Vinyl Chloride	2.553	62	245135	101.414	ug/l	98
5) Bromomethane	2.977	94	137967	101.996	ug/l	99
6) Chloroethane	3.147	64	152150	97.447	ug/l	97
7) Trichlorofluoromethane	3.524	101	312394	98.917	ug/l	99
8) Diethyl Ether	3.983	74	139844	101.631	ug/l	98
9) 1,1,2-Trichlorotrifluo...	4.389	101	198739	100.163	ug/l	99
10) Methyl Iodide	4.612	142	262771	102.162	ug/l	99
11) Tert butyl alcohol	5.547	59	327987	496.062	ug/l	99
12) 1,1-Dichloroethene	4.359	96	200212	98.838	ug/l	99
13) Acrolein	4.200	56	94171	449.961	ug/l	98
14) Allyl chloride	5.047	41	329412	98.055	ug/l	99
15) Acrylonitrile	5.735	53	770470	498.552	ug/l	99
16) Acetone	4.447	43	608178	470.648	ug/l	98
17) Carbon Disulfide	4.736	76	544511	97.179	ug/l	100
18) Methyl Acetate	5.047	43	381808	101.390	ug/l	100
19) Methyl tert-butyl Ether	5.818	73	735694	100.306	ug/l	100
20) Methylene Chloride	5.294	84	228800	94.575	ug/l	97
21) trans-1,2-Dichloroethene	5.806	96	215177	95.476	ug/l	99
22) Diisopropyl ether	6.688	45	696881	98.409	ug/l	98
23) Vinyl Acetate	6.618	43	2919350	487.936	ug/l	100
24) 1,1-Dichloroethane	6.588	63	404207	99.186	ug/l	99
25) 2-Butanone	7.494	43	1043385	496.743	ug/l	100
26) 2,2-Dichloropropane	7.500	77	331995	104.729	ug/l	99
27) cis-1,2-Dichloroethene	7.500	96	265248	98.410	ug/l	99
28) Bromochloromethane	7.824	49	188050	93.851	ug/l	99
29) Tetrahydrofuran	7.847	42	677636	495.240	ug/l	99
30) Chloroform	7.977	83	394927	97.040	ug/l	99
31) Cyclohexane	8.265	56	374933	94.869	ug/l	99
32) 1,1,1-Trichloroethane	8.177	97	336819	97.306	ug/l	99
36) 1,1-Dichloropropene	8.382	75	289907	100.156	ug/l	99
37) Ethyl Acetate	7.571	43	370652	100.591	ug/l	99
38) Carbon Tetrachloride	8.371	117	285325	100.404	ug/l	98
39) Methylcyclohexane	9.606	83	395603	99.758	ug/l	99
40) Benzene	8.612	78	927083	97.946	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060625\
 Data File : VN086866.D
 Acq On : 06 Jun 2025 14:26
 Operator : JC\MD
 Sample : VSTDICC100
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025

Quant Time: Jun 07 01:58:16 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M

Quant Title : SW846 8260

QLast Update : Sat Jun 07 01:51:02 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.794	41	195946	94.688	ug/l	95
42) 1,2-Dichloroethane	8.676	62	282015	98.232	ug/l	100
43) Isopropyl Acetate	8.694	43	579464	97.981	ug/l	99
44) Trichloroethene	9.359	130	222822	99.280	ug/l	97
45) 1,2-Dichloropropane	9.629	63	230454	100.077	ug/l	98
46) Dibromomethane	9.712	93	155530	101.685	ug/l	99
47) Bromodichloromethane	9.894	83	316375	100.533	ug/l	100
48) Methyl methacrylate	9.682	41	275710	101.288	ug/l	99
49) 1,4-Dioxane	9.706	88	104771	2115.744	ug/l #	99
51) 4-Methyl-2-Pentanone	10.447	43	1843274	517.729	ug/l	99
52) Toluene	10.635	92	578978	100.092	ug/l	100
53) t-1,3-Dichloropropene	10.841	75	359200	102.077	ug/l	99
54) cis-1,3-Dichloropropene	10.318	75	383030	101.730	ug/l	99
55) 1,1,2-Trichloroethane	11.017	97	219922	98.808	ug/l	98
56) Ethyl methacrylate	10.882	69	389802	109.955	ug/l	99
57) 1,3-Dichloropropane	11.165	76	382510	99.069	ug/l	99
58) 2-Chloroethyl Vinyl ether	10.165	63	1113523	526.621	ug/l	100
59) 2-Hexanone	11.200	43	1301539	567.535	ug/l	99
60) Dibromochloromethane	11.359	129	237965	102.616	ug/l	100
61) 1,2-Dibromoethane	11.470	107	232214	101.787	ug/l	98
64) Tetrachloroethene	11.106	164	181568	98.824	ug/l	94
65) Chlorobenzene	11.894	112	632563	98.793	ug/l	98
66) 1,1,1,2-Tetrachloroethane	11.959	131	206895	100.521	ug/l	100
67) Ethyl Benzene	11.964	91	1110803	100.717	ug/l	99
68) m/p-Xylenes	12.070	106	854557	202.411	ug/l	99
69) o-Xylene	12.400	106	412719	102.070	ug/l	100
70) Styrene	12.412	104	713548	103.127	ug/l	100
71) Bromoform	12.576	173	165781	108.707	ug/l #	99
73) Isopropylbenzene	12.694	105	1065297	99.410	ug/l	100
74) N-amyl acetate	12.506	43	403587	113.475	ug/l	98
75) 1,1,2,2-Tetrachloroethane	12.941	83	354386	97.616	ug/l	99
76) 1,2,3-Trichloropropane	12.994	75	353334m	100.738	ug/l	
77) Bromobenzene	12.982	156	246413	100.266	ug/l	99
78) n-propylbenzene	13.035	91	1301631	99.948	ug/l	100
79) 2-Chlorotoluene	13.123	91	770308	98.630	ug/l	99
80) 1,3,5-Trimethylbenzene	13.170	105	901317	101.873	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.735	75	160736	105.820	ug/l	92
82) 4-Chlorotoluene	13.223	91	783657	99.166	ug/l	100
83) tert-Butylbenzene	13.435	119	807528	99.710	ug/l	100
84) 1,2,4-Trimethylbenzene	13.482	105	904794	101.983	ug/l	100
85) sec-Butylbenzene	13.617	105	1182155	100.513	ug/l	99
86) p-Isopropyltoluene	13.729	119	990273	101.857	ug/l	100
87) 1,3-Dichlorobenzene	13.735	146	474162	98.063	ug/l	100
88) 1,4-Dichlorobenzene	13.811	146	483003	97.978	ug/l	99
89) n-Butylbenzene	14.053	91	937913	99.598	ug/l	100
90) Hexachloroethane	14.335	117	169772	103.019	ug/l	99
91) 1,2-Dichlorobenzene	14.106	146	458049	98.602	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	14.717	75	83243	95.877	ug/l	96
93) 1,2,4-Trichlorobenzene	15.394	180	298870	100.751	ug/l	99
94) Hexachlorobutadiene	15.500	225	112496	101.789	ug/l	98
95) Naphthalene	15.641	128	1129434	102.290	ug/l	100
96) 1,2,3-Trichlorobenzene	15.841	180	294127	99.796	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060625\
Data File : VN086866.D
Acq On : 06 Jun 2025 14:26
Operator : JC\MD
Sample : VSTDICC100
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Manual Integrations
APPROVED

Reviewed By : John Carlone 06/09/2025
Supervised By : Mahesh Dadoda 06/09/2025

Quant Time: Jun 07 01:58:16 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 01:51:02 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060625\
Data File : VN086866.D
Acq On : 06 Jun 2025 14:26
Operator : JC\MD
Sample : VSTDICC100
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 6 Sample Multiplier: 1


Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED

Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025


Quant Time: Jun 07 01:58:16 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 01:51:02 2025
Response via : Initial Calibration





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.235 min Scan# 1068
Delta R.T. -0.001 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100



Tgt Ion:168 Resp: 181990
Ion Ratio Lower Upper
168 100
99 57.9 49.1 73.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#2
Dichlorodifluoromethane
Concen: 107.350 ug/l
RT: 2.153 min Scan# 34
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26


Tgt Ion: 85 Resp: 194799
Ion Ratio Lower Upper
85 100
87 32.3 16.2 48.6





#3
Chloromethane
Concen: 95.833 ug/l
RT: 2.395 min Scan# 75
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100



Tgt Ion: 50 Resp: 224558
Ion Ratio Lower Upper
50 100
52 33.4 25.7 38.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#4
Vinyl Chloride
Concen: 101.414 ug/l
RT: 2.553 min Scan# 102
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26


Tgt Ion: 62 Resp: 245135
Ion Ratio Lower Upper
62 100
64 31.5 26.0 39.0





#5
Bromomethane
Concen: 101.996 ug/l
RT: 2.977 min Scan# 11
Delta R.T. -0.018 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100



Tgt Ion: 94 Resp: 13796
Ion Ratio Lower Upper
94 100
96 96.8 77.9 116.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#6
Chloroethane
Concen: 97.447 ug/l
RT: 3.147 min Scan# 203
Delta R.T. -0.006 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26


Tgt Ion: 64 Resp: 152150
Ion Ratio Lower Upper
64 100
66 32.5 24.6 36.8





#7
Trichlorofluoromethane
Concen: 98.917 ug/l
RT: 3.524 min Scan# 267
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100



Tgt Ion:101 Resp: 312394
Ion Ratio Lower Upper
101 100
103 64.7 52.2 78.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#8
Diethyl Ether
Concen: 101.631 ug/l
RT: 3.983 min Scan# 345
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26

Tgt Ion: 74 Resp: 139844
Ion Ratio Lower Upper
74 100
45 89.5 45.5 136.5









#9
1,1,2-Trichlorotrifluoroethane
Concen: 100.163 ug/l
RT: 4.389 min Scan# 415
Delta R.T. -0.006 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Tgt Ion:101 Resp: 198739
Ion Ratio Lower Upper
101 100
85 42.2 33.8 50.8
151 74.2 58.7 88.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#10
Methyl Iodide
Concen: 102.162 ug/l
RT: 4.612 min Scan# 452
Delta R.T. 0.006 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26

Tgt Ion:142 Resp: 262771
Ion Ratio Lower Upper
142 100
127 37.9 30.9 46.3
141 13.6 10.4 15.6





#11

Tert butyl alcohol

Concen: 496.062 ug/l

RT: 5.547 min Scan# 611

Delta R.T. -0.000 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC100

Tgt Ion: 59 Resp: 32798

Ion Ratio Lower Upper

59 100





57 9.7 8.1 12.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#12

1,1-Dichloroethene

Concen: 98.838 ug/l

RT: 4.359 min Scan# 409

Delta R.T. -0.000 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26




Tgt Ion: 96 Resp: 200212


Ion Ratio Lower Upper

96 100

61 156.9 125.8 188.6

98 63.7 52.5 78.7





#13

Acrolein

Concen: 449.961 ug/l

RT: 4.200 min Scan# 382

Delta R.T. -0.000 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100

Tgt Ion: 56 Resp: 9417

Ion Ratio Lower Upper

56 100





55 70.9 55.6 83.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#14

Allyl chloride

Concen: 98.055 ug/l

RT: 5.047 min Scan# 526

Delta R.T. 0.006 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26




Tgt Ion: 41 Resp: 329412


Ion Ratio Lower Upper

41 100

39 63.7 51.3 76.9


76 39.7 31.4 47.0





#15
Acrylonitrile
Concen: 498.552 ug/l
RT: 5.735 min Scan# 644
Delta R.T. -0.006 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100



Tgt Ion: 53 Resp: 770470
Ion Ratio Lower Upper
53 100
52 82.2 65.0 97.6
51 34.4 27.2 40.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#16
Acetone
Concen: 470.648 ug/l
RT: 4.447 min Scan# 424
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26

Tgt Ion: 43 Resp: 608178
Ion Ratio Lower Upper
43 100
58 36.2 28.0 42.0









#17
Carbon Disulfide
Concen: 97.179 ug/l
RT: 4.736 min Scan# 473
Delta R.T. 0.006 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Tgt Ion: 76 Resp: 54451
Ion Ratio Lower Upper
76 100
78 9.1 7.3 10.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#18
Methyl Acetate
Concen: 101.390 ug/l
RT: 5.047 min Scan# 526
Delta R.T. 0.006 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26

Tgt Ion: 43 Resp: 381808
Ion Ratio Lower Upper
43 100
74 26.1 20.7 31.1





#19

Methyl tert-butyl Ether

Concen: 100.306 ug/l

RT: 5.818 min Scan# 61

Delta R.T. 0.006 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC100

Tgt Ion: 73 Resp: 735694

Ion Ratio Lower Upper

73 100





57 21.6 17.4 26.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#20

Methylene Chloride

Concen: 94.575 ug/l

RT: 5.294 min Scan# 568

Delta R.T. -0.000 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26

Tgt Ion: 84 Resp: 228800




Ion Ratio Lower Upper


84 100

49 111.4 90.5 135.7

51 34.2 28.5 42.7

86 63.5 48.1 72.1





#21

trans-1,2-Dichloroethene

Concen: 95.476 ug/l

RT: 5.806 min Scan# 61

Delta R.T. -0.000 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100

Tgt Ion: 96 Resp: 21517

Ion Ratio Lower Upper

96 100

61 134.4 109.2 163.8





98 62.5 49.8 74.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#22

Diisopropyl ether

Concen: 98.409 ug/l

RT: 6.688 min Scan# 805

Delta R.T. -0.000 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26

Tgt Ion: 45 Resp: 696881




Ion Ratio Lower Upper


45 100

43 51.1 41.9 62.9

87 29.6 22.4 33.6


59 12.2 9.8 14.6





#23
Vinyl Acetate
Concen: 487.936 ug/l
RT: 6.618 min Scan# 793
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100



Tgt Ion: 43 Resp: 2919350
Ion Ratio Lower Upper
43 100
86 12.1 9.5 14.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#24
1,1-Dichloroethane
Concen: 99.186 ug/l
RT: 6.588 min Scan# 788
Delta R.T. 0.006 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26


Tgt Ion: 63 Resp: 404207
Ion Ratio Lower Upper
63 100
98 7.3 3.6 10.8
100 4.7 2.6 7.8





#25
2-Butanone
Concen: 496.743 ug/l
RT: 7.494 min Scan# 942
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100



Tgt Ion: 43 Resp: 104338
Ion Ratio Lower Upper
43 100
72 28.5 23.0 34.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#26
2,2-Dichloropropane
Concen: 104.729 ug/l
RT: 7.500 min Scan# 943
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26

Tgt Ion: 77 Resp: 331995
Ion Ratio Lower Upper
77 100
97 23.4 11.6 34.7





#27

cis-1,2-Dichloroethene

Concen: 98.410 ug/l

RT: 7.500 min Scan# 943

Delta R.T. -0.000 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100

Tgt Ion: 96 Resp: 26524

Ion Ratio Lower Upper

96 100

61 137.9 0.0 278.0





98 64.0 0.0 128.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#28

Bromochloromethane

Concen: 93.851 ug/l

RT: 7.824 min Scan# 998

Delta R.T. -0.000 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26




Tgt Ion: 49 Resp: 188050


Ion Ratio Lower Upper

49 100

129 1.8 0.0 3.8

130 78.9 62.2 93.2





#29

Tetrahydrofuran

Concen: 495.240 ug/l

RT: 7.847 min Scan# 1002

Delta R.T. -0.006 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC100

Tgt Ion: 42 Resp: 677630

Ion Ratio Lower Upper

42 100

72 47.9 37.8 56.6





71 44.8 35.6 53.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#30

Chloroform

Concen: 97.040 ug/l

RT: 7.977 min Scan# 1024

Delta R.T. -0.000 min

Lab File: VN086866.D




Acq: 06 Jun 2025 14:26


Tgt Ion: 83 Resp: 394927

Ion Ratio Lower Upper

83 100


85 65.1 52.9 79.3





#31
Cyclohexane
Concen: 94.869 ug/l
RT: 8.265 min Scan# 1073
Delta R.T. -0.006 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100



Tgt Ion: 56 Resp: 37493
Ion Ratio Lower Upper
56 100
69 31.3 25.4 38.0
84 86.8 68.8 103.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#32
1,1,1-Trichloroethane
Concen: 97.306 ug/l
RT: 8.177 min Scan# 1058
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26

Tgt Ion: 97 Resp: 336819
Ion Ratio Lower Upper
97 100
99 65.7 52.0 78.0
61 44.2 35.9 53.9





#33

1,2-Dichloroethane-d4

Concen: 98.065 ug/l

RT: 8.588 min Scan# 1128

Delta R.T. -0.000 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC100

Tgt Ion: 65 Resp: 23895

Ion Ratio Lower Upper

65 100





67 54.8 0.0 105.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.106 min Scan# 1216

Delta R.T. -0.000 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26




Tgt Ion:114 Resp: 327782


Ion Ratio Lower Upper

114 100

63 19.7 0.0 39.6

88 15.4 0.0 30.2





#35

Dibromofluoromethane

Concen: 100.495 ug/l

RT: 8.177 min Scan# 1057

Delta R.T. 0.006 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC100

Tgt Ion:113 Resp: 19521

Ion Ratio Lower Upper

113 100

111 102.9 84.2 126.2





192 18.2 14.2 21.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#36

1,1-Dichloropropene

Concen: 100.156 ug/l

RT: 8.382 min Scan# 1093

Delta R.T. 0.006 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26




Tgt Ion: 75 Resp: 289907


Ion Ratio Lower Upper

75 100

110 34.9 17.6 52.9

77 31.1 25.3 37.9





#37

Ethyl Acetate

Concen: 100.591 ug/l

RT: 7.571 min Scan# 91

Delta R.T. -0.006 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC100

Tgt Ion: 43 Resp: 370652

Ion Ratio Lower Upper

43 100

61 15.3 12.1 18.1





70 12.6 9.8 14.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#38

Carbon Tetrachloride

Concen: 100.404 ug/l

RT: 8.371 min Scan# 1091

Delta R.T. -0.000 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26




Tgt Ion:117 Resp: 285325


Ion Ratio Lower Upper

117 100

119 95.4 75.0 112.6

121 28.6 24.9 37.3





#39

Methylcyclohexane

Concen: 99.758 ug/l

RT: 9.606 min Scan# 11301

Delta R.T. -0.000 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC100

Tgt Ion: 83 Resp: 39560

Ion Ratio Lower Upper

83 100

55 76.8 61.0 91.6





98 47.6 38.9 58.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#40

Benzene

Concen: 97.946 ug/l

RT: 8.612 min Scan# 1132

Delta R.T. -0.000 min

Lab File: VN086866.D




Acq: 06 Jun 2025 14:26


Tgt Ion: 78 Resp: 927083

Ion Ratio Lower Upper

78 100


77 23.6 18.7 28.1





#41
Methacrylonitrile
Concen: 94.688 ug/l
RT: 7.794 min Scan# 992
Delta R.T. 0.006 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100



Tgt Ion: 41 Resp: 195940
Ion Ratio Lower Upper
41 100
39 54.3 41.2 61.8
67 82.5 61.4 92.2
52 30.6 22.7 34.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#42
1,2-Dichloroethane
Concen: 98.232 ug/l
RT: 8.676 min Scan# 1143
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26

Tgt Ion: 62 Resp: 282015
Ion Ratio Lower Upper
62 100
98 10.6 0.0 21.2





#43

Isopropyl Acetate

Concen: 97.981 ug/l

RT: 8.694 min Scan# 1146

Delta R.T. -0.000 min

Lab File: VN086866.D


Acq: 06 Jun 2025 14:26

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100



Tgt Ion: 43 Resp: 57946

Ion Ratio Lower Upper

43 100

61 28.1 22.1 33.1




87 15.0 11.8 17.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#44

Trichloroethene

Concen: 99.280 ug/l

RT: 9.359 min Scan# 1259

Delta R.T. -0.000 min

Lab File: VN086866.D




Acq: 06 Jun 2025 14:26


Tgt Ion:130 Resp: 222822

Ion Ratio Lower Upper

130 100

95 101.4 0.0 196.2





#45

1,2-Dichloropropane

Concen: 100.077 ug/l

RT: 9.629 min Scan# 1319

Delta R.T. -0.000 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100

Tgt Ion: 63 Resp: 230454

Ion Ratio Lower Upper

63 100





65 30.6 25.5 38.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#46

Dibromomethane

Concen: 101.685 ug/l

RT: 9.712 min Scan# 1319

Delta R.T. -0.006 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26




Tgt Ion: 93 Resp: 155530


Ion Ratio Lower Upper

93 100

95 82.9 67.1 100.7


174 91.6 74.2 111.4





#47
Bromodichloromethane
Concen: 100.533 ug/l
RT: 9.894 min Scan# 1350
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100



Tgt Ion: 83 Resp: 31637
Ion Ratio Lower Upper
83 100
85 63.8 50.9 76.3
127 8.3 6.6 10.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#48
Methyl methacrylate
Concen: 101.288 ug/l
RT: 9.682 min Scan# 1314
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26

Tgt Ion: 41 Resp: 275710
Ion Ratio Lower Upper
41 100
69 92.4 74.6 112.0
39 50.1 40.2 60.4





#49

1,4-Dioxane

Concen: 2115.744 ug/l

RT: 9.706 min Scan# 11

Delta R.T. 0.006 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100

Tgt Ion: 88 Resp: 10477

Ion Ratio Lower Upper

88 100

43 0.0 0.0 0.0





58 71.5 57.7 86.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#50

Toluene-d8

Concen: 100.419 ug/l

RT: 10.570 min Scan# 1465

Delta R.T. -0.000 min

Lab File: VN086866.D




Acq: 06 Jun 2025 14:26


Tgt Ion: 98 Resp: 772210

Ion Ratio Lower Upper

98 100


100 67.6 53.4 80.0





#51
4-Methyl-2-Pentanone
Concen: 517.729 ug/l
RT: 10.447 min Scan# 1444
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100



Tgt Ion: 43 Resp: 1843274
Ion Ratio Lower Upper
43 100
58 43.6 34.2 51.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#52
Toluene
Concen: 100.092 ug/l
RT: 10.635 min Scan# 1476
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26


Tgt Ion: 92 Resp: 578978
Ion Ratio Lower Upper
92 100
91 170.6 136.5 204.7





#53
t-1,3-Dichloropropene
Concen: 102.077 ug/l
RT: 10.841 min Scan# 1511
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100



Tgt Ion: 75 Resp: 359200
Ion Ratio Lower Upper
75 100
77 32.3 26.4 39.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#54
cis-1,3-Dichloropropene
Concen: 101.730 ug/l
RT: 10.318 min Scan# 1422
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26


Tgt Ion: 75 Resp: 383030
Ion Ratio Lower Upper
75 100
77 32.3 26.2 39.4
39 40.1 32.3 48.5





#55
1,1,2-Trichloroethane
Concen: 98.808 ug/l
RT: 11.017 min Scan# 1541
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100



Tgt Ion: 97 Resp: 21992
Ion Ratio Lower Upper
97 100
83 84.5 69.7 104.5
85 55.6 45.1 67.7
99 61.9 50.6 75.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#56
Ethyl methacrylate
Concen: 109.955 ug/l
RT: 10.882 min Scan# 1518
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26


Tgt Ion: 69 Resp: 389802
Ion Ratio Lower Upper
69 100
41 58.3 47.2 70.8
39 27.5 23.0 34.6





#57
1,3-Dichloropropane
Concen: 99.069 ug/l
RT: 11.165 min Scan# 1566
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100



Tgt Ion: 76 Resp: 382510
Ion Ratio Lower Upper
76 100
78 32.2 25.3 37.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#58
2-Chloroethyl Vinyl ether
Concen: 526.621 ug/l
RT: 10.165 min Scan# 1396
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26


Tgt Ion: 63 Resp: 1113523
Ion Ratio Lower Upper
63 100
106 29.1 23.0 34.6





#59
2-Hexanone
Concen: 567.535 ug/l
RT: 11.200 min Scan# 1572
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100



Tgt Ion: 43 Resp: 1301539
Ion Ratio Lower Upper
43 100
58 61.4 30.4 91.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#60
Dibromochloromethane
Concen: 102.616 ug/l
RT: 11.359 min Scan# 1599
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26


Tgt Ion:129 Resp: 237965
Ion Ratio Lower Upper
129 100
127 77.1 38.6 115.8





#61
1,2-Dibromoethane
Concen: 101.787 ug/l
RT: 11.470 min Scan# 1618
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100



Tgt Ion:107 Resp: 232214
Ion Ratio Lower Upper
107 100
109 93.8 76.3 114.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#62
4-Bromofluorobenzene
Concen: 102.339 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26

Tgt Ion: 95 Resp: 292385
Ion Ratio Lower Upper
95 100
174 71.6 0.0 141.8
176 68.6 0.0 132.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.870 min Scan# 1686
Delta R.T. 0.006 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Tgt Ion:117 Resp: 29031
Ion Ratio Lower Upper
117 100
82 54.3 44.6 67.0
119 31.9 25.5 38.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#64
Tetrachloroethene
Concen: 98.824 ug/l
RT: 11.106 min Scan# 1556
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26

Tgt Ion:164 Resp: 181568
Ion Ratio Lower Upper
164 100
166 121.8 101.8 152.8
129 96.8 83.0 124.4
131 91.5 78.2 117.2









#65
Chlorobenzene
Concen: 98.793 ug/l
RT: 11.894 min Scan# 11
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Tgt Ion:112 Resp: 63256
Ion Ratio Lower Upper
112 100
114 31.3 26.2 39.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#66
1,1,1,2-Tetrachloroethane
Concen: 100.521 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. -0.006 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26

Tgt Ion:131 Resp: 206895
Ion Ratio Lower Upper
131 100
133 96.4 48.3 144.8
119 68.2 34.0 101.8









#67
Ethyl Benzene
Concen: 100.717 ug/l
RT: 11.964 min Scan# 1702
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Tgt Ion: 91 Resp: 111080
Ion Ratio Lower Upper
91 100
106 31.5 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#68
m/p-Xylenes
Concen: 202.411 ug/l
RT: 12.070 min Scan# 1720
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26

Tgt Ion:106 Resp: 854557
Ion Ratio Lower Upper
106 100
91 200.4 159.4 239.0









#69
o-Xylene
Concen: 102.070 ug/l
RT: 12.400 min Scan# 1776
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Tgt Ion:106 Resp: 412719
Ion Ratio Lower Upper
106 100
91 212.8 106.1 318.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#70
Styrene
Concen: 103.127 ug/l
RT: 12.412 min Scan# 1778
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26


Tgt Ion:104 Resp: 713548
Ion Ratio Lower Upper
104 100
78 48.4 38.5 57.7
103 54.0 43.1 64.7





#71
Bromoform
Concen: 108.707 ug/l
RT: 12.576 min Scan# 1806
Delta R.T. -0.006 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100



Tgt Ion:173 Resp: 16578
Ion Ratio Lower Upper
173 100
175 47.9 24.1 72.4
254 0.0 0.0 0.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.794 min Scan# 2013
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26


Tgt Ion:152 Resp: 147098
Ion Ratio Lower Upper
152 100
115 60.1 30.1 90.5
150 190.4 0.0 345.0





#73
Isopropylbenzene
Concen: 99.410 ug/l
RT: 12.694 min Scan# 1826
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100



Tgt Ion:105 Resp: 106529
Ion Ratio Lower Upper
105 100
120 27.0 13.5 40.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#74
N-ethyl acetate
Concen: 113.475 ug/l
RT: 12.506 min Scan# 1794
Delta R.T. -0.006 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26


Tgt Ion: 43 Resp: 403587
Ion Ratio Lower Upper
43 100
70 47.7 37.6 56.4
55 25.0 19.0 28.6
61 27.3 20.8 31.2





#75
1,1,2,2-Tetrachloroethane
Concen: 97.616 ug/l
RT: 12.941 min Scan# 1868
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100



Tgt Ion: 83 Resp: 354380
Ion Ratio Lower Upper
83 100
131 9.9 4.9 14.7
85 64.9 32.0 96.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#76
1,2,3-Trichloropropane
Concen: 100.738 ug/l m
RT: 12.994 min Scan# 1877
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26

Tgt Ion: 75 Resp: 353334
Ion Ratio Lower Upper
75 100
77 162.2 79.3 237.9









#77
Bromobenzene
Concen: 100.266 ug/l
RT: 12.982 min Scan# 1875
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Tgt Ion:156 Resp: 24641
Ion Ratio Lower Upper
156 100
77 232.6 117.7 353.1
158 96.4 48.1 144.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#78
n-propylbenzene
Concen: 99.948 ug/l
RT: 13.035 min Scan# 1884
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26

Tgt Ion: 91 Resp: 1301631
Ion Ratio Lower Upper
91 100
120 22.9 11.4 34.2









#79
2-Chlorotoluene
Concen: 98.630 ug/l
RT: 13.123 min Scan# 1899
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Tgt Ion: 91 Resp: 770308
Ion Ratio Lower Upper
91 100
126 32.8 16.2 48.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#80
1,3,5-Trimethylbenzene
Concen: 101.873 ug/l
RT: 13.170 min Scan# 1907
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26


Tgt Ion:105 Resp: 901317
Ion Ratio Lower Upper
105 100
120 49.7 24.9 74.6





#81
trans-1,4-Dichloro-2-butene
Concen: 105.820 ug/l
RT: 12.735 min Scan# 1833
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100



Tgt Ion: 75 Resp: 160730
Ion Ratio Lower Upper
75 100
53 76.0 66.6 99.8
89 41.1 36.7 55.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#82
4-Chlorotoluene
Concen: 99.166 ug/l
RT: 13.223 min Scan# 1916
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26

Tgt Ion: 91 Resp: 783657
Ion Ratio Lower Upper
91 100
126 33.2 16.7 50.1









#83
tert-Butylbenzene
Concen: 99.710 ug/l
RT: 13.435 min Scan# 1952
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Tgt Ion:119 Resp: 807528
Ion Ratio Lower Upper
119 100
91 64.6 32.1 96.5
134 26.6 13.4 40.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#84
1,2,4-Trimethylbenzene
Concen: 101.983 ug/l
RT: 13.482 min Scan# 1960
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26


Tgt Ion:105 Resp: 904794
Ion Ratio Lower Upper
105 100
120 46.7 23.2 69.6





#85
sec-Butylbenzene
Concen: 100.513 ug/l
RT: 13.617 min Scan# 1983
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100



Tgt Ion:105 Resp: 118215
Ion Ratio Lower Upper
105 100
134 20.2 10.0 30.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#86
p-Isopropyltoluene
Concen: 101.857 ug/l
RT: 13.729 min Scan# 2002
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26

Tgt Ion:119 Resp: 990273
Ion Ratio Lower Upper
119 100
134 26.7 13.5 40.4
91 22.6 11.5 34.4





#87

1,3-Dichlorobenzene

Concen: 98.063 ug/l

RT: 13.735 min Scan# 2003

Delta R.T. -0.000 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC100

Tgt Ion:146 Resp: 474162

Ion Ratio Lower Upper

146 100

111 42.3 20.9 62.8





148 64.4 32.0 96.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#88

1,4-Dichlorobenzene

Concen: 97.978 ug/l

RT: 13.811 min Scan# 2016

Delta R.T. -0.000 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26




Tgt Ion:146 Resp: 483003


Ion Ratio Lower Upper

146 100

111 39.3 19.9 59.7

148 63.0 31.8 95.4





#89

n-Butylbenzene

Concen: 99.598 ug/l

RT: 14.053 min Scan# 2057

Delta R.T. -0.000 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC100





Tgt Ion:	91	Resp:	93791
Ion	Ratio	Lower	Upper
91	100		
92	54.7	27.5	82.4
134	26.2	13.0	39.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#90

Hexachloroethane

Concen: 103.019 ug/l




RT: 14.335 min Scan# 2105


Delta R.T. -0.000 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26

Tgt Ion:	117	Resp:	169772
Ion	Ratio	Lower	Upper
117	100		
201	65.8	33.2	99.6





#91

1,2-Dichlorobenzene

Concen: 98.602 ug/l

RT: 14.106 min Scan# 2066

Delta R.T. -0.000 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC100

Tgt Ion:146 Resp: 458049

Ion Ratio Lower Upper

146 100

111 43.5 21.6 64.6





148 63.5 31.7 95.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#92

1,2-Dibromo-3-Chloropropane

Concen: 95.877 ug/l

RT: 14.717 min Scan# 2170

Delta R.T. -0.000 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26




Tgt Ion: 75 Resp: 83243


Ion Ratio Lower Upper

75 100

155 86.1 41.1 123.4

157 109.8 52.7 158.1





#93

1,2,4-Trichlorobenzene

Concen: 100.751 ug/l

RT: 15.394 min Scan# 2284

Delta R.T. 0.006 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC100

Tgt Ion:180 Resp: 298870

Ion Ratio Lower Upper

180 100

182 94.6 46.9 140.7





145 32.2 16.0 47.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#94

Hexachlorobutadiene

Concen: 101.789 ug/l

RT: 15.500 min Scan# 2303

Delta R.T. -0.000 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26




Tgt Ion:225 Resp: 112496


Ion Ratio Lower Upper

225 100

223 64.2 33.5 100.4


227 62.1 31.1 93.5





#95
Naphthalene
Concen: 102.290 ug/l
RT: 15.641 min Scan# 2327
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100




Tgt Ion:128 Resp: 1129434
Ion Ratio Lower Upper
128 100
127 12.6 10.2 15.2
129 11.0 8.8 13.2

Manual Integrations
APPROVED



Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#96
1,2,3-Trichlorobenzene
Concen: 99.796 ug/l
RT: 15.841 min Scan# 2361
Delta R.T. 0.006 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26



Tgt Ion:180 Resp: 294127
Ion Ratio Lower Upper
180 100
182 95.3 47.0 141.0
145 33.5 16.6 49.8



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060625\
 Data File : VN086867.D
 Acq On : 06 Jun 2025 14:49
 Operator : JC\MD
 Sample : VSTDICC150
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC150

Manual Integrations
 APPROVED

Reviewed By : John Carlone 06/09/2025
 Supervised By : Mahesh Dadoda 06/09/2025

Quant Time: Jun 07 01:59:15 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 01:51:02 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	172832	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	307937	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.864	117	278552	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	136630	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	389297	168.233	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 336.460%#			
35) Dibromofluoromethane	8.176	113	324474	177.803	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 355.600%#			
50) Toluene-d8	10.570	98	1272348	176.120	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 352.240%#			
62) 4-Bromofluorobenzene	12.847	95	481421	179.364	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 358.720%#			
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	2.153	85	262143	152.122	ug/l	100
3) Chloromethane	2.394	50	304558	136.865	ug/l	100
4) Vinyl Chloride	2.553	62	335834	146.303	ug/l	100
5) Bromomethane	2.965	94	190786	148.522	ug/l	98
6) Chloroethane	3.136	64	208568	140.664	ug/l	96
7) Trichlorofluoromethane	3.518	101	427871	142.666	ug/l	100
8) Diethyl Ether	3.983	74	191555	146.593	ug/l	99
9) 1,1,2-Trichlorotrifluo...	4.388	101	269810	143.192	ug/l	100
10) Methyl Iodide	4.606	142	352020	144.118	ug/l	99
11) Tert butyl alcohol	5.553	59	430866	686.213	ug/l	100
12) 1,1-Dichloroethene	4.353	96	273388	142.118	ug/l	98
13) Acrolein	4.194	56	171625	863.526	ug/l	99
14) Allyl chloride	5.035	41	490852	153.858	ug/l	92
15) Acrylonitrile	5.735	53	1048372	714.344	ug/l	99
16) Acetone	4.447	43	819041	667.435	ug/l	100
17) Carbon Disulfide	4.730	76	742788	139.594	ug/l	100
18) Methyl Acetate	5.041	43	523996	146.527	ug/l	100
19) Methyl tert-butyl Ether	5.818	73	998781	143.396	ug/l	99
20) Methylene Chloride	5.300	84	311597	135.628	ug/l	99
21) trans-1,2-Dichloroethene	5.806	96	290960	135.947	ug/l	97
22) Diisopropyl ether	6.688	45	947638	140.915	ug/l	98
23) Vinyl Acetate	6.618	43	3932872	692.189	ug/l	99
24) 1,1-Dichloroethane	6.582	63	540881	139.761	ug/l	99
25) 2-Butanone	7.494	43	1381940	692.809	ug/l	99
26) 2,2-Dichloropropane	7.500	77	450112	149.519	ug/l	99
27) cis-1,2-Dichloroethene	7.500	96	363398	141.974	ug/l	99
28) Bromochloromethane	7.824	49	290594	152.718	ug/l	98
29) Tetrahydrofuran	7.847	42	898159	691.211	ug/l	98
30) Chloroform	7.976	83	534038	138.179	ug/l	97
31) Cyclohexane	8.265	56	505795	134.766	ug/l	96
32) 1,1,1-Trichloroethane	8.176	97	462854	140.807	ug/l	99
36) 1,1-Dichloropropene	8.376	75	393166	144.584	ug/l	100
37) Ethyl Acetate	7.571	43	494000	142.706	ug/l	99
38) Carbon Tetrachloride	8.371	117	388595	145.557	ug/l	98
39) Methylcyclohexane	9.606	83	544437	146.136	ug/l	97
40) Benzene	8.612	78	1266139	142.388	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060625\
 Data File : VN086867.D
 Acq On : 06 Jun 2025 14:49
 Operator : JC\MD
 Sample : VSTDICC150
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC150

Manual Integrations
 APPROVED

Reviewed By : John Carlone 06/09/2025
 Supervised By : Mahesh Dadoda 06/09/2025

Quant Time: Jun 07 01:59:15 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 01:51:02 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.788	41	276400	142.174	ug/l	98
42) 1,2-Dichloroethane	8.676	62	381516	141.455	ug/l	100
43) Isopropyl Acetate	8.694	43	780791	140.531	ug/l	99
44) Trichloroethene	9.359	130	302548	143.490	ug/l	96
45) 1,2-Dichloropropane	9.623	63	309047	142.856	ug/l	100
46) Dibromomethane	9.712	93	208461	145.074	ug/l	100
47) Bromodichloromethane	9.894	83	430011	145.449	ug/l	100
48) Methyl methacrylate	9.688	41	366967	143.502	ug/l	99
49) 1,4-Dioxane	9.712	88	138035	2967.116	ug/l #	100
51) 4-Methyl-2-Pentanone	10.447	43	2438311	728.995	ug/l	98
52) Toluene	10.635	92	793270	145.976	ug/l	100
53) t-1,3-Dichloropropene	10.841	75	489809	148.163	ug/l	97
54) cis-1,3-Dichloropropene	10.318	75	518092	146.469	ug/l	99
55) 1,1,2-Trichloroethane	11.017	97	298151	142.588	ug/l	99
56) Ethyl methacrylate	10.876	69	533912	160.311	ug/l	98
57) 1,3-Dichloropropane	11.165	76	516391	142.363	ug/l	98
58) 2-Chloroethyl Vinyl ether	10.165	63	1869543	941.147	ug/l	99
59) 2-Hexanone	11.200	43	1735777	805.662	ug/l	99
60) Dibromochloromethane	11.365	129	322064	147.832	ug/l	98
61) 1,2-Dibromoethane	11.470	107	314606	146.790	ug/l	99
64) Tetrachloroethene	11.106	164	245122	139.048	ug/l	97
65) Chlorobenzene	11.894	112	860345	140.041	ug/l	98
66) 1,1,1,2-Tetrachloroethane	11.959	131	282739	143.171	ug/l	99
67) Ethyl Benzene	11.964	91	1517883	143.438	ug/l	98
68) m/p-Xylenes	12.070	106	1175378	290.155	ug/l	99
69) o-Xylene	12.400	106	566301	145.966	ug/l	100
70) Styrene	12.411	104	972318	146.459	ug/l	100
71) Bromoform	12.582	173	223200	152.538	ug/l #	99
73) Isopropylbenzene	12.694	105	1453651	146.042	ug/l	100
74) N-amyl acetate	12.500	43	543747	164.596	ug/l	96
75) 1,1,2,2-Tetrachloroethane	12.941	83	474426	140.693	ug/l	100
76) 1,2,3-Trichloropropane	12.994	75	468041m	143.665	ug/l	
77) Bromobenzene	12.982	156	335734	147.077	ug/l	98
78) n-propylbenzene	13.035	91	1769694	146.301	ug/l	99
79) 2-Chlorotoluene	13.123	91	1046916	144.317	ug/l	99
80) 1,3,5-Trimethylbenzene	13.170	105	1210325	147.279	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.735	75	216253	153.277	ug/l	95
82) 4-Chlorotoluene	13.223	91	1063581	144.900	ug/l	100
83) tert-Butylbenzene	13.435	119	1093406	145.352	ug/l	100
84) 1,2,4-Trimethylbenzene	13.482	105	1216529	147.626	ug/l	100
85) sec-Butylbenzene	13.617	105	1582762	144.885	ug/l	99
86) p-Isopropyltoluene	13.729	119	1322780	146.482	ug/l	99
87) 1,3-Dichlorobenzene	13.735	146	641966	142.939	ug/l	100
88) 1,4-Dichlorobenzene	13.811	146	646188	141.124	ug/l	99
89) n-Butylbenzene	14.053	91	1251819	143.117	ug/l	100
90) Hexachloroethane	14.335	117	230195	150.387	ug/l	98
91) 1,2-Dichlorobenzene	14.105	146	613177	142.108	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	14.717	75	110677	137.242	ug/l	96
93) 1,2,4-Trichlorobenzene	15.394	180	407621	147.940	ug/l	99
94) Hexachlorobutadiene	15.500	225	151172	147.264	ug/l	98
95) Naphthalene	15.641	128	1545981	150.743	ug/l	100
96) 1,2,3-Trichlorobenzene	15.841	180	404450	147.743	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060625\
Data File : VN086867.D
Acq On : 06 Jun 2025 14:49
Operator : JC\MD
Sample : VSTDICC150
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC150

Manual Integrations
APPROVED

Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025

Quant Time: Jun 07 01:59:15 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 01:51:02 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC150

Manual Integrations APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.229 min Scan# 1068
Delta R.T. -0.007 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC150



Tgt Ion:168 Resp: 17283
Ion Ratio Lower Upper
168 100
99 61.8 49.1 73.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#2
Dichlorodifluoromethane
Concen: 152.122 ug/l
RT: 2.153 min Scan# 34
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49


Tgt Ion: 85 Resp: 262143
Ion Ratio Lower Upper
85 100
87 32.3 16.2 48.6





#3
Chloromethane
Concen: 136.865 ug/l
RT: 2.394 min Scan# 75
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC150



Tgt Ion: 50 Resp: 304558
Ion Ratio Lower Upper
50 100
52 32.2 25.7 38.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#4
Vinyl Chloride
Concen: 146.303 ug/l
RT: 2.553 min Scan# 102
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49

Tgt Ion: 62 Resp: 335834
Ion Ratio Lower Upper
62 100
64 32.6 26.0 39.0





#5

Bromomethane

Concen: 148.522 ug/l

RT: 2.965 min Scan# 11

Delta R.T. -0.030 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC150

Tgt Ion: 94 Resp: 190780

Ion Ratio Lower Upper

94 100





96 95.5 77.9 116.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#6

Chloroethane

Concen: 140.664 ug/l

RT: 3.136 min Scan# 201

Delta R.T. -0.018 min

Lab File: VN086867.D




Acq: 06 Jun 2025 14:49


Tgt Ion: 64 Resp: 208568

Ion Ratio Lower Upper

64 100


66 32.7 24.6 36.8





#7
Trichlorofluoromethane
Concen: 142.666 ug/l
RT: 3.518 min Scan# 266
Delta R.T. -0.006 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150



Tgt Ion:101 Resp: 42787
Ion Ratio Lower Upper
101 100
103 65.5 52.2 78.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#8
Diethyl Ether
Concen: 146.593 ug/l
RT: 3.983 min Scan# 345
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49

Tgt Ion: 74 Resp: 191555
Ion Ratio Lower Upper
74 100
45 90.3 45.5 136.5









#9
1,1,2-Trichlorotrifluoroethane
Concen: 143.192 ug/l
RT: 4.388 min Scan# 415
Delta R.T. -0.006 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC150

Tgt Ion:101 Resp: 269810
Ion Ratio Lower Upper
101 100
85 42.3 33.8 50.8
151 73.5 58.7 88.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#10
Methyl Iodide
Concen: 144.118 ug/l
RT: 4.606 min Scan# 451
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49

Tgt Ion:142 Resp: 352020
Ion Ratio Lower Upper
142 100
127 39.1 30.9 46.3
141 13.5 10.4 15.6





#11

Tert butyl alcohol

Concen: 686.213 ug/l

RT: 5.553 min Scan# 611

Delta R.T. 0.006 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC150

Tgt Ion: 59 Resp: 43086

Ion Ratio Lower Upper

59 100





57 9.9 8.1 12.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#12

1,1-Dichloroethene

Concen: 142.118 ug/l

RT: 4.353 min Scan# 408

Delta R.T. -0.006 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49




Tgt Ion: 96 Resp: 273388


Ion Ratio Lower Upper

96 100

61 154.5 125.8 188.6

98 64.9 52.5 78.7





#13

Acrolein

Concen: 863.526 ug/l

RT: 4.194 min Scan# 381

Delta R.T. -0.006 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC150

Tgt Ion: 56 Resp: 17162

Ion Ratio Lower Upper

56 100





55 70.4 55.6 83.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#14

Allyl chloride

Concen: 153.858 ug/l

RT: 5.035 min Scan# 524

Delta R.T. -0.006 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49




Tgt Ion: 41 Resp: 490852


Ion Ratio Lower Upper

41 100

39 56.4 51.3 76.9


76 35.7 31.4 47.0





#15
Acrylonitrile
Concen: 714.344 ug/l
RT: 5.735 min Scan# 644
Delta R.T. -0.006 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC150



Tgt Ion: 53 Resp: 1048372
Ion Ratio Lower Upper
53 100
52 80.7 65.0 97.6
51 33.8 27.2 40.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#16
Acetone
Concen: 667.435 ug/l
RT: 4.447 min Scan# 424
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49

Tgt Ion: 43 Resp: 819041
Ion Ratio Lower Upper
43 100
58 34.7 28.0 42.0









#17
Carbon Disulfide
Concen: 139.594 ug/l
RT: 4.730 min Scan# 472
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC150

Tgt Ion: 76 Resp: 742788
Ion Ratio Lower Upper
76 100
78 9.3 7.3 10.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#18
Methyl Acetate
Concen: 146.527 ug/l
RT: 5.041 min Scan# 525
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49


Tgt Ion: 43 Resp: 523996
Ion Ratio Lower Upper
43 100
74 26.1 20.7 31.1





#19
Methyl tert-butyl Ether
Concen: 143.396 ug/l
RT: 5.818 min Scan# 656
Delta R.T. 0.006 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC150



Tgt Ion: 73 Resp: 99878
Ion Ratio Lower Upper
73 100
57 22.2 17.4 26.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#20
Methylene Chloride
Concen: 135.628 ug/l
RT: 5.300 min Scan# 569
Delta R.T. 0.006 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49


Tgt Ion: 84 Resp: 311597
Ion Ratio Lower Upper
84 100
49 113.0 90.5 135.7
51 36.1 28.5 42.7
86 62.3 48.1 72.1





#21
trans-1,2-Dichloroethene
Concen: 135.947 ug/l
RT: 5.806 min Scan# 61
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC150



Tgt Ion: 96 Resp: 290960
Ion Ratio Lower Upper
96 100
61 132.7 109.2 163.8
98 65.0 49.8 74.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#22
Diisopropyl ether
Concen: 140.915 ug/l
RT: 6.688 min Scan# 805
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49


Tgt Ion: 45 Resp: 947638
Ion Ratio Lower Upper
45 100
43 53.1 41.9 62.9
87 29.9 22.4 33.6
59 13.2 9.8 14.6





#23
Vinyl Acetate
Concen: 692.189 ug/l
RT: 6.618 min Scan# 793
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC150



Tgt Ion: 43 Resp: 393287
Ion Ratio Lower Upper
43 100
86 12.1 9.5 14.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#24
1,1-Dichloroethane
Concen: 139.761 ug/l
RT: 6.582 min Scan# 787
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49

Tgt Ion: 63 Resp: 540881
Ion Ratio Lower Upper
63 100
98 7.3 3.6 10.8
100 4.4 2.6 7.8





#25

2-Butanone

Concen: 692.809 ug/l

RT: 7.494 min Scan# 942

Delta R.T. -0.000 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC150

Tgt Ion: 43 Resp: 1381940

Ion Ratio Lower Upper

43 100





72 29.4 23.0 34.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#26

2,2-Dichloropropane

Concen: 149.519 ug/l

RT: 7.500 min Scan# 943

Delta R.T. -0.000 min

Lab File: VN086867.D




Acq: 06 Jun 2025 14:49


Tgt Ion: 77 Resp: 450112

Ion Ratio Lower Upper

77 100

97 23.4 11.6 34.7





#27

cis-1,2-Dichloroethene

Concen: 141.974 ug/l

RT: 7.500 min Scan# 943

Delta R.T. -0.000 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC150





Tgt Ion:	96	Resp:	363398
Ion	Ratio	Lower	Upper
96	100		
61	138.0	0.0	278.0
98	64.5	0.0	128.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#28

Bromochloromethane

Concen: 152.718 ug/l




RT: 7.824 min Scan# 998


Delta R.T. -0.000 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49

Tgt Ion:	49	Resp:	290594
Ion	Ratio	Lower	Upper
49	100		
129	2.0	0.0	3.8
130	79.8	62.2	93.2





#29

Tetrahydrofuran

Concen: 691.211 ug/l

RT: 7.847 min Scan# 1002

Delta R.T. -0.006 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC150





Tgt Ion:	42	Resp:	89815
Ion	Ratio	Lower	Upper
42	100		
72	48.5	37.8	56.6
71	45.4	35.6	53.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#30

Chloroform

Concen: 138.179 ug/l




RT: 7.976 min Scan# 1024


Delta R.T. -0.000 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49


Tgt Ion:	83	Resp:	534038
Ion	Ratio	Lower	Upper
83	100		
85	63.4	52.9	79.3





#31
Cyclohexane
Concen: 134.766 ug/l
RT: 8.265 min Scan# 1073
Delta R.T. -0.006 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150



Tgt Ion: 56 Resp: 50579
Ion Ratio Lower Upper
56 100
69 32.0 25.4 38.0
84 90.6 68.8 103.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#32
1,1,1-Trichloroethane
Concen: 140.807 ug/l
RT: 8.176 min Scan# 1058
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49

Tgt Ion: 97 Resp: 462854
Ion Ratio Lower Upper
97 100
99 64.9 52.0 78.0
61 43.2 35.9 53.9





#33

1,2-Dichloroethane-d4

Concen: 168.233 ug/l

RT: 8.588 min Scan# 1128

Delta R.T. -0.000 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC150

Tgt Ion: 65 Resp: 38929

Ion Ratio Lower Upper

65 100





67 54.9 0.0 105.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.106 min Scan# 1216

Delta R.T. -0.000 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49




Tgt Ion:114 Resp: 307937


Ion Ratio Lower Upper

114 100

63 19.8 0.0 39.6


88 16.5 0.0 30.2





#35
Dibromofluoromethane
Concen: 177.803 ug/l
RT: 8.176 min Scan# 1057
Delta R.T. 0.006 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150



Tgt Ion: 113 Resp: 324474
Ion Ratio Lower Upper
113 100
111 102.5 84.2 126.2
192 17.7 14.2 21.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#36
1,1-Dichloropropene
Concen: 144.584 ug/l
RT: 8.376 min Scan# 1092
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49

Tgt Ion: 75 Resp: 393166
Ion Ratio Lower Upper
75 100
110 35.6 17.6 52.9
77 31.4 25.3 37.9





#37

Ethyl Acetate

Concen: 142.706 ug/l

RT: 7.571 min Scan# 955

Delta R.T. -0.006 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC150

Tgt Ion: 43 Resp: 494000

Ion Ratio Lower Upper

43 100

61 15.4 12.1 18.1





70 12.7 9.8 14.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#38

Carbon Tetrachloride

Concen: 145.557 ug/l

RT: 8.371 min Scan# 1091

Delta R.T. -0.000 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49




Tgt Ion:117 Resp: 388595


Ion Ratio Lower Upper

117 100

119 94.9 75.0 112.6

121 29.1 24.9 37.3





#39

Methylcyclohexane

Concen: 146.136 ug/l

RT: 9.606 min Scan# 1130

Delta R.T. -0.000 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC150

Tgt Ion: 83 Resp: 54443

Ion Ratio Lower Upper

83 100

55 73.6 61.0 91.6





98 47.4 38.9 58.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#40

Benzene

Concen: 142.388 ug/l

RT: 8.612 min Scan# 1132

Delta R.T. -0.000 min

Lab File: VN086867.D




Acq: 06 Jun 2025 14:49


Tgt Ion: 78 Resp: 1266139

Ion Ratio Lower Upper

78 100


77 23.6 18.7 28.1





#41
Methacrylonitrile
Concen: 142.174 ug/l
RT: 7.788 min Scan# 992
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150



Tgt Ion: 41 Resp: 276400
Ion Ratio Lower Upper
41 100
39 50.9 41.2 61.8
67 79.4 61.4 92.2
52 29.0 22.7 34.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#42
1,2-Dichloroethane
Concen: 141.455 ug/l
RT: 8.676 min Scan# 1143
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49

Tgt Ion: 62 Resp: 381516
Ion Ratio Lower Upper
62 100
98 10.5 0.0 21.2





#43

Isopropyl Acetate

Concen: 140.531 ug/l

RT: 8.694 min Scan# 1146

Delta R.T. -0.000 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC150

Tgt Ion: 43 Resp: 78079

Ion Ratio Lower Upper

43

100

61

28.3

87

15.3

22.1

11.8

33.1





17.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#44

Trichloroethene

Concen: 143.490 ug/l

RT: 9.359 min Scan# 1259

Delta R.T. -0.000 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49

Tgt Ion:130 Resp: 302548

Ion Ratio Lower Upper

130




100


95

101.6

0.0

196.2





#45

1,2-Dichloropropane

Concen: 142.856 ug/l

RT: 9.623 min Scan# 1304

Delta R.T. -0.006 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC150

Tgt Ion: 63 Resp: 30904

Ion Ratio Lower Upper

63 100





65 31.9 25.5 38.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#46

Dibromomethane

Concen: 145.074 ug/l

RT: 9.712 min Scan# 1319

Delta R.T. -0.006 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49




Tgt Ion: 93 Resp: 208461


Ion Ratio Lower Upper

93 100

95 84.8 67.1 100.7

174 92.8 74.2 111.4





#47

Bromodichloromethane

Concen: 145.449 ug/l

RT: 9.894 min Scan# 1350

Delta R.T. -0.000 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC150





Tgt Ion:	83	Resp:	43001
Ion	Ratio	Lower	Upper
83	100		
85	63.6	50.9	76.3
127	8.1	6.6	10.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#48

Methyl methacrylate

Concen: 143.502 ug/l




RT: 9.688 min Scan# 1315


Delta R.T. 0.006 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49

Tgt Ion:	41	Resp:	366967
Ion	Ratio	Lower	Upper
41	100		
69	94.2	74.6	112.0
39	50.8	40.2	60.4





#49

1,4-Dioxane

Concen: 2967.116 ug/l

RT: 9.712 min Scan# 1317

Delta R.T. 0.011 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC150

Tgt Ion: 88 Resp: 138035

Ion Ratio Lower Upper

88 100

43 0.0 0.0 0.0





58 72.2 57.7 86.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#50

Toluene-d8

Concen: 176.120 ug/l

RT: 10.570 min Scan# 1465

Delta R.T. -0.000 min

Lab File: VN086867.D




Acq: 06 Jun 2025 14:49


Tgt Ion: 98 Resp: 1272348

Ion Ratio Lower Upper

98 100


100 67.5 53.4 80.0





#51
4-Methyl-2-Pentanone
Concen: 728.995 ug/l
RT: 10.447 min Scan# 1444
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC150



Tgt Ion: 43 Resp: 243831
Ion Ratio Lower Upper
43 100
58 44.0 34.2 51.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#52
Toluene
Concen: 145.976 ug/l
RT: 10.635 min Scan# 1476
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49


Tgt Ion: 92 Resp: 793270
Ion Ratio Lower Upper
92 100
91 170.2 136.5 204.7





#53
t-1,3-Dichloropropene
Concen: 148.163 ug/l
RT: 10.841 min Scan# 1511
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC150



Tgt Ion: 75 Resp: 489809
Ion Ratio Lower Upper
75 100
77 31.4 26.4 39.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#54
cis-1,3-Dichloropropene
Concen: 146.469 ug/l
RT: 10.318 min Scan# 1422
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49


Tgt Ion: 75 Resp: 518092
Ion Ratio Lower Upper
75 100
77 31.4 26.2 39.4
39 40.6 32.3 48.5





#55
1,1,2-Trichloroethane
Concen: 142.588 ug/l
RT: 11.017 min Scan# 1541
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150




Tgt Ion: 97 Resp: 29815
Ion Ratio Lower Upper
97 100
83 85.6 69.7 104.5
85 55.2 45.1 67.7
99 63.1 50.6 75.8

Manual Integrations
APPROVED



Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025




#56
Ethyl methacrylate
Concen: 160.311 ug/l
RT: 10.876 min Scan# 1517
Delta R.T. -0.006 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49



Tgt Ion: 69 Resp: 533912
Ion Ratio Lower Upper
69 100
41 57.5 47.2 70.8
39 27.0 23.0 34.6





#57

1,3-Dichloropropane

Concen: 142.363 ug/l

RT: 11.165 min Scan# 1566

Delta R.T. -0.000 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC150

Tgt Ion: 76 Resp: 51639

Ion Ratio Lower Upper

76 100





78 32.5 25.3 37.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#58

2-Chloroethyl Vinyl ether

Concen: 941.147 ug/l

RT: 10.165 min Scan# 1396

Delta R.T. -0.000 min

Lab File: VN086867.D




Acq: 06 Jun 2025 14:49


Tgt Ion: 63 Resp: 1869543

Ion Ratio Lower Upper

63 100

106 29.2 23.0 34.6





#59

2-Hexanone

Concen: 805.662 ug/l

RT: 11.200 min Scan# 11

Delta R.T. -0.000 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC150

Tgt Ion: 43 Resp: 173577

Ion Ratio Lower Upper

43 100





58 61.6 30.4 91.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#60

Dibromochloromethane

Concen: 147.832 ug/l

RT: 11.365 min Scan# 1600

Delta R.T. 0.006 min

Lab File: VN086867.D




Acq: 06 Jun 2025 14:49


Tgt Ion:129 Resp: 322064

Ion Ratio Lower Upper

129 100


127 78.6 38.6 115.8





#61
1,2-Dibromoethane
Concen: 146.790 ug/l
RT: 11.470 min Scan# 1618
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150



Tgt Ion:107 Resp: 314600
Ion Ratio Lower Upper
107 100
109 94.0 76.3 114.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#62
4-Bromofluorobenzene
Concen: 179.364 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49


Tgt Ion: 95 Resp: 481421
Ion Ratio Lower Upper
95 100
174 70.7 0.0 141.8
176 67.8 0.0 132.6





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.864 min Scan# 1685
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150



Tgt Ion:117 Resp: 278552
Ion Ratio Lower Upper
117 100
82 54.7 44.6 67.0
119 31.3 25.5 38.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#64
Tetrachloroethene
Concen: 139.048 ug/l
RT: 11.106 min Scan# 1556
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49


Tgt Ion:164 Resp: 245122
Ion Ratio Lower Upper
164 100
166 126.7 101.8 152.8
129 96.6 83.0 124.4
131 94.9 78.2 117.2





#65
Chlorobenzene
Concen: 140.041 ug/l
RT: 11.894 min Scan# 110
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC150



Tgt Ion:112 Resp: 860345
Ion Ratio Lower Upper
112 100
114 31.8 26.2 39.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#66
1,1,1,2-Tetrachloroethane
Concen: 143.171 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. -0.006 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49

Tgt Ion:131 Resp: 282739
Ion Ratio Lower Upper
131 100
133 95.2 48.3 144.8
119 67.8 34.0 101.8









#67
Ethyl Benzene
Concen: 143.438 ug/l
RT: 11.964 min Scan# 1702
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49

Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150

Tgt Ion: 91 Resp: 151788
Ion Ratio Lower Upper
91 100
106 30.9 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#68
m/p-Xylenes
Concen: 290.155 ug/l
RT: 12.070 min Scan# 1720
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49

Tgt Ion:106 Resp: 1175378
Ion Ratio Lower Upper
106 100
91 200.8 159.4 239.0





#69

o-Xylene

Concen: 145.966 ug/l

RT: 12.400 min Scan# 1776

Delta R.T. -0.000 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC150

Tgt Ion:106 Resp: 56630

Ion Ratio Lower Upper

106 100





91 211.9 106.1 318.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#70

Styrene

Concen: 146.459 ug/l

RT: 12.411 min Scan# 1778

Delta R.T. -0.000 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49




Tgt Ion:104 Resp: 972318


Ion Ratio Lower Upper

104 100

78 48.4 38.5 57.7


103 54.0 43.1 64.7





#71
Bromoform
Concen: 152.538 ug/l
RT: 12.582 min Scan# 1807
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150




Tgt Ion:173 Resp: 223200
Ion Ratio Lower Upper
173 100
175 48.8 24.1 72.4
254 0.0 0.0 0.0

Manual Integrations
APPROVED



Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.794 min Scan# 2013
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49



Tgt Ion:152 Resp: 136630
Ion Ratio Lower Upper
152 100
115 61.1 30.1 90.5
150 206.7 0.0 345.0









#73
Isopropylbenzene
Concen: 146.042 ug/l
RT: 12.694 min Scan# 11
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC150

Tgt Ion:105 Resp: 145365
Ion Ratio Lower Upper
105 100
120 26.9 13.5 40.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#74
N-ethyl acetate
Concen: 164.596 ug/l
RT: 12.500 min Scan# 1793
Delta R.T. -0.012 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49


Tgt Ion: 43 Resp: 543747
Ion Ratio Lower Upper
43 100
70 50.0 37.6 56.4
55 26.1 19.0 28.6
61 27.6 20.8 31.2





#75
1,1,2,2-Tetrachloroethane
Concen: 140.693 ug/l
RT: 12.941 min Scan# 1868
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150



Tgt Ion: 83 Resp: 474420
Ion Ratio Lower Upper
83 100
131 9.9 4.9 14.7
85 64.3 32.0 96.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#76
1,2,3-Trichloropropane
Concen: 143.665 ug/l m
RT: 12.994 min Scan# 1877
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49

Tgt Ion: 75 Resp: 468041
Ion Ratio Lower Upper
75 100
77 166.0 79.3 237.9









#77
Bromobenzene
Concen: 147.077 ug/l
RT: 12.982 min Scan# 1875
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49

Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150

Tgt Ion:156 Resp: 335734
Ion Ratio Lower Upper
156 100
77 231.5 117.7 353.1
158 96.3 48.1 144.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#78
n-propylbenzene
Concen: 146.301 ug/l
RT: 13.035 min Scan# 1884
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49

Tgt Ion: 91 Resp: 1769694
Ion Ratio Lower Upper
91 100
120 23.2 11.4 34.2









#79
2-Chlorotoluene
Concen: 144.317 ug/l
RT: 13.123 min Scan# 1899
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC150

Tgt Ion: 91 Resp: 1046910
Ion Ratio Lower Upper
91 100
126 32.8 16.2 48.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#80
1,3,5-Trimethylbenzene
Concen: 147.279 ug/l
RT: 13.170 min Scan# 1907
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49


Tgt Ion:105 Resp: 1210325
Ion Ratio Lower Upper
105 100
120 50.0 24.9 74.6





#81
trans-1,4-Dichloro-2-butene
Concen: 153.277 ug/l
RT: 12.735 min Scan# 1833
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150



Tgt Ion: 75 Resp: 21625
Ion Ratio Lower Upper
75 100
53 81.3 66.6 99.8
89 40.0 36.7 55.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#82
4-Chlorotoluene
Concen: 144.900 ug/l
RT: 13.223 min Scan# 1916
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49


Tgt Ion: 91 Resp: 1063581
Ion Ratio Lower Upper
91 100
126 33.4 16.7 50.1





#83
tert-Butylbenzene
Concen: 145.352 ug/l
RT: 13.435 min Scan# 1952
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150



Tgt Ion:119 Resp: 1093400
Ion Ratio Lower Upper
119 100
91 63.8 32.1 96.5
134 26.7 13.4 40.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#84
1,2,4-Trimethylbenzene
Concen: 147.626 ug/l
RT: 13.482 min Scan# 1960
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49

Tgt Ion:105 Resp: 1216529
Ion Ratio Lower Upper
105 100
120 46.7 23.2 69.6









#85
sec-Butylbenzene
Concen: 144.885 ug/l
RT: 13.617 min Scan# 1983
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC150

Tgt Ion:105 Resp: 1582762
Ion Ratio Lower Upper
105 100
134 20.3 10.0 30.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#86
p-Isopropyltoluene
Concen: 146.482 ug/l
RT: 13.729 min Scan# 2002
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49

Tgt Ion:119 Resp: 1322780
Ion Ratio Lower Upper
119 100
134 26.4 13.5 40.4
91 22.7 11.5 34.4





#87

1,3-Dichlorobenzene

Concen: 142.939 ug/l

RT: 13.735 min Scan# 2003

Delta R.T. -0.000 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC150

Tgt Ion:146 Resp: 641960

Ion Ratio Lower Upper

146 100

111 42.1 20.9 62.8





148 63.8 32.0 96.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#88

1,4-Dichlorobenzene

Concen: 141.124 ug/l

RT: 13.811 min Scan# 2016

Delta R.T. -0.000 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49




Tgt Ion:146 Resp: 646188


Ion Ratio Lower Upper

146 100

111 40.2 19.9 59.7

148 63.2 31.8 95.4





#89

n-Butylbenzene

Concen: 143.117 ug/l

RT: 14.053 min Scan# 2057

Delta R.T. -0.000 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC150

Tgt Ion: 91 Resp: 1251819

Ion Ratio Lower Upper

91 100

92 54.6 27.5 82.4





134 26.1 13.0 39.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#90

Hexachloroethane

Concen: 150.387 ug/l

RT: 14.335 min Scan# 2105

Delta R.T. -0.000 min

Lab File: VN086867.D




Acq: 06 Jun 2025 14:49


Tgt Ion:117 Resp: 230195

Ion Ratio Lower Upper

117 100

201 64.5 33.2 99.6









#91
1,2-Dichlorobenzene
Concen: 142.108 ug/l
RT: 14.105 min Scan# 2066
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49

Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150

Tgt Ion:146 Resp: 61317
Ion Ratio Lower Upper
146 100
111 43.4 21.6 64.6
148 63.8 31.7 95.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#92
1,2-Dibromo-3-Chloropropane
Concen: 137.242 ug/l
RT: 14.717 min Scan# 2170
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49

Tgt Ion: 75 Resp: 110677
Ion Ratio Lower Upper
75 100
155 85.5 41.1 123.4
157 110.2 52.7 158.1





#93

1,2,4-Trichlorobenzene

Concen: 147.940 ug/l

RT: 15.394 min Scan# 2285

Delta R.T. 0.006 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC150

Tgt Ion:180 Resp: 40762

Ion Ratio Lower Upper

180 100

182 95.0 46.9 140.7





145 32.0 16.0 47.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#94

Hexachlorobutadiene

Concen: 147.264 ug/l

RT: 15.500 min Scan# 2303

Delta R.T. -0.000 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49




Tgt Ion:225 Resp: 151172


Ion Ratio Lower Upper

225 100

223 64.8 33.5 100.4


227 63.0 31.1 93.5





#95
Naphthalene
Concen: 150.743 ug/l
RT: 15.641 min Scan# 2327
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC150



Tgt Ion:128 Resp: 154598
Ion Ratio Lower Upper
128 100
127 12.6 10.2 15.2
129 10.9 8.8 13.2




Manual Integrations
APPROVED

Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#96
1,2,3-Trichlorobenzene
Concen: 147.743 ug/l
RT: 15.841 min Scan# 2361
Delta R.T. 0.006 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49

Tgt Ion:180 Resp: 404450
Ion Ratio Lower Upper
180 100
182 94.7 47.0 141.0
145 33.6 16.6 49.8



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060625\
 Data File : VN086869.D
 Acq On : 06 Jun 2025 15:54
 Operator : JC\MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ICVVN060625

Manual Integrations
 APPROVED

Reviewed By :John Carlone 06/09/2025
 Supervised By :Mahesh Dadoda 06/09/2025

Quant Time: Jun 07 02:16:19 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	218122	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	394140	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	348935	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	172710	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	152255	52.135	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 104.260%			
35) Dibromofluoromethane	8.177	113	127102	54.416	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 108.840%			
50) Toluene-d8	10.571	98	491265	53.129	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 106.260%			
62) 4-Bromofluorobenzene	12.847	95	185792	54.081	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 108.160%			
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	2.154	85	120237	55.286	ug/l	99
3) Chloromethane	2.401	50	139027	49.505	ug/l	99
4) Vinyl Chloride	2.554	62	151951	52.451	ug/l	99
5) Bromomethane	2.989	94	89638	55.292	ug/l	95
6) Chloroethane	3.154	64	95654	51.117	ug/l	96
7) Trichlorofluoromethane	3.524	101	196138	51.819	ug/l	97
8) Diethyl Ether	3.983	74	87405	53.001	ug/l	100
9) 1,1,2-Trichlorotrifluo...	4.395	101	127498	53.615	ug/l	99
10) Methyl Iodide	4.612	142	164080	53.227	ug/l	97
11) Tert butyl alcohol	5.542	59	207768	262.193	ug/l	99
12) 1,1-Dichloroethene	4.359	96	125461	51.678	ug/l	99
13) Acrolein	4.195	56	60336	240.545	ug/l	98
14) Allyl chloride	5.042	41	203344	50.504	ug/l	98
15) Acrylonitrile	5.736	53	479502	258.885	ug/l	100
16) Acetone	4.448	43	382146	246.750	ug/l	98
17) Carbon Disulfide	4.736	76	339466	50.550	ug/l	100
18) Methyl Acetate	5.042	43	234290	51.912	ug/l	100
19) Methyl tert-butyl Ether	5.818	73	454641	51.720	ug/l	100
20) Methylene Chloride	5.295	84	143862	49.617	ug/l	97
21) trans-1,2-Dichloroethene	5.806	96	133535	49.437	ug/l	99
22) Diisopropyl ether	6.689	45	440080	51.853	ug/l	99
23) Vinyl Acetate	6.618	43	1828919	255.055	ug/l	99
24) 1,1-Dichloroethane	6.589	63	251473	51.487	ug/l	98
25) 2-Butanone	7.494	43	644194	255.897	ug/l	99
26) 2,2-Dichloropropane	7.500	77	216678	57.031	ug/l	99
27) cis-1,2-Dichloroethene	7.500	96	164985	51.073	ug/l	99
28) Bromochloromethane	7.824	49	108810	45.310	ug/l	99
29) Tetrahydrofuran	7.847	42	420891	256.656	ug/l	100
30) Chloroform	7.977	83	248429	50.933	ug/l	97
31) Cyclohexane	8.265	56	237836	50.212	ug/l	97
32) 1,1,1-Trichloroethane	8.177	97	209927	50.603	ug/l	97
36) 1,1-Dichloropropene	8.377	75	181414	52.123	ug/l	99
37) Ethyl Acetate	7.571	43	222090	50.125	ug/l	99
38) Carbon Tetrachloride	8.371	117	179281	52.466	ug/l	98
39) Methylcyclohexane	9.606	83	251231	52.686	ug/l	98
40) Benzene	8.612	78	579667	50.931	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060625\
 Data File : VN086869.D
 Acq On : 06 Jun 2025 15:54
 Operator : JC\MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ICVVN060625

Manual Integrations
 APPROVED

Reviewed By : John Carlone 06/09/2025
 Supervised By : Mahesh Dadoda 06/09/2025

Quant Time: Jun 07 02:16:19 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.789	41	130903	52.607	ug/l	100
42) 1,2-Dichloroethane	8.677	62	177746	51.489	ug/l	100
43) Isopropyl Acetate	8.694	43	359378	50.536	ug/l	100
44) Trichloroethene	9.359	130	139763	51.788	ug/l	100
45) 1,2-Dichloropropane	9.624	63	142463	51.450	ug/l	99
46) Dibromomethane	9.712	93	97199	52.849	ug/l	98
47) Bromodichloromethane	9.894	83	196104	51.824	ug/l	100
48) Methyl methacrylate	9.688	41	168001	51.328	ug/l	98
49) 1,4-Dioxane	9.706	88	66102	1110.123	ug/l #	99
51) 4-Methyl-2-Pentanone	10.447	43	1146622	267.835	ug/l	99
52) Toluene	10.635	92	361963	52.040	ug/l	99
53) t-1,3-Dichloropropene	10.835	75	224839	53.137	ug/l	99
54) cis-1,3-Dichloropropene	10.318	75	237774	52.519	ug/l	99
55) 1,1,2-Trichloroethane	11.018	97	137998	51.562	ug/l	99
56) Ethyl methacrylate	10.882	69	241649	56.688	ug/l	98
57) 1,3-Dichloropropane	11.165	76	238572	51.386	ug/l	99
58) 2-Chloroethyl Vinyl ether	10.165	63	625479	246.006	ug/l	100
59) 2-Hexanone	11.200	43	781717	283.478	ug/l	100
60) Dibromochloromethane	11.359	129	146749	52.627	ug/l	99
61) 1,2-Dibromoethane	11.471	107	142338	51.887	ug/l	99
64) Tetrachloroethene	11.106	164	111409	50.450	ug/l	98
65) Chlorobenzene	11.894	112	391661	50.892	ug/l	98
66) 1,1,1,2-Tetrachloroethane	11.959	131	129045	52.164	ug/l	99
67) Ethyl Benzene	11.965	91	683879	51.590	ug/l	98
68) m/p-Xylenes	12.071	106	527293	103.912	ug/l	98
69) o-Xylene	12.400	106	254600	52.387	ug/l	100
70) Styrene	12.412	104	438751	52.758	ug/l	99
71) Bromoform	12.582	173	100357	54.751	ug/l #	99
73) Isopropylbenzene	12.694	105	652045	51.823	ug/l	100
74) N-amyl acetate	12.506	43	238355	54.212	ug/l	97
75) 1,1,2,2-Tetrachloroethane	12.941	83	220643	51.763	ug/l	100
76) 1,2,3-Trichloropropane	12.994	75	220957m	53.821	ug/l	
77) Bromobenzene	12.982	156	150950	52.313	ug/l	100
78) n-propylbenzene	13.035	91	801180	52.397	ug/l	100
79) 2-Chlorotoluene	13.123	91	473629	51.650	ug/l	99
80) 1,3,5-Trimethylbenzene	13.171	105	548347	52.787	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.735	75	99341	55.702	ug/l	97
82) 4-Chlorotoluene	13.223	91	481403	51.884	ug/l	99
83) tert-Butylbenzene	13.441	119	499519	52.532	ug/l	99
84) 1,2,4-Trimethylbenzene	13.482	105	552333	53.024	ug/l	100
85) sec-Butylbenzene	13.618	105	723986	52.428	ug/l	100
86) p-Isopropyltoluene	13.729	119	606979	53.174	ug/l	99
87) 1,3-Dichlorobenzene	13.735	146	290810	51.224	ug/l	100
88) 1,4-Dichlorobenzene	13.812	146	298412	51.557	ug/l	100
89) n-Butylbenzene	14.059	91	579636	52.424	ug/l	100
90) Hexachloroethane	14.335	117	102254	52.847	ug/l	100
91) 1,2-Dichlorobenzene	14.106	146	280025	51.340	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	14.723	75	50115	49.161	ug/l	98
93) 1,2,4-Trichlorobenzene	15.394	180	181064	51.986	ug/l	99
94) Hexachlorobutadiene	15.500	225	68570	52.843	ug/l	99
95) Naphthalene	15.641	128	674187	52.005	ug/l	100
96) 1,2,3-Trichlorobenzene	15.841	180	174888	50.539	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060625\
Data File : VN086869.D
Acq On : 06 Jun 2025 15:54
Operator : JC\MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ICVVN060625

Manual Integrations
APPROVED

Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025

Quant Time: Jun 07 02:16:19 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 02:12:50 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :
MSVOA_N
ClientSampleId :
ICVVN060625

Manual Integrations APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.230 min Scan# 1068
Delta R.T. -0.006 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54




Instrument :
MSVOA_N
ClientSampleId :
ICVVN060625



Tgt Ion:168 Resp: 21812
Ion Ratio Lower Upper
168 100
99 58.6 49.1 73.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#2
Dichlorodifluoromethane
Concen: 55.286 ug/l
RT: 2.154 min Scan# 34
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54


Tgt Ion: 85 Resp: 120237
Ion Ratio Lower Upper
85 100
87 33.2 16.2 48.6





#3
Chloromethane
Concen: 49.505 ug/l
RT: 2.401 min Scan# 75
Delta R.T. 0.006 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54




Instrument :
MSVOA_N
ClientSampleId :
ICVVN060625



Tgt Ion: 50 Resp: 13902
Ion Ratio Lower Upper
50 100
52 32.9 25.7 38.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#4
Vinyl Chloride
Concen: 52.451 ug/l
RT: 2.554 min Scan# 102
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54


Tgt Ion: 62 Resp: 151951
Ion Ratio Lower Upper
62 100
64 32.0 26.0 39.0





#5
Bromomethane
Concen: 55.292 ug/l
RT: 2.989 min Scan# 176
Delta R.T. -0.006 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54




Instrument :
MSVOA_N
ClientSampleId :
ICVVN060625



Tgt Ion: 94 Resp: 89638
Ion Ratio Lower Upper
94 100
96 92.5 77.9 116.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#6
Chloroethane
Concen: 51.117 ug/l
RT: 3.154 min Scan# 204
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54


Tgt Ion: 64 Resp: 95654
Ion Ratio Lower Upper
64 100
66 32.6 24.6 36.8





#7
Trichlorofluoromethane
Concen: 51.819 ug/l
RT: 3.524 min Scan# 267
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54




Instrument :
MSVOA_N
ClientSampleId :
ICVVN060625



Tgt Ion:101 Resp: 196138
Ion Ratio Lower Upper
101 100
103 62.9 52.2 78.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#8
Diethyl Ether
Concen: 53.001 ug/l
RT: 3.983 min Scan# 345
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54

Tgt Ion: 74 Resp: 87405
Ion Ratio Lower Upper
74 100
45 90.9 45.5 136.5





#9

1,1,2-Trichlorotrifluoroethane

Concen: 53.615 ug/l

RT: 4.395 min Scan# 415

Delta R.T. -0.000 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54

Instrument :

MSVOA_N

ClientSampleId :

ICVVN060625

Tgt Ion:101 Resp: 127498

Ion Ratio Lower Upper

101 100

85 41.9 33.8 50.8





151 72.8 58.7 88.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#10

Methyl Iodide

Concen: 53.227 ug/l

RT: 4.612 min Scan# 452

Delta R.T. 0.006 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54




Tgt Ion:142 Resp: 164080


Ion Ratio Lower Upper

142 100

127 36.6 30.9 46.3

141 14.0 10.4 15.6





#11

Tert butyl alcohol

Concen: 262.193 ug/l

RT: 5.542 min Scan# 611

Delta R.T. -0.006 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54

Instrument :

MSVOA_N

ClientSampleId :

ICVVN060625

Tgt Ion: 59 Resp: 207768

Ion Ratio Lower Upper

59 100





57 9.8 8.1 12.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#12

1,1-Dichloroethene

Concen: 51.678 ug/l

RT: 4.359 min Scan# 409

Delta R.T. -0.000 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54




Tgt Ion: 96 Resp: 125461


Ion Ratio Lower Upper

96 100

61 158.0 125.8 188.6

98 65.2 52.5 78.7





#13

Acrolein

Concen: 240.545 ug/l

RT: 4.195 min Scan# 381

Delta R.T. -0.006 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54

Instrument :

MSVOA_N

ClientSampleId :

ICVVN060625

Tgt Ion: 56 Resp: 60330

Ion Ratio Lower Upper

56 100





55 71.1 55.6 83.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#14

Allyl chloride

Concen: 50.504 ug/l

RT: 5.042 min Scan# 525

Delta R.T. -0.000 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54




Tgt Ion: 41 Resp: 203344


Ion Ratio Lower Upper

41 100

39 62.3 51.3 76.9


76 40.4 31.4 47.0





#15
Acrylonitrile
Concen: 258.885 ug/l
RT: 5.736 min Scan# 644
Delta R.T. -0.006 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54




Instrument :
MSVOA_N
ClientSampleId :
ICVVN060625



Tgt Ion: 53 Resp: 479502
Ion Ratio Lower Upper
53 100
52 81.3 65.0 97.6
51 34.5 27.2 40.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#16
Acetone
Concen: 246.750 ug/l
RT: 4.448 min Scan# 424
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54


Tgt Ion: 43 Resp: 382146
Ion Ratio Lower Upper
43 100
58 35.9 28.0 42.0





#17
Carbon Disulfide
Concen: 50.550 ug/l
RT: 4.736 min Scan# 473
Delta R.T. 0.006 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54




Instrument :
MSVOA_N
ClientSampleId :
ICVVN060625



Tgt Ion: 76 Resp: 33946
Ion Ratio Lower Upper
76 100
78 9.1 7.3 10.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#18
Methyl Acetate
Concen: 51.912 ug/l
RT: 5.042 min Scan# 525
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54

Tgt Ion: 43 Resp: 234290
Ion Ratio Lower Upper
43 100
74 25.9 20.7 31.1





#19

Methyl tert-butyl Ether

Concen: 51.720 ug/l

RT: 5.818 min Scan# 656

Delta R.T. 0.006 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54

Instrument :

MSVOA_N

ClientSampleId :

ICVVN060625

Tgt Ion: 73 Resp: 45464

Ion Ratio Lower Upper

73 100





57 21.8 17.4 26.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#20

Methylene Chloride

Concen: 49.617 ug/l

RT: 5.295 min Scan# 568

Delta R.T. -0.000 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54

Tgt Ion: 84 Resp: 143862




Ion Ratio Lower Upper


84 100

49 110.3 90.5 135.7

51 31.8 28.5 42.7

86 60.8 48.1 72.1





#21

trans-1,2-Dichloroethene

Concen: 49.437 ug/l

RT: 5.806 min Scan# 61

Delta R.T. -0.000 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54

Instrument :

MSVOA_N

ClientSampleId :

ICVVN060625

Tgt Ion: 96 Resp: 13353

Ion Ratio Lower Upper

96 100

61 136.8 109.2 163.8





98 65.2 49.8 74.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#22

Diisopropyl ether

Concen: 51.853 ug/l

RT: 6.689 min Scan# 805

Delta R.T. -0.000 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54

Tgt Ion: 45 Resp: 440080




Ion Ratio Lower Upper


45 100

43 52.4 41.9 62.9

87 29.5 22.4 33.6


59 12.4 9.8 14.6





#23
Vinyl Acetate
Concen: 255.055 ug/l
RT: 6.618 min Scan# 793
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54




Instrument :
MSVOA_N
ClientSampleId :
ICVVN060625



Tgt Ion: 43 Resp: 1828919
Ion Ratio Lower Upper
43 100
86 12.1 9.5 14.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#24
1,1-Dichloroethane
Concen: 51.487 ug/l
RT: 6.589 min Scan# 788
Delta R.T. 0.006 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54


Tgt Ion: 63 Resp: 251473
Ion Ratio Lower Upper
63 100
98 7.7 3.6 10.8
100 4.6 2.6 7.8





#25
2-Butanone
Concen: 255.897 ug/l
RT: 7.494 min Scan# 942
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54




Instrument :
MSVOA_N
ClientSampleId :
ICVVN060625



Tgt Ion: 43 Resp: 644194
Ion Ratio Lower Upper
43 100
72 29.5 23.0 34.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#26
2,2-Dichloropropane
Concen: 57.031 ug/l
RT: 7.500 min Scan# 943
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54

Tgt Ion: 77 Resp: 216678
Ion Ratio Lower Upper
77 100
97 23.7 11.6 34.7





#27

cis-1,2-Dichloroethene

Concen: 51.073 ug/l

RT: 7.500 min Scan# 943

Delta R.T. -0.000 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54

Instrument :

MSVOA_N

ClientSampleId :

ICVVN060625





Tgt Ion:	96	Resp:	16498
Ion	Ratio	Lower	Upper
96	100		
61	140.3	0.0	278.0
98	64.5	0.0	128.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#28

Bromochloromethane

Concen: 45.310 ug/l




RT: 7.824 min Scan# 998


Delta R.T. -0.000 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54

Tgt Ion:	49	Resp:	108810
Ion	Ratio	Lower	Upper
49	100		
129	1.9	0.0	3.8
130	78.5	62.2	93.2





#29

Tetrahydrofuran

Concen: 256.656 ug/l

RT: 7.847 min Scan# 1002

Delta R.T. -0.006 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54

Instrument :

MSVOA_N

ClientSampleId :

ICVVN060625

Tgt Ion: 42 Resp: 42089

Ion Ratio Lower Upper

42 100

72 47.5 37.8 56.6





71 44.5 35.6 53.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#30

Chloroform

Concen: 50.933 ug/l

RT: 7.977 min Scan# 1024

Delta R.T. -0.000 min

Lab File: VN086869.D




Acq: 06 Jun 2025 15:54


Tgt Ion: 83 Resp: 248429

Ion Ratio Lower Upper

83 100

85 63.5 52.9 79.3





#31

Cyclohexane

Concen: 50.212 ug/l

RT: 8.265 min Scan# 1073

Delta R.T. -0.006 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54

Instrument :

MSVOA_N

ClientSampleId :

ICVVN060625

Tgt Ion: 56 Resp: 237830

Ion Ratio Lower Upper

56 100

69 33.9 25.4 38.0





84 88.4 68.8 103.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#32

1,1,1-Trichloroethane

Concen: 50.603 ug/l

RT: 8.177 min Scan# 1058

Delta R.T. -0.000 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54




Tgt Ion: 97 Resp: 209927


Ion Ratio Lower Upper

97 100

99 61.7 52.0 78.0

61 45.4 35.9 53.9





#33

1,2-Dichloroethane-d4

Concen: 52.135 ug/l

RT: 8.588 min Scan# 1128

Delta R.T. -0.000 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54

Instrument :

MSVOA_N

ClientSampleId :

ICVVN060625

Tgt Ion: 65 Resp: 15225

Ion Ratio Lower Upper

65 100





67 54.6 0.0 105.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.106 min Scan# 1216

Delta R.T. -0.000 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54




Tgt Ion:114 Resp: 394140


Ion Ratio Lower Upper

114 100

63 20.2 0.0 39.6

88 16.4 0.0 30.2





#35

Dibromofluoromethane

Concen: 54.416 ug/l

RT: 8.177 min Scan# 1057

Delta R.T. 0.006 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54

Instrument :

MSVOA_N

ClientSampleId :

ICVVN060625

Tgt Ion:113 Resp: 12710

Ion Ratio Lower Upper

113 100

111 103.4 84.2 126.2





192 18.0 14.2 21.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#36

1,1-Dichloropropene

Concen: 52.123 ug/l

RT: 8.377 min Scan# 1092

Delta R.T. -0.000 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54




Tgt Ion: 75 Resp: 181414


Ion Ratio Lower Upper

75 100

110 35.0 17.6 52.9

77 31.0 25.3 37.9





#37

Ethyl Acetate

Concen: 50.125 ug/l

RT: 7.571 min Scan# 955

Delta R.T. -0.006 min

Lab File: VN086869.D


Acq: 06 Jun 2025 15:54

Instrument :

MSVOA_N

ClientSampleId :

ICVVN060625



Tgt Ion: 43 Resp: 222090

Ion Ratio Lower Upper

43 100

61 15.6 12.1 18.1




70 12.9 9.8 14.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#38

Carbon Tetrachloride

Concen: 52.466 ug/l

RT: 8.371 min Scan# 1091

Delta R.T. -0.000 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54




Tgt Ion:117 Resp: 179281


Ion Ratio Lower Upper

117 100

119 94.8 75.0 112.6

121 28.5 24.9 37.3





#39

Methylcyclohexane

Concen: 52.686 ug/l

RT: 9.606 min Scan# 1131

Delta R.T. -0.000 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54

Instrument :

MSVOA_N

ClientSampleId :

ICVVN060625

Tgt Ion: 83 Resp: 25123

Ion Ratio Lower Upper

83 100

55 74.5 61.0 91.6





98 47.2 38.9 58.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#40

Benzene

Concen: 50.931 ug/l

RT: 8.612 min Scan# 1132

Delta R.T. -0.000 min

Lab File: VN086869.D




Acq: 06 Jun 2025 15:54


Tgt Ion: 78 Resp: 579667

Ion Ratio Lower Upper

78 100


77 23.7 18.7 28.1





#41
Methacrylonitrile
Concen: 52.607 ug/l
RT: 7.789 min Scan# 992
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54




Instrument :
MSVOA_N
ClientSampleId :
ICVVN060625



Tgt Ion: 41 Resp: 130900
Ion Ratio Lower Upper
41 100
39 51.3 41.2 61.8
67 76.9 61.4 92.2
52 28.6 22.7 34.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#42
1,2-Dichloroethane
Concen: 51.489 ug/l
RT: 8.677 min Scan# 1143
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54

Tgt Ion: 62 Resp: 177746
Ion Ratio Lower Upper
62 100
98 10.6 0.0 21.2





#43

Isopropyl Acetate

Concen: 50.536 ug/l

RT: 8.694 min Scan# 1146

Delta R.T. -0.000 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54

Instrument :

MSVOA_N

ClientSampleId :

ICVVN060625

Tgt Ion: 43 Resp: 359378

Ion Ratio Lower Upper

43 100

61 27.9 22.1 33.1





87 14.8 11.8 17.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#44

Trichloroethene

Concen: 51.788 ug/l

RT: 9.359 min Scan# 1259

Delta R.T. -0.000 min

Lab File: VN086869.D




Acq: 06 Jun 2025 15:54


Tgt Ion:130 Resp: 139763

Ion Ratio Lower Upper

130 100

95 98.0 0.0 196.2





#45

1,2-Dichloropropane

Concen: 51.450 ug/l

RT: 9.624 min Scan# 1319

Delta R.T. -0.006 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54

Instrument :

MSVOA_N

ClientSampleId :

ICVVN060625

Tgt Ion: 63 Resp: 14246

Ion Ratio Lower Upper

63 100





65 32.7 25.5 38.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#46

Dibromomethane

Concen: 52.849 ug/l

RT: 9.712 min Scan# 1319

Delta R.T. -0.006 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54




Tgt Ion: 93 Resp: 97199


Ion Ratio Lower Upper

93 100

95 82.2 67.1 100.7


174 90.9 74.2 111.4





#47
Bromodichloromethane
Concen: 51.824 ug/l
RT: 9.894 min Scan# 1350
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54




Instrument :
MSVOA_N
ClientSampleId :
ICVVN060625



Tgt Ion: 83 Resp: 196104
Ion Ratio Lower Upper
83 100
85 63.8 50.9 76.3
127 8.2 6.6 10.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#48
Methyl methacrylate
Concen: 51.328 ug/l
RT: 9.688 min Scan# 1315
Delta R.T. 0.006 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54

Tgt Ion: 41 Resp: 168001
Ion Ratio Lower Upper
41 100
69 95.2 74.6 112.0
39 51.0 40.2 60.4





#49

1,4-Dioxane

Concen: 1110.123 ug/l

RT: 9.706 min Scan# 11

Delta R.T. 0.006 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54

Instrument :

MSVOA_N

ClientSampleId :

ICVVN060625

Tgt Ion: 88 Resp: 66102

Ion Ratio Lower Upper

88 100

43 0.0 0.0 0.0





58 72.7 57.7 86.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#50

Toluene-d8

Concen: 53.129 ug/l

RT: 10.571 min Scan# 1465

Delta R.T. -0.000 min

Lab File: VN086869.D




Acq: 06 Jun 2025 15:54


Tgt Ion: 98 Resp: 491265

Ion Ratio Lower Upper

98 100


100 67.6 53.4 80.0





#51
4-Methyl-2-Pentanone
Concen: 267.835 ug/l
RT: 10.447 min Scan# 1444
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54




Instrument :
MSVOA_N
ClientSampleId :
ICVVN060625



Tgt Ion: 43 Resp: 114662
Ion Ratio Lower Upper
43 100
58 43.2 34.2 51.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#52
Toluene
Concen: 52.040 ug/l
RT: 10.635 min Scan# 1476
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54


Tgt Ion: 92 Resp: 361963
Ion Ratio Lower Upper
92 100
91 169.8 136.5 204.7





#53
t-1,3-Dichloropropene
Concen: 53.137 ug/l
RT: 10.835 min Scan# 1510
Delta R.T. -0.006 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54




Instrument :
MSVOA_N
ClientSampleId :
ICVVN060625



Tgt Ion: 75 Resp: 224839
Ion Ratio Lower Upper
75 100
77 32.3 26.4 39.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#54
cis-1,3-Dichloropropene
Concen: 52.519 ug/l
RT: 10.318 min Scan# 1422
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54


Tgt Ion: 75 Resp: 237774
Ion Ratio Lower Upper
75 100
77 32.1 26.2 39.4
39 41.3 32.3 48.5





#55
1,1,2-Trichloroethane
Concen: 51.562 ug/l
RT: 11.018 min Scan# 1541
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54




Instrument :
MSVOA_N
ClientSampleId :
ICVVN060625



Tgt Ion: 97 Resp: 137998
Ion Ratio Lower Upper
97 100
83 87.4 69.7 104.5
85 54.1 45.1 67.7
99 62.0 50.6 75.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#56
Ethyl methacrylate
Concen: 56.688 ug/l
RT: 10.882 min Scan# 1518
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54


Tgt Ion: 69 Resp: 241649
Ion Ratio Lower Upper
69 100
41 57.7 47.2 70.8
39 27.4 23.0 34.6





#57
1,3-Dichloropropane
Concen: 51.386 ug/l
RT: 11.165 min Scan# 1566
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54




Instrument :
MSVOA_N
ClientSampleId :
ICVVN060625



Tgt Ion: 76 Resp: 238572
Ion Ratio Lower Upper
76 100
78 32.0 25.3 37.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#58
2-Chloroethyl Vinyl ether
Concen: 246.006 ug/l
RT: 10.165 min Scan# 1396
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54


Tgt Ion: 63 Resp: 625479
Ion Ratio Lower Upper
63 100
106 28.9 23.0 34.6





#59
2-Hexanone
Concen: 283.478 ug/l
RT: 11.200 min Scan# 1572
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54




Instrument :
MSVOA_N
ClientSampleId :
ICVVN060625



Tgt Ion: 43 Resp: 78171
Ion Ratio Lower Upper
43 100
58 61.2 30.4 91.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#60
Dibromochloromethane
Concen: 52.627 ug/l
RT: 11.359 min Scan# 1599
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54


Tgt Ion:129 Resp: 146749
Ion Ratio Lower Upper
129 100
127 78.1 38.6 115.8





#61
1,2-Dibromoethane
Concen: 51.887 ug/l
RT: 11.471 min Scan# 1618
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54




Instrument :
MSVOA_N
ClientSampleId :
ICVVN060625



Tgt Ion:107 Resp: 142338
Ion Ratio Lower Upper
107 100
109 94.5 76.3 114.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#62
4-Bromofluorobenzene
Concen: 54.081 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54


Tgt Ion: 95 Resp: 185792
Ion Ratio Lower Upper
95 100
174 71.4 0.0 141.8
176 68.1 0.0 132.6





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 1685
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54




Instrument :
MSVOA_N
ClientSampleId :
ICVVN060625



Tgt Ion:117 Resp: 34893
Ion Ratio Lower Upper
117 100
82 56.5 44.6 67.0
119 32.5 25.5 38.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#64
Tetrachloroethene
Concen: 50.450 ug/l
RT: 11.106 min Scan# 1556
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54

Tgt Ion:164 Resp: 111409
Ion Ratio Lower Upper
164 100
166 128.5 101.8 152.8
129 99.9 83.0 124.4
131 97.9 78.2 117.2









#65
Chlorobenzene
Concen: 50.892 ug/l
RT: 11.894 min Scan# 1690
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54

Instrument :
MSVOA_N
ClientSampleId :
ICVVN060625

Tgt Ion:112 Resp: 39166
Ion Ratio Lower Upper
112 100
114 31.3 26.2 39.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#66
1,1,1,2-Tetrachloroethane
Concen: 52.164 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. -0.006 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54


Tgt Ion:131 Resp: 129045
Ion Ratio Lower Upper
131 100
133 94.5 48.3 144.8
119 67.4 34.0 101.8





#67
Ethyl Benzene
Concen: 51.590 ug/l
RT: 11.965 min Scan# 1702
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54




Instrument :
MSVOA_N
ClientSampleId :
ICVVN060625



Tgt Ion: 91 Resp: 683879
Ion Ratio Lower Upper
91 100
106 30.9 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#68
m/p-Xylenes
Concen: 103.912 ug/l
RT: 12.071 min Scan# 1720
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54


Tgt Ion:106 Resp: 527293
Ion Ratio Lower Upper
106 100
91 202.0 159.4 239.0





#69
o-Xylene
Concen: 52.387 ug/l
RT: 12.400 min Scan# 1776
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54




Instrument :
MSVOA_N
ClientSampleId :
ICVVN060625



Tgt Ion:106 Resp: 254600
Ion Ratio Lower Upper
106 100
91 212.8 106.1 318.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#70
Styrene
Concen: 52.758 ug/l
RT: 12.412 min Scan# 1778
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54


Tgt Ion:104 Resp: 438751
Ion Ratio Lower Upper
104 100
78 48.7 38.5 57.7
103 53.7 43.1 64.7





#71
Bromoform
Concen: 54.751 ug/l
RT: 12.582 min Scan# 1807
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54




Instrument :
MSVOA_N
ClientSampleId :
ICVVN060625




Tgt Ion:173 Resp: 10035
Ion Ratio Lower Upper
173 100
175 48.7 24.1 72.4
254 0.0 0.0 0.0

Manual Integrations
APPROVED



Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.794 min Scan# 2013
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54




Tgt Ion:152 Resp: 172710
Ion Ratio Lower Upper
152 100
115 60.6 30.1 90.5
150 174.0 0.0 345.0





#73
Isopropylbenzene
Concen: 51.823 ug/l
RT: 12.694 min Scan# 1826
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54




Instrument :
MSVOA_N
ClientSampleId :
ICVVN060625



Tgt Ion:105 Resp: 652045
Ion Ratio Lower Upper
105 100
120 27.1 13.5 40.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#74
N-ethyl acetate
Concen: 54.212 ug/l
RT: 12.506 min Scan# 1794
Delta R.T. -0.006 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54


Tgt Ion: 43 Resp: 238355
Ion Ratio Lower Upper
43 100
70 49.3 37.6 56.4
55 25.3 19.0 28.6
61 25.2 20.8 31.2





#75
1,1,2,2-Tetrachloroethane
Concen: 51.763 ug/l
RT: 12.941 min Scan# 1868
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54




Instrument :
MSVOA_N
ClientSampleId :
ICVVN060625



Tgt Ion: 83 Resp: 22064
Ion Ratio Lower Upper
83 100
131 9.8 4.9 14.7
85 63.9 32.0 96.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#76
1,2,3-Trichloropropane
Concen: 53.821 ug/l m
RT: 12.994 min Scan# 1877
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54


Tgt Ion: 75 Resp: 220957
Ion Ratio Lower Upper
75 100
77 160.5 79.3 237.9





#77
Bromobenzene
Concen: 52.313 ug/l
RT: 12.982 min Scan# 1875
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54




Instrument :
MSVOA_N
ClientSampleId :
ICVVN060625




Tgt Ion:156 Resp: 150950
Ion Ratio Lower Upper
156 100
77 234.9 117.7 353.1
158 96.5 48.1 144.4

Manual Integrations
APPROVED



Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025




#78
n-propylbenzene
Concen: 52.397 ug/l
RT: 13.035 min Scan# 1884
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54



Tgt Ion: 91 Resp: 801180
Ion Ratio Lower Upper
91 100
120 22.9 11.4 34.2









#79
2-Chlorotoluene
Concen: 51.650 ug/l
RT: 13.123 min Scan# 1899
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54

Instrument :
MSVOA_N
ClientSampleId :
ICVVN060625

Tgt Ion: 91 Resp: 473629
Ion Ratio Lower Upper
91 100
126 32.8 16.2 48.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#80
1,3,5-Trimethylbenzene
Concen: 52.787 ug/l
RT: 13.171 min Scan# 1907
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54


Tgt Ion:105 Resp: 548347
Ion Ratio Lower Upper
105 100
120 50.6 24.9 74.6





#81
trans-1,4-Dichloro-2-butene
Concen: 55.702 ug/l
RT: 12.735 min Scan# 1833
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54




Instrument :
MSVOA_N
ClientSampleId :
ICVVN060625



Tgt Ion: 75 Resp: 9934
Ion Ratio Lower Upper
75 100
53 83.7 66.6 99.8
89 40.4 36.7 55.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#82
4-Chlorotoluene
Concen: 51.884 ug/l
RT: 13.223 min Scan# 1916
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54

Tgt Ion: 91 Resp: 481403
Ion Ratio Lower Upper
91 100
126 33.1 16.7 50.1





#83

tert-Butylbenzene

Concen: 52.532 ug/l

RT: 13.441 min Scan# 1952

Delta R.T. 0.006 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54

Instrument :

MSVOA_N

ClientSampleId :

ICVVN060625

Tgt Ion:119 Resp: 499519

Ion Ratio Lower Upper

119 100

91 63.9 32.1 96.5





134 26.4 13.4 40.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#84

1,2,4-Trimethylbenzene

Concen: 53.024 ug/l

RT: 13.482 min Scan# 1960

Delta R.T. -0.000 min

Lab File: VN086869.D




Acq: 06 Jun 2025 15:54


Tgt Ion:105 Resp: 552333

Ion Ratio Lower Upper

105 100


120 46.6 23.2 69.6





#85
sec-Butylbenzene
Concen: 52.428 ug/l
RT: 13.618 min Scan# 1983
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54




Instrument :
MSVOA_N
ClientSampleId :
ICVVN060625



Tgt Ion:105 Resp: 723980
Ion Ratio Lower Upper
105 100
134 20.2 10.0 30.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#86
p-Isopropyltoluene
Concen: 53.174 ug/l
RT: 13.729 min Scan# 2002
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54

Tgt Ion:119 Resp: 606979
Ion Ratio Lower Upper
119 100
134 26.5 13.5 40.4
91 22.7 11.5 34.4





#87

1,3-Dichlorobenzene

Concen: 51.224 ug/l

RT: 13.735 min Scan# 2003

Delta R.T. -0.000 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54

Instrument :

MSVOA_N

ClientSampleId :

ICVVN060625

Tgt Ion:146 Resp: 290810

Ion Ratio Lower Upper

146 100

111 42.1 20.9 62.8





148 64.3 32.0 96.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#88

1,4-Dichlorobenzene

Concen: 51.557 ug/l

RT: 13.812 min Scan# 2016

Delta R.T. -0.000 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54




Tgt Ion:146 Resp: 298412


Ion Ratio Lower Upper

146 100

111 40.0 19.9 59.7

148 63.3 31.8 95.4









#89
n-Butylbenzene
Concen: 52.424 ug/l
RT: 14.059 min Scan# 2058
Delta R.T. 0.006 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54

Instrument :
MSVOA_N
ClientSampleId :
ICVVN060625

Tgt Ion: 91 Resp: 579630
Ion Ratio Lower Upper
91 100
92 55.0 27.5 82.4
134 25.5 13.0 39.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#90
Hexachloroethane
Concen: 52.847 ug/l
RT: 14.335 min Scan# 2105
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54


Tgt Ion:117 Resp: 102254
Ion Ratio Lower Upper
117 100
201 66.4 33.2 99.6





#91
1,2-Dichlorobenzene
Concen: 51.340 ug/l
RT: 14.106 min Scan# 2066
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54




Instrument :
MSVOA_N
ClientSampleId :
ICVVN060625



Tgt Ion:146 Resp: 280025
Ion Ratio Lower Upper
146 100
111 43.5 21.6 64.6
148 63.4 31.7 95.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#92
1,2-Dibromo-3-Chloropropane
Concen: 49.161 ug/l
RT: 14.723 min Scan# 2171
Delta R.T. 0.006 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54

Tgt Ion: 75 Resp: 50115
Ion Ratio Lower Upper
75 100
155 83.5 41.1 123.4
157 108.6 52.7 158.1





#93

1,2,4-Trichlorobenzene

Concen: 51.986 ug/l

RT: 15.394 min Scan# 2284

Delta R.T. 0.006 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54

Instrument :

MSVOA_N

ClientSampleId :

ICVVN060625





Tgt Ion:	180	Resp:	181064
Ion Ratio	Lower	Upper	
180	100		
182	94.6	46.9	140.7
145	30.8	16.0	47.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#94

Hexachlorobutadiene

Concen: 52.843 ug/l




RT: 15.500 min Scan# 2303


Delta R.T. -0.000 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54


Tgt Ion:	225	Resp:	68570
Ion Ratio	Lower	Upper	
225	100		
223	65.3	33.5	100.4
227	62.5	31.1	93.5





#95
Naphthalene
Concen: 52.005 ug/l
RT: 15.641 min Scan# 2327
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54




Instrument :
MSVOA_N
ClientSampleId :
ICVVN060625



Tgt Ion:128 Resp: 67418
Ion Ratio Lower Upper
128 100
127 12.8 10.2 15.2
129 10.9 8.8 13.2




Manual Integrations
APPROVED

Reviewed By :John Carlone 06/09/2025
Supervised By :Mahesh Dadoda 06/09/2025



#96
1,2,3-Trichlorobenzene
Concen: 50.539 ug/l
RT: 15.841 min Scan# 2361
Delta R.T. 0.006 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54

Tgt Ion:180 Resp: 174888
Ion Ratio Lower Upper
180 100
182 95.2 47.0 141.0
145 33.8 16.6 49.8



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060625\
 Data File : VN086869.D
 Acq On : 06 Jun 2025 15:54
 Operator : JC\MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ICVVN060625

Quant Time: Jun 07 02:16:19 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	105	0.00
2 T	Dichlorodifluoromethane	0.499	0.551	-10.4	116	0.00
3 P	Chloromethane	0.644	0.637	1.1	112	0.00
4 C	Vinyl Chloride	0.664	0.697	-5.0#	114	0.00
5 T	Bromomethane	0.372	0.411	-10.5	121	0.00
6 T	Chloroethane	0.429	0.439	-2.3	113	0.00
7 T	Trichlorofluoromethane	0.868	0.899	-3.6	113	0.00
8 T	Diethyl Ether	0.378	0.401	-6.1	116	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.545	0.585	-7.3	118	0.00
10 T	Methyl Iodide	0.707	0.752	-6.4	116	0.00
11 T	Tert butyl alcohol	0.182	0.191	-4.9	114	0.00
12 CM	1,1-Dichloroethene	0.557	0.575	-3.2#	113	0.00
13 T	Acrolein	0.057	0.055	3.5	128	0.00
14 T	Allyl chloride	0.923	0.932	-1.0	113	0.00
15 T	Acrylonitrile	0.425	0.440	-3.5	114	0.00
16 T	Acetone	0.355	0.350	1.4	114	0.00
17 T	Carbon Disulfide	1.539	1.556	-1.1	115	0.00
18 T	Methyl Acetate	1.035	1.074	-3.8	115	0.00
19 T	Methyl tert-butyl Ether	2.015	2.084	-3.4	113	0.00
20 T	Methylene Chloride	0.665	0.660	0.8	115	0.00
21 T	trans-1,2-Dichloroethene	0.619	0.612	1.1	113	0.00
22 T	Diisopropyl ether	1.945	2.018	-3.8	114	0.00
23 T	Vinyl Acetate	1.644	1.677	-2.0	111	0.00
24 P	1,1-Dichloroethane	1.120	1.153	-2.9	114	0.00
25 T	2-Butanone	0.577	0.591	-2.4	113	0.00
26 T	2,2-Dichloropropane	0.871	0.993	-14.0	116	0.00
27 T	cis-1,2-Dichloroethene	0.740	0.756	-2.2	114	0.00
28 T	Bromochloromethane	0.550	0.499	9.3	113	0.00
29 T	Tetrahydrofuran	0.376	0.386	-2.7	113	0.00
30 C	Chloroform	1.118	1.139	-1.9#	113	0.00
31 T	Cyclohexane	1.086	1.090	-0.4	114	0.00
32 T	1,1,1-Trichloroethane	0.951	0.962	-1.2	113	0.00
33 S	1,2-Dichloroethane-d4	0.669	0.698	-4.3	147	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	104	0.00
35 S	Dibromofluoromethane	0.296	0.322	-8.8	153#	0.00
36 T	1,1-Dichloropropene	0.442	0.460	-4.1	115	0.00
37 T	Ethyl Acetate	0.562	0.563	-0.2	109	0.00
38 T	Carbon Tetrachloride	0.433	0.455	-5.1	116	0.00
39 T	Methylcyclohexane	0.605	0.637	-5.3	116	0.00
40 TM	Benzene	1.444	1.471	-1.9	114	0.00
41 T	Methacrylonitrile	0.316	0.332	-5.1	114	0.00
42 TM	1,2-Dichloroethane	0.438	0.451	-3.0	114	0.00
43 T	Isopropyl Acetate	0.902	0.912	-1.1	111	0.00
44 TM	Trichloroethene	0.342	0.355	-3.8	113	0.00
45 C	1,2-Dichloropropane	0.351	0.361	-2.8#	113	0.00
46 T	Dibromomethane	0.233	0.247	-6.0	116	0.00
47 T	Bromodichloromethane	0.480	0.498	-3.8	113	0.00
48 T	Methyl methacrylate	0.415	0.426	-2.7	113	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060625\
 Data File : VN086869.D
 Acq On : 06 Jun 2025 15:54
 Operator : JC\MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ICSVN060625

Quant Time: Jun 07 02:16:19 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.008	0.008	0.0	112	0.00
50 S	Toluene-d8	1.173	1.246	-6.2	151#	0.00
51 T	4-Methyl-2-Pentanone	0.543	0.582	-7.2	113	0.00
52 CM	Toluene	0.882	0.918	-4.1#	114	0.00
53 T	t-1,3-Dichloropropene	0.537	0.570	-6.1	114	0.00
54 T	cis-1,3-Dichloropropene	0.574	0.603	-5.1	114	0.00
55 T	1,1,2-Trichloroethane	0.340	0.350	-2.9	113	0.00
56 T	Ethyl methacrylate	0.541	0.613	-13.3	115	0.00
57 T	1,3-Dichloropropane	0.589	0.605	-2.7	112	0.00
58 T	2-Chloroethyl Vinyl ether	0.323	0.317	1.9	120	0.00
59 T	2-Hexanone	0.350	0.397	-13.4	112	0.00
60 T	Dibromochloromethane	0.354	0.372	-5.1	114	0.00
61 T	1,2-Dibromoethane	0.348	0.361	-3.7	114	0.00
62 S	4-Bromofluorobenzene	0.436	0.471	-8.0	151#	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	105	0.00
64 T	Tetrachloroethene	0.316	0.319	-0.9	114	0.00
65 PM	Chlorobenzene	1.103	1.122	-1.7	115	0.00
66 T	1,1,1,2-Tetrachloroethane	0.354	0.370	-4.5	115	0.00
67 C	Ethyl Benzene	1.899	1.960	-3.2#	114	0.00
68 T	m/p-Xylenes	0.727	0.756	-4.0	113	0.00
69 T	o-Xylene	0.696	0.730	-4.9	114	0.00
70 T	Styrene	1.192	1.257	-5.5	113	0.00
71 P	Bromoform	0.263	0.288	-9.5	114	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	103	0.00
73 T	Isopropylbenzene	3.643	3.775	-3.6	114	0.00
74 T	N-amyl acetate	1.273	1.380	-8.4	112	0.00
75 P	1,1,2,2-Tetrachloroethane	1.234	1.278	-3.6	112	0.00
76 T	1,2,3-Trichloropropane	1.189	1.279	-7.6	112	0.00
77 T	Bromobenzene	0.835	0.874	-4.7	114	0.00
78 T	n-propylbenzene	4.427	4.639	-4.8	114	0.00
79 T	2-Chlorotoluene	2.655	2.742	-3.3	113	0.00
80 T	1,3,5-Trimethylbenzene	3.007	3.175	-5.6	113	0.00
81 T	trans-1,4-Dichloro-2-butene	0.516	0.575	-11.4	125	0.00
82 T	4-Chlorotoluene	2.686	2.787	-3.8	114	0.00
83 T	tert-Butylbenzene	2.753	2.892	-5.0	114	0.00
84 T	1,2,4-Trimethylbenzene	3.016	3.198	-6.0	113	0.00
85 T	sec-Butylbenzene	3.998	4.192	-4.9	113	0.00
86 T	p-Isopropyltoluene	3.305	3.514	-6.3	114	0.00
87 T	1,3-Dichlorobenzene	1.644	1.684	-2.4	112	0.00
88 T	1,4-Dichlorobenzene	1.676	1.728	-3.1	113	0.00
89 T	n-Butylbenzene	3.201	3.356	-4.8	113	0.00
90 T	Hexachloroethane	0.560	0.592	-5.7	114	0.00
91 T	1,2-Dichlorobenzene	1.579	1.621	-2.7	111	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.295	0.290	1.7	110	0.00
93 T	1,2,4-Trichlorobenzene	1.008	1.048	-4.0	112	0.00
94 T	Hexachlorobutadiene	0.376	0.397	-5.6	113	0.00
95 T	Naphthalene	3.753	3.904	-4.0	112	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060625\
Data File : VN086869.D
Acq On : 06 Jun 2025 15:54
Operator : JC\MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ICVVN060625

Quant Time: Jun 07 02:16:19 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 02:12:50 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.002	1.013	-1.1	110	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060625\
 Data File : VN086869.D
 Acq On : 06 Jun 2025 15:54
 Operator : JC\MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ICVVN060625

Quant Time: Jun 07 02:16:19 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	105	0.00
2 T	Dichlorodifluoromethane	50.000	55.286	-10.6	116	0.00
3 P	Chloromethane	50.000	49.505	1.0	112	0.00
4 C	Vinyl Chloride	50.000	52.451	-4.9#	114	0.00
5 T	Bromomethane	50.000	55.292	-10.6	121	0.00
6 T	Chloroethane	50.000	51.117	-2.2	113	0.00
7 T	Trichlorofluoromethane	50.000	51.819	-3.6	113	0.00
8 T	Diethyl Ether	50.000	53.001	-6.0	116	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	53.615	-7.2	118	0.00
10 T	Methyl Iodide	50.000	53.227	-6.5	116	0.00
11 T	Tert butyl alcohol	250.000	262.193	-4.9	114	0.00
12 CM	1,1-Dichloroethene	50.000	51.678	-3.4#	113	0.00
13 T	Acrolein	250.000	240.545	3.8	128	0.00
14 T	Allyl chloride	50.000	50.504	-1.0	113	0.00
15 T	Acrylonitrile	250.000	258.885	-3.6	114	0.00
16 T	Acetone	250.000	246.750	1.3	114	0.00
17 T	Carbon Disulfide	50.000	50.550	-1.1	115	0.00
18 T	Methyl Acetate	50.000	51.912	-3.8	115	0.00
19 T	Methyl tert-butyl Ether	50.000	51.720	-3.4	113	0.00
20 T	Methylene Chloride	50.000	49.617	0.8	115	0.00
21 T	trans-1,2-Dichloroethene	50.000	49.437	1.1	113	0.00
22 T	Diisopropyl ether	50.000	51.853	-3.7	114	0.00
23 T	Vinyl Acetate	250.000	255.055	-2.0	111	0.00
24 P	1,1-Dichloroethane	50.000	51.487	-3.0	114	0.00
25 T	2-Butanone	250.000	255.897	-2.4	113	0.00
26 T	2,2-Dichloropropane	50.000	57.031	-14.1	116	0.00
27 T	cis-1,2-Dichloroethene	50.000	51.073	-2.1	114	0.00
28 T	Bromochloromethane	50.000	45.310	9.4	113	0.00
29 T	Tetrahydrofuran	250.000	256.656	-2.7	113	0.00
30 C	Chloroform	50.000	50.933	-1.9#	113	0.00
31 T	Cyclohexane	50.000	50.212	-0.4	114	0.00
32 T	1,1,1-Trichloroethane	50.000	50.603	-1.2	113	0.00
33 S	1,2-Dichloroethane-d4	50.000	52.135	-4.3	147	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	104	0.00
35 S	Dibromofluoromethane	50.000	54.416	-8.8	153	0.00
36 T	1,1-Dichloropropene	50.000	52.123	-4.2	115	0.00
37 T	Ethyl Acetate	50.000	50.125	-0.3	109	0.00
38 T	Carbon Tetrachloride	50.000	52.466	-4.9	116	0.00
39 T	Methylcyclohexane	50.000	52.686	-5.4	116	0.00
40 TM	Benzene	50.000	50.931	-1.9	114	0.00
41 T	Methacrylonitrile	50.000	52.607	-5.2	114	0.00
42 TM	1,2-Dichloroethane	50.000	51.489	-3.0	114	0.00
43 T	Isopropyl Acetate	50.000	50.536	-1.1	111	0.00
44 TM	Trichloroethene	50.000	51.788	-3.6	113	0.00
45 C	1,2-Dichloropropane	50.000	51.450	-2.9#	113	0.00
46 T	Dibromomethane	50.000	52.849	-5.7	116	0.00
47 T	Bromodichloromethane	50.000	51.824	-3.6	113	0.00
48 T	Methyl methacrylate	50.000	51.328	-2.7	113	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060625\
 Data File : VN086869.D
 Acq On : 06 Jun 2025 15:54
 Operator : JC\MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ICSVN060625

Quant Time: Jun 07 02:16:19 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	1110.123	-11.0	112	0.00
50 S	Toluene-d8	50.000	53.129	-6.3	151	0.00
51 T	4-Methyl-2-Pentanone	250.000	267.835	-7.1	113	0.00
52 CM	Toluene	50.000	52.040	-4.1#	114	0.00
53 T	t-1,3-Dichloropropene	50.000	53.137	-6.3	114	0.00
54 T	cis-1,3-Dichloropropene	50.000	52.519	-5.0	114	0.00
55 T	1,1,2-Trichloroethane	50.000	51.562	-3.1	113	0.00
56 T	Ethyl methacrylate	50.000	56.688	-13.4	115	0.00
57 T	1,3-Dichloropropane	50.000	51.386	-2.8	112	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	246.006	1.6	120	0.00
59 T	2-Hexanone	250.000	283.478	-13.4	112	0.00
60 T	Dibromochloromethane	50.000	52.627	-5.3	114	0.00
61 T	1,2-Dibromoethane	50.000	51.887	-3.8	114	0.00
62 S	4-Bromofluorobenzene	50.000	54.081	-8.2	151	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	105	0.00
64 T	Tetrachloroethene	50.000	50.450	-0.9	114	0.00
65 PM	Chlorobenzene	50.000	50.892	-1.8	115	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	52.164	-4.3	115	0.00
67 C	Ethyl Benzene	50.000	51.590	-3.2#	114	0.00
68 T	m/p-Xylenes	100.000	103.912	-3.9	113	0.00
69 T	o-Xylene	50.000	52.387	-4.8	114	0.00
70 T	Styrene	50.000	52.758	-5.5	113	0.00
71 P	Bromoform	50.000	54.751	-9.5	114	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	103	0.00
73 T	Isopropylbenzene	50.000	51.823	-3.6	114	0.00
74 T	N-ethyl acetate	50.000	54.212	-8.4	112	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	51.763	-3.5	112	0.00
76 T	1,2,3-Trichloropropane	50.000	53.821	-7.6	112	0.00
77 T	Bromobenzene	50.000	52.313	-4.6	114	0.00
78 T	n-propylbenzene	50.000	52.397	-4.8	114	0.00
79 T	2-Chlorotoluene	50.000	51.650	-3.3	113	0.00
80 T	1,3,5-Trimethylbenzene	50.000	52.787	-5.6	113	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	55.702	-11.4	125	0.00
82 T	4-Chlorotoluene	50.000	51.884	-3.8	114	0.00
83 T	tert-Butylbenzene	50.000	52.532	-5.1	114	0.00
84 T	1,2,4-Trimethylbenzene	50.000	53.024	-6.0	113	0.00
85 T	sec-Butylbenzene	50.000	52.428	-4.9	113	0.00
86 T	p-Isopropyltoluene	50.000	53.174	-6.3	114	0.00
87 T	1,3-Dichlorobenzene	50.000	51.224	-2.4	112	0.00
88 T	1,4-Dichlorobenzene	50.000	51.557	-3.1	113	0.00
89 T	n-Butylbenzene	50.000	52.424	-4.8	113	0.00
90 T	Hexachloroethane	50.000	52.847	-5.7	114	0.00
91 T	1,2-Dichlorobenzene	50.000	51.340	-2.7	111	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	49.161	1.7	110	0.00
93 T	1,2,4-Trichlorobenzene	50.000	51.986	-4.0	112	0.00
94 T	Hexachlorobutadiene	50.000	52.843	-5.7	113	0.00
95 T	Naphthalene	50.000	52.005	-4.0	112	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060625\
Data File : VN086869.D
Acq On : 06 Jun 2025 15:54
Operator : JC\MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ICVVN060625

Quant Time: Jun 07 02:16:19 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 02:12:50 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	50.539	-1.1	110	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: Q2263 SAS No.: Q2263 SDG No.: Q2263

Instrument ID: MSVOA_N Calibration Date/Time: 06/11/2025 11:32

Lab File ID: VN086940.D Init. Calib. Date(s): 06/06/2025 06/06/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 12:44 14:49

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.499	0.538		7.82	20
Chloromethane	0.644	0.555	0.1	-13.82	20
Vinyl Chloride	0.664	0.701		5.57	20
Bromomethane	0.372	0.312		-16.13	20
Chloroethane	0.429	0.457		6.53	20
Trichlorofluoromethane	0.868	0.930		7.14	20
1,1,2-Trichlorotrifluoroethane	0.545	0.582		6.79	20
1,1-Dichloroethene	0.557	0.593		6.46	20
Acetone	0.355	0.384		8.17	20
Carbon Disulfide	1.539	1.505		-2.21	20
Methyl tert-butyl Ether	2.015	2.172		7.79	20
Methyl Acetate	1.035	1.085		4.83	20
Methylene Chloride	0.665	0.673		1.2	20
trans-1,2-Dichloroethene	0.619	0.633		2.26	20
1,1-Dichloroethane	1.120	1.178	0.1	5.18	20
Cyclohexane	1.086	1.017		-6.35	20
2-Butanone	0.577	0.568		-1.56	20
Carbon Tetrachloride	0.433	0.472		9.01	20
cis-1,2-Dichloroethene	0.740	0.793		7.16	20
Bromochloromethane	0.550	0.578		5.09	20
Chloroform	1.118	1.176		5.19	20
1,1,1-Trichloroethane	0.951	0.976		2.63	20
Methylcyclohexane	0.605	0.576		-4.79	20
Benzene	1.444	1.554		7.55	20
1,2-Dichloroethane	0.438	0.472		7.76	20
Trichloroethene	0.342	0.382		11.7	20
1,2-Dichloropropane	0.351	0.384		9.4	20
Bromodichloromethane	0.480	0.532		10.83	20
4-Methyl-2-Pentanone	0.543	0.590		8.66	20
Toluene	0.882	0.978		10.88	20
t-1,3-Dichloropropene	0.537	0.609		13.41	20
cis-1,3-Dichloropropene	0.574	0.661		15.16	20
1,1,2-Trichloroethane	0.340	0.383		12.65	20
2-Hexanone	0.350	0.395		12.86	20
Dibromochloromethane	0.354	0.414		16.95	20
1,2-Dibromoethane	0.348	0.385		10.63	20
Tetrachloroethene	0.316	0.337		6.65	20
Chlorobenzene	1.103	1.222	0.3	10.79	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15
 Lab Code: CHEM Case No.: Q2263 SAS No.: Q2263 SDG No.: Q2263
 Instrument ID: MSVOA_N Calibration Date/Time: 06/11/2025 11:32
 Lab File ID: VN086940.D Init. Calib. Date(s): 06/06/2025 06/06/2025
 Heated Purge: (Y/N) N Init. Calib. Time(s): 12:44 14:49
 GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.900	2.059		8.37	20
m/p-Xylenes	0.727	0.802		10.32	20
o-Xylene	0.696	0.782		12.36	20
Styrene	1.192	1.358		13.93	20
Bromoform	0.263	0.319	0.1	21.29	20
Isopropylbenzene	3.643	3.898		7	20
1,1,2,2-Tetrachloroethane	1.234	1.402	0.3	13.61	20
1,3-Dichlorobenzene	1.644	1.843		12.1	20
1,4-Dichlorobenzene	1.676	1.887		12.59	20
1,2-Dichlorobenzene	1.579	1.761		11.53	20
1,2-Dibromo-3-Chloropropane	0.295	0.311		5.42	20
1,2,4-Trichlorobenzene	1.008	1.052		4.36	20
1,2,3-Trichlorobenzene	1.002	0.995		-0.7	20
1,2-Dichloroethane-d4	0.669	0.622		-7.03	20
Dibromofluoromethane	0.296	0.309		4.39	20
Toluene-d8	1.173	1.181		0.68	20
4-Bromofluorobenzene	0.436	0.440		0.92	20

All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061125\
 Data File : VN086940.D
 Acq On : 11 Jun 2025 11:32
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050

Manual Integrations
 APPROVED

Reviewed By : John Carlone 06/12/2025
 Supervised By : Semsettin Yesilyurt 06/12/2025

Quant Time: Jun 12 01:28:53 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	205098	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	357643	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	312173	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	152136	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	127629	46.478	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	92.960%
35) Dibromofluoromethane	8.177	113	110532	52.151	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.300%
50) Toluene-d8	10.565	98	422391	50.342	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.680%
62) 4-Bromofluorobenzene	12.847	95	157388	50.489	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	100.980%

Target Compounds					Qvalue	
2) Dichlorodifluoromethane	2.154	85	110255	53.916	ug/l	100
3) Chloromethane	2.395	50	113731	43.069	ug/l	94
4) Vinyl Chloride	2.554	62	143755	52.773	ug/l	99
5) Bromomethane	2.989	94	63914	41.928	ug/l	96
6) Chloroethane	3.154	64	93702	53.253	ug/l	99
7) Trichlorofluoromethane	3.524	101	190657	53.570	ug/l	99
8) Diethyl Ether	3.983	74	85722	55.281	ug/l	95
9) 1,1,2-Trichlorotrifluo...	4.401	101	119322	53.364	ug/l	99
10) Methyl Iodide	4.618	142	115825	39.959	ug/l	97
11) Tert butyl alcohol	5.542	59	183049	245.667	ug/l	99
12) 1,1-Dichloroethene	4.365	96	121690	53.307	ug/l	94
13) Acrolein	4.201	56	68006	288.340	ug/l	99
14) Allyl chloride	5.048	41	187099	49.420	ug/l	96
15) Acrylonitrile	5.736	53	456903	262.349	ug/l	99
16) Acetone	4.442	43	393802	270.423	ug/l	97
17) Carbon Disulfide	4.736	76	308706	48.889	ug/l	100
18) Methyl Acetate	5.042	43	222452	52.419	ug/l	98
19) Methyl tert-butyl Ether	5.812	73	445459	53.894	ug/l	98
20) Methylene Chloride	5.295	84	138117	50.660	ug/l	95
21) trans-1,2-Dichloroethene	5.806	96	129921	51.154	ug/l	98
22) Diisopropyl ether	6.683	45	412097	51.639	ug/l	99
23) Vinyl Acetate	6.618	43	1766271	261.960	ug/l	98
24) 1,1-Dichloroethane	6.589	63	241575	52.602	ug/l	99
25) 2-Butanone	7.489	43	582343	246.018	ug/l	95
26) 2,2-Dichloropropane	7.506	77	215080	60.206	ug/l	99
27) cis-1,2-Dichloroethene	7.500	96	162658	53.551	ug/l	98
28) Bromochloromethane	7.824	49	118461	52.462	ug/l	95
29) Tetrahydrofuran	7.847	42	380103	246.502	ug/l	96
30) Chloroform	7.977	83	241191	52.589	ug/l	98
31) Cyclohexane	8.265	56	208644	46.846	ug/l	97
32) 1,1,1-Trichloroethane	8.177	97	200139	51.307	ug/l	97
36) 1,1-Dichloropropene	8.377	75	172642	54.664	ug/l	99
37) Ethyl Acetate	7.571	43	211755	52.670	ug/l	99
38) Carbon Tetrachloride	8.371	117	168940	54.485	ug/l	94
39) Methylcyclohexane	9.606	83	206018	47.613	ug/l	96
40) Benzene	8.612	78	555597	53.798	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061125\
 Data File : VN086940.D
 Acq On : 11 Jun 2025 11:32
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025

Quant Time: Jun 12 01:28:53 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M

Quant Title : SW846 8260

QLast Update : Sat Jun 07 02:12:50 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.789	41	115191	51.017	ug/l	93
42) 1,2-Dichloroethane	8.677	62	168630	53.833	ug/l	99
43) Isopropyl Acetate	8.694	43	338399	52.442	ug/l	97
44) Trichloroethene	9.353	130	136748	55.842	ug/l	98
45) 1,2-Dichloropropane	9.624	63	137250	54.626	ug/l	100
46) Dibromomethane	9.712	93	94620	56.697	ug/l	99
47) Bromodichloromethane	9.888	83	190362	55.440	ug/l	100
48) Methyl methacrylate	9.683	41	154138	51.898	ug/l	96
49) 1,4-Dioxane	9.700	88	60969	1128.409	ug/l #	96
51) 4-Methyl-2-Pentanone	10.447	43	1054797	271.529	ug/l	98
52) Toluene	10.630	92	349873	55.435	ug/l	98
53) t-1,3-Dichloropropene	10.835	75	217752	56.714	ug/l	100
54) cis-1,3-Dichloropropene	10.312	75	236296	57.518	ug/l	97
55) 1,1,2-Trichloroethane	11.018	97	137099	56.454	ug/l	99
56) Ethyl methacrylate	10.877	69	231226	59.778	ug/l	95
57) 1,3-Dichloropropane	11.165	76	232746	55.247	ug/l	98
58) 2-Chloroethyl Vinyl ether	10.159	63	665780	288.579	ug/l	99
59) 2-Hexanone	11.200	43	705660	282.011	ug/l	97
60) Dibromochloromethane	11.359	129	148069	58.520	ug/l	99
61) 1,2-Dibromoethane	11.471	107	137779	55.351	ug/l	99
64) Tetrachloroethene	11.106	164	105084	53.190	ug/l	95
65) Chlorobenzene	11.888	112	381598	55.424	ug/l	97
66) 1,1,1,2-Tetrachloroethane	11.959	131	124964	56.463	ug/l	99
67) Ethyl Benzene	11.965	91	642857	54.206	ug/l	99
68) m/p-Xylenes	12.071	106	500895	110.334	ug/l	99
69) o-Xylene	12.400	106	244038	56.127	ug/l	98
70) Styrene	12.412	104	424060	56.996	ug/l	100
71) Bromoform	12.576	173	99616	60.747	ug/l #	100
73) Isopropylbenzene	12.694	105	593080	53.511	ug/l	99
74) N-amyl acetate	12.512	43	208229	53.765	ug/l	94
75) 1,1,2,2-Tetrachloroethane	12.935	83	213329	56.816	ug/l	99
76) 1,2,3-Trichloropropane	12.994	75	179987m	49.771	ug/l	
77) Bromobenzene	12.976	156	146181	57.512	ug/l	98
78) n-propylbenzene	13.035	91	717012	53.234	ug/l	99
79) 2-Chlorotoluene	13.123	91	439522	54.413	ug/l	98
80) 1,3,5-Trimethylbenzene	13.171	105	490727	53.628	ug/l	98
81) trans-1,4-Dichloro-2-b...	12.735	75	93790	59.701	ug/l #	87
82) 4-Chlorotoluene	13.218	91	446237	54.598	ug/l	99
83) tert-Butylbenzene	13.435	119	448106	53.498	ug/l	97
84) 1,2,4-Trimethylbenzene	13.482	105	495936	54.048	ug/l	99
85) sec-Butylbenzene	13.612	105	617792	50.788	ug/l	99
86) p-Isopropyltoluene	13.729	119	524440	52.156	ug/l	98
87) 1,3-Dichlorobenzene	13.729	146	280357	56.062	ug/l	99
88) 1,4-Dichlorobenzene	13.812	146	287122	56.315	ug/l	99
89) n-Butylbenzene	14.053	91	477306	49.007	ug/l	100
90) Hexachloroethane	14.329	117	89886	52.738	ug/l	99
91) 1,2-Dichlorobenzene	14.106	146	267932	55.766	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	14.717	75	47326	52.704	ug/l	92
93) 1,2,4-Trichlorobenzene	15.388	180	160030	52.161	ug/l	99
94) Hexachlorobutadiene	15.500	225	51073	44.682	ug/l	99
95) Naphthalene	15.635	128	624421	54.679	ug/l	100
96) 1,2,3-Trichlorobenzene	15.835	180	151411	49.672	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061125\
Data File : VN086940.D
Acq On : 11 Jun 2025 11:32
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED

Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025

Quant Time: Jun 12 01:28:53 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 02:12:50 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061125\
Data File : VN086940.D
Acq On : 11 Jun 2025 11:32
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1


Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Manual Integrations APPROVED

Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025


Quant Time: Jun 12 01:28:53 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 02:12:50 2025
Response via : Initial Calibration





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.230 min Scan# 1068
Delta R.T. -0.006 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion:168 Resp: 205098
Ion Ratio Lower Upper
168 100
99 57.7 49.1 73.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#2
Dichlorodifluoromethane
Concen: 53.916 ug/l
RT: 2.154 min Scan# 34
Delta R.T. -0.000 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32


Tgt Ion: 85 Resp: 110255
Ion Ratio Lower Upper
85 100
87 32.4 16.2 48.6





#3
Chloromethane
Concen: 43.069 ug/l
RT: 2.395 min Scan# 75
Delta R.T. -0.000 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion: 50 Resp: 11373
Ion Ratio Lower Upper
50 100
52 35.2 25.7 38.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#4
Vinyl Chloride
Concen: 52.773 ug/l
RT: 2.554 min Scan# 102
Delta R.T. -0.000 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32


Tgt Ion: 62 Resp: 143755
Ion Ratio Lower Upper
62 100
64 31.8 26.0 39.0





#5
Bromomethane
Concen: 41.928 ug/l
RT: 2.989 min Scan# 176
Delta R.T. -0.006 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion: 94 Resp: 63914
Ion Ratio Lower Upper
94 100
96 93.9 77.9 116.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#6
Chloroethane
Concen: 53.253 ug/l
RT: 3.154 min Scan# 204
Delta R.T. -0.000 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32


Tgt Ion: 64 Resp: 93702
Ion Ratio Lower Upper
64 100
66 31.2 24.6 36.8





#7
Trichlorofluoromethane
Concen: 53.570 ug/l
RT: 3.524 min Scan# 20
Delta R.T. -0.000 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion:101 Resp: 19065
Ion Ratio Lower Upper
101 100
103 64.5 52.2 78.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#8
Diethyl Ether
Concen: 55.281 ug/l
RT: 3.983 min Scan# 345
Delta R.T. -0.000 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32

Tgt Ion: 74 Resp: 85722
Ion Ratio Lower Upper
74 100
45 85.9 45.5 136.5









#9
1,1,2-Trichlorotrifluoroethane
Concen: 53.364 ug/l
RT: 4.401 min Scan# 415
Delta R.T. 0.006 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Tgt Ion:101 Resp: 11932
Ion Ratio Lower Upper
101 100
85 42.3 33.8 50.8
151 74.7 58.7 88.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#10
Methyl Iodide
Concen: 39.959 ug/l
RT: 4.618 min Scan# 453
Delta R.T. 0.012 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32

Tgt Ion:142 Resp: 115825
Ion Ratio Lower Upper
142 100
127 36.0 30.9 46.3
141 13.1 10.4 15.6





#11

Tert butyl alcohol

Concen: 245.667 ug/l

RT: 5.542 min Scan# 611

Delta R.T. -0.006 min

Lab File: VN086940.D

Acq: 11 Jun 2025 11:32

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion: 59 Resp: 183049

Ion Ratio Lower Upper

59 100





57 9.9 8.1 12.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025



#12

1,1-Dichloroethene

Concen: 53.307 ug/l

RT: 4.365 min Scan# 410

Delta R.T. 0.006 min

Lab File: VN086940.D

Acq: 11 Jun 2025 11:32




Tgt Ion: 96 Resp: 121690


Ion Ratio Lower Upper

96 100

61 148.3 125.8 188.6


98 62.3 52.5 78.7





#13
Acrolein
Concen: 288.340 ug/l
RT: 4.201 min Scan# 382
Delta R.T. -0.000 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion: 56 Resp: 68000
Ion Ratio Lower Upper
56 100
55 70.2 55.6 83.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#14
Allyl chloride
Concen: 49.420 ug/l
RT: 5.048 min Scan# 526
Delta R.T. 0.006 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32


Tgt Ion: 41 Resp: 187099
Ion Ratio Lower Upper
41 100
39 62.0 51.3 76.9
76 42.2 31.4 47.0





#15
Acrylonitrile
Concen: 262.349 ug/l
RT: 5.736 min Scan# 644
Delta R.T. -0.006 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion: 53 Resp: 456903
Ion Ratio Lower Upper
53 100
52 80.2 65.0 97.6
51 33.6 27.2 40.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#16
Acetone
Concen: 270.423 ug/l
RT: 4.442 min Scan# 423
Delta R.T. -0.006 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32


Tgt Ion: 43 Resp: 393802
Ion Ratio Lower Upper
43 100
58 36.9 28.0 42.0





#17
Carbon Disulfide
Concen: 48.889 ug/l
RT: 4.736 min Scan# 472
Delta R.T. 0.006 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion: 76 Resp: 308700
Ion Ratio Lower Upper
76 100
78 9.2 7.3 10.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#18
Methyl Acetate
Concen: 52.419 ug/l
RT: 5.042 min Scan# 525
Delta R.T. -0.000 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32

Tgt Ion: 43 Resp: 222452
Ion Ratio Lower Upper
43 100
74 26.7 20.7 31.1





#19

Methyl tert-butyl Ether

Concen: 53.894 ug/l

RT: 5.812 min Scan# 656

Delta R.T. -0.000 min

Lab File: VN086940.D

Acq: 11 Jun 2025 11:32

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion: 73 Resp: 445458

Ion Ratio Lower Upper

73 100





57 22.7 17.4 26.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025



#20

Methylene Chloride

Concen: 50.660 ug/l

RT: 5.295 min Scan# 568

Delta R.T. -0.000 min

Lab File: VN086940.D

Acq: 11 Jun 2025 11:32

Tgt Ion: 84 Resp: 138117




Ion Ratio Lower Upper


84 100

49 106.1 90.5 135.7

51 34.5 28.5 42.7

86 62.4 48.1 72.1





#21

trans-1,2-Dichloroethene

Concen: 51.154 ug/l

RT: 5.806 min Scan# 61

Delta R.T. -0.000 min

Lab File: VN086940.D

Acq: 11 Jun 2025 11:32

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion: 96 Resp: 12992

Ion Ratio Lower Upper

96 100

61 132.3 109.2 163.8





98 62.4 49.8 74.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025



#22

Diisopropyl ether

Concen: 51.639 ug/l

RT: 6.683 min Scan# 804

Delta R.T. -0.006 min

Lab File: VN086940.D

Acq: 11 Jun 2025 11:32

Tgt Ion: 45 Resp: 412097




Ion Ratio Lower Upper


45 100

43 52.4 41.9 62.9

87 30.3 22.4 33.6


59 12.3 9.8 14.6





#23
 Vinyl Acetate
 Concen: 261.960 ug/l
 RT: 6.618 min Scan# 793
 Delta R.T. -0.000 min
 Lab File: VN086940.D
 Acq: 11 Jun 2025 11:32




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050




Tgt Ion: 43 Resp: 176627
 Ion Ratio Lower Upper
 43 100
 86 12.8 9.5 14.3

Manual Integrations
 APPROVED



Reviewed By :John Carlone 06/12/2025
 Supervised By :Semsettin Yesilyurt 06/12/2025




#24
 1,1-Dichloroethane
 Concen: 52.602 ug/l
 RT: 6.589 min Scan# 788
 Delta R.T. 0.006 min
 Lab File: VN086940.D
 Acq: 11 Jun 2025 11:32




Tgt Ion: 63 Resp: 241575
 Ion Ratio Lower Upper
 63 100
 98 7.4 3.6 10.8
 100 4.5 2.6 7.8





#25
2-Butanone
Concen: 246.018 ug/l
RT: 7.489 min Scan# 941
Delta R.T. -0.006 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050




Tgt Ion: 43 Resp: 58234
Ion Ratio Lower Upper
43 100
72 31.5 23.0 34.6

Manual Integrations
APPROVED



Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025




#26
2,2-Dichloropropane
Concen: 60.206 ug/l
RT: 7.506 min Scan# 944
Delta R.T. 0.006 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32




Tgt Ion: 77 Resp: 215080
Ion Ratio Lower Upper
77 100
97 23.4 11.6 34.7





#27
cis-1,2-Dichloroethene
Concen: 53.551 ug/l
RT: 7.500 min Scan# 943
Delta R.T. -0.000 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion: 96 Resp: 162658
Ion Ratio Lower Upper
96 100
61 136.9 0.0 278.0
98 65.1 0.0 128.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#28
Bromochloromethane
Concen: 52.462 ug/l
RT: 7.824 min Scan# 998
Delta R.T. -0.000 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32


Tgt Ion: 49 Resp: 118461
Ion Ratio Lower Upper
49 100
129 2.2 0.0 3.8
130 82.4 62.2 93.2





#29
Tetrahydrofuran
Concen: 246.502 ug/l
RT: 7.847 min Scan# 1002
Delta R.T. -0.006 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion: 42 Resp: 38010
Ion Ratio Lower Upper
42 100
72 50.1 37.8 56.6
71 47.0 35.6 53.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#30
Chloroform
Concen: 52.589 ug/l
RT: 7.977 min Scan# 1024
Delta R.T. -0.000 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32


Tgt Ion: 83 Resp: 241191
Ion Ratio Lower Upper
83 100
85 64.5 52.9 79.3





#31
Cyclohexane
Concen: 46.846 ug/l
RT: 8.265 min Scan# 1073
Delta R.T. -0.006 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion: 56 Resp: 208644
Ion Ratio Lower Upper
56 100
69 30.7 25.4 38.0
84 88.6 68.8 103.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#32
1,1,1-Trichloroethane
Concen: 51.307 ug/l
RT: 8.177 min Scan# 1058
Delta R.T. -0.000 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32

Tgt Ion: 97 Resp: 200139
Ion Ratio Lower Upper
97 100
99 62.4 52.0 78.0
61 43.5 35.9 53.9





#33

1,2-Dichloroethane-d4

Concen: 46.478 ug/l

RT: 8.583 min Scan# 1127

Delta R.T. -0.006 min

Lab File: VN086940.D

Acq: 11 Jun 2025 11:32

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion: 65 Resp: 127629

Ion Ratio Lower Upper

65 100





67 55.0 0.0 105.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.106 min Scan# 1216

Delta R.T. -0.000 min

Lab File: VN086940.D

Acq: 11 Jun 2025 11:32




Tgt Ion:114 Resp: 357643


Ion Ratio Lower Upper

114 100

63 20.1 0.0 39.6

88 16.2 0.0 30.2





#35

Dibromofluoromethane

Concen: 52.151 ug/l

RT: 8.177 min Scan# 1057

Delta R.T. 0.006 min

Lab File: VN086940.D

Acq: 11 Jun 2025 11:32

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion:113 Resp: 11053

Ion Ratio Lower Upper

113 100

111 101.8 84.2 126.2





192 17.9 14.2 21.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025



#36

1,1-Dichloropropene

Concen: 54.664 ug/l

RT: 8.377 min Scan# 1092

Delta R.T. -0.000 min

Lab File: VN086940.D

Acq: 11 Jun 2025 11:32




Tgt Ion: 75 Resp: 172642


Ion Ratio Lower Upper

75 100

110 35.6 17.6 52.9

77 31.1 25.3 37.9





#37

Ethyl Acetate

Concen: 52.670 ug/l

RT: 7.571 min Scan# 91

Delta R.T. -0.006 min

Lab File: VN086940.D


Acq: 11 Jun 2025 11:32

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050



Tgt Ion: 43 Resp: 21175

Ion Ratio Lower Upper

43 100

61 15.2 12.1 18.1




70 12.8 9.8 14.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025



#38

Carbon Tetrachloride

Concen: 54.485 ug/l

RT: 8.371 min Scan# 1091

Delta R.T. -0.000 min

Lab File: VN086940.D

Acq: 11 Jun 2025 11:32




Tgt Ion:117 Resp: 168940


Ion Ratio Lower Upper

117 100

119 100.6 75.0 112.6


121 30.4 24.9 37.3





#39
Methylcyclohexane
Concen: 47.613 ug/l
RT: 9.606 min Scan# 1301
Delta R.T. -0.000 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion: 83 Resp: 206018
Ion Ratio Lower Upper
83 100
55 71.2 61.0 91.6
98 48.5 38.9 58.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#40
Benzene
Concen: 53.798 ug/l
RT: 8.612 min Scan# 1132
Delta R.T. -0.000 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32


Tgt Ion: 78 Resp: 555597
Ion Ratio Lower Upper
78 100
77 24.3 18.7 28.1





#41
Methacrylonitrile
Concen: 51.017 ug/l
RT: 7.789 min Scan# 992
Delta R.T. -0.000 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion: 41 Resp: 11519
Ion Ratio Lower Upper
41 100
39 53.8 41.2 61.8
67 85.5 61.4 92.2
52 30.7 22.7 34.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#42
1,2-Dichloroethane
Concen: 53.833 ug/l
RT: 8.677 min Scan# 1143
Delta R.T. -0.000 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32


Tgt Ion: 62 Resp: 168630
Ion Ratio Lower Upper
62 100
98 11.0 0.0 21.2





#43
Isopropyl Acetate
Concen: 52.442 ug/l
RT: 8.694 min Scan# 1146
Delta R.T. -0.000 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion: 43 Resp: 338399
Ion Ratio Lower Upper
43 100
61 29.2 22.1 33.1
87 15.7 11.8 17.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#44
Trichloroethene
Concen: 55.842 ug/l
RT: 9.353 min Scan# 1258
Delta R.T. -0.006 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32

Tgt Ion:130 Resp: 136748
Ion Ratio Lower Upper
130 100
95 96.1 0.0 196.2





#45

1,2-Dichloropropane

Concen: 54.626 ug/l

RT: 9.624 min Scan# 1304

Delta R.T. -0.006 min

Lab File: VN086940.D

Acq: 11 Jun 2025 11:32

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion: 63 Resp: 137250

Ion Ratio Lower Upper

63 100





65 32.0 25.5 38.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025



#46

Dibromomethane

Concen: 56.697 ug/l

RT: 9.712 min Scan# 1319

Delta R.T. -0.006 min

Lab File: VN086940.D

Acq: 11 Jun 2025 11:32




Tgt Ion: 93 Resp: 94620


Ion Ratio Lower Upper

93 100

95 82.9 67.1 100.7

174 94.1 74.2 111.4





#47

Bromodichloromethane

Concen: 55.440 ug/l

RT: 9.888 min Scan# 1349

Delta R.T. -0.006 min

Lab File: VN086940.D

Acq: 11 Jun 2025 11:32

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion: 83 Resp: 190362

Ion Ratio Lower Upper

83 100

85 63.3 50.9 76.3





127 7.9 6.6 10.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025



#48

Methyl methacrylate

Concen: 51.898 ug/l

RT: 9.683 min Scan# 1314

Delta R.T. -0.000 min

Lab File: VN086940.D

Acq: 11 Jun 2025 11:32




Tgt Ion: 41 Resp: 154138


Ion Ratio Lower Upper

41 100

69 98.4 74.6 112.0

39 49.7 40.2 60.4





#49

1,4-Dioxane

Concen: 1128.409 ug/l

RT: 9.700 min Scan# 1317

Delta R.T. -0.000 min

Lab File: VN086940.D

Acq: 11 Jun 2025 11:32

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion: 88 Resp: 60969

Ion Ratio Lower Upper

88 100

43 0.0 0.0 0.0





58 68.8 57.7 86.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025



#50

Toluene-d8

Concen: 50.342 ug/l

RT: 10.565 min Scan# 1464

Delta R.T. -0.006 min

Lab File: VN086940.D




Acq: 11 Jun 2025 11:32


Tgt Ion: 98 Resp: 422391

Ion Ratio Lower Upper

98 100


100 68.2 53.4 80.0





#51
4-Methyl-2-Pentanone
Concen: 271.529 ug/l
RT: 10.447 min Scan# 1444
Delta R.T. -0.000 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion: 43 Resp: 105479
Ion Ratio Lower Upper
43 100
58 44.3 34.2 51.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#52
Toluene
Concen: 55.435 ug/l
RT: 10.630 min Scan# 1475
Delta R.T. -0.006 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32


Tgt Ion: 92 Resp: 349873
Ion Ratio Lower Upper
92 100
91 167.2 136.5 204.7





#53
t-1,3-Dichloropropene
Concen: 56.714 ug/l
RT: 10.835 min Scan# 11
Delta R.T. -0.006 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion: 75 Resp: 217752
Ion Ratio Lower Upper
75 100
77 32.7 26.4 39.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#54
cis-1,3-Dichloropropene
Concen: 57.518 ug/l
RT: 10.312 min Scan# 1421
Delta R.T. -0.006 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32


Tgt Ion: 75 Resp: 236296
Ion Ratio Lower Upper
75 100
77 31.6 26.2 39.4
39 38.5 32.3 48.5





#55
1,1,2-Trichloroethane
Concen: 56.454 ug/l
RT: 11.018 min Scan# 1541
Delta R.T. -0.000 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion: 97 Resp: 137099
Ion Ratio Lower Upper
97 100
83 86.6 69.7 104.5
85 54.1 45.1 67.7
99 62.2 50.6 75.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#56
Ethyl methacrylate
Concen: 59.778 ug/l
RT: 10.877 min Scan# 1517
Delta R.T. -0.006 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32


Tgt Ion: 69 Resp: 231226
Ion Ratio Lower Upper
69 100
41 55.6 47.2 70.8
39 26.0 23.0 34.6





#57
1,3-Dichloropropane
Concen: 55.247 ug/l
RT: 11.165 min Scan# 1566
Delta R.T. -0.000 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion: 76 Resp: 232740
Ion Ratio Lower Upper
76 100
78 32.8 25.3 37.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#58
2-Chloroethyl Vinyl ether
Concen: 288.579 ug/l
RT: 10.159 min Scan# 1395
Delta R.T. -0.006 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32


Tgt Ion: 63 Resp: 665780
Ion Ratio Lower Upper
63 100
106 29.5 23.0 34.6





#59
2-Hexanone
Concen: 282.011 ug/l
RT: 11.200 min Scan# 1572
Delta R.T. -0.000 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion: 43 Resp: 705660
Ion Ratio Lower Upper
43 100
58 63.5 30.4 91.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#60
Dibromochloromethane
Concen: 58.520 ug/l
RT: 11.359 min Scan# 1599
Delta R.T. -0.000 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32


Tgt Ion:129 Resp: 148069
Ion Ratio Lower Upper
129 100
127 76.4 38.6 115.8





#61
1,2-Dibromoethane
Concen: 55.351 ug/l
RT: 11.471 min Scan# 1618
Delta R.T. -0.000 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion:107 Resp: 137779
Ion Ratio Lower Upper
107 100
109 96.1 76.3 114.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#62
4-Bromofluorobenzene
Concen: 50.489 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. -0.000 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32

Tgt Ion: 95 Resp: 157388
Ion Ratio Lower Upper
95 100
174 71.7 0.0 141.8
176 68.2 0.0 132.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 1685
Delta R.T. -0.000 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Tgt Ion:117 Resp: 31217
Ion Ratio Lower Upper
117 100
82 53.3 44.6 67.0
119 31.8 25.5 38.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#64
Tetrachloroethene
Concen: 53.190 ug/l
RT: 11.106 min Scan# 1556
Delta R.T. -0.000 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32


Tgt Ion:164 Resp: 105084
Ion Ratio Lower Upper
164 100
166 124.8 101.8 152.8
129 95.6 83.0 124.4
131 93.8 78.2 117.2





#65
Chlorobenzene
Concen: 55.424 ug/l
RT: 11.888 min Scan# 1689
Delta R.T. -0.006 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion:112 Resp: 381598
Ion Ratio Lower Upper
112 100
114 31.1 26.2 39.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#66
1,1,1,2-Tetrachloroethane
Concen: 56.463 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. -0.006 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32


Tgt Ion:131 Resp: 124964
Ion Ratio Lower Upper
131 100
133 95.4 48.3 144.8
119 68.5 34.0 101.8





#67
Ethyl Benzene
Concen: 54.206 ug/l
RT: 11.965 min Scan# 1702
Delta R.T. -0.000 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion: 91 Resp: 64285
Ion Ratio Lower Upper
91 100
106 32.5 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#68
m/p-Xylenes
Concen: 110.334 ug/l
RT: 12.071 min Scan# 1720
Delta R.T. -0.000 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32


Tgt Ion:106 Resp: 500895
Ion Ratio Lower Upper
106 100
91 198.0 159.4 239.0





#69
o-Xylene
Concen: 56.127 ug/l
RT: 12.400 min Scan# 1776
Delta R.T. -0.000 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion:106 Resp: 244038
Ion Ratio Lower Upper
106 100
91 208.5 106.1 318.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#70
Styrene
Concen: 56.996 ug/l
RT: 12.412 min Scan# 1778
Delta R.T. -0.000 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32


Tgt Ion:104 Resp: 424060
Ion Ratio Lower Upper
104 100
78 47.4 38.5 57.7
103 53.9 43.1 64.7





#71
Bromoform
Concen: 60.747 ug/l
RT: 12.576 min Scan# 11
Delta R.T. -0.006 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050




Tgt Ion:173 Resp: 99610
Ion Ratio Lower Upper
173 100
175 48.1 24.1 72.4
254 0.0 0.0 0.0

Manual Integrations
APPROVED



Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.788 min Scan# 2012
Delta R.T. -0.006 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32




Tgt Ion:152 Resp: 152136
Ion Ratio Lower Upper
152 100
115 60.1 30.1 90.5
150 176.5 0.0 345.0





#73
Isopropylbenzene
Concen: 53.511 ug/l
RT: 12.694 min Scan# 1826
Delta R.T. -0.000 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion: 105 Resp: 593080
Ion Ratio Lower Upper
105 100
120 27.5 13.5 40.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#74
N-ethyl acetate
Concen: 53.765 ug/l
RT: 12.512 min Scan# 1795
Delta R.T. -0.000 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32


Tgt Ion: 43 Resp: 208229
Ion Ratio Lower Upper
43 100
70 52.6 37.6 56.4
55 26.6 19.0 28.6
61 26.4 20.8 31.2





#75
1,1,2,2-Tetrachloroethane
Concen: 56.816 ug/l
RT: 12.935 min Scan# 1867
Delta R.T. -0.006 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion: 83 Resp: 213329
Ion Ratio Lower Upper
83 100
131 9.8 4.9 14.7
85 64.5 32.0 96.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#76
1,2,3-Trichloropropane
Concen: 49.771 ug/l m
RT: 12.994 min Scan# 1877
Delta R.T. -0.000 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32

Tgt Ion: 75 Resp: 179987
Ion Ratio Lower Upper
75 100
77 187.3 79.3 237.9









#77
Bromobenzene
Concen: 57.512 ug/l
RT: 12.976 min Scan# 1874
Delta R.T. -0.006 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Tgt Ion:156 Resp: 14618
Ion Ratio Lower Upper
156 100
77 230.6 117.7 353.1
158 97.1 48.1 144.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#78
n-propylbenzene
Concen: 53.234 ug/l
RT: 13.035 min Scan# 1884
Delta R.T. -0.000 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32


Tgt Ion: 91 Resp: 717012
Ion Ratio Lower Upper
91 100
120 23.3 11.4 34.2





#79
2-Chlorotoluene
Concen: 54.413 ug/l
RT: 13.123 min Scan# 1899
Delta R.T. -0.000 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion: 91 Resp: 43952
Ion Ratio Lower Upper
91 100
126 33.4 16.2 48.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#80
1,3,5-Trimethylbenzene
Concen: 53.628 ug/l
RT: 13.171 min Scan# 1907
Delta R.T. -0.000 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32


Tgt Ion:105 Resp: 490727
Ion Ratio Lower Upper
105 100
120 50.8 24.9 74.6





#81
trans-1,4-Dichloro-2-butene
Concen: 59.701 ug/l
RT: 12.735 min Scan# 1833
Delta R.T. -0.000 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion: 75 Resp: 93790
Ion Ratio Lower Upper
75 100
53 74.1 66.6 99.8
89 57.1 36.7 55.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#82
4-Chlorotoluene
Concen: 54.598 ug/l
RT: 13.218 min Scan# 1915
Delta R.T. -0.006 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32


Tgt Ion: 91 Resp: 446237
Ion Ratio Lower Upper
91 100
126 34.0 16.7 50.1





#83
tert-Butylbenzene
Concen: 53.498 ug/l
RT: 13.435 min Scan# 1952
Delta R.T. -0.000 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion:119 Resp: 448100
Ion Ratio Lower Upper
119 100
91 61.2 32.1 96.5
134 26.3 13.4 40.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#84
1,2,4-Trimethylbenzene
Concen: 54.048 ug/l
RT: 13.482 min Scan# 1960
Delta R.T. -0.000 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32

Tgt Ion:105 Resp: 495936
Ion Ratio Lower Upper
105 100
120 47.2 23.2 69.6









#85
sec-Butylbenzene
Concen: 50.788 ug/l
RT: 13.612 min Scan# 1983
Delta R.T. -0.006 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Tgt Ion:105 Resp: 617792
Ion Ratio Lower Upper
105 100
134 20.6 10.0 30.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#86
p-Isopropyltoluene
Concen: 52.156 ug/l
RT: 13.729 min Scan# 2002
Delta R.T. -0.000 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32

Tgt Ion:119 Resp: 524440
Ion Ratio Lower Upper
119 100
134 26.4 13.5 40.4
91 21.8 11.5 34.4





#87

1,3-Dichlorobenzene

Concen: 56.062 ug/l

RT: 13.729 min Scan# 2002

Delta R.T. -0.006 min

Lab File: VN086940.D

Acq: 11 Jun 2025 11:32

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion:146 Resp: 28035

Ion Ratio Lower Upper

146 100

111 40.9 20.9 62.8





148 63.8 32.0 96.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025



#88

1,4-Dichlorobenzene

Concen: 56.315 ug/l

RT: 13.812 min Scan# 2016

Delta R.T. -0.000 min

Lab File: VN086940.D

Acq: 11 Jun 2025 11:32




Tgt Ion:146 Resp: 287122


Ion Ratio Lower Upper

146 100

111 39.5 19.9 59.7

148 62.7 31.8 95.4





#89

n-Butylbenzene

Concen: 49.007 ug/l

RT: 14.053 min Scan# 2057

Delta R.T. -0.000 min

Lab File: VN086940.D

Acq: 11 Jun 2025 11:32

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion: 91 Resp: 477300

Ion Ratio Lower Upper

91 100

92 54.9 27.5 82.4





134 26.6 13.0 39.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025



#90

Hexachloroethane

Concen: 52.738 ug/l

RT: 14.329 min Scan# 2104

Delta R.T. -0.006 min

Lab File: VN086940.D




Acq: 11 Jun 2025 11:32


Tgt Ion:117 Resp: 89886

Ion Ratio Lower Upper

117 100

201 67.6 33.2 99.6









#91
1,2-Dichlorobenzene
Concen: 55.766 ug/l
RT: 14.106 min Scan# 2066
Delta R.T. -0.000 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Tgt Ion:146 Resp: 267932
Ion Ratio Lower Upper
146 100
111 43.2 21.6 64.6
148 63.3 31.7 95.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#92
1,2-Dibromo-3-Chloropropane
Concen: 52.704 ug/l
RT: 14.717 min Scan# 2170
Delta R.T. -0.000 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32


Tgt Ion: 75 Resp: 47326
Ion Ratio Lower Upper
75 100
155 87.3 41.1 123.4
157 114.9 52.7 158.1





#93
1,2,4-Trichlorobenzene
Concen: 52.161 ug/l
RT: 15.388 min Scan# 210
Delta R.T. -0.000 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion:180 Resp: 160030
Ion Ratio Lower Upper
180 100
182 92.1 46.9 140.7
145 31.7 16.0 47.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#94
Hexachlorobutadiene
Concen: 44.682 ug/l
RT: 15.500 min Scan# 2303
Delta R.T. -0.000 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32


Tgt Ion:225 Resp: 51073
Ion Ratio Lower Upper
225 100
223 65.0 33.5 100.4
227 62.2 31.1 93.5





#95
Naphthalene
Concen: 54.679 ug/l
RT: 15.635 min Scan# 2327
Delta R.T. -0.006 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion:128 Resp: 62442
Ion Ratio Lower Upper
128 100
127 12.6 10.2 15.2
129 10.8 8.8 13.2




Manual Integrations
APPROVED

Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#96
1,2,3-Trichlorobenzene
Concen: 49.672 ug/l
RT: 15.835 min Scan# 2360
Delta R.T. -0.000 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32

Tgt Ion:180 Resp: 151411
Ion Ratio Lower Upper
180 100
182 94.5 47.0 141.0
145 33.0 16.6 49.8



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061125\
 Data File : VN086940.D
 Acq On : 11 Jun 2025 11:32
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 12 01:28:53 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	99	0.00
2 T	Dichlorodifluoromethane	0.499	0.538	-7.8	106	0.00
3 P	Chloromethane	0.644	0.555	13.8	92	0.00
4 C	Vinyl Chloride	0.664	0.701	-5.6#	108	0.00
5 T	Bromomethane	0.372	0.312	16.1	86	0.00
6 T	Chloroethane	0.429	0.457	-6.5	111	0.00
7 T	Trichlorofluoromethane	0.868	0.930	-7.1	110	0.00
8 T	Diethyl Ether	0.378	0.418	-10.6	114	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.545	0.582	-6.8	111	0.00
10 T	Methyl Iodide	0.707	0.565	20.1	82	0.01
11 T	Tert butyl alcohol	0.182	0.178	2.2	100	0.00
12 CM	1,1-Dichloroethene	0.557	0.593	-6.5#	110	0.00
13 T	Acrolein	0.057	0.066	-15.8	145	0.00
14 T	Allyl chloride	0.923	0.912	1.2	104	0.00
15 T	Acrylonitrile	0.425	0.446	-4.9	108	0.00
16 T	Acetone	0.355	0.384	-8.2	118	0.00
17 T	Carbon Disulfide	1.539	1.505	2.2	104	0.00
18 T	Methyl Acetate	1.035	1.085	-4.8	109	0.00
19 T	Methyl tert-butyl Ether	2.015	2.172	-7.8	111	0.00
20 T	Methylene Chloride	0.665	0.673	-1.2	110	0.00
21 T	trans-1,2-Dichloroethene	0.619	0.633	-2.3	110	0.00
22 T	Diisopropyl ether	1.945	2.009	-3.3	107	0.00
23 T	Vinyl Acetate	1.644	1.722	-4.7	107	0.00
24 P	1,1-Dichloroethane	1.120	1.178	-5.2	110	0.00
25 T	2-Butanone	0.577	0.568	1.6	102	0.00
26 T	2,2-Dichloropropane	0.871	1.049	-20.4	115	0.00
27 T	cis-1,2-Dichloroethene	0.740	0.793	-7.2	112	0.00
28 T	Bromochloromethane	0.550	0.578	-5.1	122	0.00
29 T	Tetrahydrofuran	0.376	0.371	1.3	102	0.00
30 C	Chloroform	1.118	1.176	-5.2#	110	0.00
31 T	Cyclohexane	1.086	1.017	6.4	100	0.00
32 T	1,1,1-Trichloroethane	0.951	0.976	-2.6	108	0.00
33 S	1,2-Dichloroethane-d4	0.669	0.622	7.0	123	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	94	0.00
35 S	Dibromofluoromethane	0.296	0.309	-4.4	133	0.00
36 T	1,1-Dichloropropene	0.442	0.483	-9.3	109	0.00
37 T	Ethyl Acetate	0.562	0.592	-5.3	104	0.00
38 T	Carbon Tetrachloride	0.433	0.472	-9.0	109	0.00
39 T	Methylcyclohexane	0.605	0.576	4.8	95	0.00
40 TM	Benzene	1.444	1.553	-7.5	109	0.00
41 T	Methacrylonitrile	0.316	0.322	-1.9	100	0.00
42 TM	1,2-Dichloroethane	0.438	0.472	-7.8	108	0.00
43 T	Isopropyl Acetate	0.902	0.946	-4.9	105	0.00
44 TM	Trichloroethene	0.342	0.382	-11.7	110	0.00
45 C	1,2-Dichloropropane	0.351	0.384	-9.4#	109	0.00
46 T	Dibromomethane	0.233	0.265	-13.7	113	0.00
47 T	Bromodichloromethane	0.480	0.532	-10.8	110	0.00
48 T	Methyl methacrylate	0.415	0.431	-3.9	103	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061125\
 Data File : VN086940.D
 Acq On : 11 Jun 2025 11:32
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 12 01:28:53 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.008	0.009	-12.5	103	0.00
50 S	Toluene-d8	1.173	1.181	-0.7	130	0.00
51 T	4-Methyl-2-Pentanone	0.543	0.590	-8.7	104	0.00
52 CM	Toluene	0.882	0.978	-10.9#	111	0.00
53 T	t-1,3-Dichloropropene	0.537	0.609	-13.4	110	0.00
54 T	cis-1,3-Dichloropropene	0.574	0.661	-15.2	113	0.00
55 T	1,1,2-Trichloroethane	0.340	0.383	-12.6	112	0.00
56 T	Ethyl methacrylate	0.541	0.647	-19.6	110	0.00
57 T	1,3-Dichloropropane	0.589	0.651	-10.5	110	0.00
58 T	2-Chloroethyl Vinyl ether	0.323	0.372	-15.2	128	0.00
59 T	2-Hexanone	0.350	0.395	-12.9	101	0.00
60 T	Dibromochloromethane	0.354	0.414	-16.9	115	0.00
61 T	1,2-Dibromoethane	0.348	0.385	-10.6	111	0.00
62 S	4-Bromofluorobenzene	0.436	0.440	-0.9	128	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	94	0.00
64 T	Tetrachloroethene	0.316	0.337	-6.6	108	0.00
65 PM	Chlorobenzene	1.103	1.222	-10.8	112	0.00
66 T	1,1,1,2-Tetrachloroethane	0.354	0.400	-13.0	111	0.00
67 C	Ethyl Benzene	1.899	2.059	-8.4#	108	0.00
68 T	m/p-Xylenes	0.727	0.802	-10.3	107	0.00
69 T	o-Xylene	0.696	0.782	-12.4	109	0.00
70 T	Styrene	1.192	1.358	-13.9	110	0.00
71 P	Bromoform	0.263	0.319	-21.3	113	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	91	0.00
73 T	Isopropylbenzene	3.643	3.898	-7.0	103	0.00
74 T	N-amyl acetate	1.273	1.369	-7.5	98	0.00
75 P	1,1,2,2-Tetrachloroethane	1.234	1.402	-13.6	108	0.00
76 T	1,2,3-Trichloropropane	1.189	1.183	0.5	92	0.00
77 T	Bromobenzene	0.835	0.961	-15.1	110	0.00
78 T	n-propylbenzene	4.427	4.713	-6.5	102	0.00
79 T	2-Chlorotoluene	2.655	2.889	-8.8	105	0.00
80 T	1,3,5-Trimethylbenzene	3.007	3.226	-7.3	101	0.00
81 T	trans-1,4-Dichloro-2-butene	0.516	0.616	-19.4	118	0.00
82 T	4-Chlorotoluene	2.686	2.933	-9.2	105	0.00
83 T	tert-Butylbenzene	2.753	2.945	-7.0	102	0.00
84 T	1,2,4-Trimethylbenzene	3.016	3.260	-8.1	102	0.00
85 T	sec-Butylbenzene	3.998	4.061	-1.6	96	0.00
86 T	p-Isopropyltoluene	3.305	3.447	-4.3	98	0.00
87 T	1,3-Dichlorobenzene	1.644	1.843	-12.1	108	0.00
88 T	1,4-Dichlorobenzene	1.676	1.887	-12.6	109	0.00
89 T	n-Butylbenzene	3.201	3.137	2.0	93	0.00
90 T	Hexachloroethane	0.560	0.591	-5.5	100	0.00
91 T	1,2-Dichlorobenzene	1.579	1.761	-11.5	107	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.295	0.311	-5.4	104	0.00
93 T	1,2,4-Trichlorobenzene	1.008	1.052	-4.4	99	0.00
94 T	Hexachlorobutadiene	0.376	0.336	10.6	84	0.00
95 T	Naphthalene	3.753	4.104	-9.4	104	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061125\
Data File : VN086940.D
Acq On : 11 Jun 2025 11:32
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Jun 12 01:28:53 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 02:12:50 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.002	0.995	0.7	95	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061125\
 Data File : VN086940.D
 Acq On : 11 Jun 2025 11:32
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 12 01:28:53 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	99	0.00
2 T	Dichlorodifluoromethane	50.000	53.916	-7.8	106	0.00
3 P	Chloromethane	50.000	43.069	13.9	92	0.00
4 C	Vinyl Chloride	50.000	52.773	-5.5#	108	0.00
5 T	Bromomethane	50.000	41.928	16.1	86	0.00
6 T	Chloroethane	50.000	53.253	-6.5	111	0.00
7 T	Trichlorofluoromethane	50.000	53.570	-7.1	110	0.00
8 T	Diethyl Ether	50.000	55.281	-10.6	114	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	53.364	-6.7	111	0.00
10 T	Methyl Iodide	50.000	39.959	20.1	82	0.01
11 T	Tert butyl alcohol	250.000	245.667	1.7	100	0.00
12 CM	1,1-Dichloroethene	50.000	53.307	-6.6#	110	0.00
13 T	Acrolein	250.000	288.340	-15.3	145	0.00
14 T	Allyl chloride	50.000	49.420	1.2	104	0.00
15 T	Acrylonitrile	250.000	262.349	-4.9	108	0.00
16 T	Acetone	250.000	270.423	-8.2	118	0.00
17 T	Carbon Disulfide	50.000	48.889	2.2	104	0.00
18 T	Methyl Acetate	50.000	52.419	-4.8	109	0.00
19 T	Methyl tert-butyl Ether	50.000	53.894	-7.8	111	0.00
20 T	Methylene Chloride	50.000	50.660	-1.3	110	0.00
21 T	trans-1,2-Dichloroethene	50.000	51.154	-2.3	110	0.00
22 T	Diisopropyl ether	50.000	51.639	-3.3	107	0.00
23 T	Vinyl Acetate	250.000	261.960	-4.8	107	0.00
24 P	1,1-Dichloroethane	50.000	52.602	-5.2	110	0.00
25 T	2-Butanone	250.000	246.018	1.6	102	0.00
26 T	2,2-Dichloropropane	50.000	60.206	-20.4	115	0.00
27 T	cis-1,2-Dichloroethene	50.000	53.551	-7.1	112	0.00
28 T	Bromochloromethane	50.000	52.462	-4.9	122	0.00
29 T	Tetrahydrofuran	250.000	246.502	1.4	102	0.00
30 C	Chloroform	50.000	52.589	-5.2#	110	0.00
31 T	Cyclohexane	50.000	46.846	6.3	100	0.00
32 T	1,1,1-Trichloroethane	50.000	51.307	-2.6	108	0.00
33 S	1,2-Dichloroethane-d4	50.000	46.478	7.0	123	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	94	0.00
35 S	Dibromofluoromethane	50.000	52.151	-4.3	133	0.00
36 T	1,1-Dichloropropene	50.000	54.664	-9.3	109	0.00
37 T	Ethyl Acetate	50.000	52.670	-5.3	104	0.00
38 T	Carbon Tetrachloride	50.000	54.485	-9.0	109	0.00
39 T	Methylcyclohexane	50.000	47.613	4.8	95	0.00
40 TM	Benzene	50.000	53.798	-7.6	109	0.00
41 T	Methacrylonitrile	50.000	51.017	-2.0	100	0.00
42 TM	1,2-Dichloroethane	50.000	53.833	-7.7	108	0.00
43 T	Isopropyl Acetate	50.000	52.442	-4.9	105	0.00
44 TM	Trichloroethene	50.000	55.842	-11.7	110	0.00
45 C	1,2-Dichloropropane	50.000	54.626	-9.3#	109	0.00
46 T	Dibromomethane	50.000	56.697	-13.4	113	0.00
47 T	Bromodichloromethane	50.000	55.440	-10.9	110	0.00
48 T	Methyl methacrylate	50.000	51.898	-3.8	103	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061125\
 Data File : VN086940.D
 Acq On : 11 Jun 2025 11:32
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 12 01:28:53 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	1128.409	-12.8	103	0.00
50 S	Toluene-d8	50.000	50.342	-0.7	130	0.00
51 T	4-Methyl-2-Pentanone	250.000	271.529	-8.6	104	0.00
52 CM	Toluene	50.000	55.435	-10.9#	111	0.00
53 T	t-1,3-Dichloropropene	50.000	56.714	-13.4	110	0.00
54 T	cis-1,3-Dichloropropene	50.000	57.518	-15.0	113	0.00
55 T	1,1,2-Trichloroethane	50.000	56.454	-12.9	112	0.00
56 T	Ethyl methacrylate	50.000	59.778	-19.6	110	0.00
57 T	1,3-Dichloropropane	50.000	55.247	-10.5	110	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	288.579	-15.4	128	0.00
59 T	2-Hexanone	250.000	282.011	-12.8	101	0.00
60 T	Dibromochloromethane	50.000	58.520	-17.0	115	0.00
61 T	1,2-Dibromoethane	50.000	55.351	-10.7	111	0.00
62 S	4-Bromofluorobenzene	50.000	50.489	-1.0	128	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	94	0.00
64 T	Tetrachloroethene	50.000	53.190	-6.4	108	0.00
65 PM	Chlorobenzene	50.000	55.424	-10.8	112	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	56.463	-12.9	111	0.00
67 C	Ethyl Benzene	50.000	54.206	-8.4#	108	0.00
68 T	m/p-Xylenes	100.000	110.334	-10.3	107	0.00
69 T	o-Xylene	50.000	56.127	-12.3	109	0.00
70 T	Styrene	50.000	56.996	-14.0	110	0.00
71 P	Bromoform	50.000	60.747	-21.5	113	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	91	0.00
73 T	Isopropylbenzene	50.000	53.511	-7.0	103	0.00
74 T	N-ethyl acetate	50.000	53.765	-7.5	98	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	56.816	-13.6	108	0.00
76 T	1,2,3-Trichloropropane	50.000	49.771	0.5	92	0.00
77 T	Bromobenzene	50.000	57.512	-15.0	110	0.00
78 T	n-propylbenzene	50.000	53.234	-6.5	102	0.00
79 T	2-Chlorotoluene	50.000	54.413	-8.8	105	0.00
80 T	1,3,5-Trimethylbenzene	50.000	53.628	-7.3	101	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	59.701	-19.4	118	0.00
82 T	4-Chlorotoluene	50.000	54.598	-9.2	105	0.00
83 T	tert-Butylbenzene	50.000	53.498	-7.0	102	0.00
84 T	1,2,4-Trimethylbenzene	50.000	54.048	-8.1	102	0.00
85 T	sec-Butylbenzene	50.000	50.788	-1.6	96	0.00
86 T	p-Isopropyltoluene	50.000	52.156	-4.3	98	0.00
87 T	1,3-Dichlorobenzene	50.000	56.062	-12.1	108	0.00
88 T	1,4-Dichlorobenzene	50.000	56.315	-12.6	109	0.00
89 T	n-Butylbenzene	50.000	49.007	2.0	93	0.00
90 T	Hexachloroethane	50.000	52.738	-5.5	100	0.00
91 T	1,2-Dichlorobenzene	50.000	55.766	-11.5	107	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	52.704	-5.4	104	0.00
93 T	1,2,4-Trichlorobenzene	50.000	52.161	-4.3	99	0.00
94 T	Hexachlorobutadiene	50.000	44.682	10.6	84	0.00
95 T	Naphthalene	50.000	54.679	-9.4	104	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061125\
Data File : VN086940.D
Acq On : 11 Jun 2025 11:32
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Jun 12 01:28:53 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 02:12:50 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	49.672	0.7	95	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: Q2263 SAS No.: Q2263 SDG No.: Q2263

Instrument ID: MSVOA_N Calibration Date/Time: 06/11/2025 21:23

Lab File ID: VN086965.D Init. Calib. Date(s): 06/06/2025 06/06/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 12:44 14:49

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.499	0.552		10.62	50
Chloromethane	0.644	0.581	0.1	-9.78	50
Vinyl Chloride	0.664	0.712		7.23	50
Bromomethane	0.372	0.246		-33.87	50
Chloroethane	0.429	0.477		11.19	50
Trichlorofluoromethane	0.868	0.969		11.64	50
1,1,2-Trichlorotrifluoroethane	0.545	0.587		7.71	50
1,1-Dichloroethene	0.557	0.599		7.54	50
Acetone	0.355	0.373		5.07	50
Carbon Disulfide	1.539	1.513		-1.75	50
Methyl tert-butyl Ether	2.015	2.289		13.6	50
Methyl Acetate	1.035	1.207		16.62	50
Methylene Chloride	0.665	0.707		6.32	50
trans-1,2-Dichloroethene	0.619	0.653		5.49	50
1,1-Dichloroethane	1.120	1.255	0.1	12.05	50
Cyclohexane	1.086	1.022		-5.89	50
2-Butanone	0.577	0.628		8.84	50
Carbon Tetrachloride	0.433	0.462		6.7	50
cis-1,2-Dichloroethene	0.740	0.824		11.35	50
Bromochloromethane	0.550	0.630		14.55	50
Chloroform	1.118	1.234		10.38	50
1,1,1-Trichloroethane	0.951	1.024		7.68	50
Methylcyclohexane	0.605	0.540		-10.74	50
Benzene	1.444	1.551		7.41	50
1,2-Dichloroethane	0.438	0.472		7.76	50
Trichloroethene	0.342	0.366		7.02	50
1,2-Dichloropropane	0.351	0.381		8.55	50
Bromodichloromethane	0.480	0.537		11.88	50
4-Methyl-2-Pentanone	0.543	0.610		12.34	50
Toluene	0.882	0.953		8.05	50
t-1,3-Dichloropropene	0.537	0.588		9.5	50
cis-1,3-Dichloropropene	0.574	0.623		8.54	50
1,1,2-Trichloroethane	0.340	0.379		11.47	50
2-Hexanone	0.350	0.405		15.71	50
Dibromochloromethane	0.354	0.407		14.97	50
1,2-Dibromoethane	0.348	0.391		12.36	50
Tetrachloroethene	0.316	0.320		1.27	50
Chlorobenzene	1.103	1.173	0.3	6.35	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15
 Lab Code: CHEM Case No.: Q2263 SAS No.: Q2263 SDG No.: Q2263
 Instrument ID: MSVOA_N Calibration Date/Time: 06/11/2025 21:23
 Lab File ID: VN086965.D Init. Calib. Date(s): 06/06/2025 06/06/2025
 Heated Purge: (Y/N) N Init. Calib. Time(s): 12:44 14:49
 GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.900	2.012		5.89	50
m/p-Xylenes	0.727	0.775		6.6	50
o-Xylene	0.696	0.766		10.06	50
Styrene	1.192	1.318		10.57	50
Bromoform	0.263	0.313	0.1	19.01	50
Isopropylbenzene	3.643	3.761		3.24	50
1,1,2,2-Tetrachloroethane	1.234	1.414	0.3	14.59	50
1,3-Dichlorobenzene	1.644	1.784		8.52	50
1,4-Dichlorobenzene	1.676	1.781		6.26	50
1,2-Dichlorobenzene	1.579	1.721		8.99	50
1,2-Dibromo-3-Chloropropane	0.295	0.323		9.49	50
1,2,4-Trichlorobenzene	1.008	0.998		-0.99	50
1,2,3-Trichlorobenzene	1.002	0.977		-2.49	50
1,2-Dichloroethane-d4	0.669	0.697		4.18	50
Dibromofluoromethane	0.296	0.314		6.08	50
Toluene-d8	1.173	1.171		-0.17	50
4-Bromofluorobenzene	0.436	0.449		2.98	50

All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061125\
 Data File : VN086965.D
 Acq On : 11 Jun 2025 21:23
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC

Manual Integrations
 APPROVED

Reviewed By : John Carlone 06/12/2025
 Supervised By : Semsettin Yesilyurt 06/12/2025

Quant Time: Jun 12 01:48:05 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	177574	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	326287	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	286630	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	142002	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	123795	52.069	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 104.140%			
35) Dibromofluoromethane	8.177	113	102484	53.000	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 106.000%			
50) Toluene-d8	10.571	98	382080	49.914	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 99.820%			
62) 4-Bromofluorobenzene	12.847	95	146464	51.499	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 103.000%			
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	2.153	85	97946	55.320	ug/l	98
3) Chloromethane	2.395	50	103222	45.148	ug/l	99
4) Vinyl Chloride	2.553	62	126441	53.612	ug/l	100
5) Bromomethane	2.983	94	43618	33.049	ug/l	94
6) Chloroethane	3.147	64	84712	55.606	ug/l	97
7) Trichlorofluoromethane	3.524	101	172024	55.827	ug/l	99
8) Diethyl Ether	3.983	74	75627	56.330	ug/l	99
9) 1,1,2-Trichlorotrifluo...	4.400	101	104252	53.851	ug/l	99
10) Methyl Iodide	4.606	142	46802	18.649	ug/l	98
11) Tert butyl alcohol	5.547	59	180885	280.391	ug/l	99
12) 1,1-Dichloroethene	4.365	96	106380	53.824	ug/l	96
13) Acrolein	4.200	56	152464	746.633	ug/l	100
14) Allyl chloride	5.047	41	166911	50.921	ug/l	97
15) Acrylonitrile	5.741	53	422921	280.476	ug/l	100
16) Acetone	4.447	43	331168	262.661	ug/l	100
17) Carbon Disulfide	4.736	76	268580	49.127	ug/l	99
18) Methyl Acetate	5.047	43	214312	58.328	ug/l	99
19) Methyl tert-butyl Ether	5.818	73	406405	56.790	ug/l	99
20) Methylene Chloride	5.300	84	125496	53.166	ug/l	96
21) trans-1,2-Dichloroethene	5.806	96	115879	52.697	ug/l	95
22) Diisopropyl ether	6.688	45	382028	55.291	ug/l	99
23) Vinyl Acetate	6.618	43	1623707	278.142	ug/l	99
24) 1,1-Dichloroethane	6.588	63	222894	56.057	ug/l	99
25) 2-Butanone	7.494	43	557966	272.256	ug/l	97
26) 2,2-Dichloropropane	7.500	77	149440	48.315	ug/l	98
27) cis-1,2-Dichloroethene	7.500	96	146360	55.654	ug/l	97
28) Bromochloromethane	7.824	49	111849	57.211	ug/l	98
29) Tetrahydrofuran	7.853	42	361046	270.436	ug/l	97
30) Chloroform	7.977	83	219091	55.175	ug/l	97
31) Cyclohexane	8.265	56	181510	47.071	ug/l	98
32) 1,1,1-Trichloroethane	8.177	97	181857	53.846	ug/l	99
36) 1,1-Dichloropropene	8.377	75	153059	53.121	ug/l	99
37) Ethyl Acetate	7.571	43	196171	53.483	ug/l	99
38) Carbon Tetrachloride	8.371	117	150782	53.303	ug/l	96
39) Methylcyclohexane	9.606	83	176079	44.605	ug/l	97
40) Benzene	8.612	78	505922	53.695	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061125\
 Data File : VN086965.D
 Acq On : 11 Jun 2025 21:23
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC

Manual Integrations
 APPROVED

Reviewed By : John Carlone 06/12/2025
 Supervised By : Semsettin Yesilyurt 06/12/2025

Quant Time: Jun 12 01:48:05 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.788	41	112567	54.646	ug/l	97
42) 1,2-Dichloroethane	8.677	62	153959	53.873	ug/l	100
43) Isopropyl Acetate	8.694	43	317059	53.857	ug/l	99
44) Trichloroethene	9.359	130	119433	53.458	ug/l	98
45) 1,2-Dichloropropane	9.624	63	124231	54.196	ug/l	100
46) Dibromomethane	9.712	93	85919	56.431	ug/l	99
47) Bromodichloromethane	9.894	83	175093	55.894	ug/l	99
48) Methyl methacrylate	9.682	41	143734	53.046	ug/l	98
49) 1,4-Dioxane	9.706	88	58859	1194.044	ug/l #	96
51) 4-Methyl-2-Pentanone	10.447	43	995984	281.028	ug/l	99
52) Toluene	10.629	92	310797	53.976	ug/l	99
53) t-1,3-Dichloropropene	10.835	75	191765	54.745	ug/l	96
54) cis-1,3-Dichloropropene	10.318	75	203300	54.242	ug/l	97
55) 1,1,2-Trichloroethane	11.018	97	123723	55.842	ug/l	98
56) Ethyl methacrylate	10.882	69	209075	59.246	ug/l	96
57) 1,3-Dichloropropane	11.165	76	212617	55.319	ug/l	98
58) 2-Chloroethyl Vinyl ether	10.165	63	621634	295.338	ug/l	100
59) 2-Hexanone	11.200	43	661460	289.751	ug/l	94
60) Dibromochloromethane	11.359	129	132918	57.580	ug/l	100
61) 1,2-Dibromoethane	11.470	107	127444	56.119	ug/l	98
64) Tetrachloroethene	11.106	164	91667	50.534	ug/l	95
65) Chlorobenzene	11.894	112	336350	53.206	ug/l	100
66) 1,1,1,2-Tetrachloroethane	11.959	131	113463	55.835	ug/l	100
67) Ethyl Benzene	11.965	91	576593	52.952	ug/l	100
68) m/p-Xylenes	12.070	106	444183	106.561	ug/l	100
69) o-Xylene	12.400	106	219424	54.963	ug/l	98
70) Styrene	12.412	104	377725	55.293	ug/l	99
71) Bromoform	12.576	173	89676	59.559	ug/l #	100
73) Isopropylbenzene	12.694	105	534031	51.622	ug/l	100
74) N-amyl acetate	12.512	43	188759	52.216	ug/l	97
75) 1,1,2,2-Tetrachloroethane	12.935	83	200820	57.301	ug/l	99
76) 1,2,3-Trichloropropane	12.994	75	174388m	51.664	ug/l	
77) Bromobenzene	12.982	156	131869	55.583	ug/l	99
78) n-propylbenzene	13.035	91	643398	51.178	ug/l	99
79) 2-Chlorotoluene	13.123	91	394854	52.371	ug/l	98
80) 1,3,5-Trimethylbenzene	13.170	105	443398	51.914	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.735	75	72331	49.328	ug/l	97
82) 4-Chlorotoluene	13.223	91	406685	53.310	ug/l	100
83) tert-Butylbenzene	13.435	119	396615	50.730	ug/l	97
84) 1,2,4-Trimethylbenzene	13.482	105	447506	52.251	ug/l	99
85) sec-Butylbenzene	13.611	105	549887	48.432	ug/l	100
86) p-Isopropyltoluene	13.729	119	459716	48.982	ug/l	99
87) 1,3-Dichlorobenzene	13.735	146	253292	54.264	ug/l	99
88) 1,4-Dichlorobenzene	13.811	146	252880	53.138	ug/l	99
89) n-Butylbenzene	14.053	91	422802	46.509	ug/l	99
90) Hexachloroethane	14.329	117	83746	52.642	ug/l	98
91) 1,2-Dichlorobenzene	14.106	146	244439	54.507	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	14.717	75	45867	54.724	ug/l	97
93) 1,2,4-Trichlorobenzene	15.388	180	141753	49.501	ug/l	100
94) Hexachlorobutadiene	15.500	225	44722	41.918	ug/l	100
95) Naphthalene	15.641	128	576105	54.049	ug/l	100
96) 1,2,3-Trichlorobenzene	15.841	180	138678	48.742	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061125\
Data File : VN086965.D
Acq On : 11 Jun 2025 21:23
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Manual Integrations
APPROVED

Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025

Quant Time: Jun 12 01:48:05 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 02:12:50 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


```
Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061125\  
Data File : VN086965.D  
Acq On    : 11 Jun 2025  21:23  
Operator  : JC\MD  
Sample    : VSTDCCC050  
Misc      : 5.0mL/MSVOA_N/WATER  
ALS Vial  : 27   Sample Multiplier: 1
```


Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Manual Integrations APPROVED

Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025


Quant Time: Jun 12 01:48:05 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 02:12:50 2025
Response via : Initial Calibration





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.235 min Scan# 1068
Delta R.T. -0.001 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC



Tgt Ion:168 Resp: 177574
Ion Ratio Lower Upper
168 100
99 58.2 49.1 73.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#2
Dichlorodifluoromethane
Concen: 55.320 ug/l
RT: 2.153 min Scan# 34
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23


Tgt Ion: 85 Resp: 97946
Ion Ratio Lower Upper
85 100
87 31.5 16.2 48.6





#3
Chloromethane
Concen: 45.148 ug/l
RT: 2.395 min Scan# 75
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 50 Resp: 10322
Ion Ratio Lower Upper
50 100
52 32.5 25.7 38.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#4
Vinyl Chloride
Concen: 53.612 ug/l
RT: 2.553 min Scan# 102
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23


Tgt Ion: 62 Resp: 126441
Ion Ratio Lower Upper
62 100
64 32.3 26.0 39.0





#5
Bromomethane
Concen: 33.049 ug/l
RT: 2.983 min Scan# 11
Delta R.T. -0.012 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 94 Resp: 43618
Ion Ratio Lower Upper
94 100
96 91.4 77.9 116.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#6
Chloroethane
Concen: 55.606 ug/l
RT: 3.147 min Scan# 203
Delta R.T. -0.006 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23

Tgt Ion: 64 Resp: 84712
Ion Ratio Lower Upper
64 100
66 32.4 24.6 36.8









#7
Trichlorofluoromethane
Concen: 55.827 ug/l
RT: 3.524 min Scan# 267
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Tgt Ion:101 Resp: 172024
Ion Ratio Lower Upper
101 100
103 64.7 52.2 78.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#8
Diethyl Ether
Concen: 56.330 ug/l
RT: 3.983 min Scan# 345
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23

Tgt Ion: 74 Resp: 75627
Ion Ratio Lower Upper
74 100
45 89.7 45.5 136.5









#9
1,1,2-Trichlorotrifluoroethane
Concen: 53.851 ug/l
RT: 4.400 min Scan# 415
Delta R.T. 0.006 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Tgt Ion:101 Resp: 10425
Ion Ratio Lower Upper
101 100
85 41.8 33.8 50.8
151 72.4 58.7 88.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#10
Methyl Iodide
Concen: 18.649 ug/l
RT: 4.606 min Scan# 451
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23

Tgt Ion:142 Resp: 46802
Ion Ratio Lower Upper
142 100
127 39.6 30.9 46.3
141 12.5 10.4 15.6





#11

Tert butyl alcohol

Concen: 280.391 ug/l

RT: 5.547 min Scan# 611

Delta R.T. -0.000 min

Lab File: VN086965.D

Acq: 11 Jun 2025 21:23

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 59 Resp: 18088

Ion Ratio Lower Upper

59 100





57 9.7 8.1 12.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025



#12

1,1-Dichloroethene

Concen: 53.824 ug/l

RT: 4.365 min Scan# 410

Delta R.T. 0.006 min

Lab File: VN086965.D

Acq: 11 Jun 2025 21:23




Tgt Ion: 96 Resp: 106380


Ion Ratio Lower Upper

96 100

61 152.7 125.8 188.6


98 62.5 52.5 78.7





#13
Acrolein
Concen: 746.633 ug/l
RT: 4.200 min Scan# 382
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 56 Resp: 152464
Ion Ratio Lower Upper
56 100
55 69.4 55.6 83.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#14
Allyl chloride
Concen: 50.921 ug/l
RT: 5.047 min Scan# 526
Delta R.T. 0.006 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23


Tgt Ion: 41 Resp: 166911
Ion Ratio Lower Upper
41 100
39 61.7 51.3 76.9
76 40.8 31.4 47.0





#15
Acrylonitrile
Concen: 280.476 ug/l
RT: 5.741 min Scan# 644
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 53 Resp: 422921
Ion Ratio Lower Upper
53 100
52 81.6 65.0 97.6
51 34.0 27.2 40.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#16
Acetone
Concen: 262.661 ug/l
RT: 4.447 min Scan# 424
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23


Tgt Ion: 43 Resp: 331168
Ion Ratio Lower Upper
43 100
58 35.0 28.0 42.0





#17
Carbon Disulfide
Concen: 49.127 ug/l
RT: 4.736 min Scan# 41
Delta R.T. 0.006 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 76 Resp: 268580
Ion Ratio Lower Upper
76 100
78 9.4 7.3 10.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#18
Methyl Acetate
Concen: 58.328 ug/l
RT: 5.047 min Scan# 526
Delta R.T. 0.006 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23

Tgt Ion: 43 Resp: 214312
Ion Ratio Lower Upper
43 100
74 26.6 20.7 31.1





#19

Methyl tert-butyl Ether

Concen: 56.790 ug/l

RT: 5.818 min Scan# 61

Delta R.T. 0.006 min

Lab File: VN086965.D

Acq: 11 Jun 2025 21:23

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 73 Resp: 40640

Ion Ratio Lower Upper

73 100





57 21.3 17.4 26.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025



#20

Methylene Chloride

Concen: 53.166 ug/l

RT: 5.300 min Scan# 569

Delta R.T. 0.006 min

Lab File: VN086965.D

Acq: 11 Jun 2025 21:23

Tgt Ion: 84 Resp: 125496




Ion Ratio Lower Upper


84 100

49 107.5 90.5 135.7

51 34.3 28.5 42.7


86 62.4 48.1 72.1





#21
trans-1,2-Dichloroethene
Concen: 52.697 ug/l
RT: 5.806 min Scan# 61
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 96 Resp: 115879
Ion Ratio Lower Upper
96 100
61 128.3 109.2 163.8
98 60.9 49.8 74.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#22
Diisopropyl ether
Concen: 55.291 ug/l
RT: 6.688 min Scan# 805
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23

Tgt Ion: 45 Resp: 382028
Ion Ratio Lower Upper
45 100
43 52.3 41.9 62.9
87 29.5 22.4 33.6
59 12.9 9.8 14.6









#23
Vinyl Acetate
Concen: 278.142 ug/l
RT: 6.618 min Scan# 793
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 43 Resp: 162370
Ion Ratio Lower Upper
43 100
86 12.1 9.5 14.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#24
1,1-Dichloroethane
Concen: 56.057 ug/l
RT: 6.588 min Scan# 788
Delta R.T. 0.006 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23

Tgt Ion: 63 Resp: 222894
Ion Ratio Lower Upper
63 100
98 7.0 3.6 10.8
100 4.3 2.6 7.8









#25
2-Butanone
Concen: 272.256 ug/l
RT: 7.494 min Scan# 942
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 43 Resp: 557960
Ion Ratio Lower Upper
43 100
72 30.6 23.0 34.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#26
2,2-Dichloropropane
Concen: 48.315 ug/l
RT: 7.500 min Scan# 943
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23

Tgt Ion: 77 Resp: 149440
Ion Ratio Lower Upper
77 100
97 24.1 11.6 34.7





#27

cis-1,2-Dichloroethene

Concen: 55.654 ug/l

RT: 7.500 min Scan# 943

Delta R.T. -0.000 min

Lab File: VN086965.D

Acq: 11 Jun 2025 21:23

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 96 Resp: 146360

Ion Ratio Lower Upper

96 100

61 132.9 0.0 278.0





98 64.2 0.0 128.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025



#28

Bromochloromethane

Concen: 57.211 ug/l

RT: 7.824 min Scan# 998

Delta R.T. -0.000 min

Lab File: VN086965.D

Acq: 11 Jun 2025 21:23




Tgt Ion: 49 Resp: 111849


Ion Ratio Lower Upper

49 100

129 2.0 0.0 3.8

130 79.9 62.2 93.2





#29

Tetrahydrofuran

Concen: 270.436 ug/l

RT: 7.853 min Scan# 1003

Delta R.T. -0.000 min

Lab File: VN086965.D

Acq: 11 Jun 2025 21:23

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 42 Resp: 361040

Ion Ratio Lower Upper

42 100

72 49.5 37.8 56.6





71 46.0 35.6 53.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025



#30

Chloroform

Concen: 55.175 ug/l

RT: 7.977 min Scan# 1024

Delta R.T. -0.000 min

Lab File: VN086965.D




Acq: 11 Jun 2025 21:23


Tgt Ion: 83 Resp: 219091

Ion Ratio Lower Upper

83 100


85 63.6 52.9 79.3





#31
Cyclohexane
Concen: 47.071 ug/l
RT: 8.265 min Scan# 1073
Delta R.T. -0.006 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 56 Resp: 181510
Ion Ratio Lower Upper
56 100
69 31.3 25.4 38.0
84 88.0 68.8 103.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#32
1,1,1-Trichloroethane
Concen: 53.846 ug/l
RT: 8.177 min Scan# 1058
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23

Tgt Ion: 97 Resp: 181857
Ion Ratio Lower Upper
97 100
99 65.4 52.0 78.0
61 44.6 35.9 53.9





#33

1,2-Dichloroethane-d4

Concen: 52.069 ug/l

RT: 8.588 min Scan# 1128

Delta R.T. -0.000 min

Lab File: VN086965.D

Acq: 11 Jun 2025 21:23

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 65 Resp: 12379

Ion Ratio Lower Upper

65 100





67 54.0 0.0 105.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.106 min Scan# 1216

Delta R.T. -0.000 min

Lab File: VN086965.D

Acq: 11 Jun 2025 21:23




Tgt Ion:114 Resp: 326287


Ion Ratio Lower Upper

114 100

63 19.3 0.0 39.6

88 15.8 0.0 30.2





#35

Dibromofluoromethane

Concen: 53.000 ug/l

RT: 8.177 min Scan# 1057

Delta R.T. 0.006 min

Lab File: VN086965.D

Acq: 11 Jun 2025 21:23

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC

Tgt Ion:113 Resp: 102484

Ion Ratio Lower Upper

113 100

111 103.4 84.2 126.2





192 17.7 14.2 21.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025



#36

1,1-Dichloropropene

Concen: 53.121 ug/l

RT: 8.377 min Scan# 1092

Delta R.T. -0.000 min

Lab File: VN086965.D

Acq: 11 Jun 2025 21:23




Tgt Ion: 75 Resp: 153059


Ion Ratio Lower Upper

75 100

110 35.4 17.6 52.9

77 31.0 25.3 37.9





#37

Ethyl Acetate

Concen: 53.483 ug/l

RT: 7.571 min Scan# 91

Delta R.T. -0.006 min

Lab File: VN086965.D


Acq: 11 Jun 2025 21:23

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC



Tgt Ion: 43 Resp: 19617

Ion Ratio Lower Upper

43 100

61 15.7 12.1 18.1




70 12.7 9.8 14.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025



#38

Carbon Tetrachloride

Concen: 53.303 ug/l

RT: 8.371 min Scan# 1091

Delta R.T. -0.000 min

Lab File: VN086965.D

Acq: 11 Jun 2025 21:23




Tgt Ion:117 Resp: 150782


Ion Ratio Lower Upper

117 100

119 97.9 75.0 112.6

121 32.3 24.9 37.3





#39

Methylcyclohexane

Concen: 44.605 ug/l

RT: 9.606 min Scan# 1130

Delta R.T. -0.000 min

Lab File: VN086965.D

Acq: 11 Jun 2025 21:23

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 83 Resp: 176079

Ion Ratio Lower Upper

83 100

55 71.5 61.0 91.6





98 48.7 38.9 58.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025



#40

Benzene

Concen: 53.695 ug/l

RT: 8.612 min Scan# 1132

Delta R.T. -0.000 min

Lab File: VN086965.D




Acq: 11 Jun 2025 21:23


Tgt Ion: 78 Resp: 505922

Ion Ratio Lower Upper

78 100


77 23.4 18.7 28.1





#41
Methacrylonitrile
Concen: 54.646 ug/l
RT: 7.788 min Scan# 992
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 41 Resp: 11256
Ion Ratio Lower Upper
41 100
39 51.6 41.2 61.8
67 80.7 61.4 92.2
52 30.1 22.7 34.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#42
1,2-Dichloroethane
Concen: 53.873 ug/l
RT: 8.677 min Scan# 1143
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23


Tgt Ion: 62 Resp: 153959
Ion Ratio Lower Upper
62 100
98 10.5 0.0 21.2





#43
Isopropyl Acetate
Concen: 53.857 ug/l
RT: 8.694 min Scan# 1146
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 43 Resp: 317059
Ion Ratio Lower Upper
43 100
61 28.1 22.1 33.1
87 15.2 11.8 17.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#44
Trichloroethene
Concen: 53.458 ug/l
RT: 9.359 min Scan# 1259
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23

Tgt Ion:130 Resp: 119433
Ion Ratio Lower Upper
130 100
95 96.2 0.0 196.2





#45

1,2-Dichloropropane

Concen: 54.196 ug/l

RT: 9.624 min Scan# 1304

Delta R.T. -0.006 min

Lab File: VN086965.D

Acq: 11 Jun 2025 21:23

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 63 Resp: 12423

Ion Ratio Lower Upper

63 100





65 31.6 25.5 38.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025



#46

Dibromomethane

Concen: 56.431 ug/l

RT: 9.712 min Scan# 1319

Delta R.T. -0.006 min

Lab File: VN086965.D

Acq: 11 Jun 2025 21:23




Tgt Ion: 93 Resp: 85919


Ion Ratio Lower Upper

93 100

95 85.6 67.1 100.7


174 93.3 74.2 111.4





#47
Bromodichloromethane
Concen: 55.894 ug/l
RT: 9.894 min Scan# 1350
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 83 Resp: 175093
Ion Ratio Lower Upper
83 100
85 62.6 50.9 76.3
127 7.9 6.6 10.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#48
Methyl methacrylate
Concen: 53.046 ug/l
RT: 9.682 min Scan# 1314
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23

Tgt Ion: 41 Resp: 143734
Ion Ratio Lower Upper
41 100
69 96.2 74.6 112.0
39 49.3 40.2 60.4





#49

1,4-Dioxane

Concen: 1194.044 ug/l

RT: 9.706 min Scan# 11

Delta R.T. 0.006 min

Lab File: VN086965.D

Acq: 11 Jun 2025 21:23

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 88 Resp: 5885

Ion Ratio Lower Upper

88 100

43 0.0 0.0 0.0





58 69.1 57.7 86.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025



#50

Toluene-d8

Concen: 49.914 ug/l

RT: 10.571 min Scan# 1465

Delta R.T. -0.000 min

Lab File: VN086965.D




Acq: 11 Jun 2025 21:23


Tgt Ion: 98 Resp: 382080

Ion Ratio Lower Upper

98 100


100 68.3 53.4 80.0





#51
4-Methyl-2-Pentanone
Concen: 281.028 ug/l
RT: 10.447 min Scan# 1444
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 43 Resp: 995984
Ion Ratio Lower Upper
43 100
58 43.6 34.2 51.4

Manual Integrations
APPROVED



Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025




#52
Toluene
Concen: 53.976 ug/l
RT: 10.629 min Scan# 1475
Delta R.T. -0.006 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23




Tgt Ion: 92 Resp: 310797
Ion Ratio Lower Upper
92 100
91 169.1 136.5 204.7





#53
t-1,3-Dichloropropene
Concen: 54.745 ug/l
RT: 10.835 min Scan# 1510
Delta R.T. -0.006 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 75 Resp: 19176
Ion Ratio Lower Upper
75 100
77 30.9 26.4 39.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#54
cis-1,3-Dichloropropene
Concen: 54.242 ug/l
RT: 10.318 min Scan# 1422
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23


Tgt Ion: 75 Resp: 203300
Ion Ratio Lower Upper
75 100
77 31.0 26.2 39.4
39 39.0 32.3 48.5





#55
1,1,2-Trichloroethane
Concen: 55.842 ug/l
RT: 11.018 min Scan# 1541
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 97 Resp: 123723
Ion Ratio Lower Upper
97 100
83 89.6 69.7 104.5
85 57.7 45.1 67.7
99 65.2 50.6 75.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#56
Ethyl methacrylate
Concen: 59.246 ug/l
RT: 10.882 min Scan# 1518
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23


Tgt Ion: 69 Resp: 209075
Ion Ratio Lower Upper
69 100
41 55.6 47.2 70.8
39 26.5 23.0 34.6





#57
1,3-Dichloropropane
Concen: 55.319 ug/l
RT: 11.165 min Scan# 1566
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 76 Resp: 212617
Ion Ratio Lower Upper
76 100
78 32.9 25.3 37.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#58
2-Chloroethyl Vinyl ether
Concen: 295.338 ug/l
RT: 10.165 min Scan# 1396
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23


Tgt Ion: 63 Resp: 621634
Ion Ratio Lower Upper
63 100
106 29.1 23.0 34.6





#59
2-Hexanone
Concen: 289.751 ug/l
RT: 11.200 min Scan# 1572
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 43 Resp: 661460
Ion Ratio Lower Upper
43 100
58 65.1 30.4 91.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#60
Dibromochloromethane
Concen: 57.580 ug/l
RT: 11.359 min Scan# 1599
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23


Tgt Ion:129 Resp: 132918
Ion Ratio Lower Upper
129 100
127 77.1 38.6 115.8





#61
1,2-Dibromoethane
Concen: 56.119 ug/l
RT: 11.470 min Scan# 1618
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC



Tgt Ion:107 Resp: 127444
Ion Ratio Lower Upper
107 100
109 93.8 76.3 114.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#62
4-Bromofluorobenzene
Concen: 51.499 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23


Tgt Ion: 95 Resp: 146464
Ion Ratio Lower Upper
95 100
174 72.0 0.0 141.8
176 69.3 0.0 132.6





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 1685
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC



Tgt Ion:117 Resp: 286630
Ion Ratio Lower Upper
117 100
82 55.2 44.6 67.0
119 32.2 25.5 38.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#64
Tetrachloroethene
Concen: 50.534 ug/l
RT: 11.106 min Scan# 1556
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23

Tgt Ion:164 Resp: 91667
Ion Ratio Lower Upper
164 100
166 121.1 101.8 152.8
129 98.1 83.0 124.4
131 95.3 78.2 117.2









#65
Chlorobenzene
Concen: 53.206 ug/l
RT: 11.894 min Scan# 110
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Tgt Ion:112 Resp: 336350
Ion Ratio Lower Upper
112 100
114 32.8 26.2 39.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#66
1,1,1,2-Tetrachloroethane
Concen: 55.835 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. -0.006 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23

Tgt Ion:131 Resp: 113463
Ion Ratio Lower Upper
131 100
133 95.7 48.3 144.8
119 67.8 34.0 101.8









#67
Ethyl Benzene
Concen: 52.952 ug/l
RT: 11.965 min Scan# 1702
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 91 Resp: 576593
Ion Ratio Lower Upper
91 100
106 31.5 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#68
m/p-Xylenes
Concen: 106.561 ug/l
RT: 12.070 min Scan# 1720
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23

Tgt Ion:106 Resp: 444183
Ion Ratio Lower Upper
106 100
91 199.8 159.4 239.0









#69
o-Xylene
Concen: 54.963 ug/l
RT: 12.400 min Scan# 1776
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Tgt Ion:106 Resp: 219424
Ion Ratio Lower Upper
106 100
91 208.2 106.1 318.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#70
Styrene
Concen: 55.293 ug/l
RT: 12.412 min Scan# 1778
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23


Tgt Ion:104 Resp: 377725
Ion Ratio Lower Upper
104 100
78 47.3 38.5 57.7
103 53.9 43.1 64.7





#71
Bromoform
Concen: 59.559 ug/l
RT: 12.576 min Scan# 1806
Delta R.T. -0.006 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC



Tgt Ion:173 Resp: 89670
Ion Ratio Lower Upper
173 100
175 48.5 24.1 72.4
254 0.0 0.0 0.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.794 min Scan# 2013
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23


Tgt Ion:152 Resp: 142002
Ion Ratio Lower Upper
152 100
115 60.8 30.1 90.5
150 176.0 0.0 345.0





#73
Isopropylbenzene
Concen: 51.622 ug/l
RT: 12.694 min Scan# 1826
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 105 Resp: 53403
Ion Ratio Lower Upper
105 100
120 27.0 13.5 40.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#74
N-ethyl acetate
Concen: 52.216 ug/l
RT: 12.512 min Scan# 1795
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23


Tgt Ion: 43 Resp: 188759
Ion Ratio Lower Upper
43 100
70 49.1 37.6 56.4
55 26.3 19.0 28.6
61 27.2 20.8 31.2





#75
1,1,2,2-Tetrachloroethane
Concen: 57.301 ug/l
RT: 12.935 min Scan# 1867
Delta R.T. -0.006 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 83 Resp: 200820
Ion Ratio Lower Upper
83 100
131 9.6 4.9 14.7
85 64.7 32.0 96.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#76
1,2,3-Trichloropropane
Concen: 51.664 ug/l m
RT: 12.994 min Scan# 1877
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23


Tgt Ion: 75 Resp: 174388
Ion Ratio Lower Upper
75 100
77 175.4 79.3 237.9





#77
Bromobenzene
Concen: 55.583 ug/l
RT: 12.982 min Scan# 1875
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC



Tgt Ion:156 Resp: 131869
Ion Ratio Lower Upper
156 100
77 232.0 117.7 353.1
158 96.4 48.1 144.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#78
n-propylbenzene
Concen: 51.178 ug/l
RT: 13.035 min Scan# 1884
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23


Tgt Ion: 91 Resp: 643398
Ion Ratio Lower Upper
91 100
120 23.2 11.4 34.2





#79
2-Chlorotoluene
Concen: 52.371 ug/l
RT: 13.123 min Scan# 1899
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 91 Resp: 394854
Ion Ratio Lower Upper
91 100
126 33.6 16.2 48.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#80
1,3,5-Trimethylbenzene
Concen: 51.914 ug/l
RT: 13.170 min Scan# 1907
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23


Tgt Ion:105 Resp: 443398
Ion Ratio Lower Upper
105 100
120 50.5 24.9 74.6





#81
trans-1,4-Dichloro-2-butene
Concen: 49.328 ug/l
RT: 12.735 min Scan# 1833
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 75 Resp: 7233
Ion Ratio Lower Upper
75 100
53 81.4 66.6 99.8
89 43.7 36.7 55.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#82
4-Chlorotoluene
Concen: 53.310 ug/l
RT: 13.223 min Scan# 1916
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23


Tgt Ion: 91 Resp: 406685
Ion Ratio Lower Upper
91 100
126 33.7 16.7 50.1





#83
tert-Butylbenzene
Concen: 50.730 ug/l
RT: 13.435 min Scan# 1952
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC



Tgt Ion:119 Resp: 396615
Ion Ratio Lower Upper
119 100
91 62.4 32.1 96.5
134 25.1 13.4 40.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#84
1,2,4-Trimethylbenzene
Concen: 52.251 ug/l
RT: 13.482 min Scan# 1960
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23

Tgt Ion:105 Resp: 447506
Ion Ratio Lower Upper
105 100
120 47.0 23.2 69.6









#85
 sec-Butylbenzene
 Concen: 48.432 ug/l
 RT: 13.611 min Scan# 1983
 Delta R.T. -0.006 min
 Lab File: VN086965.D
 Acq: 11 Jun 2025 21:23

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC

Tgt Ion:105 Resp: 54988
 Ion Ratio Lower Upper
 105 100
 134 20.1 10.0 30.0




Manual Integrations
 APPROVED


Reviewed By :John Carlone 06/12/2025
 Supervised By :Semsettin Yesilyurt 06/12/2025



#86
 p-Isopropyltoluene
 Concen: 48.982 ug/l
 RT: 13.729 min Scan# 2002
 Delta R.T. -0.000 min
 Lab File: VN086965.D
 Acq: 11 Jun 2025 21:23

Tgt Ion:119 Resp: 459716
 Ion Ratio Lower Upper
 119 100
 134 26.2 13.5 40.4
 91 22.3 11.5 34.4





#87

1,3-Dichlorobenzene

Concen: 54.264 ug/l

RT: 13.735 min Scan# 2003

Delta R.T. -0.000 min

Lab File: VN086965.D

Acq: 11 Jun 2025 21:23

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC

Tgt Ion:146 Resp: 253292

Ion Ratio Lower Upper

146 100

111 41.0 20.9 62.8





148 62.9 32.0 96.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025



#88

1,4-Dichlorobenzene

Concen: 53.138 ug/l

RT: 13.811 min Scan# 2016

Delta R.T. -0.000 min

Lab File: VN086965.D

Acq: 11 Jun 2025 21:23




Tgt Ion:146 Resp: 252880


Ion Ratio Lower Upper

146 100

111 40.2 19.9 59.7

148 64.2 31.8 95.4





#89

n-Butylbenzene

Concen: 46.509 ug/l

RT: 14.053 min Scan# 2057

Delta R.T. -0.000 min

Lab File: VN086965.D

Acq: 11 Jun 2025 21:23

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 91 Resp: 42280

Ion Ratio Lower Upper

91 100

92 53.9 27.5 82.4





134 26.1 13.0 39.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025



#90

Hexachloroethane

Concen: 52.642 ug/l

RT: 14.329 min Scan# 2104

Delta R.T. -0.006 min

Lab File: VN086965.D




Acq: 11 Jun 2025 21:23


Tgt Ion:117 Resp: 83746

Ion Ratio Lower Upper

117 100


201 64.9 33.2 99.6





#91
1,2-Dichlorobenzene
Concen: 54.507 ug/l
RT: 14.106 min Scan# 2066
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC



Tgt Ion:146 Resp: 244439
Ion Ratio Lower Upper
146 100
111 43.5 21.6 64.6
148 63.6 31.7 95.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#92
1,2-Dibromo-3-Chloropropane
Concen: 54.724 ug/l
RT: 14.717 min Scan# 2170
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23

Tgt Ion: 75 Resp: 45867
Ion Ratio Lower Upper
75 100
155 84.4 41.1 123.4
157 108.4 52.7 158.1





#93

1,2,4-Trichlorobenzene

Concen: 49.501 ug/l

RT: 15.388 min Scan# 21

Delta R.T. -0.000 min

Lab File: VN086965.D


Acq: 11 Jun 2025 21:23

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC



Tgt Ion:180 Resp: 14175

Ion Ratio Lower Upper

180 100

182 93.7 46.9 140.7




145 32.3 16.0 47.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025



#94

Hexachlorobutadiene


Concen: 41.918 ug/l

RT: 15.500 min Scan# 2303

Delta R.T. -0.000 min

Lab File: VN086965.D

Acq: 11 Jun 2025 21:23





Tgt Ion:225 Resp: 44722


Ion Ratio Lower Upper

225 100

223 66.3 33.5 100.4


227 62.3 31.1 93.5





#95
Naphthalene
Concen: 54.049 ug/l
RT: 15.641 min Scan# 2327
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC



Tgt Ion:128 Resp: 57610
Ion Ratio Lower Upper
128 100
127 12.6 10.2 15.2
129 10.8 8.8 13.2




Manual Integrations
APPROVED

Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#96
1,2,3-Trichlorobenzene
Concen: 48.742 ug/l
RT: 15.841 min Scan# 2361
Delta R.T. 0.006 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23

Tgt Ion:180 Resp: 138678
Ion Ratio Lower Upper
180 100
182 95.2 47.0 141.0
145 33.4 16.6 49.8



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061125\
 Data File : VN086965.D
 Acq On : 11 Jun 2025 21:23
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 12 01:48:05 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	86	0.00
2 T	Dichlorodifluoromethane	0.499	0.552	-10.6	94	0.00
3 P	Chloromethane	0.644	0.581	9.8	83	0.00
4 C	Vinyl Chloride	0.664	0.712	-7.2#	95	0.00
5 T	Bromomethane	0.372	0.246	33.9#	59	-0.01
6 T	Chloroethane	0.429	0.477	-11.2	100	0.00
7 T	Trichlorofluoromethane	0.868	0.969	-11.6	100	0.00
8 T	Diethyl Ether	0.378	0.426	-12.7	101	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.545	0.587	-7.7	97	0.00
10 T	Methyl Iodide	0.707	0.264	62.7#	33#	0.00
11 T	Tert butyl alcohol	0.182	0.204	-12.1	99	0.00
12 CM	1,1-Dichloroethene	0.557	0.599	-7.5#	96	0.00
13 T	Acrolein	0.057	0.172	-201.8#	324#	0.00
14 T	Allyl chloride	0.923	0.940	-1.8	93	0.00
15 T	Acrylonitrile	0.425	0.476	-12.0	100	0.00
16 T	Acetone	0.355	0.373	-5.1	99	0.00
17 T	Carbon Disulfide	1.539	1.512	1.8	91	0.00
18 T	Methyl Acetate	1.035	1.207	-16.6	105	0.00
19 T	Methyl tert-butyl Ether	2.015	2.289	-13.6	101	0.00
20 T	Methylene Chloride	0.665	0.707	-6.3	100	0.00
21 T	trans-1,2-Dichloroethene	0.619	0.653	-5.5	98	0.00
22 T	Diisopropyl ether	1.945	2.151	-10.6	99	0.00
23 T	Vinyl Acetate	1.644	1.829	-11.3	98	0.00
24 P	1,1-Dichloroethane	1.120	1.255	-12.1	101	0.00
25 T	2-Butanone	0.577	0.628	-8.8	98	0.00
26 T	2,2-Dichloropropane	0.871	0.842	3.3	80	0.00
27 T	cis-1,2-Dichloroethene	0.740	0.824	-11.4	101	0.00
28 T	Bromochloromethane	0.550	0.630	-14.5	116	0.00
29 T	Tetrahydrofuran	0.376	0.407	-8.2	97	0.00
30 C	Chloroform	1.118	1.234	-10.4#	100	0.00
31 T	Cyclohexane	1.086	1.022	5.9	87	0.00
32 T	1,1,1-Trichloroethane	0.951	1.024	-7.7	98	0.00
33 S	1,2-Dichloroethane-d4	0.669	0.697	-4.2	119	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	86	0.00
35 S	Dibromofluoromethane	0.296	0.314	-6.1	123	0.00
36 T	1,1-Dichloropropene	0.442	0.469	-6.1	97	0.00
37 T	Ethyl Acetate	0.562	0.601	-6.9	97	0.00
38 T	Carbon Tetrachloride	0.433	0.462	-6.7	97	0.00
39 T	Methylcyclohexane	0.605	0.540	10.7	82	0.00
40 TM	Benzene	1.444	1.551	-7.4	99	0.00
41 T	Methacrylonitrile	0.316	0.345	-9.2	98	0.00
42 TM	1,2-Dichloroethane	0.438	0.472	-7.8	99	0.00
43 T	Isopropyl Acetate	0.902	0.972	-7.8	98	0.00
44 TM	Trichloroethene	0.342	0.366	-7.0	96	0.00
45 C	1,2-Dichloropropane	0.351	0.381	-8.5#	99	0.00
46 T	Dibromomethane	0.233	0.263	-12.9	103	0.00
47 T	Bromodichloromethane	0.480	0.537	-11.9	101	0.00
48 T	Methyl methacrylate	0.415	0.441	-6.3	96	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061125\
 Data File : VN086965.D
 Acq On : 11 Jun 2025 21:23
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 12 01:48:05 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.008	0.009	-12.5	100	0.00
50 S	Toluene-d8	1.173	1.171	0.2	117	0.00
51 T	4-Methyl-2-Pentanone	0.543	0.610	-12.3	98	0.00
52 CM	Toluene	0.882	0.953	-8.0#	98	0.00
53 T	t-1,3-Dichloropropene	0.537	0.588	-9.5	97	0.00
54 T	cis-1,3-Dichloropropene	0.574	0.623	-8.5	97	0.00
55 T	1,1,2-Trichloroethane	0.340	0.379	-11.5	101	0.00
56 T	Ethyl methacrylate	0.541	0.641	-18.5	99	0.00
57 T	1,3-Dichloropropane	0.589	0.652	-10.7	100	0.00
58 T	2-Chloroethyl Vinyl ether	0.323	0.381	-18.0	120	0.00
59 T	2-Hexanone	0.350	0.405	-15.7	95	0.00
60 T	Dibromochloromethane	0.354	0.407	-15.0	103	0.00
61 T	1,2-Dibromoethane	0.348	0.391	-12.4	102	0.00
62 S	4-Bromofluorobenzene	0.436	0.449	-3.0	119	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	86	0.00
64 T	Tetrachloroethene	0.316	0.320	-1.3	94	0.00
65 PM	Chlorobenzene	1.103	1.173	-6.3	99	0.00
66 T	1,1,1,2-Tetrachloroethane	0.354	0.396	-11.9	101	0.00
67 C	Ethyl Benzene	1.899	2.012	-6.0#	96	0.00
68 T	m/p-Xylenes	0.727	0.775	-6.6	95	0.00
69 T	o-Xylene	0.696	0.766	-10.1	98	0.00
70 T	Styrene	1.192	1.318	-10.6	98	0.00
71 P	Bromoform	0.263	0.313	-19.0	102	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	85	0.00
73 T	Isopropylbenzene	3.643	3.761	-3.2	93	0.00
74 T	N-amyl acetate	1.273	1.329	-4.4	89	0.00
75 P	1,1,2,2-Tetrachloroethane	1.234	1.414	-14.6	102	0.00
76 T	1,2,3-Trichloropropane	1.189	1.228	-3.3	89	0.00
77 T	Bromobenzene	0.835	0.929	-11.3	100	0.00
78 T	n-propylbenzene	4.427	4.531	-2.3	91	0.00
79 T	2-Chlorotoluene	2.655	2.781	-4.7	94	0.00
80 T	1,3,5-Trimethylbenzene	3.007	3.122	-3.8	91	0.00
81 T	trans-1,4-Dichloro-2-butene	0.516	0.509	1.4	91	0.00
82 T	4-Chlorotoluene	2.686	2.864	-6.6	96	0.00
83 T	tert-Butylbenzene	2.753	2.793	-1.5	90	0.00
84 T	1,2,4-Trimethylbenzene	3.016	3.151	-4.5	92	0.00
85 T	sec-Butylbenzene	3.998	3.872	3.2	86	0.00
86 T	p-Isopropyltoluene	3.305	3.237	2.1	86	0.00
87 T	1,3-Dichlorobenzene	1.644	1.784	-8.5	97	0.00
88 T	1,4-Dichlorobenzene	1.676	1.781	-6.3	96	0.00
89 T	n-Butylbenzene	3.201	2.977	7.0	82	0.00
90 T	Hexachloroethane	0.560	0.590	-5.4	93	0.00
91 T	1,2-Dichlorobenzene	1.579	1.721	-9.0	97	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.295	0.323	-9.5	101	0.00
93 T	1,2,4-Trichlorobenzene	1.008	0.998	1.0	87	0.00
94 T	Hexachlorobutadiene	0.376	0.315	16.2	73	0.00
95 T	Naphthalene	3.753	4.057	-8.1	96	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061125\
Data File : VN086965.D
Acq On : 11 Jun 2025 21:23
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Jun 12 01:48:05 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 02:12:50 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.002	0.977	2.5	87	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061125\
 Data File : VN086965.D
 Acq On : 11 Jun 2025 21:23
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 12 01:48:05 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	86	0.00
2 T	Dichlorodifluoromethane	50.000	55.320	-10.6	94	0.00
3 P	Chloromethane	50.000	45.148	9.7	83	0.00
4 C	Vinyl Chloride	50.000	53.612	-7.2#	95	0.00
5 T	Bromomethane	50.000	33.049	33.9#	59	-0.01
6 T	Chloroethane	50.000	55.606	-11.2	100	0.00
7 T	Trichlorofluoromethane	50.000	55.827	-11.7	100	0.00
8 T	Diethyl Ether	50.000	56.330	-12.7	101	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	53.851	-7.7	97	0.00
10 T	Methyl Iodide	50.000	18.649	62.7#	33	0.00
11 T	Tert butyl alcohol	250.000	280.391	-12.2	99	0.00
12 CM	1,1-Dichloroethene	50.000	53.824	-7.6#	96	0.00
13 T	Acrolein	250.000	746.633	-198.7#	324	0.00
14 T	Allyl chloride	50.000	50.921	-1.8	93	0.00
15 T	Acrylonitrile	250.000	280.476	-12.2	100	0.00
16 T	Acetone	250.000	262.661	-5.1	99	0.00
17 T	Carbon Disulfide	50.000	49.127	1.7	91	0.00
18 T	Methyl Acetate	50.000	58.328	-16.7	105	0.00
19 T	Methyl tert-butyl Ether	50.000	56.790	-13.6	101	0.00
20 T	Methylene Chloride	50.000	53.166	-6.3	100	0.00
21 T	trans-1,2-Dichloroethene	50.000	52.697	-5.4	98	0.00
22 T	Diisopropyl ether	50.000	55.291	-10.6	99	0.00
23 T	Vinyl Acetate	250.000	278.142	-11.3	98	0.00
24 P	1,1-Dichloroethane	50.000	56.057	-12.1	101	0.00
25 T	2-Butanone	250.000	272.256	-8.9	98	0.00
26 T	2,2-Dichloropropane	50.000	48.315	3.4	80	0.00
27 T	cis-1,2-Dichloroethene	50.000	55.654	-11.3	101	0.00
28 T	Bromochloromethane	50.000	57.211	-14.4	116	0.00
29 T	Tetrahydrofuran	250.000	270.436	-8.2	97	0.00
30 C	Chloroform	50.000	55.175	-10.3#	100	0.00
31 T	Cyclohexane	50.000	47.071	5.9	87	0.00
32 T	1,1,1-Trichloroethane	50.000	53.846	-7.7	98	0.00
33 S	1,2-Dichloroethane-d4	50.000	52.069	-4.1	119	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	86	0.00
35 S	Dibromofluoromethane	50.000	53.000	-6.0	123	0.00
36 T	1,1-Dichloropropene	50.000	53.121	-6.2	97	0.00
37 T	Ethyl Acetate	50.000	53.483	-7.0	97	0.00
38 T	Carbon Tetrachloride	50.000	53.303	-6.6	97	0.00
39 T	Methylcyclohexane	50.000	44.605	10.8	82	0.00
40 TM	Benzene	50.000	53.695	-7.4	99	0.00
41 T	Methacrylonitrile	50.000	54.646	-9.3	98	0.00
42 TM	1,2-Dichloroethane	50.000	53.873	-7.7	99	0.00
43 T	Isopropyl Acetate	50.000	53.857	-7.7	98	0.00
44 TM	Trichloroethene	50.000	53.458	-6.9	96	0.00
45 C	1,2-Dichloropropane	50.000	54.196	-8.4#	99	0.00
46 T	Dibromomethane	50.000	56.431	-12.9	103	0.00
47 T	Bromodichloromethane	50.000	55.894	-11.8	101	0.00
48 T	Methyl methacrylate	50.000	53.046	-6.1	96	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061125\
 Data File : VN086965.D
 Acq On : 11 Jun 2025 21:23
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 12 01:48:05 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	1194.044	-19.4	100	0.00
50 S	Toluene-d8	50.000	49.914	0.2	117	0.00
51 T	4-Methyl-2-Pentanone	250.000	281.028	-12.4	98	0.00
52 CM	Toluene	50.000	53.976	-8.0#	98	0.00
53 T	t-1,3-Dichloropropene	50.000	54.745	-9.5	97	0.00
54 T	cis-1,3-Dichloropropene	50.000	54.242	-8.5	97	0.00
55 T	1,1,2-Trichloroethane	50.000	55.842	-11.7	101	0.00
56 T	Ethyl methacrylate	50.000	59.246	-18.5	99	0.00
57 T	1,3-Dichloropropane	50.000	55.319	-10.6	100	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	295.338	-18.1	120	0.00
59 T	2-Hexanone	250.000	289.751	-15.9	95	0.00
60 T	Dibromochloromethane	50.000	57.580	-15.2	103	0.00
61 T	1,2-Dibromoethane	50.000	56.119	-12.2	102	0.00
62 S	4-Bromofluorobenzene	50.000	51.499	-3.0	119	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	86	0.00
64 T	Tetrachloroethene	50.000	50.534	-1.1	94	0.00
65 PM	Chlorobenzene	50.000	53.206	-6.4	99	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	55.835	-11.7	101	0.00
67 C	Ethyl Benzene	50.000	52.952	-5.9#	96	0.00
68 T	m/p-Xylenes	100.000	106.561	-6.6	95	0.00
69 T	o-Xylene	50.000	54.963	-9.9	98	0.00
70 T	Styrene	50.000	55.293	-10.6	98	0.00
71 P	Bromoform	50.000	59.559	-19.1	102	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	85	0.00
73 T	Isopropylbenzene	50.000	51.622	-3.2	93	0.00
74 T	N-amyl acetate	50.000	52.216	-4.4	89	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	57.301	-14.6	102	0.00
76 T	1,2,3-Trichloropropane	50.000	51.664	-3.3	89	0.00
77 T	Bromobenzene	50.000	55.583	-11.2	100	0.00
78 T	n-propylbenzene	50.000	51.178	-2.4	91	0.00
79 T	2-Chlorotoluene	50.000	52.371	-4.7	94	0.00
80 T	1,3,5-Trimethylbenzene	50.000	51.914	-3.8	91	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	49.328	1.3	91	0.00
82 T	4-Chlorotoluene	50.000	53.310	-6.6	96	0.00
83 T	tert-Butylbenzene	50.000	50.730	-1.5	90	0.00
84 T	1,2,4-Trimethylbenzene	50.000	52.251	-4.5	92	0.00
85 T	sec-Butylbenzene	50.000	48.432	3.1	86	0.00
86 T	p-Isopropyltoluene	50.000	48.982	2.0	86	0.00
87 T	1,3-Dichlorobenzene	50.000	54.264	-8.5	97	0.00
88 T	1,4-Dichlorobenzene	50.000	53.138	-6.3	96	0.00
89 T	n-Butylbenzene	50.000	46.509	7.0	82	0.00
90 T	Hexachloroethane	50.000	52.642	-5.3	93	0.00
91 T	1,2-Dichlorobenzene	50.000	54.507	-9.0	97	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	54.724	-9.4	101	0.00
93 T	1,2,4-Trichlorobenzene	50.000	49.501	1.0	87	0.00
94 T	Hexachlorobutadiene	50.000	41.918	16.2	73	0.00
95 T	Naphthalene	50.000	54.049	-8.1	96	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061125\
Data File : VN086965.D
Acq On : 11 Jun 2025 21:23
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Jun 12 01:48:05 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 02:12:50 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	48.742	2.5	87	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6





QC SAMPLE DATA

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060625\
Data File : VN086861.D
Acq On : 06 Jun 2025 07:59
Operator : JC\MD
Sample : BFB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Title : SW846 8260
Last Update : Sat Jun 07 02:12:50 2025



AutoFind: Scans 1851, 1852, 1853; Background Corrected with Scan 1843



Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.3	11610	PASS
75	95	30	60	48.1	32235	PASS
95	95	100	100	100.0	67037	PASS
96	95	5	9	6.4	4284	PASS
173	174	0.00	2	1.0	453	PASS
174	95	50	100	66.6	44632	PASS
175	174	5	9	7.1	3184	PASS
176	174	95	101	98.1	43805	PASS
177	176	5	9	6.8	2979	PASS

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061125\
Data File : VN086939.D
Acq On : 11 Jun 2025 10:22
Operator : JC\MD
Sample : BFB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Title : SW846 8260
Last Update : Sat Jun 07 02:12:50 2025



AutoFind: Scans 1851, 1852, 1853; Background Corrected with Scan 1843

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	15.7	10335	PASS
75	95	30	60	46.5	30525	PASS
95	95	100	100	100.0	65627	PASS
96	95	5	9	7.1	4656	PASS
173	174	0.00	2	1.0	446	PASS
174	95	50	100	71.1	46675	PASS
175	174	5	9	7.5	3520	PASS
176	174	95	101	96.8	45165	PASS
177	176	5	9	7.0	3156	PASS

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VN0611WBL01		SDG No.:	Q2263
Lab Sample ID:	VN0611WBL01		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086942.D	1		06/11/25 12:28	VN061125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	3.80	U	1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.22	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VN0611WBL01		SDG No.:	Q2263
Lab Sample ID:	VN0611WBL01		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086942.D	1		06/11/25 12:28	VN061125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	48.3		81 - 118		97%	SPK: 50
1868-53-7	Dibromofluoromethane	49.3		80 - 119		99%	SPK: 50
2037-26-5	Toluene-d8	51.7		89 - 112		103%	SPK: 50
460-00-4	4-Bromofluorobenzene	49.4		85 - 114		99%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	328000	8.23				
540-36-3	1,4-Difluorobenzene	619000	9.106				
3114-55-4	Chlorobenzene-d5	543000	11.865				
3855-82-1	1,4-Dichlorobenzene-d4	257000	13.788				



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VN0611WBL01		SDG No.:	Q2263
Lab Sample ID:	VN0611WBL01		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086942.D	1		06/11/25 12:28	VN061125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061125\
Data File : VN086942.D
Acq On : 11 Jun 2025 12:28
Operator : JC\MD
Sample : VN0611WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0611WBL01

Quant Time: Jun 12 01:30:13 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 02:12:50 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	328120	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	618651	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	543433	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	257257	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	212062	48.271	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	96.540%
35) Dibromofluoromethane	8.177	113	180701	49.287	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.580%
50) Toluene-d8	10.565	98	749967	51.673	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.340%
62) 4-Bromofluorobenzene	12.847	95	266545	49.431	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	98.860%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061125\
Data File : VN086942.D
Acq On : 11 Jun 2025 12:28
Operator : JC\MD
Sample : VN0611WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0611WBL01

Quant Time: Jun 12 01:30:13 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 02:12:50 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.230 min Scan# 1068
Delta R.T. -0.006 min
Lab File: VN086942.D
Acq: 11 Jun 2025 12:28




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBL01


Tgt Ion:168 Resp: 328120
Ion Ratio Lower Upper
168 100
99 58.0 49.1 73.7



#33
1,2-Dichloroethane-d4
Concen: 48.271 ug/l
RT: 8.583 min Scan# 1127
Delta R.T. -0.006 min
Lab File: VN086942.D
Acq: 11 Jun 2025 12:28

Tgt Ion: 65 Resp: 212062
Ion Ratio Lower Upper
65 100
67 55.4 0.0 105.6





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.106 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086942.D

Acq: 11 Jun 2025 12:28

Instrument :

MSVOA_N

ClientSampleId :

VN0611WBL01





Tgt Ion:114 Resp: 618651

Ion Ratio Lower Upper

114 100

63 19.9 0.0 39.6

88 16.5 0.0 30.2



#35

Dibromofluoromethane

Concen: 49.287 ug/l

RT: 8.177 min Scan# 1058

Delta R.T. 0.006 min

Lab File: VN086942.D

Acq: 11 Jun 2025 12:28




Tgt Ion:113 Resp: 180701


Ion Ratio Lower Upper

113 100

111 103.1 84.2 126.2

192 18.0 14.2 21.4









#50
Toluene-d8
Concen: 51.673 ug/l
RT: 10.565 min Scan# 1464
Delta R.T. -0.006 min
Lab File: VN086942.D
Acq: 11 Jun 2025 12:28




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBL01


Tgt Ion: 98 Resp: 749967
Ion Ratio Lower Upper
98 100
100 66.0 53.4 80.0



#62
4-Bromofluorobenzene
Concen: 49.431 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. -0.000 min
Lab File: VN086942.D
Acq: 11 Jun 2025 12:28

Tgt Ion: 95 Resp: 266545
Ion Ratio Lower Upper
95 100
174 72.3 0.0 141.8
176 68.8 0.0 132.6





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.865 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086942.D

Acq: 11 Jun 2025 12:28

Instrument :

MSVOA_N

ClientSampleId :

VN0611WBL01





Tgt Ion:117 Resp: 543433

Ion Ratio Lower Upper

117 100

82 55.0 44.6 67.0

119 31.9 25.5 38.3



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.788 min Scan# 2012

Delta R.T. -0.006 min

Lab File: VN086942.D

Acq: 11 Jun 2025 12:28




Tgt Ion:152 Resp: 257257

Ion Ratio Lower Upper

152 100

115 61.2 30.1 90.5

150 157.9 0.0 345.0



Report of Analysis

Client:	AECOM Technical Services, Inc.	Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258	Date Received:	
Client Sample ID:	VN0611WBS02	SDG No.:	Q2263
Lab Sample ID:	VN0611WBS02	Matrix:	Water
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086944.D	1		06/11/25 13:27	VN061125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	19.7		0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	16.1		0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	18.9		0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	16.6		1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	19.4		0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	19.2		0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	19.6		0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	19.2		0.23	0.75	1.00	ug/L
67-64-1	Acetone	99.0		1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	17.8		0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	19.6		0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	19.0		0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	19.0		0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	18.5		0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	19.4		0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	17.6		1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	90.4		0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	19.2		0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	19.5		0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	19.9		0.22	0.50	1.00	ug/L
67-66-3	Chloroform	19.5		0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	19.0		0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	16.9		0.16	0.50	1.00	ug/L
71-43-2	Benzene	19.3		0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	19.5		0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	19.9		0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	19.6		0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	19.6		0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	98.0		0.68	2.50	5.00	ug/L
108-88-3	Toluene	19.5		0.14	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.	Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258	Date Received:	
Client Sample ID:	VN0611WBS02	SDG No.:	Q2263
Lab Sample ID:	VN0611WBS02	Matrix:	Water
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086944.D	1		06/11/25 13:27	VN061125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	19.9		0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	20.1		0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	20.4		0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	89.2		0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	20.5		0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	19.7		0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	19.2		0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	20.0		0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	19.6		0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	39.6		0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	20.2		0.12	0.50	1.00	ug/L
100-42-5	Styrene	20.1		0.15	0.50	1.00	ug/L
75-25-2	Bromoform	21.0		0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	19.3		0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	21.0		0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	20.3		0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	20.4		0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	20.0		0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	19.2		0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	18.4		0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	17.8		0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	47.1		81 - 118		94%	SPK: 50
1868-53-7	Dibromofluoromethane	51.1		80 - 119		102%	SPK: 50
2037-26-5	Toluene-d8	48.7		89 - 112		97%	SPK: 50
460-00-4	4-Bromofluorobenzene	50.0		85 - 114		100%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	207000	8.235				
540-36-3	1,4-Difluorobenzene	371000	9.106				
3114-55-4	Chlorobenzene-d5	321000	11.865				
3855-82-1	1,4-Dichlorobenzene-d4	157000	13.788				



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VN0611WBS02		SDG No.:	Q2263
Lab Sample ID:	VN0611WBS02		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086944.D	1		06/11/25 13:27	VN061125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061125\
 Data File : VN086944.D
 Acq On : 11 Jun 2025 13:27
 Operator : JC\MD
 Sample : VN0611WBS02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0611WBS02

Manual Integrations
 APPROVED

Reviewed By : John Carlone 06/12/2025
 Supervised By : Semsettin Yesilyurt 06/12/2025

Quant Time: Jun 12 01:31:27 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	206799	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	370524	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	321188	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	157403	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	130342	47.075	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	94.160%		
35) Dibromofluoromethane	8.177	113	112198	51.096	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	102.200%		
50) Toluene-d8	10.571	98	423380	48.706	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	97.420%		
62) 4-Bromofluorobenzene	12.847	95	161321	49.951	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	99.900%		
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	2.153	85	40672	19.725	ug/l	100
3) Chloromethane	2.401	50	42812	16.079	ug/l	97
4) Vinyl Chloride	2.553	62	51780	18.852	ug/l	95
5) Bromomethane	3.000	94	25577	16.641	ug/l	90
6) Chloroethane	3.153	64	34500	19.446	ug/l	99
7) Trichlorofluoromethane	3.530	101	68752	19.159	ug/l	99
8) Diethyl Ether	3.983	74	30914	19.772	ug/l	94
9) 1,1,2-Trichlorotrifluo...	4.400	101	44117	19.568	ug/l	99
10) Methyl Iodide	4.612	142	36924	12.634	ug/l	100
11) Tert butyl alcohol	5.541	59	69743	92.831	ug/l	99
12) 1,1-Dichloroethene	4.365	96	44269	19.233	ug/l	97
13) Acrolein	4.206	56	22290	93.730	ug/l	98
14) Allyl chloride	5.047	41	68027	17.821	ug/l	97
15) Acrylonitrile	5.741	53	166971	95.084	ug/l	100
16) Acetone	4.447	43	145322	98.972	ug/l	98
17) Carbon Disulfide	4.742	76	113132	17.769	ug/l	100
18) Methyl Acetate	5.047	43	81357	19.013	ug/l	99
19) Methyl tert-butyl Ether	5.812	73	163338	19.599	ug/l	98
20) Methylene Chloride	5.294	84	52196	18.988	ug/l	96
21) trans-1,2-Dichloroethene	5.812	96	47284	18.464	ug/l	95
22) Diisopropyl ether	6.688	45	155508	19.326	ug/l	97
23) Vinyl Acetate	6.618	43	653278	96.092	ug/l	98
24) 1,1-Dichloroethane	6.588	63	89930	19.421	ug/l	100
25) 2-Butanone	7.494	43	215760	90.401	ug/l	96
26) 2,2-Dichloropropane	7.500	77	78031	21.663	ug/l	98
27) cis-1,2-Dichloroethene	7.500	96	59578	19.453	ug/l	99
28) Bromochloromethane	7.824	49	45329	19.909	ug/l	95
29) Tetrahydrofuran	7.853	42	143866	92.532	ug/l	97
30) Chloroform	7.977	83	90272	19.521	ug/l	98
31) Cyclohexane	8.265	56	79050	17.603	ug/l	97
32) 1,1,1-Trichloroethane	8.182	97	74861	19.033	ug/l	97
36) 1,1-Dichloropropene	8.377	75	63577	19.431	ug/l	99
37) Ethyl Acetate	7.571	43	80415	19.306	ug/l	99
38) Carbon Tetrachloride	8.371	117	61747	19.222	ug/l	97
39) Methylcyclohexane	9.606	83	75971	16.947	ug/l	98
40) Benzene	8.612	78	206255	19.277	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061125\
 Data File : VN086944.D
 Acq On : 11 Jun 2025 13:27
 Operator : JC\MD
 Sample : VN0611WBS02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0611WBS02

Manual Integrations
 APPROVED

Reviewed By :John Carlone 06/12/2025
 Supervised By :Semsettin Yesilyurt 06/12/2025

Quant Time: Jun 12 01:31:27 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.788	41	42617	18.218	ug/l	93
42) 1,2-Dichloroethane	8.677	62	63379	19.530	ug/l	98
43) Isopropyl Acetate	8.694	43	126914	18.984	ug/l	98
44) Trichloroethene	9.359	130	50593	19.942	ug/l	99
45) 1,2-Dichloropropane	9.624	63	51091	19.628	ug/l	95
46) Dibromomethane	9.712	93	34741	20.093	ug/l	99
47) Bromodichloromethane	9.894	83	69845	19.634	ug/l	98
48) Methyl methacrylate	9.682	41	57757	18.771	ug/l	99
49) 1,4-Dioxane	9.706	88	22914	409.347	ug/l #	97
51) 4-Methyl-2-Pentanone	10.447	43	394396	97.997	ug/l	99
52) Toluene	10.629	92	127563	19.509	ug/l	99
53) t-1,3-Dichloropropene	10.835	75	79224	19.917	ug/l	100
54) cis-1,3-Dichloropropene	10.318	75	85586	20.109	ug/l	99
55) 1,1,2-Trichloroethane	11.018	97	51421	20.438	ug/l	94
56) Ethyl methacrylate	10.882	69	80942	20.198	ug/l	94
57) 1,3-Dichloropropane	11.165	76	85602	19.613	ug/l	99
58) 2-Chloroethyl Vinyl ether	10.165	63	243584	101.910	ug/l	99
59) 2-Hexanone	11.206	43	231264	89.210	ug/l	95
60) Dibromochloromethane	11.359	129	53655	20.468	ug/l	99
61) 1,2-Dibromoethane	11.471	107	50832	19.711	ug/l	99
64) Tetrachloroethene	11.106	164	38926	19.150	ug/l	98
65) Chlorobenzene	11.894	112	141721	20.006	ug/l	98
66) 1,1,1,2-Tetrachloroethane	11.959	131	46404	20.378	ug/l	99
67) Ethyl Benzene	11.965	91	239733	19.647	ug/l	100
68) m/p-Xylenes	12.070	106	185026	39.613	ug/l	99
69) o-Xylene	12.400	106	90240	20.172	ug/l	99
70) Styrene	12.412	104	154149	20.137	ug/l	99
71) Bromoform	12.582	173	35350	20.952	ug/l #	96
73) Isopropylbenzene	12.694	105	221545	19.320	ug/l	99
74) N-amyl acetate	12.529	43	63289	15.794	ug/l #	85
75) 1,1,2,2-Tetrachloroethane	12.941	83	81476	20.973	ug/l	100
76) 1,2,3-Trichloropropane	12.994	75	71582m	19.132	ug/l	
77) Bromobenzene	12.982	156	55176	20.981	ug/l	98
78) n-propylbenzene	13.035	91	267384	19.187	ug/l	100
79) 2-Chlorotoluene	13.123	91	164813	19.721	ug/l	99
80) 1,3,5-Trimethylbenzene	13.170	105	181773	19.200	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.735	75	30121	18.532	ug/l #	82
82) 4-Chlorotoluene	13.217	91	167991	19.866	ug/l	100
83) tert-Butylbenzene	13.435	119	161593	18.646	ug/l	99
84) 1,2,4-Trimethylbenzene	13.482	105	182431	19.216	ug/l	98
85) sec-Butylbenzene	13.617	105	228907	18.189	ug/l	100
86) p-Isopropyltoluene	13.729	119	192991	18.551	ug/l	99
87) 1,3-Dichlorobenzene	13.735	146	104883	20.271	ug/l	98
88) 1,4-Dichlorobenzene	13.812	146	107750	20.426	ug/l	99
89) n-Butylbenzene	14.053	91	177482	17.613	ug/l	100
90) Hexachloroethane	14.329	117	33456	18.972	ug/l	99
91) 1,2-Dichlorobenzene	14.106	146	99394	19.995	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	14.717	75	17796	19.155	ug/l	98
93) 1,2,4-Trichlorobenzene	15.394	180	58509	18.433	ug/l	98
94) Hexachlorobutadiene	15.500	225	19592	16.567	ug/l	97
95) Naphthalene	15.635	128	232760	19.700	ug/l	100
96) 1,2,3-Trichlorobenzene	15.841	180	56144	17.802	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061125\
Data File : VN086944.D
Acq On : 11 Jun 2025 13:27
Operator : JC\MD
Sample : VN0611WBS02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0611WBS02

Manual Integrations
APPROVED

Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025

Quant Time: Jun 12 01:31:27 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 02:12:50 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


```
Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061125\  
Data File : VN086944.D  
Acq On    : 11 Jun 2025  13:27  
Operator  : JC\MD  
Sample    : VN0611WBS02  
Misc      : 5.0mL/MSVOA_N/WATER  
ALS Vial  : 6      Sample Multiplier: 1
```


Instrument :
MSVOA_N
ClientSampleId :
VN0611WBS02

Manual Integrations

Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025


Quant Time: Jun 12 01:31:27 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 02:12:50 2025
Response via : Initial Calibration





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.235 min Scan# 1068
Delta R.T. -0.001 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBS02



Tgt Ion:168 Resp: 206799
Ion Ratio Lower Upper
168 100
99 57.5 49.1 73.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#2
Dichlorodifluoromethane
Concen: 19.725 ug/l
RT: 2.153 min Scan# 34
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27

Tgt Ion: 85 Resp: 40672
Ion Ratio Lower Upper
85 100
87 32.5 16.2 48.6









#3
Chloromethane
Concen: 16.079 ug/l
RT: 2.401 min Scan# 75
Delta R.T. 0.006 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27

Instrument :
MSVOA_N
ClientSampleId :
VN0611WBS02

Tgt Ion: 50 Resp: 4281
Ion Ratio Lower Upper
50 100
52 33.8 25.7 38.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#4
Vinyl Chloride
Concen: 18.852 ug/l
RT: 2.553 min Scan# 102
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27


Tgt Ion: 62 Resp: 51780
Ion Ratio Lower Upper
62 100
64 29.7 26.0 39.0





#5
Bromomethane
Concen: 16.641 ug/l
RT: 3.000 min Scan# 11
Delta R.T. 0.006 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBS02



Tgt Ion: 94 Resp: 2557
Ion Ratio Lower Upper
94 100
96 87.8 77.9 116.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#6
Chloroethane
Concen: 19.446 ug/l
RT: 3.153 min Scan# 204
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27

Tgt Ion: 64 Resp: 34500
Ion Ratio Lower Upper
64 100
66 30.3 24.6 36.8









#7
Trichlorofluoromethane
Concen: 19.159 ug/l
RT: 3.530 min Scan# 267
Delta R.T. 0.006 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27

Instrument :
MSVOA_N
ClientSampleId :
VN0611WBS02

Tgt Ion:101 Resp: 68752
Ion Ratio Lower Upper
101 100
103 64.8 52.2 78.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#8
Diethyl Ether
Concen: 19.772 ug/l
RT: 3.983 min Scan# 345
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27

Tgt Ion: 74 Resp: 30914
Ion Ratio Lower Upper
74 100
45 85.5 45.5 136.5









#9
1,1,2-Trichlorotrifluoroethane
Concen: 19.568 ug/l
RT: 4.400 min Scan# 415
Delta R.T. 0.006 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27

Instrument :
MSVOA_N
ClientSampleId :
VN0611WBS02

Tgt Ion:101 Resp: 4411
Ion Ratio Lower Upper
101 100
85 41.5 33.8 50.8
151 74.0 58.7 88.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#10
Methyl Iodide
Concen: 12.634 ug/l
RT: 4.612 min Scan# 452
Delta R.T. 0.006 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27

Tgt Ion:142 Resp: 36924
Ion Ratio Lower Upper
142 100
127 38.9 30.9 46.3
141 12.9 10.4 15.6





#11

Tert butyl alcohol

Concen: 92.831 ug/l

RT: 5.541 min Scan# 611

Delta R.T. -0.006 min

Lab File: VN086944.D

Acq: 11 Jun 2025 13:27

Instrument :

MSVOA_N

ClientSampleId :

VN0611WBS02

Tgt Ion: 59 Resp: 6974

Ion Ratio Lower Upper

59 100





57 9.8 8.1 12.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025



#12

1,1-Dichloroethene

Concen: 19.233 ug/l

RT: 4.365 min Scan# 410

Delta R.T. 0.006 min

Lab File: VN086944.D

Acq: 11 Jun 2025 13:27




Tgt Ion: 96 Resp: 44269


Ion Ratio Lower Upper

96 100

61 153.8 125.8 188.6


98 61.2 52.5 78.7





#13
Acrolein
Concen: 93.730 ug/l
RT: 4.206 min Scan# 382
Delta R.T. 0.006 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBS02



Tgt Ion: 56 Resp: 22290
Ion Ratio Lower Upper
56 100
55 70.9 55.6 83.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#14
Allyl chloride
Concen: 17.821 ug/l
RT: 5.047 min Scan# 526
Delta R.T. 0.006 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27


Tgt Ion: 41 Resp: 68027
Ion Ratio Lower Upper
41 100
39 62.5 51.3 76.9
76 41.4 31.4 47.0





#15
Acrylonitrile
Concen: 95.084 ug/l
RT: 5.741 min Scan# 644
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBS02



Tgt Ion: 53 Resp: 166972
Ion Ratio Lower Upper
53 100
52 81.1 65.0 97.6
51 34.5 27.2 40.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#16
Acetone
Concen: 98.972 ug/l
RT: 4.447 min Scan# 424
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27


Tgt Ion: 43 Resp: 145322
Ion Ratio Lower Upper
43 100
58 35.9 28.0 42.0





#17
Carbon Disulfide
Concen: 17.769 ug/l
RT: 4.742 min Scan# 41
Delta R.T. 0.012 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBS02



Tgt Ion: 76 Resp: 11313
Ion Ratio Lower Upper
76 100
78 9.0 7.3 10.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#18
Methyl Acetate
Concen: 19.013 ug/l
RT: 5.047 min Scan# 526
Delta R.T. 0.006 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27

Tgt Ion: 43 Resp: 81357
Ion Ratio Lower Upper
43 100
74 26.2 20.7 31.1





#19

Methyl tert-butyl Ether

Concen: 19.599 ug/l

RT: 5.812 min Scan# 656

Delta R.T. -0.000 min

Lab File: VN086944.D

Acq: 11 Jun 2025 13:27

Instrument :

MSVOA_N

ClientSampleId :

VN0611WBS02

Tgt Ion: 73 Resp: 163338

Ion Ratio Lower Upper

73 100





57 21.0 17.4 26.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025



#20

Methylene Chloride

Concen: 18.988 ug/l

RT: 5.294 min Scan# 568

Delta R.T. -0.000 min

Lab File: VN086944.D

Acq: 11 Jun 2025 13:27

Tgt Ion: 84 Resp: 52196




Ion Ratio Lower Upper


84 100

49 108.4 90.5 135.7

51 35.0 28.5 42.7


86 64.6 48.1 72.1





#21
trans-1,2-Dichloroethene
Concen: 18.464 ug/l
RT: 5.812 min Scan# 61
Delta R.T. 0.006 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBS02



Tgt Ion: 96 Resp: 47284
Ion Ratio Lower Upper
96 100
61 127.9 109.2 163.8
98 61.9 49.8 74.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#22
Diisopropyl ether
Concen: 19.326 ug/l
RT: 6.688 min Scan# 805
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27


Tgt Ion: 45 Resp: 155508
Ion Ratio Lower Upper
45 100
43 50.5 41.9 62.9
87 29.8 22.4 33.6
59 12.1 9.8 14.6





#23
Vinyl Acetate
Concen: 96.092 ug/l
RT: 6.618 min Scan# 793
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBS02



Tgt Ion: 43 Resp: 653278
Ion Ratio Lower Upper
43 100
86 12.6 9.5 14.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#24
1,1-Dichloroethane
Concen: 19.421 ug/l
RT: 6.588 min Scan# 788
Delta R.T. 0.006 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27


Tgt Ion: 63 Resp: 89930
Ion Ratio Lower Upper
63 100
98 7.3 3.6 10.8
100 5.3 2.6 7.8





#25
2-Butanone
Concen: 90.401 ug/l
RT: 7.494 min Scan# 942
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBS02



Tgt Ion: 43 Resp: 215760
Ion Ratio Lower Upper
43 100
72 30.8 23.0 34.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#26
2,2-Dichloropropane
Concen: 21.663 ug/l
RT: 7.500 min Scan# 943
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27


Tgt Ion: 77 Resp: 78031
Ion Ratio Lower Upper
77 100
97 24.3 11.6 34.7





#27
 cis-1,2-Dichloroethene
 Concen: 19.453 ug/l
 RT: 7.500 min Scan# 943
 Delta R.T. -0.000 min
 Lab File: VN086944.D
 Acq: 11 Jun 2025 13:27




Instrument :
 MSVOA_N
 ClientSampleId :
 VN0611WBS02



Tgt Ion: 96 Resp: 59578
 Ion Ratio Lower Upper
 96 100
 61 138.3 0.0 278.0
 98 64.6 0.0 128.2




Manual Integrations
 APPROVED


Reviewed By :John Carlone 06/12/2025
 Supervised By :Semsettin Yesilyurt 06/12/2025



#28
 Bromochloromethane
 Concen: 19.909 ug/l
 RT: 7.824 min Scan# 998
 Delta R.T. -0.000 min
 Lab File: VN086944.D
 Acq: 11 Jun 2025 13:27

Tgt Ion: 49 Resp: 45329
 Ion Ratio Lower Upper
 49 100
 129 1.6 0.0 3.8
 130 82.2 62.2 93.2





#29

Tetrahydrofuran

Concen: 92.532 ug/l

RT: 7.853 min Scan# 1003

Delta R.T. -0.000 min

Lab File: VN086944.D

Acq: 11 Jun 2025 13:27

Instrument :

MSVOA_N

ClientSampleId :

VN0611WBS02

Tgt Ion: 42 Resp: 14386

Ion Ratio Lower Upper

42 100

72 49.4 37.8 56.6





71 46.2 35.6 53.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025



#30

Chloroform

Concen: 19.521 ug/l

RT: 7.977 min Scan# 1024

Delta R.T. -0.000 min

Lab File: VN086944.D




Acq: 11 Jun 2025 13:27


Tgt Ion: 83 Resp: 90272

Ion Ratio Lower Upper

83 100


85 64.2 52.9 79.3





#31
Cyclohexane
Concen: 17.603 ug/l
RT: 8.265 min Scan# 1073
Delta R.T. -0.006 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBS02




Tgt Ion: 56 Resp: 79050
Ion Ratio Lower Upper
56 100
69 35.7 25.4 38.0
84 87.1 68.8 103.2

Manual Integrations
APPROVED



Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025




#32
1,1,1-Trichloroethane
Concen: 19.033 ug/l
RT: 8.182 min Scan# 1059
Delta R.T. 0.006 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27



Tgt Ion: 97 Resp: 74861
Ion Ratio Lower Upper
97 100
99 64.2 52.0 78.0
61 47.9 35.9 53.9





#33

1,2-Dichloroethane-d4

Concen: 47.075 ug/l

RT: 8.588 min Scan# 1128

Delta R.T. -0.000 min

Lab File: VN086944.D

Acq: 11 Jun 2025 13:27

Instrument :

MSVOA_N

ClientSampleId :

VN0611WBS02

Tgt Ion: 65 Resp: 13034

Ion Ratio Lower Upper

65 100





67 54.7 0.0 105.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.106 min Scan# 1216

Delta R.T. -0.000 min

Lab File: VN086944.D

Acq: 11 Jun 2025 13:27




Tgt Ion:114 Resp: 370524


Ion Ratio Lower Upper

114 100

63 18.7 0.0 39.6

88 16.4 0.0 30.2





#35

Dibromofluoromethane

Concen: 51.096 ug/l

RT: 8.177 min Scan# 1057

Delta R.T. 0.006 min

Lab File: VN086944.D

Acq: 11 Jun 2025 13:27

Instrument :

MSVOA_N

ClientSampleId :

VN0611WBS02

Tgt Ion:113 Resp: 112198

Ion Ratio Lower Upper

113 100

111 102.8 84.2 126.2





192 17.9 14.2 21.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025



#36

1,1-Dichloropropene

Concen: 19.431 ug/l

RT: 8.377 min Scan# 1092

Delta R.T. -0.000 min

Lab File: VN086944.D

Acq: 11 Jun 2025 13:27




Tgt Ion: 75 Resp: 63577


Ion Ratio Lower Upper

75 100

110 35.1 17.6 52.9

77 30.1 25.3 37.9





#37

Ethyl Acetate

Concen: 19.306 ug/l

RT: 7.571 min Scan# 91

Delta R.T. -0.006 min

Lab File: VN086944.D

Acq: 11 Jun 2025 13:27

Instrument :

MSVOA_N

ClientSampleId :

VN0611WBS02





Tgt Ion:	43	Resp:	8041
Ion	Ratio	Lower	Upper
43	100		
61	15.4	12.1	18.1
70	12.3	9.8	14.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025



#38

Carbon Tetrachloride

Concen: 19.222 ug/l




RT: 8.371 min Scan# 1091


Delta R.T. -0.000 min

Lab File: VN086944.D

Acq: 11 Jun 2025 13:27

Tgt Ion:	117	Resp:	61747
Ion	Ratio	Lower	Upper
117	100		
119	97.1	75.0	112.6
121	32.6	24.9	37.3





#39

Methylcyclohexane

Concen: 16.947 ug/l

RT: 9.606 min Scan# 11301

Delta R.T. -0.000 min

Lab File: VN086944.D

Acq: 11 Jun 2025 13:27

Instrument :

MSVOA_N

ClientSampleId :

VN0611WBS02





Tgt Ion:	83	Resp:	7597
Ion	Ratio	Lower	Upper
83	100		
55	73.2	61.0	91.6
98	48.9	38.9	58.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025



#40

Benzene

Concen: 19.277 ug/l




RT: 8.612 min Scan# 1132


Delta R.T. -0.000 min

Lab File: VN086944.D

Acq: 11 Jun 2025 13:27


Tgt Ion:	78	Resp:	206255
Ion	Ratio	Lower	Upper
78	100		
77	23.1	18.7	28.1





#41
Methacrylonitrile
Concen: 18.218 ug/l
RT: 7.788 min Scan# 992
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBS02



Tgt Ion: 41 Resp: 4261
Ion Ratio Lower Upper
41 100
39 53.5 41.2 61.8
67 85.5 61.4 92.2
52 31.4 22.7 34.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#42
1,2-Dichloroethane
Concen: 19.530 ug/l
RT: 8.677 min Scan# 1143
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27


Tgt Ion: 62 Resp: 63379
Ion Ratio Lower Upper
62 100
98 11.4 0.0 21.2





#43
Isopropyl Acetate
Concen: 18.984 ug/l
RT: 8.694 min Scan# 1146
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBS02



Tgt Ion: 43 Resp: 126914
Ion Ratio Lower Upper
43 100
61 28.5 22.1 33.1
87 15.6 11.8 17.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#44
Trichloroethene
Concen: 19.942 ug/l
RT: 9.359 min Scan# 1259
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27

Tgt Ion:130 Resp: 50593
Ion Ratio Lower Upper
130 100
95 99.6 0.0 196.2





#45

1,2-Dichloropropane

Concen: 19.628 ug/l

RT: 9.624 min Scan# 1304

Delta R.T. -0.006 min

Lab File: VN086944.D

Acq: 11 Jun 2025 13:27

Instrument :

MSVOA_N

ClientSampleId :

VN0611WBS02

Tgt Ion: 63 Resp: 5109

Ion Ratio Lower Upper

63 100





65 34.4 25.5 38.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025



#46

Dibromomethane

Concen: 20.093 ug/l

RT: 9.712 min Scan# 1319

Delta R.T. -0.006 min

Lab File: VN086944.D

Acq: 11 Jun 2025 13:27




Tgt Ion: 93 Resp: 34741


Ion Ratio Lower Upper

93 100

95 85.1 67.1 100.7


174 94.0 74.2 111.4





#47
Bromodichloromethane
Concen: 19.634 ug/l
RT: 9.894 min Scan# 1350
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBS02



Tgt Ion: 83 Resp: 6984
Ion Ratio Lower Upper
83 100
85 65.0 50.9 76.3
127 8.1 6.6 10.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#48
Methyl methacrylate
Concen: 18.771 ug/l
RT: 9.682 min Scan# 1314
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27

Tgt Ion: 41 Resp: 57757
Ion Ratio Lower Upper
41 100
69 93.3 74.6 112.0
39 48.5 40.2 60.4





#49

1,4-Dioxane

Concen: 409.347 ug/l

RT: 9.706 min Scan# 11

Delta R.T. 0.006 min

Lab File: VN086944.D

Acq: 11 Jun 2025 13:27

Instrument :

MSVOA_N

ClientSampleId :

VN0611WBS02

Tgt Ion: 88 Resp: 22914

Ion Ratio Lower Upper

88 100

43 0.0 0.0 0.0





58 69.9 57.7 86.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025



#50

Toluene-d8

Concen: 48.706 ug/l

RT: 10.571 min Scan# 1465

Delta R.T. -0.000 min

Lab File: VN086944.D




Acq: 11 Jun 2025 13:27


Tgt Ion: 98 Resp: 423380

Ion Ratio Lower Upper

98 100


100 67.3 53.4 80.0





#51
4-Methyl-2-Pentanone
Concen: 97.997 ug/l
RT: 10.447 min Scan# 1444
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBS02



Tgt Ion: 43 Resp: 394390
Ion Ratio Lower Upper
43 100
58 42.4 34.2 51.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#52
Toluene
Concen: 19.509 ug/l
RT: 10.629 min Scan# 1475
Delta R.T. -0.006 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27


Tgt Ion: 92 Resp: 127563
Ion Ratio Lower Upper
92 100
91 171.7 136.5 204.7





#53
t-1,3-Dichloropropene
Concen: 19.917 ug/l
RT: 10.835 min Scan# 1510
Delta R.T. -0.006 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBS02



Tgt Ion: 75 Resp: 79224
Ion Ratio Lower Upper
75 100
77 33.3 26.4 39.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#54
cis-1,3-Dichloropropene
Concen: 20.109 ug/l
RT: 10.318 min Scan# 1422
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27


Tgt Ion: 75 Resp: 85586
Ion Ratio Lower Upper
75 100
77 33.6 26.2 39.4
39 39.5 32.3 48.5





#55
1,1,2-Trichloroethane
Concen: 20.438 ug/l
RT: 11.018 min Scan# 1541
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBS02



Tgt Ion: 97 Resp: 5142
Ion Ratio Lower Upper
97 100
83 79.7 69.7 104.5
85 51.4 45.1 67.7
99 61.3 50.6 75.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#56
Ethyl methacrylate
Concen: 20.198 ug/l
RT: 10.882 min Scan# 1518
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27

Tgt Ion: 69 Resp: 80942
Ion Ratio Lower Upper
69 100
41 55.5 47.2 70.8
39 24.3 23.0 34.6









#57
1,3-Dichloropropane
Concen: 19.613 ug/l
RT: 11.165 min Scan# 1566
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27

Instrument :
MSVOA_N
ClientSampleId :
VN0611WBS02

Tgt Ion: 76 Resp: 85602
Ion Ratio Lower Upper
76 100
78 32.4 25.3 37.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#58
2-Chloroethyl Vinyl ether
Concen: 101.910 ug/l
RT: 10.165 min Scan# 1396
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27


Tgt Ion: 63 Resp: 243584
Ion Ratio Lower Upper
63 100
106 29.2 23.0 34.6





#59
2-Hexanone
Concen: 89.210 ug/l
RT: 11.206 min Scan# 1572
Delta R.T. 0.006 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBS02



Tgt Ion: 43 Resp: 231264
Ion Ratio Lower Upper
43 100
58 65.0 30.4 91.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#60
Dibromochloromethane
Concen: 20.468 ug/l
RT: 11.359 min Scan# 1599
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27


Tgt Ion:129 Resp: 53655
Ion Ratio Lower Upper
129 100
127 76.3 38.6 115.8





#61
1,2-Dibromoethane
Concen: 19.711 ug/l
RT: 11.471 min Scan# 1618
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBS02



Tgt Ion:107 Resp: 50832
Ion Ratio Lower Upper
107 100
109 96.3 76.3 114.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#62
4-Bromofluorobenzene
Concen: 49.951 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27

Tgt Ion: 95 Resp: 161321
Ion Ratio Lower Upper
95 100
174 71.0 0.0 141.8
176 67.2 0.0 132.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 1685
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27

Instrument :
MSVOA_N
ClientSampleId :
VN0611WBS02

Tgt Ion:117 Resp: 32118
Ion Ratio Lower Upper
117 100
82 54.7 44.6 67.0
119 31.9 25.5 38.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#64
Tetrachloroethene
Concen: 19.150 ug/l
RT: 11.106 min Scan# 1556
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27


Tgt Ion:164 Resp: 38926
Ion Ratio Lower Upper
164 100
166 128.0 101.8 152.8
129 106.5 83.0 124.4
131 100.0 78.2 117.2





#65
Chlorobenzene
Concen: 20.006 ug/l
RT: 11.894 min Scan# 111
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBS02



Tgt Ion:112 Resp: 14172
Ion Ratio Lower Upper
112 100
114 31.4 26.2 39.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#66
1,1,1,2-Tetrachloroethane
Concen: 20.378 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. -0.006 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27

Tgt Ion:131 Resp: 46404
Ion Ratio Lower Upper
131 100
133 94.2 48.3 144.8
119 68.0 34.0 101.8









#67
Ethyl Benzene
Concen: 19.647 ug/l
RT: 11.965 min Scan# 1702
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27

Instrument :
MSVOA_N
ClientSampleId :
VN0611WBS02

Tgt Ion: 91 Resp: 239733
Ion Ratio Lower Upper
91 100
106 31.9 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#68
m/p-Xylenes
Concen: 39.613 ug/l
RT: 12.070 min Scan# 1720
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27


Tgt Ion:106 Resp: 185026
Ion Ratio Lower Upper
106 100
91 197.2 159.4 239.0





#69
o-Xylene
Concen: 20.172 ug/l
RT: 12.400 min Scan# 1776
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBS02



Tgt Ion:106 Resp: 90240
Ion Ratio Lower Upper
106 100
91 210.1 106.1 318.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#70
Styrene
Concen: 20.137 ug/l
RT: 12.412 min Scan# 1778
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27


Tgt Ion:104 Resp: 154149
Ion Ratio Lower Upper
104 100
78 47.5 38.5 57.7
103 54.3 43.1 64.7





#71
Bromoform
Concen: 20.952 ug/l
RT: 12.582 min Scan# 11
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBS02




Tgt Ion:173 Resp: 35350
Ion Ratio Lower Upper
173 100
175 51.2 24.1 72.4
254 0.0 0.0 0.0

Manual Integrations
APPROVED



Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.788 min Scan# 2012
Delta R.T. -0.006 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27




Tgt Ion:152 Resp: 157403
Ion Ratio Lower Upper
152 100
115 61.4 30.1 90.5
150 164.6 0.0 345.0





#73
Isopropylbenzene
Concen: 19.320 ug/l
RT: 12.694 min Scan# 11
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBS02




Tgt Ion:105 Resp: 22154
Ion Ratio Lower Upper
105 100
120 27.3 13.5 40.4

Manual Integrations
APPROVED



Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025




#74
N-ethyl acetate
Concen: 15.794 ug/l
RT: 12.529 min Scan# 1798
Delta R.T. 0.017 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27




Tgt Ion: 43 Resp: 63289
Ion Ratio Lower Upper
43 100
70 34.9 37.6 56.4#
55 19.9 19.0 28.6
61 18.0 20.8 31.2#





#75
1,1,2,2-Tetrachloroethane
Concen: 20.973 ug/l
RT: 12.941 min Scan# 1868
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBS02



Tgt Ion: 83 Resp: 81470
Ion Ratio Lower Upper
83 100
131 9.3 4.9 14.7
85 64.3 32.0 96.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#76
1,2,3-Trichloropropane
Concen: 19.132 ug/l m
RT: 12.994 min Scan# 1877
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27

Tgt Ion: 75 Resp: 71582
Ion Ratio Lower Upper
75 100
77 177.6 79.3 237.9









#77
Bromobenzene
Concen: 20.981 ug/l
RT: 12.982 min Scan# 1875
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27

Instrument :
MSVOA_N
ClientSampleId :
VN0611WBS02

Tgt Ion:156 Resp: 55170
Ion Ratio Lower Upper
156 100
77 230.4 117.7 353.1
158 95.7 48.1 144.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#78
n-propylbenzene
Concen: 19.187 ug/l
RT: 13.035 min Scan# 1884
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27

Tgt Ion: 91 Resp: 267384
Ion Ratio Lower Upper
91 100
120 22.9 11.4 34.2









#79
2-Chlorotoluene
Concen: 19.721 ug/l
RT: 13.123 min Scan# 1899
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27

Instrument :
MSVOA_N
ClientSampleId :
VN0611WBS02

Tgt Ion: 91 Resp: 16481
Ion Ratio Lower Upper
91 100
126 32.6 16.2 48.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#80
1,3,5-Trimethylbenzene
Concen: 19.200 ug/l
RT: 13.170 min Scan# 1907
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27


Tgt Ion:105 Resp: 181773
Ion Ratio Lower Upper
105 100
120 50.4 24.9 74.6





#81
trans-1,4-Dichloro-2-butene
Concen: 18.532 ug/l
RT: 12.735 min Scan# 1833
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBS02



Tgt Ion: 75 Resp: 3012
Ion Ratio Lower Upper
75 100
53 94.2 66.6 99.8
89 63.9 36.7 55.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#82
4-Chlorotoluene
Concen: 19.866 ug/l
RT: 13.217 min Scan# 1915
Delta R.T. -0.006 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27


Tgt Ion: 91 Resp: 167991
Ion Ratio Lower Upper
91 100
126 33.4 16.7 50.1





#83
tert-Butylbenzene
Concen: 18.646 ug/l
RT: 13.435 min Scan# 1952
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBS02




Tgt Ion:119 Resp: 16159
Ion Ratio Lower Upper
119 100
91 63.1 32.1 96.5
134 26.9 13.4 40.1

Manual Integrations
APPROVED



Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025




#84
1,2,4-Trimethylbenzene
Concen: 19.216 ug/l
RT: 13.482 min Scan# 1960
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27



Tgt Ion:105 Resp: 182431
Ion Ratio Lower Upper
105 100
120 47.6 23.2 69.6









#85
sec-Butylbenzene
Concen: 18.189 ug/l
RT: 13.617 min Scan# 1983
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27

Instrument :
MSVOA_N
ClientSampleId :
VN0611WBS02

Tgt Ion:105 Resp: 22890
Ion Ratio Lower Upper
105 100
134 19.9 10.0 30.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#86
p-Isopropyltoluene
Concen: 18.551 ug/l
RT: 13.729 min Scan# 2002
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27

Tgt Ion:119 Resp: 192991
Ion Ratio Lower Upper
119 100
134 26.2 13.5 40.4
91 22.1 11.5 34.4





#87

1,3-Dichlorobenzene

Concen: 20.271 ug/l

RT: 13.735 min Scan# 2003

Delta R.T. -0.000 min

Lab File: VN086944.D

Acq: 11 Jun 2025 13:27

Instrument :

MSVOA_N

ClientSampleId :

VN0611WBS02

Tgt Ion:146 Resp: 10488

Ion Ratio Lower Upper

146 100

111 41.3 20.9 62.8





148 62.6 32.0 96.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025



#88

1,4-Dichlorobenzene

Concen: 20.426 ug/l

RT: 13.812 min Scan# 2016

Delta R.T. -0.000 min

Lab File: VN086944.D

Acq: 11 Jun 2025 13:27




Tgt Ion:146 Resp: 107750


Ion Ratio Lower Upper

146 100

111 40.7 19.9 59.7

148 63.1 31.8 95.4





#89

n-Butylbenzene

Concen: 17.613 ug/l

RT: 14.053 min Scan# 2057

Delta R.T. -0.000 min

Lab File: VN086944.D

Acq: 11 Jun 2025 13:27

Instrument :

MSVOA_N

ClientSampleId :

VN0611WBS02

Tgt Ion: 91 Resp: 17748

Ion Ratio Lower Upper

91 100

92 54.9 27.5 82.4





134 26.7 13.0 39.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025



#90

Hexachloroethane

Concen: 18.972 ug/l

RT: 14.329 min Scan# 2104

Delta R.T. -0.006 min

Lab File: VN086944.D




Acq: 11 Jun 2025 13:27


Tgt Ion:117 Resp: 33456

Ion Ratio Lower Upper

117 100

201 67.1 33.2 99.6









#91
1,2-Dichlorobenzene
Concen: 19.995 ug/l
RT: 14.106 min Scan# 2066
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27

Instrument :
MSVOA_N
ClientSampleId :
VN0611WBS02

Tgt Ion:146 Resp: 99394
Ion Ratio Lower Upper
146 100
111 44.5 21.6 64.6
148 64.5 31.7 95.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#92
1,2-Dibromo-3-Chloropropane
Concen: 19.155 ug/l
RT: 14.717 min Scan# 2170
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27


Tgt Ion: 75 Resp: 17796
Ion Ratio Lower Upper
75 100
155 85.2 41.1 123.4
157 106.8 52.7 158.1





#93
1,2,4-Trichlorobenzene
Concen: 18.433 ug/l
RT: 15.394 min Scan# 21
Delta R.T. 0.006 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBS02




Tgt Ion:180 Resp: 58509
Ion Ratio Lower Upper
180 100
182 96.7 46.9 140.7
145 32.3 16.0 47.9

Manual Integrations
APPROVED



Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025




#94
Hexachlorobutadiene
Concen: 16.567 ug/l
RT: 15.500 min Scan# 2303
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27




Tgt Ion:225 Resp: 19592
Ion Ratio Lower Upper
225 100
223 64.0 33.5 100.4
227 60.0 31.1 93.5





#95
Naphthalene
Concen: 19.700 ug/l
RT: 15.635 min Scan# 2327
Delta R.T. -0.006 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBS02



Tgt Ion:128 Resp: 232760
Ion Ratio Lower Upper
128 100
127 12.6 10.2 15.2
129 10.9 8.8 13.2




Manual Integrations
APPROVED

Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#96
1,2,3-Trichlorobenzene
Concen: 17.802 ug/l
RT: 15.841 min Scan# 2361
Delta R.T. 0.006 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27

Tgt Ion:180 Resp: 56144
Ion Ratio Lower Upper
180 100
182 96.0 47.0 141.0
145 33.6 16.6 49.8





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VN0611WBSD02		SDG No.:	Q2263
Lab Sample ID:	VN0611WBSD02		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086945.D	1		06/11/25 14:01	VN061125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	19.1		0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	15.5		0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	18.7		0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	16.2		1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	18.9		0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	19.0		0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	18.7		0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	19.1		0.23	0.75	1.00	ug/L
67-64-1	Acetone	92.4		1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	17.2		0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	20.2		0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	20.1		0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	19.1		0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	18.6		0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	19.3		0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	17.2		1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	94.9		0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	19.5		0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	19.6		0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	21.1		0.22	0.50	1.00	ug/L
67-66-3	Chloroform	19.4		0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	18.8		0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	16.6		0.16	0.50	1.00	ug/L
71-43-2	Benzene	19.7		0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	20.3		0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	20.1		0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	20.0		0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	20.7		0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	100		0.68	2.50	5.00	ug/L
108-88-3	Toluene	19.9		0.14	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VN0611WBSD02		SDG No.:	Q2263
Lab Sample ID:	VN0611WBSD02		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086945.D	1		06/11/25 14:01	VN061125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	20.7		0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	20.7		0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	20.9		0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	93.8		0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	21.2		0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	20.7		0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	18.5		0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	20.2		0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	19.3		0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	39.5		0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	20.0		0.12	0.50	1.00	ug/L
100-42-5	Styrene	20.6		0.15	0.50	1.00	ug/L
75-25-2	Bromoform	22.4		0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	19.3		0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	21.7		0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	20.5		0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	20.4		0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	20.6		0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	20.8		0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	18.7		0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	18.1		0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	47.9		81 - 118		96%	SPK: 50
1868-53-7	Dibromofluoromethane	52.3		80 - 119		105%	SPK: 50
2037-26-5	Toluene-d8	49.5		89 - 112		99%	SPK: 50
460-00-4	4-Bromofluorobenzene	50.1		85 - 114		100%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	200000	8.229				
540-36-3	1,4-Difluorobenzene	354000	9.106				
3114-55-4	Chlorobenzene-d5	311000	11.865				
3855-82-1	1,4-Dichlorobenzene-d4	152000	13.788				

Report of Analysis

Client:	AECOM Technical Services, Inc.			Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258			Date Received:	
Client Sample ID:	VN0611WBSD02			SDG No.:	Q2263
Lab Sample ID:	VN0611WBSD02			Matrix:	Water
Analytical Method:	8260D			% Solid:	0
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000 uL
Soil Aliquot Vol:			uL	Test:	VOCMS Group1
GC Column:	RXI-624	ID :	0.25	Level :	LOW
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086945.D	1		06/11/25 14:01	VN061125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
() = Laboratory InHouse Limit
A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061125\
 Data File : VN086945.D
 Acq On : 11 Jun 2025 14:01
 Operator : JC\MD
 Sample : VN0611WBSD02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0611WBSD02

Manual Integrations
 APPROVED

Reviewed By : John Carlone 06/12/2025
 Supervised By : Semsettin Yesilyurt 06/12/2025

Quant Time: Jun 12 01:32:16 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	199904	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	354069	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	310598	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	151691	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	128166	47.886	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	95.780%		
35) Dibromofluoromethane	8.177	113	109657	52.260	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	104.520%		
50) Toluene-d8	10.571	98	411227	49.506	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	99.020%		
62) 4-Bromofluorobenzene	12.847	95	154682	50.121	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	100.240%		
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	2.153	85	38145	19.138	ug/l	96
3) Chloromethane	2.400	50	39948	15.521	ug/l	99
4) Vinyl Chloride	2.553	62	49695	18.717	ug/l	97
5) Bromomethane	3.000	94	24041	16.181	ug/l	95
6) Chloroethane	3.153	64	32368	18.874	ug/l	100
7) Trichlorofluoromethane	3.530	101	66071	19.047	ug/l	96
8) Diethyl Ether	3.977	74	29872	19.765	ug/l	99
9) 1,1,2-Trichlorotrifluo...	4.394	101	40831	18.735	ug/l	97
10) Methyl Iodide	4.612	142	36500	12.919	ug/l	100
11) Tert butyl alcohol	5.541	59	71180	98.012	ug/l	100
12) 1,1-Dichloroethene	4.365	96	42515	19.108	ug/l	92
13) Acrolein	4.200	56	22554	98.112	ug/l	100
14) Allyl chloride	5.053	41	65845	17.844	ug/l	98
15) Acrylonitrile	5.736	53	167429	98.634	ug/l	100
16) Acetone	4.447	43	131189	92.428	ug/l	97
17) Carbon Disulfide	4.736	76	105790	17.189	ug/l	98
18) Methyl Acetate	5.047	43	83025	20.072	ug/l	99
19) Methyl tert-butyl Ether	5.818	73	162549	20.177	ug/l	99
20) Methylene Chloride	5.300	84	50840	19.132	ug/l	96
21) trans-1,2-Dichloroethene	5.806	96	46018	18.589	ug/l	99
22) Diisopropyl ether	6.688	45	152785	19.643	ug/l	98
23) Vinyl Acetate	6.618	43	656574	99.908	ug/l	98
24) 1,1-Dichloroethane	6.588	63	86331	19.287	ug/l	99
25) 2-Butanone	7.494	43	218999	94.923	ug/l	99
26) 2,2-Dichloropropane	7.506	77	74316	21.343	ug/l	98
27) cis-1,2-Dichloroethene	7.500	96	57947	19.573	ug/l	97
28) Bromochloromethane	7.824	49	46411	21.088	ug/l	97
29) Tetrahydrofuran	7.853	42	145366	96.721	ug/l	98
30) Chloroform	7.977	83	86604	19.374	ug/l	100
31) Cyclohexane	8.265	56	74510	17.164	ug/l	98
32) 1,1,1-Trichloroethane	8.177	97	71441	18.790	ug/l	98
36) 1,1-Dichloropropene	8.382	75	60912	19.481	ug/l	99
37) Ethyl Acetate	7.571	43	80010	20.102	ug/l	99
38) Carbon Tetrachloride	8.371	117	59839	19.494	ug/l	96
39) Methylcyclohexane	9.606	83	71058	16.588	ug/l	97
40) Benzene	8.612	78	200940	19.653	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061125\
 Data File : VN086945.D
 Acq On : 11 Jun 2025 14:01
 Operator : JC\MD
 Sample : VN0611WBSD02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0611WBSD02

Manual Integrations
 APPROVED

Reviewed By :John Carlone 06/12/2025
 Supervised By :Semsettin Yesilyurt 06/12/2025

Quant Time: Jun 12 01:32:16 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.788	41	46047	20.600	ug/l	98
42) 1,2-Dichloroethane	8.677	62	63066	20.336	ug/l	100
43) Isopropyl Acetate	8.694	43	127320	19.930	ug/l	99
44) Trichloroethene	9.359	130	48830	20.141	ug/l	97
45) 1,2-Dichloropropane	9.624	63	49815	20.027	ug/l	95
46) Dibromomethane	9.712	93	35176	21.290	ug/l	99
47) Bromodichloromethane	9.894	83	70222	20.658	ug/l	99
48) Methyl methacrylate	9.688	41	56099	19.079	ug/l	95
49) 1,4-Dioxane	9.700	88	24300	454.282	ug/l #	93
51) 4-Methyl-2-Pentanone	10.447	43	399735	103.940	ug/l	99
52) Toluene	10.629	92	124298	19.893	ug/l	100
53) t-1,3-Dichloropropene	10.841	75	78850	20.744	ug/l	100
54) cis-1,3-Dichloropropene	10.318	75	84225	20.709	ug/l	97
55) 1,1,2-Trichloroethane	11.018	97	50307	20.924	ug/l	97
56) Ethyl methacrylate	10.882	69	79449	20.747	ug/l	97
57) 1,3-Dichloropropane	11.165	76	85222	20.434	ug/l	98
58) 2-Chloroethyl Vinyl ether	10.165	63	241856	105.889	ug/l	99
59) 2-Hexanone	11.206	43	232463	93.840	ug/l	95
60) Dibromochloromethane	11.359	129	53154	21.220	ug/l	100
61) 1,2-Dibromoethane	11.470	107	51056	20.718	ug/l	99
64) Tetrachloroethene	11.106	164	36363	18.499	ug/l	98
65) Chlorobenzene	11.894	112	138224	20.178	ug/l	99
66) 1,1,1,2-Tetrachloroethane	11.965	131	45014	20.442	ug/l	99
67) Ethyl Benzene	11.965	91	227138	19.250	ug/l	100
68) m/p-Xylenes	12.070	106	178421	39.501	ug/l	98
69) o-Xylene	12.400	106	86537	20.004	ug/l	100
70) Styrene	12.412	104	152236	20.565	ug/l	98
71) Bromoform	12.582	173	36528	22.388	ug/l #	98
73) Isopropylbenzene	12.694	105	213019	19.276	ug/l	100
74) N-amyl acetate	12.529	43	69705	18.051	ug/l #	88
75) 1,1,2,2-Tetrachloroethane	12.935	83	81389	21.740	ug/l	100
76) 1,2,3-Trichloropropane	12.994	75	71147m	19.732	ug/l	
77) Bromobenzene	12.976	156	54030	21.319	ug/l	98
78) n-propylbenzene	13.035	91	257426	19.168	ug/l	99
79) 2-Chlorotoluene	13.123	91	157627	19.571	ug/l	99
80) 1,3,5-Trimethylbenzene	13.170	105	175428	19.228	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.735	75	29703	18.963	ug/l	95
82) 4-Chlorotoluene	13.223	91	162421	19.931	ug/l	100
83) tert-Butylbenzene	13.435	119	156588	18.749	ug/l	98
84) 1,2,4-Trimethylbenzene	13.482	105	177241	19.373	ug/l	99
85) sec-Butylbenzene	13.611	105	219639	18.109	ug/l	99
86) p-Isopropyltoluene	13.729	119	182853	18.238	ug/l	99
87) 1,3-Dichlorobenzene	13.729	146	102302	20.517	ug/l	100
88) 1,4-Dichlorobenzene	13.811	146	103614	20.382	ug/l	98
89) n-Butylbenzene	14.053	91	171576	17.668	ug/l	99
90) Hexachloroethane	14.329	117	31966	18.810	ug/l	100
91) 1,2-Dichlorobenzene	14.106	146	98864	20.638	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.723	75	18610	20.786	ug/l	99
93) 1,2,4-Trichlorobenzene	15.388	180	57083	18.660	ug/l	99
94) Hexachlorobutadiene	15.500	225	17672	15.506	ug/l	98
95) Naphthalene	15.635	128	230927	20.281	ug/l	99
96) 1,2,3-Trichlorobenzene	15.835	180	55026	18.105	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061125\
Data File : VN086945.D
Acq On : 11 Jun 2025 14:01
Operator : JC\MD
Sample : VN0611WBSD02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0611WBSD02

Manual Integrations
APPROVED

Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025

Quant Time: Jun 12 01:32:16 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 02:12:50 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :
MSVOA_N
ClientSampleId :
VN0611WBSD02

Manual Integrations


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.229 min Scan# 1068
Delta R.T. -0.007 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBSD02



Tgt Ion:168 Resp: 199904
Ion Ratio Lower Upper
168 100
99 58.1 49.1 73.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#2
Dichlorodifluoromethane
Concen: 19.138 ug/l
RT: 2.153 min Scan# 34
Delta R.T. -0.000 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01


Tgt Ion: 85 Resp: 38145
Ion Ratio Lower Upper
85 100
87 30.2 16.2 48.6





#3
Chloromethane
Concen: 15.521 ug/l
RT: 2.400 min Scan# 75
Delta R.T. 0.006 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBSD02



Tgt Ion: 50 Resp: 39948
Ion Ratio Lower Upper
50 100
52 32.7 25.7 38.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#4
Vinyl Chloride
Concen: 18.717 ug/l
RT: 2.553 min Scan# 102
Delta R.T. -0.000 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01


Tgt Ion: 62 Resp: 49695
Ion Ratio Lower Upper
62 100
64 30.8 26.0 39.0





#5
Bromomethane
Concen: 16.181 ug/l
RT: 3.000 min Scan# 177
Delta R.T. 0.006 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBSD02



Tgt Ion: 94 Resp: 2404
Ion Ratio Lower Upper
94 100
96 92.1 77.9 116.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#6
Chloroethane
Concen: 18.874 ug/l
RT: 3.153 min Scan# 204
Delta R.T. -0.000 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01


Tgt Ion: 64 Resp: 32368
Ion Ratio Lower Upper
64 100
66 30.6 24.6 36.8





#7
Trichlorofluoromethane
Concen: 19.047 ug/l
RT: 3.530 min Scan# 20
Delta R.T. 0.006 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBSD02



Tgt Ion:101 Resp: 6607
Ion Ratio Lower Upper
101 100
103 62.1 52.2 78.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#8
Diethyl Ether
Concen: 19.765 ug/l
RT: 3.977 min Scan# 344
Delta R.T. -0.006 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01

Tgt Ion: 74 Resp: 29872
Ion Ratio Lower Upper
74 100
45 90.0 45.5 136.5









#9
1,1,2-Trichlorotrifluoroethane
Concen: 18.735 ug/l
RT: 4.394 min Scan# 415
Delta R.T. -0.000 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01

Instrument :
MSVOA_N
ClientSampleId :
VN0611WBSD02

Tgt Ion:101 Resp: 4083
Ion Ratio Lower Upper
101 100
85 39.9 33.8 50.8
151 75.2 58.7 88.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#10
Methyl Iodide
Concen: 12.919 ug/l
RT: 4.612 min Scan# 452
Delta R.T. 0.006 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01

Tgt Ion:142 Resp: 36500
Ion Ratio Lower Upper
142 100
127 38.4 30.9 46.3
141 13.1 10.4 15.6





#11

Tert butyl alcohol

Concen: 98.012 ug/l

RT: 5.541 min Scan# 611

Delta R.T. -0.006 min

Lab File: VN086945.D

Acq: 11 Jun 2025 14:01

Instrument :

MSVOA_N

ClientSampleId :

VN0611WBSD02

Tgt Ion: 59 Resp: 71180

Ion Ratio Lower Upper

59 100





57 10.2 8.1 12.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025



#12

1,1-Dichloroethene

Concen: 19.108 ug/l

RT: 4.365 min Scan# 410

Delta R.T. 0.006 min

Lab File: VN086945.D

Acq: 11 Jun 2025 14:01




Tgt Ion: 96 Resp: 42515


Ion Ratio Lower Upper

96 100

61 144.2 125.8 188.6


98 62.9 52.5 78.7





#13
Acrolein
Concen: 98.112 ug/l
RT: 4.200 min Scan# 382
Delta R.T. -0.000 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBSD02



Tgt Ion: 56 Resp: 22554
Ion Ratio Lower Upper
56 100
55 69.3 55.6 83.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#14
Allyl chloride
Concen: 17.844 ug/l
RT: 5.053 min Scan# 527
Delta R.T. 0.012 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01


Tgt Ion: 41 Resp: 65845
Ion Ratio Lower Upper
41 100
39 64.2 51.3 76.9
76 42.0 31.4 47.0





#15
Acrylonitrile
Concen: 98.634 ug/l
RT: 5.736 min Scan# 644
Delta R.T. -0.006 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBSD02



Tgt Ion: 53 Resp: 167429
Ion Ratio Lower Upper
53 100
52 81.7 65.0 97.6
51 33.9 27.2 40.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#16
Acetone
Concen: 92.428 ug/l
RT: 4.447 min Scan# 424
Delta R.T. -0.000 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01


Tgt Ion: 43 Resp: 131189
Ion Ratio Lower Upper
43 100
58 36.9 28.0 42.0





#17
Carbon Disulfide
Concen: 17.189 ug/l
RT: 4.736 min Scan# 473
Delta R.T. 0.006 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBSD02



Tgt Ion: 76 Resp: 105790
Ion Ratio Lower Upper
76 100
78 8.4 7.3 10.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#18
Methyl Acetate
Concen: 20.072 ug/l
RT: 5.047 min Scan# 526
Delta R.T. 0.006 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01

Tgt Ion: 43 Resp: 83025
Ion Ratio Lower Upper
43 100
74 26.6 20.7 31.1





#19

Methyl tert-butyl Ether

Concen: 20.177 ug/l

RT: 5.818 min Scan# 61

Delta R.T. 0.006 min

Lab File: VN086945.D

Acq: 11 Jun 2025 14:01

Instrument :

MSVOA_N

ClientSampleId :

VN0611WBSD02

Tgt Ion: 73 Resp: 162549

Ion Ratio Lower Upper

73 100





57 22.2 17.4 26.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025



#20

Methylene Chloride

Concen: 19.132 ug/l

RT: 5.300 min Scan# 569

Delta R.T. 0.006 min

Lab File: VN086945.D

Acq: 11 Jun 2025 14:01

Tgt Ion: 84 Resp: 50840




Ion Ratio Lower Upper


84 100

49 110.8 90.5 135.7

51 34.5 28.5 42.7


86 66.6 48.1 72.1





#21
trans-1,2-Dichloroethene
Concen: 18.589 ug/l
RT: 5.806 min Scan# 611
Delta R.T. -0.000 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBSD02



Tgt Ion: 96 Resp: 46018
Ion Ratio Lower Upper
96 100
61 136.9 109.2 163.8
98 65.3 49.8 74.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#22
Diisopropyl ether
Concen: 19.643 ug/l
RT: 6.688 min Scan# 805
Delta R.T. -0.000 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01


Tgt Ion: 45 Resp: 152785
Ion Ratio Lower Upper
45 100
43 53.3 41.9 62.9
87 30.6 22.4 33.6
59 12.1 9.8 14.6





#23
Vinyl Acetate
Concen: 99.908 ug/l
RT: 6.618 min Scan# 793
Delta R.T. -0.000 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBSD02



Tgt Ion: 43 Resp: 656574
Ion Ratio Lower Upper
43 100
86 12.5 9.5 14.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#24
1,1-Dichloroethane
Concen: 19.287 ug/l
RT: 6.588 min Scan# 788
Delta R.T. 0.006 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01


Tgt Ion: 63 Resp: 86331
Ion Ratio Lower Upper
63 100
98 7.6 3.6 10.8
100 4.8 2.6 7.8





#25
2-Butanone
Concen: 94.923 ug/l
RT: 7.494 min Scan# 942
Delta R.T. -0.000 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBSD02



Tgt Ion: 43 Resp: 21899
Ion Ratio Lower Upper
43 100
72 29.1 23.0 34.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#26
2,2-Dichloropropane
Concen: 21.343 ug/l
RT: 7.506 min Scan# 944
Delta R.T. 0.006 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01

Tgt Ion: 77 Resp: 74316
Ion Ratio Lower Upper
77 100
97 24.0 11.6 34.7





#27

cis-1,2-Dichloroethene

Concen: 19.573 ug/l

RT: 7.500 min Scan# 943

Delta R.T. -0.000 min

Lab File: VN086945.D

Acq: 11 Jun 2025 14:01

Instrument :

MSVOA_N

ClientSampleId :

VN0611WBSD02





Tgt Ion:	96	Resp:	5794
Ion	Ratio	Lower	Upper
96	100		
61	135.2	0.0	278.0
98	65.0	0.0	128.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025



#28

Bromochloromethane

Concen: 21.088 ug/l




RT: 7.824 min Scan# 998


Delta R.T. -0.000 min

Lab File: VN086945.D

Acq: 11 Jun 2025 14:01

Tgt Ion:	49	Resp:	46411
Ion	Ratio	Lower	Upper
49	100		
129	1.7	0.0	3.8
130	80.4	62.2	93.2





#29

Tetrahydrofuran

Concen: 96.721 ug/l

RT: 7.853 min Scan# 1003

Delta R.T. -0.000 min

Lab File: VN086945.D

Acq: 11 Jun 2025 14:01

Instrument :

MSVOA_N

ClientSampleId :

VN0611WBSD02

Tgt Ion: 42 Resp: 145360

Ion Ratio Lower Upper

42 100

72 48.9 37.8 56.6





71 46.0 35.6 53.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025



#30

Chloroform

Concen: 19.374 ug/l

RT: 7.977 min Scan# 1024

Delta R.T. -0.000 min

Lab File: VN086945.D




Acq: 11 Jun 2025 14:01


Tgt Ion: 83 Resp: 86604

Ion Ratio Lower Upper

83 100

85 65.8 52.9 79.3





#31

Cyclohexane

Concen: 17.164 ug/l

RT: 8.265 min Scan# 1073

Delta R.T. -0.006 min

Lab File: VN086945.D

Acq: 11 Jun 2025 14:01

Instrument :

MSVOA_N

ClientSampleId :

VN0611WBSD02

Tgt Ion: 56 Resp: 74510

Ion Ratio Lower Upper

56 100

69 32.4 25.4 38.0





84 88.3 68.8 103.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025



#32

1,1,1-Trichloroethane

Concen: 18.790 ug/l

RT: 8.177 min Scan# 1058

Delta R.T. -0.000 min

Lab File: VN086945.D

Acq: 11 Jun 2025 14:01




Tgt Ion: 97 Resp: 71441


Ion Ratio Lower Upper

97 100

99 63.6 52.0 78.0

61 42.8 35.9 53.9





#33

1,2-Dichloroethane-d4

Concen: 47.886 ug/l

RT: 8.588 min Scan# 1128

Delta R.T. -0.000 min

Lab File: VN086945.D

Acq: 11 Jun 2025 14:01

Instrument :

MSVOA_N

ClientSampleId :

VN0611WBSD02

Tgt Ion: 65 Resp: 12816

Ion Ratio Lower Upper

65 100





67 54.8 0.0 105.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.106 min Scan# 1216

Delta R.T. -0.000 min

Lab File: VN086945.D

Acq: 11 Jun 2025 14:01




Tgt Ion:114 Resp: 354069


Ion Ratio Lower Upper

114 100

63 19.8 0.0 39.6

88 16.1 0.0 30.2





#35

Dibromofluoromethane

Concen: 52.260 ug/l

RT: 8.177 min Scan# 1057

Delta R.T. 0.006 min

Lab File: VN086945.D

Acq: 11 Jun 2025 14:01

Instrument :

MSVOA_N

ClientSampleId :

VN0611WBSD02

Tgt Ion:113 Resp: 10965

Ion Ratio Lower Upper

113 100

111 101.1 84.2 126.2





192 17.6 14.2 21.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025



#36

1,1-Dichloropropene

Concen: 19.481 ug/l

RT: 8.382 min Scan# 1093

Delta R.T. 0.006 min

Lab File: VN086945.D

Acq: 11 Jun 2025 14:01




Tgt Ion: 75 Resp: 60912


Ion Ratio Lower Upper

75 100

110 36.1 17.6 52.9

77 31.6 25.3 37.9





#37

Ethyl Acetate

Concen: 20.102 ug/l

RT: 7.571 min Scan# 91

Delta R.T. -0.006 min

Lab File: VN086945.D

Acq: 11 Jun 2025 14:01

Instrument :

MSVOA_N

ClientSampleId :

VN0611WBSD02

Tgt Ion: 43 Resp: 80010

Ion Ratio Lower Upper

43 100

61 15.9 12.1 18.1





70 12.4 9.8 14.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025



#38

Carbon Tetrachloride

Concen: 19.494 ug/l

RT: 8.371 min Scan# 1091

Delta R.T. -0.000 min

Lab File: VN086945.D

Acq: 11 Jun 2025 14:01




Tgt Ion:117 Resp: 59839


Ion Ratio Lower Upper

117 100

119 98.4 75.0 112.6

121 32.4 24.9 37.3





#39

Methylcyclohexane

Concen: 16.588 ug/l

RT: 9.606 min Scan# 1130

Delta R.T. -0.000 min

Lab File: VN086945.D

Acq: 11 Jun 2025 14:01

Instrument :

MSVOA_N

ClientSampleId :

VN0611WBSD02

Tgt Ion: 83 Resp: 71058

Ion Ratio Lower Upper

83 100

55 72.5 61.0 91.6





98 47.5 38.9 58.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025



#40

Benzene

Concen: 19.653 ug/l

RT: 8.612 min Scan# 1132

Delta R.T. -0.000 min

Lab File: VN086945.D




Acq: 11 Jun 2025 14:01


Tgt Ion: 78 Resp: 200940

Ion Ratio Lower Upper

78 100


77 22.6 18.7 28.1





#41
Methacrylonitrile
Concen: 20.600 ug/l
RT: 7.788 min Scan# 992
Delta R.T. -0.000 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBSD02



Tgt Ion: 41 Resp: 4604
Ion Ratio Lower Upper
41 100
39 50.5 41.2 61.8
67 78.7 61.4 92.2
52 28.7 22.7 34.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#42
1,2-Dichloroethane
Concen: 20.336 ug/l
RT: 8.677 min Scan# 1143
Delta R.T. -0.000 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01


Tgt Ion: 62 Resp: 63066
Ion Ratio Lower Upper
62 100
98 10.5 0.0 21.2





#43
Isopropyl Acetate
Concen: 19.930 ug/l
RT: 8.694 min Scan# 1146
Delta R.T. -0.000 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBSD02



Tgt Ion: 43 Resp: 127320
Ion Ratio Lower Upper
43 100
61 28.5 22.1 33.1
87 15.1 11.8 17.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#44
Trichloroethene
Concen: 20.141 ug/l
RT: 9.359 min Scan# 1259
Delta R.T. -0.000 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01

Tgt Ion:130 Resp: 48830
Ion Ratio Lower Upper
130 100
95 94.9 0.0 196.2





#45

1,2-Dichloropropane

Concen: 20.027 ug/l

RT: 9.624 min Scan# 1304

Delta R.T. -0.006 min

Lab File: VN086945.D

Acq: 11 Jun 2025 14:01

Instrument :

MSVOA_N

ClientSampleId :

VN0611WBSD02

Tgt Ion: 63 Resp: 4981

Ion Ratio Lower Upper

63 100





65 29.1 25.5 38.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025



#46

Dibromomethane

Concen: 21.290 ug/l

RT: 9.712 min Scan# 1319

Delta R.T. -0.006 min

Lab File: VN086945.D

Acq: 11 Jun 2025 14:01




Tgt Ion: 93 Resp: 35176


Ion Ratio Lower Upper

93 100

95 83.2 67.1 100.7

174 92.2 74.2 111.4





#47

Bromodichloromethane

Concen: 20.658 ug/l

RT: 9.894 min Scan# 1350

Delta R.T. -0.000 min

Lab File: VN086945.D

Acq: 11 Jun 2025 14:01

Instrument :

MSVOA_N

ClientSampleId :

VN0611WBSD02





Tgt Ion:	83	Resp:	7022
Ion	Ratio	Lower	Upper
83	100		
85	62.7	50.9	76.3
127	8.5	6.6	10.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025



#48

Methyl methacrylate

Concen: 19.079 ug/l




RT: 9.688 min Scan# 1315


Delta R.T. 0.006 min

Lab File: VN086945.D

Acq: 11 Jun 2025 14:01

Tgt Ion:	41	Resp:	56099
Ion	Ratio	Lower	Upper
41	100		
69	99.4	74.6	112.0
39	51.4	40.2	60.4





#49

1,4-Dioxane

Concen: 454.282 ug/l

RT: 9.700 min Scan# 1317

Delta R.T. -0.000 min

Lab File: VN086945.D

Acq: 11 Jun 2025 14:01

Instrument :

MSVOA_N

ClientSampleId :

VN0611WBSD02

Tgt Ion: 88 Resp: 24300

Ion Ratio Lower Upper

88 100

43 22.4 0.0 0.0





58 66.1 57.7 86.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025



#50

Toluene-d8

Concen: 49.506 ug/l

RT: 10.571 min Scan# 1465

Delta R.T. -0.000 min

Lab File: VN086945.D




Acq: 11 Jun 2025 14:01


Tgt Ion: 98 Resp: 411227

Ion Ratio Lower Upper

98 100


100 66.6 53.4 80.0





#51
4-Methyl-2-Pentanone
Concen: 103.940 ug/l
RT: 10.447 min Scan# 1444
Delta R.T. -0.000 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBSD02




Tgt Ion: 43 Resp: 399735
Ion Ratio Lower Upper
43 100
58 42.4 34.2 51.4

Manual Integrations
APPROVED



Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025




#52
Toluene
Concen: 19.893 ug/l
RT: 10.629 min Scan# 1475
Delta R.T. -0.006 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01




Tgt Ion: 92 Resp: 124298
Ion Ratio Lower Upper
92 100
91 170.3 136.5 204.7





#53
t-1,3-Dichloropropene
Concen: 20.744 ug/l
RT: 10.841 min Scan# 1511
Delta R.T. -0.000 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBSD02



Tgt Ion: 75 Resp: 78850
Ion Ratio Lower Upper
75 100
77 32.9 26.4 39.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#54
cis-1,3-Dichloropropene
Concen: 20.709 ug/l
RT: 10.318 min Scan# 1422
Delta R.T. -0.000 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01


Tgt Ion: 75 Resp: 84225
Ion Ratio Lower Upper
75 100
77 30.9 26.2 39.4
39 39.0 32.3 48.5





#55
1,1,2-Trichloroethane
Concen: 20.924 ug/l
RT: 11.018 min Scan# 1541
Delta R.T. -0.000 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBSD02



Tgt Ion: 97 Resp: 5030
Ion Ratio Lower Upper
97 100
83 88.4 69.7 104.5
85 52.7 45.1 67.7
99 64.8 50.6 75.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#56
Ethyl methacrylate
Concen: 20.747 ug/l
RT: 10.882 min Scan# 1518
Delta R.T. -0.000 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01


Tgt Ion: 69 Resp: 79449
Ion Ratio Lower Upper
69 100
41 58.8 47.2 70.8
39 25.0 23.0 34.6





#57
1,3-Dichloropropane
Concen: 20.434 ug/l
RT: 11.165 min Scan# 1566
Delta R.T. -0.000 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBSD02



Tgt Ion: 76 Resp: 8522
Ion Ratio Lower Upper
76 100
78 32.6 25.3 37.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#58
2-Chloroethyl Vinyl ether
Concen: 105.889 ug/l
RT: 10.165 min Scan# 1396
Delta R.T. -0.000 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01


Tgt Ion: 63 Resp: 241856
Ion Ratio Lower Upper
63 100
106 29.1 23.0 34.6





#59
2-Hexanone
Concen: 93.840 ug/l
RT: 11.206 min Scan# 1572
Delta R.T. 0.006 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBSD02



Tgt Ion: 43 Resp: 23246
Ion Ratio Lower Upper
43 100
58 64.8 30.4 91.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#60
Dibromochloromethane
Concen: 21.220 ug/l
RT: 11.359 min Scan# 1599
Delta R.T. -0.000 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01


Tgt Ion:129 Resp: 53154
Ion Ratio Lower Upper
129 100
127 77.2 38.6 115.8





#61
1,2-Dibromoethane
Concen: 20.718 ug/l
RT: 11.470 min Scan# 1618
Delta R.T. -0.000 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBSD02




Tgt Ion:107 Resp: 51050
Ion Ratio Lower Upper
107 100
109 94.8 76.3 114.5

Manual Integrations
APPROVED



Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025




#62
4-Bromofluorobenzene
Concen: 50.121 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. -0.000 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01



Tgt Ion: 95 Resp: 154682
Ion Ratio Lower Upper
95 100
174 72.6 0.0 141.8
176 68.6 0.0 132.6





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.865 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086945.D

Acq: 11 Jun 2025 14:01

Instrument :

MSVOA_N

ClientSampleId :

VN0611WBSD02

Tgt Ion:117 Resp: 310598

Ion Ratio Lower Upper

117 100

82 54.9 44.6 67.0





119 32.6 25.5 38.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025



#64

Tetrachloroethene

Concen: 18.499 ug/l

RT: 11.106 min Scan# 1556

Delta R.T. -0.000 min

Lab File: VN086945.D

Acq: 11 Jun 2025 14:01

Tgt Ion:164 Resp: 36363




Ion Ratio Lower Upper


164 100

166 125.7 101.8 152.8

129 101.3 83.0 124.4


131 93.9 78.2 117.2





#65
Chlorobenzene
Concen: 20.178 ug/l
RT: 11.894 min Scan# 1690
Delta R.T. -0.000 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBSD02



Tgt Ion:112 Resp: 138224
Ion Ratio Lower Upper
112 100
114 33.4 26.2 39.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#66
1,1,1,2-Tetrachloroethane
Concen: 20.442 ug/l
RT: 11.965 min Scan# 1702
Delta R.T. -0.000 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01


Tgt Ion:131 Resp: 45014
Ion Ratio Lower Upper
131 100
133 96.4 48.3 144.8
119 70.3 34.0 101.8





#67
Ethyl Benzene
Concen: 19.250 ug/l
RT: 11.965 min Scan# 1702
Delta R.T. -0.000 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBSD02




Tgt Ion: 91 Resp: 227138
Ion Ratio Lower Upper
91 100
106 31.9 25.4 38.2

Manual Integrations
APPROVED



Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025




#68
m/p-Xylenes
Concen: 39.501 ug/l
RT: 12.070 min Scan# 1720
Delta R.T. -0.000 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01




Tgt Ion:106 Resp: 178421
Ion Ratio Lower Upper
106 100
91 196.9 159.4 239.0





#69
o-Xylene
Concen: 20.004 ug/l
RT: 12.400 min Scan# 1776
Delta R.T. -0.000 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBSD02



Tgt Ion:106 Resp: 8653
Ion Ratio Lower Upper
106 100
91 212.7 106.1 318.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#70
Styrene
Concen: 20.565 ug/l
RT: 12.412 min Scan# 1778
Delta R.T. -0.000 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01


Tgt Ion:104 Resp: 152236
Ion Ratio Lower Upper
104 100
78 46.8 38.5 57.7
103 53.0 43.1 64.7





#71
Bromoform
Concen: 22.388 ug/l
RT: 12.582 min Scan# 11
Delta R.T. -0.000 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBSD02




Tgt Ion:173 Resp: 3652
Ion Ratio Lower Upper
173 100
175 46.8 24.1 72.4
254 0.0 0.0 0.0

Manual Integrations
APPROVED



Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.788 min Scan# 2012
Delta R.T. -0.006 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01




Tgt Ion:152 Resp: 151691
Ion Ratio Lower Upper
152 100
115 61.0 30.1 90.5
150 166.9 0.0 345.0





#73
Isopropylbenzene
Concen: 19.276 ug/l
RT: 12.694 min Scan# 1826
Delta R.T. -0.000 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBSD02




Tgt Ion:105 Resp: 213019
Ion Ratio Lower Upper
105 100
120 26.8 13.5 40.4

Manual Integrations
APPROVED



Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025




#74
N-ethyl acetate
Concen: 18.051 ug/l
RT: 12.529 min Scan# 1798
Delta R.T. 0.017 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01




Tgt Ion: 43 Resp: 69705
Ion Ratio Lower Upper
43 100
70 38.7 37.6 56.4
55 18.5 19.0 28.6#
61 19.5 20.8 31.2#





#75
1,1,2,2-Tetrachloroethane
Concen: 21.740 ug/l
RT: 12.935 min Scan# 1867
Delta R.T. -0.006 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBSD02



Tgt Ion: 83 Resp: 81389
Ion Ratio Lower Upper
83 100
131 9.7 4.9 14.7
85 63.6 32.0 96.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#76
1,2,3-Trichloropropane
Concen: 19.732 ug/l m
RT: 12.994 min Scan# 1877
Delta R.T. -0.000 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01

Tgt Ion: 75 Resp: 71147
Ion Ratio Lower Upper
75 100
77 175.2 79.3 237.9









#77
Bromobenzene
Concen: 21.319 ug/l
RT: 12.976 min Scan# 1874
Delta R.T. -0.006 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01

Instrument :
MSVOA_N
ClientSampleId :
VN0611WBSD02

Tgt Ion:156 Resp: 54030
Ion Ratio Lower Upper
156 100
77 230.7 117.7 353.1
158 95.1 48.1 144.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#78
n-propylbenzene
Concen: 19.168 ug/l
RT: 13.035 min Scan# 1884
Delta R.T. -0.000 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01

Tgt Ion: 91 Resp: 257426
Ion Ratio Lower Upper
91 100
120 23.2 11.4 34.2









#79
2-Chlorotoluene
Concen: 19.571 ug/l
RT: 13.123 min Scan# 1899
Delta R.T. -0.000 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01

Instrument :
MSVOA_N
ClientSampleId :
VN0611WBSD02

Tgt Ion: 91 Resp: 15762
Ion Ratio Lower Upper
91 100
126 33.1 16.2 48.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#80
1,3,5-Trimethylbenzene
Concen: 19.228 ug/l
RT: 13.170 min Scan# 1907
Delta R.T. -0.000 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01


Tgt Ion:105 Resp: 175428
Ion Ratio Lower Upper
105 100
120 50.7 24.9 74.6





#81
trans-1,4-Dichloro-2-butene
Concen: 18.963 ug/l
RT: 12.735 min Scan# 1833
Delta R.T. -0.000 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBSD02



Tgt Ion: 75 Resp: 2970
Ion Ratio Lower Upper
75 100
53 89.6 66.6 99.8
89 47.5 36.7 55.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#82
4-Chlorotoluene
Concen: 19.931 ug/l
RT: 13.223 min Scan# 1916
Delta R.T. -0.000 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01


Tgt Ion: 91 Resp: 162421
Ion Ratio Lower Upper
91 100
126 33.2 16.7 50.1





#83
tert-Butylbenzene
Concen: 18.749 ug/l
RT: 13.435 min Scan# 1952
Delta R.T. -0.000 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBSD02



Tgt Ion:119 Resp: 156588
Ion Ratio Lower Upper
119 100
91 63.5 32.1 96.5
134 24.9 13.4 40.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#84
1,2,4-Trimethylbenzene
Concen: 19.373 ug/l
RT: 13.482 min Scan# 1960
Delta R.T. -0.000 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01

Tgt Ion:105 Resp: 177241
Ion Ratio Lower Upper
105 100
120 46.8 23.2 69.6









#85
 sec-Butylbenzene
 Concen: 18.109 ug/l
 RT: 13.611 min Scan# 1983
 Delta R.T. -0.006 min
 Lab File: VN086945.D
 Acq: 11 Jun 2025 14:01

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0611WBSD02

Tgt Ion:105 Resp: 219639
 Ion Ratio Lower Upper
 105 100
 134 20.3 10.0 30.0




Manual Integrations
 APPROVED


Reviewed By :John Carlone 06/12/2025
 Supervised By :Semsettin Yesilyurt 06/12/2025



#86
 p-Isopropyltoluene
 Concen: 18.238 ug/l
 RT: 13.729 min Scan# 2002
 Delta R.T. -0.000 min
 Lab File: VN086945.D
 Acq: 11 Jun 2025 14:01


Tgt Ion:119 Resp: 182853
 Ion Ratio Lower Upper
 119 100
 134 26.9 13.5 40.4
 91 22.2 11.5 34.4





#87
1,3-Dichlorobenzene
Concen: 20.517 ug/l
RT: 13.729 min Scan# 2002
Delta R.T. -0.006 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBSD02



Tgt Ion:146 Resp: 102302
Ion Ratio Lower Upper
146 100
111 41.7 20.9 62.8
148 63.8 32.0 96.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#88
1,4-Dichlorobenzene
Concen: 20.382 ug/l
RT: 13.811 min Scan# 2016
Delta R.T. -0.000 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01

Tgt Ion:146 Resp: 103614
Ion Ratio Lower Upper
146 100
111 42.1 19.9 59.7
148 63.3 31.8 95.4





#89

n-Butylbenzene

Concen: 17.668 ug/l

RT: 14.053 min Scan# 2057

Delta R.T. -0.000 min

Lab File: VN086945.D

Acq: 11 Jun 2025 14:01

Instrument :

MSVOA_N

ClientSampleId :

VN0611WBSD02

Tgt Ion: 91 Resp: 171570

Ion Ratio Lower Upper

91 100

92 54.4 27.5 82.4





134 25.4 13.0 39.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025



#90

Hexachloroethane

Concen: 18.810 ug/l

RT: 14.329 min Scan# 2104

Delta R.T. -0.006 min

Lab File: VN086945.D




Acq: 11 Jun 2025 14:01


Tgt Ion:117 Resp: 31966

Ion Ratio Lower Upper

117 100


201 66.7 33.2 99.6





#91
1,2-Dichlorobenzene
Concen: 20.638 ug/l
RT: 14.106 min Scan# 2066
Delta R.T. -0.000 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBSD02



Tgt Ion:146 Resp: 98864
Ion Ratio Lower Upper
146 100
111 44.0 21.6 64.6
148 62.9 31.7 95.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#92
1,2-Dibromo-3-Chloropropane
Concen: 20.786 ug/l
RT: 14.723 min Scan# 2171
Delta R.T. 0.006 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01

Tgt Ion: 75 Resp: 18610
Ion Ratio Lower Upper
75 100
155 81.3 41.1 123.4
157 105.2 52.7 158.1





#93

1,2,4-Trichlorobenzene

Concen: 18.660 ug/l

RT: 15.388 min Scan# 21

Delta R.T. -0.000 min

Lab File: VN086945.D

Acq: 11 Jun 2025 14:01

Instrument :

MSVOA_N

ClientSampleId :

VN0611WBSD02

Tgt Ion:180 Resp: 5708

Ion Ratio Lower Upper

180 100

182 94.8 46.9 140.7





145 32.7 16.0 47.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025



#94

Hexachlorobutadiene

Concen: 15.506 ug/l

RT: 15.500 min Scan# 2303

Delta R.T. -0.000 min

Lab File: VN086945.D

Acq: 11 Jun 2025 14:01




Tgt Ion:225 Resp: 17672


Ion Ratio Lower Upper

225 100

223 68.8 33.5 100.4


227 61.4 31.1 93.5





#95
Naphthalene
Concen: 20.281 ug/l
RT: 15.635 min Scan# 2327
Delta R.T. -0.006 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBSD02



Tgt Ion:128 Resp: 23092
Ion Ratio Lower Upper
128 100
127 12.8 10.2 15.2
129 10.6 8.8 13.2




Manual Integrations
APPROVED

Reviewed By :John Carlone 06/12/2025
Supervised By :Semsettin Yesilyurt 06/12/2025



#96
1,2,3-Trichlorobenzene
Concen: 18.105 ug/l
RT: 15.835 min Scan# 2360
Delta R.T. -0.000 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01

Tgt Ion:180 Resp: 55026
Ion Ratio Lower Upper
180 100
182 95.9 47.0 141.0
145 34.9 16.6 49.8



Manual Integration Report

Sequence:	vn060625	Instrument	MSVOA_n

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
VSTDICC001	VN086862.D	1,1,2-Trichlorotrifluoroethane	JOHN	6/9/2025 8:02:09 AM	MMDadoda	6/9/2025 1:13:18 PM	Peak Integrated by Software
VSTDICC001	VN086862.D	1,2,3-Trichloropropane	JOHN	6/9/2025 8:02:09 AM	MMDadoda	6/9/2025 1:13:18 PM	Peak Integrated by Software
VSTDICC001	VN086862.D	1,4-Dichlorobenzene	JOHN	6/9/2025 8:02:09 AM	MMDadoda	6/9/2025 1:13:18 PM	Peak Integrated by Software
VSTDICC001	VN086862.D	2-Hexanone	JOHN	6/9/2025 8:02:09 AM	MMDadoda	6/9/2025 1:13:18 PM	Peak Integrated by Software
VSTDICC001	VN086862.D	N-amyl acetate	JOHN	6/9/2025 8:02:09 AM	MMDadoda	6/9/2025 1:13:18 PM	Peak Integrated by Software
VSTDICC005	VN086863.D	1,2,3-Trichloropropane	JOHN	6/9/2025 8:02:13 AM	MMDadoda	6/9/2025 1:13:19 PM	Peak Integrated by Software
VSTDICC005	VN086863.D	N-amyl acetate	JOHN	6/9/2025 8:02:13 AM	MMDadoda	6/9/2025 1:13:19 PM	Peak Integrated by Software
VSTDICC020	VN086864.D	1,2,3-Trichloropropane	JOHN	6/9/2025 8:02:19 AM	MMDadoda	6/9/2025 1:13:21 PM	Peak Integrated by Software
VSTDICCC050	VN086865.D	1,2,3-Trichloropropane	JOHN	6/9/2025 8:02:25 AM	MMDadoda	6/9/2025 1:13:23 PM	Peak Integrated by Software
VSTDICC100	VN086866.D	1,2,3-Trichloropropane	JOHN	6/9/2025 8:02:29 AM	MMDadoda	6/9/2025 1:13:28 PM	Peak Integrated by Software
VSTDICC150	VN086867.D	1,2,3-Trichloropropane	JOHN	6/9/2025 8:02:34 AM	MMDadoda	6/9/2025 1:13:30 PM	Peak Integrated by Software
VSTDICV050	VN086869.D	1,2,3-Trichloropropane	JOHN	6/9/2025 8:02:38 AM	MMDadoda	6/9/2025 1:13:34 PM	Peak Integrated by Software
VSTDCCC050	VN086886.D	1,2,3-Trichloropropane	JOHN	6/9/2025 8:02:55 AM	MMDadoda	6/9/2025 1:13:44 PM	Peak Integrated by Software



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:

vn060625

Instrument

MSVOA_n

Sample ID

File ID

Parameter

Review By

Review On

Supervised
By

Supervised On

Reason

Manual Integration Report

Sequence:

vn061125

Instrument

MSVOA_n

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
VSTDCCC050	VN086940.D	1,2,3-Trichloropropane	JOHN	6/12/2025 9:47:09 AM	Sam	6/12/2025 3:54:50 PM	Peak Integrated by Software
VN0611WBS02	VN086944.D	1,2,3-Trichloropropane	JOHN	6/12/2025 9:47:21 AM	Sam	6/12/2025 3:54:50 PM	Peak Integrated by Software
VN0611WBSD02	VN086945.D	1,2,3-Trichloropropane	JOHN	6/12/2025 9:47:26 AM	Sam	6/12/2025 3:54:52 PM	Peak Integrated by Software
VSTDCCC050	VN086965.D	1,2,3-Trichloropropane	JOHN	6/12/2025 9:53:57 AM	Sam	6/12/2025 3:54:56 PM	Peak Integrated by Software

Instrument ID: **MSVOA_N**

Daily Analysis Runlog For Sequence/QC Batch ID # VN060625

Review By	John Carlone	Review On	6/9/2025 8:08:23 AM
Supervise By	Maresh Dadoda	Supervise On	6/9/2025 1:13:51 PM
SubDirectory	VN060625	HP Acquire Method	HP Processing Method 82N060625W.M
STD. NAME	STD REF.#		
Tune/Reschk	VP134155		
Initial Calibration Stds	VP134242,VP134243,VP134244,VP134245,VP134246,VP134247		
CCC	VP134156		
Internal Standard/PEM			
ICV/I.BLK	VP134248		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VN086861.D	06 Jun 2025 07:59	JC\MD	Ok
2	VSTDIC001	VN086862.D	06 Jun 2025 12:44	JC\MD	Ok,M
3	VSTDIC005	VN086863.D	06 Jun 2025 13:17	JC\MD	Ok,M
4	VSTDIC020	VN086864.D	06 Jun 2025 13:40	JC\MD	Ok,M
5	VSTDIC050	VN086865.D	06 Jun 2025 14:03	JC\MD	Ok,M
6	VSTDIC100	VN086866.D	06 Jun 2025 14:26	JC\MD	Ok,M
7	VSTDIC150	VN086867.D	06 Jun 2025 14:49	JC\MD	Ok,M
8	IBLK	VN086868.D	06 Jun 2025 15:12	JC\MD	Ok
9	VSTDICV050	VN086869.D	06 Jun 2025 15:54	JC\MD	Ok,M
10	VN0606WBL01	VN086870.D	06 Jun 2025 16:47	JC\MD	Ok
11	VN0606WBL02	VN086871.D	06 Jun 2025 17:10	JC\MD	Ok
12	VN0606WBS01	VN086872.D	06 Jun 2025 17:33	JC\MD	Ok,M
13	VN0606WBSD01	VN086873.D	06 Jun 2025 17:56	JC\MD	Ok,M
14	Q2254-01	VN086874.D	06 Jun 2025 18:19	JC\MD	Not Ok
15	Q2237-02	VN086875.D	06 Jun 2025 18:42	JC\MD	Ok
16	Q2216-02	VN086876.D	06 Jun 2025 19:05	JC\MD	Ok
17	Q2216-03	VN086877.D	06 Jun 2025 19:28	JC\MD	Ok
18	Q2216-04	VN086878.D	06 Jun 2025 19:51	JC\MD	Ok
19	Q2216-05	VN086879.D	06 Jun 2025 20:13	JC\MD	Not Ok
20	Q2216-06	VN086880.D	06 Jun 2025 20:36	JC\MD	Not Ok
21	Q2206-04	VN086881.D	06 Jun 2025 20:59	JC\MD	Not Ok

Instrument ID: **MSVOA_N**

Daily Analysis Runlog For Sequence/QCBatch ID # VN060625

Review By	John Carlone	Review On	6/9/2025 8:08:23 AM
Supervise By	Maresh Dadoda	Supervise On	6/9/2025 1:13:51 PM
SubDirectory	VN060625	HP Acquire Method	HP Processing Method 82N060625W.M
STD. NAME	STD REF.#		
Tune/Reschk	VP134155		
Initial Calibration Stds	VP134242,VP134243,VP134244,VP134245,VP134246,VP134247		
CCC	VP134156		
Internal Standard/PEM			
ICV/I.BLK	VP134248		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	Q2242-04	VN086882.D	06 Jun 2025 21:21	JC\MD	Not Ok
23	Q2192-01	VN086883.D	06 Jun 2025 21:44	JC\MD	Not Ok
24	Q2198-02	VN086884.D	06 Jun 2025 22:07	JC\MD	Not Ok
25	Q2198-04	VN086885.D	06 Jun 2025 22:29	JC\MD	Not Ok
26	VSTDCCC050	VN086886.D	06 Jun 2025 22:52	JC\MD	Not Ok

M : Manual Integration

Instrument ID: **MSVOA_N**

Daily Analysis Runlog For Sequence/QCBatch ID # VN061125

Review By	John Carlone	Review On	6/12/2025 9:57:37 AM
Supervise By	Semsettin Yesilyurt	Supervise On	6/12/2025 3:55:33 PM
SubDirectory	VN061125	HP Acquire Method	HP Processing Method 82N060625W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP134284		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134285,VP134286		

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VN086939.D	11 Jun 2025 10:22	JC\MD	Ok
2	VSTDCCC050	VN086940.D	11 Jun 2025 11:32	JC\MD	Ok,M
3	VN0611MBL01	VN086941.D	11 Jun 2025 12:06	JC\MD	Ok
4	VN0611WBL01	VN086942.D	11 Jun 2025 12:28	JC\MD	Ok
5	VN0611WBS01	VN086943.D	11 Jun 2025 12:50	JC\MD	Not Ok
6	VN0611WBS02	VN086944.D	11 Jun 2025 13:27	JC\MD	Ok,M
7	VN0611WBSD02	VN086945.D	11 Jun 2025 14:01	JC\MD	Ok,M
8	Q2251-06	VN086946.D	11 Jun 2025 14:22	JC\MD	Ok
9	Q2202-03DL	VN086947.D	11 Jun 2025 14:44	JC\MD	Ok,M
10	Q2250-01DL	VN086948.D	11 Jun 2025 15:06	JC\MD	Ok
11	Q2202-06	VN086949.D	11 Jun 2025 15:28	JC\MD	Ok
12	Q2189-01	VN086950.D	11 Jun 2025 15:50	JC\MD	Ok
13	Q2202-05	VN086951.D	11 Jun 2025 16:12	JC\MD	Ok
14	Q2202-04	VN086952.D	11 Jun 2025 16:35	JC\MD	Ok
15	Q2231-01	VN086953.D	11 Jun 2025 16:57	JC\MD	Ok
16	Q2231-02	VN086954.D	11 Jun 2025 17:19	JC\MD	Ok,M
17	Q2231-03	VN086955.D	11 Jun 2025 17:41	JC\MD	Ok
18	Q2231-04	VN086956.D	11 Jun 2025 18:04	JC\MD	Ok
19	Q2231-05	VN086957.D	11 Jun 2025 18:26	JC\MD	Ok
20	Q2231-06	VN086958.D	11 Jun 2025 18:48	JC\MD	Ok
21	Q2210-01	VN086959.D	11 Jun 2025 19:10	JC\MD	Ok

Instrument ID: **MSVOA_N**

Daily Analysis Runlog For Sequence/QC Batch ID # VN061125

Review By	John Carlone	Review On	6/12/2025 9:57:37 AM
Supervise By	Semsettin Yesilyurt	Supervise On	6/12/2025 3:55:33 PM
SubDirectory	VN061125	HP Acquire Method	HP Processing Method 82N060625W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP134284		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134285,VP134286		

22	Q2209-01	VN086960.D	11 Jun 2025 19:33	JC\MD	Ok
23	IBLK	VN086961.D	11 Jun 2025 19:55	JC\MD	Ok
24	IBLK	VN086962.D	11 Jun 2025 20:17	JC\MD	Ok
25	Q2263-01	VN086963.D	11 Jun 2025 20:39	JC\MD	Ok
26	Q2263-02	VN086964.D	11 Jun 2025 21:01	JC\MD	Ok
27	VSTDCCC050	VN086965.D	11 Jun 2025 21:23	JC\MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN060625

Review By	John Carlone	Review On	6/9/2025 8:08:23 AM
Supervise By	Mahesh Dadoda	Supervise On	6/9/2025 1:13:51 PM
SubDirectory	VN060625	HP Acquire Method	HP Processing Method 82N060625W.M

STD. NAME	STD REF.#
Tune/Reschk	VP134155
Initial Calibration Stds	VP134242,VP134243,VP134244,VP134245,VP134246,VP134247
CCC	VP134156
Internal Standard/PEM	
ICV/I.BLK	VP134248
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN086861.D	06 Jun 2025 07:59		JC\MD	Ok
2	VSTDIC001	VSTDIC001	VN086862.D	06 Jun 2025 12:44	Method failed for com.#13	JC\MD	Ok,M
3	VSTDIC005	VSTDIC005	VN086863.D	06 Jun 2025 13:17		JC\MD	Ok,M
4	VSTDIC020	VSTDIC020	VN086864.D	06 Jun 2025 13:40		JC\MD	Ok,M
5	VSTDIC050	VSTDIC050	VN086865.D	06 Jun 2025 14:03		JC\MD	Ok,M
6	VSTDIC100	VSTDIC100	VN086866.D	06 Jun 2025 14:26		JC\MD	Ok,M
7	VSTDIC150	VSTDIC150	VN086867.D	06 Jun 2025 14:49		JC\MD	Ok,M
8	IBLK	IBLK	VN086868.D	06 Jun 2025 15:12		JC\MD	Ok
9	VSTDICV050	ICVVN060625	VN086869.D	06 Jun 2025 15:54		JC\MD	Ok,M
10	VN0606WBL01	VN0606WBL01	VN086870.D	06 Jun 2025 16:47		JC\MD	Ok
11	VN0606WBL02	VN0606WBL02	VN086871.D	06 Jun 2025 17:10		JC\MD	Ok
12	VN0606WBS01	VN0606WBS01	VN086872.D	06 Jun 2025 17:33		JC\MD	Ok,M
13	VN0606WBSD01	VN0606WBSD01	VN086873.D	06 Jun 2025 17:56		JC\MD	Ok,M
14	Q2254-01	BP-VPB-182-GW-810-8	VN086874.D	06 Jun 2025 18:19	vial A pH<2 endccc out of tune	JC\MD	Not Ok
15	Q2237-02	TW-WTS-10	VN086875.D	06 Jun 2025 18:42	vial A pH<2	JC\MD	Ok
16	Q2216-02	3887	VN086876.D	06 Jun 2025 19:05	vial A pH<2	JC\MD	Ok
17	Q2216-03	3888	VN086877.D	06 Jun 2025 19:28	vial A pH<2	JC\MD	Ok
18	Q2216-04	3864	VN086878.D	06 Jun 2025 19:51	vial A pH<2	JC\MD	Ok

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN060625

Review By	John Carlone	Review On	6/9/2025 8:08:23 AM
Supervise By	Maresh Dadoda	Supervise On	6/9/2025 1:13:51 PM
SubDirectory	VN060625	HP Acquire Method	HP Processing Method 82N060625W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP134155 VP134242,VP134243,VP134244,VP134245,VP134246,VP134247		
CCC Internal Standard/PEM	VP134156		
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134248		

19	Q2216-05	3865	VN086879.D	06 Jun 2025 20:13	vial A pH<2 Out of Tune	JC\MD	Not Ok
20	Q2216-06	3851	VN086880.D	06 Jun 2025 20:36	vial A pH<2 Out of Tune	JC\MD	Not Ok
21	Q2206-04	TP-1	VN086881.D	06 Jun 2025 20:59	vial A pH<2 Out of Tune	JC\MD	Not Ok
22	Q2242-04	TP09-MHJ	VN086882.D	06 Jun 2025 21:21	vial A pH<2 Out of Tune	JC\MD	Not Ok
23	Q2192-01	SB-1	VN086883.D	06 Jun 2025 21:44	vial A pH<2 Out of Tune	JC\MD	Not Ok
24	Q2198-02	B-202-SB02	VN086884.D	06 Jun 2025 22:07	vial A pH<2 Out of Tune	JC\MD	Not Ok
25	Q2198-04	B-207-SB02	VN086885.D	06 Jun 2025 22:29	vial A pH<2 Out of Tune	JC\MD	Not Ok
26	VSTDCCC050	VSTDCCC050EC	VN086886.D	06 Jun 2025 22:52	Out of Tune	JC\MD	Not Ok

M : Manual Integration

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN061125

Review By	John Carlone	Review On	6/12/2025 9:57:37 AM
Supervise By	Semsettin Yesilyurt	Supervise On	6/12/2025 3:55:33 PM
SubDirectory	VN061125	HP Acquire Method	HP Processing Method 82N060625W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP134284
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134285,VP134286

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN086939.D	11 Jun 2025 10:22		JC\MD	Ok
2	VSTDCCC050	VSTDCCC050	VN086940.D	11 Jun 2025 11:32	pH#Lot#V12668	JC\MD	Ok,M
3	VN0611MBL01	VN0611MBL01	VN086941.D	11 Jun 2025 12:06		JC\MD	Ok
4	VN0611WBL01	VN0611WBL01	VN086942.D	11 Jun 2025 12:28		JC\MD	Ok
5	VN0611WBS01	VN0611WBS01	VN086943.D	11 Jun 2025 12:50	Recovery fail	JC\MD	Not Ok
6	VN0611WBS02	VN0611WBS02	VN086944.D	11 Jun 2025 13:27		JC\MD	Ok,M
7	VN0611WBSD02	VN0611WBSD02	VN086945.D	11 Jun 2025 14:01		JC\MD	Ok,M
8	Q2251-06	VPB182-HYD-2025060	VN086946.D	11 Jun 2025 14:22	vial A pH<2	JC\MD	Ok
9	Q2202-03DL	MW-12-20250603DL	VN086947.D	11 Jun 2025 14:44	vial B pH<2	JC\MD	Ok,M
10	Q2250-01DL	MW-11A-13.5-060525D	VN086948.D	11 Jun 2025 15:06	vial B pH<2	JC\MD	Ok
11	Q2202-06	TB-20250603	VN086949.D	11 Jun 2025 15:28	vial A pH<2 TB	JC\MD	Ok
12	Q2189-01	MW-7-20250602	VN086950.D	11 Jun 2025 15:50	vial B pH<2	JC\MD	Ok
13	Q2202-05	MW-1A-20250603	VN086951.D	11 Jun 2025 16:12	vial A pH<2	JC\MD	Ok
14	Q2202-04	MW-130-20250603	VN086952.D	11 Jun 2025 16:35	vial A pH<2	JC\MD	Ok
15	Q2231-01	MW-10D-20250604	VN086953.D	11 Jun 2025 16:57	vial A pH<2	JC\MD	Ok
16	Q2231-02	MW-14-20250604	VN086954.D	11 Jun 2025 17:19	vial A pH<2	JC\MD	Ok,M
17	Q2231-03	MW-15-20250604	VN086955.D	11 Jun 2025 17:41	vial A pH<2	JC\MD	Ok
18	Q2231-04	MW-16D-20250604	VN086956.D	11 Jun 2025 18:04	vial A pH<2	JC\MD	Ok

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN061125

Review By	John Carlone	Review On	6/12/2025 9:57:37 AM
Supervise By	Semsettin Yesilyurt	Supervise On	6/12/2025 3:55:33 PM
SubDirectory	VN061125	HP Acquire Method	HP Processing Method 82N060625W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP134284		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134285,VP134286		

19	Q2231-05	MW-17-20250604	VN086957.D	11 Jun 2025 18:26	vial A pH<2	JC\MD	Ok
20	Q2231-06	TB-20250604	VN086958.D	11 Jun 2025 18:48	vial A pH<2 TB	JC\MD	Ok
21	Q2210-01	TW1	VN086959.D	11 Jun 2025 19:10	vial A pH<2	JC\MD	Ok
22	Q2209-01	P01W	VN086960.D	11 Jun 2025 19:33	vial A pH<2	JC\MD	Ok
23	IBLK	IBLK	VN086961.D	11 Jun 2025 19:55		JC\MD	Ok
24	IBLK	IBLK	VN086962.D	11 Jun 2025 20:17		JC\MD	Ok
25	Q2263-01	RW9-MW01D3-202506	VN086963.D	11 Jun 2025 20:39	vial A pH<2	JC\MD	Ok
26	Q2263-02	RW9-MW01S-2025060	VN086964.D	11 Jun 2025 21:01	vial A pH<2	JC\MD	Ok
27	VSTDCCC050	VSTDCCC050EC	VN086965.D	11 Jun 2025 21:23		JC\MD	Ok,M

M : Manual Integration

Prep Standard - Chemical Standard Summary

Order ID : Q2263

Test : VOCMS Group1

Prepbatch ID :

Sequence ID/Qc Batch ID: vn061125,

Standard ID :

VP133174,VP133887,VP133935,VP133953,VP133991,VP134142,VP134149,VP134284,VP134285,VP134286,

Chemical ID :

V13391,V13706,V14290,V14432,V14435,V14503,V14504,V14525,V14526,V14613,V14620,V14626,V14636,V14637,V14638,V14639,V14668,V14671,V14673,V14675,V14711,V14717,V14718,V14721,V14749,V14750,V14811,V14812,V14843,V14921,V14929,V14944,V14945,V14946,V14947,W3112,

VOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
617	8260 Surrogate, 400PPM	VP133174	02/27/2025	08/27/2025	Semsettin Yesilyurt	None	None	Maresh Dadoda
								03/04/2025

FROM 0.40000ml of V13706 + 24.60000ml of V14613 = Final Quantity: 25.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
257	8260 Calibration Working STD Mix-First source, 160PPM	VP133887	05/12/2025	06/23/2025	Semsettin Yesilyurt	None	None	Maresh Dadoda
								05/14/2025

FROM 0.40000ml of V14843 + 1.00000ml of V14432 + 1.00000ml of V14435 + 1.00000ml of V14503 + 1.00000ml of V14504 + 1.00000ml of V14525 + 1.00000ml of V14526 + 1.00000ml of V14711 + 1.00000ml of V14717 + 1.00000ml of V14718 + 1.00000ml of V14721 + 1.00000ml of V14749 + 1.00000ml of V14750 + 1.00000ml of V14811 + 1.00000ml of V14812 + 10.60000ml of V14921 = Final Quantity: 25.000 ml

VOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
247	8260 Internal Standard, 250PPM	VP133935	05/16/2025	11/12/2025	Semsettin Yesilyurt	None	None	Maresh Dadoda
								05/21/2025

FROM 0.25000ml of V14290 + 24.75000ml of V14921 = Final Quantity: 25.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
218	BFB, 25PPM	VP133953	05/19/2025	11/09/2025	Semsettin Yesilyurt	None	None	Maresh Dadoda
								05/21/2025

FROM 0.25000ml of V13391 + 24.75000ml of V14626 = Final Quantity: 25.000 ml



VOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
51	8260 Working STD (Acrolein) -first source, 800PPM	VP133991	05/22/2025	06/19/2025	Semsettin Yesilyurt	None	None	Mahesh Dadoda 05/24/2025

FROM 1.00000ml of V14944 + 1.00000ml of V14945 + 1.00000ml of V14946 + 1.00000ml of V14947 + 21.00000ml of V14620 = Final
Quantity: 25.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
719	8260 Working STD (BCM)-First source, 400PPM	VP134142	06/06/2025	12/06/2025	Semsettin Yesilyurt	None	None	Mahesh Dadoda 06/10/2025

FROM 1.00000ml of V14668 + 1.00000ml of V14671 + 1.00000ml of V14673 + 1.00000ml of V14675 + 16.00000ml of V14929 = Final
Quantity: 20.000 ml



<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1810	8260 Working Std(2-CVE)-800ppm	VP134149	06/06/2025	12/06/2025	Semsettin Yesilyurt	None	None	Mahesh Dadoda 06/10/2025
<u>FROM</u>	1.00000ml of V14636 + 1.00000ml of V14637 + 1.00000ml of V14638 + 1.00000ml of V14639 + 46.00000ml of V14929 = Final Quantity: 50.000 ml							

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
589	BFB TUNE CHECK	VP134284	06/11/2025	06/12/2025	John Carlone	None	None	Mahesh Dadoda 06/12/2025
<u>FROM</u> 39.98400ml of W3112 + 0.01600ml of VP133953 = Final Quantity: 40.000 ml								



VOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
620	50 PPB CCC, 8260-Water	VP134285	06/11/2025	06/12/2025	John Carlone	None	None	Mahesh Dadoda 06/12/2025
<u>FROM</u>	39.94450ml of W3112 + 0.00500ml of VP133174 + 0.00500ml of VP134142 + 0.00800ml of VP133935 + 0.01250ml of VP133887 + 0.01250ml of VP133991 + 0.01250ml of VP134149 = Final Quantity: 40.000 ml							

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
620	50 PPB CCC, 8260-Water	VP134286	06/11/2025	06/12/2025	John Carlone	None	None	Mahesh Dadoda 06/12/2025
<u>FROM</u>	39.94450ml of W3112 + 0.00500ml of VP133174 + 0.00500ml of VP134142 + 0.00800ml of VP133935 + 0.01250ml of VP133887 + 0.01250ml of VP133991 + 0.01250ml of VP134149 = Final Quantity: 40.000 ml							

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30067 / BFB tuneing solution	A0191805	11/22/2025	11/22/2024 / SAM	01/13/2023 / SAM	V13391

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555582 / Custom Mixture, 8260 A/B Surrogate Mix [CS 5179-2]	A0196865	02/27/2026	02/27/2025 / SAM	04/12/2023 / SAM	V13706

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555581 / Custom Standard, 8260 Internal Std [CS 5179-1]	A0210184	12/12/2025	12/12/2024 / SAM	04/15/2024 / SAM	V14290

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30489 / VOA Mix, 8260B Acetates Mix, P&TM, 1mL	A0209618	09/30/2025	05/12/2025 / SAM	08/15/2024 / SAM	V14432

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30489 / VOA Mix, 8260B Acetates Mix, P&TM, 1mL	A0209618	09/20/2025	03/20/2025 / SAM	08/15/2024 / SAM	V14435

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	95317 / Universal VOA Mega Mix (Min order = 5)	021624	11/12/2025	05/12/2025 / SAM	09/17/2024 / SAM	V14503

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	95317 / Universal VOA Mega Mix (Min order = 5)	021624	11/12/2025	05/12/2025 / SAM	09/17/2024 / SAM	V14504

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	95319 / Revised Additions Mix (Min = 5)	091724	11/12/2025	05/12/2025 / SAM	09/18/2024 / SAM	V14525

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	95319 / Revised Additions Mix (Min = 5)	091724	11/12/2025	05/12/2025 / SAM	09/18/2024 / SAM	V14526

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA9077-02 / Methanol, Purge/Trap (cs=6x1L)	22L0562016	08/27/2025	02/27/2025 / SAM	11/26/2024 / SAM	V14613

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA9077-02 / Methanol, Purge/Trap (cs=6x1L)	22L0562016	10/25/2025	05/09/2025 / SAM	11/26/2024 / SAM	V14620

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA9077-02 / Methanol, Purge/Trap (cs=6x1L)	23I0762004	11/09/2025	05/09/2025 / SAM	11/26/2024 / SAM	V14626

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	/ 2-Chloroethyl vinyl ether	120524	12/06/2025	06/06/2025 / SAM	12/06/2024 / SAM	V14636

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	/ 2-Chloroethyl vinyl ether	120524	12/06/2025	06/06/2025 / SAM	12/06/2024 / SAM	V14637

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	/ 2-Chloroethyl vinyl ether	120524	12/06/2025	06/06/2025 / SAM	12/06/2024 / SAM	V14638

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	/ 2-Chloroethyl vinyl ether	120524	12/06/2025	06/06/2025 / SAM	12/06/2024 / SAM	V14639

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30225 / VOA Mix, bromochloromethane, 2000ug/mL, P&TM, 1mL/ampul	A0214960	12/06/2025	06/06/2025 / SAM	12/09/2024 / SAM	V14668

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30225 / VOA Mix, bromochloromethane, 2000ug/mL, P&TM, 1mL/ampul	A0214960	12/06/2025	06/06/2025 / SAM	12/09/2024 / SAM	V14671

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30225 / VOA Mix, bromochloromethane, 2000ug/mL, P&TM, 1mL/ampul	A0214960	12/06/2025	06/06/2025 / SAM	12/09/2024 / SAM	V14673

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30225 / VOA Mix, bromochloromethane, 2000ug/mL, P&TM, 1mL/ampul	A0214960	12/06/2025	06/06/2025 / SAM	12/09/2024 / SAM	V14675

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30006 / VOA Mix, CLP method Calibration Std #1 ketones 5000uq/ml, PTM, 1ml	A02110618	11/12/2025	05/12/2025 / SAM	12/17/2024 / SAM	V14711

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30006 / VOA Mix, CLP method Calibration Std #1 ketones 5000uq/ml, PTM, 1ml	A02110618	11/12/2025	05/12/2025 / SAM	12/17/2024 / SAM	V14717

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30006 / VOA Mix, CLP method Calibration Std #1 ketones 5000uq/ml, PTM, 1ml	A02110618	11/12/2025	05/12/2025 / SAM	12/17/2024 / SAM	V14718

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30006 / VOA Mix, CLP method Calibration Std #1 ketones 5000uq/ml, PTM, 1ml	A02110618	11/12/2025	05/12/2025 / SAM	12/17/2024 / SAM	V14721

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30042 / VOA Mix,500 series method 502.2 Calibration Std #1 gases, 2000uq/ml, PTM, 1ml	A0216826	11/13/2025	05/12/2025 / SAM	12/17/2024 / SAM	V14749

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30042 / VOA Mix,500 series method 502.2 Calibration Std #1 gases, 2000uq/ml, PTM, 1ml	A0216826	11/12/2025	05/12/2025 / SAM	12/17/2024 / SAM	V14750

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555408 / Custom Standard, Vinyl Acetate Standard w/ Grav [CS 5066-6] TWO SEPARATE	A0220471	11/12/2025	05/12/2025 / SAM	01/08/2025 / SAM	V14811

LOTS

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555408 / Custom Standard, Vinyl Acetate Standard w/ Grav [CS 5066-6] TWO SEPARATE	A0220471	06/30/2026	05/12/2025 / SAM	01/08/2025 / SAM	V14812

LOTS

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30470 / VOA Stock Solution, tert-butanol std, 1mL, P&TM	A0217535	11/12/2025	05/12/2025 / SAM	01/21/2025 / SAM	V14843

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA9077-02 / Methanol, Purge/Trap (cs=6x1L)	24G0262002	11/12/2025	05/12/2025 / SAM	05/09/2025 / SAM	V14921

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA9077-02 / Methanol, Purge/Trap (cs=6x1L)	24G0262002	12/06/2025	06/06/2025 / SAM	05/09/2025 / SAM	V14929

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	91980 / Acrolin Std (Min = 5)	051925	06/19/2025	05/22/2025 / SAM	05/21/2025 / SAM	V14944

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	91980 / Acrolin Std (Min = 5)	051925	06/19/2025	05/22/2025 / SAM	05/21/2025 / SAM	V14945

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	91980 / Acrolin Std (Min = 5)	051925	06/19/2025	05/22/2025 / SAM	05/21/2025 / SAM	V14946

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	91980 / Acrolin Std (Min = 5)	051925	06/19/2025	05/22/2025 / SAM	05/21/2025 / SAM	V14947

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	DIW / DI Water	Daily Lab-Certified	07/03/2029	07/03/2024 / lwona	07/03/2024 / lwona	W3112

Methanol
ULTRA RESI-ANALYZED
For Purge and Trap Analysis



Material No.: 9077-02
Batch No.: 23I0762004
Manufactured Date: 2023-08-11
Expiration Date: 2026-08-10
Revision No.: 0

Certificate of Analysis


Test	Specification	Result
Assay (CH ₃ OH) (by GC, corrected for water)	≥ 99.9 %	100.0 %
Residue after Evaporation	≤ 1.0 ppm	0.5 ppm
Titration Acid (μeq/g)	≤ 0.3	0.2
Titration Base (μeq/g)	≤ 0.10	0.01
Water (by KF, coulometric)	≤ 0.08 %	< 0.01 %
Volatile Organic Trace Analysis – Below EPA 8260B CRQL	Conforms	Conforms

For Laboratory, Research, or Manufacturing Use
Performance Tested for Use in EPA Methods
500 Series for Drinking Water
600 Series for Wastewater
846 for Solid Waste

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

Ken Koehnlein
Sr. Manager, Quality Assurance

Methanol
ULTRA RESI-ANALYZED
For Purge and Trap Analysis



Material No.: 9077-02
Batch No.: 22L0562016
Manufactured Date: 2022-10-26
Expiration Date: 2025-10-25
Revision No.: 0

Certificate of Analysis


Test	Specification	Result
Assay (CH ₃ OH) (by GC, corrected for water)	≥ 99.9 %	100.0 %
Residue after Evaporation	≤ 1.0 ppm	0.2 ppm
Titration Acid (μeq/g)	≤ 0.3	0.2
Titration Base (μeq/g)	≤ 0.10	0.03
Water (by KF, coulometric)	≤ 0.08 %	< 0.01 %
Volatile Organic Trace Analysis – Below EPA 8260B CRQL	Conforms	Conforms

For Laboratory, Research, or Manufacturing Use
Performance Tested for Use in EPA Methods
500 Series for Drinking Water
600 Series for Wastewater
846 for Solid Waste

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

Jamie Ethier
Vice President Global Quality

Methanol
ULTRA RESI-ANALYZED
For Purge and Trap Analysis



Material No.: 9077-02
Batch No.: 22L0562016
Manufactured Date: 2022-10-26
Expiration Date: 2025-10-25
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay (CH ₃ OH) (by GC, corrected for water)	≥ 99.9 %	100.0 %
Residue after Evaporation	≤ 1.0 ppm	0.2 ppm
Titration Acid (μeq/g)	≤ 0.3	0.2
Titration Base (μeq/g)	≤ 0.10	0.03
Water (by KF, coulometric)	≤ 0.08 %	< 0.01 %
Volatile Organic Trace Analysis – Below EPA 8260B CRQL	Conforms	Conforms

For Laboratory, Research, or Manufacturing Use
Performance Tested for Use in EPA Methods
500 Series for Drinking Water
600 Series for Wastewater
846 for Solid Waste

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

Jamie Ethier
Vice President Global Quality



Ree 03117/24

CERTIFIED WEIGHT REPORT

Part Number: 95317

Lot Number: 021624

Description: Universal VOA Megamix

69 components

Expiration Date: 021627

Recommended Storage: Freezer (0 °C)

Nominal Concentration (µg/mL): 2000

NIST Test ID#: BUTB

Solvent(s): Lot#
Methanol EG359-USQ12Weight(s) shown below were combined and diluted to (mL): 100.0 0.021 Balance Uncertainty
Flask Uncertainty

Formulated By:	Prashant Chauhan	021624
Reviewed By:	Pedro L. Renteria	021624
DATE		

Compound	(KMe) Part Number	Lot Number	DR Factor	Initial Vol. (mL)	Initial Conc. (µg/mL)	Nominal Conc. (µg/mL)	Purity (%)	Purity Uncertainty	Uncertainty Pipette (mL)	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
														CAS#	OSHA PEL (TWA)	LD50
1. Acetonitrile	(0324)	021644	NA	NA	NA	2000	99.99	0.2	NA	0.20007	0.20020	2001.3	8.1	75-05-8	40 ppm (70mg/m3/8H)	or-rat 2400mg/kg
2. Allyl chloride (3-Chloropropene)	(0325)	102396	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20221	2001.4	8.2	107-05-1	1 ppm (3mg/m3/8H)	or-rat 700mg/kg
3. Carbon disulfide	(0060)	MKCR8561	NA	NA	NA	2000	99.99	0.2	NA	0.20007	0.20023	2001.6	8.1	75-15-0	4 ppm (12mg/m3) (skin)	or-rat 1200mg/kg
4. cis-1,4-Dichloro-2-butene	(1196)	14718EF	NA	NA	NA	2000	95	0.2	NA	0.21056	0.21069	2001.1	8.5	1478-11-6	N/A	N/A
5. trans-1,4-Dichloro-2-butene	(0486)	MKBP9041V	NA	NA	NA	2000	96.5	0.2	NA	0.20731	0.20746	2001.7	8.4	110-57-6	N/A	N/A
6. Diethyl ether	(0153)	IK18CAS000C	NA	NA	NA	2000	99.9	0.2	NA	0.20025	0.20040	2001.5	8.1	60-29-7	N/A	N/A
7. Ethyl methacrylate	(0381)	06128PX	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20230	2002.3	8.2	97-63-2	N/A	N/A
8. Iodomethane	(0489)	SHBF8718V	NA	NA	NA	2000	99.5	0.2	NA	0.20106	0.20121	2001.5	8.2	74-88-4	5 ppm (28mg/m3/8H) (skin)	or-rat 14800mg/kg
9. 2-Methyl-1-propanol	(0445)	15241EB	NA	NA	NA	2000	99.5	0.2	NA	0.20106	0.20120	2001.4	8.1	78-83-1	50 ppm (150mg/m3/8H)	or-rat 3460mg/kg
10. Methacrylonitrile	(0472)	00427ET	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20221	2001.4	8.2	128-98-7	1 ppm (3mg/m3/8H) (skin)	or-rat 120mg/kg
11. Methyl acrylate	(1075)	SHBK0679	NA	NA	NA	2000	99.9	0.2	NA	0.20025	0.20040	2001.5	8.1	96-33-3	10 ppm (35mg/m3/8H) (skin)	or-rat 277mg/kg
12. Methyl methacrylate	(0404)	MKGW6137V	NA	NA	NA	2000	99.9	0.2	NA	0.20025	0.20041	2001.6	8.1	80-62-6	100 ppm (410mg/m3/8H)	or-rat 7872mg/kg
13. Nitrobenzene	(0228)	01213TV	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20220	2001.3	8.2	98-95-3	1 ppm (5mg/m3/8H) (skin)	or-rat 780mg/kg
14. 2-Nitropropane	(0461)	HG002JK	NA	NA	NA	2000	97.3	0.2	NA	0.20560	0.20577	2001.6	8.3	78-46-9	10 ppm (35mg/m3/8H)	or-rat 720mg/kg
15. Pentachloroethane	(0460)	HGA01	NA	NA	NA	2000	98	0.2	NA	0.20413	0.20430	2001.6	8.3	78-01-7	N/A	N/A
16. 1,1,2-Trichlorotrifluoroethane	(0474)	18990	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20225	2001.8	8.2	78-13-1	1000 ppm (7600mg/m3/8H)	or-rat 435g/kg
17. Bromodichloromethane	35171	101623	0.05	5.00	40001.7	2000	NA	NA	0.017	NA	NA	1999.6	22.9	75-27-4	NA	or-rat 915mg/kg
18. Dibromochloromethane	35171	101623	0.05	5.00	40002.1	2000	NA	NA	0.017	NA	NA	1999.6	23.0	124-48-1	N/A	or-rat 848mg/kg
19. cis-1,2-Dichloroethene	35171	101623	0.05	5.00	40003.1	2000	NA	NA	0.017	NA	NA	1999.7	22.9	156-59-2	N/A	N/A
20. trans-1,2-Dichloroethene	35171	101623	0.05	5.00	40003.4	2000	NA	NA	0.017	NA	NA	1999.6	23.0	156-60-5	N/A	or-rat 1235mg/kg
21. Methylene chloride	35171	101623	0.05	5.00	40002.8	2000	NA	NA	0.017	NA	NA	1999.6	22.9	75-09-2	500 ppm	or-rat 820mg/kg
22. 1,1-Dichloroethane	32251	102023	0.10	10.00	20001.6	2000	NA	NA	0.042	NA	NA	1999.7	20.4	75-35-4	1 ppm (4mg/m3/8H)	or-rat 200mg/kg
23. Bromoform	95321	020724	0.10	10.00	20003.2	2000	NA	NA	0.042	NA	NA	1999.8	20.5	75-25-2	0.5 ppm (5mg/m3) (skin)	or-rat 930mg/kg
24. Carbon tetrachloride	95321	020724	0.10	10.00	20003.4	2000	NA	NA	0.042	NA	NA	1999.8	20.4	56-23-5	2 ppm (12.6mg/m3/8H)	or-rat 2350mg/kg
25. Chloroform	95321	020724	0.10	10.00	20004.0	2000	NA	NA	0.042	NA	NA	1999.8	20.5	67-68-3	60 ppm (240mg/m3) (CL)	or-rat 908mg/kg
26. Dibromomethane	95321	020724	0.10	10.00	20002.9	2000	NA	NA	0.042	NA	NA	1999.8	20.5	74-95-3	N/A	or-rat 108mg/kg
27. 1,1-Dichloroethane	95321	020724	0.10	10.00	20003.4	2000	NA	NA	0.042	NA	NA	1999.8	20.4	594-20-7	N/A	N/A
28. 2,2-Dichloropropane	95321	020724	0.10	10.00	20003.4	2000	NA	NA	0.042	NA	NA	1999.8	20.4	594-20-7	N/A	N/A
29. Trichloroethene	95321	020724	0.10	10.00	20201.1	2000	NA	NA	0.042	NA	NA	2019.8	20.8	127-18-4	25 ppm (170mg/m3/8H) (final)	or-rat 2629mg/kg
30. 1,1,1-Trichloroethane	95321	020724	0.10	10.00	20003.0	2000	NA	NA	0.042	NA	NA	1999.8	20.5	71-55-6	350 ppm (1900mg/m3/8H)	or-rat 10000mg/kg
31. 1,2-Dibromo-3-chloropropane	35161	112322	0.05	5.00	40016.5	2000	NA	NA	0.017	NA	NA	2000.3	22.9	96-12-8	0.001 ppm	or-rat 170mg/kg
32. 1,2-Dibromoethane	35161	112322	0.05	5.00	40024.8	2000	NA	NA	0.017	NA	NA	2000.7	22.9	106-93-4	20 ppm (8H)	or-rat 108mg/kg
33. 1,2-Dichloroethane	35161	112322	0.05	5.00	40018.0	2000	NA	NA	0.017	NA	NA	2000.4	22.9	107-06-2	50 ppm (8H)	or-rat 670mg/kg
34. 1,2-Dichloropropane	35161	112322	0.05	5.00	40051.0	2000	NA	NA	0.017	NA	NA	2002.0	22.9	78-87-5	75 ppm (350mg/m3/8H)	or-rat 1847mg/kg
35. 1,3-Dichloropropane	35161	112322	0.05	5.00	40005.9	2000	NA	NA	0.017	NA	NA	1999.8	22.9	142-28-9	N/A	or-rat 3600mg/kg
36. 1,1-Dichloropropene	35161	112322	0.05	5.00	40012.1	2000	NA	NA	0.017	NA	NA	2000.1	28.7	563-58-6	N/A	N/A
37. cis-1,3-Dichloropropene	35161	112322	0.05	5.00	40010.0	2000	NA	NA	0.017	NA	NA	2000.0	23.0	10061-01-5	N/A	N/A
38. trans-1,3-Dichloropropene	35161	112322	0.05	5.00	40017.6	2000	NA	NA	0.017	NA	NA	2000.4	23.0	10061-02-6	N/A	N/A
39. Hexachloro-1,3-butadiene	35161	112322	0.05	5.00	40021.9	2000	NA	NA	0.017	NA	NA	2000.6	29.7	87-68-3	0.02 ppm (0.24mg/m3/8H)	or-rat 82mg/kg
40. 1,1,1,2-Tetrachloroethane	35161	112322	0.05	5.00	40011.9	2000	NA	NA	0.017	NA	NA	2000.1	22.9	830-20-6	N/A	or-rat 670mg/kg
41. 1,1,2,2-Tetrachloroethane	35161	112322	0.05	5.00	40007.5	2000	NA	NA	0.017	NA	NA	1999.9	22.9	79-34-5	5 ppm (35mg/m3/8H) (skin)	or-rat 800mg/kg
42. 1,1,2-Trichloroethane	35161	112322	0.05	5.00	40006.6	2000	NA	NA	0.017	NA	NA	1999.8	23.0	79-00-5	10 ppm (45mg/m3/8H) (skin)	or-rat 936mg/kg
43. Trichloroethene	35161	112322	0.05	5.00	40029.0	2000	NA	NA	0.017	NA	NA	2000.9	22.9	79-01-6	50 ppm (270mg/m3/8H)	or-mus 2402mg/kg
44. 1,2,3-Trichloropropane	35161	112322	0.05	5.00	40007.5	2000	NA	NA	0.017	NA	NA	1999.9	22.9	96-18-4	10 ppm (60mg/m3/8H)	or-rat 149.8mg/kg
45. Benzene	35162	050823	0.05	5.00	40007.5	2000	NA	NA	0.017	NA	NA	1999.7	22.9	71-43-2	1 ppm	or-rat 4894mg/kg
46. Bromobenzene	35162	050823	0.05	5.00	40006.0	2000	NA	NA	0.017	NA	NA	1999.7	22.9	104-51-8	N/A	N/A
47. n-Butyl benzene	35162	050823	0.05	5.00	40003.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	100-41-4	100 ppm (435mg/m3/8H)	or-rat >2000mg/kg
48. Ethyl benzene	35162	050823	0.05	5.00	40004.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	99-87-6	N/A	or-rat 4750mg/kg
49. p-Isopropyl toluene	35162	050823	0.05	5.00	40005.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	91-20-3	10 ppm (50mg/m3/8H)	or-rat 490mg/kg
50. Naphthalene	35162	050823	0.05	5.00	40005.2	2000	NA	NA	0.017	NA	NA	1999.8	22.9	100-42-5	100 ppm	or-rat 5000mg/kg
51. Styrene	35162	050823	0.05	5.00	40004.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	108-98-3	200 ppm	or-rat 5000mg/kg
52. Toluene	35162	050823	0.05	5.00	40006.2	2000	NA	NA	0.017	NA	NA	1999.7	22.9	87-81-6	N/A	or-rat 1390mg/kg
53. 1,2,3-Trichlorobenzene	35162	050823	0.05	5.00	40003.1	2000	NA	NA	0.017	NA	NA	1999.8	22.9	120-82-1	5 ppm (CL) (40mg/m3)	or-rat 756mg/kg
54. 1,2,4-Trichlorobenzene	35162	050823	0.05	5.00	40006.8	2000	NA	NA	0.017	NA	NA	1999.8	23.0	95-63-6	N/A	or-rat 5g/kg
55. 1,2,4-Trimethylbenzene	35162	050823	0.05	5.00	40001.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	108-87-8	N/A	or-rat 5000mg/kg
56. 1,3,5-Trimethylbenzene	35162	050823	0.05	5.00	40006.7	2000	NA	NA	0.017	NA	NA	1999.8	22.9	108-38-3	100 ppm (435mg/m3/8H)	or-rat 5g/kg
57. m-Xylene	35162	050823	0.05	5.00	40005.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	96-06-6	N/A	N/A
58. tert-Butyl benzene	35163	101923	0.05	5.00	40001.2	2000	NA	NA	0.017	NA	NA	1999.8	22.9	135-98-8	N/A	or-rat 2240mg/kg
59. sec-Butyl benzene	35163	101923	0.05	5.00	40002.4	2000	NA	NA	0.017	NA	NA	1999.7	22.9	108-90-7	75 ppm (350mg/m3/8H)	or-rat 2290mg/kg
60. Chlorobenzene	35163	101923	0.05	5.00	40003.8	2000	NA	NA	0.017	NA	NA	1999.5	22.9	95-49-8	50 ppm (250mg/m3/8H)	or-rat 3900mg/kg
61. 2-Chlorotoluene	35163	101923	0.05	5.00	40003.3	2000	NA	NA	0.017	NA	NA	1999.7	22.9	106-43-4	N/A	or-rat 2100mg/kg
62. 4-Chlorotoluene	35163	101923	0.05	5.00	40003.3	2000	NA	NA	0.017	NA	NA	1999.7	22.9	95-50-1	50 ppm (300mg/m3) (CL)	or-rat 500mg/kg
63. 1,2-Dichlorobenzene	35163	101923	0.05	5.00	40001.7	2000	NA	NA	0.017	NA	NA	1999.6	23.0	541-73-1	N/A	or-mus 1062mg/kg
64. 1,3-Dichlorobenzene	35163	101923	0.05	5.00	40001.7	2000	NA	NA	0.017	NA	NA	1999.6	22.9	108-46-7	75 ppm (450mg/m3/8H)	or-rat




Run 16, "P95317 L021624 (2000µg/mL in MeOH)"

Run Length: 60.00 min, 35998 points at 10 points/second.
Created: Sat, Feb 17, 2024 at 8:56:46 AM.
Sampled: Sequence "021624-GC5M1", Method "GC5-M1".
Analyzed using Method "GC5-M1".

Comments

GC5-M1 Analysis by Candice Warren
Column ID SPB-Vocol 105 meter X 0.53mm X 3.0µm film thickness
Flow rates: Total flow=290mL/min., Helium (carrier)=10mL/min.,
Helium(make-up)=10mL/min., Hydrogen(make-up)=40mL/min., Air(make-up)=230mL/min.
Oven Profile: Temp. 1=35°C (Time 1=10 min.), Temp 2=200°C (Time 2=8.75 min.),
Rate = 4°C/min., Total run time=60 min. Injector temp.=200°C, FID Temp.=200°C.
FID Signal = Edaq Channel 1
Standard injection = 0.5µL, Range=3



Peak #	Name	FID RT (min.)
1	Ether	9.97
2	1,1,2-Trichloro-1,2,2-trifluoroethane	10.53
3	1,1-Dichloroethane	11.10
4	Acetone	12.00
5	Isodimethane	12.21
6	Allyl chloride	12.56
7	Carbon disulfide/Methylene chloride	13.04
8	trans-1,2-Dichloroethane	14.07
9	1,1-Dichloroethane	15.74
10	2,2-Dichloropropane	17.74
11	cis-1,2-Dichloroethane	18.00
12	Methyl methacrylate/Methyl acrylate/Chloroform	18.49
13	Isobutene/1,1,1-Trichloroethane	18.91
14	1,1-Dichloropropane	20.46
15	Carbon tetrachloride	20.79
16	Benzene/1,2-Dichloroethane	21.48
17	Trichloroethane	23.88
18	1,2-Dichloropropane	24.52
19	Methyl methacrylate	25.13
20	Bromochloropropane	25.46
21	Dibromomethane/2,2-Dichloropropane	27.07
22	cis-1,2-Dichloroethane	28.03
23	Toluene	28.73
24	Ethyl methacrylate/Trans-1,2-Dichloropropane	29.34
25	1,1,2-Trichloroethane	30.34
26	Tetrachloroethane/1,2-Dichloropropane	31.16
27	Dibromochloromethane	31.84
28	1,2-Dibromomethane	33.06
29	Chlorobenzene	33.40
30	Ethylbenzene/1,1,1,2-Tetrachloroethane	33.85
31	m-Xylene/p-Xylene	35.33
32	o-Xylene	35.70
33	Styrene	35.70
34	Isopropylbenzene/Bromofarm	36.48
35	cis-1,4-Dichloro-3-butene	36.80
36	1,1,2-Trichloropropane	37.23
37	n-Propylbenzene	37.77
38	trans-1,4-Dichloro-3-butene	37.92
39	Bromobenzene	38.05
40	Bromobenzene	38.14
41	1,3,5-Trimethylbenzene	38.50
42	Chlorobenzene	38.63
43	4-Chlorobenzene	38.77
44	tert-Butylbenzene	39.76
45	1,2,4-Trimethylbenzene	39.91
46	Pentachloroethane	40.17
47	sec-Butylbenzene	40.52
48	p-Isopropylbenzene	41.02
49	1,3-Trichlorobenzene	41.42
50	1,4-Dichlorobenzene	41.83
51	n-Butylbenzene	42.18
52	1,2-Dichlorobenzene	42.18
53	1,2-Dibromo-3-chloropropane	46.12
54	Nitrobenzene	46.40
55	1,2,4-Trifluorobenzene	49.26
56	Hexachlorocyclopentadiene	49.72
57	Naphthalene	50.56
58	1,2,3-Trichlorobenzene	51.16

Safety Data Sheet (SDS)

GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY ANALYTICAL STANDARD DISSOLVED IN METHANOL

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	1-800-535-5053
Address	44 Rossotto Dr. Hamden CT, 06514	Emergency Telephone International	1-352-323-3500
		Date Prepared/Revised	January 1, 2024

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

H225 H370 P271 P302,332	Highly Flammable Liquid and Vapor Cause damage to organs Use in ventilated area If on skin, wash with soap and water	H301, 311, 331 H351 P280 P305,351,338	Toxic if swallowed, skin contact, inhaled Suspected of causing cancer Use gloves, eye protection/face shield If in eyes, remove contacts, rinse with water



Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s))	CAS#	% (optional)
Methanol METHYL ALCOHOL	67-56-1	> 97

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance. Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Flammability	Flammable in the presence of a source of ignition when the temperature is above the flash point. Keep away from heat/sparks/open flame/hot surface. No smoking.
Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist.
Storage Conditions	Use ventilation. Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge. Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Methanol	67-56-1 TWA 200 ppm
Skin notation	TWA 200 ppm
Potential for skin absorption	ingestion and inhalation.
Personal protective equipment	Respiratory protection Handle with gloves. Gloves must be inspected prior to use. Eye protection.
Avoid contact with skin, eyes and clothing.	Wash hands thoroughly after handling the product.

Section IX - Physical/Chemical Characteristics

Boiling Point	65°C	Specific Gravity (H ₂ O = 1)	0.79
Vapor Pressure (mm Hg)	96	Melting Point	-98°C
Vapor Density (AIR = 1)	1.11	Evaporation rate (Butyl Acetate = 1)	4.6
Solubility in Water	COMPLETE		
Appearance and Odor	CLEAR, COLORLESS LIQUID WITH CHARACTERISTIC PUNGENT ODOR.		

Section X. STABILITY AND REACTIVITY

Chemical stability	Stable under recommended storage conditions.
Possibility of hazardous reactions	Vapours may form explosive mixture with air.
Conditions to avoid	Heat, flames, sparks, extreme temperature and sunlight.
Materials to avoid	Acid chlorides, Acid anhydrides, Oxidizing agents, Alkali metals, Reducing agents, Acids
Hazardous decomposition products formed under fire conditions.	- Carbon oxides

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - rat - 5,628 mg/kg
LC50 Inhalation - rat - 4 h - 64000 ppm
LD50 Dermal - rabbit - 15,800 mg/kg
Toxic if absorbed through skin. Causes skin irritation.
Eye damage/eye irritation
Toxic if inhaled. Causes respiratory tract irritation.
Toxic if swallowed.

Section XII. ECOLOGICAL INFORMATION FOR REPORTABLE QUANTITY OF 5000 lbs.

LC50 15,400 mg/l - 96 h
EC50 24,500.00 mg/l - 48 h
EC100 10,000.00 mg/l - 24 h

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)	IATA
UN number: 1230 Class: 3 Packing group: II	UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol	Proper shipping name: Methanol

Section XV. REGULATORY INFORMATION

OSHA Hazards Flammable liquid, Target Organ Effect, Toxic by inhalation., Toxic by ingestion, Toxic by skin absorption, Irritant
SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.



Ree 03117/24

CERTIFIED WEIGHT REPORT

Part Number: 95317

Lot Number: 021624

Description: Universal VOA Megamix

69 components

Expiration Date: 021627

Recommended Storage: Freezer (0 °C)

Nominal Concentration (µg/mL): 2000

NIST Test ID#: BUTB

Solvent(s): Lot#
Methanol EG359-USQ12Weight(s) shown below were combined and diluted to (mL): 100.0 0.021 Balance Uncertainty
Flask Uncertainty

Formulated By:	Prashant Chauhan	021624
Reviewed By:	Pedro L. Renteria	021624
DATE		


Compound	(K049)	Lot	Dil.	Initial	Initial	Nominal	Purity	Purity	Uncertainty	Target	Actual	Actual	Expanded	SDS Information		
	Part Number	Number	Factor	Vol. (mL)	Conc. (µg/mL)	Conc. (µg/mL)	(%)	Uncertainty	Pipette (mL)	Weight(g)	Weight(g)	Conc (µg/mL)	Uncertainty	(Solvent Safety Info. On Attached pg.)	OSHA PEL (TWA)	LD50
1. Acetonitrile	(0324)	021644	NA	NA	NA	2000	99.99	0.2	NA	0.20007	0.20020	2001.3	8.1	75-05-8	40 ppm (70mg/m3/8H)	or-rat 2480mg/kg
2. Allyl chloride (3-Chloropropene)	(0325)	102396	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20221	2001.4	8.2	107-05-1	1 ppm (3mg/m3/8H)	or-rat 700mg/kg
3. Carbon disulfide	(0060)	MKCR8561	NA	NA	NA	2000	99.99	0.2	NA	0.20007	0.20023	2001.6	8.1	75-15-0	4 ppm (12mg/m3) (skin)	or-rat 1200mg/kg
4. cis-1,4-Dichloro-2-butene	(1196)	14718EF	NA	NA	NA	2000	95	0.2	NA	0.21058	0.21069	2001.1	8.5	1478-11-6	N/A	N/A
5. trans-1,4-Dichloro-2-butene	(0486)	MKBP6041V	NA	NA	NA	2000	96.5	0.2	NA	0.20731	0.20746	2001.7	8.4	110-57-6	N/A	N/A
6. Diethyl ether	(0153)	IK18CAS000C	NA	NA	NA	2000	99.9	0.2	NA	0.20025	0.20040	2001.5	8.1	60-29-7	N/A	N/A
7. Ethyl methacrylate	(0381)	06128PX	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20230	2002.3	8.2	97-63-2	N/A	or-rat 14800mg/kg
8. Iodomethane	(0489)	SHBF8718V	NA	NA	NA	2000	99.5	0.2	NA	0.20106	0.20121	2001.5	8.2	74-88-4	5 ppm (28mg/m3/8H) (skin)	or-rat 75mg/kg
9. 2-Methyl-1-propanol	(0445)	15241EB	NA	NA	NA	2000	99.5	0.2	NA	0.20106	0.20120	2001.4	8.1	78-83-1	50 ppm (150mg/m3/8H)	or-rat 2480mg/kg
10. Methacrylonitrile	(0442)	00427ET	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20221	2001.4	8.2	128-98-7	1 ppm (3mg/m3/8H) (skin)	or-rat 127mg/kg
11. Methyl acrylate	(1075)	SHBK0679	NA	NA	NA	2000	99.9	0.2	NA	0.20025	0.20040	2001.5	8.1	96-33-3	10 ppm (35mg/m3/8H) (skin)	or-rat 277mg/kg
12. Methyl methacrylate	(0404)	MKGW6137V	NA	NA	NA	2000	99.9	0.2	NA	0.20025	0.20041	2001.6	8.1	80-62-6	100 ppm (410mg/m3/8H)	or-rat 7872mg/kg
13. Nitrobenzene	(0228)	01213TV	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20220	2001.3	8.2	98-95-3	1 ppm (5mg/m3/8H) (skin)	or-rat 780mg/kg
14. 2-Nitropropane	(0461)	HG002JK	NA	NA	NA	2000	97.3	0.2	NA	0.20560	0.20577	2001.6	8.3	78-46-9	10 ppm (35mg/m3/8H)	or-rat 720mg/kg
15. Perchloroethane	(0460)	HGA01	NA	NA	NA	2000	98	0.2	NA	0.20413	0.20430	2001.6	8.3	78-01-7	N/A	N/A
16. 1,1,2-Trichlorotrifluoroethane	(0474)	18990	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20225	2001.8	8.2	78-13-1	1000 ppm (7600mg/m3/8H)	or-rat 435g/kg
17. Bromodichloromethane	35171	101623	0.05	5.00	40001.7	2000	NA	NA	0.017	NA	NA	1999.6	22.9	75-27-4	N/A	or-rat 915mg/kg
18. Dibromochloromethane	35171	101623	0.05	5.00	40002.1	2000	NA	NA	0.017	NA	NA	1999.6	23.0	124-48-1	N/A	or-rat 848mg/kg
19. cis-1,2-Dichloroethene	35171	101623	0.05	5.00	40003.1	2000	NA	NA	0.017	NA	NA	1999.7	22.9	158-59-2	N/A	N/A
20. trans-1,2-Dichloroethene	35171	101623	0.05	5.00	40003.2	2000	NA	NA	0.017	NA	NA	1999.6	23.0	158-60-5	N/A	or-rat 1235mg/kg
21. Methylene chloride	35171	101623	0.05	5.00	40002.8	2000	NA	NA	0.017	NA	NA	1999.6	22.9	75-09-2	500 ppm	or-rat 820mg/kg
22. 1,1-Dichloroethane	32261	102023	0.10	10.00	20001.6	2000	NA	NA	0.042	NA	NA	1999.8	20.5	75-35-4	1 ppm (4mg/m3/8H)	or-rat 200mg/kg
23. Bromoform	95321	020724	0.10	10.00	20003.2	2000	NA	NA	0.042	NA	NA	1999.8	20.5	75-25-2	0.5 ppm (5mg/m3) (skin)	or-rat 933mg/kg
24. Carbon tetrachloride	95321	020724	0.10	10.00	20003.4	2000	NA	NA	0.042	NA	NA	1999.8	20.4	58-23-5	2 ppm (12.6mg/m3/8H)	or-rat 2350mg/kg
25. Chloroform	95321	020724	0.10	10.00	20024.0	2000	NA	NA	0.042	NA	NA	2001.3	20.5	67-66-3	60 ppm (240mg/m3) (CL)	or-rat 908mg/kg
26. Dibromomethane	95321	020724	0.10	10.00	20002.9	2000	NA	NA	0.042	NA	NA	1999.8	20.5	74-95-3	N/A	or-rat 108mg/kg
27. 1,1-Dichloroethane	95321	020724	0.10	10.00	20003.4	2000	NA	NA	0.042	NA	NA	1999.8	20.5	75-34-3	100 ppm	or-rat 725mg/kg
28. 2,2-Dichloropropane	95321	020724	0.10	10.00	20003.4	2000	NA	NA	0.042	NA	NA	1999.8	20.4	594-20-7	N/A	N/A
29. Tetrachloroethene	95321	020724	0.10	10.00	20201.1	2000	NA	NA	0.042	NA	NA	2019.8	20.8	127-18-4	25 ppm (170mg/m3/8H) (final)	or-rat 2629mg/kg
30. 1,1,1-Trichloroethane	95321	020724	0.10	10.00	20003.0	2000	NA	NA	0.042	NA	NA	1999.8	20.5	71-55-6	350 ppm (1800mg/m3/8H)	or-rat 10300mg/kg
31. 1,2-Dibromo-3-chloropropane	35161	112322	0.05	5.00	40018.5	2000	NA	NA	0.017	NA	NA	2000.3	22.9	86-12-8	0.001 ppm	or-rat 170mg/kg
32. 1,2-Dibromoethane	35161	112322	0.05	5.00	40024.8	2000	NA	NA	0.017	NA	NA	2000.7	22.9	108-93-4	20 ppm (8H)	or-rat 108mg/kg
33. 1,2-Dichloroethane	35161	112322	0.05	5.00	40018.0	2000	NA	NA	0.017	NA	NA	2000.4	22.9	107-06-2	50 ppm (8H)	or-rat 670mg/kg
34. 1,2-Dichloropropane	35161	112322	0.05	5.00	40051.0	2000	NA	NA	0.017	NA	NA	2002.0	22.9	78-87-5	75 ppm (350mg/m3/8H)	or-rat 1847mg/kg
35. 1,3-Dichloropropane	35161	112322	0.05	5.00	40005.9	2000	NA	NA	0.017	NA	NA	1999.8	22.9	142-28-9	N/A	or-rat 3600mg/kg
36. 1,1-Dichloropropene	35161	112322	0.05	5.00	40012.1	2000	NA	NA	0.017	NA	NA	2000.1	28.7	563-58-6	N/A	N/A
37. cis-1,3-Dichloropropene	35161	112322	0.05	5.00	40010.0	2000	NA	NA	0.017	NA	NA	2000.0	23.0	10061-01-5	N/A	N/A
38. trans-1,3-Dichloropropene	35161	112322	0.05	5.00	40017.6	2000	NA	NA	0.017	NA	NA	2000.4	23.0	10061-02-6	N/A	N/A
39. Hexachloro-1,3-butadiene	35161	112322	0.05	5.00	40021.9	2000	NA	NA	0.017	NA	NA	2000.6	29.7	87-68-3	0.02 ppm (0.24mg/m3/8H)	or-rat 82mg/kg
40. 1,1,1,2-Tetrachloroethane	35161	112322	0.05	5.00	40011.9	2000	NA	NA	0.017	NA	NA	2000.1	22.9	830-20-6	N/A	or-rat 670mg/kg
41. 1,1,2,2-Tetrachloroethane	35161	112322	0.05	5.00	40007.5	2000	NA	NA	0.017	NA	NA	1999.9	22.9	79-34-5	5 ppm (35mg/m3/8H) (skin)	or-rat 800mg/kg
42. 1,1,2-Trichloroethane	35161	112322	0.05	5.00	40006.6	2000	NA	NA	0.017	NA	NA	1999.8	23.0	79-00-5	10 ppm (45mg/m3/8H) (skin)	or-rat 936mg/kg
43. Trichloroethene	35161	112322	0.05	5.00	40029.0	2000	NA	NA	0.017	NA	NA	2000.9	22.9	79-01-6	50 ppm (270mg/m3/8H)	or-mus 2402mg/kg
44. 1,2,3-Trichloropropane	35161	112322	0.05	5.00	40007.5	2000	NA	NA	0.017	NA	NA	1999.9	22.9	98-18-4	10 ppm (60mg/m3/8H)	or-rat 149.8mg/kg
45. Benzene	35162	050823	0.05	5.00	40006.0	2000	NA	NA	0.017	NA	NA	1999.7	22.9	71-43-2	1 ppm	or-rat 4894mg/kg
46. Bromobenzene	35162	050823	0.05	5.00	40003.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	104-51-8	N/A	or-rat 2699mg/kg
47. n-Butyl benzene	35162	050823	0.05	5.00	40004.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	100-41-4	100 ppm (435mg/m3/8H)	or-rat >2000mg/kg
48. Ethyl benzene	35162	050823	0.05	5.00	40004.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	99-87-6	N/A	or-rat 4750mg/kg
49. p-Isopropyl toluene	35162	050823	0.05	5.00	40005.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	91-20-3	10 ppm (50mg/m3/8H)	or-rat 490mg/kg
50. Naphthalene	35162	050823	0.05	5.00	40005.2	2000	NA	NA	0.017	NA	NA	1999.8	22.9	100-42-5	100 ppm	or-rat 5000mg/kg
51. Styrene	35162	050823	0.05	5.00	40004.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	108-98-3	200 ppm	or-rat 5000mg/kg
52. Toluene	35162	050823	0.05	5.00	40008.2	2000	NA	NA	0.017	NA	NA	1999.7	22.9	87-81-6	N/A	or-mus 1390mg/kg
53. 1,2,3-Trichlorobenzene	35162	050823	0.05	5.00	40003.1	2000	NA	NA	0.017	NA	NA	1999.8	22.9	120-82-1	5 ppm (CL) (40mg/m3)	or-rat 756mg/kg
54. 1,2,4-Trichlorobenzene	35162	050823	0.05	5.00	40006.8	2000	NA	NA	0.017	NA	NA	1999.6	23.0	95-63-6	N/A	or-rat 5g/kg
55. 1,2,4-Trimethylbenzene	35162	050823	0.05	5.00	40001.6	2000	NA	NA	0.017	NA	NA	1999.8	22.9	108-87-8	N/A	or-rat 5000mg/kg
56. 1,3,5-Trimethylbenzene	35162	050823	0.05	5.00	40006.7	2000	NA	NA	0.017	NA	NA	1999.8	22.9	108-38-3	100 ppm (435mg/m3/8H)	or-rat 5g/kg
57. m-Xylene	35162	050823	0.05	5.00	40005.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	96-06-6	N/A	N/A
58. tert-Butyl benzene	35163	101923	0.05	5.00	40001.2	2000	NA	NA	0.017	NA	NA	1999.8	22.9	135-98-8	N/A	or-rat 2240mg/kg
59. sec-Butyl benzene	35163	101923	0.05	5.00	40002.4	2000	NA	NA	0.017	NA	NA	1999.8	22.9	108-90-7	75 ppm (350mg/m3/8H)	or-rat 2290mg/kg
60. Chlorobenzene	35163	101923	0.05	5.00	40003.8	2000	NA	NA	0.017	NA	NA	1999.5	22.9	95-49-8	50 ppm (250mg/m3/8H)	or-rat 3900mg/kg
61. 2-Chlorotoluene	35163	101923	0.05	5.00	40003.3	2000	NA	NA	0.017	NA	NA	1999.7	22.9	106-43-4	N/A	or-rat 2100mg/kg
62. 4-Chlorotoluene	35163	101923	0.05	5.00	40003.3	2000	NA	NA	0.017	NA	NA	1999.7	22.9	95-50-1	50 ppm (300mg/m3) (CL)	or-rat 500mg/kg
63. 1,2-Dichlorobenzene	35163	101923	0.05	5.00	40001.7	2000	NA	NA	0.017	NA	NA	1999.6	23.0	541-73-1	N/A	or-mus 1062mg/kg
64. 1,3-Dichlorobenzene	35163	101923	0.05	5.00	40001.7	2000	NA	NA	0.017	NA	NA	1999.6	22.9	108-46-7	75 ppm (450mg/m3/8H)	or-rat 500mg/kg
65. 1,4-Dichlorobenzene	35															



Run 16, "P95317 L021624 (2000µg/mL in MeOH)"

Run Length: 60.00 min, 35998 points at 10 points/second.
Created: Sat, Feb 17, 2024 at 8:56:46 AM.
Sampled: Sequence "021624-GC5M1", Method "GC5-M1".
Analyzed using Method "GC5-M1".

Comments
GC5-M1 Analysis by Candice Warren
Column ID SPB-Vocol 105 meter X 0.53mm X 3.0µm film thickness
Flow rates: Total flow=290mL/min., Helium (carrier)=10mL/min.,
Helium(make-up)=10mL/min., Hydrogen(make-up)=40mL/min., Air(make-up)=230mL/min.
Oven Profile: Temp. 1=35°C (Time 1=10 min.), Temp 2=200°C (Time 2=8.75 min.),
Rate = 4°C/min., Total run time=60 min. Injector temp.=200°C, FID Temp.=200°C.
FID Signal = Edaq Channel 1
Standard injection = 0.5µL, Range=3



Peak #	Name	FID RT (min.)
1	Ether	9.97
2	1,1,2-Trichloro-1,2,2-trifluoroethane	10.53
3	1,1-Dichloroethane	11.10
4	Acetonitrile	12.60
5	Iodomethane	12.91
6	Allyl chloride	12.96
7	Carbon disulfide/Methylene chloride	13.04
8	trans-1,2-Dichloroethene	14.07
9	1,1-Dichloroethane	15.74
10	2,2-Dichloropropane	17.74
11	cis-1,2-Dichloroethene	18.00
12	Methoxyvinylbenzene/Methyl acrylate/Chloroform	18.49
13	Isobutene/1,1,1-Trichloroethane	18.91
14	1,1-Dichloropropane	20.46
15	Carbon tetrachloride	20.79
16	Benzene/1,2-Dichloroethane	21.48
17	Trichloroethene	23.88
18	1,2-Dichloropropane	24.52
19	Methyl methacrylate	25.13
20	Bromochloropropane	25.46
21	Dibromomethane/2,2-Dichloropropane	27.07
22	cis-1,2-Dichloroethene	28.03
23	Toluene	28.73
24	Ethyl methacrylate/Trans-1,2-Dichloropropane	29.34
25	1,1,2-Trichloroethane	30.34
26	Tetrachloroethane/1,2-Dichloropropane	31.16
27	Dibromochloromethane	32.84
28	1,2-Dibromomethane	33.26
29	Chlorobenzene	33.40
30	Ethylbenzene/1,1,1,2-Tetrachloroethane	33.85
31	m-Xylene/p-Xylene	35.33
32	o-Xylene	35.70
33	Styrene	35.70
34	Isopropylbenzene/Bromofarm	36.48
35	cis-1,4-Dichloro-3-butene	36.80
36	1,1,2-Trichloropropane	37.23
37	n-Propylbenzene	37.77
38	trans-1,4-Dichloro-3-butene	37.92
39	Bromobenzene	38.05
40	1,3,5-Trimethylbenzene	38.14
41	1,3,5-Trimethylbenzene	38.50
42	Chlorobenzene	38.63
43	4-Chlorobenzene	38.77
44	tert-Butylbenzene	39.76
45	1,2,4-Trimethylbenzene	39.91
46	Pentachloroethane	40.17
47	sec-Butylbenzene	40.52
48	p-Isopropylbenzene	41.02
49	1,3-Trichlorobenzene	41.42
50	1,4-Dichlorobenzene	41.83
51	n-Butylbenzene	42.52
52	1,2-Dichlorobenzene	42.18
53	1,2-Dibromo-3-chloropropane	46.12
54	Nitrobenzene	46.40
55	1,2,4-Trifluorobenzene	49.28
56	Hexachlorocyclopentadiene	49.72
57	Naphthalene	50.56
58	1,2,3-Trichlorobenzene	51.16

Safety Data Sheet (SDS)

GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY ANALYTICAL STANDARD DISSOLVED IN METHANOL

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	1-800-535-5053
Address	44 Rossotto Dr. Hamden CT, 06514	Emergency Telephone International	1-352-323-3500
		Date Prepared/Revised	January 1, 2024

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

H225 H370 P271 P302,332	Highly Flammable Liquid and Vapor Cause damage to organs Use in ventilated area If on skin, wash with soap and water	H301, 311, 331 H351 P280 P305,351,338	Toxic if swallowed, skin contact, inhaled Suspected of causing cancer Use gloves, eye protection/face shield If in eyes, remove contacts, rinse with water



Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s))	CAS#	% (optional)
Methanol METHYL ALCOHOL	67-56-1	> 97

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance. Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Flammability	Flammable in the presence of a source of ignition when the temperature is above the flash point. Keep away from heat/sparks/open flame/hot surface. No smoking.
Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist.
Storage Conditions	Use ventilation. Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge. Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Methanol	67-56-1 TWA 200 ppm
Skin notation	TWA 200 ppm
Potential for skin absorption	ingestion and inhalation.
Personal protective equipment	Respiratory protection Handle with gloves. Gloves must be inspected prior to use. Eye protection.
Avoid contact with skin, eyes and clothing.	Wash hands thoroughly after handling the product.

Section IX - Physical/Chemical Characteristics

Boiling Point	65°C	Specific Gravity (H ₂ O = 1)	0.79
Vapor Pressure (mm Hg)	96	Melting Point	-98°C
Vapor Density (AIR = 1)	1.11	Evaporation rate (Butyl Acetate = 1)	4.6
Solubility in Water	COMPLETE		
Appearance and Odor	CLEAR, COLORLESS LIQUID WITH CHARACTERISTIC PUNGENT ODOR.		

Section X. STABILITY AND REACTIVITY

Chemical stability	Stable under recommended storage conditions.
Possibility of hazardous reactions	Vapours may form explosive mixture with air.
Conditions to avoid	Heat, flames, sparks, extreme temperature and sunlight.
Materials to avoid	Acid chlorides, Acid anhydrides, Oxidizing agents, Alkali metals, Reducing agents, Acids
Hazardous decomposition products formed under fire conditions.	- Carbon oxides

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - rat - 5,628 mg/kg
LC50 Inhalation - rat - 4 h - 64000 ppm
LD50 Dermal - rabbit - 15,800 mg/kg
Toxic if absorbed through skin. Causes skin irritation.
Eye damage/eye irritation
Toxic if inhaled. Causes respiratory tract irritation.
Toxic if swallowed.

Section XII. ECOLOGICAL INFORMATION FOR REPORTABLE QUANTITY OF 5000 lbs.

LC50 15,400 mg/l - 96 h
EC50 24,500.00 mg/l - 48 h
EC100 10,000.00 mg/l - 24 h

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)	IATA
UN number: 1230 Class: 3 Packing group: II	UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol	Proper shipping name: Methanol

Section XV. REGULATORY INFORMATION

OSHA Hazards Flammable liquid, Target Organ Effect, Toxic by inhalation., Toxic by ingestion, Toxic by skin absorption, Irritant
SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.



Certified Reference Material CRM

Dec 09/17/24

2 vial

ANAB ISO 17034 Accredited
AR-1539 Certificate Number
https://Absolutestandards.com

CERTIFIED WEIGHT REPORT

Part Number: **91980**
Lot Number: **091424**
Description: **Acrolein**Solvent(s):
Water Lot#
072324QExpiration Date: 101424
Recommended Storage: Refrigerate (4 °C)
Nominal Concentration (µg/mL): 5000
NIST Test ID#: 6UTBWeight(s) shown below were combined and diluted to (mL): 10.0
5E-05 Balance Uncertainty
0.001 Flask Uncertainty

<i>Justin Dippold</i>		091424
Formulated By:	Justin Dippold	DATE
<i>Pedro L. Rentas</i>		091424
Reviewed By:	Pedro L. Rentas	DATE

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. Acrolein	5	103755V10F	5000	97	0.5	0.05166	0.05175	5008.9	52.5	107-02-8	0.1 ppm	ori-rat 46mg/kg

Method: GC6MSD-1. Detector: Mass Selective Detector (Scan mode). Column: Vocol (60m X 0.25mm ID X 1.5µm film thickness). Oven Profile: Temp. 1 = 35°C (Time 1 = 10min.), Temp. 2=200°C (Time 2 = 8.75 min.), Rate = 4°C/min., Injector Temp. = 200°C, Detector Temp. = 220°C. Analyst: Pedro Rentas. NOTE: Due to the instability of acrolein in solution, all solutions of acrolein, and any dilutions thereof, should be used immediately. Long term storage is not recommended. Please contact our technical department if further information is required.

Abundance TIC: [BSB2]79005.D

Abundance

Scan 232 (8.927 min): [BSB2]79005.D

250000 8.93

200000

150000

100000

50000

Time--> 0

10.00 15.00 20.00 25.00 30.00 35.00 40.00 45.00 50.00 55.00 60.00



m/z--> 0

20 30 40 50 60 70 80 90 100 110 120 130 140 150 160 170

- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).



Certified Reference Material CRM

Dec 09/17/24

2 vial

ANAB ISO 17034 Accredited
AR-1539 Certificate Number
https://Absolutestandards.com

CERTIFIED WEIGHT REPORT

Part Number: **91980**
Lot Number: **091424**
Description: **Acrolein**Solvent(s):
Water Lot#
072324QExpiration Date: 101424
Recommended Storage: Refrigerate (4 °C)
Nominal Concentration (µg/mL): 5000
NIST Test ID#: 6UTBWeight(s) shown below were combined and diluted to (mL): 10.0
5E-05 Balance Uncertainty
0.001 Flask Uncertainty

<i>Justin Dippold</i>		091424
Formulated By:	Justin Dippold	DATE
<i>Pedro L. Rentas</i>		091424
Reviewed By:	Pedro L. Rentas	DATE

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. Acrolein	5	103755V10F	5000	97	0.5	0.05166	0.05175	5008.9	52.5	107-02-8	0.1 ppm	ori-rat 46mg/kg

Method: GC6MSD-1. Detector: Mass Selective Detector (Scan mode). Column: Vocol (60m X 0.25mm ID X 1.5µm film thickness). Oven Profile: Temp. 1 = 35°C (Time 1 = 10min.), Temp. 2=200°C (Time 2 = 8.75 min.), Rate = 4°C/min., Injector Temp. = 200°C, Detector Temp. = 220°C. Analyst: Pedro Rentas. NOTE: Due to the instability of acrolein in solution, all solutions of acrolein, and any dilutions thereof, should be used immediately. Long term storage is not recommended. Please contact our technical department if further information is required.

Abundance TIC: [BSB2]79005.D

Abundance Scan 232 (8.927 min): [BSB2]79005.D

250000 8.93

200000

150000

100000

50000

Time--> 0

10.00 15.00 20.00 25.00 30.00 35.00 40.00 45.00 50.00 55.00 60.00



m/z--> 0

20 30 40 50 60 70 80 90 100 110 120 130 140 150 160 170

27

60000

50000

40000

30000

20000

10000

37

44

65 75 85

119

158 169

- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).



Certified Reference Material CRM



CERTIFIED WEIGHT REPORT

Part Number: **95318**
Lot Number: **120524**
Description: **2-Chloroethyl vinyl ether**

Solvent(s):
Methanol

Lot#
EJ143-US

Expiration Date: **120527**
Recommended Storage: **Refrigerate (4 °C)**
Nominal Concentration (µg/mL): **10000**
NIST Test ID#: **6UTB**

See 1216124
20 vial
✓ 14630 to
✓ 14649



Weight(s) shown below were combined and diluted to (mL): **50.0**

5E-05 Balance Uncertainty
0.001 Flask Uncertainty

<i>Prashant Chauhan</i>		120524
Formulated By:	Prashant Chauhan	DATE
<i>Pedro L. Rentas</i>		120524
Reviewed By:	Pedro L. Rentas	DATE

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight (g)	Actual Weight (g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. 2-Chloroethyl vinyl ether	74	MKCD0033	10000	99	0.2	0.50536	0.50550	10002.9	40.5	110-75-8	N/A	ori-rat 250mg/kg

Method: GC6MSD-1.M. Detector: MSD. Column: (60m X 0.25mm X 1.5 µm). Oven Profile: Temp 1 = 35°C (Time 1=10min.), Temp 2 = 200°C (Time 2=8.75 min.), Rate = 4°C/min., Injector B Temp = 200°C, Detector B Temp = 220°C. Analyst: Candice Warren.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY ANALYTICAL STANDARD DISSOLVED IN METHANOL

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	1-800-535-5053
Address	44 Rosotto Dr. Hamden CT. 06514	Emergency Telephone International Date Prepared/Revised	1-352-323-3500 January 1, 2024

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

H225	Highly Flammable Liquid and Vapor	H301, 311, 331	Toxic if swallowed, skin contact, inhaled
H370	Cause damage to organs	H351	Suspected of causing cancer
P271	Use in ventilated area	P280	Use gloves, eye protection/face shield
P302,332	If on skin, wash with soap and water	P305,351,338	If in eyes, remove contacts, rinse with water



Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s))	% (optional)
Methanol METHYL ALCOHOL CAS#: 67-56-1	> 97

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance. Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Flammability	Flammable in the presence of a source of ignition when the temperature is above the flash point. Keep away from heat/sparks/open flame/hot surface. No smoking.
Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist. Use ventilation Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge.
Storage Conditions	Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Methanol	67-56-1 TWA 200 ppm		
Skin notation	TWA 200 ppm		
Potential for skin absorption, ingestion and inhalation.			
Personal protective equipment	Respiratory protection	Handle with gloves. Gloves must be inspected prior to use.	Eye protection.
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.			

Section IX - Physical/Chemical Characteristics

Boiling Point	65°C	Specific Gravity (H ₂ O = 1)	0.79
Vapor Pressure (mm Hg)	96	Melting Point	-98°C
Vapor Density (AIR = 1)	1.11	Evaporation rate (Butyl Acetate = 1)	4.6
Solubility in Water	COMPLETE		
Appearance and Odor	CLEAR, COLORLESS LIQUID WITH CHARACTERISTIC PUNGENT ODOR.		

Section X. STABILITY AND REACTIVITY

Chemical stability	Stable under recommended storage conditions.
Possibility of hazardous reactions	Vapours may form explosive mixture with air.
Conditions to avoid	Heat, flames, sparks, extreme temperature and sunlight.
Materials to avoid	Acid chlorides, Acid anhydrides, Oxidizing agents, Alkali metals, Reducing agents, Acids
Hazardous decomposition products formed under fire conditions.	- Carbon oxides

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - rat - 5,628 mg/kg
LC50 Inhalation - rat - 4 h - 64000 ppm
LD50 Dermal - rabbit - 15,800 mg/kg
Toxic if absorbed through skin. Causes skin irritation.
Eye damage/eye irritation
Toxic if inhaled. Causes respiratory tract irritation.
Toxic if swallowed.

Section XII. ECOLOGICAL INFORMATION FOR REPORTABLE QUANTITY OF 5000 lbs.

LC50 15,400 mg/l - 96 h
EC50 24,500.00 mg/l - 48 h
EC100 10,000.00 mg/l - 24 h

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)	IATA
UN number: 1230 Class: 3 Packing group: II	UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol	Proper shipping name: Methanol

Section XV. REGULATORY INFORMATION

OSHA Hazards Flammable liquid, Target Organ Effect, Toxic by inhalation., Toxic by ingestion, Toxic by skin absorption, Irritant
SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.



Certified Reference Material CRM



CERTIFIED WEIGHT REPORT

Part Number: **95318**
Lot Number: **120524**
Description: **2-Chloroethyl vinyl ether**

Solvent(s):
Methanol

Lot#
EJ143-US

Expiration Date: **120527**
Recommended Storage: **Refrigerate (4 °C)**
Nominal Concentration (µg/mL): **10000**
NIST Test ID#: **6UTB**

See 1216124
20 vial
✓ 14630 to
✓ 14649



Weight(s) shown below were combined and diluted to (mL): **50.0**

5E-05 Balance Uncertainty
0.001 Flask Uncertainty

<i>Prashant Chauhan</i>		120524
Formulated By:	Prashant Chauhan	DATE
<i>Pedro L. Rentas</i>		120524
Reviewed By:	Pedro L. Rentas	DATE

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight (g)	Actual Weight (g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. 2-Chloroethyl vinyl ether	74	MKCD0033	10000	99	0.2	0.50536	0.50550	10002.9	40.5	110-75-8	N/A	ori-rat 250mg/kg

Method: GC6MSD-1.M. Detector: MSD. Column: (60m X 0.25mm X 1.5 µm). Oven Profile: Temp 1 = 35°C (Time 1=10min.), Temp 2 = 200°C (Time 2=8.75 min.), Rate = 4°C/min., Injector B Temp = 200°C, Detector B Temp = 220°C. Analyst: Candice Warren.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY ANALYTICAL STANDARD DISSOLVED IN METHANOL

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	1-800-535-5053
Address	44 Rosotto Dr. Hamden CT. 06514	Emergency Telephone International Date Prepared/Revised	1-352-323-3500 January 1, 2024

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

H225	Highly Flammable Liquid and Vapor	H301, 311, 331	Toxic if swallowed, skin contact, inhaled
H370	Cause damage to organs	H351	Suspected of causing cancer
P271	Use in ventilated area	P280	Use gloves, eye protection/face shield
P302, 332	If on skin, wash with soap and water	P305, 351, 338	If in eyes, remove contacts, rinse with water



Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s))	CAS#	% (optional)
Methanol	METHYL ALCOHOL	CAS#: 67-56-1
		> 97

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance. Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Flammability	Flammable in the presence of a source of ignition when the temperature is above the flash point. Keep away from heat/sparks/open flame/hot surface. No smoking.
Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist. Use ventilation. Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge.
Storage Conditions	Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Methanol	67-56-1 TWA 200 ppm		
Skin notation	TWA 200 ppm		
Potential for skin absorption, ingestion and inhalation.			
Personal protective equipment	Respiratory protection	Handle with gloves. Gloves must be inspected prior to use.	Eye protection.
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.			

Section IX - Physical/Chemical Characteristics

Boiling Point	65°C	Specific Gravity (H ₂ O = 1)	0.79
Vapor Pressure (mm Hg)	96	Melting Point	-98°C
Vapor Density (AIR = 1)	1.11	Evaporation rate (Butyl Acetate = 1)	4.6
Solubility in Water	COMPLETE		
Appearance and Odor	CLEAR, COLORLESS LIQUID WITH CHARACTERISTIC PUNGENT ODOR.		

Section X. STABILITY AND REACTIVITY

Chemical stability	Stable under recommended storage conditions.
Possibility of hazardous reactions	Vapours may form explosive mixture with air.
Conditions to avoid	Heat, flames, sparks, extreme temperature and sunlight.
Materials to avoid	Acid chlorides, Acid anhydrides, Oxidizing agents, Alkali metals, Reducing agents, Acids
Hazardous decomposition products formed under fire conditions.	- Carbon oxides

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - rat - 5,628 mg/kg
LC50 Inhalation - rat - 4 h - 64000 ppm
LD50 Dermal - rabbit - 15,800 mg/kg
Toxic if absorbed through skin. Causes skin irritation.
Eye damage/eye irritation
Toxic if inhaled. Causes respiratory tract irritation.
Toxic if swallowed.

Section XII. ECOLOGICAL INFORMATION FOR REPORTABLE QUANTITY OF 5000 lbs.

LC50 15,400 mg/l - 96 h
EC50 24,500.00 mg/l - 48 h
EC100 10,000.00 mg/l - 24 h

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)	IATA
UN number: 1230 Class: 3 Packing group: II	UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol	Proper shipping name: Methanol

Section XV. REGULATORY INFORMATION

OSHA Hazards Flammable liquid, Target Organ Effect, Toxic by inhalation., Toxic by ingestion, Toxic by skin absorption, Irritant
SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.



Certified Reference Material CRM



CERTIFIED WEIGHT REPORT

Part Number: **95318**
Lot Number: **120524**
Description: **2-Chloroethyl vinyl ether**

Solvent(s):
Methanol

Lot#
EJ143-US

Expiration Date: **120527**
Recommended Storage: **Refrigerate (4 °C)**
Nominal Concentration (µg/mL): **10000**
NIST Test ID#: **6UTB**

See 1216124
20 vial
✓ 14630 to
✓ 14649



Weight(s) shown below were combined and diluted to (mL): **50.0**

5E-05 Balance Uncertainty
0.001 Flask Uncertainty

<i>Prashant Chauhan</i>		120524
Formulated By:	Prashant Chauhan	DATE
<i>Pedro L. Rentas</i>		120524
Reviewed By:	Pedro L. Rentas	DATE

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight (g)	Actual Weight (g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. 2-Chloroethyl vinyl ether	74	MKCD0033	10000	99	0.2	0.50536	0.50550	10002.9	40.5	110-75-8	N/A	ori-rat 250mg/kg

Method: GC6MSD-1.M. Detector: MSD. Column: (60m X 0.25mm X 1.5 µm). Oven Profile: Temp 1 = 35°C (Time 1=10min.), Temp 2 = 200°C (Time 2=8.75 min.), Rate = 4°C/min., Injector B Temp = 200°C, Detector B Temp = 220°C. Analyst: Candice Warren.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Safety Data Sheet (SDS)

GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY ANALYTICAL STANDARD DISSOLVED IN METHANOL

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	1-800-535-5053
Address	44 Rossotto Dr.	Emergency Telephone International	1-352-323-3500
	Hamden CT, 06514	Date Prepared/Revised	January 1, 2024

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

H225	Highly Flammable Liquid and Vapor	H301, 311, 331	Toxic if swallowed, skin contact, inhaled
H370	Cause damage to organs	H351	Suspected of causing cancer
P271	Use in ventilated area	P280	Use gloves, eye protection/face shield
P302,332	If on skin, wash with soap and water	P305,351,338	If in eyes, remove contacts, rinse with water



Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s))		% (optional)
Methanol	METHYL ALCOHOL	> 97

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance. Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Flammability	Flammable in the presence of a source of ignition when the temperature is above the flash point. Keep away from heat/sparks/open flame/hot surface. No smoking.
Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist. Use ventilation. Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge.
Storage Conditions	Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Methanol	67-56-1	67-56-1	TWA 200 ppm
Skin notation			TWA 200 ppm
Potential for skin absorption, ingestion and inhalation.			
Personal protective equipment	Respiratory protection	Handle with gloves. Gloves must be inspected prior to use.	Eye protection.
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.			

Section IX - Physical/Chemical Characteristics

Boiling Point	65°C	Specific Gravity (H ₂ O = 1)	0.79
Vapor Pressure (mm Hg)	96	Melting Point	-98°C
Vapor Density (AIR = 1)	1.11	Evaporation rate (Butyl Acetate = 1)	4.6
Solubility in Water	COMPLETE		
Appearance and Odor	CLEAR, COLORLESS LIQUID WITH CHARACTERISTIC PUNGENT ODOR.		

Section X. STABILITY AND REACTIVITY

Chemical stability	Stable under recommended storage conditions.
Possibility of hazardous reactions	Vapours may form explosive mixture with air.
Conditions to avoid	Heat, flames, sparks, extreme temperature and sunlight.
Materials to avoid	Acid chlorides, Acid anhydrides, Oxidizing agents, Alkali metals, Reducing agents, Acids
Hazardous decomposition products formed under fire conditions.	- Carbon oxides

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - rat - 5,628 mg/kg
LC50 Inhalation - rat - 4 h - 64000 ppm
LD50 Dermal - rabbit - 15,800 mg/kg
Toxic if absorbed through skin. Causes skin irritation.
Eye damage/eye irritation
Toxic if inhaled. Causes respiratory tract irritation.
Toxic if swallowed.

Section XII. ECOLOGICAL INFORMATION FOR REPORTABLE QUANTITY OF 5000 lbs.

LC50 15,400 mg/l - 96 h
EC50 24,500.00 mg/l - 48 h
EC100 10,000.00 mg/l - 24 h

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)	IATA
UN number: 1230 Class: 3 Packing group: II	UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol	Proper shipping name: Methanol

Section XV. REGULATORY INFORMATION

OSHA Hazards Flammable liquid, Target Organ Effect, Toxic by inhalation., Toxic by ingestion, Toxic by skin absorption, Irritant
SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.



CERTIFIED WEIGHT REPORT

Part Number: **95318**
Lot Number: **120524**
Description: **2-Chloroethyl vinyl ether**

Solvent(s):
Methanol

Lot#
EJ143-US

Expiration Date: **120527**
Recommended Storage: **Refrigerate (4 °C)**
Nominal Concentration (µg/mL): **10000**
NIST Test ID#: **6UTB**

See 1216124
20 vial
✓ 14630 to
✓ 14649



Weight(s) shown below were combined and diluted to (mL): **50.0**

5E-05 Balance Uncertainty
0.001 Flask Uncertainty

<i>Prashant Chauhan</i>		120524
Formulated By:	Prashant Chauhan	DATE
<i>Pedro L. Rentas</i>		120524
Reviewed By:	Pedro L. Rentas	DATE

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight (g)	Actual Weight (g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. 2-Chloroethyl vinyl ether	74	MKCD0033	10000	99	0.2	0.50536	0.50550	10002.9	40.5	110-75-8	N/A	ori-rat 250mg/kg

Method: GC6MSD-1.M. Detector: MSD. Column: (60m X 0.25mm X 1.5 µm). Oven Profile: Temp 1 = 35°C (Time 1=10min.), Temp 2 = 200°C (Time 2=8.75 min.), Rate = 4°C/min., Injector B Temp = 200°C, Detector B Temp = 220°C. Analyst: Candice Warren.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Safety Data Sheet (SDS)

GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY ANALYTICAL STANDARD DISSOLVED IN METHANOL

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	1-800-535-5053
Address	44 Rossotto Dr.	Emergency Telephone International	1-352-323-3500
	Hamden CT, 06514	Date Prepared/Revised	January 1, 2024

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

H225	Highly Flammable Liquid and Vapor	H301, 311, 331	Toxic if swallowed, skin contact, inhaled
H370	Cause damage to organs	H351	Suspected of causing cancer
P271	Use in ventilated area	P280	Use gloves, eye protection/face shield
P302,332	If on skin, wash with soap and water	P305,351,338	If in eyes, remove contacts, rinse with water



Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s))		% (optional)
Methanol	METHYL ALCOHOL	> 97

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance. Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Flammability	Flammable in the presence of a source of ignition when the temperature is above the flash point. Keep away from heat/sparks/open flame/hot surface. No smoking.
Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist. Use ventilation. Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge.
Storage Conditions	Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Methanol	67-56-1	67-56-1	TWA 200 ppm
Skin notation			TWA 200 ppm
Potential for skin absorption, ingestion and inhalation.			
Personal protective equipment	Respiratory protection	Handle with gloves. Gloves must be inspected prior to use.	Eye protection.
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.			

Section IX - Physical/Chemical Characteristics

Boiling Point	65°C	Specific Gravity (H ₂ O = 1)	0.79
Vapor Pressure (mm Hg)	96	Melting Point	-98°C
Vapor Density (AIR = 1)	1.11	Evaporation rate (Butyl Acetate = 1)	4.6
Solubility in Water	COMPLETE		
Appearance and Odor	CLEAR, COLORLESS LIQUID WITH CHARACTERISTIC PUNGENT ODOR.		

Section X. STABILITY AND REACTIVITY

Chemical stability	Stable under recommended storage conditions.
Possibility of hazardous reactions	Vapours may form explosive mixture with air.
Conditions to avoid	Heat, flames, sparks, extreme temperature and sunlight.
Materials to avoid	Acid chlorides, Acid anhydrides, Oxidizing agents, Alkali metals, Reducing agents, Acids
Hazardous decomposition products formed under fire conditions.	- Carbon oxides

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - rat - 5,628 mg/kg
LC50 Inhalation - rat - 4 h - 64000 ppm
LD50 Dermal - rabbit - 15,800 mg/kg
Toxic if absorbed through skin. Causes skin irritation.
Eye damage/eye irritation
Toxic if inhaled. Causes respiratory tract irritation.
Toxic if swallowed.

Section XII. ECOLOGICAL INFORMATION FOR REPORTABLE QUANTITY OF 5000 lbs.

LC50 15,400 mg/l - 96 h
EC50 24,500.00 mg/l - 48 h
EC100 10,000.00 mg/l - 24 h

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)	IATA
UN number: 1230 Class: 3 Packing group: II	UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol	Proper shipping name: Methanol

Section XV. REGULATORY INFORMATION

OSHA Hazards Flammable liquid, Target Organ Effect, Toxic by inhalation., Toxic by ingestion, Toxic by skin absorption, Irritant
SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30067 **Lot No.:** A0191805

Description : 4-Bromofluorobenzene Standard

4-Bromofluorobenzene Standard 2,500µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : November 30, 2027 **Storage:** 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1-Bromo-4-fluorobenzene (BFB)	460-00-4	184975	99%	2,483.9 µg/mL	+/- 139.5488

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:


40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Alicia Leathers - Operation Technician I

Date Mixed: 17-Nov-2022

Balance Serial # B251644995


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 21-Nov-2022

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555582 **Lot No.:** A0196865

Description : Custom 8260A/B Surrogate Mix
Custom 8260A/B Surrogate Mix 25,000µg/mL, P&T Methanol,
1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : April 30, 2026 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,2-Dichloroethane-d4	17060-07-0	PR-32845	99%	25,036.0 µg/mL	+/- 1,417.9179
2	1-Bromo-4-fluorobenzene (BFB)	460-00-4	184975	99%	25,132.0 µg/mL	+/- 1,423.3549
3	Dibromofluoromethane	1868-53-7	022013	99%	25,040.0 µg/mL	+/- 1,418.1445
4	Toluene-d8	2037-26-5	PR-33397	99%	25,028.0 µg/mL	+/- 1,417.4648

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Russ Bookhamer - Operations Technician I

Date Mixed: 11-Apr-2023

Balance: 1127510105

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30489 **Lot No.:** A0209618

Description : 8260B Acetates Mix

8260B Acetates Mix 2,000 µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : September 30, 2025 **Storage:** -20°C or colder

Handling: This product is photosensitive. **Ship:** On Ice

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Methyl acetate	79-20-9	SHBP3100	99%	2,019.3 µg/mL	+/- 69.7974
2	Vinyl acetate	108-05-4	RP231030CTH	98%	2,016.8 µg/mL	+/- 69.7112
3	Ethyl acetate	141-78-6	SHBQ9682	99%	2,010.7 µg/mL	+/- 69.4979
4	Isopropyl acetate	108-21-4	BCCG7069	99%	2,016.0 µg/mL	+/- 69.6822
5	Propyl acetate	109-60-4	P8XLN	99%	2,008.0 µg/mL	+/- 69.4057
6	Butyl acetate	123-86-4	SHBP6314	99%	2,007.3 µg/mL	+/- 69.3826
7	Amyl acetate	628-63-7	41325/1	97%	2,004.7 µg/mL	+/- 69.2905

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Tech Tips:

Vinyl acetate is a volatile organic ester included in the target lists of several US EPA and other methods. Under acidic conditions, esters react with alcohols to form new esters (transesterification). Methanol-based mixes containing halogenated compounds are slightly acidic, so it is important to minimize exposure of vinyl acetate to mixes of halogenated compounds in methanol. For this

reason, we offer vinyl acetate in individual solution, and suggest that it be introduced into the working level calibration solution immediately before use. This will minimize problems and ensure more consistent results.

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Sam Moodler
Sam Moodler - Operations Tech I

Date Mixed: 28-Mar-2024

Balance Serial # B707717271

Dylan Murphy
Dylan Murphy - Operations Technician I

Date Passed: 01-Apr-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30489 **Lot No.:** A0209618

Description : 8260B Acetates Mix
8260B Acetates Mix 2,000 µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : September 30, 2025 **Storage:** -20°C or colder

Handling: This product is photosensitive. **Ship:** On Ice

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Methyl acetate	79-20-9	SHBP3100	99%	2,019.3 µg/mL	+/- 69.7974
2	Vinyl acetate	108-05-4	RP231030CTH	98%	2,016.8 µg/mL	+/- 69.7112
3	Ethyl acetate	141-78-6	SHBQ9682	99%	2,010.7 µg/mL	+/- 69.4979
4	Isopropyl acetate	108-21-4	BCCG7069	99%	2,016.0 µg/mL	+/- 69.6822
5	Propyl acetate	109-60-4	P8XLN	99%	2,008.0 µg/mL	+/- 69.4057
6	Butyl acetate	123-86-4	SHBP6314	99%	2,007.3 µg/mL	+/- 69.3826
7	Amyl acetate	628-63-7	41325/1	97%	2,004.7 µg/mL	+/- 69.2905

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Tech Tips:

Vinyl acetate is a volatile organic ester included in the target lists of several US EPA and other methods. Under acidic conditions, esters react with alcohols to form new esters (transesterification). Methanol-based mixes containing halogenated compounds are slightly acidic, so it is important to minimize exposure of vinyl acetate to mixes of halogenated compounds in methanol. For this

reason, we offer vinyl acetate in individual solution, and suggest that it be introduced into the working level calibration solution immediately before use. This will minimize problems and ensure more consistent results.

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Sam Moodler
Sam Moodler - Operations Tech I

Date Mixed: 28-Mar-2024

Balance Serial # B707717271

Dillon Murphy
Dillon Murphy - Operations Technician I

Date Passed: 01-Apr-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555581 **Lot No.:** A0210184

Description : Custom 8260 Internal Standard Mix

Custom 8260 Internal Standard Mix 25,000µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL


Expiration Date : April 30, 2027 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,4-Dichlorobenzene-d4	3855-82-1	PR-30447	99%	25,212.0 µg/mL	+/- 1,427.8857
2	1,4-Difluorobenzene	540-36-3	MKCS8657	99%	25,220.0 µg/mL	+/- 1,428.3388
3	Chlorobenzene-d5	3114-55-4	PR-31132	99%	25,116.0 µg/mL	+/- 1,422.4487
4	Pentafluorobenzene	363-72-4	MKCR9383	99%	25,180.0 µg/mL	+/- 1,426.0734

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%


John Friedline - Operations Technician I

Date Mixed: 11-Apr-2024

Balance: 1127510105

REVISED
By Restek Review of SDS per April 18, 2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

Dec 12/17/24
30 v. 4
CERTIFIED REFERENCE MATERIAL

Certificate of Analysis
chromatographic plus
V14697-to-14726



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30006 **Lot No.:** A0210618

Description : VOA Calibration Mix #1

VOA Calibration Mix #1 5,000µg/mL, P&T Methanol/Water(90:10), 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2027 **Storage:** 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Acetone	67-64-1	SHBQ8504	99%	5,014.8 µg/mL	+/- 173.2883
2	2-Butanone (MEK)	78-93-3	SHBQ4704	99%	5,012.4 µg/mL	+/- 173.2054
3	4-Methyl-2-pentanone (MIBK)	108-10-1	SHBP9200	99%	5,011.6 µg/mL	+/- 173.1777
4	2-Hexanone	591-78-6	MKCQ6663	99%	5,013.0 µg/mL	+/- 173.2261

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol/Water (90:10)
CAS # 67-56-1/7732-18-5
Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Dakota Parson - Operations Technician I

Date Mixed: 22-Apr-2024

Balance Serial # B707717271


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 24-Apr-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

Dec 12/17/24
30 v. 4
CERTIFIED REFERENCE MATERIAL

Certificate of Analysis
chromatographic plus
V14697-to-14726



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30006 **Lot No.:** A0210618
Description : VOA Calibration Mix #1
VOA Calibration Mix #1 5,000µg/mL, P&T Methanol/Water(90:10), 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : July 31, 2027 **Storage:** 0°C or colder
Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Acetone	67-64-1	SHBQ8504	99%	5,014.8 µg/mL	+/- 173.2883
2	2-Butanone (MEK)	78-93-3	SHBQ4704	99%	5,012.4 µg/mL	+/- 173.2054
3	4-Methyl-2-pentanone (MIBK)	108-10-1	SHBP9200	99%	5,011.6 µg/mL	+/- 173.1777
4	2-Hexanone	591-78-6	MKCQ6663	99%	5,013.0 µg/mL	+/- 173.2261

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol/Water (90:10)
CAS # 67-56-1/7732-18-5
Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Dakota Parson - Operations Technician I

Date Mixed: 22-Apr-2024

Balance Serial # B707717271


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 24-Apr-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

Dec 12/17/24
30 v. 4
CERTIFIED REFERENCE MATERIAL

Certificate of Analysis
chromatographic plus
V14697-to-14726



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30006 **Lot No.:** A0210618
Description : VOA Calibration Mix #1
VOA Calibration Mix #1 5,000µg/mL, P&T Methanol/Water(90:10), 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : July 31, 2027 **Storage:** 0°C or colder
Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Acetone	67-64-1	SHBQ8504	99%	5,014.8 µg/mL	+/- 173.2883
2	2-Butanone (MEK)	78-93-3	SHBQ4704	99%	5,012.4 µg/mL	+/- 173.2054
3	4-Methyl-2-pentanone (MIBK)	108-10-1	SHBP9200	99%	5,011.6 µg/mL	+/- 173.1777
4	2-Hexanone	591-78-6	MKCQ6663	99%	5,013.0 µg/mL	+/- 173.2261

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol/Water (90:10)
CAS # 67-56-1/7732-18-5
Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Dakota Parson - Operations Technician I

Date Mixed: 22-Apr-2024

Balance Serial # B707717271


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 24-Apr-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

Dec 12/17/24
30 v. 4
CERTIFIED REFERENCE MATERIAL

Certificate of Analysis
chromatographic plus
V14697-to-14726



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30006 **Lot No.:** A0210618
Description : VOA Calibration Mix #1
VOA Calibration Mix #1 5,000µg/mL, P&T Methanol/Water(90:10), 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : July 31, 2027 **Storage:** 0°C or colder
Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Acetone	67-64-1	SHBQ8504	99%	5,014.8 µg/mL	+/- 173.2883
2	2-Butanone (MEK)	78-93-3	SHBQ4704	99%	5,012.4 µg/mL	+/- 173.2054
3	4-Methyl-2-pentanone (MIBK)	108-10-1	SHBP9200	99%	5,011.6 µg/mL	+/- 173.1777
4	2-Hexanone	591-78-6	MKCQ6663	99%	5,013.0 µg/mL	+/- 173.2261

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol/Water (90:10)
CAS # 67-56-1/7732-18-5
Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Dakota Parson - Operations Technician I

Date Mixed: 22-Apr-2024

Balance Serial # B707717271


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 24-Apr-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

10 vial.
CERTIFIED REFERENCE MATERIAL

Dec: 12/09/24

Certificate of Analysis

chromatographic plus

V14667-5
V14676



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30225 **Lot No.:** A0214960
Description : Bromochloromethane Standard
Bromochloromethane 2000µg/mL, P&T Methanol, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : August 31, 2029 **Storage:** 0°C or colder
Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Bromochloromethane	74-97-5	SYN240416CTH	99%	2,012.0 µg/mL	+/- 113.0519

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Stacey Wanner - Operations Technician I

Date Mixed: 08-Aug-2024

Balance Serial # 1127510105

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 14-Aug-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/ μ ECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

10 vial.
Dec: 12/09/24
CERTIFIED REFERENCE MATERIAL

Certificate of Analysis
chromatographic plus

V14667-5
V14676



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30225 **Lot No.:** A0214960
Description : Bromochloromethane Standard
Bromochloromethane 2000µg/mL, P&T Methanol, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : August 31, 2029 **Storage:** 0°C or colder
Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Bromochloromethane	74-97-5	SYN240416CTH	99%	2,012.0 µg/mL	+/- 113.0519

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Stacey Wanner - Operations Technician I

Date Mixed: 08-Aug-2024

Balance Serial # 1127510105

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 14-Aug-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/ μ ECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

10 vial.
CERTIFIED REFERENCE MATERIAL

Dec: 12/09/24

Certificate of Analysis

chromatographic plus

V14667-5

V14676



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30225 **Lot No.:** A0214960

Description : Bromochloromethane Standard

Bromochloromethane 2000µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : August 31, 2029 **Storage:** 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Bromochloromethane	74-97-5	SYN240416CTH	99%	2,012.0 µg/mL	+/- 113.0519

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol

CAS # 67-56-1

Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Stacey Wanner - Operations Technician I

Date Mixed: 08-Aug-2024

Balance Serial # 1127510105

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 14-Aug-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/ μ ECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

10 vial.
Dec: 12/09/24
CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus
V14667-5
V14676



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30225 **Lot No.:** A0214960
Description : Bromochloromethane Standard
Bromochloromethane 2000µg/mL, P&T Methanol, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : August 31, 2029 **Storage:** 0°C or colder
Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Bromochloromethane	74-97-5	SYN240416CTH	99%	2,012.0 µg/mL	+/- 113.0519

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Stacey Wanner - Operations Technician I

Date Mixed: 08-Aug-2024

Balance Serial # 1127510105

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 14-Aug-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/ μ ECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

Rec 12/17/24
30 ml
CERTIFIED REFERENCE MATERIAL

Certificate of Analysis
chromatographic plus

V14727 to
V14756



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30042 **Lot No.:** A0216826

Description : 502.2 Calibration Mix #1
502.2 Calibration Mix #1 2,000µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : May 31, 2031 **Storage:** 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Dichlorodifluoromethane (CFC-12)	75-71-8	00022922	99%	2,000.9 µg/mL	+/- 112.4144
2	Chloromethane (methyl chloride)	74-87-3	00022694	99%	2,000.7 µg/mL	+/- 112.3998
3	Vinyl chloride	75-01-4	00015559	99%	2,000.3 µg/mL	+/- 112.3779
4	Bromomethane (methyl bromide)	74-83-9	00017022	99%	2,001.8 µg/mL	+/- 112.4650
5	Chloroethane (ethyl chloride)	75-00-3	107-401039114-1	99%	2,000.1 µg/mL	+/- 112.3700
6	Trichlorofluoromethane (CFC-11)	75-69-4	MKCJ8658	99%	2,000.7 µg/mL	+/- 112.3992

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:

60m x 0.25mm x 1.4µm
Rtx-502.2 (cat.#10916)

Carrier Gas:

helium-constant flow 2.0 mL/min.

Temp. Program:

40°C (hold 6 min.) to 100°C
@ 6°C/min.

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


MSD

Split Vent:

Split ratio 10:1

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.



Tom Suckal - Mix Technician

Date Mixed: 23-Sep-2024

Balance Serial # B707717271



Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 04-Oct-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

Rec 12/17/24
30 ml
CERTIFIED REFERENCE MATERIAL

Certificate of Analysis
chromatographic plus

V14727 to
V14756



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30042 **Lot No.:** A0216826
Description : 502.2 Calibration Mix #1
502.2 Calibration Mix #1 2,000µg/mL, P&T Methanol, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : May 31, 2031 **Storage:** 0°C or colder
Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Dichlorodifluoromethane (CFC-12)	75-71-8	00022922	99%	2,000.9 µg/mL	+/- 112.4144
2	Chloromethane (methyl chloride)	74-87-3	00022694	99%	2,000.7 µg/mL	+/- 112.3998
3	Vinyl chloride	75-01-4	00015559	99%	2,000.3 µg/mL	+/- 112.3779
4	Bromomethane (methyl bromide)	74-83-9	00017022	99%	2,001.8 µg/mL	+/- 112.4650
5	Chloroethane (ethyl chloride)	75-00-3	107-401039114-1	99%	2,000.1 µg/mL	+/- 112.3700
6	Trichlorofluoromethane (CFC-11)	75-69-4	MKCJ8658	99%	2,000.7 µg/mL	+/- 112.3992

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:

60m x 0.25mm x 1.4µm
Rtx-502.2 (cat.#10916)

Carrier Gas:

helium-constant flow 2.0 mL/min.

Temp. Program:

40°C (hold 6 min.) to 100°C
@ 6°C/min.

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


MSD

Split Vent:

Split ratio 10:1

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.



Tom Suckal - Mix Technician

Date Mixed: 23-Sep-2024

Balance Serial # B707717271



Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 04-Oct-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

✓ 14842 to 14846



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30470 **Lot No.:** A0217535

Description : tert-Butanol Standard

tert-Butanol Std 50,000µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : October 31, 2027 **Storage:** 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	tert-Butanol (TBA)	75-65-0	SHBQ8002-1	99%	50,007.5 µg/mL	+/- 717.6137

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

A. O. E.
Aaron Enyart - Operations Tech I

Date Mixed: 07-Oct-2024

Balance Serial # B251644995

Brittany Federinko
Brittany Federinko - Operations Tech I

Date Passed: 09-Oct-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/ μ ECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

2014 Dec 01/08/21
CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic

V14803 - V14822



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555408-SL **Lot No.:** A0220471
Description : Custom Vinyl Acetate Standard
Custom Vinyl Acetate Standard 8,000µg/mL, P&T Methanol, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : June 30, 2026 **Storage:** -20°C or colder
Handling: This product is photosensitive. **Ship:** On Ice

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Vinyl acetate	108-05-4	RD240423RSR	99%	8,066.0 µg/mL	+/- 278.7979

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Tech Tips:

Vinyl acetate is a volatile organic ester included in the target lists of several US EPA and other methods. Under acidic conditions, esters react with alcohols to form new esters (transesterification). Methanol-based mixes containing halogenated compounds are slightly acidic, so it is important to minimize exposure of vinyl acetate to mixes of halogenated compounds in methanol. For this reason, we offer vinyl acetate in individual solution, and suggest that it be introduced into the working level calibration solution immediately before use. This will minimize problems and ensure more consistent results.

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Ethan Winiarski - Operations Tech I

Date Mixed: 24-Dec-2024

Balance Serial # 1127510105

Dillan Murphy - Operations Technician I

Date Passed: 02-Jan-2025

REVIEWED
By Jennifer Pollock at 7:12 am, Jan 05, 2025

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/ μ ECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

2014 Dec 01/08/21
CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic

V14803 - V14822



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555408-SL **Lot No.:** A0220471
Description : Custom Vinyl Acetate Standard
Custom Vinyl Acetate Standard 8,000µg/mL, P&T Methanol, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : June 30, 2026 **Storage:** -20°C or colder
Handling: This product is photosensitive. **Ship:** On Ice

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Vinyl acetate	108-05-4	RD240423RSR	99%	8,066.0 µg/mL	+/- 278.7979

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Tech Tips:

Vinyl acetate is a volatile organic ester included in the target lists of several US EPA and other methods. Under acidic conditions, esters react with alcohols to form new esters (transesterification). Methanol-based mixes containing halogenated compounds are slightly acidic, so it is important to minimize exposure of vinyl acetate to mixes of halogenated compounds in methanol. For this reason, we offer vinyl acetate in individual solution, and suggest that it be introduced into the working level calibration solution immediately before use. This will minimize problems and ensure more consistent results.

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Ethan Winiarski - Operations Tech I

Date Mixed: 24-Dec-2024

Balance Serial # 1127510105

Dillan Murphy - Operations Technician I

Date Passed: 02-Jan-2025

REVIEWED
By Jennifer Pollock at 7:12 am, Jan 05, 2025

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/ μ ECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

Methanol
ULTRA RESI-ANALYZED
For Purge and Trap Analysis

Rev 05/09/25
Avantor™



V 14921 to
V 14938

Material No.: 9077-02
Batch No.: 24G0262002
Manufactured Date: 2024-05-14
Expiration Date: 2027-05-14
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay (CH ₃ OH) (by GC, corrected for water)	≥ 99.9 %	100.0 %
Residue after Evaporation	≤ 1.0 ppm	0.3 ppm
Titration Acid (μeq/g)	≤ 0.3	0.3
Titration Base (μeq/g)	≤ 0.10	0.03
Water (by KF, coulometric)	≤ 0.08 %	< 0.01 %
Volatile Organic Trace Analysis – Below EPA 8260B CRQL	Conforms	Conforms

For Laboratory, Research, or Manufacturing Use
Performance Tested for Use in EPA Methods
500 Series for Drinking Water
600 Series for Wastewater
846 for Solid Waste

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

Jamie Croak
Director Quality Operations, Bioscience Production

Methanol
ULTRA RESI-ANALYZED
For Purge and Trap Analysis

Rev 05/09/25
Avantor™



V 14921 to
V 14938

Material No.: 9077-02
Batch No.: 24G0262002
Manufactured Date: 2024-05-14
Expiration Date: 2027-05-14
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay (CH ₃ OH) (by GC, corrected for water)	≥ 99.9 %	100.0 %
Residue after Evaporation	≤ 1.0 ppm	0.3 ppm
Titration Acid (μeq/g)	≤ 0.3	0.3
Titration Base (μeq/g)	≤ 0.10	0.03
Water (by KF, coulometric)	≤ 0.08 %	< 0.01 %
Volatile Organic Trace Analysis – Below EPA 8260B CRQL	Conforms	Conforms

For Laboratory, Research, or Manufacturing Use
Performance Tested for Use in EPA Methods
500 Series for Drinking Water
600 Series for Wastewater
846 for Solid Waste

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

Jamie Croak
Director Quality Operations, Bioscience Production



Certified Reference Material CRM

see 05/21/25



CERTIFIED WEIGHT REPORT

Part Number: **91980**
Lot Number: **051925**
Description: **Acrolein**

Solvent(s):
Water Lot# 041725Q

5.019

114944-114948

		051925
Formulated By:	Lawrence Barry	DATE
		051925
Reviewed By:	Pedro L. Rentas	DATE



Expiration Date: 061925
Recommended Storage: Refrigerate (4 °C)
Nominal Concentration (µg/mL): 5000
NIST Test ID#: 6UTB

5E-05 Balance Uncertainty
0.001 Flask Uncertainty

Weight(s) shown below were combined and diluted to (mL): 10.0

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. Acrolein	5	103755V10F	5000	97	0.5	0.05166	0.05170	5004.1	52.5	107-02-8	0.1 ppm	ori-rat 46mg/kg

Method: GC6MSD-1. Detector: Mass Selective Detector (Scan mode). Column: Vocol (60m X 0.25mm ID X 1.5µm film thickness). Oven Profile: Temp. 1 = 35°C (Time 1 = 10min.), Temp. 2=200°C (Time 2 = 8.75 min.) Rate = 4°C/min., Injector Temp. = 200°C, Detector Temp. = 220°C. Analyst: Pedro Rentas. NOTE: Due to the instability of acrolein in solution, all solutions of acrolein, and any dilutions thereof, should be used immediately. Long term storage is not recommended. Please contact our technical department if further information is required.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
 - Standards are prepared gravimetrically using balances that are calibrated by an ISO 17025 certified organization with weights traceable through NIST to the SI kilogram (see above).
 - Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
 - All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
 - Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).
- Rev 1.0, 2/25/2025



Certified Reference Material CRM

see 05/21/25



CERTIFIED WEIGHT REPORT

Part Number: **91980**
Lot Number: **051925**
Description: **Acrolein**

Solvent(s):
Water Lot# 041725Q

5.019

114944-114948

		051925
Formulated By:	Lawrence Barry	DATE
		051925
Reviewed By:	Pedro L. Rentas	DATE



Expiration Date: 061925
Recommended Storage: Refrigerate (4 °C)
Nominal Concentration (µg/mL): 5000
NIST Test ID#: 6UTB

5E-05 Balance Uncertainty
0.001 Flask Uncertainty

Weight(s) shown below were combined and diluted to (mL): 10.0

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. Acrolein	5	103755V10F	5000	97	0.5	0.05166	0.05170	5004.1	52.5	107-02-8	0.1 ppm	ori-rat 46mg/kg

Method: GC6MSD-1. Detector: Mass Selective Detector (Scan mode). Column: Vocol (60m X 0.25mm ID X 1.5µm film thickness). Oven Profile: Temp. 1 = 35°C (Time 1 = 10min.), Temp. 2=200°C (Time 2 = 8.75 min.) Rate = 4°C/min., Injector Temp. = 200°C, Detector Temp. = 220°C. Analyst: Pedro Rentas. NOTE: Due to the instability of acrolein in solution, all solutions of acrolein, and any dilutions thereof, should be used immediately. Long term storage is not recommended. Please contact our technical department if further information is required.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
 - Standards are prepared gravimetrically using balances that are calibrated by an ISO 17025 certified organization with weights traceable through NIST to the SI kilogram (see above).
 - Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
 - All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
 - Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).
- Rev 1.0, 2/25/2025



Certified Reference Material CRM

see 05/21/25



CERTIFIED WEIGHT REPORT

Part Number: 91980
Lot Number: 051925
Description: Acrolein

Solvent(s): Water
Lot#: 041725Q

5.019

114944-114948

		051925
Formulated By:	Lawrence Barry	DATE
		051925
Reviewed By:	Pedro L. Rentas	DATE



Expiration Date: 061925
Recommended Storage: Refrigerate (4 °C)
Nominal Concentration (µg/mL): 5000
NIST Test ID#: 6UTB

5E-05 Balance Uncertainty
0.001 Flask Uncertainty

Weight(s) shown below were combined and diluted to (mL): 10.0

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. Acrolein	5	103755V10F	5000	97	0.5	0.05166	0.05170	5004.1	52.5	107-02-8	0.1 ppm	ori-rat 46mg/kg

Method: GC6MSD-1. Detector: Mass Selective Detector (Scan mode). Column: Vocol (60m X 0.25mm ID X 1.5µm film thickness). Oven Profile: Temp. 1 = 35°C (Time 1 = 10min.), Temp. 2=200°C (Time 2 = 8.75 min.) Rate = 4°C/min., Injector Temp. = 200°C, Detector Temp. = 220°C. Analyst: Pedro Rentas. NOTE: Due to the instability of acrolein in solution, all solutions of acrolein, and any dilutions thereof, should be used immediately. Long term storage is not recommended. Please contact our technical department if further information is required.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
 - Standards are prepared gravimetrically using balances that are calibrated by an ISO 17025 certified organization with weights traceable through NIST to the SI kilogram (see above).
 - Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
 - All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
 - Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).
- Rev 1.0, 2/25/2025



Certified Reference Material CRM

see 05/21/25



CERTIFIED WEIGHT REPORT

Part Number: 91980
Lot Number: 051925
Description: Acrolein

Solvent(s): Water
Lot#: 041725Q

5.019

114944-114948

		051925
Formulated By:	Lawrence Barry	DATE
		051925
Reviewed By:	Pedro L. Rentas	DATE



Expiration Date: 061925
Recommended Storage: Refrigerate (4 °C)
Nominal Concentration (µg/mL): 5000
NIST Test ID#: 6UTB

5E-05 Balance Uncertainty
0.001 Flask Uncertainty

Weight(s) shown below were combined and diluted to (mL): 10.0

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. Acrolein	5	103755V10F	5000	97	0.5	0.05166	0.05170	5004.1	52.5	107-02-8	0.1 ppm	ori-rat 46mg/kg

Method: GC6MSD-1. Detector: Mass Selective Detector (Scan mode). Column: Vocol (60m X 0.25mm ID X 1.5µm film thickness). Oven Profile: Temp. 1 = 35°C (Time 1 = 10min.), Temp. 2=200°C (Time 2 = 8.75 min.) Rate = 4°C/min., Injector Temp. = 200°C, Detector Temp. = 220°C. Analyst: Pedro Rentas. NOTE: Due to the instability of acrolein in solution, all solutions of acrolein, and any dilutions thereof, should be used immediately. Long term storage is not recommended. Please contact our technical department if further information is required.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
 - Standards are prepared gravimetrically using balances that are calibrated by an ISO 17025 certified organization with weights traceable through NIST to the SI kilogram (see above).
 - Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
 - All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
 - Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).
- Rev 1.0, 2/25/2025



SHIPPING DOCUMENTS

CLIENT INFORMATION

REPORT TO BE SENT TO:

COMPANY: Resolution Consultants

ADDRESS:

CITY STATE: ZIP:

ATTENTION: Eleanor Vivandon

PHONE: FAX:

CLIENT PROJECT INFORMATION

PROJECT NAME: NWIRP BETHPAGE

PROJECT NO.: 60731872 LOCATION: BETHPAGE

PROJECT MANAGER: Eleanor Vivandon

e-mail: Eleanor.vivandon@aecom.com

PHONE: FAX:

CLIENT BILLING INFORMATION

BILL TO: Eleanor V. PO#:

ADDRESS: 27 Ellis Place

CITY OSSINING STATE: NY ZIP: 10563

ATTENTION: PHONE:

ANALYSIS

DATA TURNAROUND INFORMATION

FAX (RUSH) DAYS*

HARDCOPY (DATA PACKAGE): Standard DAYS*

EDD: Standard DAYS*

*TO BE APPROVED BY CHEMTECH

STANDARD HARDCOPY TURNAROUND TIME IS 10 BUSINESS

DATA DELIVERABLE INFORMATION

☐ Level 1 (Results Only) ☐ Level 4 (QC + Full Raw Data)

☐ Level 2 (Results + QC) ☐ NJ Reduced ☐ US EPA CLP

☐ Level 3 (Results + QC) ☐ NYS ASP A ☐ NYS ASP B

+ Raw Data) ☐ Other

☐ EDD FORMAT

VOCS P260
2,4 dioxene 827050

PRESERVATIVES

COMMENTS

ALLIANCE SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE		SAMPLE COLLECTION		# OF BOTTLES										← Specify Preservatives A-HCl B-HNO3 C-H2SO4 D-NaOH E-ICE F-OTHER
			COMP	GRAB	DATE	TIME		A	F								
1.	RW9-MW01D3-20250606	GW		X	06/06/25	0550	3	2	1								
2.	RW9-MW01S-20250606	GW		X	↓	1015	3	2	1								
3.																	
4.																	
5.																	
6.																	
7.																	
8.																	
9.																	
10.																	

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER:	DATE/TIME: 1215 06/06/25	RECEIVED BY: 1215 6-6-25	Conditions of bottles or coolers at receipt: ☐ COMPLIANT ☐ NON COMPLIANT ☐ COOLER TEMP 2.1 °C
1. [Signature]		1. [Signature]	Comments:
RELINQUISHED BY SAMPLER:	DATE/TIME:	RECEIVED BY:	
2.		2.	
RELINQUISHED BY SAMPLER:	DATE/TIME: 1642 6-6-25	RECEIVED BY:	
3. [Signature]		3.	

Page 1 of 1

CLIENT: ☐ Hand Delivered ☐ Other

Shipment Complete

☐ YES ☐ NO

Laboratory Certification

Certified By	License No.
CAS EPA CLP Contract	68HERH20D0011
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255424 Rev 1
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	T104704488

LOGIN REPORT/SAMPLE TRANSFER

Order ID : Q2263	AECO15	Order Date : 6/6/2025 12:42:00 PM	Project Mgr :
Client Name : AECOM Technical Services		Project Name : NAVFAC NWIRP Bethpage	Report Type : Results Only NYS ASP B
Client Contact : Eleanor Vivadou		Receive DateTime : 6/6/2025 12:00:00 AM	EDD Type : EQUIS
Invoice Name : AECOM Technical Services		Purchase Order : 1642	Hard Copy Date :
Invoice Contact : Eleanor Vivadou			Date Signoff :

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
Q2263-01	RW9-MW01D3-20250606	Water	06/06/2025	09:50					
					VOCMS Group1		8260-Low	10 Bus. Days	
Q2263-02	RW9-MW01D3-20250606 01S	Water	06/06/2025	10:15					
					VOCMS Group1		8260-Low	10 Bus. Days	

Relinquished By : 
Date / Time : 6/9/25 0845

SAMPLES RECEIVED ON 6/6/25 @ 1642
PLACED IN SM-REF-2

Received By : 
Date / Time : 06/09/25 8.45

Storage Area : VOA Refridgerator Room

Pg 4

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060625\
 Data File : VN086862.D
 Acq On : 06 Jun 2025 12:44
 Operator : JC\MD
 Sample : VSTDICC001
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC001

Quant Time: Jun 07 01:54:26 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 01:51:02 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	212682	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	393457	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	348010	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	167399	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.594	65	3429	1.204	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	2.400%#
35) Dibromofluoromethane	8.171	113	2626	1.126	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	2.260%#
50) Toluene-d8	10.570	98	10645	1.153	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	2.300%#
62) 4-Bromofluorobenzene	12.853	95	4252	1.240	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	2.480%#

Target Compounds				Qvalue		
2) Dichlorodifluoromethane	2.153	85	1985	0.936	ug/l	72
3) Chloromethane	2.400	50	3241	1.184	ug/l	98
4) Vinyl Chloride	2.559	62	2848	1.008	ug/l #	88
5) Bromomethane	3.006	94	1886	1.193	ug/l	93
6) Chloroethane	3.159	64	1957	1.073	ug/l #	79
7) Trichlorofluoromethane	3.530	101	3752	1.017	ug/l	97
8) Diethyl Ether	3.994	74	1583	0.984	ug/l	76
9) 1,1,2-Trichlorotrifluo...	4.394	101	1332	0.574	ug/l #	55
10) Methyl Iodide	4.624	142	3165	1.053	ug/l #	67
11) Tert butyl alcohol	5.547	59	4268	5.524	ug/l #	77
12) 1,1-Dichloroethene	4.365	96	2436	1.029	ug/l	85
13) Acrolein	4.194	56	1467	5.998	ug/l	85
14) Allyl chloride	5.041	41	4189	1.067	ug/l #	75
15) Acrylonitrile	5.735	53	9222	5.106	ug/l	96
16) Acetone	4.459	43	9069	6.006	ug/l	95
17) Carbon Disulfide	4.730	76	7306	1.116	ug/l	97
18) Methyl Acetate	5.053	43	4402	1.000	ug/l #	66
19) Methyl tert-butyl Ether	5.824	73	9019	1.052	ug/l	92
20) Methylene Chloride	5.294	84	3496	1.237	ug/l	93
21) trans-1,2-Dichloroethene	5.800	96	2979	1.131	ug/l	87
22) Diisopropyl ether	6.682	45	8584	1.037	ug/l #	96
23) Vinyl Acetate	6.618	43	37075	5.303	ug/l	95
24) 1,1-Dichloroethane	6.588	63	5071	1.065	ug/l #	81
25) 2-Butanone	7.500	43	12841	5.231	ug/l	93
26) 2,2-Dichloropropane	7.500	77	4112	1.110	ug/l	94
27) cis-1,2-Dichloroethene	7.494	96	3345	1.062	ug/l	95
28) Bromochloromethane	7.824	49	2464	1.052	ug/l #	88
29) Tetrahydrofuran	7.853	42	8301	5.191	ug/l	100
30) Chloroform	7.971	83	5255	1.105	ug/l	97
31) Cyclohexane	8.247	56	9143	1.980	ug/l #	28
32) 1,1,1-Trichloroethane	8.171	97	4378	1.082	ug/l #	50
36) 1,1-Dichloropropene	8.376	75	3678	1.059	ug/l	96
37) Ethyl Acetate	7.577	43	4282	0.968	ug/l #	78
38) Carbon Tetrachloride	8.376	117	3563	1.045	ug/l #	93
39) Methylcyclohexane	9.606	83	4981	1.046	ug/l #	90
40) Benzene	8.612	78	12498	1.100	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060625\
 Data File : VN086862.D
 Acq On : 06 Jun 2025 12:44
 Operator : JC\MD
 Sample : VSTDIC001
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC001

Quant Time: Jun 07 01:54:26 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 01:51:02 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.782	41	2799	1.127	ug/l	89
42) 1,2-Dichloroethane	8.682	62	3725	1.081	ug/l	82
43) Isopropyl Acetate	8.694	43	7673	1.081	ug/l #	96
44) Trichloroethene	9.365	130	2822	1.047	ug/l	71
45) 1,2-Dichloropropane	9.629	63	2883	1.043	ug/l	95
46) Dibromomethane	9.712	93	1835	0.999	ug/l	94
47) Bromodichloromethane	9.888	83	4017	1.063	ug/l #	96
48) Methyl methacrylate	9.688	41	3492	1.069	ug/l	88
49) 1,4-Dioxane	9.706	88	1004	16.891	ug/l #	74
51) 4-Methyl-2-Pentanone	10.453	43	19886	4.653	ug/l	99
52) Toluene	10.635	92	7227	1.041	ug/l	95
53) t-1,3-Dichloropropene	10.841	75	4494	1.064	ug/l	97
54) cis-1,3-Dichloropropene	10.312	75	4792	1.060	ug/l	97
55) 1,1,2-Trichloroethane	11.023	97	2828	1.058	ug/l	89
56) Ethyl methacrylate	10.882	69	3406	0.800	ug/l #	85
57) 1,3-Dichloropropane	11.170	76	5062	1.092	ug/l	98
58) 2-Chloroethyl Vinyl ether	10.165	63	10924	4.304	ug/l	97
59) 2-Hexanone	11.235	43	5016	1.822	ug/l #	35
60) Dibromochloromethane	11.359	129	2843	1.021	ug/l	96
61) 1,2-Dibromoethane	11.470	107	2777	1.014	ug/l	96
64) Tetrachloroethene	11.106	164	2473	1.123	ug/l	94
65) Chlorobenzene	11.894	112	8581	1.118	ug/l #	87
66) 1,1,1,2-Tetrachloroethane	11.959	131	2523	1.023	ug/l #	63
67) Ethyl Benzene	11.970	91	13845	1.047	ug/l	94
68) m/p-Xylenes	12.070	106	10163	2.008	ug/l	95
69) o-Xylene	12.400	106	4970	1.025	ug/l	93
70) Styrene	12.417	104	8189	0.987	ug/l	98
71) Bromoform	12.582	173	1508	0.825	ug/l #	78
73) Isopropylbenzene	12.694	105	12938	1.061	ug/l	98
75) 1,1,2,2-Tetrachloroethane	12.935	83	4326	1.047	ug/l #	93
76) 1,2,3-Trichloropropane	13.000	75	5925	1.484	ug/l	65
77) Bromobenzene	12.982	156	2913	1.042	ug/l	96
78) n-propylbenzene	13.035	91	15168	1.023	ug/l	98
79) 2-Chlorotoluene	13.123	91	9550	1.074	ug/l	97
80) 1,3,5-Trimethylbenzene	13.170	105	10058	0.999	ug/l	96
81) trans-1,4-Dichloro-2-b...	12.741	75	1392	0.805	ug/l	92
82) 4-Chlorotoluene	13.223	91	9706	1.079	ug/l	96
83) tert-Butylbenzene	13.435	119	9849	1.069	ug/l	98
84) 1,2,4-Trimethylbenzene	13.482	105	9882	0.979	ug/l	99
85) sec-Butylbenzene	13.617	105	13775	1.029	ug/l	100
86) p-Isopropyltoluene	13.729	119	11065	1.000	ug/l	97
87) 1,3-Dichlorobenzene	13.735	146	5903	1.073	ug/l	94
88) 1,4-Dichlorobenzene	13.735	146	5903	1.052	ug/l	93
89) n-Butylbenzene	14.058	91	11627	1.085	ug/l	94
90) Hexachloroethane	14.329	117	1942	1.036	ug/l	93
91) 1,2-Dichlorobenzene	14.106	146	5607	1.061	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.717	75	1136	1.150	ug/l	78
93) 1,2,4-Trichlorobenzene	15.388	180	3489	1.034	ug/l	98
94) Hexachlorobutadiene	15.500	225	1219	0.969	ug/l	88
95) Naphthalene	15.635	128	12789	1.018	ug/l	98
96) 1,2,3-Trichlorobenzene	15.835	180	3594	1.072	ug/l	96

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060625\
Data File : VN086862.D
Acq On : 06 Jun 2025 12:44
Operator : JC\MD
Sample : VSTDICC001
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001

Quant Time: Jun 07 01:54:26 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 01:51:02 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#)=qualifier out of range (m)=manual integration (+)=signals summed						

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.235 min Scan# 1068
Delta R.T. -0.001 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion:168 Resp: 212682
Ion Ratio Lower Upper
168 100
99 58.1 49.1 73.7



#2
Dichlorodifluoromethane
Concen: 0.936 ug/l
RT: 2.153 min Scan# 34
Delta R.T. -0.000 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44

Tgt Ion: 85 Resp: 1985
Ion Ratio Lower Upper
85 100
87 48.3 16.2 48.6









#3
Chloromethane
Concen: 1.184 ug/l
RT: 2.400 min Scan# 75
Delta R.T. 0.006 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion: 50 Resp: 3241
Ion Ratio Lower Upper
50 100
52 30.7 25.7 38.5



#4
Vinyl Chloride
Concen: 1.008 ug/l
RT: 2.559 min Scan# 103
Delta R.T. 0.006 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44

Tgt Ion: 62 Resp: 2848
Ion Ratio Lower Upper
62 100
64 25.9 26.0 39.0#









#5
Bromomethane
Concen: 1.193 ug/l
RT: 3.006 min Scan# 11
Delta R.T. 0.012 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001


Tgt Ion: 94 Resp: 1886
Ion Ratio Lower Upper
94 100
96 90.7 77.9 116.9



#6
Chloroethane
Concen: 1.073 ug/l
RT: 3.159 min Scan# 205
Delta R.T. 0.006 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44

Tgt Ion: 64 Resp: 1957
Ion Ratio Lower Upper
64 100
66 42.3 24.6 36.8#









#7
 Trichlorofluoromethane
 Concen: 1.017 ug/l
 RT: 3.530 min Scan# 20
 Delta R.T. 0.006 min
 Lab File: VN086862.D
 Acq: 06 Jun 2025 12:44




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC001


Tgt Ion:101 Resp: 3752
 Ion Ratio Lower Upper
 101 100
 103 67.2 52.2 78.2



#8
 Diethyl Ether
 Concen: 0.984 ug/l
 RT: 3.994 min Scan# 347
 Delta R.T. 0.011 min
 Lab File: VN086862.D
 Acq: 06 Jun 2025 12:44

Tgt Ion: 74 Resp: 1583
 Ion Ratio Lower Upper
 74 100
 45 113.8 45.5 136.5









#9
1,1,2-Trichlorotrifluoroethane
Concen: 0.574 ug/l
RT: 4.394 min Scan# 41
Delta R.T. -0.000 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion:101 Resp: 1332
Ion Ratio Lower Upper
101 100
85 49.2 33.8 50.8
151 127.6 58.7 88.1#



#10
Methyl Iodide
Concen: 1.053 ug/l
RT: 4.624 min Scan# 454
Delta R.T. 0.017 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44

Tgt Ion:142 Resp: 3165
Ion Ratio Lower Upper
142 100
127 18.2 30.9 46.3#
141 0.0 10.4 15.6#









#11
Tert butyl alcohol
Concen: 5.524 ug/l
RT: 5.547 min Scan# 611
Delta R.T. -0.000 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion: 59 Resp: 4268
Ion Ratio Lower Upper
59 100
57 1.4 8.1 12.1#



#12
1,1-Dichloroethene
Concen: 1.029 ug/l
RT: 4.365 min Scan# 410
Delta R.T. 0.006 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44

Tgt Ion: 96 Resp: 2436
Ion Ratio Lower Upper
96 100
61 137.1 125.8 188.6
98 55.8 52.5 78.7









#13
Acrolein
Concen: 5.998 ug/l
RT: 4.194 min Scan# 381
Delta R.T. -0.006 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion: 56 Resp: 1467
Ion Ratio Lower Upper
56 100
55 56.9 55.6 83.4



#14
Allyl chloride
Concen: 1.067 ug/l
RT: 5.041 min Scan# 525
Delta R.T. -0.000 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44

Tgt Ion: 41 Resp: 4189
Ion Ratio Lower Upper
41 100
39 35.3 51.3 76.9#
76 36.4 31.4 47.0









#15
Acrylonitrile
Concen: 5.106 ug/l
RT: 5.735 min Scan# 644
Delta R.T. -0.006 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001


Tgt Ion: 53	Resp:	9222	
Ion	Ratio	Lower	Upper
53	100		
52	85.4	65.0	97.6
51	32.7	27.2	40.8



#16
Acetone
Concen: 6.006 ug/l
RT: 4.459 min Scan# 426
Delta R.T. 0.011 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44

Tgt Ion: 43	Resp:	9069	
Ion	Ratio	Lower	Upper
43	100		
58	38.0	28.0	42.0









#17
Carbon Disulfide
Concen: 1.116 ug/l
RT: 4.730 min Scan# 41
Delta R.T. -0.000 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001


Tgt Ion: 76 Resp: 7306
Ion Ratio Lower Upper
76 100
78 10.3 7.3 10.9



#18
Methyl Acetate
Concen: 1.000 ug/l
RT: 5.053 min Scan# 527
Delta R.T. 0.011 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44

Tgt Ion: 43 Resp: 4402
Ion Ratio Lower Upper
43 100
74 8.7 20.7 31.1#









#19
Methyl tert-butyl Ether
Concen: 1.052 ug/l
RT: 5.824 min Scan# 619
Delta R.T. 0.011 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion: 73 Resp: 9019
Ion Ratio Lower Upper
73 100
57 18.0 17.4 26.2



#20
Methylene Chloride
Concen: 1.237 ug/l
RT: 5.294 min Scan# 568
Delta R.T. -0.000 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44


Tgt Ion: 84 Resp: 3496
Ion Ratio Lower Upper
84 100
49 125.7 90.5 135.7
51 35.9 28.5 42.7
86 60.5 48.1 72.1








#21
trans-1,2-Dichloroethene
Concen: 1.131 ug/l
RT: 5.800 min Scan# 61
Delta R.T. -0.006 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44

Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001




Tgt Ion: 96 Resp: 2979

Ion	Ratio	Lower	Upper
96	100		
61	119.4	109.2	163.8
98	54.3	49.8	74.8






#22
Diisopropyl ether
Concen: 1.037 ug/l
RT: 6.682 min Scan# 804
Delta R.T. -0.006 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44



Tgt Ion: 45 Resp: 8584

Ion	Ratio	Lower	Upper
45	100		
43	51.5	41.9	62.9
87	24.5	22.4	33.6
59	8.4	9.8	14.6









#23
 Vinyl Acetate
 Concen: 5.303 ug/l
 RT: 6.618 min Scan# 793
 Delta R.T. -0.000 min
 Lab File: VN086862.D
 Acq: 06 Jun 2025 12:44




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC001


Tgt Ion: 43 Resp: 37075
 Ion Ratio Lower Upper
 43 100
 86 10.0 9.5 14.3



#24
 1,1-Dichloroethane
 Concen: 1.065 ug/l
 RT: 6.588 min Scan# 788
 Delta R.T. 0.006 min
 Lab File: VN086862.D
 Acq: 06 Jun 2025 12:44

Tgt Ion: 63 Resp: 5071
 Ion Ratio Lower Upper
 63 100
 98 0.0 3.6 10.8#
 100 0.0 2.6 7.8#









#25
2-Butanone
Concen: 5.231 ug/l
RT: 7.500 min Scan# 943
Delta R.T. 0.006 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001


Tgt Ion: 43 Resp: 12841
Ion Ratio Lower Upper
43 100
72 32.7 23.0 34.6



#26
2,2-Dichloropropane
Concen: 1.110 ug/l
RT: 7.500 min Scan# 943
Delta R.T. -0.000 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44

Tgt Ion: 77 Resp: 4112
Ion Ratio Lower Upper
77 100
97 20.1 11.6 34.7









#27
 cis-1,2-Dichloroethene
 Concen: 1.062 ug/l
 RT: 7.494 min Scan# 942
 Delta R.T. -0.006 min
 Lab File: VN086862.D
 Acq: 06 Jun 2025 12:44




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC001


Tgt Ion: 96 Resp: 3345
 Ion Ratio Lower Upper
 96 100
 61 134.6 0.0 278.0
 98 58.7 0.0 128.2



#28
 Bromochloromethane
 Concen: 1.052 ug/l
 RT: 7.824 min Scan# 998
 Delta R.T. -0.000 min
 Lab File: VN086862.D
 Acq: 06 Jun 2025 12:44

Tgt Ion: 49 Resp: 2464
 Ion Ratio Lower Upper
 49 100
 129 0.0 0.0 3.8
 130 67.5 62.2 93.2





#29

Tetrahydrofuran

Concen: 5.191 ug/l

RT: 7.853 min Scan# 1003

Delta R.T. -0.000 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC001





Tgt Ion: 42 Resp: 8301

Ion Ratio Lower Upper

42 100

72 47.2 37.8 56.6

71 44.4 35.6 53.4



#30

Chloroform

Concen: 1.105 ug/l

RT: 7.971 min Scan# 1023

Delta R.T. -0.006 min

Lab File: VN086862.D




Acq: 06 Jun 2025 12:44


Tgt Ion: 83 Resp: 5255

Ion Ratio Lower Upper

83 100

85 68.6 52.9 79.3









#31
Cyclohexane
Concen: 1.980 ug/l
RT: 8.247 min Scan# 1074
Delta R.T. -0.024 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion: 56 Resp: 9143
Ion Ratio Lower Upper
56 100
69 124.2 25.4 38.0#
84 120.5 68.8 103.2#



#32
1,1,1-Trichloroethane
Concen: 1.082 ug/l
RT: 8.171 min Scan# 1057
Delta R.T. -0.006 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44

Tgt Ion: 97 Resp: 4378
Ion Ratio Lower Upper
97 100
99 0.0 52.0 78.0#
61 42.5 35.9 53.9





#33

1,2-Dichloroethane-d4

Concen: 1.204 ug/l

RT: 8.594 min Scan# 1129

Delta R.T. 0.006 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC001

Tgt Ion: 65 Resp: 3429

Ion Ratio Lower Upper

65 100

67 55.9 0.0 105.6



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.106 min Scan# 1216

Delta R.T. -0.000 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44




Tgt Ion: 114 Resp: 393457


Ion Ratio Lower Upper

114 100

63 20.2 0.0 39.6

88 16.0 0.0 30.2









#35
Dibromofluoromethane
Concen: 1.126 ug/l
RT: 8.171 min Scan# 1057
Delta R.T. -0.000 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001


Tgt Ion:113 Resp: 2626
Ion Ratio Lower Upper
113 100
111 94.8 84.2 126.2
192 7.7 14.2 21.4#



#36
1,1-Dichloropropene
Concen: 1.059 ug/l
RT: 8.376 min Scan# 1092
Delta R.T. -0.000 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44

Tgt Ion: 75 Resp: 3678
Ion Ratio Lower Upper
75 100
110 31.8 17.6 52.9
77 30.2 25.3 37.9





#37

Ethyl Acetate

Concen: 0.968 ug/l

RT: 7.577 min Scan# 91

Delta R.T. -0.000 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001





Tgt Ion: 43 Resp: 4282

Ion Ratio Lower Upper

43 100

61 0.0 12.1 18.1#

70 14.1 9.8 14.6



#38

Carbon Tetrachloride

Concen: 1.045 ug/l

RT: 8.376 min Scan# 1092

Delta R.T. 0.006 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44




Tgt Ion: 117 Resp: 3563


Ion Ratio Lower Upper

117 100

119 89.0 75.0 112.6

121 23.8 24.9 37.3#





#39

Methylcyclohexane

Concen: 1.046 ug/l

RT: 9.606 min Scan# 1130

Delta R.T. -0.000 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001





Tgt Ion: 83 Resp: 4981

Ion Ratio Lower Upper

83 100

55 79.7 61.0 91.6

98 61.6 38.9 58.3#



#40

Benzene

Concen: 1.100 ug/l

RT: 8.612 min Scan# 1132

Delta R.T. -0.000 min

Lab File: VN086862.D




Acq: 06 Jun 2025 12:44


Tgt Ion: 78 Resp: 12498

Ion Ratio Lower Upper

78 100

77 23.2 18.7 28.1









#41
Methacrylonitrile
Concen: 1.127 ug/l
RT: 7.782 min Scan# 991
Delta R.T. -0.006 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001


Tgt Ion:	41	Resp:	2799
Ion	Ratio	Lower	Upper
41	100		
39	50.3	41.2	61.8
67	62.4	61.4	92.2
52	22.9	22.7	34.1



#42
1,2-Dichloroethane
Concen: 1.081 ug/l
RT: 8.682 min Scan# 1144
Delta R.T. 0.006 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44

Tgt Ion:	62	Resp:	3725
Ion	Ratio	Lower	Upper
62	100		
98	3.9	0.0	21.2









#43
Isopropyl Acetate
Concen: 1.081 ug/l
RT: 8.694 min Scan# 1146
Delta R.T. -0.000 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001


Tgt Ion: 43 Resp: 7673
Ion Ratio Lower Upper
43 100
61 26.2 22.1 33.1
87 11.7 11.8 17.6#



#44
Trichloroethene
Concen: 1.047 ug/l
RT: 9.365 min Scan# 1260
Delta R.T. 0.006 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44

Tgt Ion:130 Resp: 2822
Ion Ratio Lower Upper
130 100
95 69.1 0.0 196.2





#45

1,2-Dichloropropane

Concen: 1.043 ug/l

RT: 9.629 min Scan# 1319

Delta R.T. -0.000 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44

Instrument :

MSVOA_N

ClientSampleId :





VSTDICC001

Tgt Ion: 63 Resp: 2883

Ion Ratio Lower Upper

63 100

65 29.0 25.5 38.3



#46

Dibromomethane

Concen: 0.999 ug/l

RT: 9.712 min Scan# 1319

Delta R.T. -0.006 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44




Tgt Ion: 93 Resp: 1835


Ion Ratio Lower Upper

93 100

95 87.6 67.1 100.7

174 100.5 74.2 111.4





#47

Bromodichloromethane

Concen: 1.063 ug/l

RT: 9.888 min Scan# 1349

Delta R.T. -0.006 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001





Tgt Ion: 83 Resp: 4017

Ion Ratio Lower Upper

83 100

85 62.7 50.9 76.3

127 0.0 6.6 10.0#



#48

Methyl methacrylate

Concen: 1.069 ug/l

RT: 9.688 min Scan# 1315

Delta R.T. 0.006 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44




Tgt Ion: 41 Resp: 3492


Ion Ratio Lower Upper

41 100

69 79.4 74.6 112.0

39 44.9 40.2 60.4





#49

1,4-Dioxane

Concen: 16.891 ug/l

RT: 9.706 min Scan# 1

Delta R.T. 0.006 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC001





Tgt Ion: 88 Resp: 1004

Ion Ratio Lower Upper

88 100

43 0.0 0.0 0.0

58 94.0 57.7 86.5#



#50

Toluene-d8

Concen: 1.153 ug/l

RT: 10.570 min Scan# 1465

Delta R.T. -0.000 min

Lab File: VN086862.D




Acq: 06 Jun 2025 12:44


Tgt Ion: 98 Resp: 10645

Ion Ratio Lower Upper

98 100

100 67.3 53.4 80.0





#51

4-Methyl-2-Pentanone

Concen: 4.653 ug/l

RT: 10.453 min Scan# 1444

Delta R.T. 0.006 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44

Instrument :

MSVOA_N

ClientSampleId :





VSTDICC001

Tgt Ion: 43 Resp: 19886

Ion Ratio Lower Upper

43 100

58 42.3 34.2 51.4



#52

Toluene

Concen: 1.041 ug/l

RT: 10.635 min Scan# 1476

Delta R.T. -0.000 min

Lab File: VN086862.D




Acq: 06 Jun 2025 12:44


Tgt Ion: 92 Resp: 7227

Ion Ratio Lower Upper

92 100

91 178.1 136.5 204.7









#53
t-1,3-Dichloropropene
Concen: 1.064 ug/l
RT: 10.841 min Scan# 1511
Delta R.T. -0.000 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001


Tgt Ion: 75 Resp: 4494
Ion Ratio Lower Upper
75 100
77 31.4 26.4 39.6



#54
cis-1,3-Dichloropropene
Concen: 1.060 ug/l
RT: 10.312 min Scan# 1421
Delta R.T. -0.006 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44

Tgt Ion: 75 Resp: 4792
Ion Ratio Lower Upper
75 100
77 35.9 26.2 39.4
39 41.1 32.3 48.5









#55
1,1,2-Trichloroethane
Concen: 1.058 ug/l
RT: 11.023 min Scan# 11
Delta R.T. 0.006 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion: 97	Resp: 2828
Ion Ratio	Lower Upper
97 100	
83 80.6	69.7 104.5
85 46.5	45.1 67.7
99 52.5	50.6 75.8



#56
Ethyl methacrylate
Concen: 0.800 ug/l
RT: 10.882 min Scan# 1518
Delta R.T. -0.000 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44

Tgt Ion: 69	Resp: 3406
Ion Ratio	Lower Upper
69 100	
41 73.5	47.2 70.8#
39 31.6	23.0 34.6





#57

1,3-Dichloropropane

Concen: 1.092 ug/l

RT: 11.170 min Scan# 1566

Delta R.T. 0.006 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC001

Tgt Ion: 76 Resp: 5062

Ion Ratio Lower Upper

76 100

78 32.8 25.3 37.9



#58

2-Chloroethyl Vinyl ether

Concen: 4.304 ug/l

RT: 10.165 min Scan# 1396

Delta R.T. -0.000 min

Lab File: VN086862.D




Acq: 06 Jun 2025 12:44


Tgt Ion: 63 Resp: 10924

Ion Ratio Lower Upper

63 100

106 30.4 23.0 34.6









#59
2-Hexanone
Concen: 1.822 ug/l
RT: 11.235 min Scan# 1578
Delta R.T. 0.035 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion: 43 Resp: 5016
Ion Ratio Lower Upper
43 100
58 11.5 30.4 91.3#



#60
Dibromochloromethane
Concen: 1.021 ug/l
RT: 11.359 min Scan# 1599
Delta R.T. -0.000 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44

Tgt Ion: 129 Resp: 2843
Ion Ratio Lower Upper
129 100
127 73.8 38.6 115.8









#61
1,2-Dibromoethane
Concen: 1.014 ug/l
RT: 11.470 min Scan# 1618
Delta R.T. -0.000 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion:107 Resp: 2777
Ion Ratio Lower Upper
107 100
109 91.2 76.3 114.5



#62
4-Bromofluorobenzene
Concen: 1.240 ug/l
RT: 12.853 min Scan# 1853
Delta R.T. 0.006 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44

Tgt Ion: 95 Resp: 4252
Ion Ratio Lower Upper
95 100
174 68.3 0.0 141.8
176 65.0 0.0 132.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 117
Delta R.T. -0.000 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion:117 Resp: 348010
Ion Ratio Lower Upper
117 100
82 57.5 44.6 67.0
119 32.5 25.5 38.3



#64
Tetrachloroethene
Concen: 1.123 ug/l
RT: 11.106 min Scan# 1556
Delta R.T. -0.000 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44

Tgt Ion:164 Resp: 2473
Ion Ratio Lower Upper
164 100
166 121.4 101.8 152.8
129 104.4 83.0 124.4
131 84.3 78.2 117.2









#65
Chlorobenzene
Concen: 1.118 ug/l
RT: 11.894 min Scan# 11
Delta R.T. -0.000 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion:112 Resp: 8581
Ion Ratio Lower Upper
112 100
114 25.7 26.2 39.2#



#66
1,1,1,2-Tetrachloroethane
Concen: 1.023 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. -0.006 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44

Tgt Ion:131 Resp: 2523
Ion Ratio Lower Upper
131 100
133 100.9 48.3 144.8
119 0.0 34.0 101.8#





#67

Ethyl Benzene

Concen: 1.047 ug/l

RT: 11.970 min Scan# 1703

Delta R.T. 0.006 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC001

Tgt Ion: 91 Resp: 13845

Ion Ratio Lower Upper

91 100

106 28.7 25.4 38.2



#68

m/p-Xylenes

Concen: 2.008 ug/l

RT: 12.070 min Scan# 1720

Delta R.T. -0.000 min

Lab File: VN086862.D




Acq: 06 Jun 2025 12:44


Tgt Ion: 106 Resp: 10163

Ion Ratio Lower Upper

106 100

91 207.3 159.4 239.0





#69

o-Xylene

Concen: 1.025 ug/l

RT: 12.400 min Scan# 1776

Delta R.T. -0.000 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC001

Tgt Ion:106 Resp: 4970

Ion Ratio Lower Upper

106 100

91 223.6 106.1 318.1



#70

Styrene

Concen: 0.987 ug/l

RT: 12.417 min Scan# 1779

Delta R.T. 0.006 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44




Tgt Ion:104 Resp: 8189


Ion Ratio Lower Upper

104 100

78 49.6 38.5 57.7

103 52.5 43.1 64.7









#71
Bromoform
Concen: 0.825 ug/l
RT: 12.582 min Scan# 1807
Delta R.T. -0.000 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion:	173	Resp:	1508
Ion Ratio	Lower	Upper	
173	100		
175	63.1	24.1	72.4
254	0.0	0.0	0.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.794 min Scan# 2013
Delta R.T. -0.000 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44

Tgt Ion:	152	Resp:	167399
Ion Ratio	Lower	Upper	
152	100		
115	61.1	30.1	90.5
150	158.0	0.0	345.0





#73

Isopropylbenzene

Concen: 1.061 ug/l

RT: 12.694 min Scan# 1826

Delta R.T. -0.000 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC001

Tgt Ion:105 Resp: 12938

Ion Ratio Lower Upper

105 100

120 28.1 13.5 40.4



#75

1,1,2,2-Tetrachloroethane

Concen: 1.047 ug/l

RT: 12.935 min Scan# 1867

Delta R.T. -0.006 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44




Tgt Ion: 83 Resp: 4326


Ion Ratio Lower Upper

83 100

131 4.1 4.9 14.7#

85 68.6 32.0 96.0









#76
1,2,3-Trichloropropane
Concen: 1.484 ug/l
RT: 13.000 min Scan# 1878
Delta R.T. 0.006 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion: 75 Resp: 5925
Ion Ratio Lower Upper
75 100
77 113.2 79.3 237.9



#77
Bromobenzene
Concen: 1.042 ug/l
RT: 12.982 min Scan# 1875
Delta R.T. -0.000 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44

Tgt Ion:156 Resp: 2913
Ion Ratio Lower Upper
156 100
77 230.2 117.7 353.1
158 102.5 48.1 144.4









#78
n-propylbenzene
Concen: 1.023 ug/l
RT: 13.035 min Scan# 1884
Delta R.T. -0.000 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion: 91 Resp: 15168
Ion Ratio Lower Upper
91 100
120 21.9 11.4 34.2



#79
2-Chlorotoluene
Concen: 1.074 ug/l
RT: 13.123 min Scan# 1899
Delta R.T. -0.000 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44

Tgt Ion: 91 Resp: 9550
Ion Ratio Lower Upper
91 100
126 30.4 16.2 48.5









#80
1,3,5-Trimethylbenzene
Concen: 0.999 ug/l
RT: 13.170 min Scan# 1907
Delta R.T. -0.000 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion:105 Resp: 10058
Ion Ratio Lower Upper
105 100
120 52.7 24.9 74.6



#81
trans-1,4-Dichloro-2-butene
Concen: 0.805 ug/l
RT: 12.741 min Scan# 1834
Delta R.T. 0.006 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44

Tgt Ion: 75 Resp: 1392
Ion Ratio Lower Upper
75 100
53 93.8 66.6 99.8
89 46.6 36.7 55.1









#82
4-Chlorotoluene
Concen: 1.079 ug/l
RT: 13.223 min Scan# 1916
Delta R.T. -0.000 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion: 91 Resp: 9706
Ion Ratio Lower Upper
91 100
126 31.1 16.7 50.1



#83
tert-Butylbenzene
Concen: 1.069 ug/l
RT: 13.435 min Scan# 1952
Delta R.T. -0.000 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44

Tgt Ion: 119 Resp: 9849
Ion Ratio Lower Upper
119 100
91 65.3 32.1 96.5
134 23.9 13.4 40.1









#84
1,2,4-Trimethylbenzene
Concen: 0.979 ug/l
RT: 13.482 min Scan# 1960
Delta R.T. -0.000 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion:105 Resp: 9882
Ion Ratio Lower Upper
105 100
120 47.2 23.2 69.6



#85
sec-Butylbenzene
Concen: 1.029 ug/l
RT: 13.617 min Scan# 1983
Delta R.T. -0.000 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44

Tgt Ion:105 Resp: 13775
Ion Ratio Lower Upper
105 100
134 20.2 10.0 30.0





#86

p-Isopropyltoluene

Concen: 1.000 ug/l

RT: 13.729 min Scan# 2002

Delta R.T. -0.000 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001





Tgt Ion:119 Resp: 11065

Ion Ratio Lower Upper

119 100

134 25.4 13.5 40.4

91 24.1 11.5 34.4



#87

1,3-Dichlorobenzene

Concen: 1.073 ug/l

RT: 13.735 min Scan# 2003

Delta R.T. -0.000 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44




Tgt Ion:146 Resp: 5903


Ion Ratio Lower Upper

146 100

111 45.8 20.9 62.8

148 59.7 32.0 96.2









#88
1,4-Dichlorobenzene
Concen: 1.052 ug/l
RT: 13.735 min Scan# 2016
Delta R.T. -0.077 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001


Tgt Ion:146 Resp: 5903
Ion Ratio Lower Upper
146 100
111 45.8 19.9 59.7
148 59.7 31.8 95.4



#89
n-Butylbenzene
Concen: 1.085 ug/l
RT: 14.058 min Scan# 2058
Delta R.T. 0.006 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44

Tgt Ion: 91 Resp: 11627
Ion Ratio Lower Upper
91 100
92 50.8 27.5 82.4
134 22.6 13.0 39.0









#90
Hexachloroethane
Concen: 1.036 ug/l
RT: 14.329 min Scan# 2104
Delta R.T. -0.006 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion:117 Resp: 1942
Ion Ratio Lower Upper
117 100
201 60.6 33.2 99.6



#91
1,2-Dichlorobenzene
Concen: 1.061 ug/l
RT: 14.106 min Scan# 2066
Delta R.T. -0.000 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44

Tgt Ion:146 Resp: 5607
Ion Ratio Lower Upper
146 100
111 43.1 21.6 64.6
148 64.7 31.7 95.1









#92
 1,2-Dibromo-3-Chloropropane
 Concen: 1.150 ug/l
 RT: 14.717 min Scan# 2170
 Delta R.T. -0.000 min
 Lab File: VN086862.D
 Acq: 06 Jun 2025 12:44




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC001


Tgt Ion: 75 Resp: 1136
 Ion Ratio Lower Upper
 75 100
 155 63.6 41.1 123.4
 157 81.0 52.7 158.1



#93
 1,2,4-Trichlorobenzene
 Concen: 1.034 ug/l
 RT: 15.388 min Scan# 2284
 Delta R.T. -0.000 min
 Lab File: VN086862.D
 Acq: 06 Jun 2025 12:44

Tgt Ion: 180 Resp: 3489
 Ion Ratio Lower Upper
 180 100
 182 92.1 46.9 140.7
 145 30.4 16.0 47.9









#94
Hexachlorobutadiene
Concen: 0.969 ug/l
RT: 15.500 min Scan# 21
Delta R.T. -0.000 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion:	225	Resp:	1219
Ion	Ratio	Lower	Upper
225	100		
223	53.6	33.5	100.4
227	57.5	31.1	93.5



#95
Naphthalene
Concen: 1.018 ug/l
RT: 15.635 min Scan# 2326
Delta R.T. -0.006 min
Lab File: VN086862.D
Acq: 06 Jun 2025 12:44


Tgt Ion:	128	Resp:	12789
Ion	Ratio	Lower	Upper
128	100		
127	12.4	10.2	15.2
129	9.5	8.8	13.2







#96
 1,2,3-Trichlorobenzene
 Concen: 1.072 ug/l
 RT: 15.835 min Scan# 21
 Delta R.T. -0.000 min
 Lab File: VN086862.D
 Acq: 06 Jun 2025 12:44

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC001



Tgt Ion:	180	Resp:	3594
Ion	Ratio	Lower	Upper
180	100		
182	91.6	47.0	141.0
145	29.0	16.6	49.8



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060625\
 Data File : VN086863.D
 Acq On : 06 Jun 2025 13:17
 Operator : JC\MD
 Sample : VSTDIC005
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC005

Quant Time: Jun 07 01:55:22 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 01:51:02 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	229701	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	423980	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	371851	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	181065	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	16818	5.468	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	10.940%#
35) Dibromofluoromethane	8.177	113	12867	5.121	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	10.240%#
50) Toluene-d8	10.571	98	52794	5.308	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	10.620%#
62) 4-Bromofluorobenzene	12.847	95	18697	5.059	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	10.120%#

Target Compounds				Qvalue		
2) Dichlorodifluoromethane	2.154	85	10196	4.452	ug/l	93
3) Chloromethane	2.395	50	15025	5.080	ug/l	98
4) Vinyl Chloride	2.554	62	15384	5.043	ug/l	97
5) Bromomethane	2.989	94	8606	5.041	ug/l	97
6) Chloroethane	3.154	64	10190	5.171	ug/l	96
7) Trichlorofluoromethane	3.518	101	20740	5.203	ug/l	99
8) Diethyl Ether	3.971	74	9121	5.252	ug/l	95
9) 1,1,2-Trichlorotrifluo...	4.401	101	13034	5.205	ug/l	97
10) Methyl Iodide	4.606	142	16530	5.092	ug/l	98
11) Tert butyl alcohol	5.542	59	22000	26.363	ug/l	99
12) 1,1-Dichloroethene	4.359	96	13610	5.323	ug/l	94
13) Acrolein	4.195	56	8401	31.804	ug/l	99
14) Allyl chloride	5.042	41	20925	4.935	ug/l	99
15) Acrylonitrile	5.736	53	49826	25.545	ug/l	99
16) Acetone	4.448	43	41981	25.741	ug/l	99
17) Carbon Disulfide	4.736	76	37246	5.267	ug/l	97
18) Methyl Acetate	5.048	43	24096	5.070	ug/l	97
19) Methyl tert-butyl Ether	5.812	73	46823	5.058	ug/l	95
20) Methylene Chloride	5.300	84	15809	5.178	ug/l	96
21) trans-1,2-Dichloroethene	5.806	96	15480	5.442	ug/l #	81
22) Diisopropyl ether	6.689	45	46396	5.191	ug/l	99
23) Vinyl Acetate	6.618	43	194334	25.735	ug/l	100
24) 1,1-Dichloroethane	6.583	63	26473	5.147	ug/l	99
25) 2-Butanone	7.494	43	68629	25.888	ug/l	99
26) 2,2-Dichloropropane	7.500	77	18288	4.571	ug/l	100
27) cis-1,2-Dichloroethene	7.500	96	17594	5.172	ug/l	98
28) Bromochloromethane	7.830	49	12958	5.124	ug/l #	97
29) Tetrahydrofuran	7.853	42	44749	25.912	ug/l	98
30) Chloroform	7.977	83	26450	5.149	ug/l	95
31) Cyclohexane	8.271	56	29938	6.002	ug/l	95
32) 1,1,1-Trichloroethane	8.177	97	22857	5.232	ug/l #	50
36) 1,1-Dichloropropene	8.377	75	19410	5.184	ug/l	98
37) Ethyl Acetate	7.571	43	26073	5.470	ug/l	97
38) Carbon Tetrachloride	8.371	117	19056	5.184	ug/l	99
39) Methylcyclohexane	9.606	83	27353	5.333	ug/l	95
40) Benzene	8.612	78	63646	5.199	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060625\
 Data File : VN086863.D
 Acq On : 06 Jun 2025 13:17
 Operator : JC\MD
 Sample : VSTDICC005
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC005

Quant Time: Jun 07 01:55:22 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 01:51:02 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.783	41	13518	5.050	ug/l	99
42) 1,2-Dichloroethane	8.683	62	19329	5.205	ug/l	98
43) Isopropyl Acetate	8.700	43	40268	5.264	ug/l	98
44) Trichloroethene	9.359	130	15245	5.251	ug/l	96
45) 1,2-Dichloropropane	9.624	63	15632	5.248	ug/l	99
46) Dibromomethane	9.718	93	10241	5.176	ug/l	99
47) Bromodichloromethane	9.888	83	20512	5.039	ug/l	97
48) Methyl methacrylate	9.688	41	17591	4.996	ug/l	94
49) 1,4-Dioxane	9.700	88	6141	95.874	ug/l #	93
51) 4-Methyl-2-Pentanone	10.447	43	116434	25.283	ug/l	97
52) Toluene	10.630	92	38744	5.178	ug/l	99
53) t-1,3-Dichloropropene	10.841	75	22285	4.896	ug/l	98
54) cis-1,3-Dichloropropene	10.318	75	24477	5.026	ug/l	93
55) 1,1,2-Trichloroethane	11.018	97	15068	5.234	ug/l	97
56) Ethyl methacrylate	10.888	69	21893	4.774	ug/l	95
57) 1,3-Dichloropropane	11.165	76	25452	5.096	ug/l	98
58) 2-Chloroethyl Vinyl ether	10.165	63	64722	23.664	ug/l	98
59) 2-Hexanone	11.212	43	59884	20.188	ug/l	99
60) Dibromochloromethane	11.359	129	14884	4.962	ug/l	98
61) 1,2-Dibromoethane	11.471	107	15172	5.141	ug/l	99
64) Tetrachloroethene	11.106	164	12322	5.236	ug/l	92
65) Chlorobenzene	11.894	112	42195	5.145	ug/l	98
66) 1,1,1,2-Tetrachloroethane	11.965	131	13897	5.271	ug/l	97
67) Ethyl Benzene	11.965	91	73443	5.199	ug/l	97
68) m/p-Xylenes	12.071	106	55859	10.330	ug/l	100
69) o-Xylene	12.400	106	26004	5.021	ug/l	96
70) Styrene	12.412	104	44380	5.008	ug/l	98
71) Bromoform	12.582	173	9890	5.063	ug/l #	96
73) Isopropylbenzene	12.694	105	67878	5.146	ug/l	100
74) N-amyl acetate	12.559	43	7051	1.611	ug/l #	53
75) 1,1,2,2-Tetrachloroethane	12.935	83	23516	5.262	ug/l	98
76) 1,2,3-Trichloropropane	12.994	75	31004	7.181	ug/l	74
77) Bromobenzene	12.976	156	15485	5.119	ug/l	92
78) n-propylbenzene	13.035	91	83781	5.226	ug/l	99
79) 2-Chlorotoluene	13.124	91	49297	5.128	ug/l	100
80) 1,3,5-Trimethylbenzene	13.171	105	55961	5.139	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.741	75	9234	4.939	ug/l	94
82) 4-Chlorotoluene	13.223	91	49924	5.132	ug/l	100
83) tert-Butylbenzene	13.435	119	50874	5.103	ug/l	98
84) 1,2,4-Trimethylbenzene	13.482	105	56525	5.176	ug/l	100
85) sec-Butylbenzene	13.618	105	74711	5.161	ug/l	99
86) p-Isopropyltoluene	13.729	119	62016	5.182	ug/l	99
87) 1,3-Dichlorobenzene	13.735	146	30949	5.200	ug/l	98
88) 1,4-Dichlorobenzene	13.812	146	32342	5.330	ug/l	96
89) n-Butylbenzene	14.053	91	59601	5.142	ug/l	99
90) Hexachloroethane	14.335	117	9725	4.794	ug/l	95
91) 1,2-Dichlorobenzene	14.106	146	29897	5.228	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.717	75	5731	5.363	ug/l	93
93) 1,2,4-Trichlorobenzene	15.394	180	18778	5.143	ug/l	99
94) Hexachlorobutadiene	15.500	225	7072	5.199	ug/l	99
95) Naphthalene	15.641	128	67481	4.965	ug/l	99
96) 1,2,3-Trichlorobenzene	15.841	180	18262	5.034	ug/l	97


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060625\
Data File : VN086863.D
Acq On : 06 Jun 2025 13:17
Operator : JC\MD
Sample : VSTDICC005
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Quant Time: Jun 07 01:55:22 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 01:51:02 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.230 min Scan# 11
Delta R.T. -0.006 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005


Tgt Ion:168 Resp: 229701
Ion Ratio Lower Upper
168 100
99 59.2 49.1 73.7



#2
Dichlorodifluoromethane
Concen: 4.452 ug/l
RT: 2.154 min Scan# 34
Delta R.T. -0.000 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17

Tgt Ion: 85 Resp: 10196
Ion Ratio Lower Upper
85 100
87 28.6 16.2 48.6









#3
Chloromethane
Concen: 5.080 ug/l
RT: 2.395 min Scan# 75
Delta R.T. 0.000 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005


Tgt Ion: 50 Resp: 15025
Ion Ratio Lower Upper
50 100
52 30.9 25.7 38.5



#4
Vinyl Chloride
Concen: 5.043 ug/l
RT: 2.554 min Scan# 102
Delta R.T. 0.000 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17

Tgt Ion: 62 Resp: 15384
Ion Ratio Lower Upper
62 100
64 34.3 26.0 39.0









#5
Bromomethane
Concen: 5.041 ug/l
RT: 2.989 min Scan# 176
Delta R.T. -0.006 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005


Tgt Ion: 94 Resp: 8606
Ion Ratio Lower Upper
94 100
96 100.6 77.9 116.9



#6
Chloroethane
Concen: 5.171 ug/l
RT: 3.154 min Scan# 204
Delta R.T. -0.000 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17

Tgt Ion: 64 Resp: 10190
Ion Ratio Lower Upper
64 100
66 28.8 24.6 36.8









#7
Trichlorofluoromethane
Concen: 5.203 ug/l
RT: 3.518 min Scan# 207
Delta R.T. -0.006 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005


Tgt Ion:101 Resp: 20740
Ion Ratio Lower Upper
101 100
103 64.8 52.2 78.2



#8
Diethyl Ether
Concen: 5.252 ug/l
RT: 3.971 min Scan# 343
Delta R.T. -0.012 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17

Tgt Ion: 74 Resp: 9121
Ion Ratio Lower Upper
74 100
45 95.8 45.5 136.5









#9
1,1,2-Trichlorotrifluoroethane
Concen: 5.205 ug/l
RT: 4.401 min Scan# 415
Delta R.T. 0.006 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005


Tgt Ion:101 Resp: 13034
Ion Ratio Lower Upper
101 100
85 43.3 33.8 50.8
151 70.8 58.7 88.1



#10
Methyl Iodide
Concen: 5.092 ug/l
RT: 4.606 min Scan# 451
Delta R.T. -0.000 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17

Tgt Ion:142 Resp: 16530
Ion Ratio Lower Upper
142 100
127 39.8 30.9 46.3
141 14.5 10.4 15.6









#11
Tert butyl alcohol
Concen: 26.363 ug/l
RT: 5.542 min Scan# 611
Delta R.T. -0.006 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005


Tgt Ion: 59 Resp: 22000
Ion Ratio Lower Upper
59 100
57 9.6 8.1 12.1



#12
1,1-Dichloroethene
Concen: 5.323 ug/l
RT: 4.359 min Scan# 409
Delta R.T. -0.000 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17

Tgt Ion: 96 Resp: 13610
Ion Ratio Lower Upper
96 100
61 148.1 125.8 188.6
98 67.5 52.5 78.7





#13

Acrolein

Concen: 31.804 ug/l

RT: 4.195 min Scan# 381

Delta R.T. -0.006 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17

Instrument :

MSVOA_N

ClientSampleId :





VSTDICC005

Tgt Ion: 56 Resp: 8401

Ion Ratio Lower Upper

56 100

55 70.2 55.6 83.4



#14

Allyl chloride

Concen: 4.935 ug/l

RT: 5.042 min Scan# 525

Delta R.T. -0.000 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17




Tgt Ion: 41 Resp: 20925


Ion Ratio Lower Upper

41 100

39 63.2 51.3 76.9

76 39.0 31.4 47.0









#15
Acrylonitrile
Concen: 25.545 ug/l
RT: 5.736 min Scan# 644
Delta R.T. -0.006 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005


Tgt Ion: 53	Resp: 49826
Ion Ratio	Lower Upper
53 100	
52 82.4	65.0 97.6
51 32.9	27.2 40.8



#16
Acetone
Concen: 25.741 ug/l
RT: 4.448 min Scan# 424
Delta R.T. -0.000 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17

Tgt Ion: 43	Resp: 41981
Ion Ratio	Lower Upper
43 100	
58 34.1	28.0 42.0









#17
Carbon Disulfide
Concen: 5.267 ug/l
RT: 4.736 min Scan# 41
Delta R.T. 0.006 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005


Tgt Ion: 76 Resp: 37246
Ion Ratio Lower Upper
76 100
78 8.0 7.3 10.9



#18
Methyl Acetate
Concen: 5.070 ug/l
RT: 5.048 min Scan# 526
Delta R.T. 0.006 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17

Tgt Ion: 43 Resp: 24096
Ion Ratio Lower Upper
43 100
74 24.4 20.7 31.1









#19
Methyl tert-butyl Ether
Concen: 5.058 ug/l
RT: 5.812 min Scan# 61
Delta R.T. -0.000 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005


Tgt Ion: 73 Resp: 46823
Ion Ratio Lower Upper
73 100
57 24.3 17.4 26.2



#20
Methylene Chloride
Concen: 5.178 ug/l
RT: 5.300 min Scan# 569
Delta R.T. 0.006 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17

Tgt Ion: 84 Resp: 15809
Ion Ratio Lower Upper
84 100
49 106.8 90.5 135.7
51 36.6 28.5 42.7
86 60.4 48.1 72.1









#21
trans-1,2-Dichloroethene
Concen: 5.442 ug/l
RT: 5.806 min Scan# 61
Delta R.T. -0.000 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005


Tgt Ion: 96 Resp: 15480
Ion Ratio Lower Upper
96 100
61 105.6 109.2 163.8#
98 58.4 49.8 74.8



#22
Diisopropyl ether
Concen: 5.191 ug/l
RT: 6.689 min Scan# 805
Delta R.T. -0.000 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17

Tgt Ion: 45 Resp: 46396
Ion Ratio Lower Upper
45 100
43 53.2 41.9 62.9
87 27.9 22.4 33.6
59 12.6 9.8 14.6









#23
 Vinyl Acetate
 Concen: 25.735 ug/l
 RT: 6.618 min Scan# 793
 Delta R.T. -0.000 min
 Lab File: VN086863.D
 Acq: 06 Jun 2025 13:17




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC005


Tgt Ion: 43 Resp: 194334
 Ion Ratio Lower Upper
 43 100
 86 11.8 9.5 14.3



#24
 1,1-Dichloroethane
 Concen: 5.147 ug/l
 RT: 6.583 min Scan# 787
 Delta R.T. 0.000 min
 Lab File: VN086863.D
 Acq: 06 Jun 2025 13:17

Tgt Ion: 63 Resp: 26473
 Ion Ratio Lower Upper
 63 100
 98 6.4 3.6 10.8
 100 5.1 2.6 7.8









#25
2-Butanone
Concen: 25.888 ug/l
RT: 7.494 min Scan# 942
Delta R.T. 0.000 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005


Tgt Ion: 43 Resp: 68629
Ion Ratio Lower Upper
43 100
72 28.1 23.0 34.6



#26
2,2-Dichloropropane
Concen: 4.571 ug/l
RT: 7.500 min Scan# 943
Delta R.T. 0.000 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17

Tgt Ion: 77 Resp: 18288
Ion Ratio Lower Upper
77 100
97 23.3 11.6 34.7









#27
cis-1,2-Dichloroethene
Concen: 5.172 ug/l
RT: 7.500 min Scan# 943
Delta R.T. 0.000 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005


Tgt Ion:	96	Resp:	17594
Ion	Ratio	Lower	Upper
96	100		
61	136.0	0.0	278.0
98	64.0	0.0	128.2



#28
Bromochloromethane
Concen: 5.124 ug/l
RT: 7.830 min Scan# 999
Delta R.T. 0.006 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17

Tgt Ion:	49	Resp:	12958
Ion	Ratio	Lower	Upper
49	100		
129	0.0	0.0	3.8
130	75.2	62.2	93.2





#29

Tetrahydrofuran

Concen: 25.912 ug/l

RT: 7.853 min Scan# 1003

Delta R.T. 0.000 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC005





Tgt Ion: 42 Resp: 44749

Ion Ratio Lower Upper

42 100

72 48.6 37.8 56.6

71 43.5 35.6 53.4



#30

Chloroform

Concen: 5.149 ug/l

RT: 7.977 min Scan# 1024

Delta R.T. -0.000 min

Lab File: VN086863.D




Acq: 06 Jun 2025 13:17


Tgt Ion: 83 Resp: 26450

Ion Ratio Lower Upper

83 100

85 69.9 52.9 79.3









#31
Cyclohexane
Concen: 6.002 ug/l
RT: 8.271 min Scan# 1074
Delta R.T. -0.000 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005


Tgt Ion:	56	Resp:	29938
Ion Ratio	Lower	Upper	
56	100		
69	33.2	25.4	38.0
84	81.1	68.8	103.2



#32
1,1,1-Trichloroethane
Concen: 5.232 ug/l
RT: 8.177 min Scan# 1058
Delta R.T. -0.000 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17

Tgt Ion:	97	Resp:	22857
Ion Ratio	Lower	Upper	
97	100		
99	0.0	52.0	78.0#
61	42.9	35.9	53.9





#33

1,2-Dichloroethane-d4

Concen: 5.468 ug/l

RT: 8.588 min Scan# 1128

Delta R.T. 0.000 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17

Instrument :

MSVOA_N

ClientSampleId :





VSTDICC005

Tgt Ion: 65 Resp: 16818

Ion Ratio Lower Upper

65 100

67 54.4 0.0 105.6



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.106 min Scan# 1216

Delta R.T. -0.000 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17




Tgt Ion: 114 Resp: 423980


Ion Ratio Lower Upper

114 100

63 20.8 0.0 39.6

88 15.9 0.0 30.2





#35

Dibromofluoromethane

Concen: 5.121 ug/l

RT: 8.177 min Scan# 1057

Delta R.T. 0.006 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005





Tgt Ion:113 Resp: 12867

Ion Ratio Lower Upper

113 100

111 106.5 84.2 126.2

192 18.0 14.2 21.4



#36

1,1-Dichloropropene

Concen: 5.184 ug/l

RT: 8.377 min Scan# 1092

Delta R.T. 0.000 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17




Tgt Ion: 75 Resp: 19410


Ion Ratio Lower Upper

75 100

110 37.4 17.6 52.9

77 31.7 25.3 37.9









#37
Ethyl Acetate
Concen: 5.470 ug/l
RT: 7.571 min Scan# 91
Delta R.T. -0.006 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005


Tgt Ion: 43 Resp: 26073
Ion Ratio Lower Upper
43 100
61 14.1 12.1 18.1
70 10.9 9.8 14.6



#38
Carbon Tetrachloride
Concen: 5.184 ug/l
RT: 8.371 min Scan# 1091
Delta R.T. 0.000 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17

Tgt Ion: 117 Resp: 19056
Ion Ratio Lower Upper
117 100
119 93.8 75.0 112.6
121 28.5 24.9 37.3





#39

Methylcyclohexane

Concen: 5.333 ug/l

RT: 9.606 min Scan# 1130

Delta R.T. -0.000 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005





Tgt Ion: 83 Resp: 27353

Ion Ratio Lower Upper

83 100

55 73.6 61.0 91.6

98 42.6 38.9 58.3



#40

Benzene

Concen: 5.199 ug/l

RT: 8.612 min Scan# 1132

Delta R.T. -0.000 min

Lab File: VN086863.D




Acq: 06 Jun 2025 13:17


Tgt Ion: 78 Resp: 63646

Ion Ratio Lower Upper

78 100

77 22.6 18.7 28.1









#41
Methacrylonitrile
Concen: 5.050 ug/l
RT: 7.783 min Scan# 991
Delta R.T. -0.006 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005


Tgt Ion:	41	Resp:	13518
Ion	Ratio	Lower	Upper
41	100		
39	53.1	41.2	61.8
67	77.0	61.4	92.2
52	29.8	22.7	34.1



#42
1,2-Dichloroethane
Concen: 5.205 ug/l
RT: 8.683 min Scan# 1144
Delta R.T. 0.006 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17

Tgt Ion:	62	Resp:	19329
Ion	Ratio	Lower	Upper
62	100		
98	9.9	0.0	21.2









#43
Isopropyl Acetate
Concen: 5.264 ug/l
RT: 8.700 min Scan# 1146
Delta R.T. 0.006 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005


Tgt Ion: 43 Resp: 40268
Ion Ratio Lower Upper
43 100
61 26.2 22.1 33.1
87 14.1 11.8 17.6



#44
Trichloroethene
Concen: 5.251 ug/l
RT: 9.359 min Scan# 1259
Delta R.T. 0.000 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17

Tgt Ion: 130 Resp: 15245
Ion Ratio Lower Upper
130 100
95 101.6 0.0 196.2





#45

1,2-Dichloropropane

Concen: 5.248 ug/l

RT: 9.624 min Scan# 1304

Delta R.T. -0.006 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC005

Tgt Ion: 63 Resp: 15632

Ion Ratio Lower Upper

63 100

65 32.7 25.5 38.3



#46

Dibromomethane

Concen: 5.176 ug/l

RT: 9.718 min Scan# 1320

Delta R.T. -0.000 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17




Tgt Ion: 93 Resp: 10241


Ion Ratio Lower Upper

93 100

95 84.9 67.1 100.7

174 93.4 74.2 111.4









#47
Bromodichloromethane
Concen: 5.039 ug/l
RT: 9.888 min Scan# 1349
Delta R.T. -0.006 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005


Tgt Ion: 83 Resp: 20512
Ion Ratio Lower Upper
83 100
85 61.3 50.9 76.3
127 7.0 6.6 10.0



#48
Methyl methacrylate
Concen: 4.996 ug/l
RT: 9.688 min Scan# 1315
Delta R.T. 0.006 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17

Tgt Ion: 41 Resp: 17591
Ion Ratio Lower Upper
41 100
69 85.2 74.6 112.0
39 52.3 40.2 60.4





#49

1,4-Dioxane

Concen: 95.874 ug/l

RT: 9.700 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005





Tgt Ion: 88 Resp: 6141

Ion Ratio Lower Upper

88 100

43 0.0 0.0 0.0

58 77.9 57.7 86.5



#50

Toluene-d8

Concen: 5.308 ug/l

RT: 10.571 min Scan# 1465

Delta R.T. -0.000 min

Lab File: VN086863.D




Acq: 06 Jun 2025 13:17


Tgt Ion: 98 Resp: 52794

Ion Ratio Lower Upper

98 100

100 68.1 53.4 80.0









#51
4-Methyl-2-Pentanone
Concen: 25.283 ug/l
RT: 10.447 min Scan# 1444
Delta R.T. 0.000 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005


Tgt Ion: 43 Resp: 116434
Ion Ratio Lower Upper
43 100
58 40.9 34.2 51.4



#52
Toluene
Concen: 5.178 ug/l
RT: 10.630 min Scan# 1475
Delta R.T. -0.006 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17

Tgt Ion: 92 Resp: 38744
Ion Ratio Lower Upper
92 100
91 169.5 136.5 204.7









#53
t-1,3-Dichloropropene
Concen: 4.896 ug/l
RT: 10.841 min Scan# 1511
Delta R.T. -0.000 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005


Tgt Ion: 75 Resp: 22285
Ion Ratio Lower Upper
75 100
77 31.9 26.4 39.6



#54
cis-1,3-Dichloropropene
Concen: 5.026 ug/l
RT: 10.318 min Scan# 1422
Delta R.T. 0.000 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17

Tgt Ion: 75 Resp: 24477
Ion Ratio Lower Upper
75 100
77 34.5 26.2 39.4
39 46.5 32.3 48.5





#55

1,1,2-Trichloroethane

Concen: 5.234 ug/l

RT: 11.018 min Scan# 1541

Delta R.T. 0.000 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005

Tgt Ion: 97 Resp: 15068





Ion Ratio Lower Upper

97 100

83 82.5 69.7 104.5

85 53.5 45.1 67.7

99 62.6 50.6 75.8



#56

Ethyl methacrylate

Concen: 4.774 ug/l

RT: 10.888 min Scan# 1519

Delta R.T. 0.006 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17




Tgt Ion: 69 Resp: 21893


Ion Ratio Lower Upper

69 100

41 63.5 47.2 70.8

39 29.9 23.0 34.6





#57

1,3-Dichloropropane

Concen: 5.096 ug/l

RT: 11.165 min Scan# 1566

Delta R.T. 0.000 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC005

Tgt Ion: 76 Resp: 25452

Ion Ratio Lower Upper

76 100

78 32.8 25.3 37.9



#58

2-Chloroethyl Vinyl ether

Concen: 23.664 ug/l

RT: 10.165 min Scan# 1396

Delta R.T. 0.000 min

Lab File: VN086863.D




Acq: 06 Jun 2025 13:17


Tgt Ion: 63 Resp: 64722

Ion Ratio Lower Upper

63 100

106 27.9 23.0 34.6









#59
2-Hexanone
Concen: 20.188 ug/l
RT: 11.212 min Scan# 1574
Delta R.T. 0.012 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005


Tgt Ion: 43 Resp: 59884
Ion Ratio Lower Upper
43 100
58 60.5 30.4 91.3



#60
Dibromochloromethane
Concen: 4.962 ug/l
RT: 11.359 min Scan# 1599
Delta R.T. -0.000 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17

Tgt Ion: 129 Resp: 14884
Ion Ratio Lower Upper
129 100
127 75.9 38.6 115.8









#61
1,2-Dibromoethane
Concen: 5.141 ug/l
RT: 11.471 min Scan# 1618
Delta R.T. 0.000 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005


Tgt Ion:107 Resp: 15172
Ion Ratio Lower Upper
107 100
109 94.2 76.3 114.5



#62
4-Bromofluorobenzene
Concen: 5.059 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. -0.000 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17

Tgt Ion: 95 Resp: 18697
Ion Ratio Lower Upper
95 100
174 75.5 0.0 141.8
176 70.2 0.0 132.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 117
Delta R.T. -0.000 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005


Tgt Ion:117 Resp: 371851
Ion Ratio Lower Upper
117 100
82 57.3 44.6 67.0
119 32.3 25.5 38.3



#64
Tetrachloroethene
Concen: 5.236 ug/l
RT: 11.106 min Scan# 1556
Delta R.T. 0.000 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17

Tgt Ion:164 Resp: 12322
Ion Ratio Lower Upper
164 100
166 117.9 101.8 152.8
129 96.6 83.0 124.4
131 88.0 78.2 117.2









#65
Chlorobenzene
Concen: 5.145 ug/l
RT: 11.894 min Scan# 11195
Delta R.T. 0.000 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005


Tgt Ion:112 Resp: 42195
Ion Ratio Lower Upper
112 100
114 31.5 26.2 39.2



#66
1,1,1,2-Tetrachloroethane
Concen: 5.271 ug/l
RT: 11.965 min Scan# 1702
Delta R.T. 0.000 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17

Tgt Ion:131 Resp: 13897
Ion Ratio Lower Upper
131 100
133 91.7 48.3 144.8
119 69.1 34.0 101.8









#67
Ethyl Benzene
Concen: 5.199 ug/l
RT: 11.965 min Scan# 1702
Delta R.T. 0.000 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005


Tgt Ion: 91 Resp: 73443
Ion Ratio Lower Upper
91 100
106 30.3 25.4 38.2



#68
m/p-Xylenes
Concen: 10.330 ug/l
RT: 12.071 min Scan# 1720
Delta R.T. -0.000 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17

Tgt Ion:106 Resp: 55859
Ion Ratio Lower Upper
106 100
91 199.0 159.4 239.0









#69
o-Xylene
Concen: 5.021 ug/l
RT: 12.400 min Scan# 1776
Delta R.T. 0.000 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005


Tgt Ion:106 Resp: 26004
Ion Ratio Lower Upper
106 100
91 218.1 106.1 318.1



#70
Styrene
Concen: 5.008 ug/l
RT: 12.412 min Scan# 1778
Delta R.T. -0.000 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17

Tgt Ion:104 Resp: 44380
Ion Ratio Lower Upper
104 100
78 50.4 38.5 57.7
103 54.4 43.1 64.7





#71

Bromoform

Concen: 5.063 ug/l

RT: 12.582 min Scan# 11

Delta R.T. -0.000 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC005





Tgt Ion:173 Resp: 9890

Ion Ratio Lower Upper

173 100

175 45.8 24.1 72.4

254 0.0 0.0 0.0



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.794 min Scan# 2013

Delta R.T. -0.000 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17




Tgt Ion:152 Resp: 181065


Ion Ratio Lower Upper

152 100

115 61.2 30.1 90.5

150 157.8 0.0 345.0









#73
Isopropylbenzene
Concen: 5.146 ug/l
RT: 12.694 min Scan# 11
Delta R.T. 0.000 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005


Tgt Ion:105 Resp: 67878
Ion Ratio Lower Upper
105 100
120 26.8 13.5 40.4



#74
N-ethyl acetate
Concen: 1.611 ug/l
RT: 12.559 min Scan# 1803
Delta R.T. 0.047 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17

Tgt Ion: 43 Resp: 7051
Ion Ratio Lower Upper
43 100
70 0.0 37.6 56.4#
55 0.0 19.0 28.6#
61 28.5 20.8 31.2





#75

1,1,2,2-Tetrachloroethane

Concen: 5.262 ug/l

RT: 12.935 min Scan# 1867

Delta R.T. -0.006 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC005





Tgt Ion: 83 Resp: 23516

Ion Ratio Lower Upper

83 100

131 8.4 4.9 14.7

85 65.1 32.0 96.0



#76

1,2,3-Trichloropropane

Concen: 7.181 ug/l

RT: 12.994 min Scan# 1877

Delta R.T. -0.000 min

Lab File: VN086863.D




Acq: 06 Jun 2025 13:17


Tgt Ion: 75 Resp: 31004

Ion Ratio Lower Upper

75 100

77 124.8 79.3 237.9









#77
Bromobenzene
Concen: 5.119 ug/l
RT: 12.976 min Scan# 1874
Delta R.T. -0.006 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005


Tgt Ion:156 Resp: 15485
Ion Ratio Lower Upper
156 100
77 249.9 117.7 353.1
158 102.9 48.1 144.4



#78
n-propylbenzene
Concen: 5.226 ug/l
RT: 13.035 min Scan# 1884
Delta R.T. 0.000 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17

Tgt Ion: 91 Resp: 83781
Ion Ratio Lower Upper
91 100
120 22.1 11.4 34.2









#79
2-Chlorotoluene
Concen: 5.128 ug/l
RT: 13.124 min Scan# 1899
Delta R.T. 0.000 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005


Tgt Ion: 91 Resp: 49297
Ion Ratio Lower Upper
91 100
126 32.5 16.2 48.5



#80
1,3,5-Trimethylbenzene
Concen: 5.139 ug/l
RT: 13.171 min Scan# 1907
Delta R.T. -0.000 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17

Tgt Ion: 105 Resp: 55961
Ion Ratio Lower Upper
105 100
120 49.8 24.9 74.6









#81
trans-1,4-Dichloro-2-butene
Concen: 4.939 ug/l
RT: 12.741 min Scan# 1834
Delta R.T. 0.006 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005


Tgt Ion: 75 Resp: 9234
Ion Ratio Lower Upper
75 100
53 88.6 66.6 99.8
89 41.4 36.7 55.1



#82
4-Chlorotoluene
Concen: 5.132 ug/l
RT: 13.223 min Scan# 1916
Delta R.T. -0.000 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17

Tgt Ion: 91 Resp: 49924
Ion Ratio Lower Upper
91 100
126 33.3 16.7 50.1









#83
tert-Butylbenzene
Concen: 5.103 ug/l
RT: 13.435 min Scan# 1952
Delta R.T. -0.000 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005


Tgt Ion:119 Resp: 50874
Ion Ratio Lower Upper
119 100
91 65.3 32.1 96.5
134 24.5 13.4 40.1



#84
1,2,4-Trimethylbenzene
Concen: 5.176 ug/l
RT: 13.482 min Scan# 1960
Delta R.T. 0.000 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17

Tgt Ion:105 Resp: 56525
Ion Ratio Lower Upper
105 100
120 46.6 23.2 69.6









#85
 sec-Butylbenzene
 Concen: 5.161 ug/l
 RT: 13.618 min Scan# 1983
 Delta R.T. 0.000 min
 Lab File: VN086863.D
 Acq: 06 Jun 2025 13:17




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC005


Tgt Ion:105 Resp: 74711
 Ion Ratio Lower Upper
 105 100
 134 19.5 10.0 30.0



#86
 p-Isopropyltoluene
 Concen: 5.182 ug/l
 RT: 13.729 min Scan# 2002
 Delta R.T. -0.000 min
 Lab File: VN086863.D
 Acq: 06 Jun 2025 13:17

Tgt Ion:119 Resp: 62016
 Ion Ratio Lower Upper
 119 100
 134 26.4 13.5 40.4
 91 22.9 11.5 34.4





#87

1,3-Dichlorobenzene

Concen: 5.200 ug/l

RT: 13.735 min Scan# 2003

Delta R.T. -0.000 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC005





Tgt Ion:146 Resp: 30949

Ion Ratio Lower Upper

146 100

111 41.6 20.9 62.8

148 61.1 32.0 96.2



#88

1,4-Dichlorobenzene

Concen: 5.330 ug/l

RT: 13.812 min Scan# 2016

Delta R.T. -0.000 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17




Tgt Ion:146 Resp: 32342


Ion Ratio Lower Upper

146 100

111 46.1 19.9 59.7

148 63.3 31.8 95.4









#89
n-Butylbenzene
Concen: 5.142 ug/l
RT: 14.053 min Scan# 2057
Delta R.T. 0.000 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005


Tgt Ion:	91	Resp:	59601
Ion	Ratio	Lower	Upper
91	100		
92	54.1	27.5	82.4
134	24.9	13.0	39.0



#90
Hexachloroethane
Concen: 4.794 ug/l
RT: 14.335 min Scan# 2105
Delta R.T. 0.000 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17

Tgt Ion:	117	Resp:	9725
Ion	Ratio	Lower	Upper
117	100		
201	70.6	33.2	99.6









#91
1,2-Dichlorobenzene
Concen: 5.228 ug/l
RT: 14.106 min Scan# 2066
Delta R.T. -0.000 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005


Tgt Ion:146 Resp: 29897
Ion Ratio Lower Upper
146 100
111 44.0 21.6 64.6
148 64.3 31.7 95.1



#92
1,2-Dibromo-3-Chloropropane
Concen: 5.363 ug/l
RT: 14.717 min Scan# 2170
Delta R.T. -0.000 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17

Tgt Ion: 75 Resp: 5731
Ion Ratio Lower Upper
75 100
155 76.0 41.1 123.4
157 98.8 52.7 158.1









#93
1,2,4-Trichlorobenzene
Concen: 5.143 ug/l
RT: 15.394 min Scan# 2284
Delta R.T. 0.006 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005


Tgt Ion:180 Resp: 18778
Ion Ratio Lower Upper
180 100
182 92.6 46.9 140.7
145 31.7 16.0 47.9



#94
Hexachlorobutadiene
Concen: 5.199 ug/l
RT: 15.500 min Scan# 2303
Delta R.T. 0.000 min
Lab File: VN086863.D
Acq: 06 Jun 2025 13:17

Tgt Ion:225 Resp: 7072
Ion Ratio Lower Upper
225 100
223 66.4 33.5 100.4
227 61.7 31.1 93.5





#95

Naphthalene

Concen: 4.965 ug/l

RT: 15.641 min Scan# 2327

Delta R.T. 0.000 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC005





Tgt Ion:128 Resp: 67481

Ion Ratio Lower Upper

128 100

127 13.0 10.2 15.2

129 10.3 8.8 13.2



#96

1,2,3-Trichlorobenzene

Concen: 5.034 ug/l

RT: 15.841 min Scan# 2361

Delta R.T. 0.006 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17




Tgt Ion:180 Resp: 18262

Ion Ratio Lower Upper

180 100

182 90.9 47.0 141.0

145 35.3 16.6 49.8



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060625\
 Data File : VN086864.D
 Acq On : 06 Jun 2025 13:40
 Operator : JC\MD
 Sample : VSTDICC020
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC020

Quant Time: Jun 07 01:56:20 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 01:51:02 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	224006	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	418154	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	363172	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	181178	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	63390	21.136	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	42.280%#
35) Dibromofluoromethane	8.177	113	51797	20.902	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	41.800%#
50) Toluene-d8	10.571	98	201268	20.516	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	41.040%#
62) 4-Bromofluorobenzene	12.853	95	74622	20.474	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	40.940%#

Target Compounds					Qvalue	
2) Dichlorodifluoromethane	2.159	85	48254	21.605	ug/l	92
3) Chloromethane	2.401	50	57778	20.033	ug/l	93
4) Vinyl Chloride	2.559	62	61254	20.589	ug/l	99
5) Bromomethane	3.006	94	34022	20.435	ug/l	95
6) Chloroethane	3.159	64	39625	20.619	ug/l	99
7) Trichlorofluoromethane	3.524	101	80978	20.832	ug/l	94
8) Diethyl Ether	3.983	74	34288	20.245	ug/l	98
9) 1,1,2-Trichlorotrifluo...	4.395	101	50458	20.661	ug/l	98
10) Methyl Iodide	4.612	142	65360	20.646	ug/l	98
11) Tert butyl alcohol	5.547	59	87102	107.031	ug/l	99
12) 1,1-Dichloroethene	4.359	96	50485	20.249	ug/l	96
13) Acrolein	4.200	56	22866	88.767	ug/l	97
14) Allyl chloride	5.047	41	82924	20.055	ug/l	98
15) Acrylonitrile	5.742	53	199498	104.881	ug/l	99
16) Acetone	4.447	43	163979	103.099	ug/l	98
17) Carbon Disulfide	4.736	76	138195	20.038	ug/l	97
18) Methyl Acetate	5.047	43	96547	20.830	ug/l	99
19) Methyl tert-butyl Ether	5.818	73	183780	20.358	ug/l	98
20) Methylene Chloride	5.294	84	57611	19.348	ug/l	98
21) trans-1,2-Dichloroethene	5.806	96	55643	20.059	ug/l	94
22) Diisopropyl ether	6.689	45	182472	20.935	ug/l	98
23) Vinyl Acetate	6.624	43	768556	104.365	ug/l	99
24) 1,1-Dichloroethane	6.583	63	103578	20.650	ug/l	99
25) 2-Butanone	7.494	43	270477	104.621	ug/l	100
26) 2,2-Dichloropropane	7.506	77	69692	17.862	ug/l	99
27) cis-1,2-Dichloroethene	7.500	96	68292	20.585	ug/l	98
28) Bromochloromethane	7.824	49	55196	22.381	ug/l	99
29) Tetrahydrofuran	7.853	42	178571	106.031	ug/l	99
30) Chloroform	7.977	83	102615	20.485	ug/l	93
31) Cyclohexane	8.271	56	100026	20.563	ug/l	98
32) 1,1,1-Trichloroethane	8.177	97	86810	20.376	ug/l	98
36) 1,1-Dichloropropene	8.383	75	73252	19.838	ug/l	100
37) Ethyl Acetate	7.571	43	96467	20.522	ug/l	98
38) Carbon Tetrachloride	8.377	117	72628	20.034	ug/l	97
39) Methylcyclohexane	9.606	83	98425	19.455	ug/l	99
40) Benzene	8.612	78	241444	19.996	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060625\
 Data File : VN086864.D
 Acq On : 06 Jun 2025 13:40
 Operator : JC\MD
 Sample : VSTDICC020
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC020

Quant Time: Jun 07 01:56:20 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 01:51:02 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.788	41	53244	20.169	ug/l	98
42) 1,2-Dichloroethane	8.677	62	74238	20.270	ug/l	100
43) Isopropyl Acetate	8.694	43	151601	20.094	ug/l	99
44) Trichloroethene	9.359	130	57110	19.946	ug/l	96
45) 1,2-Dichloropropane	9.630	63	59293	20.184	ug/l	97
46) Dibromomethane	9.712	93	40290	20.648	ug/l	98
47) Bromodichloromethane	9.894	83	80353	20.015	ug/l	98
48) Methyl methacrylate	9.682	41	70457	20.290	ug/l	97
49) 1,4-Dioxane	9.706	88	28231	446.886	ug/l #	98
51) 4-Methyl-2-Pentanone	10.447	43	481741	106.066	ug/l	99
52) Toluene	10.635	92	148000	20.056	ug/l	98
53) t-1,3-Dichloropropene	10.841	75	87683	19.532	ug/l	99
54) cis-1,3-Dichloropropene	10.318	75	94315	19.636	ug/l	99
55) 1,1,2-Trichloroethane	11.018	97	57195	20.143	ug/l	99
56) Ethyl methacrylate	10.882	69	94829	20.968	ug/l	98
57) 1,3-Dichloropropane	11.165	76	98205	19.938	ug/l	98
58) 2-Chloroethyl Vinyl ether	10.165	63	278986	103.426	ug/l	99
59) 2-Hexanone	11.206	43	303686	103.803	ug/l	95
60) Dibromochloromethane	11.359	129	59961	20.268	ug/l	99
61) 1,2-Dibromoethane	11.471	107	59139	20.320	ug/l	97
64) Tetrachloroethene	11.106	164	45374	19.742	ug/l	96
65) Chlorobenzene	11.894	112	160803	20.076	ug/l	97
66) 1,1,1,2-Tetrachloroethane	11.959	131	52069	20.223	ug/l	99
67) Ethyl Benzene	11.965	91	277087	20.083	ug/l	98
68) m/p-Xylenes	12.071	106	215378	40.780	ug/l	100
69) o-Xylene	12.400	106	102050	20.175	ug/l	98
70) Styrene	12.412	104	178061	20.572	ug/l	99
71) Bromoform	12.576	173	40070	21.004	ug/l #	99
73) Isopropylbenzene	12.694	105	264442	20.035	ug/l	100
74) N-amyl acetate	12.523	43	86003	19.633	ug/l	98
75) 1,1,2,2-Tetrachloroethane	12.941	83	92264	20.634	ug/l	99
76) 1,2,3-Trichloropropane	12.994	75	122897	28.448	ug/l	69
77) Bromobenzene	12.982	156	60877	20.111	ug/l	98
78) n-propylbenzene	13.035	91	322436	20.102	ug/l	100
79) 2-Chlorotoluene	13.123	91	193774	20.144	ug/l	100
80) 1,3,5-Trimethylbenzene	13.170	105	219937	20.183	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.735	75	37808	20.209	ug/l	92
82) 4-Chlorotoluene	13.223	91	193544	19.885	ug/l	99
83) tert-Butylbenzene	13.435	119	198303	19.880	ug/l	99
84) 1,2,4-Trimethylbenzene	13.482	105	221990	20.315	ug/l	100
85) sec-Butylbenzene	13.617	105	292563	20.196	ug/l	100
86) p-Isopropyltoluene	13.729	119	240682	20.099	ug/l	100
87) 1,3-Dichlorobenzene	13.735	146	120099	20.166	ug/l	99
88) 1,4-Dichlorobenzene	13.812	146	120102	19.780	ug/l	97
89) n-Butylbenzene	14.053	91	227536	19.617	ug/l	100
90) Hexachloroethane	14.335	117	41202	20.299	ug/l	100
91) 1,2-Dichlorobenzene	14.106	146	115645	20.212	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.717	75	21025	19.661	ug/l	98
93) 1,2,4-Trichlorobenzene	15.388	180	71826	19.659	ug/l	98
94) Hexachlorobutadiene	15.500	225	27887	20.486	ug/l	97
95) Naphthalene	15.641	128	272584	20.043	ug/l	100
96) 1,2,3-Trichlorobenzene	15.841	180	71929	19.815	ug/l	99


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060625\
Data File : VN086864.D
Acq On : 06 Jun 2025 13:40
Operator : JC\MD
Sample : VSTDICC020
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Quant Time: Jun 07 01:56:20 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 01:51:02 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.235 min Scan# 11
Delta R.T. -0.001 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020


Tgt Ion:168 Resp: 224006
Ion Ratio Lower Upper
168 100
99 61.1 49.1 73.7



#2
Dichlorodifluoromethane
Concen: 21.605 ug/l
RT: 2.159 min Scan# 35
Delta R.T. 0.006 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40

Tgt Ion: 85 Resp: 48254
Ion Ratio Lower Upper
85 100
87 37.1 16.2 48.6









#3
Chloromethane
Concen: 20.033 ug/l
RT: 2.401 min Scan# 75
Delta R.T. 0.006 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020


Tgt Ion: 50 Resp: 57778
Ion Ratio Lower Upper
50 100
52 35.8 25.7 38.5



#4
Vinyl Chloride
Concen: 20.589 ug/l
RT: 2.559 min Scan# 103
Delta R.T. 0.006 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40

Tgt Ion: 62 Resp: 61254
Ion Ratio Lower Upper
62 100
64 31.7 26.0 39.0









#5
Bromomethane
Concen: 20.435 ug/l
RT: 3.006 min Scan# 11
Delta R.T. 0.012 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020


Tgt Ion: 94 Resp: 34022
Ion Ratio Lower Upper
94 100
96 92.9 77.9 116.9



#6
Chloroethane
Concen: 20.619 ug/l
RT: 3.159 min Scan# 205
Delta R.T. 0.006 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40

Tgt Ion: 64 Resp: 39625
Ion Ratio Lower Upper
64 100
66 30.4 24.6 36.8









#7
Trichlorofluoromethane
Concen: 20.832 ug/l
RT: 3.524 min Scan# 20
Delta R.T. -0.000 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020


Tgt Ion:101 Resp: 80978
Ion Ratio Lower Upper
101 100
103 60.4 52.2 78.2



#8
Diethyl Ether
Concen: 20.245 ug/l
RT: 3.983 min Scan# 345
Delta R.T. -0.000 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40

Tgt Ion: 74 Resp: 34288
Ion Ratio Lower Upper
74 100
45 92.9 45.5 136.5









#9
1,1,2-Trichlorotrifluoroethane
Concen: 20.661 ug/l
RT: 4.395 min Scan# 415
Delta R.T. -0.000 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020


Tgt Ion:101 Resp: 50458
Ion Ratio Lower Upper
101 100
85 41.7 33.8 50.8
151 71.3 58.7 88.1



#10
Methyl Iodide
Concen: 20.646 ug/l
RT: 4.612 min Scan# 452
Delta R.T. 0.006 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40

Tgt Ion:142 Resp: 65360
Ion Ratio Lower Upper
142 100
127 39.2 30.9 46.3
141 14.3 10.4 15.6









#11
Tert butyl alcohol
Concen: 107.031 ug/l
RT: 5.547 min Scan# 611
Delta R.T. -0.000 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020


Tgt Ion: 59 Resp: 87102
Ion Ratio Lower Upper
59 100
57 9.9 8.1 12.1



#12
1,1-Dichloroethene
Concen: 20.249 ug/l
RT: 4.359 min Scan# 409
Delta R.T. -0.000 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40

Tgt Ion: 96 Resp: 50485
Ion Ratio Lower Upper
96 100
61 153.1 125.8 188.6
98 61.0 52.5 78.7





#13

Acrolein

Concen: 88.767 ug/l

RT: 4.200 min Scan# 382

Delta R.T. -0.000 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40

Instrument :

MSVOA_N

ClientSampleId :





VSTDICC020

Tgt Ion: 56 Resp: 22866

Ion Ratio Lower Upper

56 100

55 71.6 55.6 83.4



#14

Allyl chloride

Concen: 20.055 ug/l

RT: 5.047 min Scan# 526

Delta R.T. 0.006 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40




Tgt Ion: 41 Resp: 82924


Ion Ratio Lower Upper

41 100

39 62.6 51.3 76.9

76 38.1 31.4 47.0









#15
Acrylonitrile
Concen: 104.881 ug/l
RT: 5.742 min Scan# 644
Delta R.T. -0.000 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020


Tgt Ion: 53 Resp: 199498
Ion Ratio Lower Upper
53 100
52 80.3 65.0 97.6
51 33.7 27.2 40.8



#16
Acetone
Concen: 103.099 ug/l
RT: 4.447 min Scan# 424
Delta R.T. -0.000 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40

Tgt Ion: 43 Resp: 163979
Ion Ratio Lower Upper
43 100
58 35.9 28.0 42.0









#17
Carbon Disulfide
Concen: 20.038 ug/l
RT: 4.736 min Scan# 473
Delta R.T. 0.006 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020


Tgt Ion: 76 Resp: 138195
Ion Ratio Lower Upper
76 100
78 7.8 7.3 10.9



#18
Methyl Acetate
Concen: 20.830 ug/l
RT: 5.047 min Scan# 526
Delta R.T. 0.006 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40

Tgt Ion: 43 Resp: 96547
Ion Ratio Lower Upper
43 100
74 26.2 20.7 31.1









#19
Methyl tert-butyl Ether
Concen: 20.358 ug/l
RT: 5.818 min Scan# 61
Delta R.T. 0.006 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020


Tgt Ion: 73 Resp: 183780
Ion Ratio Lower Upper
73 100
57 22.9 17.4 26.2



#20
Methylene Chloride
Concen: 19.348 ug/l
RT: 5.294 min Scan# 568
Delta R.T. -0.000 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40

Tgt Ion: 84 Resp: 57611
Ion Ratio Lower Upper
84 100
49 114.5 90.5 135.7
51 35.9 28.5 42.7
86 63.0 48.1 72.1









#21
trans-1,2-Dichloroethene
Concen: 20.059 ug/l
RT: 5.806 min Scan# 611
Delta R.T. -0.000 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020


Tgt Ion: 96 Resp: 55643
Ion Ratio Lower Upper
96 100
61 126.9 109.2 163.8
98 61.8 49.8 74.8



#22
Diisopropyl ether
Concen: 20.935 ug/l
RT: 6.689 min Scan# 805
Delta R.T. -0.000 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40

Tgt Ion: 45 Resp: 182472
Ion Ratio Lower Upper
45 100
43 50.7 41.9 62.9
87 27.0 22.4 33.6
59 13.1 9.8 14.6









#23
 Vinyl Acetate
 Concen: 104.365 ug/l
 RT: 6.624 min Scan# 794
 Delta R.T. 0.006 min
 Lab File: VN086864.D
 Acq: 06 Jun 2025 13:40




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC020


Tgt Ion: 43 Resp: 768556
 Ion Ratio Lower Upper
 43 100
 86 11.4 9.5 14.3



#24
 1,1-Dichloroethane
 Concen: 20.650 ug/l
 RT: 6.583 min Scan# 787
 Delta R.T. -0.000 min
 Lab File: VN086864.D
 Acq: 06 Jun 2025 13:40

Tgt Ion: 63 Resp: 103578
 Ion Ratio Lower Upper
 63 100
 98 7.5 3.6 10.8
 100 4.7 2.6 7.8









#25
2-Butanone
Concen: 104.621 ug/l
RT: 7.494 min Scan# 942
Delta R.T. -0.000 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020


Tgt Ion: 43 Resp: 270477
Ion Ratio Lower Upper
43 100
72 28.6 23.0 34.6



#26
2,2-Dichloropropane
Concen: 17.862 ug/l
RT: 7.506 min Scan# 944
Delta R.T. 0.006 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40

Tgt Ion: 77 Resp: 69692
Ion Ratio Lower Upper
77 100
97 23.7 11.6 34.7









#27
 cis-1,2-Dichloroethene
 Concen: 20.585 ug/l
 RT: 7.500 min Scan# 943
 Delta R.T. -0.000 min
 Lab File: VN086864.D
 Acq: 06 Jun 2025 13:40




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC020


Tgt Ion: 96 Resp: 68292
 Ion Ratio Lower Upper
 96 100
 61 135.5 0.0 278.0
 98 63.0 0.0 128.2



#28
 Bromochloromethane
 Concen: 22.381 ug/l
 RT: 7.824 min Scan# 998
 Delta R.T. -0.000 min
 Lab File: VN086864.D
 Acq: 06 Jun 2025 13:40

Tgt Ion: 49 Resp: 55196
 Ion Ratio Lower Upper
 49 100
 129 1.9 0.0 3.8
 130 77.1 62.2 93.2









#29
Tetrahydrofuran
Concen: 106.031 ug/l
RT: 7.853 min Scan# 1003
Delta R.T. -0.000 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020


Tgt Ion: 42 Resp: 178571
Ion Ratio Lower Upper
42 100
72 46.6 37.8 56.6
71 43.5 35.6 53.4



#30
Chloroform
Concen: 20.485 ug/l
RT: 7.977 min Scan# 1024
Delta R.T. -0.000 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40

Tgt Ion: 83 Resp: 102615
Ion Ratio Lower Upper
83 100
85 60.7 52.9 79.3









#31
Cyclohexane
Concen: 20.563 ug/l
RT: 8.271 min Scan# 1074
Delta R.T. -0.000 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020


Tgt Ion: 56 Resp: 100026
Ion Ratio Lower Upper
56 100
69 29.0 25.4 38.0
84 85.6 68.8 103.2



#32
1,1,1-Trichloroethane
Concen: 20.376 ug/l
RT: 8.177 min Scan# 1058
Delta R.T. -0.000 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40

Tgt Ion: 97 Resp: 86810
Ion Ratio Lower Upper
97 100
99 63.8 52.0 78.0
61 43.3 35.9 53.9





#33

1,2-Dichloroethane-d4

Concen: 21.136 ug/l

RT: 8.588 min Scan# 1128

Delta R.T. -0.000 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC020

Tgt Ion: 65 Resp: 63390

Ion Ratio Lower Upper

65 100

67 55.3 0.0 105.6



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.106 min Scan# 1216

Delta R.T. -0.000 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40




Tgt Ion: 114 Resp: 418154


Ion Ratio Lower Upper

114 100

63 20.6 0.0 39.6

88 15.8 0.0 30.2





#35

Dibromofluoromethane

Concen: 20.902 ug/l

RT: 8.177 min Scan# 1057

Delta R.T. 0.006 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020





Tgt Ion:113 Resp: 51797

Ion Ratio Lower Upper

113 100

111 105.1 84.2 126.2

192 17.7 14.2 21.4



#36

1,1-Dichloropropene

Concen: 19.838 ug/l

RT: 8.383 min Scan# 1093

Delta R.T. 0.006 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40




Tgt Ion: 75 Resp: 73252


Ion Ratio Lower Upper

75 100

110 35.6 17.6 52.9

77 31.8 25.3 37.9









#37
Ethyl Acetate
Concen: 20.522 ug/l
RT: 7.571 min Scan# 91
Delta R.T. -0.006 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020


Tgt Ion: 43 Resp: 96467
Ion Ratio Lower Upper
43 100
61 13.9 12.1 18.1
70 12.6 9.8 14.6



#38
Carbon Tetrachloride
Concen: 20.034 ug/l
RT: 8.377 min Scan# 1092
Delta R.T. 0.006 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40

Tgt Ion: 117 Resp: 72628
Ion Ratio Lower Upper
117 100
119 91.3 75.0 112.6
121 28.1 24.9 37.3









#39
Methylcyclohexane
Concen: 19.455 ug/l
RT: 9.606 min Scan# 1130
Delta R.T. -0.000 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020


Tgt Ion: 83 Resp: 98425
Ion Ratio Lower Upper
83 100
55 75.7 61.0 91.6
98 47.4 38.9 58.3



#40
Benzene
Concen: 19.996 ug/l
RT: 8.612 min Scan# 1132
Delta R.T. -0.000 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40

Tgt Ion: 78 Resp: 241444
Ion Ratio Lower Upper
78 100
77 23.8 18.7 28.1









#41
Methacrylonitrile
Concen: 20.169 ug/l
RT: 7.788 min Scan# 992
Delta R.T. -0.000 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020


Tgt Ion:	41	Resp:	53244
Ion	Ratio	Lower	Upper
41	100		
39	49.5	41.2	61.8
67	77.3	61.4	92.2
52	29.6	22.7	34.1



#42
1,2-Dichloroethane
Concen: 20.270 ug/l
RT: 8.677 min Scan# 1143
Delta R.T. -0.000 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40

Tgt Ion:	62	Resp:	74238
Ion	Ratio	Lower	Upper
62	100		
98	10.6	0.0	21.2









#43
Isopropyl Acetate
Concen: 20.094 ug/l
RT: 8.694 min Scan# 1146
Delta R.T. -0.000 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020


Tgt Ion: 43 Resp: 151601
Ion Ratio Lower Upper
43 100
61 27.9 22.1 33.1
87 14.4 11.8 17.6



#44
Trichloroethene
Concen: 19.946 ug/l
RT: 9.359 min Scan# 1259
Delta R.T. -0.000 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40

Tgt Ion: 130 Resp: 57110
Ion Ratio Lower Upper
130 100
95 102.1 0.0 196.2





#45

1,2-Dichloropropane

Concen: 20.184 ug/l

RT: 9.630 min Scan# 1305

Delta R.T. -0.000 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40

Instrument :

MSVOA_N

ClientSampleId :





VSTDICC020

Tgt Ion: 63 Resp: 59293

Ion Ratio Lower Upper

63 100

65 30.4 25.5 38.3



#46

Dibromomethane

Concen: 20.648 ug/l

RT: 9.712 min Scan# 1319

Delta R.T. -0.006 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40




Tgt Ion: 93 Resp: 40290


Ion Ratio Lower Upper

93 100

95 84.1 67.1 100.7

174 90.1 74.2 111.4









#47
Bromodichloromethane
Concen: 20.015 ug/l
RT: 9.894 min Scan# 1350
Delta R.T. -0.000 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020


Tgt Ion: 83 Resp: 80353
Ion Ratio Lower Upper
83 100
85 62.3 50.9 76.3
127 7.6 6.6 10.0



#48
Methyl methacrylate
Concen: 20.290 ug/l
RT: 9.682 min Scan# 1314
Delta R.T. -0.000 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40

Tgt Ion: 41 Resp: 70457
Ion Ratio Lower Upper
41 100
69 89.4 74.6 112.0
39 50.3 40.2 60.4





#49

1,4-Dioxane

Concen: 446.886 ug/l

RT: 9.706 min Scan# 1

Delta R.T. 0.006 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC020





Tgt Ion: 88 Resp: 28231

Ion Ratio Lower Upper

88 100

43 19.7 0.0 0.0#

58 70.1 57.7 86.5



#50

Toluene-d8

Concen: 20.516 ug/l

RT: 10.571 min Scan# 1465

Delta R.T. -0.000 min

Lab File: VN086864.D




Acq: 06 Jun 2025 13:40


Tgt Ion: 98 Resp: 201268

Ion Ratio Lower Upper

98 100

100 67.8 53.4 80.0









#51
4-Methyl-2-Pentanone
Concen: 106.066 ug/l
RT: 10.447 min Scan# 1444
Delta R.T. -0.000 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020


Tgt Ion: 43 Resp: 481741
Ion Ratio Lower Upper
43 100
58 42.0 34.2 51.4



#52
Toluene
Concen: 20.056 ug/l
RT: 10.635 min Scan# 1476
Delta R.T. -0.000 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40

Tgt Ion: 92 Resp: 148000
Ion Ratio Lower Upper
92 100
91 172.9 136.5 204.7









#53
t-1,3-Dichloropropene
Concen: 19.532 ug/l
RT: 10.841 min Scan# 1511
Delta R.T. -0.000 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020


Tgt Ion: 75 Resp: 87683
Ion Ratio Lower Upper
75 100
77 32.7 26.4 39.6



#54
cis-1,3-Dichloropropene
Concen: 19.636 ug/l
RT: 10.318 min Scan# 1422
Delta R.T. -0.000 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40

Tgt Ion: 75 Resp: 94315
Ion Ratio Lower Upper
75 100
77 32.5 26.2 39.4
39 41.3 32.3 48.5





#55

1,1,2-Trichloroethane

Concen: 20.143 ug/l

RT: 11.018 min Scan# 1541

Delta R.T. -0.000 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020

Tgt Ion: 97 Resp: 57195





Ion Ratio Lower Upper

97 100

83 87.1 69.7 104.5

85 54.0 45.1 67.7

99 62.2 50.6 75.8



#56

Ethyl methacrylate

Concen: 20.968 ug/l

RT: 10.882 min Scan# 1518

Delta R.T. -0.000 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40




Tgt Ion: 69 Resp: 94829


Ion Ratio Lower Upper

69 100

41 58.9 47.2 70.8

39 25.9 23.0 34.6





#57

1,3-Dichloropropane

Concen: 19.938 ug/l

RT: 11.165 min Scan# 1566

Delta R.T. -0.000 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40

Instrument :

MSVOA_N

ClientSampleId :





VSTDICC020

Tgt Ion: 76 Resp: 98205

Ion Ratio Lower Upper

76 100

78 32.4 25.3 37.9



#58

2-Chloroethyl Vinyl ether

Concen: 103.426 ug/l

RT: 10.165 min Scan# 1396

Delta R.T. -0.000 min

Lab File: VN086864.D




Acq: 06 Jun 2025 13:40


Tgt Ion: 63 Resp: 278986

Ion Ratio Lower Upper

63 100

106 28.1 23.0 34.6









#59
2-Hexanone
Concen: 103.803 ug/l
RT: 11.206 min Scan# 1572
Delta R.T. 0.006 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020


Tgt Ion: 43 Resp: 303686
Ion Ratio Lower Upper
43 100
58 64.8 30.4 91.3



#60
Dibromochloromethane
Concen: 20.268 ug/l
RT: 11.359 min Scan# 1599
Delta R.T. -0.000 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40

Tgt Ion:129 Resp: 59961
Ion Ratio Lower Upper
129 100
127 76.3 38.6 115.8









#61
1,2-Dibromoethane
Concen: 20.320 ug/l
RT: 11.471 min Scan# 1618
Delta R.T. -0.000 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020


Tgt Ion:107 Resp: 59139
Ion Ratio Lower Upper
107 100
109 93.0 76.3 114.5



#62
4-Bromofluorobenzene
Concen: 20.474 ug/l
RT: 12.853 min Scan# 1853
Delta R.T. 0.006 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40

Tgt Ion: 95 Resp: 74622
Ion Ratio Lower Upper
95 100
174 71.7 0.0 141.8
176 68.0 0.0 132.6





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.865 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020





Tgt Ion:117 Resp: 363172

Ion Ratio Lower Upper

117 100

82 58.0 44.6 67.0

119 32.6 25.5 38.3



#64

Tetrachloroethene

Concen: 19.742 ug/l

RT: 11.106 min Scan# 1556

Delta R.T. -0.000 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40

Tgt Ion:164 Resp: 45374




Ion Ratio Lower Upper


164 100

166 124.4 101.8 152.8

129 100.4 83.0 124.4

131 92.1 78.2 117.2









#65
Chlorobenzene
Concen: 20.076 ug/l
RT: 11.894 min Scan# 11
Delta R.T. -0.000 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020


Tgt Ion:112 Resp: 160803
Ion Ratio Lower Upper
112 100
114 30.8 26.2 39.2



#66
1,1,1,2-Tetrachloroethane
Concen: 20.223 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. -0.006 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40

Tgt Ion:131 Resp: 52069
Ion Ratio Lower Upper
131 100
133 96.3 48.3 144.8
119 69.6 34.0 101.8









#67
Ethyl Benzene
Concen: 20.083 ug/l
RT: 11.965 min Scan# 1702
Delta R.T. -0.000 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020


Tgt Ion: 91 Resp: 277087
Ion Ratio Lower Upper
91 100
106 30.5 25.4 38.2



#68
m/p-Xylenes
Concen: 40.780 ug/l
RT: 12.071 min Scan# 1720
Delta R.T. -0.000 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40

Tgt Ion: 106 Resp: 215378
Ion Ratio Lower Upper
106 100
91 199.8 159.4 239.0





#69

o-Xylene

Concen: 20.175 ug/l

RT: 12.400 min Scan# 1776

Delta R.T. -0.000 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC020

Tgt Ion:106 Resp: 102050

Ion Ratio Lower Upper

106 100

91 215.9 106.1 318.1



#70

Styrene

Concen: 20.572 ug/l

RT: 12.412 min Scan# 1778

Delta R.T. -0.000 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40




Tgt Ion:104 Resp: 178061


Ion Ratio Lower Upper

104 100

78 48.3 38.5 57.7

103 54.4 43.1 64.7





#71

Bromoform

Concen: 21.004 ug/l

RT: 12.576 min Scan# 11

Delta R.T. -0.006 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC020





Tgt Ion:173 Resp: 40070

Ion Ratio Lower Upper

173 100

175 47.6 24.1 72.4

254 0.0 0.0 0.0



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.788 min Scan# 2012

Delta R.T. -0.006 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40




Tgt Ion:152 Resp: 181178


Ion Ratio Lower Upper

152 100

115 61.4 30.1 90.5

150 164.5 0.0 345.0









#73
Isopropylbenzene
Concen: 20.035 ug/l
RT: 12.694 min Scan# 11
Delta R.T. -0.000 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020


Tgt Ion:105 Resp: 264442
Ion Ratio Lower Upper
105 100
120 27.0 13.5 40.4



#74
N-ethyl acetate
Concen: 19.633 ug/l
RT: 12.523 min Scan# 1797
Delta R.T. 0.012 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40

Tgt Ion: 43 Resp: 86003
Ion Ratio Lower Upper
43 100
70 47.7 37.6 56.4
55 26.6 19.0 28.6
61 25.9 20.8 31.2









#75
 1,1,2-Tetrachloroethane
 Concen: 20.634 ug/l
 RT: 12.941 min Scan# 1868
 Delta R.T. -0.000 min
 Lab File: VN086864.D
 Acq: 06 Jun 2025 13:40




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC020


Tgt Ion: 83 Resp: 92264
 Ion Ratio Lower Upper
 83 100
 131 9.8 4.9 14.7
 85 64.5 32.0 96.0



#76
 1,2,3-Trichloropropane
 Concen: 28.448 ug/l
 RT: 12.994 min Scan# 1877
 Delta R.T. -0.000 min
 Lab File: VN086864.D
 Acq: 06 Jun 2025 13:40

Tgt Ion: 75 Resp: 122897
 Ion Ratio Lower Upper
 75 100
 77 117.5 79.3 237.9





#77

Bromobenzene

Concen: 20.111 ug/l

RT: 12.982 min Scan# 1875

Delta R.T. -0.000 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC020





Tgt Ion:156 Resp: 60877

Ion Ratio Lower Upper

156 100

77 237.3 117.7 353.1

158 99.0 48.1 144.4



#78

n-propylbenzene

Concen: 20.102 ug/l

RT: 13.035 min Scan# 1884

Delta R.T. -0.000 min

Lab File: VN086864.D




Acq: 06 Jun 2025 13:40


Tgt Ion: 91 Resp: 322436

Ion Ratio Lower Upper

91 100

120 22.9 11.4 34.2









#79
2-Chlorotoluene
Concen: 20.144 ug/l
RT: 13.123 min Scan# 1899
Delta R.T. -0.000 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020


Tgt Ion: 91 Resp: 193774
Ion Ratio Lower Upper
91 100
126 32.3 16.2 48.5



#80
1,3,5-Trimethylbenzene
Concen: 20.183 ug/l
RT: 13.170 min Scan# 1907
Delta R.T. -0.000 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40

Tgt Ion:105 Resp: 219937
Ion Ratio Lower Upper
105 100
120 49.6 24.9 74.6









#81
trans-1,4-Dichloro-2-butene
Concen: 20.209 ug/l
RT: 12.735 min Scan# 1833
Delta R.T. -0.000 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020


Tgt Ion: 75 Resp: 37808
Ion Ratio Lower Upper
75 100
53 77.3 66.6 99.8
89 39.6 36.7 55.1



#82
4-Chlorotoluene
Concen: 19.885 ug/l
RT: 13.223 min Scan# 1916
Delta R.T. -0.000 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40

Tgt Ion: 91 Resp: 193544
Ion Ratio Lower Upper
91 100
126 33.7 16.7 50.1









#83
 tert-Butylbenzene
 Concen: 19.880 ug/l
 RT: 13.435 min Scan# 1952
 Delta R.T. -0.000 min
 Lab File: VN086864.D
 Acq: 06 Jun 2025 13:40




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC020


Tgt Ion:119 Resp: 198303
 Ion Ratio Lower Upper
 119 100
 91 65.2 32.1 96.5
 134 26.5 13.4 40.1



#84
 1,2,4-Trimethylbenzene
 Concen: 20.315 ug/l
 RT: 13.482 min Scan# 1960
 Delta R.T. -0.000 min
 Lab File: VN086864.D
 Acq: 06 Jun 2025 13:40

Tgt Ion:105 Resp: 221990
 Ion Ratio Lower Upper
 105 100
 120 46.1 23.2 69.6









#85
 sec-Butylbenzene
 Concen: 20.196 ug/l
 RT: 13.617 min Scan# 1983
 Delta R.T. -0.000 min
 Lab File: VN086864.D
 Acq: 06 Jun 2025 13:40




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC020


Tgt Ion:105 Resp: 292563
 Ion Ratio Lower Upper
 105 100
 134 20.0 10.0 30.0



#86
 p-Isopropyltoluene
 Concen: 20.099 ug/l
 RT: 13.729 min Scan# 2002
 Delta R.T. -0.000 min
 Lab File: VN086864.D
 Acq: 06 Jun 2025 13:40

Tgt Ion:119 Resp: 240682
 Ion Ratio Lower Upper
 119 100
 134 27.1 13.5 40.4
 91 22.7 11.5 34.4









#87
 1,3-Dichlorobenzene
 Concen: 20.166 ug/l
 RT: 13.735 min Scan# 2003
 Delta R.T. -0.000 min
 Lab File: VN086864.D
 Acq: 06 Jun 2025 13:40




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC020


Tgt Ion:146 Resp: 120099
 Ion Ratio Lower Upper
 146 100
 111 41.5 20.9 62.8
 148 62.5 32.0 96.2



#88
 1,4-Dichlorobenzene
 Concen: 19.780 ug/l
 RT: 13.812 min Scan# 2016
 Delta R.T. -0.000 min
 Lab File: VN086864.D
 Acq: 06 Jun 2025 13:40

Tgt Ion:146 Resp: 120102
 Ion Ratio Lower Upper
 146 100
 111 42.5 19.9 59.7
 148 64.7 31.8 95.4









#89
n-Butylbenzene
Concen: 19.617 ug/l
RT: 14.053 min Scan# 2057
Delta R.T. -0.000 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020


Tgt Ion:	91	Resp:	227536
Ion	Ratio	Lower	Upper
91	100		
92	54.5	27.5	82.4
134	26.2	13.0	39.0



#90
Hexachloroethane
Concen: 20.299 ug/l
RT: 14.335 min Scan# 2105
Delta R.T. -0.000 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40

Tgt Ion:	117	Resp:	41202
Ion	Ratio	Lower	Upper
117	100		
201	66.2	33.2	99.6









#91
1,2-Dichlorobenzene
Concen: 20.212 ug/l
RT: 14.106 min Scan# 2066
Delta R.T. -0.000 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020


Tgt Ion:146 Resp: 115645
Ion Ratio Lower Upper
146 100
111 42.9 21.6 64.6
148 63.9 31.7 95.1



#92
1,2-Dibromo-3-Chloropropane
Concen: 19.661 ug/l
RT: 14.717 min Scan# 2170
Delta R.T. -0.000 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40

Tgt Ion: 75 Resp: 21025
Ion Ratio Lower Upper
75 100
155 81.8 41.1 123.4
157 109.0 52.7 158.1









#93
1,2,4-Trichlorobenzene
Concen: 19.659 ug/l
RT: 15.388 min Scan# 21
Delta R.T. -0.000 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020


Tgt Ion:180 Resp: 71826
Ion Ratio Lower Upper
180 100
182 95.4 46.9 140.7
145 33.3 16.0 47.9



#94
Hexachlorobutadiene
Concen: 20.486 ug/l
RT: 15.500 min Scan# 2303
Delta R.T. -0.000 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40

Tgt Ion:225 Resp: 27887
Ion Ratio Lower Upper
225 100
223 63.0 33.5 100.4
227 63.3 31.1 93.5









#95
Naphthalene
Concen: 20.043 ug/l
RT: 15.641 min Scan# 2327
Delta R.T. -0.000 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Tgt Ion:128 Resp: 272584
Ion Ratio Lower Upper
128 100
127 12.7 10.2 15.2
129 11.0 8.8 13.2



#96
1,2,3-Trichlorobenzene
Concen: 19.815 ug/l
RT: 15.841 min Scan# 2361
Delta R.T. 0.006 min
Lab File: VN086864.D
Acq: 06 Jun 2025 13:40

Tgt Ion:180 Resp: 71929
Ion Ratio Lower Upper
180 100
182 95.3 47.0 141.0
145 33.1 16.6 49.8



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060625\
 Data File : VN086865.D
 Acq On : 06 Jun 2025 14:03
 Operator : JC\MD
 Sample : VSTDICCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICCC050

Quant Time: Jun 07 01:57:17 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 01:51:02 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	207354	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	378587	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	332766	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	167532	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	103736	37.366	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	74.740%
35) Dibromofluoromethane	8.171	113	83058	37.020	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	74.040%#
50) Toluene-d8	10.571	98	326105	36.716	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	73.440%#
62) 4-Bromofluorobenzene	12.847	95	122965	37.264	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	74.520%#
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	2.154	85	103958	50.283	ug/l	100
3) Chloromethane	2.395	50	123866	46.397	ug/l	100
4) Vinyl Chloride	2.554	62	132783	48.215	ug/l	100
5) Bromomethane	2.995	94	73974	47.999	ug/l	100
6) Chloroethane	3.154	64	84505	47.504	ug/l	100
7) Trichlorofluoromethane	3.524	101	172860	48.041	ug/l	100
8) Diethyl Ether	3.983	74	75193	47.963	ug/l	100
9) 1,1,2-Trichlorotrifluo...	4.395	101	107695	47.640	ug/l	100
10) Methyl Iodide	4.606	142	141678	48.346	ug/l	100
11) Tert butyl alcohol	5.548	59	182315	242.020	ug/l	100
12) 1,1-Dichloroethene	4.359	96	110549	47.900	ug/l	100
13) Acrolein	4.201	56	47026	197.217	ug/l	100
14) Allyl chloride	5.042	41	179313	46.848	ug/l	100
15) Acrylonitrile	5.742	53	421941	239.638	ug/l	100
16) Acetone	4.448	43	333857	226.765	ug/l	100
17) Carbon Disulfide	4.730	76	295753	46.328	ug/l	100
18) Methyl Acetate	5.042	43	204538	47.673	ug/l	100
19) Methyl tert-butyl Ether	5.812	73	400775	47.960	ug/l	100
20) Methylene Chloride	5.295	84	125498	45.531	ug/l	100
21) trans-1,2-Dichloroethene	5.806	96	117660	45.822	ug/l	100
22) Diisopropyl ether	6.689	45	384939	47.711	ug/l	100
23) Vinyl Acetate	6.618	43	1649023	241.910	ug/l	100
24) 1,1-Dichloroethane	6.583	63	220477	47.485	ug/l	100
25) 2-Butanone	7.494	43	571221	238.693	ug/l	100
26) 2,2-Dichloropropane	7.500	77	187567	51.933	ug/l	100
27) cis-1,2-Dichloroethene	7.500	96	144911	47.189	ug/l	100
28) Bromochloromethane	7.824	49	96707	42.362	ug/l	100
29) Tetrahydrofuran	7.853	42	371357	238.210	ug/l	100
30) Chloroform	7.977	83	220093	47.467	ug/l	100
31) Cyclohexane	8.271	56	208105	46.217	ug/l	100
32) 1,1,1-Trichloroethane	8.177	97	185502	47.037	ug/l	100
36) 1,1-Dichloropropene	8.377	75	158341	47.362	ug/l	100
37) Ethyl Acetate	7.577	43	203091	47.720	ug/l	100
38) Carbon Tetrachloride	8.371	117	154677	47.126	ug/l	100
39) Methylcyclohexane	9.606	83	215849	47.126	ug/l	100
40) Benzene	8.612	78	509334	46.590	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060625\
 Data File : VN086865.D
 Acq On : 06 Jun 2025 14:03
 Operator : JC\MD
 Sample : VSTDICCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICCC050

Quant Time: Jun 07 01:57:17 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 01:51:02 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.789	41	114727	48.000	ug/l	100
42) 1,2-Dichloroethane	8.677	62	155713	46.960	ug/l	100
43) Isopropyl Acetate	8.694	43	322741	47.248	ug/l	100
44) Trichloroethene	9.359	130	123847	47.776	ug/l	100
45) 1,2-Dichloropropane	9.630	63	125673	47.251	ug/l	100
46) Dibromomethane	9.718	93	83811	47.442	ug/l	100
47) Bromodichloromethane	9.894	83	173200	47.651	ug/l	100
48) Methyl methacrylate	9.683	41	149019	47.399	ug/l	100
49) 1,4-Dioxane	9.700	88	59060	1032.607	ug/l #	100
51) 4-Methyl-2-Pentanone	10.447	43	1017688	247.484	ug/l	100
52) Toluene	10.635	92	316219	47.331	ug/l	100
53) t-1,3-Dichloropropene	10.841	75	197475	48.587	ug/l	100
54) cis-1,3-Dichloropropene	10.318	75	208523	47.950	ug/l	100
55) 1,1,2-Trichloroethane	11.018	97	121969	47.445	ug/l	100
56) Ethyl methacrylate	10.882	69	210467	51.401	ug/l	100
57) 1,3-Dichloropropane	11.165	76	212243	47.593	ug/l	100
58) 2-Chloroethyl Vinyl ether	10.165	63	519139	212.570	ug/l	100
59) 2-Hexanone	11.200	43	696567	262.977	ug/l	100
60) Dibromochloromethane	11.359	129	128715	48.056	ug/l	100
61) 1,2-Dibromoethane	11.471	107	124522	47.258	ug/l	100
64) Tetrachloroethene	11.106	164	97679	46.382	ug/l	100
65) Chlorobenzene	11.894	112	340439	46.386	ug/l	100
66) 1,1,1,2-Tetrachloroethane	11.965	131	112333	47.615	ug/l	100
67) Ethyl Benzene	11.965	91	597593	47.271	ug/l	100
68) m/p-Xylenes	12.071	106	466620	96.424	ug/l	100
69) o-Xylene	12.400	106	224299	48.395	ug/l	100
70) Styrene	12.412	104	386596	48.745	ug/l	100
71) Bromoform	12.582	173	88127	50.415	ug/l #	100
73) Isopropylbenzene	12.694	105	573905	47.023	ug/l	100
74) N-amyl acetate	12.512	43	212064	52.353	ug/l	100
75) 1,1,2,2-Tetrachloroethane	12.941	83	197360	47.732	ug/l	100
76) 1,2,3-Trichloropropane	12.994	75	267483	66.959	ug/l	68
77) Bromobenzene	12.982	156	132381	47.296	ug/l	100
78) n-propylbenzene	13.035	91	705524	47.567	ug/l	100
79) 2-Chlorotoluene	13.124	91	419986	47.216	ug/l	100
80) 1,3,5-Trimethylbenzene	13.171	105	485499	48.181	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.735	75	79732	46.089	ug/l	100
82) 4-Chlorotoluene	13.224	91	424066	47.117	ug/l	100
83) tert-Butylbenzene	13.435	119	438417	47.531	ug/l	100
84) 1,2,4-Trimethylbenzene	13.482	105	488189	48.314	ug/l	100
85) sec-Butylbenzene	13.618	105	641523	47.893	ug/l	100
86) p-Isopropyltoluene	13.729	119	533366	48.169	ug/l	100
87) 1,3-Dichlorobenzene	13.735	146	260310	47.269	ug/l	100
88) 1,4-Dichlorobenzene	13.812	146	263358	46.907	ug/l	100
89) n-Butylbenzene	14.053	91	512519	47.787	ug/l	100
90) Hexachloroethane	14.335	117	89896	47.896	ug/l	100
91) 1,2-Dichlorobenzene	14.106	146	251235	47.486	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	14.718	75	45535	46.049	ug/l	100
93) 1,2,4-Trichlorobenzene	15.388	180	162374	48.061	ug/l	100
94) Hexachlorobutadiene	15.500	225	60866	48.356	ug/l	100
95) Naphthalene	15.641	128	603088	47.958	ug/l	100
96) 1,2,3-Trichlorobenzene	15.835	180	159113	47.402	ug/l	100


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060625\
Data File : VN086865.D
Acq On : 06 Jun 2025 14:03
Operator : JC\MD
Sample : VSTDICCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Quant Time: Jun 07 01:57:17 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 01:51:02 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.236 min Scan# 1068
Delta R.T. -0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tgt Ion:168 Resp: 207354
Ion Ratio Lower Upper
168 100
99 61.4 49.1 73.7



#2
Dichlorodifluoromethane
Concen: 50.283 ug/l
RT: 2.154 min Scan# 34
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03

Tgt Ion: 85 Resp: 103958
Ion Ratio Lower Upper
85 100
87 32.4 16.2 48.6









#3
Chloromethane
Concen: 46.397 ug/l
RT: 2.395 min Scan# 75
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tgt Ion: 50 Resp: 123866
Ion Ratio Lower Upper
50 100
52 32.1 25.7 38.5



#4
Vinyl Chloride
Concen: 48.215 ug/l
RT: 2.554 min Scan# 102
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03

Tgt Ion: 62 Resp: 132783
Ion Ratio Lower Upper
62 100
64 32.5 26.0 39.0









#5
Bromomethane
Concen: 47.999 ug/l
RT: 2.995 min Scan# 177
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tgt Ion: 94 Resp: 73974
Ion Ratio Lower Upper
94 100
96 97.4 77.9 116.9



#6
Chloroethane
Concen: 47.504 ug/l
RT: 3.154 min Scan# 204
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03

Tgt Ion: 64 Resp: 84505
Ion Ratio Lower Upper
64 100
66 30.7 24.6 36.8









#7
 Trichlorofluoromethane
 Concen: 48.041 ug/l
 RT: 3.524 min Scan# 267
 Delta R.T. 0.000 min
 Lab File: VN086865.D
 Acq: 06 Jun 2025 14:03




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICCC050


Tgt Ion:101 Resp: 172860
 Ion Ratio Lower Upper
 101 100
 103 65.2 52.2 78.2



#8
 Diethyl Ether
 Concen: 47.963 ug/l
 RT: 3.983 min Scan# 345
 Delta R.T. 0.000 min
 Lab File: VN086865.D
 Acq: 06 Jun 2025 14:03

Tgt Ion: 74 Resp: 75193
 Ion Ratio Lower Upper
 74 100
 45 91.0 45.5 136.5









#9
1,1,2-Trichlorotrifluoroethane
Concen: 47.640 ug/l
RT: 4.395 min Scan# 415
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tgt Ion:101 Resp: 107695
Ion Ratio Lower Upper
101 100
85 42.3 33.8 50.8
151 73.4 58.7 88.1



#10
Methyl Iodide
Concen: 48.346 ug/l
RT: 4.606 min Scan# 451
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03

Tgt Ion:142 Resp: 141678
Ion Ratio Lower Upper
142 100
127 38.6 30.9 46.3
141 13.0 10.4 15.6









#11
Tert butyl alcohol
Concen: 242.020 ug/l
RT: 5.548 min Scan# 611
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tgt Ion: 59 Resp: 182315
Ion Ratio Lower Upper
59 100
57 10.1 8.1 12.1



#12
1,1-Dichloroethene
Concen: 47.900 ug/l
RT: 4.359 min Scan# 409
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03

Tgt Ion: 96 Resp: 110549
Ion Ratio Lower Upper
96 100
61 157.2 125.8 188.6
98 65.6 52.5 78.7





#13

Acrolein

Concen: 197.217 ug/l

RT: 4.201 min Scan# 382

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03

Instrument :

MSVOA_N

ClientSampleId :





VSTDICCC050

Tgt Ion: 56 Resp: 47026

Ion Ratio Lower Upper

56 100

55 69.5 55.6 83.4



#14

Allyl chloride

Concen: 46.848 ug/l

RT: 5.042 min Scan# 525

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03




Tgt Ion: 41 Resp: 179313


Ion Ratio Lower Upper

41 100

39 64.1 51.3 76.9

76 39.2 31.4 47.0









#15
Acrylonitrile
Concen: 239.638 ug/l
RT: 5.742 min Scan# 644
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tgt Ion: 53 Resp: 421941
Ion Ratio Lower Upper
53 100
52 81.3 65.0 97.6
51 34.0 27.2 40.8



#16
Acetone
Concen: 226.765 ug/l
RT: 4.448 min Scan# 424
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03

Tgt Ion: 43 Resp: 333857
Ion Ratio Lower Upper
43 100
58 35.0 28.0 42.0





#17

Carbon Disulfide

Concen: 46.328 ug/l

RT: 4.730 min Scan# 472

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03

Instrument :

MSVOA_N

ClientSampleId :





VSTDICCC050

Tgt Ion: 76 Resp: 295753

Ion Ratio Lower Upper

76 100

78 9.1 7.3 10.9



#18

Methyl Acetate

Concen: 47.673 ug/l

RT: 5.042 min Scan# 525

Delta R.T. 0.000 min

Lab File: VN086865.D




Acq: 06 Jun 2025 14:03


Tgt Ion: 43 Resp: 204538

Ion Ratio Lower Upper

43 100

74 25.9 20.7 31.1









#19
Methyl tert-butyl Ether
Concen: 47.960 ug/l
RT: 5.812 min Scan# 61
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tgt Ion: 73 Resp: 400775
Ion Ratio Lower Upper
73 100
57 21.8 17.4 26.2



#20
Methylene Chloride
Concen: 45.531 ug/l
RT: 5.295 min Scan# 568
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03

Tgt Ion: 84 Resp: 125498
Ion Ratio Lower Upper
84 100
49 113.1 90.5 135.7
51 35.6 28.5 42.7
86 60.1 48.1 72.1









#21
trans-1,2-Dichloroethene
Concen: 45.822 ug/l
RT: 5.806 min Scan# 611
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tgt Ion: 96 Resp: 117660
Ion Ratio Lower Upper
96 100
61 136.5 109.2 163.8
98 62.3 49.8 74.8



#22
Diisopropyl ether
Concen: 47.711 ug/l
RT: 6.689 min Scan# 805
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03

Tgt Ion: 45 Resp: 384939
Ion Ratio Lower Upper
45 100
43 52.4 41.9 62.9
87 28.0 22.4 33.6
59 12.2 9.8 14.6









#23
 Vinyl Acetate
 Concen: 241.910 ug/l
 RT: 6.618 min Scan# 793
 Delta R.T. 0.000 min
 Lab File: VN086865.D
 Acq: 06 Jun 2025 14:03




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICCC050


Tgt Ion: 43 Resp: 1649023
 Ion Ratio Lower Upper
 43 100
 86 11.9 9.5 14.3



#24
 1,1-Dichloroethane
 Concen: 47.485 ug/l
 RT: 6.583 min Scan# 787
 Delta R.T. 0.000 min
 Lab File: VN086865.D
 Acq: 06 Jun 2025 14:03

Tgt Ion: 63 Resp: 220477
 Ion Ratio Lower Upper
 63 100
 98 7.2 3.6 10.8
 100 5.2 2.6 7.8









#25
2-Butanone
Concen: 238.693 ug/l
RT: 7.494 min Scan# 942
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tgt Ion: 43 Resp: 571221
Ion Ratio Lower Upper
43 100
72 28.8 23.0 34.6



#26
2,2-Dichloropropane
Concen: 51.933 ug/l
RT: 7.500 min Scan# 943
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03

Tgt Ion: 77 Resp: 187567
Ion Ratio Lower Upper
77 100
97 23.1 11.6 34.7









#27
 cis-1,2-Dichloroethene
 Concen: 47.189 ug/l
 RT: 7.500 min Scan# 943
 Delta R.T. 0.000 min
 Lab File: VN086865.D
 Acq: 06 Jun 2025 14:03




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICCC050


Tgt Ion: 96 Resp: 144911
 Ion Ratio Lower Upper
 96 100
 61 139.0 0.0 278.0
 98 64.1 0.0 128.2



#28
 Bromochloromethane
 Concen: 42.362 ug/l
 RT: 7.824 min Scan# 998
 Delta R.T. 0.000 min
 Lab File: VN086865.D
 Acq: 06 Jun 2025 14:03

Tgt Ion: 49 Resp: 96707
 Ion Ratio Lower Upper
 49 100
 129 1.9 0.0 3.8
 130 77.7 62.2 93.2





#29

Tetrahydrofuran

Concen: 238.210 ug/l

RT: 7.853 min Scan# 1003

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050





Tgt Ion: 42 Resp: 371357

Ion Ratio Lower Upper

42 100

72 47.2 37.8 56.6

71 44.5 35.6 53.4



#30

Chloroform

Concen: 47.467 ug/l

RT: 7.977 min Scan# 1024

Delta R.T. 0.000 min

Lab File: VN086865.D




Acq: 06 Jun 2025 14:03


Tgt Ion: 83 Resp: 220093

Ion Ratio Lower Upper

83 100

85 66.1 52.9 79.3









#31
Cyclohexane
Concen: 46.217 ug/l
RT: 8.271 min Scan# 1074
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tgt Ion: 56 Resp: 208105
Ion Ratio Lower Upper
56 100
69 31.7 25.4 38.0
84 86.0 68.8 103.2



#32
1,1,1-Trichloroethane
Concen: 47.037 ug/l
RT: 8.177 min Scan# 1058
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03

Tgt Ion: 97 Resp: 185502
Ion Ratio Lower Upper
97 100
99 65.0 52.0 78.0
61 44.9 35.9 53.9





#33

1,2-Dichloroethane-d4

Concen: 37.366 ug/l

RT: 8.588 min Scan# 1128

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03

Instrument :

MSVOA_N

ClientSampleId :





VSTDICCC050

Tgt Ion: 65 Resp: 103736

Ion Ratio Lower Upper

65 100

67 52.8 0.0 105.6



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.106 min Scan# 1216

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03




Tgt Ion: 114 Resp: 378587


Ion Ratio Lower Upper

114 100

63 19.8 0.0 39.6

88 15.1 0.0 30.2









#35
Dibromofluoromethane
Concen: 37.020 ug/l
RT: 8.171 min Scan# 1057
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tgt Ion:113 Resp: 83058
Ion Ratio Lower Upper
113 100
111 105.2 84.2 126.2
192 17.8 14.2 21.4



#36
1,1-Dichloropropene
Concen: 47.362 ug/l
RT: 8.377 min Scan# 1092
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03

Tgt Ion: 75 Resp: 158341
Ion Ratio Lower Upper
75 100
110 35.3 17.6 52.9
77 31.6 25.3 37.9









#37
Ethyl Acetate
Concen: 47.720 ug/l
RT: 7.577 min Scan# 91
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tgt Ion: 43 Resp: 203091
Ion Ratio Lower Upper
43 100
61 15.1 12.1 18.1
70 12.2 9.8 14.6



#38
Carbon Tetrachloride
Concen: 47.126 ug/l
RT: 8.371 min Scan# 1091
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03

Tgt Ion: 117 Resp: 154677
Ion Ratio Lower Upper
117 100
119 93.8 75.0 112.6
121 31.1 24.9 37.3









#39
Methylcyclohexane
Concen: 47.126 ug/l
RT: 9.606 min Scan# 1130
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tgt Ion: 83 Resp: 215849
Ion Ratio Lower Upper
83 100
55 76.3 61.0 91.6
98 48.6 38.9 58.3



#40
Benzene
Concen: 46.590 ug/l
RT: 8.612 min Scan# 1132
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03

Tgt Ion: 78 Resp: 509334
Ion Ratio Lower Upper
78 100
77 23.4 18.7 28.1









#41
Methacrylonitrile
Concen: 48.000 ug/l
RT: 7.789 min Scan# 992
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tgt Ion:	41	Resp:	114727
Ion	Ratio	Lower	Upper
41	100		
39	51.5	41.2	61.8
67	76.8	61.4	92.2
52	28.4	22.7	34.1



#42
1,2-Dichloroethane
Concen: 46.960 ug/l
RT: 8.677 min Scan# 1143
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03

Tgt Ion:	62	Resp:	155713
Ion	Ratio	Lower	Upper
62	100		
98	10.6	0.0	21.2









#43
Isopropyl Acetate
Concen: 47.248 ug/l
RT: 8.694 min Scan# 1146
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tgt Ion: 43 Resp: 322741
Ion Ratio Lower Upper
43 100
61 27.6 22.1 33.1
87 14.7 11.8 17.6



#44
Trichloroethene
Concen: 47.776 ug/l
RT: 9.359 min Scan# 1259
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03

Tgt Ion:130 Resp: 123847
Ion Ratio Lower Upper
130 100
95 98.1 0.0 196.2





#45

1,2-Dichloropropane

Concen: 47.251 ug/l

RT: 9.630 min Scan# 1305

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03

Instrument :

MSVOA_N

ClientSampleId :





VSTDICCC050

Tgt Ion: 63 Resp: 125673

Ion Ratio Lower Upper

63 100

65 31.9 25.5 38.3



#46

Dibromomethane

Concen: 47.442 ug/l

RT: 9.718 min Scan# 1320

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03




Tgt Ion: 93 Resp: 83811


Ion Ratio Lower Upper

93 100

95 83.9 67.1 100.7

174 92.8 74.2 111.4









#47
Bromodichloromethane
Concen: 47.651 ug/l
RT: 9.894 min Scan# 1350
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tgt Ion: 83 Resp: 173200
Ion Ratio Lower Upper
83 100
85 63.6 50.9 76.3
127 8.3 6.6 10.0



#48
Methyl methacrylate
Concen: 47.399 ug/l
RT: 9.683 min Scan# 1314
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03

Tgt Ion: 41 Resp: 149019
Ion Ratio Lower Upper
41 100
69 93.3 74.6 112.0
39 50.3 40.2 60.4





#49

1,4-Dioxane

Concen: 1032.607 ug/l

RT: 9.700 min Scan# 11

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050





Tgt Ion: 88 Resp: 59060

Ion Ratio Lower Upper

88 100

43 0.0 0.0 0.0

58 72.1 57.7 86.5



#50

Toluene-d8

Concen: 36.716 ug/l

RT: 10.571 min Scan# 1465

Delta R.T. 0.000 min

Lab File: VN086865.D




Acq: 06 Jun 2025 14:03


Tgt Ion: 98 Resp: 326105

Ion Ratio Lower Upper

98 100

100 66.7 53.4 80.0









#51
4-Methyl-2-Pentanone
Concen: 247.484 ug/l
RT: 10.447 min Scan# 1444
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tgt Ion: 43 Resp: 1017688
Ion Ratio Lower Upper
43 100
58 42.8 34.2 51.4



#52
Toluene
Concen: 47.331 ug/l
RT: 10.635 min Scan# 1476
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03

Tgt Ion: 92 Resp: 316219
Ion Ratio Lower Upper
92 100
91 170.6 136.5 204.7









#53
 t-1,3-Dichloropropene
 Concen: 48.587 ug/l
 RT: 10.841 min Scan# 1511
 Delta R.T. 0.000 min
 Lab File: VN086865.D
 Acq: 06 Jun 2025 14:03




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICCC050


Tgt Ion: 75 Resp: 197475
 Ion Ratio Lower Upper
 75 100
 77 33.0 26.4 39.6



#54
 cis-1,3-Dichloropropene
 Concen: 47.950 ug/l
 RT: 10.318 min Scan# 1422
 Delta R.T. 0.000 min
 Lab File: VN086865.D
 Acq: 06 Jun 2025 14:03

Tgt Ion: 75 Resp: 208523
 Ion Ratio Lower Upper
 75 100
 77 32.8 26.2 39.4
 39 40.4 32.3 48.5





#55

1,1,2-Trichloroethane

Concen: 47.445 ug/l

RT: 11.018 min Scan# 1541

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion: 97 Resp: 121969





Ion Ratio Lower Upper

97 100

83 87.1 69.7 104.5

85 56.4 45.1 67.7

99 63.2 50.6 75.8



#56

Ethyl methacrylate

Concen: 51.401 ug/l

RT: 10.882 min Scan# 1518

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03




Tgt Ion: 69 Resp: 210467


Ion Ratio Lower Upper

69 100

41 59.0 47.2 70.8

39 28.8 23.0 34.6





#57

1,3-Dichloropropane

Concen: 47.593 ug/l

RT: 11.165 min Scan# 1566

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03

Instrument :

MSVOA_N

ClientSampleId :





VSTDICCC050

Tgt Ion: 76 Resp: 212243

Ion Ratio Lower Upper

76 100

78 31.6 25.3 37.9



#58

2-Chloroethyl Vinyl ether

Concen: 212.570 ug/l

RT: 10.165 min Scan# 1396

Delta R.T. 0.000 min

Lab File: VN086865.D




Acq: 06 Jun 2025 14:03


Tgt Ion: 63 Resp: 519139

Ion Ratio Lower Upper

63 100

106 28.8 23.0 34.6









#59
2-Hexanone
Concen: 262.977 ug/l
RT: 11.200 min Scan# 1572
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tgt Ion: 43 Resp: 696567
Ion Ratio Lower Upper
43 100
58 60.9 30.4 91.3



#60
Dibromochloromethane
Concen: 48.056 ug/l
RT: 11.359 min Scan# 1599
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03

Tgt Ion: 129 Resp: 128715
Ion Ratio Lower Upper
129 100
127 77.2 38.6 115.8









#61
1,2-Dibromoethane
Concen: 47.258 ug/l
RT: 11.471 min Scan# 1618
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tgt Ion:107 Resp: 124522
Ion Ratio Lower Upper
107 100
109 95.4 76.3 114.5



#62
4-Bromofluorobenzene
Concen: 37.264 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03

Tgt Ion: 95 Resp: 122965
Ion Ratio Lower Upper
95 100
174 70.9 0.0 141.8
176 66.3 0.0 132.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 117
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tgt Ion:117 Resp: 332766
Ion Ratio Lower Upper
117 100
82 55.8 44.6 67.0
119 31.9 25.5 38.3



#64
Tetrachloroethene
Concen: 46.382 ug/l
RT: 11.106 min Scan# 1556
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03

Tgt Ion:164 Resp: 97679
Ion Ratio Lower Upper
164 100
166 127.3 101.8 152.8
129 103.7 83.0 124.4
131 97.7 78.2 117.2









#65
Chlorobenzene
Concen: 46.386 ug/l
RT: 11.894 min Scan# 1690
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tgt Ion:112 Resp: 340439
Ion Ratio Lower Upper
112 100
114 32.7 26.2 39.2



#66
1,1,1,2-Tetrachloroethane
Concen: 47.615 ug/l
RT: 11.965 min Scan# 1702
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03

Tgt Ion:131 Resp: 112333
Ion Ratio Lower Upper
131 100
133 96.5 48.3 144.8
119 67.9 34.0 101.8









#67
Ethyl Benzene
Concen: 47.271 ug/l
RT: 11.965 min Scan# 1702
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tgt Ion: 91 Resp: 597593
Ion Ratio Lower Upper
91 100
106 31.8 25.4 38.2



#68
m/p-Xylenes
Concen: 96.424 ug/l
RT: 12.071 min Scan# 1720
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03

Tgt Ion:106 Resp: 466620
Ion Ratio Lower Upper
106 100
91 199.2 159.4 239.0





#69

o-Xylene

Concen: 48.395 ug/l

RT: 12.400 min Scan# 1776

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03

Instrument :

MSVOA_N

ClientSampleId :





VSTDICCC050

Tgt Ion:106 Resp: 224299

Ion Ratio Lower Upper

106 100

91 212.1 106.1 318.1



#70

Styrene

Concen: 48.745 ug/l

RT: 12.412 min Scan# 1778

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03




Tgt Ion:104 Resp: 386596


Ion Ratio Lower Upper

104 100

78 48.1 38.5 57.7

103 53.9 43.1 64.7









#71
Bromoform
Concen: 50.415 ug/l
RT: 12.582 min Scan# 1807
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tgt Ion:173 Resp: 88127
Ion Ratio Lower Upper
173 100
175 48.3 24.1 72.4
254 0.0 0.0 0.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.794 min Scan# 2013
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03

Tgt Ion:152 Resp: 167532
Ion Ratio Lower Upper
152 100
115 60.3 30.1 90.5
150 172.5 0.0 345.0









#73
Isopropylbenzene
Concen: 47.023 ug/l
RT: 12.694 min Scan# 1826
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tgt Ion:105 Resp: 573905
Ion Ratio Lower Upper
105 100
120 26.9 13.5 40.4



#74
N-ethyl acetate
Concen: 52.353 ug/l
RT: 12.512 min Scan# 1795
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03

Tgt Ion: 43 Resp: 212064
Ion Ratio Lower Upper
43 100
70 47.0 37.6 56.4
55 23.8 19.0 28.6
61 26.0 20.8 31.2









#75
1,1,2,2-Tetrachloroethane
Concen: 47.732 ug/l
RT: 12.941 min Scan# 1868
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tgt Ion: 83 Resp: 197360
Ion Ratio Lower Upper
83 100
131 9.8 4.9 14.7
85 64.0 32.0 96.0



#76
1,2,3-Trichloropropane
Concen: 66.959 ug/l
RT: 12.994 min Scan# 1877
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03

Tgt Ion: 75 Resp: 267483
Ion Ratio Lower Upper
75 100
77 116.5 79.3 237.9









#77
Bromobenzene
Concen: 47.296 ug/l
RT: 12.982 min Scan# 1875
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tgt Ion:156 Resp: 132381
Ion Ratio Lower Upper
156 100
77 235.4 117.7 353.1
158 96.3 48.1 144.4



#78
n-propylbenzene
Concen: 47.567 ug/l
RT: 13.035 min Scan# 1884
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03

Tgt Ion: 91 Resp: 705524
Ion Ratio Lower Upper
91 100
120 22.8 11.4 34.2









#79
2-Chlorotoluene
Concen: 47.216 ug/l
RT: 13.124 min Scan# 1899
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tgt Ion: 91 Resp: 419986
Ion Ratio Lower Upper
91 100
126 32.3 16.2 48.5



#80
1,3,5-Trimethylbenzene
Concen: 48.181 ug/l
RT: 13.171 min Scan# 1907
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03

Tgt Ion: 105 Resp: 485499
Ion Ratio Lower Upper
105 100
120 49.7 24.9 74.6









#81
trans-1,4-Dichloro-2-butene
Concen: 46.089 ug/l
RT: 12.735 min Scan# 1833
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tgt Ion: 75 Resp: 79732
Ion Ratio Lower Upper
75 100
53 83.2 66.6 99.8
89 45.9 36.7 55.1



#82
4-Chlorotoluene
Concen: 47.117 ug/l
RT: 13.224 min Scan# 1916
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03

Tgt Ion: 91 Resp: 424066
Ion Ratio Lower Upper
91 100
126 33.4 16.7 50.1





#83

tert-Butylbenzene

Concen: 47.531 ug/l

RT: 13.435 min Scan# 1952

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050





Tgt Ion:119 Resp: 438417

Ion Ratio Lower Upper

119 100

91 64.3 32.1 96.5

134 26.7 13.4 40.1



#84

1,2,4-Trimethylbenzene

Concen: 48.314 ug/l

RT: 13.482 min Scan# 1960

Delta R.T. 0.000 min

Lab File: VN086865.D




Acq: 06 Jun 2025 14:03


Tgt Ion:105 Resp: 488189

Ion Ratio Lower Upper

105 100

120 46.4 23.2 69.6









#85
 sec-Butylbenzene
 Concen: 47.893 ug/l
 RT: 13.618 min Scan# 1983
 Delta R.T. 0.000 min
 Lab File: VN086865.D
 Acq: 06 Jun 2025 14:03




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICCC050


Tgt Ion:105 Resp: 641523
 Ion Ratio Lower Upper
 105 100
 134 20.0 10.0 30.0



#86
 p-Isopropyltoluene
 Concen: 48.169 ug/l
 RT: 13.729 min Scan# 2002
 Delta R.T. 0.000 min
 Lab File: VN086865.D
 Acq: 06 Jun 2025 14:03

Tgt Ion:119 Resp: 533366
 Ion Ratio Lower Upper
 119 100
 134 26.9 13.5 40.4
 91 22.9 11.5 34.4





#87

1,3-Dichlorobenzene

Concen: 47.269 ug/l

RT: 13.735 min Scan# 2003

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050





Tgt Ion:146 Resp: 260310

Ion Ratio Lower Upper

146 100

111 41.9 20.9 62.8

148 64.1 32.0 96.2



#88

1,4-Dichlorobenzene

Concen: 46.907 ug/l

RT: 13.812 min Scan# 2016

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03




Tgt Ion:146 Resp: 263358


Ion Ratio Lower Upper

146 100

111 39.8 19.9 59.7

148 63.6 31.8 95.4









#89
n-Butylbenzene
Concen: 47.787 ug/l
RT: 14.053 min Scan# 2057
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tgt Ion: 91 Resp: 512519
Ion Ratio Lower Upper
91 100
92 54.9 27.5 82.4
134 26.0 13.0 39.0



#90
Hexachloroethane
Concen: 47.896 ug/l
RT: 14.335 min Scan# 2105
Delta R.T. 0.000 min
Lab File: VN086865.D
Acq: 06 Jun 2025 14:03

Tgt Ion: 117 Resp: 89896
Ion Ratio Lower Upper
117 100
201 66.4 33.2 99.6









#91
 1,2-Dichlorobenzene
 Concen: 47.486 ug/l
 RT: 14.106 min Scan# 2066
 Delta R.T. 0.000 min
 Lab File: VN086865.D
 Acq: 06 Jun 2025 14:03




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICCC050


Tgt Ion:146 Resp: 251235
 Ion Ratio Lower Upper
 146 100
 111 43.1 21.6 64.6
 148 63.4 31.7 95.1



#92
 1,2-Dibromo-3-Chloropropane
 Concen: 46.049 ug/l
 RT: 14.718 min Scan# 2170
 Delta R.T. 0.000 min
 Lab File: VN086865.D
 Acq: 06 Jun 2025 14:03

Tgt Ion: 75 Resp: 45535
 Ion Ratio Lower Upper
 75 100
 155 82.3 41.1 123.4
 157 105.4 52.7 158.1





#93

1,2,4-Trichlorobenzene

Concen: 48.061 ug/l

RT: 15.388 min Scan# 21

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050





Tgt Ion:180 Resp: 162374

Ion Ratio Lower Upper

180 100

182 93.8 46.9 140.7

145 31.9 16.0 47.9



#94

Hexachlorobutadiene

Concen: 48.356 ug/l

RT: 15.500 min Scan# 2303

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03




Tgt Ion:225 Resp: 60866


Ion Ratio Lower Upper

225 100

223 66.9 33.5 100.4

227 62.3 31.1 93.5





#95

Naphthalene

Concen: 47.958 ug/l

RT: 15.641 min Scan# 2327

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050





Tgt Ion:128 Resp: 603088

Ion Ratio Lower Upper

128 100

127 12.7 10.2 15.2

129 11.0 8.8 13.2



#96

1,2,3-Trichlorobenzene

Concen: 47.402 ug/l

RT: 15.835 min Scan# 2360

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03




Tgt Ion:180 Resp: 159113

Ion Ratio Lower Upper

180 100

182 94.0 47.0 141.0

145 33.2 16.6 49.8



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060625\
 Data File : VN086866.D
 Acq On : 06 Jun 2025 14:26
 Operator : JC\MD
 Sample : VSTDICC100
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC100

Quant Time: Jun 07 01:58:16 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 01:51:02 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	181996	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	327782	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	290313	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	147098	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	238957	98.065	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 196.120%#	
35) Dibromofluoromethane	8.177	113	195212	100.495	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 200.980%#	
50) Toluene-d8	10.570	98	772210	100.419	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 200.840%#	
62) 4-Bromofluorobenzene	12.847	95	292385	102.339	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 204.680%#	
Target Compounds						Qvalue
2) Dichlorodifluoromethane	2.153	85	194799	107.350	ug/l	100
3) Chloromethane	2.395	50	224558	95.833	ug/l	98
4) Vinyl Chloride	2.553	62	245135	101.414	ug/l	98
5) Bromomethane	2.977	94	137967	101.996	ug/l	99
6) Chloroethane	3.147	64	152150	97.447	ug/l	97
7) Trichlorofluoromethane	3.524	101	312394	98.917	ug/l	99
8) Diethyl Ether	3.983	74	139844	101.631	ug/l	98
9) 1,1,2-Trichlorotrifluo...	4.389	101	198739	100.163	ug/l	99
10) Methyl Iodide	4.612	142	262771	102.162	ug/l	99
11) Tert butyl alcohol	5.547	59	327987	496.062	ug/l	99
12) 1,1-Dichloroethene	4.359	96	200212	98.838	ug/l	99
13) Acrolein	4.200	56	94171	449.961	ug/l	98
14) Allyl chloride	5.047	41	329412	98.055	ug/l	99
15) Acrylonitrile	5.735	53	770470	498.552	ug/l	99
16) Acetone	4.447	43	608178	470.648	ug/l	98
17) Carbon Disulfide	4.736	76	544511	97.179	ug/l	100
18) Methyl Acetate	5.047	43	381808	101.390	ug/l	100
19) Methyl tert-butyl Ether	5.818	73	735694	100.306	ug/l	100
20) Methylene Chloride	5.294	84	228800	94.575	ug/l	97
21) trans-1,2-Dichloroethene	5.806	96	215177	95.476	ug/l	99
22) Diisopropyl ether	6.688	45	696881	98.409	ug/l	98
23) Vinyl Acetate	6.618	43	2919350	487.936	ug/l	100
24) 1,1-Dichloroethane	6.588	63	404207	99.186	ug/l	99
25) 2-Butanone	7.494	43	1043385	496.743	ug/l	100
26) 2,2-Dichloropropane	7.500	77	331995	104.729	ug/l	99
27) cis-1,2-Dichloroethene	7.500	96	265248	98.410	ug/l	99
28) Bromochloromethane	7.824	49	188050	93.851	ug/l	99
29) Tetrahydrofuran	7.847	42	677636	495.240	ug/l	99
30) Chloroform	7.977	83	394927	97.040	ug/l	99
31) Cyclohexane	8.265	56	374933	94.869	ug/l	99
32) 1,1,1-Trichloroethane	8.177	97	336819	97.306	ug/l	99
36) 1,1-Dichloropropene	8.382	75	289907	100.156	ug/l	99
37) Ethyl Acetate	7.571	43	370652	100.591	ug/l	99
38) Carbon Tetrachloride	8.371	117	285325	100.404	ug/l	98
39) Methylcyclohexane	9.606	83	395603	99.758	ug/l	99
40) Benzene	8.612	78	927083	97.946	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060625\
 Data File : VN086866.D
 Acq On : 06 Jun 2025 14:26
 Operator : JC\MD
 Sample : VSTDICC100
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC100

Quant Time: Jun 07 01:58:16 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 01:51:02 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.794	41	195946	94.688	ug/l	95
42) 1,2-Dichloroethane	8.676	62	282015	98.232	ug/l	100
43) Isopropyl Acetate	8.694	43	579464	97.981	ug/l	99
44) Trichloroethene	9.359	130	222822	99.280	ug/l	97
45) 1,2-Dichloropropane	9.629	63	230454	100.077	ug/l	98
46) Dibromomethane	9.712	93	155530	101.685	ug/l	99
47) Bromodichloromethane	9.894	83	316375	100.533	ug/l	100
48) Methyl methacrylate	9.682	41	275710	101.288	ug/l	99
49) 1,4-Dioxane	9.706	88	104771	2115.744	ug/l #	99
51) 4-Methyl-2-Pentanone	10.447	43	1843274	517.729	ug/l	99
52) Toluene	10.635	92	578978	100.092	ug/l	100
53) t-1,3-Dichloropropene	10.841	75	359200	102.077	ug/l	99
54) cis-1,3-Dichloropropene	10.318	75	383030	101.730	ug/l	99
55) 1,1,2-Trichloroethane	11.017	97	219922	98.808	ug/l	98
56) Ethyl methacrylate	10.882	69	389802	109.955	ug/l	99
57) 1,3-Dichloropropane	11.165	76	382510	99.069	ug/l	99
58) 2-Chloroethyl Vinyl ether	10.165	63	1113523	526.621	ug/l	100
59) 2-Hexanone	11.200	43	1301539	567.535	ug/l	99
60) Dibromochloromethane	11.359	129	237965	102.616	ug/l	100
61) 1,2-Dibromoethane	11.470	107	232214	101.787	ug/l	98
64) Tetrachloroethene	11.106	164	181568	98.824	ug/l	94
65) Chlorobenzene	11.894	112	632563	98.793	ug/l	98
66) 1,1,1,2-Tetrachloroethane	11.959	131	206895	100.521	ug/l	100
67) Ethyl Benzene	11.964	91	1110803	100.717	ug/l	99
68) m/p-Xylenes	12.070	106	854557	202.411	ug/l	99
69) o-Xylene	12.400	106	412719	102.070	ug/l	100
70) Styrene	12.412	104	713548	103.127	ug/l	100
71) Bromoform	12.576	173	165781	108.707	ug/l #	99
73) Isopropylbenzene	12.694	105	1065297	99.410	ug/l	100
74) N-amyl acetate	12.506	43	403587	113.475	ug/l	98
75) 1,1,2,2-Tetrachloroethane	12.941	83	354386	97.616	ug/l	99
76) 1,2,3-Trichloropropane	12.994	75	485083	138.300	ug/l	69
77) Bromobenzene	12.982	156	246413	100.266	ug/l	99
78) n-propylbenzene	13.035	91	1301631	99.948	ug/l	100
79) 2-Chlorotoluene	13.123	91	770308	98.630	ug/l	99
80) 1,3,5-Trimethylbenzene	13.170	105	901317	101.873	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.735	75	160736	105.820	ug/l	92
82) 4-Chlorotoluene	13.223	91	783657	99.166	ug/l	100
83) tert-Butylbenzene	13.435	119	807528	99.710	ug/l	100
84) 1,2,4-Trimethylbenzene	13.482	105	904794	101.983	ug/l	100
85) sec-Butylbenzene	13.617	105	1182155	100.513	ug/l	99
86) p-Isopropyltoluene	13.729	119	990273	101.857	ug/l	100
87) 1,3-Dichlorobenzene	13.735	146	474162	98.063	ug/l	100
88) 1,4-Dichlorobenzene	13.811	146	483003	97.978	ug/l	99
89) n-Butylbenzene	14.053	91	937913	99.598	ug/l	100
90) Hexachloroethane	14.335	117	169772	103.019	ug/l	99
91) 1,2-Dichlorobenzene	14.106	146	458049	98.602	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	14.717	75	83243	95.877	ug/l	96
93) 1,2,4-Trichlorobenzene	15.394	180	298870	100.751	ug/l	99
94) Hexachlorobutadiene	15.500	225	112496	101.789	ug/l	98
95) Naphthalene	15.641	128	1129434	102.290	ug/l	100
96) 1,2,3-Trichlorobenzene	15.841	180	294127	99.796	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060625\
Data File : VN086866.D
Acq On : 06 Jun 2025 14:26
Operator : JC\MD
Sample : VSTDICC100
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 6 Sample Multiplier: 1


Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100


Quant Time: Jun 07 01:58:16 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 01:51:02 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Quant Time: Jun 07 01:58:16 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 01:51:02 2025
Response via : Initial Calibration








#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.235 min Scan# 10
Delta R.T. -0.001 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26

Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100







Tgt Ion:168 Resp: 181996
Ion Ratio Lower Upper
168 100
99 57.9 49.1 73.7



#2
Dichlorodifluoromethane
Concen: 107.350 ug/l
RT: 2.153 min Scan# 34
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26

Tgt Ion: 85 Resp: 194799
Ion Ratio Lower Upper
85 100
87 32.3 16.2 48.6









#3
Chloromethane
Concen: 95.833 ug/l
RT: 2.395 min Scan# 75
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100


Tgt Ion: 50 Resp: 224558
Ion Ratio Lower Upper
50 100
52 33.4 25.7 38.5



#4
Vinyl Chloride
Concen: 101.414 ug/l
RT: 2.553 min Scan# 102
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26

Tgt Ion: 62 Resp: 245135
Ion Ratio Lower Upper
62 100
64 31.5 26.0 39.0









#5
Bromomethane
Concen: 101.996 ug/l
RT: 2.977 min Scan# 11
Delta R.T. -0.018 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100


Tgt Ion: 94 Resp: 137967
Ion Ratio Lower Upper
94 100
96 96.8 77.9 116.9



#6
Chloroethane
Concen: 97.447 ug/l
RT: 3.147 min Scan# 203
Delta R.T. -0.006 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26

Tgt Ion: 64 Resp: 152150
Ion Ratio Lower Upper
64 100
66 32.5 24.6 36.8









#7
 Trichlorofluoromethane
 Concen: 98.917 ug/l
 RT: 3.524 min Scan# 267
 Delta R.T. -0.000 min
 Lab File: VN086866.D
 Acq: 06 Jun 2025 14:26




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC100


Tgt Ion:101 Resp: 312394
 Ion Ratio Lower Upper
 101 100
 103 64.7 52.2 78.2



#8
 Diethyl Ether
 Concen: 101.631 ug/l
 RT: 3.983 min Scan# 345
 Delta R.T. -0.000 min
 Lab File: VN086866.D
 Acq: 06 Jun 2025 14:26

Tgt Ion: 74 Resp: 139844
 Ion Ratio Lower Upper
 74 100
 45 89.5 45.5 136.5









#9
1,1,2-Trichlorotrifluoroethane
Concen: 100.163 ug/l
RT: 4.389 min Scan# 41
Delta R.T. -0.006 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100


Tgt Ion:101 Resp: 198739
Ion Ratio Lower Upper
101 100
85 42.2 33.8 50.8
151 74.2 58.7 88.1



#10
Methyl Iodide
Concen: 102.162 ug/l
RT: 4.612 min Scan# 452
Delta R.T. 0.006 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26

Tgt Ion:142 Resp: 262771
Ion Ratio Lower Upper
142 100
127 37.9 30.9 46.3
141 13.6 10.4 15.6









#11
Tert butyl alcohol
Concen: 496.062 ug/l
RT: 5.547 min Scan# 611
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100


Tgt Ion: 59 Resp: 327987
Ion Ratio Lower Upper
59 100
57 9.7 8.1 12.1



#12
1,1-Dichloroethene
Concen: 98.838 ug/l
RT: 4.359 min Scan# 409
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26

Tgt Ion: 96 Resp: 200212
Ion Ratio Lower Upper
96 100
61 156.9 125.8 188.6
98 63.7 52.5 78.7









#13
Acrolein
Concen: 449.961 ug/l
RT: 4.200 min Scan# 382
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100


Tgt Ion: 56 Resp: 94171
Ion Ratio Lower Upper
56 100
55 70.9 55.6 83.4



#14
Allyl chloride
Concen: 98.055 ug/l
RT: 5.047 min Scan# 526
Delta R.T. 0.006 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26

Tgt Ion: 41 Resp: 329412
Ion Ratio Lower Upper
41 100
39 63.7 51.3 76.9
76 39.7 31.4 47.0









#15
Acrylonitrile
Concen: 498.552 ug/l
RT: 5.735 min Scan# 644
Delta R.T. -0.006 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100


Tgt Ion: 53 Resp: 770470
Ion Ratio Lower Upper
53 100
52 82.2 65.0 97.6
51 34.4 27.2 40.8



#16
Acetone
Concen: 470.648 ug/l
RT: 4.447 min Scan# 424
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26

Tgt Ion: 43 Resp: 608178
Ion Ratio Lower Upper
43 100
58 36.2 28.0 42.0









#17
Carbon Disulfide
Concen: 97.179 ug/l
RT: 4.736 min Scan# 41
Delta R.T. 0.006 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100


Tgt Ion: 76 Resp: 544511
Ion Ratio Lower Upper
76 100
78 9.1 7.3 10.9



#18
Methyl Acetate
Concen: 101.390 ug/l
RT: 5.047 min Scan# 526
Delta R.T. 0.006 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26

Tgt Ion: 43 Resp: 381808
Ion Ratio Lower Upper
43 100
74 26.1 20.7 31.1









#19
Methyl tert-butyl Ether
Concen: 100.306 ug/l
RT: 5.818 min Scan# 61
Delta R.T. 0.006 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100


Tgt Ion: 73 Resp: 735694
Ion Ratio Lower Upper
73 100
57 21.6 17.4 26.2



#20
Methylene Chloride
Concen: 94.575 ug/l
RT: 5.294 min Scan# 568
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26

Tgt Ion: 84 Resp: 228800
Ion Ratio Lower Upper
84 100
49 111.4 90.5 135.7
51 34.2 28.5 42.7
86 63.5 48.1 72.1









#21
trans-1,2-Dichloroethene
Concen: 95.476 ug/l
RT: 5.806 min Scan# 611
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100


Tgt Ion: 96 Resp: 215177
Ion Ratio Lower Upper
96 100
61 134.4 109.2 163.8
98 62.5 49.8 74.8



#22
Diisopropyl ether
Concen: 98.409 ug/l
RT: 6.688 min Scan# 805
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26

Tgt Ion: 45 Resp: 696881
Ion Ratio Lower Upper
45 100
43 51.1 41.9 62.9
87 29.6 22.4 33.6
59 12.2 9.8 14.6









#23
 Vinyl Acetate
 Concen: 487.936 ug/l
 RT: 6.618 min Scan# 793
 Delta R.T. -0.000 min
 Lab File: VN086866.D
 Acq: 06 Jun 2025 14:26




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC100


Tgt Ion: 43 Resp: 2919350
 Ion Ratio Lower Upper
 43 100
 86 12.1 9.5 14.3



#24
 1,1-Dichloroethane
 Concen: 99.186 ug/l
 RT: 6.588 min Scan# 788
 Delta R.T. 0.006 min
 Lab File: VN086866.D
 Acq: 06 Jun 2025 14:26

Tgt Ion: 63 Resp: 404207
 Ion Ratio Lower Upper
 63 100
 98 7.3 3.6 10.8
 100 4.7 2.6 7.8









#25
2-Butanone
Concen: 496.743 ug/l
RT: 7.494 min Scan# 942
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100


Tgt Ion: 43 Resp: 1043385
Ion Ratio Lower Upper
43 100
72 28.5 23.0 34.6



#26
2,2-Dichloropropane
Concen: 104.729 ug/l
RT: 7.500 min Scan# 943
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26

Tgt Ion: 77 Resp: 331995
Ion Ratio Lower Upper
77 100
97 23.4 11.6 34.7









#27
 cis-1,2-Dichloroethene
 Concen: 98.410 ug/l
 RT: 7.500 min Scan# 943
 Delta R.T. -0.000 min
 Lab File: VN086866.D
 Acq: 06 Jun 2025 14:26




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC100


Tgt Ion: 96 Resp: 265248
 Ion Ratio Lower Upper
 96 100
 61 137.9 0.0 278.0
 98 64.0 0.0 128.2



#28
 Bromochloromethane
 Concen: 93.851 ug/l
 RT: 7.824 min Scan# 998
 Delta R.T. -0.000 min
 Lab File: VN086866.D
 Acq: 06 Jun 2025 14:26

Tgt Ion: 49 Resp: 188050
 Ion Ratio Lower Upper
 49 100
 129 1.8 0.0 3.8
 130 78.9 62.2 93.2









#29
Tetrahydrofuran
Concen: 495.240 ug/l
RT: 7.847 min Scan# 1002
Delta R.T. -0.006 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100


Tgt Ion: 42 Resp: 677636
Ion Ratio Lower Upper
42 100
72 47.9 37.8 56.6
71 44.8 35.6 53.4



#30
Chloroform
Concen: 97.040 ug/l
RT: 7.977 min Scan# 1024
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26

Tgt Ion: 83 Resp: 394927
Ion Ratio Lower Upper
83 100
85 65.1 52.9 79.3









#31
Cyclohexane
Concen: 94.869 ug/l
RT: 8.265 min Scan# 1073
Delta R.T. -0.006 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100


Tgt Ion: 56 Resp: 374933
Ion Ratio Lower Upper
56 100
69 31.3 25.4 38.0
84 86.8 68.8 103.2



#32
1,1,1-Trichloroethane
Concen: 97.306 ug/l
RT: 8.177 min Scan# 1058
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26

Tgt Ion: 97 Resp: 336819
Ion Ratio Lower Upper
97 100
99 65.7 52.0 78.0
61 44.2 35.9 53.9





#33

1,2-Dichloroethane-d4

Concen: 98.065 ug/l

RT: 8.588 min Scan# 1128

Delta R.T. -0.000 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26

Instrument :

MSVOA_N

ClientSampleId :





VSTDICC100

Tgt Ion: 65 Resp: 238957

Ion Ratio Lower Upper

65 100

67 54.8 0.0 105.6



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.106 min Scan# 1216

Delta R.T. -0.000 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26




Tgt Ion: 114 Resp: 327782


Ion Ratio Lower Upper

114 100

63 19.7 0.0 39.6

88 15.4 0.0 30.2









#35
Dibromofluoromethane
Concen: 100.495 ug/l
RT: 8.177 min Scan# 1057
Delta R.T. 0.006 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100


Tgt Ion:113 Resp: 195212
Ion Ratio Lower Upper
113 100
111 102.9 84.2 126.2
192 18.2 14.2 21.4



#36
1,1-Dichloropropene
Concen: 100.156 ug/l
RT: 8.382 min Scan# 1093
Delta R.T. 0.006 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26

Tgt Ion: 75 Resp: 289907
Ion Ratio Lower Upper
75 100
110 34.9 17.6 52.9
77 31.1 25.3 37.9









#37
Ethyl Acetate
Concen: 100.591 ug/l
RT: 7.571 min Scan# 91
Delta R.T. -0.006 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100


Tgt Ion: 43 Resp: 370652
Ion Ratio Lower Upper
43 100
61 15.3 12.1 18.1
70 12.6 9.8 14.6



#38
Carbon Tetrachloride
Concen: 100.404 ug/l
RT: 8.371 min Scan# 1091
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26

Tgt Ion: 117 Resp: 285325
Ion Ratio Lower Upper
117 100
119 95.4 75.0 112.6
121 28.6 24.9 37.3





#39

Methylcyclohexane

Concen: 99.758 ug/l

RT: 9.606 min Scan# 1130

Delta R.T. -0.000 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100





Tgt Ion: 83 Resp: 395603

Ion Ratio Lower Upper

83 100

55 76.8 61.0 91.6

98 47.6 38.9 58.3



#40

Benzene

Concen: 97.946 ug/l

RT: 8.612 min Scan# 1132

Delta R.T. -0.000 min

Lab File: VN086866.D




Acq: 06 Jun 2025 14:26


Tgt Ion: 78 Resp: 927083

Ion Ratio Lower Upper

78 100

77 23.6 18.7 28.1









#41
Methacrylonitrile
Concen: 94.688 ug/l
RT: 7.794 min Scan# 992
Delta R.T. 0.006 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26

Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100

Tgt Ion: 41 Resp: 195946




Ion	Ratio	Lower	Upper
41	100		
39	54.3	41.2	61.8
67	82.5	61.4	92.2
52	30.6	22.7	34.1




#42
1,2-Dichloroethane
Concen: 98.232 ug/l
RT: 8.676 min Scan# 1143
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26

Tgt Ion: 62 Resp: 282015

Ion	Ratio	Lower	Upper
62	100		
98	10.6	0.0	21.2









#43
Isopropyl Acetate
Concen: 97.981 ug/l
RT: 8.694 min Scan# 1146
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100


Tgt Ion: 43 Resp: 579464
Ion Ratio Lower Upper
43 100
61 28.1 22.1 33.1
87 15.0 11.8 17.6



#44
Trichloroethene
Concen: 99.280 ug/l
RT: 9.359 min Scan# 1259
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26

Tgt Ion:130 Resp: 222822
Ion Ratio Lower Upper
130 100
95 101.4 0.0 196.2





#45

1,2-Dichloropropane

Concen: 100.077 ug/l

RT: 9.629 min Scan# 1319

Delta R.T. -0.000 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC100

Tgt Ion: 63 Resp: 230454

Ion Ratio Lower Upper

63 100

65 30.6 25.5 38.3



#46

Dibromomethane

Concen: 101.685 ug/l

RT: 9.712 min Scan# 1319

Delta R.T. -0.006 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26




Tgt Ion: 93 Resp: 155530


Ion Ratio Lower Upper

93 100

95 82.9 67.1 100.7

174 91.6 74.2 111.4









#47
Bromodichloromethane
Concen: 100.533 ug/l
RT: 9.894 min Scan# 1350
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100


Tgt Ion: 83 Resp: 316375
Ion Ratio Lower Upper
83 100
85 63.8 50.9 76.3
127 8.3 6.6 10.0



#48
Methyl methacrylate
Concen: 101.288 ug/l
RT: 9.682 min Scan# 1314
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26

Tgt Ion: 41 Resp: 275710
Ion Ratio Lower Upper
41 100
69 92.4 74.6 112.0
39 50.1 40.2 60.4





#49

1,4-Dioxane

Concen: 2115.744 ug/l

RT: 9.706 min Scan# 11

Delta R.T. 0.006 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100





Tgt Ion: 88 Resp: 104771

Ion Ratio Lower Upper

88 100

43 0.0 0.0 0.0

58 71.5 57.7 86.5



#50

Toluene-d8

Concen: 100.419 ug/l

RT: 10.570 min Scan# 1465

Delta R.T. -0.000 min

Lab File: VN086866.D




Acq: 06 Jun 2025 14:26


Tgt Ion: 98 Resp: 772210

Ion Ratio Lower Upper

98 100

100 67.6 53.4 80.0









#51
4-Methyl-2-Pentanone
Concen: 517.729 ug/l
RT: 10.447 min Scan# 1444
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100


Tgt Ion: 43 Resp: 1843274
Ion Ratio Lower Upper
43 100
58 43.6 34.2 51.4



#52
Toluene
Concen: 100.092 ug/l
RT: 10.635 min Scan# 1476
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26

Tgt Ion: 92 Resp: 578978
Ion Ratio Lower Upper
92 100
91 170.6 136.5 204.7









#53
t-1,3-Dichloropropene
Concen: 102.077 ug/l
RT: 10.841 min Scan# 1511
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100


Tgt Ion: 75 Resp: 359200
Ion Ratio Lower Upper
75 100
77 32.3 26.4 39.6



#54
cis-1,3-Dichloropropene
Concen: 101.730 ug/l
RT: 10.318 min Scan# 1422
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26

Tgt Ion: 75 Resp: 383030
Ion Ratio Lower Upper
75 100
77 32.3 26.2 39.4
39 40.1 32.3 48.5









#55
 1,1,2-Trichloroethane
 Concen: 98.808 ug/l
 RT: 11.017 min Scan# 1541
 Delta R.T. -0.000 min
 Lab File: VN086866.D
 Acq: 06 Jun 2025 14:26

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC100

Tgt Ion: 97 Resp: 219922




Ion	Ratio	Lower	Upper
97	100		
83	84.5	69.7	104.5
85	55.6	45.1	67.7
99	61.9	50.6	75.8




#56
 Ethyl methacrylate
 Concen: 109.955 ug/l
 RT: 10.882 min Scan# 1518
 Delta R.T. -0.000 min
 Lab File: VN086866.D
 Acq: 06 Jun 2025 14:26

Tgt Ion: 69 Resp: 389802

Ion	Ratio	Lower	Upper
69	100		
41	58.3	47.2	70.8
39	27.5	23.0	34.6





#57

1,3-Dichloropropane

Concen: 99.069 ug/l

RT: 11.165 min Scan# 1566

Delta R.T. -0.000 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC100

Tgt Ion: 76 Resp: 382510

Ion Ratio Lower Upper

76 100

78 32.2 25.3 37.9



#58

2-Chloroethyl Vinyl ether

Concen: 526.621 ug/l

RT: 10.165 min Scan# 1396

Delta R.T. -0.000 min

Lab File: VN086866.D




Acq: 06 Jun 2025 14:26


Tgt Ion: 63 Resp: 1113523

Ion Ratio Lower Upper

63 100

106 29.1 23.0 34.6









#59
2-Hexanone
Concen: 567.535 ug/l
RT: 11.200 min Scan# 11
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100


Tgt Ion: 43 Resp: 1301539
Ion Ratio Lower Upper
43 100
58 61.4 30.4 91.3



#60
Dibromochloromethane
Concen: 102.616 ug/l
RT: 11.359 min Scan# 1599
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26

Tgt Ion: 129 Resp: 237965
Ion Ratio Lower Upper
129 100
127 77.1 38.6 115.8









#61
1,2-Dibromoethane
Concen: 101.787 ug/l
RT: 11.470 min Scan# 1618
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100


Tgt Ion:107 Resp: 232214
Ion Ratio Lower Upper
107 100
109 93.8 76.3 114.5



#62
4-Bromofluorobenzene
Concen: 102.339 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26

Tgt Ion: 95 Resp: 292385
Ion Ratio Lower Upper
95 100
174 71.6 0.0 141.8
176 68.6 0.0 132.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.870 min Scan# 1686
Delta R.T. 0.006 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100


Tgt Ion:117 Resp: 290313
Ion Ratio Lower Upper
117 100
82 54.3 44.6 67.0
119 31.9 25.5 38.3



#64
Tetrachloroethene
Concen: 98.824 ug/l
RT: 11.106 min Scan# 1556
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26

Tgt Ion:164 Resp: 181568
Ion Ratio Lower Upper
164 100
166 121.8 101.8 152.8
129 96.8 83.0 124.4
131 91.5 78.2 117.2









#65
Chlorobenzene
Concen: 98.793 ug/l
RT: 11.894 min Scan# 11
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100


Tgt Ion:112 Resp: 632563
Ion Ratio Lower Upper
112 100
114 31.3 26.2 39.2



#66
1,1,1,2-Tetrachloroethane
Concen: 100.521 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. -0.006 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26

Tgt Ion:131 Resp: 206895
Ion Ratio Lower Upper
131 100
133 96.4 48.3 144.8
119 68.2 34.0 101.8









#67
Ethyl Benzene
Concen: 100.717 ug/l
RT: 11.964 min Scan# 1702
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100


Tgt Ion: 91 Resp: 1110803
Ion Ratio Lower Upper
91 100
106 31.5 25.4 38.2



#68
m/p-Xylenes
Concen: 202.411 ug/l
RT: 12.070 min Scan# 1720
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26

Tgt Ion: 106 Resp: 854557
Ion Ratio Lower Upper
106 100
91 200.4 159.4 239.0









#69
o-Xylene
Concen: 102.070 ug/l
RT: 12.400 min Scan# 1776
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100


Tgt Ion:106 Resp: 412719
Ion Ratio Lower Upper
106 100
91 212.8 106.1 318.1



#70
Styrene
Concen: 103.127 ug/l
RT: 12.412 min Scan# 1778
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26

Tgt Ion:104 Resp: 713548
Ion Ratio Lower Upper
104 100
78 48.4 38.5 57.7
103 54.0 43.1 64.7









#71
Bromoform
Concen: 108.707 ug/l
RT: 12.576 min Scan# 11
Delta R.T. -0.006 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100


Tgt Ion:173 Resp: 165781
Ion Ratio Lower Upper
173 100
175 47.9 24.1 72.4
254 0.0 0.0 0.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.794 min Scan# 2013
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26

Tgt Ion:152 Resp: 147098
Ion Ratio Lower Upper
152 100
115 60.1 30.1 90.5
150 190.4 0.0 345.0









#73
Isopropylbenzene
Concen: 99.410 ug/l
RT: 12.694 min Scan# 11
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100


Tgt Ion:105 Resp: 1065297
Ion Ratio Lower Upper
105 100
120 27.0 13.5 40.4



#74
N-ethyl acetate
Concen: 113.475 ug/l
RT: 12.506 min Scan# 1794
Delta R.T. -0.006 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26

Tgt Ion: 43 Resp: 403587
Ion Ratio Lower Upper
43 100
70 47.7 37.6 56.4
55 25.0 19.0 28.6
61 27.3 20.8 31.2









#75
1,1,2,2-Tetrachloroethane
Concen: 97.616 ug/l
RT: 12.941 min Scan# 1868
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100


Tgt Ion: 83 Resp: 354386
Ion Ratio Lower Upper
83 100
131 9.9 4.9 14.7
85 64.9 32.0 96.0



#76
1,2,3-Trichloropropane
Concen: 138.300 ug/l
RT: 12.994 min Scan# 1877
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26

Tgt Ion: 75 Resp: 485083
Ion Ratio Lower Upper
75 100
77 118.2 79.3 237.9









#77
Bromobenzene
Concen: 100.266 ug/l
RT: 12.982 min Scan# 1875
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100


Tgt Ion:156 Resp: 246413
Ion Ratio Lower Upper
156 100
77 232.6 117.7 353.1
158 96.4 48.1 144.4



#78
n-propylbenzene
Concen: 99.948 ug/l
RT: 13.035 min Scan# 1884
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26

Tgt Ion: 91 Resp: 1301631
Ion Ratio Lower Upper
91 100
120 22.9 11.4 34.2





#79

2-Chlorotoluene

Concen: 98.630 ug/l

RT: 13.123 min Scan# 1899

Delta R.T. -0.000 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26

Instrument :

MSVOA_N

ClientSampleId :





VSTDICC100

Tgt Ion: 91 Resp: 770308

Ion Ratio Lower Upper

91 100

126 32.8 16.2 48.5



#80

1,3,5-Trimethylbenzene

Concen: 101.873 ug/l

RT: 13.170 min Scan# 1907

Delta R.T. -0.000 min

Lab File: VN086866.D




Acq: 06 Jun 2025 14:26


Tgt Ion:105 Resp: 901317

Ion Ratio Lower Upper

105 100

120 49.7 24.9 74.6









#81
trans-1,4-Dichloro-2-butene
Concen: 105.820 ug/l
RT: 12.735 min Scan# 1833
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100


Tgt Ion: 75 Resp: 160736
Ion Ratio Lower Upper
75 100
53 76.0 66.6 99.8
89 41.1 36.7 55.1



#82
4-Chlorotoluene
Concen: 99.166 ug/l
RT: 13.223 min Scan# 1916
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26

Tgt Ion: 91 Resp: 783657
Ion Ratio Lower Upper
91 100
126 33.2 16.7 50.1









#83
 tert-Butylbenzene
 Concen: 99.710 ug/l
 RT: 13.435 min Scan# 1952
 Delta R.T. -0.000 min
 Lab File: VN086866.D
 Acq: 06 Jun 2025 14:26

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC100

Tgt Ion:119 Resp: 807528




Ion	Ratio	Lower	Upper
119	100		
91	64.6	32.1	96.5
134	26.6	13.4	40.1




#84
 1,2,4-Trimethylbenzene
 Concen: 101.983 ug/l
 RT: 13.482 min Scan# 1960
 Delta R.T. -0.000 min
 Lab File: VN086866.D
 Acq: 06 Jun 2025 14:26

Tgt Ion:105 Resp: 904794

Ion	Ratio	Lower	Upper
105	100		
120	46.7	23.2	69.6









#85
 sec-Butylbenzene
 Concen: 100.513 ug/l
 RT: 13.617 min Scan# 1983
 Delta R.T. -0.000 min
 Lab File: VN086866.D
 Acq: 06 Jun 2025 14:26




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC100


Tgt Ion:105 Resp: 1182155
 Ion Ratio Lower Upper
 105 100
 134 20.2 10.0 30.0



#86
 p-Isopropyltoluene
 Concen: 101.857 ug/l
 RT: 13.729 min Scan# 2002
 Delta R.T. -0.000 min
 Lab File: VN086866.D
 Acq: 06 Jun 2025 14:26

Tgt Ion:119 Resp: 990273
 Ion Ratio Lower Upper
 119 100
 134 26.7 13.5 40.4
 91 22.6 11.5 34.4





#87

1,3-Dichlorobenzene

Concen: 98.063 ug/l

RT: 13.735 min Scan# 2003

Delta R.T. -0.000 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100





Tgt Ion:146 Resp: 474162

Ion Ratio Lower Upper

146 100

111 42.3 20.9 62.8

148 64.4 32.0 96.2



#88

1,4-Dichlorobenzene

Concen: 97.978 ug/l

RT: 13.811 min Scan# 2016

Delta R.T. -0.000 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26




Tgt Ion:146 Resp: 483003


Ion Ratio Lower Upper

146 100

111 39.3 19.9 59.7

148 63.0 31.8 95.4





#89

n-Butylbenzene

Concen: 99.598 ug/l

RT: 14.053 min Scan# 2057

Delta R.T. -0.000 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC100





Tgt Ion: 91 Resp: 937913

Ion Ratio Lower Upper

91 100

92 54.7 27.5 82.4

134 26.2 13.0 39.0



#90

Hexachloroethane

Concen: 103.019 ug/l

RT: 14.335 min Scan# 2105

Delta R.T. -0.000 min

Lab File: VN086866.D




Acq: 06 Jun 2025 14:26


Tgt Ion: 117 Resp: 169772

Ion Ratio Lower Upper

117 100

201 65.8 33.2 99.6









#91
 1,2-Dichlorobenzene
 Concen: 98.602 ug/l
 RT: 14.106 min Scan# 2066
 Delta R.T. -0.000 min
 Lab File: VN086866.D
 Acq: 06 Jun 2025 14:26




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC100


Tgt Ion:146 Resp: 458049
 Ion Ratio Lower Upper
 146 100
 111 43.5 21.6 64.6
 148 63.5 31.7 95.1



#92
 1,2-Dibromo-3-Chloropropane
 Concen: 95.877 ug/l
 RT: 14.717 min Scan# 2170
 Delta R.T. -0.000 min
 Lab File: VN086866.D
 Acq: 06 Jun 2025 14:26

Tgt Ion: 75 Resp: 83243
 Ion Ratio Lower Upper
 75 100
 155 86.1 41.1 123.4
 157 109.8 52.7 158.1





#93

1,2,4-Trichlorobenzene

Concen: 100.751 ug/l

RT: 15.394 min Scan# 2285

Delta R.T. 0.006 min

Lab File: VN086866.D


Acq: 06 Jun 2025 14:26

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100






Tgt Ion:180 Resp: 298870

Ion	Ratio	Lower	Upper

180	100		

182	94.6	46.9	140.7

145	32.2	16.0	47.9



#94

Hexachlorobutadiene

Concen: 101.789 ug/l

RT: 15.500 min Scan# 2303

Delta R.T. -0.000 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26




Tgt Ion:225 Resp: 112496


Ion	Ratio	Lower	Upper

225	100		

223	64.2	33.5	100.4

227	62.1	31.1	93.5









#95
Naphthalene
Concen: 102.290 ug/l
RT: 15.641 min Scan# 2327
Delta R.T. -0.000 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100

Tgt Ion:128 Resp: 1129434
Ion Ratio Lower Upper
128 100
127 12.6 10.2 15.2
129 11.0 8.8 13.2



#96
1,2,3-Trichlorobenzene
Concen: 99.796 ug/l
RT: 15.841 min Scan# 2361
Delta R.T. 0.006 min
Lab File: VN086866.D
Acq: 06 Jun 2025 14:26

Tgt Ion:180 Resp: 294127
Ion Ratio Lower Upper
180 100
182 95.3 47.0 141.0
145 33.5 16.6 49.8



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060625\
 Data File : VN086867.D
 Acq On : 06 Jun 2025 14:49
 Operator : JC\MD
 Sample : VSTDICC150
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC150

Quant Time: Jun 07 01:59:15 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 01:51:02 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	172832	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	307937	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.864	117	278552	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	136630	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	389297	168.233	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 336.460%#	
35) Dibromofluoromethane	8.176	113	324474	177.803	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 355.600%#	
50) Toluene-d8	10.570	98	1272348	176.120	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 352.240%#	
62) 4-Bromofluorobenzene	12.847	95	481421	179.364	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 358.720%#	

Target Compounds				Qvalue		
2) Dichlorodifluoromethane	2.153	85	262143	152.122	ug/l	100
3) Chloromethane	2.394	50	304558	136.865	ug/l	100
4) Vinyl Chloride	2.553	62	335834	146.303	ug/l	100
5) Bromomethane	2.965	94	190786	148.522	ug/l	98
6) Chloroethane	3.136	64	208568	140.664	ug/l	96
7) Trichlorofluoromethane	3.518	101	427871	142.666	ug/l	100
8) Diethyl Ether	3.983	74	191555	146.593	ug/l	99
9) 1,1,2-Trichlorotrifluo...	4.388	101	269810	143.192	ug/l	100
10) Methyl Iodide	4.606	142	352020	144.118	ug/l	99
11) Tert butyl alcohol	5.553	59	430866	686.213	ug/l	100
12) 1,1-Dichloroethene	4.353	96	273388	142.118	ug/l	98
13) Acrolein	4.194	56	171625	863.526	ug/l	99
14) Allyl chloride	5.035	41	490852	153.858	ug/l	92
15) Acrylonitrile	5.735	53	1048372	714.344	ug/l	99
16) Acetone	4.447	43	819041	667.435	ug/l	100
17) Carbon Disulfide	4.730	76	742788	139.594	ug/l	100
18) Methyl Acetate	5.041	43	523996	146.527	ug/l	100
19) Methyl tert-butyl Ether	5.818	73	998781	143.396	ug/l	99
20) Methylene Chloride	5.300	84	311597	135.628	ug/l	99
21) trans-1,2-Dichloroethene	5.806	96	290960	135.947	ug/l	97
22) Diisopropyl ether	6.688	45	947638	140.915	ug/l	98
23) Vinyl Acetate	6.618	43	3932872	692.189	ug/l	99
24) 1,1-Dichloroethane	6.582	63	540881	139.761	ug/l	99
25) 2-Butanone	7.494	43	1381940	692.809	ug/l	99
26) 2,2-Dichloropropane	7.500	77	450112	149.519	ug/l	99
27) cis-1,2-Dichloroethene	7.500	96	363398	141.974	ug/l	99
28) Bromochloromethane	7.824	49	290594	152.718	ug/l	98
29) Tetrahydrofuran	7.847	42	898159	691.211	ug/l	98
30) Chloroform	7.976	83	534038	138.179	ug/l	97
31) Cyclohexane	8.265	56	505795	134.766	ug/l	96
32) 1,1,1-Trichloroethane	8.176	97	462854	140.807	ug/l	99
36) 1,1-Dichloropropene	8.376	75	393166	144.584	ug/l	100
37) Ethyl Acetate	7.571	43	494000	142.706	ug/l	99
38) Carbon Tetrachloride	8.371	117	388595	145.557	ug/l	98
39) Methylcyclohexane	9.606	83	544437	146.136	ug/l	97
40) Benzene	8.612	78	1266139	142.388	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060625\
 Data File : VN086867.D
 Acq On : 06 Jun 2025 14:49
 Operator : JC\MD
 Sample : VSTDICC150
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC150

Quant Time: Jun 07 01:59:15 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 01:51:02 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.788	41	276400	142.174	ug/l	98
42) 1,2-Dichloroethane	8.676	62	381516	141.455	ug/l	100
43) Isopropyl Acetate	8.694	43	780791	140.531	ug/l	99
44) Trichloroethene	9.359	130	302548	143.490	ug/l	96
45) 1,2-Dichloropropane	9.623	63	309047	142.856	ug/l	100
46) Dibromomethane	9.712	93	208461	145.074	ug/l	100
47) Bromodichloromethane	9.894	83	430011	145.449	ug/l	100
48) Methyl methacrylate	9.688	41	366967	143.502	ug/l	99
49) 1,4-Dioxane	9.712	88	138035	2967.116	ug/l #	100
51) 4-Methyl-2-Pentanone	10.447	43	2438311	728.995	ug/l	98
52) Toluene	10.635	92	793270	145.976	ug/l	100
53) t-1,3-Dichloropropene	10.841	75	489809	148.163	ug/l	97
54) cis-1,3-Dichloropropene	10.318	75	518092	146.469	ug/l	99
55) 1,1,2-Trichloroethane	11.017	97	298151	142.588	ug/l	99
56) Ethyl methacrylate	10.876	69	533912	160.311	ug/l	98
57) 1,3-Dichloropropane	11.165	76	516391	142.363	ug/l	98
58) 2-Chloroethyl Vinyl ether	10.165	63	1869543	941.147	ug/l	99
59) 2-Hexanone	11.200	43	1735777	805.662	ug/l	99
60) Dibromochloromethane	11.365	129	322064	147.832	ug/l	98
61) 1,2-Dibromoethane	11.470	107	314606	146.790	ug/l	99
64) Tetrachloroethene	11.106	164	245122	139.048	ug/l	97
65) Chlorobenzene	11.894	112	860345	140.041	ug/l	98
66) 1,1,1,2-Tetrachloroethane	11.959	131	282739	143.171	ug/l	99
67) Ethyl Benzene	11.964	91	1517883	143.438	ug/l	98
68) m/p-Xylenes	12.070	106	1175378	290.155	ug/l	99
69) o-Xylene	12.400	106	566301	145.966	ug/l	100
70) Styrene	12.411	104	972318	146.459	ug/l	100
71) Bromoform	12.582	173	223200	152.538	ug/l #	99
73) Isopropylbenzene	12.694	105	1453651	146.042	ug/l	100
74) N-amyl acetate	12.500	43	543747	164.596	ug/l	96
75) 1,1,2,2-Tetrachloroethane	12.941	83	474426	140.693	ug/l	100
76) 1,2,3-Trichloropropane	12.994	75	644184	197.732	ug/l	71
77) Bromobenzene	12.982	156	335734	147.077	ug/l	98
78) n-propylbenzene	13.035	91	1769694	146.301	ug/l	99
79) 2-Chlorotoluene	13.123	91	1046916	144.317	ug/l	99
80) 1,3,5-Trimethylbenzene	13.170	105	1210325	147.279	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.735	75	216253	153.277	ug/l	95
82) 4-Chlorotoluene	13.223	91	1063581	144.900	ug/l	100
83) tert-Butylbenzene	13.435	119	1093406	145.352	ug/l	100
84) 1,2,4-Trimethylbenzene	13.482	105	1216529	147.626	ug/l	100
85) sec-Butylbenzene	13.617	105	1582762	144.885	ug/l	99
86) p-Isopropyltoluene	13.729	119	1322780	146.482	ug/l	99
87) 1,3-Dichlorobenzene	13.735	146	641966	142.939	ug/l	100
88) 1,4-Dichlorobenzene	13.811	146	646188	141.124	ug/l	99
89) n-Butylbenzene	14.053	91	1251819	143.117	ug/l	100
90) Hexachloroethane	14.335	117	230195	150.387	ug/l	98
91) 1,2-Dichlorobenzene	14.105	146	613177	142.108	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	14.717	75	110677	137.242	ug/l	96
93) 1,2,4-Trichlorobenzene	15.394	180	407621	147.940	ug/l	99
94) Hexachlorobutadiene	15.500	225	151172	147.264	ug/l	98
95) Naphthalene	15.641	128	1545981	150.743	ug/l	100
96) 1,2,3-Trichlorobenzene	15.841	180	404450	147.743	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060625\
Data File : VN086867.D
Acq On : 06 Jun 2025 14:49
Operator : JC\MD
Sample : VSTDICC150
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC150

Quant Time: Jun 07 01:59:15 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 01:51:02 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060625\
Data File : VN086867.D
Acq On : 06 Jun 2025 14:49
Operator : JC\MD
Sample : VSTDICC150
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC150

Quant Time: Jun 07 01:59:15 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 01:51:02 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.229 min Scan# 1067
Delta R.T. -0.007 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150


Tgt Ion:168 Resp: 172832
Ion Ratio Lower Upper
168 100
99 61.8 49.1 73.7



#2
Dichlorodifluoromethane
Concen: 152.122 ug/l
RT: 2.153 min Scan# 34
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49

Tgt Ion: 85 Resp: 262143
Ion Ratio Lower Upper
85 100
87 32.3 16.2 48.6









#3
Chloromethane
Concen: 136.865 ug/l
RT: 2.394 min Scan# 75
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150


Tgt Ion: 50 Resp: 304558
Ion Ratio Lower Upper
50 100
52 32.2 25.7 38.5



#4
Vinyl Chloride
Concen: 146.303 ug/l
RT: 2.553 min Scan# 102
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49

Tgt Ion: 62 Resp: 335834
Ion Ratio Lower Upper
62 100
64 32.6 26.0 39.0









#5
Bromomethane
Concen: 148.522 ug/l
RT: 2.965 min Scan# 11
Delta R.T. -0.030 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150


Tgt Ion: 94 Resp: 190786
Ion Ratio Lower Upper
94 100
96 95.5 77.9 116.9



#6
Chloroethane
Concen: 140.664 ug/l
RT: 3.136 min Scan# 201
Delta R.T. -0.018 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49

Tgt Ion: 64 Resp: 208568
Ion Ratio Lower Upper
64 100
66 32.7 24.6 36.8









#7
 Trichlorofluoromethane
 Concen: 142.666 ug/l
 RT: 3.518 min Scan# 20
 Delta R.T. -0.006 min
 Lab File: VN086867.D
 Acq: 06 Jun 2025 14:49




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC150


Tgt Ion:101 Resp: 427871
 Ion Ratio Lower Upper
 101 100
 103 65.5 52.2 78.2



#8
 Diethyl Ether
 Concen: 146.593 ug/l
 RT: 3.983 min Scan# 345
 Delta R.T. -0.000 min
 Lab File: VN086867.D
 Acq: 06 Jun 2025 14:49

Tgt Ion: 74 Resp: 191555
 Ion Ratio Lower Upper
 74 100
 45 90.3 45.5 136.5









#9
1,1,2-Trichlorotrifluoroethane
Concen: 143.192 ug/l
RT: 4.388 min Scan# 415
Delta R.T. -0.006 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150


Tgt Ion:101 Resp: 269810
Ion Ratio Lower Upper
101 100
85 42.3 33.8 50.8
151 73.5 58.7 88.1



#10
Methyl Iodide
Concen: 144.118 ug/l
RT: 4.606 min Scan# 451
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49

Tgt Ion:142 Resp: 352020
Ion Ratio Lower Upper
142 100
127 39.1 30.9 46.3
141 13.5 10.4 15.6









#11
Tert butyl alcohol
Concen: 686.213 ug/l
RT: 5.553 min Scan# 611
Delta R.T. 0.006 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150


Tgt Ion: 59 Resp: 430866
Ion Ratio Lower Upper
59 100
57 9.9 8.1 12.1



#12
1,1-Dichloroethene
Concen: 142.118 ug/l
RT: 4.353 min Scan# 408
Delta R.T. -0.006 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49

Tgt Ion: 96 Resp: 273388
Ion Ratio Lower Upper
96 100
61 154.5 125.8 188.6
98 64.9 52.5 78.7





#13

Acrolein

Concen: 863.526 ug/l

RT: 4.194 min Scan# 381

Delta R.T. -0.006 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49

Instrument :

MSVOA_N

ClientSampleId :





VSTDICC150

Tgt Ion: 56 Resp: 171625

Ion Ratio Lower Upper

56 100

55 70.4 55.6 83.4



#14

Allyl chloride

Concen: 153.858 ug/l

RT: 5.035 min Scan# 524

Delta R.T. -0.006 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49




Tgt Ion: 41 Resp: 490852


Ion Ratio Lower Upper

41 100

39 56.4 51.3 76.9

76 35.7 31.4 47.0









#15
Acrylonitrile
Concen: 714.344 ug/l
RT: 5.735 min Scan# 644
Delta R.T. -0.006 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC150

Tgt Ion: 53 Resp: 1048372




Ion	Ratio	Lower	Upper
53	100		
52	80.7	65.0	97.6
51	33.8	27.2	40.8




#16
Acetone
Concen: 667.435 ug/l
RT: 4.447 min Scan# 424
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49

Tgt Ion: 43 Resp: 819041

Ion	Ratio	Lower	Upper
43	100		
58	34.7	28.0	42.0









#17
Carbon Disulfide
Concen: 139.594 ug/l
RT: 4.730 min Scan# 472
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150


Tgt Ion: 76 Resp: 742788
Ion Ratio Lower Upper
76 100
78 9.3 7.3 10.9



#18
Methyl Acetate
Concen: 146.527 ug/l
RT: 5.041 min Scan# 525
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49

Tgt Ion: 43 Resp: 523996
Ion Ratio Lower Upper
43 100
74 26.1 20.7 31.1









#19
Methyl tert-butyl Ether
Concen: 143.396 ug/l
RT: 5.818 min Scan# 61
Delta R.T. 0.006 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150


Tgt Ion: 73 Resp: 998781
Ion Ratio Lower Upper
73 100
57 22.2 17.4 26.2



#20
Methylene Chloride
Concen: 135.628 ug/l
RT: 5.300 min Scan# 569
Delta R.T. 0.006 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49

Tgt Ion: 84 Resp: 311597
Ion Ratio Lower Upper
84 100
49 113.0 90.5 135.7
51 36.1 28.5 42.7
86 62.3 48.1 72.1









#21
trans-1,2-Dichloroethene
Concen: 135.947 ug/l
RT: 5.806 min Scan# 61
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150


Tgt Ion: 96 Resp: 290960
Ion Ratio Lower Upper
96 100
61 132.7 109.2 163.8
98 65.0 49.8 74.8



#22
Diisopropyl ether
Concen: 140.915 ug/l
RT: 6.688 min Scan# 805
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49

Tgt Ion: 45 Resp: 947638
Ion Ratio Lower Upper
45 100
43 53.1 41.9 62.9
87 29.9 22.4 33.6
59 13.2 9.8 14.6









#23
 Vinyl Acetate
 Concen: 692.189 ug/l
 RT: 6.618 min Scan# 793
 Delta R.T. -0.000 min
 Lab File: VN086867.D
 Acq: 06 Jun 2025 14:49




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC150


Tgt Ion: 43 Resp: 3932872
 Ion Ratio Lower Upper
 43 100
 86 12.1 9.5 14.3



#24
 1,1-Dichloroethane
 Concen: 139.761 ug/l
 RT: 6.582 min Scan# 787
 Delta R.T. -0.000 min
 Lab File: VN086867.D
 Acq: 06 Jun 2025 14:49

Tgt Ion: 63 Resp: 540881
 Ion Ratio Lower Upper
 63 100
 98 7.3 3.6 10.8
 100 4.4 2.6 7.8









#25
2-Butanone
Concen: 692.809 ug/l
RT: 7.494 min Scan# 942
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150


Tgt Ion: 43 Resp: 1381940
Ion Ratio Lower Upper
43 100
72 29.4 23.0 34.6



#26
2,2-Dichloropropane
Concen: 149.519 ug/l
RT: 7.500 min Scan# 943
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49

Tgt Ion: 77 Resp: 450112
Ion Ratio Lower Upper
77 100
97 23.4 11.6 34.7









#27
 cis-1,2-Dichloroethene
 Concen: 141.974 ug/l
 RT: 7.500 min Scan# 943
 Delta R.T. -0.000 min
 Lab File: VN086867.D
 Acq: 06 Jun 2025 14:49




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC150


Tgt Ion: 96 Resp: 363398
 Ion Ratio Lower Upper
 96 100
 61 138.0 0.0 278.0
 98 64.5 0.0 128.2



#28
 Bromochloromethane
 Concen: 152.718 ug/l
 RT: 7.824 min Scan# 998
 Delta R.T. -0.000 min
 Lab File: VN086867.D
 Acq: 06 Jun 2025 14:49

Tgt Ion: 49 Resp: 290594
 Ion Ratio Lower Upper
 49 100
 129 2.0 0.0 3.8
 130 79.8 62.2 93.2





#29

Tetrahydrofuran

Concen: 691.211 ug/l

RT: 7.847 min Scan# 1002

Delta R.T. -0.006 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC150





Tgt Ion: 42 Resp: 898159

Ion Ratio Lower Upper

42 100

72 48.5 37.8 56.6

71 45.4 35.6 53.4



#30

Chloroform

Concen: 138.179 ug/l

RT: 7.976 min Scan# 1024

Delta R.T. -0.000 min

Lab File: VN086867.D




Acq: 06 Jun 2025 14:49


Tgt Ion: 83 Resp: 534038

Ion Ratio Lower Upper

83 100

85 63.4 52.9 79.3









#31
Cyclohexane
Concen: 134.766 ug/l
RT: 8.265 min Scan# 1073
Delta R.T. -0.006 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150


Tgt Ion: 56 Resp: 505795
Ion Ratio Lower Upper
56 100
69 32.0 25.4 38.0
84 90.6 68.8 103.2



#32
1,1,1-Trichloroethane
Concen: 140.807 ug/l
RT: 8.176 min Scan# 1058
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49

Tgt Ion: 97 Resp: 462854
Ion Ratio Lower Upper
97 100
99 64.9 52.0 78.0
61 43.2 35.9 53.9





#33

1,2-Dichloroethane-d4

Concen: 168.233 ug/l

RT: 8.588 min Scan# 1128

Delta R.T. -0.000 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC150

Tgt Ion: 65 Resp: 389297

Ion Ratio Lower Upper

65 100

67 54.9 0.0 105.6



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.106 min Scan# 1216

Delta R.T. -0.000 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49




Tgt Ion: 114 Resp: 307937


Ion Ratio Lower Upper

114 100

63 19.8 0.0 39.6

88 16.5 0.0 30.2





#35

Dibromofluoromethane

Concen: 177.803 ug/l

RT: 8.176 min Scan# 1057

Delta R.T. 0.006 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC150





Tgt Ion:113 Resp: 324474

Ion Ratio Lower Upper

113 100

111 102.5 84.2 126.2

192 17.7 14.2 21.4



#36

1,1-Dichloropropene

Concen: 144.584 ug/l

RT: 8.376 min Scan# 1092

Delta R.T. -0.000 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49




Tgt Ion: 75 Resp: 393166


Ion Ratio Lower Upper

75 100

110 35.6 17.6 52.9

77 31.4 25.3 37.9









#37
Ethyl Acetate
Concen: 142.706 ug/l
RT: 7.571 min Scan# 91
Delta R.T. -0.006 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150


Tgt Ion: 43 Resp: 494000
Ion Ratio Lower Upper
43 100
61 15.4 12.1 18.1
70 12.7 9.8 14.6



#38
Carbon Tetrachloride
Concen: 145.557 ug/l
RT: 8.371 min Scan# 1091
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49

Tgt Ion: 117 Resp: 388595
Ion Ratio Lower Upper
117 100
119 94.9 75.0 112.6
121 29.1 24.9 37.3





#39

Methylcyclohexane

Concen: 146.136 ug/l

RT: 9.606 min Scan# 1130

Delta R.T. -0.000 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC150





Tgt Ion: 83 Resp: 544437

Ion Ratio Lower Upper

83 100

55 73.6 61.0 91.6

98 47.4 38.9 58.3



#40

Benzene

Concen: 142.388 ug/l

RT: 8.612 min Scan# 1132

Delta R.T. -0.000 min

Lab File: VN086867.D




Acq: 06 Jun 2025 14:49


Tgt Ion: 78 Resp: 1266139

Ion Ratio Lower Upper

78 100

77 23.6 18.7 28.1









#41
Methacrylonitrile
Concen: 142.174 ug/l
RT: 7.788 min Scan# 992
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49

Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150

Tgt Ion: 41 Resp: 276400




Ion	Ratio	Lower	Upper
41	100		
39	50.9	41.2	61.8
67	79.4	61.4	92.2
52	29.0	22.7	34.1




#42
1,2-Dichloroethane
Concen: 141.455 ug/l
RT: 8.676 min Scan# 1143
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49

Tgt Ion: 62 Resp: 381516

Ion	Ratio	Lower	Upper
62	100		
98	10.5	0.0	21.2









#43
Isopropyl Acetate
Concen: 140.531 ug/l
RT: 8.694 min Scan# 1146
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150


Tgt Ion: 43 Resp: 780791
Ion Ratio Lower Upper
43 100
61 28.3 22.1 33.1
87 15.3 11.8 17.6



#44
Trichloroethene
Concen: 143.490 ug/l
RT: 9.359 min Scan# 1259
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49

Tgt Ion: 130 Resp: 302548
Ion Ratio Lower Upper
130 100
95 101.6 0.0 196.2





#45

1,2-Dichloropropane

Concen: 142.856 ug/l

RT: 9.623 min Scan# 1319

Delta R.T. -0.006 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC150

Tgt Ion: 63 Resp: 309047

Ion Ratio Lower Upper

63 100

65 31.9 25.5 38.3



#46

Dibromomethane

Concen: 145.074 ug/l

RT: 9.712 min Scan# 1319

Delta R.T. -0.006 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49




Tgt Ion: 93 Resp: 208461


Ion Ratio Lower Upper

93 100

95 84.8 67.1 100.7

174 92.8 74.2 111.4









#47
Bromodichloromethane
Concen: 145.449 ug/l
RT: 9.894 min Scan# 1350
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150


Tgt Ion: 83 Resp: 430011
Ion Ratio Lower Upper
83 100
85 63.6 50.9 76.3
127 8.1 6.6 10.0



#48
Methyl methacrylate
Concen: 143.502 ug/l
RT: 9.688 min Scan# 1315
Delta R.T. 0.006 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49

Tgt Ion: 41 Resp: 366967
Ion Ratio Lower Upper
41 100
69 94.2 74.6 112.0
39 50.8 40.2 60.4





#49

1,4-Dioxane

Concen: 2967.116 ug/l

RT: 9.712 min Scan# 1319

Delta R.T. 0.011 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC150





Tgt Ion: 88 Resp: 138035

Ion Ratio Lower Upper

88 100

43 0.0 0.0 0.0

58 72.2 57.7 86.5



#50

Toluene-d8

Concen: 176.120 ug/l

RT: 10.570 min Scan# 1465

Delta R.T. -0.000 min

Lab File: VN086867.D




Acq: 06 Jun 2025 14:49


Tgt Ion: 98 Resp: 1272348

Ion Ratio Lower Upper

98 100

100 67.5 53.4 80.0









#51
4-Methyl-2-Pentanone
Concen: 728.995 ug/l
RT: 10.447 min Scan# 1444
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150


Tgt Ion: 43 Resp: 2438311
Ion Ratio Lower Upper
43 100
58 44.0 34.2 51.4



#52
Toluene
Concen: 145.976 ug/l
RT: 10.635 min Scan# 1476
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49

Tgt Ion: 92 Resp: 793270
Ion Ratio Lower Upper
92 100
91 170.2 136.5 204.7









#53
t-1,3-Dichloropropene
Concen: 148.163 ug/l
RT: 10.841 min Scan# 1511
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150


Tgt Ion: 75 Resp: 489809
Ion Ratio Lower Upper
75 100
77 31.4 26.4 39.6



#54
cis-1,3-Dichloropropene
Concen: 146.469 ug/l
RT: 10.318 min Scan# 1422
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49

Tgt Ion: 75 Resp: 518092
Ion Ratio Lower Upper
75 100
77 31.4 26.2 39.4
39 40.6 32.3 48.5









#55
1,1,2-Trichloroethane
Concen: 142.588 ug/l
RT: 11.017 min Scan# 1541
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150


Tgt Ion:	97	Resp:	298151
Ion	Ratio	Lower	Upper
97	100		
83	85.6	69.7	104.5
85	55.2	45.1	67.7
99	63.1	50.6	75.8



#56
Ethyl methacrylate
Concen: 160.311 ug/l
RT: 10.876 min Scan# 1517
Delta R.T. -0.006 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49

Tgt Ion:	69	Resp:	533912
Ion	Ratio	Lower	Upper
69	100		
41	57.5	47.2	70.8
39	27.0	23.0	34.6





#57

1,3-Dichloropropane

Concen: 142.363 ug/l

RT: 11.165 min Scan# 11

Delta R.T. -0.000 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC150

Tgt Ion: 76 Resp: 516391

Ion Ratio Lower Upper

76 100

78 32.5 25.3 37.9



#58

2-Chloroethyl Vinyl ether

Concen: 941.147 ug/l

RT: 10.165 min Scan# 1396

Delta R.T. -0.000 min

Lab File: VN086867.D




Acq: 06 Jun 2025 14:49


Tgt Ion: 63 Resp: 1869543

Ion Ratio Lower Upper

63 100

106 29.2 23.0 34.6









#59
2-Hexanone
Concen: 805.662 ug/l
RT: 11.200 min Scan# 11
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150


Tgt Ion: 43 Resp: 1735777
Ion Ratio Lower Upper
43 100
58 61.6 30.4 91.3



#60
Dibromochloromethane
Concen: 147.832 ug/l
RT: 11.365 min Scan# 1600
Delta R.T. 0.006 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49

Tgt Ion:129 Resp: 322064
Ion Ratio Lower Upper
129 100
127 78.6 38.6 115.8









#61
1,2-Dibromoethane
Concen: 146.790 ug/l
RT: 11.470 min Scan# 1618
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150


Tgt Ion:107 Resp: 314606
Ion Ratio Lower Upper
107 100
109 94.0 76.3 114.5



#62
4-Bromofluorobenzene
Concen: 179.364 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49

Tgt Ion: 95 Resp: 481421
Ion Ratio Lower Upper
95 100
174 70.7 0.0 141.8
176 67.8 0.0 132.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.864 min Scan# 1
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150


Tgt Ion:117 Resp: 278552
Ion Ratio Lower Upper
117 100
82 54.7 44.6 67.0
119 31.3 25.5 38.3



#64
Tetrachloroethene
Concen: 139.048 ug/l
RT: 11.106 min Scan# 1556
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49

Tgt Ion:164 Resp: 245122
Ion Ratio Lower Upper
164 100
166 126.7 101.8 152.8
129 96.6 83.0 124.4
131 94.9 78.2 117.2









#65
Chlorobenzene
Concen: 140.041 ug/l
RT: 11.894 min Scan# 11
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150


Tgt Ion:112 Resp: 860345
Ion Ratio Lower Upper
112 100
114 31.8 26.2 39.2



#66
1,1,1,2-Tetrachloroethane
Concen: 143.171 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. -0.006 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49

Tgt Ion:131 Resp: 282739
Ion Ratio Lower Upper
131 100
133 95.2 48.3 144.8
119 67.8 34.0 101.8









#67
Ethyl Benzene
Concen: 143.438 ug/l
RT: 11.964 min Scan# 1702
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150


Tgt Ion: 91 Resp: 1517883
Ion Ratio Lower Upper
91 100
106 30.9 25.4 38.2



#68
m/p-Xylenes
Concen: 290.155 ug/l
RT: 12.070 min Scan# 1720
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49

Tgt Ion:106 Resp: 1175378
Ion Ratio Lower Upper
106 100
91 200.8 159.4 239.0









#69
o-Xylene
Concen: 145.966 ug/l
RT: 12.400 min Scan# 1776
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150


Tgt Ion:106 Resp: 566301
Ion Ratio Lower Upper
106 100
91 211.9 106.1 318.1



#70
Styrene
Concen: 146.459 ug/l
RT: 12.411 min Scan# 1778
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49

Tgt Ion:104 Resp: 972318
Ion Ratio Lower Upper
104 100
78 48.4 38.5 57.7
103 54.0 43.1 64.7





#71

Bromoform

Concen: 152.538 ug/l

RT: 12.582 min Scan# 11

Delta R.T. -0.000 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC150





Tgt Ion:173 Resp: 223200

Ion Ratio Lower Upper

173 100

175 48.8 24.1 72.4

254 0.0 0.0 0.0



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.794 min Scan# 2013

Delta R.T. -0.000 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49




Tgt Ion:152 Resp: 136630


Ion Ratio Lower Upper

152 100

115 61.1 30.1 90.5

150 206.7 0.0 345.0









#73
Isopropylbenzene
Concen: 146.042 ug/l
RT: 12.694 min Scan# 11
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150


Tgt Ion:105 Resp: 1453651
Ion Ratio Lower Upper
105 100
120 26.9 13.5 40.4



#74
N-ethyl acetate
Concen: 164.596 ug/l
RT: 12.500 min Scan# 1793
Delta R.T. -0.012 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49

Tgt Ion: 43 Resp: 543747
Ion Ratio Lower Upper
43 100
70 50.0 37.6 56.4
55 26.1 19.0 28.6
61 27.6 20.8 31.2









#75
1,1,2,2-Tetrachloroethane
Concen: 140.693 ug/l
RT: 12.941 min Scan# 1868
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150


Tgt Ion: 83 Resp: 474426
Ion Ratio Lower Upper
83 100
131 9.9 4.9 14.7
85 64.3 32.0 96.0



#76
1,2,3-Trichloropropane
Concen: 197.732 ug/l
RT: 12.994 min Scan# 1877
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49

Tgt Ion: 75 Resp: 644184
Ion Ratio Lower Upper
75 100
77 120.6 79.3 237.9





#77

Bromobenzene

Concen: 147.077 ug/l

RT: 12.982 min Scan# 1875

Delta R.T. -0.000 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC150





Tgt Ion:156 Resp: 335734

Ion Ratio Lower Upper

156 100

77 231.5 117.7 353.1

158 96.3 48.1 144.4



#78

n-propylbenzene

Concen: 146.301 ug/l

RT: 13.035 min Scan# 1884

Delta R.T. -0.000 min

Lab File: VN086867.D




Acq: 06 Jun 2025 14:49


Tgt Ion: 91 Resp: 1769694

Ion Ratio Lower Upper

91 100

120 23.2 11.4 34.2









#79
2-Chlorotoluene
Concen: 144.317 ug/l
RT: 13.123 min Scan# 1899
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150


Tgt Ion: 91 Resp: 1046916
Ion Ratio Lower Upper
91 100
126 32.8 16.2 48.5



#80
1,3,5-Trimethylbenzene
Concen: 147.279 ug/l
RT: 13.170 min Scan# 1907
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49

Tgt Ion:105 Resp: 1210325
Ion Ratio Lower Upper
105 100
120 50.0 24.9 74.6









#81
trans-1,4-Dichloro-2-butene
Concen: 153.277 ug/l
RT: 12.735 min Scan# 1833
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150


Tgt Ion: 75 Resp: 216253
Ion Ratio Lower Upper
75 100
53 81.3 66.6 99.8
89 40.0 36.7 55.1



#82
4-Chlorotoluene
Concen: 144.900 ug/l
RT: 13.223 min Scan# 1916
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49

Tgt Ion: 91 Resp: 1063581
Ion Ratio Lower Upper
91 100
126 33.4 16.7 50.1









#83
tert-Butylbenzene
Concen: 145.352 ug/l
RT: 13.435 min Scan# 1952
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150


Tgt Ion:119 Resp: 1093406
Ion Ratio Lower Upper
119 100
91 63.8 32.1 96.5
134 26.7 13.4 40.1



#84
1,2,4-Trimethylbenzene
Concen: 147.626 ug/l
RT: 13.482 min Scan# 1960
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49

Tgt Ion:105 Resp: 1216529
Ion Ratio Lower Upper
105 100
120 46.7 23.2 69.6









#85
 sec-Butylbenzene
 Concen: 144.885 ug/l
 RT: 13.617 min Scan# 1983
 Delta R.T. -0.000 min
 Lab File: VN086867.D
 Acq: 06 Jun 2025 14:49




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC150


Tgt Ion:105 Resp: 1582762
 Ion Ratio Lower Upper
 105 100
 134 20.3 10.0 30.0



#86
 p-Isopropyltoluene
 Concen: 146.482 ug/l
 RT: 13.729 min Scan# 2002
 Delta R.T. -0.000 min
 Lab File: VN086867.D
 Acq: 06 Jun 2025 14:49

Tgt Ion:119 Resp: 1322780
 Ion Ratio Lower Upper
 119 100
 134 26.4 13.5 40.4
 91 22.7 11.5 34.4





#87

1,3-Dichlorobenzene

Concen: 142.939 ug/l

RT: 13.735 min Scan# 2003

Delta R.T. -0.000 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC150





Tgt Ion:146 Resp: 641966

Ion Ratio Lower Upper

146 100

111 42.1 20.9 62.8

148 63.8 32.0 96.2



#88

1,4-Dichlorobenzene

Concen: 141.124 ug/l

RT: 13.811 min Scan# 2016

Delta R.T. -0.000 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49




Tgt Ion:146 Resp: 646188


Ion Ratio Lower Upper

146 100

111 40.2 19.9 59.7

148 63.2 31.8 95.4









#89
n-Butylbenzene
Concen: 143.117 ug/l
RT: 14.053 min Scan# 2057
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150


Tgt Ion: 91 Resp: 1251819
Ion Ratio Lower Upper
91 100
92 54.6 27.5 82.4
134 26.1 13.0 39.0



#90
Hexachloroethane
Concen: 150.387 ug/l
RT: 14.335 min Scan# 2105
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49

Tgt Ion: 117 Resp: 230195
Ion Ratio Lower Upper
117 100
201 64.5 33.2 99.6









#91
1,2-Dichlorobenzene
Concen: 142.108 ug/l
RT: 14.105 min Scan# 2066
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150


Tgt Ion:146 Resp: 613177
Ion Ratio Lower Upper
146 100
111 43.4 21.6 64.6
148 63.8 31.7 95.1



#92
1,2-Dibromo-3-Chloropropane
Concen: 137.242 ug/l
RT: 14.717 min Scan# 2170
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49

Tgt Ion: 75 Resp: 110677
Ion Ratio Lower Upper
75 100
155 85.5 41.1 123.4
157 110.2 52.7 158.1





#93

1,2,4-Trichlorobenzene

Concen: 147.940 ug/l

RT: 15.394 min Scan# 2285

Delta R.T. 0.006 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC150





Tgt Ion:180 Resp: 407621

Ion Ratio Lower Upper

180 100

182 95.0 46.9 140.7

145 32.0 16.0 47.9



#94

Hexachlorobutadiene

Concen: 147.264 ug/l

RT: 15.500 min Scan# 2303

Delta R.T. -0.000 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49




Tgt Ion:225 Resp: 151172


Ion Ratio Lower Upper

225 100

223 64.8 33.5 100.4

227 63.0 31.1 93.5









#95
Naphthalene
Concen: 150.743 ug/l
RT: 15.641 min Scan# 2327
Delta R.T. -0.000 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150

Tgt Ion:128 Resp: 1545981
Ion Ratio Lower Upper
128 100
127 12.6 10.2 15.2
129 10.9 8.8 13.2



#96
1,2,3-Trichlorobenzene
Concen: 147.743 ug/l
RT: 15.841 min Scan# 2361
Delta R.T. 0.006 min
Lab File: VN086867.D
Acq: 06 Jun 2025 14:49

Tgt Ion:180 Resp: 404450
Ion Ratio Lower Upper
180 100
182 94.7 47.0 141.0
145 33.6 16.6 49.8



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060625\
 Data File : VN086869.D
 Acq On : 06 Jun 2025 15:54
 Operator : JC\MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ICVVN060625

Quant Time: Jun 07 02:16:19 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	218122	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	394140	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	348935	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	172710	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	152255	52.135	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 104.260%	
35) Dibromofluoromethane	8.177	113	127102	54.416	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 108.840%	
50) Toluene-d8	10.571	98	491265	53.129	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 106.260%	
62) 4-Bromofluorobenzene	12.847	95	185792	54.081	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 108.160%	

Target Compounds				Qvalue		
2) Dichlorodifluoromethane	2.154	85	120237	55.286	ug/l	99
3) Chloromethane	2.401	50	139027	49.505	ug/l	99
4) Vinyl Chloride	2.554	62	151951	52.451	ug/l	99
5) Bromomethane	2.989	94	89638	55.292	ug/l	95
6) Chloroethane	3.154	64	95654	51.117	ug/l	96
7) Trichlorofluoromethane	3.524	101	196138	51.819	ug/l	97
8) Diethyl Ether	3.983	74	87405	53.001	ug/l	100
9) 1,1,2-Trichlorotrifluo...	4.395	101	127498	53.615	ug/l	99
10) Methyl Iodide	4.612	142	164080	53.227	ug/l	97
11) Tert butyl alcohol	5.542	59	207768	262.193	ug/l	99
12) 1,1-Dichloroethene	4.359	96	125461	51.678	ug/l	99
13) Acrolein	4.195	56	60336	240.545	ug/l	98
14) Allyl chloride	5.042	41	203344	50.504	ug/l	98
15) Acrylonitrile	5.736	53	479502	258.885	ug/l	100
16) Acetone	4.448	43	382146	246.750	ug/l	98
17) Carbon Disulfide	4.736	76	339466	50.550	ug/l	100
18) Methyl Acetate	5.042	43	234290	51.912	ug/l	100
19) Methyl tert-butyl Ether	5.818	73	454641	51.720	ug/l	100
20) Methylene Chloride	5.295	84	143862	49.617	ug/l	97
21) trans-1,2-Dichloroethene	5.806	96	133535	49.437	ug/l	99
22) Diisopropyl ether	6.689	45	440080	51.853	ug/l	99
23) Vinyl Acetate	6.618	43	1828919	255.055	ug/l	99
24) 1,1-Dichloroethane	6.589	63	251473	51.487	ug/l	98
25) 2-Butanone	7.494	43	644194	255.897	ug/l	99
26) 2,2-Dichloropropane	7.500	77	216678	57.031	ug/l	99
27) cis-1,2-Dichloroethene	7.500	96	164985	51.073	ug/l	99
28) Bromochloromethane	7.824	49	108810	45.310	ug/l	99
29) Tetrahydrofuran	7.847	42	420891	256.656	ug/l	100
30) Chloroform	7.977	83	248429	50.933	ug/l	97
31) Cyclohexane	8.265	56	237836	50.212	ug/l	97
32) 1,1,1-Trichloroethane	8.177	97	209927	50.603	ug/l	97
36) 1,1-Dichloropropene	8.377	75	181414	52.123	ug/l	99
37) Ethyl Acetate	7.571	43	222090	50.125	ug/l	99
38) Carbon Tetrachloride	8.371	117	179281	52.466	ug/l	98
39) Methylcyclohexane	9.606	83	251231	52.686	ug/l	98
40) Benzene	8.612	78	579667	50.931	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060625\
 Data File : VN086869.D
 Acq On : 06 Jun 2025 15:54
 Operator : JC\MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ICVVN060625

Quant Time: Jun 07 02:16:19 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.789	41	130903	52.607	ug/l	100
42) 1,2-Dichloroethane	8.677	62	177746	51.489	ug/l	100
43) Isopropyl Acetate	8.694	43	359378	50.536	ug/l	100
44) Trichloroethene	9.359	130	139763	51.788	ug/l	100
45) 1,2-Dichloropropane	9.624	63	142463	51.450	ug/l	99
46) Dibromomethane	9.712	93	97199	52.849	ug/l	98
47) Bromodichloromethane	9.894	83	196104	51.824	ug/l	100
48) Methyl methacrylate	9.688	41	168001	51.328	ug/l	98
49) 1,4-Dioxane	9.706	88	66102	1110.123	ug/l #	99
51) 4-Methyl-2-Pentanone	10.447	43	1146622	267.835	ug/l	99
52) Toluene	10.635	92	361963	52.040	ug/l	99
53) t-1,3-Dichloropropene	10.835	75	224839	53.137	ug/l	99
54) cis-1,3-Dichloropropene	10.318	75	237774	52.519	ug/l	99
55) 1,1,2-Trichloroethane	11.018	97	137998	51.562	ug/l	99
56) Ethyl methacrylate	10.882	69	241649	56.688	ug/l	98
57) 1,3-Dichloropropane	11.165	76	238572	51.386	ug/l	99
58) 2-Chloroethyl Vinyl ether	10.165	63	625479	246.006	ug/l	100
59) 2-Hexanone	11.200	43	781717	283.478	ug/l	100
60) Dibromochloromethane	11.359	129	146749	52.627	ug/l	99
61) 1,2-Dibromoethane	11.471	107	142338	51.887	ug/l	99
64) Tetrachloroethene	11.106	164	111409	50.450	ug/l	98
65) Chlorobenzene	11.894	112	391661	50.892	ug/l	98
66) 1,1,1,2-Tetrachloroethane	11.959	131	129045	52.164	ug/l	99
67) Ethyl Benzene	11.965	91	683879	51.590	ug/l	98
68) m/p-Xylenes	12.071	106	527293	103.912	ug/l	98
69) o-Xylene	12.400	106	254600	52.387	ug/l	100
70) Styrene	12.412	104	438751	52.758	ug/l	99
71) Bromoform	12.582	173	100357	54.751	ug/l #	99
73) Isopropylbenzene	12.694	105	652045	51.823	ug/l	100
74) N-amyl acetate	12.506	43	238355	54.212	ug/l	97
75) 1,1,2,2-Tetrachloroethane	12.941	83	220643	51.763	ug/l	100
76) 1,2,3-Trichloropropane	12.994	75	302004	73.563	ug/l	69
77) Bromobenzene	12.982	156	150950	52.313	ug/l	100
78) n-propylbenzene	13.035	91	801180	52.397	ug/l	100
79) 2-Chlorotoluene	13.123	91	473629	51.650	ug/l	99
80) 1,3,5-Trimethylbenzene	13.171	105	548347	52.787	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.735	75	99341	55.702	ug/l	97
82) 4-Chlorotoluene	13.223	91	481403	51.884	ug/l	99
83) tert-Butylbenzene	13.441	119	499519	52.532	ug/l	99
84) 1,2,4-Trimethylbenzene	13.482	105	552333	53.024	ug/l	100
85) sec-Butylbenzene	13.618	105	723986	52.428	ug/l	100
86) p-Isopropyltoluene	13.729	119	606979	53.174	ug/l	99
87) 1,3-Dichlorobenzene	13.735	146	290810	51.224	ug/l	100
88) 1,4-Dichlorobenzene	13.812	146	298412	51.557	ug/l	100
89) n-Butylbenzene	14.059	91	579636	52.424	ug/l	100
90) Hexachloroethane	14.335	117	102254	52.847	ug/l	100
91) 1,2-Dichlorobenzene	14.106	146	280025	51.340	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	14.723	75	50115	49.161	ug/l	98
93) 1,2,4-Trichlorobenzene	15.394	180	181064	51.986	ug/l	99
94) Hexachlorobutadiene	15.500	225	68570	52.843	ug/l	99
95) Naphthalene	15.641	128	674187	52.005	ug/l	100
96) 1,2,3-Trichlorobenzene	15.841	180	174888	50.539	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060625\
Data File : VN086869.D
Acq On : 06 Jun 2025 15:54
Operator : JC\MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1


Instrument :
MSVOA_N
ClientSampleId :
ICVVN060625


Quant Time: Jun 07 02:16:19 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 02:12:50 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jun 07 02:16:19 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 02:12:50 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.230 min Scan# 1068
Delta R.T. -0.006 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54




Instrument :
MSVOA_N
ClientSampleId :
ICVVN060625


Tgt Ion:168 Resp: 218122
Ion Ratio Lower Upper
168 100
99 58.6 49.1 73.7



#2
Dichlorodifluoromethane
Concen: 55.286 ug/l
RT: 2.154 min Scan# 34
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54

Tgt Ion: 85 Resp: 120237
Ion Ratio Lower Upper
85 100
87 33.2 16.2 48.6









#3
Chloromethane
Concen: 49.505 ug/l
RT: 2.401 min Scan# 75
Delta R.T. 0.006 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54




Instrument :
MSVOA_N
ClientSampleId :
ICVVN060625


Tgt Ion: 50 Resp: 139027
Ion Ratio Lower Upper
50 100
52 32.9 25.7 38.5



#4
Vinyl Chloride
Concen: 52.451 ug/l
RT: 2.554 min Scan# 102
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54

Tgt Ion: 62 Resp: 151951
Ion Ratio Lower Upper
62 100
64 32.0 26.0 39.0









#5
Bromomethane
Concen: 55.292 ug/l
RT: 2.989 min Scan# 177
Delta R.T. -0.006 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54




Instrument :
MSVOA_N
ClientSampleId :
ICVVN060625


Tgt Ion: 94 Resp: 89638
Ion Ratio Lower Upper
94 100
96 92.5 77.9 116.9



#6
Chloroethane
Concen: 51.117 ug/l
RT: 3.154 min Scan# 204
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54

Tgt Ion: 64 Resp: 95654
Ion Ratio Lower Upper
64 100
66 32.6 24.6 36.8









#7
Trichlorofluoromethane
Concen: 51.819 ug/l
RT: 3.524 min Scan# 267
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54




Instrument :
MSVOA_N
ClientSampleId :
ICVVN060625


Tgt Ion:101 Resp: 196138
Ion Ratio Lower Upper
101 100
103 62.9 52.2 78.2



#8
Diethyl Ether
Concen: 53.001 ug/l
RT: 3.983 min Scan# 345
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54

Tgt Ion: 74 Resp: 87405
Ion Ratio Lower Upper
74 100
45 90.9 45.5 136.5









#9
1,1,2-Trichlorotrifluoroethane
Concen: 53.615 ug/l
RT: 4.395 min Scan# 415
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54




Instrument :
MSVOA_N
ClientSampleId :
ICVVN060625


Tgt Ion:101 Resp: 127498
Ion Ratio Lower Upper
101 100
85 41.9 33.8 50.8
151 72.8 58.7 88.1



#10
Methyl Iodide
Concen: 53.227 ug/l
RT: 4.612 min Scan# 452
Delta R.T. 0.006 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54

Tgt Ion:142 Resp: 164080
Ion Ratio Lower Upper
142 100
127 36.6 30.9 46.3
141 14.0 10.4 15.6









#11
Tert butyl alcohol
Concen: 262.193 ug/l
RT: 5.542 min Scan# 611
Delta R.T. -0.006 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54




Instrument :
MSVOA_N
ClientSampleId :
ICVVN060625


Tgt Ion: 59 Resp: 207768
Ion Ratio Lower Upper
59 100
57 9.8 8.1 12.1



#12
1,1-Dichloroethene
Concen: 51.678 ug/l
RT: 4.359 min Scan# 409
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54

Tgt Ion: 96 Resp: 125461
Ion Ratio Lower Upper
96 100
61 158.0 125.8 188.6
98 65.2 52.5 78.7





#13

Acrolein

Concen: 240.545 ug/l

RT: 4.195 min Scan# 382

Delta R.T. -0.006 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54

Instrument :

MSVOA_N

ClientSampleId :





ICVVN060625

Tgt Ion: 56 Resp: 60336

Ion Ratio Lower Upper

56 100

55 71.1 55.6 83.4



#14

Allyl chloride

Concen: 50.504 ug/l

RT: 5.042 min Scan# 525

Delta R.T. -0.000 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54




Tgt Ion: 41 Resp: 203344


Ion Ratio Lower Upper

41 100

39 62.3 51.3 76.9

76 40.4 31.4 47.0









#15
Acrylonitrile
Concen: 258.885 ug/l
RT: 5.736 min Scan# 644
Delta R.T. -0.006 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54




Instrument :
MSVOA_N
ClientSampleId :
ICVVN060625


Tgt Ion: 53 Resp: 479502
Ion Ratio Lower Upper
53 100
52 81.3 65.0 97.6
51 34.5 27.2 40.8



#16
Acetone
Concen: 246.750 ug/l
RT: 4.448 min Scan# 424
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54

Tgt Ion: 43 Resp: 382146
Ion Ratio Lower Upper
43 100
58 35.9 28.0 42.0









#17
Carbon Disulfide
Concen: 50.550 ug/l
RT: 4.736 min Scan# 473
Delta R.T. 0.006 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54




Instrument :
MSVOA_N
ClientSampleId :
ICVVN060625


Tgt Ion: 76 Resp: 339466
Ion Ratio Lower Upper
76 100
78 9.1 7.3 10.9



#18
Methyl Acetate
Concen: 51.912 ug/l
RT: 5.042 min Scan# 525
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54

Tgt Ion: 43 Resp: 234290
Ion Ratio Lower Upper
43 100
74 25.9 20.7 31.1









#19
Methyl tert-butyl Ether
Concen: 51.720 ug/l
RT: 5.818 min Scan# 61
Delta R.T. 0.006 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54




Instrument :
MSVOA_N
ClientSampleId :
ICVNVN060625


Tgt Ion: 73 Resp: 454641
Ion Ratio Lower Upper
73 100
57 21.8 17.4 26.2



#20
Methylene Chloride
Concen: 49.617 ug/l
RT: 5.295 min Scan# 568
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54

Tgt Ion: 84 Resp: 143862
Ion Ratio Lower Upper
84 100
49 110.3 90.5 135.7
51 31.8 28.5 42.7
86 60.8 48.1 72.1





#21

trans-1,2-Dichloroethene

Concen: 49.437 ug/l

RT: 5.806 min Scan# 61

Delta R.T. -0.000 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54

Instrument :

MSVOA_N

ClientSampleId :

ICVVN060625





Tgt Ion: 96 Resp: 133535

Ion Ratio Lower Upper

96 100

61 136.8 109.2 163.8

98 65.2 49.8 74.8



#22

Diisopropyl ether

Concen: 51.853 ug/l

RT: 6.689 min Scan# 805

Delta R.T. -0.000 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54

Tgt Ion: 45 Resp: 440080




Ion Ratio Lower Upper


45 100

43 52.4 41.9 62.9

87 29.5 22.4 33.6

59 12.4 9.8 14.6









#23
 Vinyl Acetate
 Concen: 255.055 ug/l
 RT: 6.618 min Scan# 793
 Delta R.T. -0.000 min
 Lab File: VN086869.D
 Acq: 06 Jun 2025 15:54




Instrument :
 MSVOA_N
 ClientSampleId :
 ICVVN060625


Tgt Ion: 43 Resp: 1828919
 Ion Ratio Lower Upper
 43 100
 86 12.1 9.5 14.3



#24
 1,1-Dichloroethane
 Concen: 51.487 ug/l
 RT: 6.589 min Scan# 788
 Delta R.T. 0.006 min
 Lab File: VN086869.D
 Acq: 06 Jun 2025 15:54

Tgt Ion: 63 Resp: 251473
 Ion Ratio Lower Upper
 63 100
 98 7.7 3.6 10.8
 100 4.6 2.6 7.8









#25
2-Butanone
Concen: 255.897 ug/l
RT: 7.494 min Scan# 942
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54




Instrument :
MSVOA_N
ClientSampleId :
ICVVN060625


Tgt Ion: 43 Resp: 644194
Ion Ratio Lower Upper
43 100
72 29.5 23.0 34.6



#26
2,2-Dichloropropane
Concen: 57.031 ug/l
RT: 7.500 min Scan# 943
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54

Tgt Ion: 77 Resp: 216678
Ion Ratio Lower Upper
77 100
97 23.7 11.6 34.7









#27
 cis-1,2-Dichloroethene
 Concen: 51.073 ug/l
 RT: 7.500 min Scan# 943
 Delta R.T. -0.000 min
 Lab File: VN086869.D
 Acq: 06 Jun 2025 15:54




Instrument :
 MSVOA_N
 ClientSampleId :
 ICVVN060625


Tgt Ion: 96 Resp: 164985
 Ion Ratio Lower Upper
 96 100
 61 140.3 0.0 278.0
 98 64.5 0.0 128.2



#28
 Bromochloromethane
 Concen: 45.310 ug/l
 RT: 7.824 min Scan# 998
 Delta R.T. -0.000 min
 Lab File: VN086869.D
 Acq: 06 Jun 2025 15:54

Tgt Ion: 49 Resp: 108810
 Ion Ratio Lower Upper
 49 100
 129 1.9 0.0 3.8
 130 78.5 62.2 93.2





#29

Tetrahydrofuran

Concen: 256.656 ug/l

RT: 7.847 min Scan# 1002

Delta R.T. -0.006 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54

Instrument :

MSVOA_N

ClientSampleId :

ICVVN060625





Tgt Ion: 42 Resp: 420891

Ion Ratio Lower Upper

42 100

72 47.5 37.8 56.6

71 44.5 35.6 53.4



#30

Chloroform

Concen: 50.933 ug/l

RT: 7.977 min Scan# 1024

Delta R.T. -0.000 min

Lab File: VN086869.D




Acq: 06 Jun 2025 15:54


Tgt Ion: 83 Resp: 248429

Ion Ratio Lower Upper

83 100

85 63.5 52.9 79.3









#31
Cyclohexane
Concen: 50.212 ug/l
RT: 8.265 min Scan# 1073
Delta R.T. -0.006 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54




Instrument :
MSVOA_N
ClientSampleId :
ICVVN060625


Tgt Ion: 56 Resp: 237836
Ion Ratio Lower Upper
56 100
69 33.9 25.4 38.0
84 88.4 68.8 103.2



#32
1,1,1-Trichloroethane
Concen: 50.603 ug/l
RT: 8.177 min Scan# 1058
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54

Tgt Ion: 97 Resp: 209927
Ion Ratio Lower Upper
97 100
99 61.7 52.0 78.0
61 45.4 35.9 53.9





#33

1,2-Dichloroethane-d4

Concen: 52.135 ug/l

RT: 8.588 min Scan# 1128

Delta R.T. -0.000 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54

Instrument :

MSVOA_N

ClientSampleId :





ICVVN060625

Tgt Ion: 65 Resp: 152255

Ion Ratio Lower Upper

65 100

67 54.6 0.0 105.6



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.106 min Scan# 1216

Delta R.T. -0.000 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54




Tgt Ion: 114 Resp: 394140


Ion Ratio Lower Upper

114 100

63 20.2 0.0 39.6

88 16.4 0.0 30.2









#35
Dibromofluoromethane
Concen: 54.416 ug/l
RT: 8.177 min Scan# 1057
Delta R.T. 0.006 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54




Instrument :
MSVOA_N
ClientSampleId :
ICVVN060625


Tgt Ion:113 Resp: 127102
Ion Ratio Lower Upper
113 100
111 103.4 84.2 126.2
192 18.0 14.2 21.4



#36
1,1-Dichloropropene
Concen: 52.123 ug/l
RT: 8.377 min Scan# 1092
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54

Tgt Ion: 75 Resp: 181414
Ion Ratio Lower Upper
75 100
110 35.0 17.6 52.9
77 31.0 25.3 37.9









#37
Ethyl Acetate
Concen: 50.125 ug/l
RT: 7.571 min Scan# 91
Delta R.T. -0.006 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54




Instrument :
MSVOA_N
ClientSampleId :
ICVVN060625


Tgt Ion: 43 Resp: 222090
Ion Ratio Lower Upper
43 100
61 15.6 12.1 18.1
70 12.9 9.8 14.6



#38
Carbon Tetrachloride
Concen: 52.466 ug/l
RT: 8.371 min Scan# 1091
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54

Tgt Ion: 117 Resp: 179281
Ion Ratio Lower Upper
117 100
119 94.8 75.0 112.6
121 28.5 24.9 37.3





#39

Methylcyclohexane

Concen: 52.686 ug/l

RT: 9.606 min Scan# 1130

Delta R.T. -0.000 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54

Instrument :

MSVOA_N

ClientSampleId :

ICVVN060625





Tgt Ion: 83 Resp: 251231

Ion Ratio Lower Upper

83 100

55 74.5 61.0 91.6

98 47.2 38.9 58.3



#40

Benzene

Concen: 50.931 ug/l

RT: 8.612 min Scan# 1132

Delta R.T. -0.000 min

Lab File: VN086869.D




Acq: 06 Jun 2025 15:54


Tgt Ion: 78 Resp: 579667

Ion Ratio Lower Upper

78 100

77 23.7 18.7 28.1









#41
Methacrylonitrile
Concen: 52.607 ug/l
RT: 7.789 min Scan# 992
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54




Instrument :
MSVOA_N
ClientSampleId :
ICVVN060625


Tgt Ion:	41	Resp:	130903
Ion	Ratio	Lower	Upper
41	100		
39	51.3	41.2	61.8
67	76.9	61.4	92.2
52	28.6	22.7	34.1



#42
1,2-Dichloroethane
Concen: 51.489 ug/l
RT: 8.677 min Scan# 1143
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54

Tgt Ion:	62	Resp:	177746
Ion	Ratio	Lower	Upper
62	100		
98	10.6	0.0	21.2









#43
Isopropyl Acetate
Concen: 50.536 ug/l
RT: 8.694 min Scan# 1146
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54




Instrument :
MSVOA_N
ClientSampleId :
ICVVN060625


Tgt Ion: 43 Resp: 359378
Ion Ratio Lower Upper
43 100
61 27.9 22.1 33.1
87 14.8 11.8 17.6



#44
Trichloroethene
Concen: 51.788 ug/l
RT: 9.359 min Scan# 1259
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54

Tgt Ion:130 Resp: 139763
Ion Ratio Lower Upper
130 100
95 98.0 0.0 196.2





#45

1,2-Dichloropropane

Concen: 51.450 ug/l

RT: 9.624 min Scan# 1304

Delta R.T. -0.006 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54

Instrument :

MSVOA_N

ClientSampleId :





ICVVN060625

Tgt Ion: 63 Resp: 142463

Ion Ratio Lower Upper

63 100

65 32.7 25.5 38.3



#46

Dibromomethane

Concen: 52.849 ug/l

RT: 9.712 min Scan# 1319

Delta R.T. -0.006 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54




Tgt Ion: 93 Resp: 97199


Ion Ratio Lower Upper

93 100

95 82.2 67.1 100.7

174 90.9 74.2 111.4









#47
Bromodichloromethane
Concen: 51.824 ug/l
RT: 9.894 min Scan# 1350
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54




Instrument :
MSVOA_N
ClientSampleId :
ICVVN060625


Tgt Ion: 83 Resp: 196104
Ion Ratio Lower Upper
83 100
85 63.8 50.9 76.3
127 8.2 6.6 10.0



#48
Methyl methacrylate
Concen: 51.328 ug/l
RT: 9.688 min Scan# 1315
Delta R.T. 0.006 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54

Tgt Ion: 41 Resp: 168001
Ion Ratio Lower Upper
41 100
69 95.2 74.6 112.0
39 51.0 40.2 60.4





#49

1,4-Dioxane

Concen: 1110.123 ug/l

RT: 9.706 min Scan# 1

Delta R.T. 0.006 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54

Instrument :

MSVOA_N

ClientSampleId :

ICVVN060625





Tgt Ion: 88 Resp: 66102

Ion Ratio Lower Upper

88 100

43 0.0 0.0 0.0

58 72.7 57.7 86.5



#50

Toluene-d8

Concen: 53.129 ug/l

RT: 10.571 min Scan# 1465

Delta R.T. -0.000 min

Lab File: VN086869.D




Acq: 06 Jun 2025 15:54


Tgt Ion: 98 Resp: 491265

Ion Ratio Lower Upper

98 100

100 67.6 53.4 80.0









#51
4-Methyl-2-Pentanone
Concen: 267.835 ug/l
RT: 10.447 min Scan# 1444
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54




Instrument :
MSVOA_N
ClientSampleId :
ICVVN060625


Tgt Ion: 43 Resp: 1146622
Ion Ratio Lower Upper
43 100
58 43.2 34.2 51.4



#52
Toluene
Concen: 52.040 ug/l
RT: 10.635 min Scan# 1476
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54

Tgt Ion: 92 Resp: 361963
Ion Ratio Lower Upper
92 100
91 169.8 136.5 204.7









#53
t-1,3-Dichloropropene
Concen: 53.137 ug/l
RT: 10.835 min Scan# 11
Delta R.T. -0.006 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54




Instrument :
MSVOA_N
ClientSampleId :
ICVVN060625


Tgt Ion: 75 Resp: 224839
Ion Ratio Lower Upper
75 100
77 32.3 26.4 39.6



#54
cis-1,3-Dichloropropene
Concen: 52.519 ug/l
RT: 10.318 min Scan# 1422
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54

Tgt Ion: 75 Resp: 237774
Ion Ratio Lower Upper
75 100
77 32.1 26.2 39.4
39 41.3 32.3 48.5









#55
1,1,2-Trichloroethane
Concen: 51.562 ug/l
RT: 11.018 min Scan# 1541
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54




Instrument :
MSVOA_N
ClientSampleId :
ICVVN060625


Tgt Ion:	97	Resp:	137998
Ion	Ratio	Lower	Upper
97	100		
83	87.4	69.7	104.5
85	54.1	45.1	67.7
99	62.0	50.6	75.8



#56
Ethyl methacrylate
Concen: 56.688 ug/l
RT: 10.882 min Scan# 1518
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54

Tgt Ion:	69	Resp:	241649
Ion	Ratio	Lower	Upper
69	100		
41	57.7	47.2	70.8
39	27.4	23.0	34.6





#57

1,3-Dichloropropane

Concen: 51.386 ug/l

RT: 11.165 min Scan# 11

Delta R.T. -0.000 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54

Instrument :

MSVOA_N

ClientSampleId :





ICVVN060625

Tgt Ion: 76 Resp: 238572

Ion Ratio Lower Upper

76 100

78 32.0 25.3 37.9



#58

2-Chloroethyl Vinyl ether

Concen: 246.006 ug/l

RT: 10.165 min Scan# 1396

Delta R.T. -0.000 min

Lab File: VN086869.D




Acq: 06 Jun 2025 15:54


Tgt Ion: 63 Resp: 625479

Ion Ratio Lower Upper

63 100

106 28.9 23.0 34.6









#59
2-Hexanone
Concen: 283.478 ug/l
RT: 11.200 min Scan# 1572
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54




Instrument :
MSVOA_N
ClientSampleId :
ICVVN060625


Tgt Ion: 43 Resp: 781717
Ion Ratio Lower Upper
43 100
58 61.2 30.4 91.3



#60
Dibromochloromethane
Concen: 52.627 ug/l
RT: 11.359 min Scan# 1599
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54

Tgt Ion: 129 Resp: 146749
Ion Ratio Lower Upper
129 100
127 78.1 38.6 115.8









#61
1,2-Dibromoethane
Concen: 51.887 ug/l
RT: 11.471 min Scan# 1618
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54




Instrument :
MSVOA_N
ClientSampleId :
ICVVN060625


Tgt Ion:107 Resp: 142338
Ion Ratio Lower Upper
107 100
109 94.5 76.3 114.5



#62
4-Bromofluorobenzene
Concen: 54.081 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54

Tgt Ion: 95 Resp: 185792
Ion Ratio Lower Upper
95 100
174 71.4 0.0 141.8
176 68.1 0.0 132.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 117
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54




Instrument :
MSVOA_N
ClientSampleId :
ICVVN060625


Tgt Ion:117 Resp: 348935
Ion Ratio Lower Upper
117 100
82 56.5 44.6 67.0
119 32.5 25.5 38.3



#64
Tetrachloroethene
Concen: 50.450 ug/l
RT: 11.106 min Scan# 1556
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54

Tgt Ion:164 Resp: 111409
Ion Ratio Lower Upper
164 100
166 128.5 101.8 152.8
129 99.9 83.0 124.4
131 97.9 78.2 117.2









#65
Chlorobenzene
Concen: 50.892 ug/l
RT: 11.894 min Scan# 111
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54




Instrument :
MSVOA_N
ClientSampleId :
ICVVN060625


Tgt Ion:112 Resp: 391661
Ion Ratio Lower Upper
112 100
114 31.3 26.2 39.2



#66
1,1,1,2-Tetrachloroethane
Concen: 52.164 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. -0.006 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54

Tgt Ion:131 Resp: 129045
Ion Ratio Lower Upper
131 100
133 94.5 48.3 144.8
119 67.4 34.0 101.8









#67
Ethyl Benzene
Concen: 51.590 ug/l
RT: 11.965 min Scan# 1702
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54




Instrument :
MSVOA_N
ClientSampleId :
ICVVN060625


Tgt Ion: 91 Resp: 683879
Ion Ratio Lower Upper
91 100
106 30.9 25.4 38.2



#68
m/p-Xylenes
Concen: 103.912 ug/l
RT: 12.071 min Scan# 1720
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54

Tgt Ion:106 Resp: 527293
Ion Ratio Lower Upper
106 100
91 202.0 159.4 239.0









#69
o-Xylene
Concen: 52.387 ug/l
RT: 12.400 min Scan# 1776
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54




Instrument :
MSVOA_N
ClientSampleId :
ICVVN060625


Tgt Ion:106 Resp: 254600
Ion Ratio Lower Upper
106 100
91 212.8 106.1 318.1



#70
Styrene
Concen: 52.758 ug/l
RT: 12.412 min Scan# 1778
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54

Tgt Ion:104 Resp: 438751
Ion Ratio Lower Upper
104 100
78 48.7 38.5 57.7
103 53.7 43.1 64.7





#71

Bromoform

Concen: 54.751 ug/l

RT: 12.582 min Scan# 1807

Delta R.T. -0.000 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54

Instrument :

MSVOA_N

ClientSampleId :

ICVVN060625





Tgt Ion:173 Resp: 100357

Ion Ratio Lower Upper

173 100

175 48.7 24.1 72.4

254 0.0 0.0 0.0



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.794 min Scan# 2013

Delta R.T. -0.000 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54




Tgt Ion:152 Resp: 172710


Ion Ratio Lower Upper

152 100

115 60.6 30.1 90.5

150 174.0 0.0 345.0









#73
Isopropylbenzene
Concen: 51.823 ug/l
RT: 12.694 min Scan# 1826
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54




Instrument :
MSVOA_N
ClientSampleId :
ICVVN060625


Tgt Ion:105 Resp: 652045
Ion Ratio Lower Upper
105 100
120 27.1 13.5 40.4



#74
N-ethyl acetate
Concen: 54.212 ug/l
RT: 12.506 min Scan# 1794
Delta R.T. -0.006 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54

Tgt Ion: 43 Resp: 238355
Ion Ratio Lower Upper
43 100
70 49.3 37.6 56.4
55 25.3 19.0 28.6
61 25.2 20.8 31.2









#75
1,1,2-Tetrachloroethane
Concen: 51.763 ug/l
RT: 12.941 min Scan# 1868
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54




Instrument :
MSVOA_N
ClientSampleId :
ICVVN060625


Tgt Ion: 83 Resp: 220643
Ion Ratio Lower Upper
83 100
131 9.8 4.9 14.7
85 63.9 32.0 96.0



#76
1,2,3-Trichloropropane
Concen: 73.563 ug/l
RT: 12.994 min Scan# 1877
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54

Tgt Ion: 75 Resp: 302004
Ion Ratio Lower Upper
75 100
77 117.4 79.3 237.9





#77

Bromobenzene

Concen: 52.313 ug/l

RT: 12.982 min Scan# 1875

Delta R.T. -0.000 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54

Instrument :

MSVOA_N

ClientSampleId :

ICVVN060625





Tgt Ion:156 Resp: 150950

Ion Ratio Lower Upper

156 100

77 234.9 117.7 353.1

158 96.5 48.1 144.4



#78

n-propylbenzene

Concen: 52.397 ug/l

RT: 13.035 min Scan# 1884

Delta R.T. -0.000 min

Lab File: VN086869.D




Acq: 06 Jun 2025 15:54


Tgt Ion: 91 Resp: 801180

Ion Ratio Lower Upper

91 100

120 22.9 11.4 34.2









#79
2-Chlorotoluene
Concen: 51.650 ug/l
RT: 13.123 min Scan# 1899
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54




Instrument :
MSVOA_N
ClientSampleId :
ICVVN060625


Tgt Ion: 91 Resp: 473629
Ion Ratio Lower Upper
91 100
126 32.8 16.2 48.5



#80
1,3,5-Trimethylbenzene
Concen: 52.787 ug/l
RT: 13.171 min Scan# 1907
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54

Tgt Ion: 105 Resp: 548347
Ion Ratio Lower Upper
105 100
120 50.6 24.9 74.6









#81
trans-1,4-Dichloro-2-butene
Concen: 55.702 ug/l
RT: 12.735 min Scan# 1833
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54




Instrument :
MSVOA_N
ClientSampleId :
ICVVN060625


Tgt Ion: 75 Resp: 99341
Ion Ratio Lower Upper
75 100
53 83.7 66.6 99.8
89 40.4 36.7 55.1



#82
4-Chlorotoluene
Concen: 51.884 ug/l
RT: 13.223 min Scan# 1916
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54

Tgt Ion: 91 Resp: 481403
Ion Ratio Lower Upper
91 100
126 33.1 16.7 50.1





#83

tert-Butylbenzene

Concen: 52.532 ug/l

RT: 13.441 min Scan# 1953

Delta R.T. 0.006 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54

Instrument :

MSVOA_N

ClientSampleId :

ICVVN060625





Tgt Ion:119 Resp: 499519

Ion Ratio Lower Upper

119 100

91 63.9 32.1 96.5

134 26.4 13.4 40.1



#84

1,2,4-Trimethylbenzene

Concen: 53.024 ug/l

RT: 13.482 min Scan# 1960

Delta R.T. -0.000 min

Lab File: VN086869.D




Acq: 06 Jun 2025 15:54


Tgt Ion:105 Resp: 552333

Ion Ratio Lower Upper

105 100

120 46.6 23.2 69.6









#85
 sec-Butylbenzene
 Concen: 52.428 ug/l
 RT: 13.618 min Scan# 1983
 Delta R.T. -0.000 min
 Lab File: VN086869.D
 Acq: 06 Jun 2025 15:54




Instrument :
 MSVOA_N
 ClientSampleId :
 ICVVN060625


Tgt Ion:105 Resp: 723986
 Ion Ratio Lower Upper
 105 100
 134 20.2 10.0 30.0



#86
 p-Isopropyltoluene
 Concen: 53.174 ug/l
 RT: 13.729 min Scan# 2002
 Delta R.T. -0.000 min
 Lab File: VN086869.D
 Acq: 06 Jun 2025 15:54

Tgt Ion:119 Resp: 606979
 Ion Ratio Lower Upper
 119 100
 134 26.5 13.5 40.4
 91 22.7 11.5 34.4





#87

1,3-Dichlorobenzene

Concen: 51.224 ug/l

RT: 13.735 min Scan# 2003

Delta R.T. -0.000 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54

Instrument :

MSVOA_N

ClientSampleId :

ICVVN060625





Tgt Ion:146 Resp: 290810

Ion Ratio Lower Upper

146 100

111 42.1 20.9 62.8

148 64.3 32.0 96.2



#88

1,4-Dichlorobenzene

Concen: 51.557 ug/l

RT: 13.812 min Scan# 2016

Delta R.T. -0.000 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54




Tgt Ion:146 Resp: 298412


Ion Ratio Lower Upper

146 100

111 40.0 19.9 59.7

148 63.3 31.8 95.4





#89

n-Butylbenzene

Concen: 52.424 ug/l

RT: 14.059 min Scan# 2057

Delta R.T. 0.006 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54

Instrument :

MSVOA_N

ClientSampleId :

ICVVN060625





Tgt Ion: 91 Resp: 579636

Ion Ratio Lower Upper

91 100

92 55.0 27.5 82.4

134 25.5 13.0 39.0



#90

Hexachloroethane

Concen: 52.847 ug/l

RT: 14.335 min Scan# 2105

Delta R.T. -0.000 min

Lab File: VN086869.D




Acq: 06 Jun 2025 15:54


Tgt Ion: 117 Resp: 102254

Ion Ratio Lower Upper

117 100

201 66.4 33.2 99.6









#91
1,2-Dichlorobenzene
Concen: 51.340 ug/l
RT: 14.106 min Scan# 2066
Delta R.T. -0.000 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54




Instrument :
MSVOA_N
ClientSampleId :
ICVVN060625


Tgt Ion:146 Resp: 280025
Ion Ratio Lower Upper
146 100
111 43.5 21.6 64.6
148 63.4 31.7 95.1



#92
1,2-Dibromo-3-Chloropropane
Concen: 49.161 ug/l
RT: 14.723 min Scan# 2171
Delta R.T. 0.006 min
Lab File: VN086869.D
Acq: 06 Jun 2025 15:54

Tgt Ion: 75 Resp: 50115
Ion Ratio Lower Upper
75 100
155 83.5 41.1 123.4
157 108.6 52.7 158.1





#93

1,2,4-Trichlorobenzene

Concen: 51.986 ug/l

RT: 15.394 min Scan# 2284

Delta R.T. 0.006 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54

Instrument :

MSVOA_N

ClientSampleId :

ICVVN060625





Tgt Ion:180 Resp: 181064

Ion Ratio Lower Upper

180 100

182 94.6 46.9 140.7

145 30.8 16.0 47.9



#94

Hexachlorobutadiene

Concen: 52.843 ug/l

RT: 15.500 min Scan# 2303

Delta R.T. -0.000 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54




Tgt Ion:225 Resp: 68570


Ion Ratio Lower Upper

225 100

223 65.3 33.5 100.4

227 62.5 31.1 93.5





#95

Naphthalene

Concen: 52.005 ug/l

RT: 15.641 min Scan# 2327

Delta R.T. -0.000 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54

Instrument :

MSVOA_N

ClientSampleId :

ICVVN060625





Tgt Ion:128 Resp: 674187

Ion Ratio Lower Upper

128 100

127 12.8 10.2 15.2

129 10.9 8.8 13.2



#96

1,2,3-Trichlorobenzene

Concen: 50.539 ug/l

RT: 15.841 min Scan# 2361

Delta R.T. 0.006 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54




Tgt Ion:180 Resp: 174888

Ion Ratio Lower Upper

180 100

182 95.2 47.0 141.0

145 33.8 16.6 49.8



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061125\
 Data File : VN086940.D
 Acq On : 11 Jun 2025 11:32
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050

Quant Time: Jun 12 01:28:53 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	205098	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	357643	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	312173	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	152136	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	127629	46.478	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	92.960%
35) Dibromofluoromethane	8.177	113	110532	52.151	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.300%
50) Toluene-d8	10.565	98	422391	50.342	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.680%
62) 4-Bromofluorobenzene	12.847	95	157388	50.489	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	100.980%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	2.154	85	110255	53.916	ug/l	100
3) Chloromethane	2.395	50	113731	43.069	ug/l	94
4) Vinyl Chloride	2.554	62	143755	52.773	ug/l	99
5) Bromomethane	2.989	94	63914	41.928	ug/l	96
6) Chloroethane	3.154	64	93702	53.253	ug/l	99
7) Trichlorofluoromethane	3.524	101	190657	53.570	ug/l	99
8) Diethyl Ether	3.983	74	85722	55.281	ug/l	95
9) 1,1,2-Trichlorotrifluo...	4.401	101	119322	53.364	ug/l	99
10) Methyl Iodide	4.618	142	115825	39.959	ug/l	97
11) Tert butyl alcohol	5.542	59	183049	245.667	ug/l	99
12) 1,1-Dichloroethene	4.365	96	121690	53.307	ug/l	94
13) Acrolein	4.201	56	68006	288.340	ug/l	99
14) Allyl chloride	5.048	41	187099	49.420	ug/l	96
15) Acrylonitrile	5.736	53	456903	262.349	ug/l	99
16) Acetone	4.442	43	393802	270.423	ug/l	97
17) Carbon Disulfide	4.736	76	308706	48.889	ug/l	100
18) Methyl Acetate	5.042	43	222452	52.419	ug/l	98
19) Methyl tert-butyl Ether	5.812	73	445459	53.894	ug/l	98
20) Methylene Chloride	5.295	84	138117	50.660	ug/l	95
21) trans-1,2-Dichloroethene	5.806	96	129921	51.154	ug/l	98
22) Diisopropyl ether	6.683	45	412097	51.639	ug/l	99
23) Vinyl Acetate	6.618	43	1766271	261.960	ug/l	98
24) 1,1-Dichloroethane	6.589	63	241575	52.602	ug/l	99
25) 2-Butanone	7.489	43	582343	246.018	ug/l	95
26) 2,2-Dichloropropane	7.506	77	215080	60.206	ug/l	99
27) cis-1,2-Dichloroethene	7.500	96	162658	53.551	ug/l	98
28) Bromochloromethane	7.824	49	118461	52.462	ug/l	95
29) Tetrahydrofuran	7.847	42	380103	246.502	ug/l	96
30) Chloroform	7.977	83	241191	52.589	ug/l	98
31) Cyclohexane	8.265	56	208644	46.846	ug/l	97
32) 1,1,1-Trichloroethane	8.177	97	200139	51.307	ug/l	97
36) 1,1-Dichloropropene	8.377	75	172642	54.664	ug/l	99
37) Ethyl Acetate	7.571	43	211755	52.670	ug/l	99
38) Carbon Tetrachloride	8.371	117	168940	54.485	ug/l	94
39) Methylcyclohexane	9.606	83	206018	47.613	ug/l	96
40) Benzene	8.612	78	555597	53.798	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061125\
 Data File : VN086940.D
 Acq On : 11 Jun 2025 11:32
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050

Quant Time: Jun 12 01:28:53 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.789	41	115191	51.017	ug/l	93
42) 1,2-Dichloroethane	8.677	62	168630	53.833	ug/l	99
43) Isopropyl Acetate	8.694	43	338399	52.442	ug/l	97
44) Trichloroethene	9.353	130	136748	55.842	ug/l	98
45) 1,2-Dichloropropane	9.624	63	137250	54.626	ug/l	100
46) Dibromomethane	9.712	93	94620	56.697	ug/l	99
47) Bromodichloromethane	9.888	83	190362	55.440	ug/l	100
48) Methyl methacrylate	9.683	41	154138	51.898	ug/l	96
49) 1,4-Dioxane	9.700	88	60969	1128.409	ug/l #	96
51) 4-Methyl-2-Pentanone	10.447	43	1054797	271.529	ug/l	98
52) Toluene	10.630	92	349873	55.435	ug/l	98
53) t-1,3-Dichloropropene	10.835	75	217752	56.714	ug/l	100
54) cis-1,3-Dichloropropene	10.312	75	236296	57.518	ug/l	97
55) 1,1,2-Trichloroethane	11.018	97	137099	56.454	ug/l	99
56) Ethyl methacrylate	10.877	69	231226	59.778	ug/l	95
57) 1,3-Dichloropropane	11.165	76	232746	55.247	ug/l	98
58) 2-Chloroethyl Vinyl ether	10.159	63	665780	288.579	ug/l	99
59) 2-Hexanone	11.200	43	705660	282.011	ug/l	97
60) Dibromochloromethane	11.359	129	148069	58.520	ug/l	99
61) 1,2-Dibromoethane	11.471	107	137779	55.351	ug/l	99
64) Tetrachloroethene	11.106	164	105084	53.190	ug/l	95
65) Chlorobenzene	11.888	112	381598	55.424	ug/l	97
66) 1,1,1,2-Tetrachloroethane	11.959	131	124964	56.463	ug/l	99
67) Ethyl Benzene	11.965	91	642857	54.206	ug/l	99
68) m/p-Xylenes	12.071	106	500895	110.334	ug/l	99
69) o-Xylene	12.400	106	244038	56.127	ug/l	98
70) Styrene	12.412	104	424060	56.996	ug/l	100
71) Bromoform	12.576	173	99616	60.747	ug/l #	100
73) Isopropylbenzene	12.694	105	593080	53.511	ug/l	99
74) N-amyl acetate	12.512	43	208229	53.765	ug/l	94
75) 1,1,2,2-Tetrachloroethane	12.935	83	213329	56.816	ug/l	99
76) 1,2,3-Trichloropropane	12.994	75	288770	79.852	ug/l	68
77) Bromobenzene	12.976	156	146181	57.512	ug/l	98
78) n-propylbenzene	13.035	91	717012	53.234	ug/l	99
79) 2-Chlorotoluene	13.123	91	439522	54.413	ug/l	98
80) 1,3,5-Trimethylbenzene	13.171	105	490727	53.628	ug/l	98
81) trans-1,4-Dichloro-2-b...	12.735	75	93790	59.701	ug/l #	87
82) 4-Chlorotoluene	13.218	91	446237	54.598	ug/l	99
83) tert-Butylbenzene	13.435	119	448106	53.498	ug/l	97
84) 1,2,4-Trimethylbenzene	13.482	105	495936	54.048	ug/l	99
85) sec-Butylbenzene	13.612	105	617792	50.788	ug/l	99
86) p-Isopropyltoluene	13.729	119	524440	52.156	ug/l	98
87) 1,3-Dichlorobenzene	13.729	146	280357	56.062	ug/l	99
88) 1,4-Dichlorobenzene	13.812	146	287122	56.315	ug/l	99
89) n-Butylbenzene	14.053	91	477306	49.007	ug/l	100
90) Hexachloroethane	14.329	117	89886	52.738	ug/l	99
91) 1,2-Dichlorobenzene	14.106	146	267932	55.766	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	14.717	75	47326	52.704	ug/l	92
93) 1,2,4-Trichlorobenzene	15.388	180	160030	52.161	ug/l	99
94) Hexachlorobutadiene	15.500	225	51073	44.682	ug/l	99
95) Naphthalene	15.635	128	624421	54.679	ug/l	100
96) 1,2,3-Trichlorobenzene	15.835	180	151411	49.672	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061125\
Data File : VN086940.D
Acq On : 11 Jun 2025 11:32
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1


Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Quant Time: Jun 12 01:28:53 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 02:12:50 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jun 12 01:28:53 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 02:12:50 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.230 min Scan# 1068
Delta R.T. -0.006 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion:168 Resp: 205098
Ion Ratio Lower Upper
168 100
99 57.7 49.1 73.7



#2
Dichlorodifluoromethane
Concen: 53.916 ug/l
RT: 2.154 min Scan# 34
Delta R.T. -0.000 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32

Tgt Ion: 85 Resp: 110255
Ion Ratio Lower Upper
85 100
87 32.4 16.2 48.6









#3
Chloromethane
Concen: 43.069 ug/l
RT: 2.395 min Scan# 75
Delta R.T. -0.000 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion: 50 Resp: 113731
Ion Ratio Lower Upper
50 100
52 35.2 25.7 38.5



#4
Vinyl Chloride
Concen: 52.773 ug/l
RT: 2.554 min Scan# 102
Delta R.T. -0.000 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32

Tgt Ion: 62 Resp: 143755
Ion Ratio Lower Upper
62 100
64 31.8 26.0 39.0









#5
Bromomethane
Concen: 41.928 ug/l
RT: 2.989 min Scan# 11
Delta R.T. -0.006 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion: 94 Resp: 63914
Ion Ratio Lower Upper
94 100
96 93.9 77.9 116.9



#6
Chloroethane
Concen: 53.253 ug/l
RT: 3.154 min Scan# 204
Delta R.T. -0.000 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32

Tgt Ion: 64 Resp: 93702
Ion Ratio Lower Upper
64 100
66 31.2 24.6 36.8









#7
 Trichlorofluoromethane
 Concen: 53.570 ug/l
 RT: 3.524 min Scan# 20
 Delta R.T. -0.000 min
 Lab File: VN086940.D
 Acq: 11 Jun 2025 11:32




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050


Tgt Ion:101 Resp: 190657
 Ion Ratio Lower Upper
 101 100
 103 64.5 52.2 78.2



#8
 Diethyl Ether
 Concen: 55.281 ug/l
 RT: 3.983 min Scan# 345
 Delta R.T. -0.000 min
 Lab File: VN086940.D
 Acq: 11 Jun 2025 11:32

Tgt Ion: 74 Resp: 85722
 Ion Ratio Lower Upper
 74 100
 45 85.9 45.5 136.5









#9
1,1,2-Trichlorotrifluoroethane
Concen: 53.364 ug/l
RT: 4.401 min Scan# 415
Delta R.T. 0.006 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion:101 Resp: 119322
Ion Ratio Lower Upper
101 100
85 42.3 33.8 50.8
151 74.7 58.7 88.1



#10
Methyl Iodide
Concen: 39.959 ug/l
RT: 4.618 min Scan# 453
Delta R.T. 0.012 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32

Tgt Ion:142 Resp: 115825
Ion Ratio Lower Upper
142 100
127 36.0 30.9 46.3
141 13.1 10.4 15.6









#11
Tert butyl alcohol
Concen: 245.667 ug/l
RT: 5.542 min Scan# 611
Delta R.T. -0.006 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion: 59 Resp: 183049
Ion Ratio Lower Upper
59 100
57 9.9 8.1 12.1



#12
1,1-Dichloroethene
Concen: 53.307 ug/l
RT: 4.365 min Scan# 410
Delta R.T. 0.006 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32

Tgt Ion: 96 Resp: 121690
Ion Ratio Lower Upper
96 100
61 148.3 125.8 188.6
98 62.3 52.5 78.7









#13
Acrolein
Concen: 288.340 ug/l
RT: 4.201 min Scan# 382
Delta R.T. -0.000 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion: 56 Resp: 68006
Ion Ratio Lower Upper
56 100
55 70.2 55.6 83.4



#14
Allyl chloride
Concen: 49.420 ug/l
RT: 5.048 min Scan# 526
Delta R.T. 0.006 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32

Tgt Ion: 41 Resp: 187099
Ion Ratio Lower Upper
41 100
39 62.0 51.3 76.9
76 42.2 31.4 47.0









#15
Acrylonitrile
Concen: 262.349 ug/l
RT: 5.736 min Scan# 644
Delta R.T. -0.006 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion: 53 Resp: 456903
Ion Ratio Lower Upper
53 100
52 80.2 65.0 97.6
51 33.6 27.2 40.8



#16
Acetone
Concen: 270.423 ug/l
RT: 4.442 min Scan# 423
Delta R.T. -0.006 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32

Tgt Ion: 43 Resp: 393802
Ion Ratio Lower Upper
43 100
58 36.9 28.0 42.0









#17
Carbon Disulfide
Concen: 48.889 ug/l
RT: 4.736 min Scan# 473
Delta R.T. 0.006 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion: 76 Resp: 308706
Ion Ratio Lower Upper
76 100
78 9.2 7.3 10.9



#18
Methyl Acetate
Concen: 52.419 ug/l
RT: 5.042 min Scan# 525
Delta R.T. -0.000 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32

Tgt Ion: 43 Resp: 222452
Ion Ratio Lower Upper
43 100
74 26.7 20.7 31.1









#19
Methyl tert-butyl Ether
Concen: 53.894 ug/l
RT: 5.812 min Scan# 61
Delta R.T. -0.000 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion: 73 Resp: 445459
Ion Ratio Lower Upper
73 100
57 22.7 17.4 26.2



#20
Methylene Chloride
Concen: 50.660 ug/l
RT: 5.295 min Scan# 568
Delta R.T. -0.000 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32

Tgt Ion: 84 Resp: 138117
Ion Ratio Lower Upper
84 100
49 106.1 90.5 135.7
51 34.5 28.5 42.7
86 62.4 48.1 72.1









#21
trans-1,2-Dichloroethene
Concen: 51.154 ug/l
RT: 5.806 min Scan# 61
Delta R.T. -0.000 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion: 96 Resp: 129921
Ion Ratio Lower Upper
96 100
61 132.3 109.2 163.8
98 62.4 49.8 74.8



#22
Diisopropyl ether
Concen: 51.639 ug/l
RT: 6.683 min Scan# 804
Delta R.T. -0.006 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32

Tgt Ion: 45 Resp: 412097
Ion Ratio Lower Upper
45 100
43 52.4 41.9 62.9
87 30.3 22.4 33.6
59 12.3 9.8 14.6









#23
 Vinyl Acetate
 Concen: 261.960 ug/l
 RT: 6.618 min Scan# 793
 Delta R.T. -0.000 min
 Lab File: VN086940.D
 Acq: 11 Jun 2025 11:32




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050


Tgt Ion: 43 Resp: 1766271
 Ion Ratio Lower Upper
 43 100
 86 12.8 9.5 14.3



#24
 1,1-Dichloroethane
 Concen: 52.602 ug/l
 RT: 6.589 min Scan# 788
 Delta R.T. 0.006 min
 Lab File: VN086940.D
 Acq: 11 Jun 2025 11:32

Tgt Ion: 63 Resp: 241575
 Ion Ratio Lower Upper
 63 100
 98 7.4 3.6 10.8
 100 4.5 2.6 7.8









#25
2-Butanone
Concen: 246.018 ug/l
RT: 7.489 min Scan# 941
Delta R.T. -0.006 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion: 43 Resp: 582343
Ion Ratio Lower Upper
43 100
72 31.5 23.0 34.6



#26
2,2-Dichloropropane
Concen: 60.206 ug/l
RT: 7.506 min Scan# 944
Delta R.T. 0.006 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32

Tgt Ion: 77 Resp: 215080
Ion Ratio Lower Upper
77 100
97 23.4 11.6 34.7









#27
 cis-1,2-Dichloroethene
 Concen: 53.551 ug/l
 RT: 7.500 min Scan# 943
 Delta R.T. -0.000 min
 Lab File: VN086940.D
 Acq: 11 Jun 2025 11:32




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050


Tgt Ion: 96 Resp: 162658
 Ion Ratio Lower Upper
 96 100
 61 136.9 0.0 278.0
 98 65.1 0.0 128.2



#28
 Bromochloromethane
 Concen: 52.462 ug/l
 RT: 7.824 min Scan# 998
 Delta R.T. -0.000 min
 Lab File: VN086940.D
 Acq: 11 Jun 2025 11:32

Tgt Ion: 49 Resp: 118461
 Ion Ratio Lower Upper
 49 100
 129 2.2 0.0 3.8
 130 82.4 62.2 93.2





#29

Tetrahydrofuran

Concen: 246.502 ug/l

RT: 7.847 min Scan# 1002

Delta R.T. -0.006 min

Lab File: VN086940.D

Acq: 11 Jun 2025 11:32

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050





Tgt Ion: 42 Resp: 380103

Ion Ratio Lower Upper

42 100

72 50.1 37.8 56.6

71 47.0 35.6 53.4



#30

Chloroform

Concen: 52.589 ug/l

RT: 7.977 min Scan# 1024

Delta R.T. -0.000 min

Lab File: VN086940.D




Acq: 11 Jun 2025 11:32


Tgt Ion: 83 Resp: 241191

Ion Ratio Lower Upper

83 100

85 64.5 52.9 79.3









#31
Cyclohexane
Concen: 46.846 ug/l
RT: 8.265 min Scan# 1073
Delta R.T. -0.006 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion: 56 Resp: 208644
Ion Ratio Lower Upper
56 100
69 30.7 25.4 38.0
84 88.6 68.8 103.2



#32
1,1,1-Trichloroethane
Concen: 51.307 ug/l
RT: 8.177 min Scan# 1058
Delta R.T. -0.000 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32

Tgt Ion: 97 Resp: 200139
Ion Ratio Lower Upper
97 100
99 62.4 52.0 78.0
61 43.5 35.9 53.9





#33

1,2-Dichloroethane-d4

Concen: 46.478 ug/l

RT: 8.583 min Scan# 1127

Delta R.T. -0.006 min

Lab File: VN086940.D

Acq: 11 Jun 2025 11:32

Instrument :

MSVOA_N

ClientSampleId :





VSTDCCC050

Tgt Ion: 65 Resp: 127629

Ion Ratio Lower Upper

65 100

67 55.0 0.0 105.6



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.106 min Scan# 1216

Delta R.T. -0.000 min

Lab File: VN086940.D

Acq: 11 Jun 2025 11:32




Tgt Ion: 114 Resp: 357643


Ion Ratio Lower Upper

114 100

63 20.1 0.0 39.6

88 16.2 0.0 30.2









#35
Dibromofluoromethane
Concen: 52.151 ug/l
RT: 8.177 min Scan# 1057
Delta R.T. 0.006 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion:113 Resp: 110532
Ion Ratio Lower Upper
113 100
111 101.8 84.2 126.2
192 17.9 14.2 21.4



#36
1,1-Dichloropropene
Concen: 54.664 ug/l
RT: 8.377 min Scan# 1092
Delta R.T. -0.000 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32

Tgt Ion: 75 Resp: 172642
Ion Ratio Lower Upper
75 100
110 35.6 17.6 52.9
77 31.1 25.3 37.9





#37

Ethyl Acetate

Concen: 52.670 ug/l

RT: 7.571 min Scan# 91

Delta R.T. -0.006 min

Lab File: VN086940.D

Acq: 11 Jun 2025 11:32

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050





Tgt Ion: 43 Resp: 211755

Ion Ratio Lower Upper

43 100

61 15.2 12.1 18.1

70 12.8 9.8 14.6



#38

Carbon Tetrachloride

Concen: 54.485 ug/l

RT: 8.371 min Scan# 1091

Delta R.T. -0.000 min

Lab File: VN086940.D

Acq: 11 Jun 2025 11:32




Tgt Ion: 117 Resp: 168940


Ion Ratio Lower Upper

117 100

119 100.6 75.0 112.6

121 30.4 24.9 37.3





#39

Methylcyclohexane

Concen: 47.613 ug/l

RT: 9.606 min Scan# 1130

Delta R.T. -0.000 min

Lab File: VN086940.D

Acq: 11 Jun 2025 11:32

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050





Tgt Ion: 83 Resp: 206018

Ion Ratio Lower Upper

83 100

55 71.2 61.0 91.6

98 48.5 38.9 58.3



#40

Benzene

Concen: 53.798 ug/l

RT: 8.612 min Scan# 1132

Delta R.T. -0.000 min

Lab File: VN086940.D




Acq: 11 Jun 2025 11:32


Tgt Ion: 78 Resp: 555597

Ion Ratio Lower Upper

78 100

77 24.3 18.7 28.1









#41
Methacrylonitrile
Concen: 51.017 ug/l
RT: 7.789 min Scan# 992
Delta R.T. -0.000 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion:	41	Resp:	115191
Ion	Ratio	Lower	Upper
41	100		
39	53.8	41.2	61.8
67	85.5	61.4	92.2
52	30.7	22.7	34.1



#42
1,2-Dichloroethane
Concen: 53.833 ug/l
RT: 8.677 min Scan# 1143
Delta R.T. -0.000 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32

Tgt Ion:	62	Resp:	168630
Ion	Ratio	Lower	Upper
62	100		
98	11.0	0.0	21.2









#43
Isopropyl Acetate
Concen: 52.442 ug/l
RT: 8.694 min Scan# 1146
Delta R.T. -0.000 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion: 43 Resp: 338399
Ion Ratio Lower Upper
43 100
61 29.2 22.1 33.1
87 15.7 11.8 17.6



#44
Trichloroethene
Concen: 55.842 ug/l
RT: 9.353 min Scan# 1258
Delta R.T. -0.006 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32

Tgt Ion:130 Resp: 136748
Ion Ratio Lower Upper
130 100
95 96.1 0.0 196.2





#45

1,2-Dichloropropane

Concen: 54.626 ug/l

RT: 9.624 min Scan# 1304

Delta R.T. -0.006 min

Lab File: VN086940.D

Acq: 11 Jun 2025 11:32

Instrument :

MSVOA_N

ClientSampleId :





VSTDCCC050

Tgt Ion: 63 Resp: 137250

Ion Ratio Lower Upper

63 100

65 32.0 25.5 38.3



#46

Dibromomethane

Concen: 56.697 ug/l

RT: 9.712 min Scan# 1319

Delta R.T. -0.006 min

Lab File: VN086940.D

Acq: 11 Jun 2025 11:32




Tgt Ion: 93 Resp: 94620


Ion Ratio Lower Upper

93 100

95 82.9 67.1 100.7

174 94.1 74.2 111.4









#47
Bromodichloromethane
Concen: 55.440 ug/l
RT: 9.888 min Scan# 11
Delta R.T. -0.006 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion: 83 Resp: 190362
Ion Ratio Lower Upper
83 100
85 63.3 50.9 76.3
127 7.9 6.6 10.0



#48
Methyl methacrylate
Concen: 51.898 ug/l
RT: 9.683 min Scan# 1314
Delta R.T. -0.000 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32

Tgt Ion: 41 Resp: 154138
Ion Ratio Lower Upper
41 100
69 98.4 74.6 112.0
39 49.7 40.2 60.4





#49

1,4-Dioxane

Concen: 1128.409 ug/l

RT: 9.700 min Scan# 11

Delta R.T. -0.000 min

Lab File: VN086940.D

Acq: 11 Jun 2025 11:32

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050





Tgt Ion: 88 Resp: 60969

Ion Ratio Lower Upper

88 100

43 0.0 0.0 0.0

58 68.8 57.7 86.5



#50

Toluene-d8

Concen: 50.342 ug/l

RT: 10.565 min Scan# 1464

Delta R.T. -0.006 min

Lab File: VN086940.D




Acq: 11 Jun 2025 11:32


Tgt Ion: 98 Resp: 422391

Ion Ratio Lower Upper

98 100

100 68.2 53.4 80.0









#51
4-Methyl-2-Pentanone
Concen: 271.529 ug/l
RT: 10.447 min Scan# 1444
Delta R.T. -0.000 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion: 43 Resp: 1054797
Ion Ratio Lower Upper
43 100
58 44.3 34.2 51.4



#52
Toluene
Concen: 55.435 ug/l
RT: 10.630 min Scan# 1475
Delta R.T. -0.006 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32

Tgt Ion: 92 Resp: 349873
Ion Ratio Lower Upper
92 100
91 167.2 136.5 204.7









#53
t-1,3-Dichloropropene
Concen: 56.714 ug/l
RT: 10.835 min Scan# 1510
Delta R.T. -0.006 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion: 75 Resp: 217752
Ion Ratio Lower Upper
75 100
77 32.7 26.4 39.6



#54
cis-1,3-Dichloropropene
Concen: 57.518 ug/l
RT: 10.312 min Scan# 1421
Delta R.T. -0.006 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32

Tgt Ion: 75 Resp: 236296
Ion Ratio Lower Upper
75 100
77 31.6 26.2 39.4
39 38.5 32.3 48.5









#55
 1,1,2-Trichloroethane
 Concen: 56.454 ug/l
 RT: 11.018 min Scan# 1541
 Delta R.T. -0.000 min
 Lab File: VN086940.D
 Acq: 11 Jun 2025 11:32

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050

Tgt Ion: 97 Resp: 137099




Ion	Ratio	Lower	Upper
97	100		
83	86.6	69.7	104.5
85	54.1	45.1	67.7
99	62.2	50.6	75.8




#56
 Ethyl methacrylate
 Concen: 59.778 ug/l
 RT: 10.877 min Scan# 1517
 Delta R.T. -0.006 min
 Lab File: VN086940.D
 Acq: 11 Jun 2025 11:32

Tgt Ion: 69 Resp: 231226

Ion	Ratio	Lower	Upper
69	100		
41	55.6	47.2	70.8
39	26.0	23.0	34.6





#57

1,3-Dichloropropane

Concen: 55.247 ug/l

RT: 11.165 min Scan# 11

Delta R.T. -0.000 min

Lab File: VN086940.D

Acq: 11 Jun 2025 11:32

Instrument :

MSVOA_N

ClientSampleId :





VSTDCCC050

Tgt Ion: 76 Resp: 232746

Ion Ratio Lower Upper

76 100

78 32.8 25.3 37.9



#58

2-Chloroethyl Vinyl ether

Concen: 288.579 ug/l

RT: 10.159 min Scan# 1395

Delta R.T. -0.006 min

Lab File: VN086940.D




Acq: 11 Jun 2025 11:32


Tgt Ion: 63 Resp: 665780

Ion Ratio Lower Upper

63 100

106 29.5 23.0 34.6









#59
2-Hexanone
Concen: 282.011 ug/l
RT: 11.200 min Scan# 1572
Delta R.T. -0.000 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion: 43 Resp: 705660
Ion Ratio Lower Upper
43 100
58 63.5 30.4 91.3



#60
Dibromochloromethane
Concen: 58.520 ug/l
RT: 11.359 min Scan# 1599
Delta R.T. -0.000 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32

Tgt Ion: 129 Resp: 148069
Ion Ratio Lower Upper
129 100
127 76.4 38.6 115.8









#61
1,2-Dibromoethane
Concen: 55.351 ug/l
RT: 11.471 min Scan# 1618
Delta R.T. -0.000 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion:107 Resp: 137779
Ion Ratio Lower Upper
107 100
109 96.1 76.3 114.5



#62
4-Bromofluorobenzene
Concen: 50.489 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. -0.000 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32

Tgt Ion: 95 Resp: 157388
Ion Ratio Lower Upper
95 100
174 71.7 0.0 141.8
176 68.2 0.0 132.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 117
Delta R.T. -0.000 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion:117 Resp: 312173
Ion Ratio Lower Upper
117 100
82 53.3 44.6 67.0
119 31.8 25.5 38.3



#64
Tetrachloroethene
Concen: 53.190 ug/l
RT: 11.106 min Scan# 1556
Delta R.T. -0.000 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32

Tgt Ion:164 Resp: 105084
Ion Ratio Lower Upper
164 100
166 124.8 101.8 152.8
129 95.6 83.0 124.4
131 93.8 78.2 117.2









#65
Chlorobenzene
Concen: 55.424 ug/l
RT: 11.888 min Scan# 1689
Delta R.T. -0.006 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion:112 Resp: 381598
Ion Ratio Lower Upper
112 100
114 31.1 26.2 39.2



#66
1,1,1,2-Tetrachloroethane
Concen: 56.463 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. -0.006 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32

Tgt Ion:131 Resp: 124964
Ion Ratio Lower Upper
131 100
133 95.4 48.3 144.8
119 68.5 34.0 101.8





#67

Ethyl Benzene

Concen: 54.206 ug/l

RT: 11.965 min Scan# 1702

Delta R.T. -0.000 min

Lab File: VN086940.D

Acq: 11 Jun 2025 11:32

Instrument :

MSVOA_N

ClientSampleId :





VSTDCCC050

Tgt Ion: 91 Resp: 642857

Ion Ratio Lower Upper

91 100

106 32.5 25.4 38.2



#68

m/p-Xylenes

Concen: 110.334 ug/l

RT: 12.071 min Scan# 1720

Delta R.T. -0.000 min

Lab File: VN086940.D




Acq: 11 Jun 2025 11:32


Tgt Ion: 106 Resp: 500895

Ion Ratio Lower Upper

106 100

91 198.0 159.4 239.0





#69

o-Xylene

Concen: 56.127 ug/l

RT: 12.400 min Scan# 1776

Delta R.T. -0.000 min

Lab File: VN086940.D

Acq: 11 Jun 2025 11:32

Instrument :

MSVOA_N

ClientSampleId :





VSTDCCC050

Tgt Ion:106 Resp: 244038

Ion Ratio Lower Upper

106 100

91 208.5 106.1 318.1



#70

Styrene

Concen: 56.996 ug/l

RT: 12.412 min Scan# 1778

Delta R.T. -0.000 min

Lab File: VN086940.D

Acq: 11 Jun 2025 11:32




Tgt Ion:104 Resp: 424060


Ion Ratio Lower Upper

104 100

78 47.4 38.5 57.7

103 53.9 43.1 64.7





#71

Bromoform

Concen: 60.747 ug/l

RT: 12.576 min Scan# 11

Delta R.T. -0.006 min

Lab File: VN086940.D

Acq: 11 Jun 2025 11:32

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050





Tgt Ion:173 Resp: 99616

Ion Ratio Lower Upper

173 100

175 48.1 24.1 72.4

254 0.0 0.0 0.0



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.788 min Scan# 2012

Delta R.T. -0.006 min

Lab File: VN086940.D

Acq: 11 Jun 2025 11:32




Tgt Ion:152 Resp: 152136


Ion Ratio Lower Upper

152 100

115 60.1 30.1 90.5

150 176.5 0.0 345.0









#73
Isopropylbenzene
Concen: 53.511 ug/l
RT: 12.694 min Scan# 1826
Delta R.T. -0.000 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion:105 Resp: 593080
Ion Ratio Lower Upper
105 100
120 27.5 13.5 40.4



#74
N-ethyl acetate
Concen: 53.765 ug/l
RT: 12.512 min Scan# 1795
Delta R.T. -0.000 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32

Tgt Ion: 43 Resp: 208229
Ion Ratio Lower Upper
43 100
70 52.6 37.6 56.4
55 26.6 19.0 28.6
61 26.4 20.8 31.2









#75
 1,1,2,2-Tetrachloroethane
 Concen: 56.816 ug/l
 RT: 12.935 min Scan# 1867
 Delta R.T. -0.006 min
 Lab File: VN086940.D
 Acq: 11 Jun 2025 11:32




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050


Tgt Ion: 83 Resp: 213329
 Ion Ratio Lower Upper
 83 100
 131 9.8 4.9 14.7
 85 64.5 32.0 96.0



#76
 1,2,3-Trichloropropane
 Concen: 79.852 ug/l
 RT: 12.994 min Scan# 1877
 Delta R.T. -0.000 min
 Lab File: VN086940.D
 Acq: 11 Jun 2025 11:32

Tgt Ion: 75 Resp: 288770
 Ion Ratio Lower Upper
 75 100
 77 116.7 79.3 237.9









#77
Bromobenzene
Concen: 57.512 ug/l
RT: 12.976 min Scan# 1874
Delta R.T. -0.006 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion:156 Resp: 146181
Ion Ratio Lower Upper
156 100
77 230.6 117.7 353.1
158 97.1 48.1 144.4



#78
n-propylbenzene
Concen: 53.234 ug/l
RT: 13.035 min Scan# 1884
Delta R.T. -0.000 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32

Tgt Ion: 91 Resp: 717012
Ion Ratio Lower Upper
91 100
120 23.3 11.4 34.2





#79

2-Chlorotoluene

Concen: 54.413 ug/l

RT: 13.123 min Scan# 1899

Delta R.T. -0.000 min

Lab File: VN086940.D

Acq: 11 Jun 2025 11:32

Instrument :

MSVOA_N

ClientSampleId :





VSTDCCC050

Tgt Ion: 91 Resp: 439522

Ion Ratio Lower Upper

91 100

126 33.4 16.2 48.5



#80

1,3,5-Trimethylbenzene

Concen: 53.628 ug/l

RT: 13.171 min Scan# 1907

Delta R.T. -0.000 min

Lab File: VN086940.D




Acq: 11 Jun 2025 11:32


Tgt Ion:105 Resp: 490727

Ion Ratio Lower Upper

105 100

120 50.8 24.9 74.6









#81
trans-1,4-Dichloro-2-butene
Concen: 59.701 ug/l
RT: 12.735 min Scan# 1833
Delta R.T. -0.000 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion: 75 Resp: 93790
Ion Ratio Lower Upper
75 100
53 74.1 66.6 99.8
89 57.1 36.7 55.1#



#82
4-Chlorotoluene
Concen: 54.598 ug/l
RT: 13.218 min Scan# 1915
Delta R.T. -0.006 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32

Tgt Ion: 91 Resp: 446237
Ion Ratio Lower Upper
91 100
126 34.0 16.7 50.1









#83
 tert-Butylbenzene
 Concen: 53.498 ug/l
 RT: 13.435 min Scan# 1952
 Delta R.T. -0.000 min
 Lab File: VN086940.D
 Acq: 11 Jun 2025 11:32




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050


Tgt Ion:119 Resp: 448106
 Ion Ratio Lower Upper
 119 100
 91 61.2 32.1 96.5
 134 26.3 13.4 40.1



#84
 1,2,4-Trimethylbenzene
 Concen: 54.048 ug/l
 RT: 13.482 min Scan# 1960
 Delta R.T. -0.000 min
 Lab File: VN086940.D
 Acq: 11 Jun 2025 11:32

Tgt Ion:105 Resp: 495936
 Ion Ratio Lower Upper
 105 100
 120 47.2 23.2 69.6









#85
 sec-Butylbenzene
 Concen: 50.788 ug/l
 RT: 13.612 min Scan# 1982
 Delta R.T. -0.006 min
 Lab File: VN086940.D
 Acq: 11 Jun 2025 11:32




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050


Tgt Ion:105 Resp: 617792
 Ion Ratio Lower Upper
 105 100
 134 20.6 10.0 30.0



#86
 p-Isopropyltoluene
 Concen: 52.156 ug/l
 RT: 13.729 min Scan# 2002
 Delta R.T. -0.000 min
 Lab File: VN086940.D
 Acq: 11 Jun 2025 11:32

Tgt Ion:119 Resp: 524440
 Ion Ratio Lower Upper
 119 100
 134 26.4 13.5 40.4
 91 21.8 11.5 34.4









#87
1,3-Dichlorobenzene
Concen: 56.062 ug/l
RT: 13.729 min Scan# 2003
Delta R.T. -0.006 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion:146 Resp: 280357
Ion Ratio Lower Upper
146 100
111 40.9 20.9 62.8
148 63.8 32.0 96.2



#88
1,4-Dichlorobenzene
Concen: 56.315 ug/l
RT: 13.812 min Scan# 2016
Delta R.T. -0.000 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32

Tgt Ion:146 Resp: 287122
Ion Ratio Lower Upper
146 100
111 39.5 19.9 59.7
148 62.7 31.8 95.4









#89
n-Butylbenzene
Concen: 49.007 ug/l
RT: 14.053 min Scan# 2057
Delta R.T. -0.000 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion: 91 Resp: 477306
Ion Ratio Lower Upper
91 100
92 54.9 27.5 82.4
134 26.6 13.0 39.0



#90
Hexachloroethane
Concen: 52.738 ug/l
RT: 14.329 min Scan# 2104
Delta R.T. -0.006 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32

Tgt Ion: 117 Resp: 89886
Ion Ratio Lower Upper
117 100
201 67.6 33.2 99.6









#91
 1,2-Dichlorobenzene
 Concen: 55.766 ug/l
 RT: 14.106 min Scan# 2066
 Delta R.T. -0.000 min
 Lab File: VN086940.D
 Acq: 11 Jun 2025 11:32




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050


Tgt Ion:146 Resp: 267932
 Ion Ratio Lower Upper
 146 100
 111 43.2 21.6 64.6
 148 63.3 31.7 95.1



#92
 1,2-Dibromo-3-Chloropropane
 Concen: 52.704 ug/l
 RT: 14.717 min Scan# 2170
 Delta R.T. -0.000 min
 Lab File: VN086940.D
 Acq: 11 Jun 2025 11:32

Tgt Ion: 75 Resp: 47326
 Ion Ratio Lower Upper
 75 100
 155 87.3 41.1 123.4
 157 114.9 52.7 158.1





#93

1,2,4-Trichlorobenzene

Concen: 52.161 ug/l

RT: 15.388 min Scan# 21

Delta R.T. -0.000 min

Lab File: VN086940.D

Acq: 11 Jun 2025 11:32

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050





Tgt Ion:180 Resp: 160030

Ion Ratio Lower Upper

180 100

182 92.1 46.9 140.7

145 31.7 16.0 47.9



#94

Hexachlorobutadiene

Concen: 44.682 ug/l

RT: 15.500 min Scan# 2303

Delta R.T. -0.000 min

Lab File: VN086940.D

Acq: 11 Jun 2025 11:32




Tgt Ion:225 Resp: 51073


Ion Ratio Lower Upper

225 100

223 65.0 33.5 100.4

227 62.2 31.1 93.5









#95
Naphthalene
Concen: 54.679 ug/l
RT: 15.635 min Scan# 21
Delta R.T. -0.006 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Tgt Ion:128 Resp: 624421
Ion Ratio Lower Upper
128 100
127 12.6 10.2 15.2
129 10.8 8.8 13.2



#96
1,2,3-Trichlorobenzene
Concen: 49.672 ug/l
RT: 15.835 min Scan# 2360
Delta R.T. -0.000 min
Lab File: VN086940.D
Acq: 11 Jun 2025 11:32

Tgt Ion:180 Resp: 151411
Ion Ratio Lower Upper
180 100
182 94.5 47.0 141.0
145 33.0 16.6 49.8



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061125\
 Data File : VN086944.D
 Acq On : 11 Jun 2025 13:27
 Operator : JC\MD
 Sample : VN0611WBS02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0611WBS02

Quant Time: Jun 12 01:31:27 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	206799	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	370524	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	321188	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	157403	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	130342	47.075	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	94.160%
35) Dibromofluoromethane	8.177	113	112198	51.096	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.200%
50) Toluene-d8	10.571	98	423380	48.706	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.420%
62) 4-Bromofluorobenzene	12.847	95	161321	49.951	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	99.900%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	2.153	85	40672	19.725	ug/l	100
3) Chloromethane	2.401	50	42812	16.079	ug/l	97
4) Vinyl Chloride	2.553	62	51780	18.852	ug/l	95
5) Bromomethane	3.000	94	25577	16.641	ug/l	90
6) Chloroethane	3.153	64	34500	19.446	ug/l	99
7) Trichlorofluoromethane	3.530	101	68752	19.159	ug/l	99
8) Diethyl Ether	3.983	74	30914	19.772	ug/l	94
9) 1,1,2-Trichlorotrifluo...	4.400	101	44117	19.568	ug/l	99
10) Methyl Iodide	4.612	142	36924	12.634	ug/l	100
11) Tert butyl alcohol	5.541	59	69743	92.831	ug/l	99
12) 1,1-Dichloroethene	4.365	96	44269	19.233	ug/l	97
13) Acrolein	4.206	56	22290	93.730	ug/l	98
14) Allyl chloride	5.047	41	68027	17.821	ug/l	97
15) Acrylonitrile	5.741	53	166971	95.084	ug/l	100
16) Acetone	4.447	43	145322	98.972	ug/l	98
17) Carbon Disulfide	4.742	76	113132	17.769	ug/l	100
18) Methyl Acetate	5.047	43	81357	19.013	ug/l	99
19) Methyl tert-butyl Ether	5.812	73	163338	19.599	ug/l	98
20) Methylene Chloride	5.294	84	52196	18.988	ug/l	96
21) trans-1,2-Dichloroethene	5.812	96	47284	18.464	ug/l	95
22) Diisopropyl ether	6.688	45	155508	19.326	ug/l	97
23) Vinyl Acetate	6.618	43	653278	96.092	ug/l	98
24) 1,1-Dichloroethane	6.588	63	89930	19.421	ug/l	100
25) 2-Butanone	7.494	43	215760	90.401	ug/l	96
26) 2,2-Dichloropropane	7.500	77	78031	21.663	ug/l	98
27) cis-1,2-Dichloroethene	7.500	96	59578	19.453	ug/l	99
28) Bromochloromethane	7.824	49	45329	19.909	ug/l	95
29) Tetrahydrofuran	7.853	42	143866	92.532	ug/l	97
30) Chloroform	7.977	83	90272	19.521	ug/l	98
31) Cyclohexane	8.265	56	79050	17.603	ug/l	97
32) 1,1,1-Trichloroethane	8.182	97	74861	19.033	ug/l	97
36) 1,1-Dichloropropene	8.377	75	63577	19.431	ug/l	99
37) Ethyl Acetate	7.571	43	80415	19.306	ug/l	99
38) Carbon Tetrachloride	8.371	117	61747	19.222	ug/l	97
39) Methylcyclohexane	9.606	83	75971	16.947	ug/l	98
40) Benzene	8.612	78	206255	19.277	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061125\
 Data File : VN086944.D
 Acq On : 11 Jun 2025 13:27
 Operator : JC\MD
 Sample : VN0611WBS02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0611WBS02

Quant Time: Jun 12 01:31:27 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.788	41	42617	18.218	ug/l	93
42) 1,2-Dichloroethane	8.677	62	63379	19.530	ug/l	98
43) Isopropyl Acetate	8.694	43	126914	18.984	ug/l	98
44) Trichloroethene	9.359	130	50593	19.942	ug/l	99
45) 1,2-Dichloropropane	9.624	63	51091	19.628	ug/l	95
46) Dibromomethane	9.712	93	34741	20.093	ug/l	99
47) Bromodichloromethane	9.894	83	69845	19.634	ug/l	98
48) Methyl methacrylate	9.682	41	57757	18.771	ug/l	99
49) 1,4-Dioxane	9.706	88	22914	409.347	ug/l #	97
51) 4-Methyl-2-Pentanone	10.447	43	394396	97.997	ug/l	99
52) Toluene	10.629	92	127563	19.509	ug/l	99
53) t-1,3-Dichloropropene	10.835	75	79224	19.917	ug/l	100
54) cis-1,3-Dichloropropene	10.318	75	85586	20.109	ug/l	99
55) 1,1,2-Trichloroethane	11.018	97	51421	20.438	ug/l	94
56) Ethyl methacrylate	10.882	69	80942	20.198	ug/l	94
57) 1,3-Dichloropropane	11.165	76	85602	19.613	ug/l	99
58) 2-Chloroethyl Vinyl ether	10.165	63	243584	101.910	ug/l	99
59) 2-Hexanone	11.206	43	231264	89.210	ug/l	95
60) Dibromochloromethane	11.359	129	53655	20.468	ug/l	99
61) 1,2-Dibromoethane	11.471	107	50832	19.711	ug/l	99
64) Tetrachloroethene	11.106	164	38926	19.150	ug/l	98
65) Chlorobenzene	11.894	112	141721	20.006	ug/l	98
66) 1,1,1,2-Tetrachloroethane	11.959	131	46404	20.378	ug/l	99
67) Ethyl Benzene	11.965	91	239733	19.647	ug/l	100
68) m/p-Xylenes	12.070	106	185026	39.613	ug/l	99
69) o-Xylene	12.400	106	90240	20.172	ug/l	99
70) Styrene	12.412	104	154149	20.137	ug/l	99
71) Bromoform	12.582	173	35350	20.952	ug/l #	96
73) Isopropylbenzene	12.694	105	221545	19.320	ug/l	99
74) N-amyl acetate	12.529	43	63289	15.794	ug/l #	85
75) 1,1,2,2-Tetrachloroethane	12.941	83	81476	20.973	ug/l	100
76) 1,2,3-Trichloropropane	12.994	75	108842	29.090	ug/l	68
77) Bromobenzene	12.982	156	55176	20.981	ug/l	98
78) n-propylbenzene	13.035	91	267384	19.187	ug/l	100
79) 2-Chlorotoluene	13.123	91	164813	19.721	ug/l	99
80) 1,3,5-Trimethylbenzene	13.170	105	181773	19.200	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.735	75	30121	18.532	ug/l #	82
82) 4-Chlorotoluene	13.217	91	167991	19.866	ug/l	100
83) tert-Butylbenzene	13.435	119	161593	18.646	ug/l	99
84) 1,2,4-Trimethylbenzene	13.482	105	182431	19.216	ug/l	98
85) sec-Butylbenzene	13.617	105	228907	18.189	ug/l	100
86) p-Isopropyltoluene	13.729	119	192991	18.551	ug/l	99
87) 1,3-Dichlorobenzene	13.735	146	104883	20.271	ug/l	98
88) 1,4-Dichlorobenzene	13.812	146	107750	20.426	ug/l	99
89) n-Butylbenzene	14.053	91	177482	17.613	ug/l	100
90) Hexachloroethane	14.329	117	33456	18.972	ug/l	99
91) 1,2-Dichlorobenzene	14.106	146	99394	19.995	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	14.717	75	17796	19.155	ug/l	98
93) 1,2,4-Trichlorobenzene	15.394	180	58509	18.433	ug/l	98
94) Hexachlorobutadiene	15.500	225	19592	16.567	ug/l	97
95) Naphthalene	15.635	128	232760	19.700	ug/l	100
96) 1,2,3-Trichlorobenzene	15.841	180	56144	17.802	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061125\
Data File : VN086944.D
Acq On : 11 Jun 2025 13:27
Operator : JC\MD
Sample : VN0611WBS02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 6 Sample Multiplier: 1


Instrument :
MSVOA_N
ClientSampleId :
VN0611WBS02


Quant Time: Jun 12 01:31:27 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 02:12:50 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jun 12 01:31:27 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 02:12:50 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.235 min Scan# 1068
Delta R.T. -0.001 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBS02


Tgt Ion:168 Resp: 206799
Ion Ratio Lower Upper
168 100
99 57.5 49.1 73.7



#2
Dichlorodifluoromethane
Concen: 19.725 ug/l
RT: 2.153 min Scan# 34
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27

Tgt Ion: 85 Resp: 40672
Ion Ratio Lower Upper
85 100
87 32.5 16.2 48.6









#3
Chloromethane
Concen: 16.079 ug/l
RT: 2.401 min Scan# 75
Delta R.T. 0.006 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBS02


Tgt Ion: 50 Resp: 42812
Ion Ratio Lower Upper
50 100
52 33.8 25.7 38.5



#4
Vinyl Chloride
Concen: 18.852 ug/l
RT: 2.553 min Scan# 102
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27

Tgt Ion: 62 Resp: 51780
Ion Ratio Lower Upper
62 100
64 29.7 26.0 39.0









#5
Bromomethane
Concen: 16.641 ug/l
RT: 3.000 min Scan# 177
Delta R.T. 0.006 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBS02


Tgt Ion: 94 Resp: 25577
Ion Ratio Lower Upper
94 100
96 87.8 77.9 116.9



#6
Chloroethane
Concen: 19.446 ug/l
RT: 3.153 min Scan# 204
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27

Tgt Ion: 64 Resp: 34500
Ion Ratio Lower Upper
64 100
66 30.3 24.6 36.8









#7
Trichlorofluoromethane
Concen: 19.159 ug/l
RT: 3.530 min Scan# 267
Delta R.T. 0.006 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBS02


Tgt Ion:101 Resp: 68752
Ion Ratio Lower Upper
101 100
103 64.8 52.2 78.2



#8
Diethyl Ether
Concen: 19.772 ug/l
RT: 3.983 min Scan# 345
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27

Tgt Ion: 74 Resp: 30914
Ion Ratio Lower Upper
74 100
45 85.5 45.5 136.5









#9
1,1,2-Trichlorotrifluoroethane
Concen: 19.568 ug/l
RT: 4.400 min Scan# 415
Delta R.T. 0.006 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBS02


Tgt Ion:101 Resp: 44117
Ion Ratio Lower Upper
101 100
85 41.5 33.8 50.8
151 74.0 58.7 88.1



#10
Methyl Iodide
Concen: 12.634 ug/l
RT: 4.612 min Scan# 452
Delta R.T. 0.006 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27

Tgt Ion:142 Resp: 36924
Ion Ratio Lower Upper
142 100
127 38.9 30.9 46.3
141 12.9 10.4 15.6









#11
Tert butyl alcohol
Concen: 92.831 ug/l
RT: 5.541 min Scan# 611
Delta R.T. -0.006 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBS02


Tgt Ion: 59 Resp: 69743
Ion Ratio Lower Upper
59 100
57 9.8 8.1 12.1



#12
1,1-Dichloroethene
Concen: 19.233 ug/l
RT: 4.365 min Scan# 410
Delta R.T. 0.006 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27

Tgt Ion: 96 Resp: 44269
Ion Ratio Lower Upper
96 100
61 153.8 125.8 188.6
98 61.2 52.5 78.7









#13
Acrolein
Concen: 93.730 ug/l
RT: 4.206 min Scan# 382
Delta R.T. 0.006 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBS02


Tgt Ion: 56 Resp: 22290
Ion Ratio Lower Upper
56 100
55 70.9 55.6 83.4



#14
Allyl chloride
Concen: 17.821 ug/l
RT: 5.047 min Scan# 526
Delta R.T. 0.006 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27


Tgt Ion: 41 Resp: 68027
Ion Ratio Lower Upper
41 100
39 62.5 51.3 76.9
76 41.4 31.4 47.0








#15
Acrylonitrile
Concen: 95.084 ug/l
RT: 5.741 min Scan# 644
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27

Instrument :
MSVOA_N
ClientSampleId :
VN0611WBS02



Tgt Ion: 53 Resp: 166971




Ion	Ratio	Lower	Upper
53	100		
52	81.1	65.0	97.6
51	34.5	27.2	40.8




#16
Acetone
Concen: 98.972 ug/l
RT: 4.447 min Scan# 424
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27

Tgt Ion: 43 Resp: 145322

Ion	Ratio	Lower	Upper
43	100		
58	35.9	28.0	42.0









#17
Carbon Disulfide
Concen: 17.769 ug/l
RT: 4.742 min Scan# 474
Delta R.T. 0.012 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBS02


Tgt Ion: 76 Resp: 113132
Ion Ratio Lower Upper
76 100
78 9.0 7.3 10.9



#18
Methyl Acetate
Concen: 19.013 ug/l
RT: 5.047 min Scan# 526
Delta R.T. 0.006 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27

Tgt Ion: 43 Resp: 81357
Ion Ratio Lower Upper
43 100
74 26.2 20.7 31.1









#19
Methyl tert-butyl Ether
Concen: 19.599 ug/l
RT: 5.812 min Scan# 61
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBS02


Tgt Ion: 73 Resp: 163338
Ion Ratio Lower Upper
73 100
57 21.0 17.4 26.2



#20
Methylene Chloride
Concen: 18.988 ug/l
RT: 5.294 min Scan# 568
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27

Tgt Ion: 84 Resp: 52196
Ion Ratio Lower Upper
84 100
49 108.4 90.5 135.7
51 35.0 28.5 42.7
86 64.6 48.1 72.1









#21
trans-1,2-Dichloroethene
Concen: 18.464 ug/l
RT: 5.812 min Scan# 61
Delta R.T. 0.006 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBS02


Tgt Ion: 96 Resp: 47284
Ion Ratio Lower Upper
96 100
61 127.9 109.2 163.8
98 61.9 49.8 74.8



#22
Diisopropyl ether
Concen: 19.326 ug/l
RT: 6.688 min Scan# 805
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27

Tgt Ion: 45 Resp: 155508
Ion Ratio Lower Upper
45 100
43 50.5 41.9 62.9
87 29.8 22.4 33.6
59 12.1 9.8 14.6









#23
 Vinyl Acetate
 Concen: 96.092 ug/l
 RT: 6.618 min Scan# 793
 Delta R.T. -0.000 min
 Lab File: VN086944.D
 Acq: 11 Jun 2025 13:27




Instrument :
 MSVOA_N
 ClientSampleId :
 VN0611WBS02


Tgt Ion: 43 Resp: 653278
 Ion Ratio Lower Upper
 43 100
 86 12.6 9.5 14.3



#24
 1,1-Dichloroethane
 Concen: 19.421 ug/l
 RT: 6.588 min Scan# 788
 Delta R.T. 0.006 min
 Lab File: VN086944.D
 Acq: 11 Jun 2025 13:27

Tgt Ion: 63 Resp: 89930
 Ion Ratio Lower Upper
 63 100
 98 7.3 3.6 10.8
 100 5.3 2.6 7.8









#25
2-Butanone
Concen: 90.401 ug/l
RT: 7.494 min Scan# 942
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBS02


Tgt Ion: 43 Resp: 215760
Ion Ratio Lower Upper
43 100
72 30.8 23.0 34.6



#26
2,2-Dichloropropane
Concen: 21.663 ug/l
RT: 7.500 min Scan# 943
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27

Tgt Ion: 77 Resp: 78031
Ion Ratio Lower Upper
77 100
97 24.3 11.6 34.7









#27
 cis-1,2-Dichloroethene
 Concen: 19.453 ug/l
 RT: 7.500 min Scan# 943
 Delta R.T. -0.000 min
 Lab File: VN086944.D
 Acq: 11 Jun 2025 13:27




Instrument :
 MSVOA_N
 ClientSampleId :
 VN0611WBS02


Tgt Ion: 96 Resp: 59578
 Ion Ratio Lower Upper
 96 100
 61 138.3 0.0 278.0
 98 64.6 0.0 128.2



#28
 Bromochloromethane
 Concen: 19.909 ug/l
 RT: 7.824 min Scan# 998
 Delta R.T. -0.000 min
 Lab File: VN086944.D
 Acq: 11 Jun 2025 13:27

Tgt Ion: 49 Resp: 45329
 Ion Ratio Lower Upper
 49 100
 129 1.6 0.0 3.8
 130 82.2 62.2 93.2





#29

Tetrahydrofuran

Concen: 92.532 ug/l

RT: 7.853 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086944.D

Acq: 11 Jun 2025 13:27

Instrument :

MSVOA_N

ClientSampleId :

VN0611WBS02





Tgt Ion: 42 Resp: 143866

Ion Ratio Lower Upper

42 100

72 49.4 37.8 56.6

71 46.2 35.6 53.4



#30

Chloroform

Concen: 19.521 ug/l

RT: 7.977 min Scan# 1024

Delta R.T. -0.000 min

Lab File: VN086944.D




Acq: 11 Jun 2025 13:27


Tgt Ion: 83 Resp: 90272

Ion Ratio Lower Upper

83 100

85 64.2 52.9 79.3









#31
Cyclohexane
Concen: 17.603 ug/l
RT: 8.265 min Scan# 1073
Delta R.T. -0.006 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBS02


Tgt Ion: 56 Resp: 79050
Ion Ratio Lower Upper
56 100
69 35.7 25.4 38.0
84 87.1 68.8 103.2



#32
1,1,1-Trichloroethane
Concen: 19.033 ug/l
RT: 8.182 min Scan# 1059
Delta R.T. 0.006 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27

Tgt Ion: 97 Resp: 74861
Ion Ratio Lower Upper
97 100
99 64.2 52.0 78.0
61 47.9 35.9 53.9









#33
1,2-Dichloroethane-d4
Concen: 47.075 ug/l
RT: 8.588 min Scan# 1128
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBS02


Tgt Ion: 65 Resp: 130342
Ion Ratio Lower Upper
65 100
67 54.7 0.0 105.6



#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 9.106 min Scan# 1216
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27

Tgt Ion: 114 Resp: 370524
Ion Ratio Lower Upper
114 100
63 18.7 0.0 39.6
88 16.4 0.0 30.2









#35
Dibromofluoromethane
Concen: 51.096 ug/l
RT: 8.177 min Scan# 1057
Delta R.T. 0.006 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBS02


Tgt Ion:113 Resp: 112198
Ion Ratio Lower Upper
113 100
111 102.8 84.2 126.2
192 17.9 14.2 21.4



#36
1,1-Dichloropropene
Concen: 19.431 ug/l
RT: 8.377 min Scan# 1092
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27

Tgt Ion: 75 Resp: 63577
Ion Ratio Lower Upper
75 100
110 35.1 17.6 52.9
77 30.1 25.3 37.9









#37
Ethyl Acetate
Concen: 19.306 ug/l
RT: 7.571 min Scan# 91
Delta R.T. -0.006 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBS02


Tgt Ion: 43 Resp: 80415
Ion Ratio Lower Upper
43 100
61 15.4 12.1 18.1
70 12.3 9.8 14.6



#38
Carbon Tetrachloride
Concen: 19.222 ug/l
RT: 8.371 min Scan# 1091
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27

Tgt Ion: 117 Resp: 61747
Ion Ratio Lower Upper
117 100
119 97.1 75.0 112.6
121 32.6 24.9 37.3





#39

Methylcyclohexane

Concen: 16.947 ug/l

RT: 9.606 min Scan# 1130

Delta R.T. -0.000 min

Lab File: VN086944.D

Acq: 11 Jun 2025 13:27

Instrument :

MSVOA_N

ClientSampleId :

VN0611WBS02





Tgt Ion: 83 Resp: 75971

Ion Ratio Lower Upper

83 100

55 73.2 61.0 91.6

98 48.9 38.9 58.3



#40

Benzene

Concen: 19.277 ug/l

RT: 8.612 min Scan# 1132

Delta R.T. -0.000 min

Lab File: VN086944.D




Acq: 11 Jun 2025 13:27


Tgt Ion: 78 Resp: 206255

Ion Ratio Lower Upper

78 100

77 23.1 18.7 28.1









#41
Methacrylonitrile
Concen: 18.218 ug/l
RT: 7.788 min Scan# 992
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBS02


Tgt Ion:	41	Resp:	42617
Ion	Ratio	Lower	Upper
41	100		
39	53.5	41.2	61.8
67	85.5	61.4	92.2
52	31.4	22.7	34.1



#42
1,2-Dichloroethane
Concen: 19.530 ug/l
RT: 8.677 min Scan# 1143
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27

Tgt Ion:	62	Resp:	63379
Ion	Ratio	Lower	Upper
62	100		
98	11.4	0.0	21.2









#43
Isopropyl Acetate
Concen: 18.984 ug/l
RT: 8.694 min Scan# 1146
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBS02


Tgt Ion: 43 Resp: 126914
Ion Ratio Lower Upper
43 100
61 28.5 22.1 33.1
87 15.6 11.8 17.6



#44
Trichloroethene
Concen: 19.942 ug/l
RT: 9.359 min Scan# 1259
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27

Tgt Ion:130 Resp: 50593
Ion Ratio Lower Upper
130 100
95 99.6 0.0 196.2





#45

1,2-Dichloropropane

Concen: 19.628 ug/l

RT: 9.624 min Scan# 1304

Delta R.T. -0.006 min

Lab File: VN086944.D

Acq: 11 Jun 2025 13:27

Instrument :

MSVOA_N

ClientSampleId :





VN0611WBS02

Tgt Ion: 63 Resp: 51091

Ion Ratio Lower Upper

63 100

65 34.4 25.5 38.3



#46

Dibromomethane

Concen: 20.093 ug/l

RT: 9.712 min Scan# 1319

Delta R.T. -0.006 min

Lab File: VN086944.D

Acq: 11 Jun 2025 13:27




Tgt Ion: 93 Resp: 34741


Ion Ratio Lower Upper

93 100

95 85.1 67.1 100.7

174 94.0 74.2 111.4





#47

Bromodichloromethane

Concen: 19.634 ug/l

RT: 9.894 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086944.D

Acq: 11 Jun 2025 13:27

Instrument :

MSVOA_N

ClientSampleId :

VN0611WBS02





Tgt Ion: 83 Resp: 69845

Ion Ratio Lower Upper

83 100

85 65.0 50.9 76.3

127 8.1 6.6 10.0



#48

Methyl methacrylate

Concen: 18.771 ug/l

RT: 9.682 min Scan# 1314

Delta R.T. -0.000 min

Lab File: VN086944.D

Acq: 11 Jun 2025 13:27




Tgt Ion: 41 Resp: 57757


Ion Ratio Lower Upper

41 100

69 93.3 74.6 112.0

39 48.5 40.2 60.4









#49
1,4-Dioxane
Concen: 409.347 ug/l
RT: 9.706 min Scan# 11
Delta R.T. 0.006 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBS02


Tgt Ion: 88 Resp: 22914
Ion Ratio Lower Upper
88 100
43 0.0 0.0 0.0
58 69.9 57.7 86.5



#50
Toluene-d8
Concen: 48.706 ug/l
RT: 10.571 min Scan# 1465
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27

Tgt Ion: 98 Resp: 423380
Ion Ratio Lower Upper
98 100
100 67.3 53.4 80.0









#51
4-Methyl-2-Pentanone
Concen: 97.997 ug/l
RT: 10.447 min Scan# 1444
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBS02


Tgt Ion: 43 Resp: 394396
Ion Ratio Lower Upper
43 100
58 42.4 34.2 51.4



#52
Toluene
Concen: 19.509 ug/l
RT: 10.629 min Scan# 1475
Delta R.T. -0.006 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27

Tgt Ion: 92 Resp: 127563
Ion Ratio Lower Upper
92 100
91 171.7 136.5 204.7









#53
 t-1,3-Dichloropropene
 Concen: 19.917 ug/l
 RT: 10.835 min Scan# 1510
 Delta R.T. -0.006 min
 Lab File: VN086944.D
 Acq: 11 Jun 2025 13:27




Instrument :
 MSVOA_N
 ClientSampleId :
 VN0611WBS02


Tgt Ion: 75 Resp: 79224
 Ion Ratio Lower Upper
 75 100
 77 33.3 26.4 39.6



#54
 cis-1,3-Dichloropropene
 Concen: 20.109 ug/l
 RT: 10.318 min Scan# 1422
 Delta R.T. -0.000 min
 Lab File: VN086944.D
 Acq: 11 Jun 2025 13:27

Tgt Ion: 75 Resp: 85586
 Ion Ratio Lower Upper
 75 100
 77 33.6 26.2 39.4
 39 39.5 32.3 48.5





#55

1,1,2-Trichloroethane

Concen: 20.438 ug/l

RT: 11.018 min Scan# 1541

Delta R.T. -0.000 min

Lab File: VN086944.D

Acq: 11 Jun 2025 13:27

Instrument :

MSVOA_N

ClientSampleId :

VN0611WBS02

Tgt Ion: 97 Resp: 51421





Ion Ratio Lower Upper

97 100

83 79.7 69.7 104.5

85 51.4 45.1 67.7

99 61.3 50.6 75.8



#56

Ethyl methacrylate

Concen: 20.198 ug/l

RT: 10.882 min Scan# 1518

Delta R.T. -0.000 min

Lab File: VN086944.D

Acq: 11 Jun 2025 13:27




Tgt Ion: 69 Resp: 80942


Ion Ratio Lower Upper

69 100

41 55.5 47.2 70.8

39 24.3 23.0 34.6





#57

1,3-Dichloropropane

Concen: 19.613 ug/l

RT: 11.165 min Scan# 11

Delta R.T. -0.000 min

Lab File: VN086944.D

Acq: 11 Jun 2025 13:27

Instrument :

MSVOA_N

ClientSampleId :





VN0611WBS02

Tgt Ion: 76 Resp: 85602

Ion Ratio Lower Upper

76 100

78 32.4 25.3 37.9



#58

2-Chloroethyl Vinyl ether

Concen: 101.910 ug/l

RT: 10.165 min Scan# 1396

Delta R.T. -0.000 min

Lab File: VN086944.D




Acq: 11 Jun 2025 13:27


Tgt Ion: 63 Resp: 243584

Ion Ratio Lower Upper

63 100

106 29.2 23.0 34.6









#59
2-Hexanone
Concen: 89.210 ug/l
RT: 11.206 min Scan# 1573
Delta R.T. 0.006 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBS02


Tgt Ion: 43 Resp: 231264
Ion Ratio Lower Upper
43 100
58 65.0 30.4 91.3



#60
Dibromochloromethane
Concen: 20.468 ug/l
RT: 11.359 min Scan# 1599
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27

Tgt Ion: 129 Resp: 53655
Ion Ratio Lower Upper
129 100
127 76.3 38.6 115.8









#61
1,2-Dibromoethane
Concen: 19.711 ug/l
RT: 11.471 min Scan# 1618
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBS02


Tgt Ion:107 Resp: 50832
Ion Ratio Lower Upper
107 100
109 96.3 76.3 114.5



#62
4-Bromofluorobenzene
Concen: 49.951 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27

Tgt Ion: 95 Resp: 161321
Ion Ratio Lower Upper
95 100
174 71.0 0.0 141.8
176 67.2 0.0 132.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 117
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBS02


Tgt Ion:117 Resp: 321188
Ion Ratio Lower Upper
117 100
82 54.7 44.6 67.0
119 31.9 25.5 38.3



#64
Tetrachloroethene
Concen: 19.150 ug/l
RT: 11.106 min Scan# 1556
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27

Tgt Ion:164 Resp: 38926
Ion Ratio Lower Upper
164 100
166 128.0 101.8 152.8
129 106.5 83.0 124.4
131 100.0 78.2 117.2









#65
Chlorobenzene
Concen: 20.006 ug/l
RT: 11.894 min Scan# 11
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBS02


Tgt Ion:112 Resp: 141721
Ion Ratio Lower Upper
112 100
114 31.4 26.2 39.2



#66
1,1,1,2-Tetrachloroethane
Concen: 20.378 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. -0.006 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27

Tgt Ion:131 Resp: 46404
Ion Ratio Lower Upper
131 100
133 94.2 48.3 144.8
119 68.0 34.0 101.8





#67

Ethyl Benzene

Concen: 19.647 ug/l

RT: 11.965 min Scan# 1702

Delta R.T. -0.000 min

Lab File: VN086944.D

Acq: 11 Jun 2025 13:27

Instrument :

MSVOA_N

ClientSampleId :





VN0611WBS02

Tgt Ion: 91 Resp: 239733

Ion Ratio Lower Upper

91 100

106 31.9 25.4 38.2



#68

m/p-Xylenes

Concen: 39.613 ug/l

RT: 12.070 min Scan# 1720

Delta R.T. -0.000 min

Lab File: VN086944.D




Acq: 11 Jun 2025 13:27


Tgt Ion: 106 Resp: 185026

Ion Ratio Lower Upper

106 100

91 197.2 159.4 239.0









#69
o-Xylene
Concen: 20.172 ug/l
RT: 12.400 min Scan# 1776
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBS02


Tgt Ion:106 Resp: 90240
Ion Ratio Lower Upper
106 100
91 210.1 106.1 318.1



#70
Styrene
Concen: 20.137 ug/l
RT: 12.412 min Scan# 1778
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27

Tgt Ion:104 Resp: 154149
Ion Ratio Lower Upper
104 100
78 47.5 38.5 57.7
103 54.3 43.1 64.7









#71
 Bromoform
 Concen: 20.952 ug/l
 RT: 12.582 min Scan# 11
 Delta R.T. -0.000 min
 Lab File: VN086944.D
 Acq: 11 Jun 2025 13:27




Instrument :
 MSVOA_N
 ClientSampleId :
 VN0611WBS02


Tgt Ion:	173	Resp:	35350
Ion Ratio	Lower	Upper	
173	100		
175	51.2	24.1	72.4
254	0.0	0.0	0.0



#72
 1,4-Dichlorobenzene-d4
 Concen: 50.000 ug/l
 RT: 13.788 min Scan# 2012
 Delta R.T. -0.006 min
 Lab File: VN086944.D
 Acq: 11 Jun 2025 13:27

Tgt Ion:	152	Resp:	157403
Ion Ratio	Lower	Upper	
152	100		
115	61.4	30.1	90.5
150	164.6	0.0	345.0









#73
Isopropylbenzene
Concen: 19.320 ug/l
RT: 12.694 min Scan# 11
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBS02


Tgt Ion:105 Resp: 221545
Ion Ratio Lower Upper
105 100
120 27.3 13.5 40.4



#74
N-ethyl acetate
Concen: 15.794 ug/l
RT: 12.529 min Scan# 1798
Delta R.T. 0.017 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27

Tgt Ion: 43 Resp: 63289
Ion Ratio Lower Upper
43 100
70 34.9 37.6 56.4#
55 19.9 19.0 28.6
61 18.0 20.8 31.2#









#75
1,1,2,2-Tetrachloroethane
Concen: 20.973 ug/l
RT: 12.941 min Scan# 1868
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBS02


Tgt Ion: 83 Resp: 81476
Ion Ratio Lower Upper
83 100
131 9.3 4.9 14.7
85 64.3 32.0 96.0



#76
1,2,3-Trichloropropane
Concen: 29.090 ug/l
RT: 12.994 min Scan# 1877
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27

Tgt Ion: 75 Resp: 108842
Ion Ratio Lower Upper
75 100
77 116.8 79.3 237.9









#77
Bromobenzene
Concen: 20.981 ug/l
RT: 12.982 min Scan# 1875
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBS02


Tgt Ion:156 Resp: 55176
Ion Ratio Lower Upper
156 100
77 230.4 117.7 353.1
158 95.7 48.1 144.4



#78
n-propylbenzene
Concen: 19.187 ug/l
RT: 13.035 min Scan# 1884
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27

Tgt Ion: 91 Resp: 267384
Ion Ratio Lower Upper
91 100
120 22.9 11.4 34.2









#79
2-Chlorotoluene
Concen: 19.721 ug/l
RT: 13.123 min Scan# 1899
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBS02


Tgt Ion: 91 Resp: 164813
Ion Ratio Lower Upper
91 100
126 32.6 16.2 48.5



#80
1,3,5-Trimethylbenzene
Concen: 19.200 ug/l
RT: 13.170 min Scan# 1907
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27

Tgt Ion:105 Resp: 181773
Ion Ratio Lower Upper
105 100
120 50.4 24.9 74.6









#81
trans-1,4-Dichloro-2-butene
Concen: 18.532 ug/l
RT: 12.735 min Scan# 1833
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBS02


Tgt Ion: 75 Resp: 30121
Ion Ratio Lower Upper
75 100
53 94.2 66.6 99.8
89 63.9 36.7 55.1#



#82
4-Chlorotoluene
Concen: 19.866 ug/l
RT: 13.217 min Scan# 1915
Delta R.T. -0.006 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27

Tgt Ion: 91 Resp: 167991
Ion Ratio Lower Upper
91 100
126 33.4 16.7 50.1









#83
 tert-Butylbenzene
 Concen: 18.646 ug/l
 RT: 13.435 min Scan# 1952
 Delta R.T. -0.000 min
 Lab File: VN086944.D
 Acq: 11 Jun 2025 13:27




Instrument :
 MSVOA_N
 ClientSampleId :
 VN0611WBS02


Tgt Ion:119 Resp: 161593
 Ion Ratio Lower Upper
 119 100
 91 63.1 32.1 96.5
 134 26.9 13.4 40.1



#84
 1,2,4-Trimethylbenzene
 Concen: 19.216 ug/l
 RT: 13.482 min Scan# 1960
 Delta R.T. -0.000 min
 Lab File: VN086944.D
 Acq: 11 Jun 2025 13:27

Tgt Ion:105 Resp: 182431
 Ion Ratio Lower Upper
 105 100
 120 47.6 23.2 69.6









#85
 sec-Butylbenzene
 Concen: 18.189 ug/l
 RT: 13.617 min Scan# 1983
 Delta R.T. -0.000 min
 Lab File: VN086944.D
 Acq: 11 Jun 2025 13:27




Instrument :
 MSVOA_N
 ClientSampleId :
 VN0611WBS02


Tgt Ion:105 Resp: 228907
 Ion Ratio Lower Upper
 105 100
 134 19.9 10.0 30.0



#86
 p-Isopropyltoluene
 Concen: 18.551 ug/l
 RT: 13.729 min Scan# 2002
 Delta R.T. -0.000 min
 Lab File: VN086944.D
 Acq: 11 Jun 2025 13:27

Tgt Ion:119 Resp: 192991
 Ion Ratio Lower Upper
 119 100
 134 26.2 13.5 40.4
 91 22.1 11.5 34.4





#87

1,3-Dichlorobenzene

Concen: 20.271 ug/l

RT: 13.735 min Scan# 2003

Delta R.T. -0.000 min

Lab File: VN086944.D

Acq: 11 Jun 2025 13:27

Instrument :

MSVOA_N

ClientSampleId :

VN0611WBS02





Tgt Ion:146 Resp: 104883

Ion Ratio Lower Upper

146 100

111 41.3 20.9 62.8

148 62.6 32.0 96.2



#88

1,4-Dichlorobenzene

Concen: 20.426 ug/l

RT: 13.812 min Scan# 2016

Delta R.T. -0.000 min

Lab File: VN086944.D

Acq: 11 Jun 2025 13:27




Tgt Ion:146 Resp: 107750


Ion Ratio Lower Upper

146 100

111 40.7 19.9 59.7

148 63.1 31.8 95.4









#89
n-Butylbenzene
Concen: 17.613 ug/l
RT: 14.053 min Scan# 2057
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBS02


Tgt Ion: 91 Resp: 177482
Ion Ratio Lower Upper
91 100
92 54.9 27.5 82.4
134 26.7 13.0 39.0



#90
Hexachloroethane
Concen: 18.972 ug/l
RT: 14.329 min Scan# 2104
Delta R.T. -0.006 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27

Tgt Ion: 117 Resp: 33456
Ion Ratio Lower Upper
117 100
201 67.1 33.2 99.6









#91
1,2-Dichlorobenzene
Concen: 19.995 ug/l
RT: 14.106 min Scan# 2066
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBS02


Tgt Ion:146 Resp: 99394
Ion Ratio Lower Upper
146 100
111 44.5 21.6 64.6
148 64.5 31.7 95.1



#92
1,2-Dibromo-3-Chloropropane
Concen: 19.155 ug/l
RT: 14.717 min Scan# 2170
Delta R.T. -0.000 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27

Tgt Ion: 75 Resp: 17796
Ion Ratio Lower Upper
75 100
155 85.2 41.1 123.4
157 106.8 52.7 158.1





#93

1,2,4-Trichlorobenzene

Concen: 18.433 ug/l

RT: 15.394 min Scan# 21

Delta R.T. 0.006 min

Lab File: VN086944.D

Acq: 11 Jun 2025 13:27

Instrument :

MSVOA_N

ClientSampleId :

VN0611WBS02





Tgt Ion:180 Resp: 58509

Ion Ratio Lower Upper

180 100

182 96.7 46.9 140.7

145 32.3 16.0 47.9



#94

Hexachlorobutadiene

Concen: 16.567 ug/l

RT: 15.500 min Scan# 2303

Delta R.T. -0.000 min

Lab File: VN086944.D

Acq: 11 Jun 2025 13:27




Tgt Ion:225 Resp: 19592


Ion Ratio Lower Upper

225 100

223 64.0 33.5 100.4

227 60.0 31.1 93.5









#95
Naphthalene
Concen: 19.700 ug/l
RT: 15.635 min Scan# 2327
Delta R.T. -0.006 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBS02

Tgt Ion:128 Resp: 232760
Ion Ratio Lower Upper
128 100
127 12.6 10.2 15.2
129 10.9 8.8 13.2



#96
1,2,3-Trichlorobenzene
Concen: 17.802 ug/l
RT: 15.841 min Scan# 2361
Delta R.T. 0.006 min
Lab File: VN086944.D
Acq: 11 Jun 2025 13:27

Tgt Ion:180 Resp: 56144
Ion Ratio Lower Upper
180 100
182 96.0 47.0 141.0
145 33.6 16.6 49.8



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061125\
 Data File : VN086945.D
 Acq On : 11 Jun 2025 14:01
 Operator : JC\MD
 Sample : VN0611WBSD02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0611WBSD02

Quant Time: Jun 12 01:32:16 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	199904	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	354069	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	310598	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	151691	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	128166	47.886	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	95.780%
35) Dibromofluoromethane	8.177	113	109657	52.260	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.520%
50) Toluene-d8	10.571	98	411227	49.506	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.020%
62) 4-Bromofluorobenzene	12.847	95	154682	50.121	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	100.240%

Target Compounds				Qvalue		
2) Dichlorodifluoromethane	2.153	85	38145	19.138	ug/l	96
3) Chloromethane	2.400	50	39948	15.521	ug/l	99
4) Vinyl Chloride	2.553	62	49695	18.717	ug/l	97
5) Bromomethane	3.000	94	24041	16.181	ug/l	95
6) Chloroethane	3.153	64	32368	18.874	ug/l	100
7) Trichlorofluoromethane	3.530	101	66071	19.047	ug/l	96
8) Diethyl Ether	3.977	74	29872	19.765	ug/l	99
9) 1,1,2-Trichlorotrifluo...	4.394	101	40831	18.735	ug/l	97
10) Methyl Iodide	4.612	142	36500	12.919	ug/l	100
11) Tert butyl alcohol	5.541	59	71180	98.012	ug/l	100
12) 1,1-Dichloroethene	4.365	96	42515	19.108	ug/l	92
13) Acrolein	4.200	56	22554	98.112	ug/l	100
14) Allyl chloride	5.053	41	65845	17.844	ug/l	98
15) Acrylonitrile	5.736	53	167429	98.634	ug/l	100
16) Acetone	4.447	43	131189	92.428	ug/l	97
17) Carbon Disulfide	4.736	76	105790	17.189	ug/l	98
18) Methyl Acetate	5.047	43	83025	20.072	ug/l	99
19) Methyl tert-butyl Ether	5.818	73	162549	20.177	ug/l	99
20) Methylene Chloride	5.300	84	50840	19.132	ug/l	96
21) trans-1,2-Dichloroethene	5.806	96	46018	18.589	ug/l	99
22) Diisopropyl ether	6.688	45	152785	19.643	ug/l	98
23) Vinyl Acetate	6.618	43	656574	99.908	ug/l	98
24) 1,1-Dichloroethane	6.588	63	86331	19.287	ug/l	99
25) 2-Butanone	7.494	43	218999	94.923	ug/l	99
26) 2,2-Dichloropropane	7.506	77	74316	21.343	ug/l	98
27) cis-1,2-Dichloroethene	7.500	96	57947	19.573	ug/l	97
28) Bromochloromethane	7.824	49	46411	21.088	ug/l	97
29) Tetrahydrofuran	7.853	42	145366	96.721	ug/l	98
30) Chloroform	7.977	83	86604	19.374	ug/l	100
31) Cyclohexane	8.265	56	74510	17.164	ug/l	98
32) 1,1,1-Trichloroethane	8.177	97	71441	18.790	ug/l	98
36) 1,1-Dichloropropene	8.382	75	60912	19.481	ug/l	99
37) Ethyl Acetate	7.571	43	80010	20.102	ug/l	99
38) Carbon Tetrachloride	8.371	117	59839	19.494	ug/l	96
39) Methylcyclohexane	9.606	83	71058	16.588	ug/l	97
40) Benzene	8.612	78	200940	19.653	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061125\
 Data File : VN086945.D
 Acq On : 11 Jun 2025 14:01
 Operator : JC\MD
 Sample : VN0611WBSD02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0611WBSD02

Quant Time: Jun 12 01:32:16 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.788	41	46047	20.600	ug/l	98
42) 1,2-Dichloroethane	8.677	62	63066	20.336	ug/l	100
43) Isopropyl Acetate	8.694	43	127320	19.930	ug/l	99
44) Trichloroethene	9.359	130	48830	20.141	ug/l	97
45) 1,2-Dichloropropane	9.624	63	49815	20.027	ug/l	95
46) Dibromomethane	9.712	93	35176	21.290	ug/l	99
47) Bromodichloromethane	9.894	83	70222	20.658	ug/l	99
48) Methyl methacrylate	9.688	41	56099	19.079	ug/l	95
49) 1,4-Dioxane	9.700	88	24300	454.282	ug/l #	93
51) 4-Methyl-2-Pentanone	10.447	43	399735	103.940	ug/l	99
52) Toluene	10.629	92	124298	19.893	ug/l	100
53) t-1,3-Dichloropropene	10.841	75	78850	20.744	ug/l	100
54) cis-1,3-Dichloropropene	10.318	75	84225	20.709	ug/l	97
55) 1,1,2-Trichloroethane	11.018	97	50307	20.924	ug/l	97
56) Ethyl methacrylate	10.882	69	79449	20.747	ug/l	97
57) 1,3-Dichloropropane	11.165	76	85222	20.434	ug/l	98
58) 2-Chloroethyl Vinyl ether	10.165	63	241856	105.889	ug/l	99
59) 2-Hexanone	11.206	43	232463	93.840	ug/l	95
60) Dibromochloromethane	11.359	129	53154	21.220	ug/l	100
61) 1,2-Dibromoethane	11.470	107	51056	20.718	ug/l	99
64) Tetrachloroethene	11.106	164	36363	18.499	ug/l	98
65) Chlorobenzene	11.894	112	138224	20.178	ug/l	99
66) 1,1,1,2-Tetrachloroethane	11.965	131	45014	20.442	ug/l	99
67) Ethyl Benzene	11.965	91	227138	19.250	ug/l	100
68) m/p-Xylenes	12.070	106	178421	39.501	ug/l	98
69) o-Xylene	12.400	106	86537	20.004	ug/l	100
70) Styrene	12.412	104	152236	20.565	ug/l	98
71) Bromoform	12.582	173	36528	22.388	ug/l #	98
73) Isopropylbenzene	12.694	105	213019	19.276	ug/l	100
74) N-amyl acetate	12.529	43	69705	18.051	ug/l #	88
75) 1,1,2,2-Tetrachloroethane	12.935	83	81389	21.740	ug/l	100
76) 1,2,3-Trichloropropane	12.994	75	108666	30.137	ug/l	67
77) Bromobenzene	12.976	156	54030	21.319	ug/l	98
78) n-propylbenzene	13.035	91	257426	19.168	ug/l	99
79) 2-Chlorotoluene	13.123	91	157627	19.571	ug/l	99
80) 1,3,5-Trimethylbenzene	13.170	105	175428	19.228	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.735	75	29703	18.963	ug/l	95
82) 4-Chlorotoluene	13.223	91	162421	19.931	ug/l	100
83) tert-Butylbenzene	13.435	119	156588	18.749	ug/l	98
84) 1,2,4-Trimethylbenzene	13.482	105	177241	19.373	ug/l	99
85) sec-Butylbenzene	13.611	105	219639	18.109	ug/l	99
86) p-Isopropyltoluene	13.729	119	182853	18.238	ug/l	99
87) 1,3-Dichlorobenzene	13.729	146	102302	20.517	ug/l	100
88) 1,4-Dichlorobenzene	13.811	146	103614	20.382	ug/l	98
89) n-Butylbenzene	14.053	91	171576	17.668	ug/l	99
90) Hexachloroethane	14.329	117	31966	18.810	ug/l	100
91) 1,2-Dichlorobenzene	14.106	146	98864	20.638	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.723	75	18610	20.786	ug/l	99
93) 1,2,4-Trichlorobenzene	15.388	180	57083	18.660	ug/l	99
94) Hexachlorobutadiene	15.500	225	17672	15.506	ug/l	98
95) Naphthalene	15.635	128	230927	20.281	ug/l	99
96) 1,2,3-Trichlorobenzene	15.835	180	55026	18.105	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061125\
Data File : VN086945.D
Acq On : 11 Jun 2025 14:01
Operator : JC\MD
Sample : VN0611WBSD02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0611WBSD02

Quant Time: Jun 12 01:32:16 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 02:12:50 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061125\
 Data File : VN086945.D
 Acq On : 11 Jun 2025 14:01
 Operator : JC\MD
 Sample : VN0611WBSD02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0611WBSD02

Quant Time: Jun 12 01:32:16 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.229 min Scan# 1
Delta R.T. -0.007 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBSD02


Tgt Ion:168 Resp: 199904
Ion Ratio Lower Upper
168 100
99 58.1 49.1 73.7



#2
Dichlorodifluoromethane
Concen: 19.138 ug/l
RT: 2.153 min Scan# 34
Delta R.T. -0.000 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01

Tgt Ion: 85 Resp: 38145
Ion Ratio Lower Upper
85 100
87 30.2 16.2 48.6









#3
Chloromethane
Concen: 15.521 ug/l
RT: 2.400 min Scan# 75
Delta R.T. 0.006 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBSD02


Tgt Ion: 50 Resp: 39948
Ion Ratio Lower Upper
50 100
52 32.7 25.7 38.5



#4
Vinyl Chloride
Concen: 18.717 ug/l
RT: 2.553 min Scan# 102
Delta R.T. -0.000 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01

Tgt Ion: 62 Resp: 49695
Ion Ratio Lower Upper
62 100
64 30.8 26.0 39.0









#5
Bromomethane
Concen: 16.181 ug/l
RT: 3.000 min Scan# 11
Delta R.T. 0.006 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBSD02


Tgt Ion: 94 Resp: 24041
Ion Ratio Lower Upper
94 100
96 92.1 77.9 116.9



#6
Chloroethane
Concen: 18.874 ug/l
RT: 3.153 min Scan# 204
Delta R.T. -0.000 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01

Tgt Ion: 64 Resp: 32368
Ion Ratio Lower Upper
64 100
66 30.6 24.6 36.8









#7
Trichlorofluoromethane
Concen: 19.047 ug/l
RT: 3.530 min Scan# 20
Delta R.T. 0.006 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBSD02


Tgt Ion:101 Resp: 66071
Ion Ratio Lower Upper
101 100
103 62.1 52.2 78.2



#8
Diethyl Ether
Concen: 19.765 ug/l
RT: 3.977 min Scan# 344
Delta R.T. -0.006 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01

Tgt Ion: 74 Resp: 29872
Ion Ratio Lower Upper
74 100
45 90.0 45.5 136.5









#9
1,1,2-Trichlorotrifluoroethane
Concen: 18.735 ug/l
RT: 4.394 min Scan# 415
Delta R.T. -0.000 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBSD02


Tgt Ion:101 Resp: 40831
Ion Ratio Lower Upper
101 100
85 39.9 33.8 50.8
151 75.2 58.7 88.1



#10
Methyl Iodide
Concen: 12.919 ug/l
RT: 4.612 min Scan# 452
Delta R.T. 0.006 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01

Tgt Ion:142 Resp: 36500
Ion Ratio Lower Upper
142 100
127 38.4 30.9 46.3
141 13.1 10.4 15.6









#11
Tert butyl alcohol
Concen: 98.012 ug/l
RT: 5.541 min Scan# 611
Delta R.T. -0.006 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBSD02


Tgt Ion: 59 Resp: 71180
Ion Ratio Lower Upper
59 100
57 10.2 8.1 12.1



#12
1,1-Dichloroethene
Concen: 19.108 ug/l
RT: 4.365 min Scan# 410
Delta R.T. 0.006 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01

Tgt Ion: 96 Resp: 42515
Ion Ratio Lower Upper
96 100
61 144.2 125.8 188.6
98 62.9 52.5 78.7





#13

Acrolein

Concen: 98.112 ug/l

RT: 4.200 min Scan# 382

Delta R.T. -0.000 min

Lab File: VN086945.D

Acq: 11 Jun 2025 14:01

Instrument :

MSVOA_N

ClientSampleId :





VN0611WBSD02

Tgt Ion: 56 Resp: 22554

Ion Ratio Lower Upper

56 100

55 69.3 55.6 83.4



#14

Allyl chloride

Concen: 17.844 ug/l

RT: 5.053 min Scan# 527

Delta R.T. 0.012 min

Lab File: VN086945.D

Acq: 11 Jun 2025 14:01




Tgt Ion: 41 Resp: 65845


Ion Ratio Lower Upper

41 100

39 64.2 51.3 76.9

76 42.0 31.4 47.0









#15
Acrylonitrile
Concen: 98.634 ug/l
RT: 5.736 min Scan# 644
Delta R.T. -0.006 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBSD02


Tgt Ion: 53 Resp: 167429
Ion Ratio Lower Upper
53 100
52 81.7 65.0 97.6
51 33.9 27.2 40.8



#16
Acetone
Concen: 92.428 ug/l
RT: 4.447 min Scan# 424
Delta R.T. -0.000 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01

Tgt Ion: 43 Resp: 131189
Ion Ratio Lower Upper
43 100
58 36.9 28.0 42.0









#17
Carbon Disulfide
Concen: 17.189 ug/l
RT: 4.736 min Scan# 473
Delta R.T. 0.006 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBSD02


Tgt Ion: 76 Resp: 105790
Ion Ratio Lower Upper
76 100
78 8.4 7.3 10.9



#18
Methyl Acetate
Concen: 20.072 ug/l
RT: 5.047 min Scan# 526
Delta R.T. 0.006 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01

Tgt Ion: 43 Resp: 83025
Ion Ratio Lower Upper
43 100
74 26.6 20.7 31.1









#19
Methyl tert-butyl Ether
Concen: 20.177 ug/l
RT: 5.818 min Scan# 61
Delta R.T. 0.006 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBSD02


Tgt Ion: 73 Resp: 162549
Ion Ratio Lower Upper
73 100
57 22.2 17.4 26.2



#20
Methylene Chloride
Concen: 19.132 ug/l
RT: 5.300 min Scan# 569
Delta R.T. 0.006 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01

Tgt Ion: 84 Resp: 50840
Ion Ratio Lower Upper
84 100
49 110.8 90.5 135.7
51 34.5 28.5 42.7
86 66.6 48.1 72.1









#21
trans-1,2-Dichloroethene
Concen: 18.589 ug/l
RT: 5.806 min Scan# 61
Delta R.T. -0.000 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBSD02


Tgt Ion: 96 Resp: 46018
Ion Ratio Lower Upper
96 100
61 136.9 109.2 163.8
98 65.3 49.8 74.8



#22
Diisopropyl ether
Concen: 19.643 ug/l
RT: 6.688 min Scan# 805
Delta R.T. -0.000 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01

Tgt Ion: 45 Resp: 152785
Ion Ratio Lower Upper
45 100
43 53.3 41.9 62.9
87 30.6 22.4 33.6
59 12.1 9.8 14.6









#23
 Vinyl Acetate
 Concen: 99.908 ug/l
 RT: 6.618 min Scan# 793
 Delta R.T. -0.000 min
 Lab File: VN086945.D
 Acq: 11 Jun 2025 14:01




Instrument :
 MSVOA_N
 ClientSampleId :
 VN0611WBSD02


Tgt Ion: 43 Resp: 656574
 Ion Ratio Lower Upper
 43 100
 86 12.5 9.5 14.3



#24
 1,1-Dichloroethane
 Concen: 19.287 ug/l
 RT: 6.588 min Scan# 788
 Delta R.T. 0.006 min
 Lab File: VN086945.D
 Acq: 11 Jun 2025 14:01

Tgt Ion: 63 Resp: 86331
 Ion Ratio Lower Upper
 63 100
 98 7.6 3.6 10.8
 100 4.8 2.6 7.8









#25
2-Butanone
Concen: 94.923 ug/l
RT: 7.494 min Scan# 94
Delta R.T. -0.000 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBSD02


Tgt Ion: 43 Resp: 218999
Ion Ratio Lower Upper
43 100
72 29.1 23.0 34.6



#26
2,2-Dichloropropane
Concen: 21.343 ug/l
RT: 7.506 min Scan# 944
Delta R.T. 0.006 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01

Tgt Ion: 77 Resp: 74316
Ion Ratio Lower Upper
77 100
97 24.0 11.6 34.7









#27
 cis-1,2-Dichloroethene
 Concen: 19.573 ug/l
 RT: 7.500 min Scan# 943
 Delta R.T. -0.000 min
 Lab File: VN086945.D
 Acq: 11 Jun 2025 14:01




Instrument :
 MSVOA_N
 ClientSampleId :
 VN0611WBSD02


Tgt Ion: 96 Resp: 57947
 Ion Ratio Lower Upper
 96 100
 61 135.2 0.0 278.0
 98 65.0 0.0 128.2



#28
 Bromochloromethane
 Concen: 21.088 ug/l
 RT: 7.824 min Scan# 998
 Delta R.T. -0.000 min
 Lab File: VN086945.D
 Acq: 11 Jun 2025 14:01

Tgt Ion: 49 Resp: 46411
 Ion Ratio Lower Upper
 49 100
 129 1.7 0.0 3.8
 130 80.4 62.2 93.2





#29

Tetrahydrofuran

Concen: 96.721 ug/l

RT: 7.853 min Scan# 1003

Delta R.T. -0.000 min

Lab File: VN086945.D

Acq: 11 Jun 2025 14:01

Instrument :

MSVOA_N

ClientSampleId :

VN0611WBSD02





Tgt Ion: 42 Resp: 145366

Ion Ratio Lower Upper

42 100

72 48.9 37.8 56.6

71 46.0 35.6 53.4



#30

Chloroform

Concen: 19.374 ug/l

RT: 7.977 min Scan# 1024

Delta R.T. -0.000 min

Lab File: VN086945.D




Acq: 11 Jun 2025 14:01


Tgt Ion: 83 Resp: 86604

Ion Ratio Lower Upper

83 100

85 65.8 52.9 79.3









#31
Cyclohexane
Concen: 17.164 ug/l
RT: 8.265 min Scan# 1073
Delta R.T. -0.006 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBSD02


Tgt Ion: 56 Resp: 74510
Ion Ratio Lower Upper
56 100
69 32.4 25.4 38.0
84 88.3 68.8 103.2



#32
1,1,1-Trichloroethane
Concen: 18.790 ug/l
RT: 8.177 min Scan# 1058
Delta R.T. -0.000 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01

Tgt Ion: 97 Resp: 71441
Ion Ratio Lower Upper
97 100
99 63.6 52.0 78.0
61 42.8 35.9 53.9





#33

1,2-Dichloroethane-d4

Concen: 47.886 ug/l

RT: 8.588 min Scan# 1128

Delta R.T. -0.000 min

Lab File: VN086945.D

Acq: 11 Jun 2025 14:01

Instrument :

MSVOA_N

ClientSampleId :





VN0611WBSD02

Tgt Ion: 65 Resp: 128166

Ion Ratio Lower Upper

65 100

67 54.8 0.0 105.6



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.106 min Scan# 1216

Delta R.T. -0.000 min

Lab File: VN086945.D

Acq: 11 Jun 2025 14:01




Tgt Ion: 114 Resp: 354069


Ion Ratio Lower Upper

114 100

63 19.8 0.0 39.6

88 16.1 0.0 30.2









#35
Dibromofluoromethane
Concen: 52.260 ug/l
RT: 8.177 min Scan# 1057
Delta R.T. 0.006 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBSD02


Tgt Ion:113 Resp: 109657
Ion Ratio Lower Upper
113 100
111 101.1 84.2 126.2
192 17.6 14.2 21.4



#36
1,1-Dichloropropene
Concen: 19.481 ug/l
RT: 8.382 min Scan# 1093
Delta R.T. 0.006 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01

Tgt Ion: 75 Resp: 60912
Ion Ratio Lower Upper
75 100
110 36.1 17.6 52.9
77 31.6 25.3 37.9









#37
Ethyl Acetate
Concen: 20.102 ug/l
RT: 7.571 min Scan# 91
Delta R.T. -0.006 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBSD02


Tgt Ion: 43 Resp: 80010
Ion Ratio Lower Upper
43 100
61 15.9 12.1 18.1
70 12.4 9.8 14.6



#38
Carbon Tetrachloride
Concen: 19.494 ug/l
RT: 8.371 min Scan# 1091
Delta R.T. -0.000 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01

Tgt Ion: 117 Resp: 59839
Ion Ratio Lower Upper
117 100
119 98.4 75.0 112.6
121 32.4 24.9 37.3





#39

Methylcyclohexane

Concen: 16.588 ug/l

RT: 9.606 min Scan# 1130

Delta R.T. -0.000 min

Lab File: VN086945.D

Acq: 11 Jun 2025 14:01

Instrument :

MSVOA_N

ClientSampleId :

VN0611WBSD02





Tgt Ion: 83 Resp: 71058

Ion Ratio Lower Upper

83 100

55 72.5 61.0 91.6

98 47.5 38.9 58.3



#40

Benzene

Concen: 19.653 ug/l

RT: 8.612 min Scan# 1132

Delta R.T. -0.000 min

Lab File: VN086945.D




Acq: 11 Jun 2025 14:01


Tgt Ion: 78 Resp: 200940

Ion Ratio Lower Upper

78 100

77 22.6 18.7 28.1









#41
Methacrylonitrile
Concen: 20.600 ug/l
RT: 7.788 min Scan# 992
Delta R.T. -0.000 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBSD02


Tgt Ion:	41	Resp:	46047
Ion	Ratio	Lower	Upper
41	100		
39	50.5	41.2	61.8
67	78.7	61.4	92.2
52	28.7	22.7	34.1



#42
1,2-Dichloroethane
Concen: 20.336 ug/l
RT: 8.677 min Scan# 1143
Delta R.T. -0.000 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01

Tgt Ion:	62	Resp:	63066
Ion	Ratio	Lower	Upper
62	100		
98	10.5	0.0	21.2









#43
Isopropyl Acetate
Concen: 19.930 ug/l
RT: 8.694 min Scan# 1146
Delta R.T. -0.000 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBSD02


Tgt Ion: 43 Resp: 127320
Ion Ratio Lower Upper
43 100
61 28.5 22.1 33.1
87 15.1 11.8 17.6



#44
Trichloroethene
Concen: 20.141 ug/l
RT: 9.359 min Scan# 1259
Delta R.T. -0.000 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01

Tgt Ion:130 Resp: 48830
Ion Ratio Lower Upper
130 100
95 94.9 0.0 196.2





#45

1,2-Dichloropropane

Concen: 20.027 ug/l

RT: 9.624 min Scan# 11

Delta R.T. -0.006 min

Lab File: VN086945.D

Acq: 11 Jun 2025 14:01

Instrument :

MSVOA_N

ClientSampleId :





VN0611WBSD02

Tgt Ion: 63 Resp: 49815

Ion Ratio Lower Upper

63 100

65 29.1 25.5 38.3



#46

Dibromomethane

Concen: 21.290 ug/l

RT: 9.712 min Scan# 1319

Delta R.T. -0.006 min

Lab File: VN086945.D

Acq: 11 Jun 2025 14:01




Tgt Ion: 93 Resp: 35176


Ion Ratio Lower Upper

93 100

95 83.2 67.1 100.7

174 92.2 74.2 111.4









#47
Bromodichloromethane
Concen: 20.658 ug/l
RT: 9.894 min Scan# 1350
Delta R.T. -0.000 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBSD02


Tgt Ion: 83 Resp: 70222
Ion Ratio Lower Upper
83 100
85 62.7 50.9 76.3
127 8.5 6.6 10.0



#48
Methyl methacrylate
Concen: 19.079 ug/l
RT: 9.688 min Scan# 1315
Delta R.T. 0.006 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01

Tgt Ion: 41 Resp: 56099
Ion Ratio Lower Upper
41 100
69 99.4 74.6 112.0
39 51.4 40.2 60.4





#49

1,4-Dioxane

Concen: 454.282 ug/l

RT: 9.700 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086945.D

Acq: 11 Jun 2025 14:01

Instrument :

MSVOA_N

ClientSampleId :

VN0611WBSD02





Tgt Ion: 88 Resp: 24300

Ion Ratio Lower Upper

88 100

43 22.4 0.0 0.0#

58 66.1 57.7 86.5



#50

Toluene-d8

Concen: 49.506 ug/l

RT: 10.571 min Scan# 1465

Delta R.T. -0.000 min

Lab File: VN086945.D




Acq: 11 Jun 2025 14:01


Tgt Ion: 98 Resp: 411227

Ion Ratio Lower Upper

98 100

100 66.6 53.4 80.0









#51
4-Methyl-2-Pentanone
Concen: 103.940 ug/l
RT: 10.447 min Scan# 1444
Delta R.T. -0.000 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBSD02


Tgt Ion: 43 Resp: 399735
Ion Ratio Lower Upper
43 100
58 42.4 34.2 51.4



#52
Toluene
Concen: 19.893 ug/l
RT: 10.629 min Scan# 1475
Delta R.T. -0.006 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01

Tgt Ion: 92 Resp: 124298
Ion Ratio Lower Upper
92 100
91 170.3 136.5 204.7









#53
 t-1,3-Dichloropropene
 Concen: 20.744 ug/l
 RT: 10.841 min Scan# 1511
 Delta R.T. -0.000 min
 Lab File: VN086945.D
 Acq: 11 Jun 2025 14:01




Instrument :
 MSVOA_N
 ClientSampleId :
 VN0611WBSD02


Tgt Ion: 75 Resp: 78850
 Ion Ratio Lower Upper
 75 100
 77 32.9 26.4 39.6



#54
 cis-1,3-Dichloropropene
 Concen: 20.709 ug/l
 RT: 10.318 min Scan# 1422
 Delta R.T. -0.000 min
 Lab File: VN086945.D
 Acq: 11 Jun 2025 14:01

Tgt Ion: 75 Resp: 84225
 Ion Ratio Lower Upper
 75 100
 77 30.9 26.2 39.4
 39 39.0 32.3 48.5









#55
1,1,2-Trichloroethane
Concen: 20.924 ug/l
RT: 11.018 min Scan# 1541
Delta R.T. -0.000 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBSD02


Tgt Ion:	97	Resp:	50307
Ion	Ratio	Lower	Upper
97	100		
83	88.4	69.7	104.5
85	52.7	45.1	67.7
99	64.8	50.6	75.8



#56
Ethyl methacrylate
Concen: 20.747 ug/l
RT: 10.882 min Scan# 1518
Delta R.T. -0.000 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01

Tgt Ion:	69	Resp:	79449
Ion	Ratio	Lower	Upper
69	100		
41	58.8	47.2	70.8
39	25.0	23.0	34.6









#57
1,3-Dichloropropane
Concen: 20.434 ug/l
RT: 11.165 min Scan# 1566
Delta R.T. -0.000 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBSD02


Tgt Ion: 76 Resp: 85222
Ion Ratio Lower Upper
76 100
78 32.6 25.3 37.9



#58
2-Chloroethyl Vinyl ether
Concen: 105.889 ug/l
RT: 10.165 min Scan# 1396
Delta R.T. -0.000 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01

Tgt Ion: 63 Resp: 241856
Ion Ratio Lower Upper
63 100
106 29.1 23.0 34.6









#59
2-Hexanone
Concen: 93.840 ug/l
RT: 11.206 min Scan# 11
Delta R.T. 0.006 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBSD02


Tgt Ion: 43 Resp: 232463
Ion Ratio Lower Upper
43 100
58 64.8 30.4 91.3



#60
Dibromochloromethane
Concen: 21.220 ug/l
RT: 11.359 min Scan# 1599
Delta R.T. -0.000 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01

Tgt Ion: 129 Resp: 53154
Ion Ratio Lower Upper
129 100
127 77.2 38.6 115.8









#61
1,2-Dibromoethane
Concen: 20.718 ug/l
RT: 11.470 min Scan# 1618
Delta R.T. -0.000 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBSD02


Tgt Ion:107 Resp: 51056
Ion Ratio Lower Upper
107 100
109 94.8 76.3 114.5



#62
4-Bromofluorobenzene
Concen: 50.121 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. -0.000 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01

Tgt Ion: 95 Resp: 154682
Ion Ratio Lower Upper
95 100
174 72.6 0.0 141.8
176 68.6 0.0 132.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 117
Delta R.T. -0.000 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBSD02


Tgt Ion:117 Resp: 310598
Ion Ratio Lower Upper
117 100
82 54.9 44.6 67.0
119 32.6 25.5 38.3



#64
Tetrachloroethene
Concen: 18.499 ug/l
RT: 11.106 min Scan# 1556
Delta R.T. -0.000 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01

Tgt Ion:164 Resp: 36363
Ion Ratio Lower Upper
164 100
166 125.7 101.8 152.8
129 101.3 83.0 124.4
131 93.9 78.2 117.2









#65
Chlorobenzene
Concen: 20.178 ug/l
RT: 11.894 min Scan# 1119
Delta R.T. -0.000 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBSD02


Tgt Ion:112 Resp: 138224
Ion Ratio Lower Upper
112 100
114 33.4 26.2 39.2



#66
1,1,1,2-Tetrachloroethane
Concen: 20.442 ug/l
RT: 11.965 min Scan# 1702
Delta R.T. -0.000 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01

Tgt Ion:131 Resp: 45014
Ion Ratio Lower Upper
131 100
133 96.4 48.3 144.8
119 70.3 34.0 101.8









#67
Ethyl Benzene
Concen: 19.250 ug/l
RT: 11.965 min Scan# 1702
Delta R.T. -0.000 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBSD02


Tgt Ion: 91 Resp: 227138
Ion Ratio Lower Upper
91 100
106 31.9 25.4 38.2



#68
m/p-Xylenes
Concen: 39.501 ug/l
RT: 12.070 min Scan# 1720
Delta R.T. -0.000 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01

Tgt Ion: 106 Resp: 178421
Ion Ratio Lower Upper
106 100
91 196.9 159.4 239.0









#69
o-Xylene
Concen: 20.004 ug/l
RT: 12.400 min Scan# 1776
Delta R.T. -0.000 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBSD02


Tgt Ion:106 Resp: 86537
Ion Ratio Lower Upper
106 100
91 212.7 106.1 318.1



#70
Styrene
Concen: 20.565 ug/l
RT: 12.412 min Scan# 1778
Delta R.T. -0.000 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01

Tgt Ion:104 Resp: 152236
Ion Ratio Lower Upper
104 100
78 46.8 38.5 57.7
103 53.0 43.1 64.7





#71

Bromoform

Concen: 22.388 ug/l

RT: 12.582 min Scan# 11

Delta R.T. -0.000 min

Lab File: VN086945.D

Acq: 11 Jun 2025 14:01

Instrument :

MSVOA_N

ClientSampleId :

VN0611WBSD02





Tgt Ion:173 Resp: 36528

Ion Ratio Lower Upper

173 100

175 46.8 24.1 72.4

254 0.0 0.0 0.0



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.788 min Scan# 2012

Delta R.T. -0.006 min

Lab File: VN086945.D

Acq: 11 Jun 2025 14:01




Tgt Ion:152 Resp: 151691


Ion Ratio Lower Upper

152 100

115 61.0 30.1 90.5

150 166.9 0.0 345.0









#73
Isopropylbenzene
Concen: 19.276 ug/l
RT: 12.694 min Scan# 11
Delta R.T. -0.000 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBSD02


Tgt Ion:105 Resp: 213019
Ion Ratio Lower Upper
105 100
120 26.8 13.5 40.4



#74
N-ethyl acetate
Concen: 18.051 ug/l
RT: 12.529 min Scan# 1798
Delta R.T. 0.017 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01

Tgt Ion: 43 Resp: 69705
Ion Ratio Lower Upper
43 100
70 38.7 37.6 56.4
55 18.5 19.0 28.6#
61 19.5 20.8 31.2#









#75
 1,1,2-Tetrachloroethane
 Concen: 21.740 ug/l
 RT: 12.935 min Scan# 1867
 Delta R.T. -0.006 min
 Lab File: VN086945.D
 Acq: 11 Jun 2025 14:01




Instrument :
 MSVOA_N
 ClientSampleId :
 VN0611WBSD02


Tgt Ion: 83 Resp: 81389
 Ion Ratio Lower Upper
 83 100
 131 9.7 4.9 14.7
 85 63.6 32.0 96.0



#76
 1,2,3-Trichloropropane
 Concen: 30.137 ug/l
 RT: 12.994 min Scan# 1877
 Delta R.T. -0.000 min
 Lab File: VN086945.D
 Acq: 11 Jun 2025 14:01

Tgt Ion: 75 Resp: 108666
 Ion Ratio Lower Upper
 75 100
 77 114.7 79.3 237.9









#77
Bromobenzene
Concen: 21.319 ug/l
RT: 12.976 min Scan# 1874
Delta R.T. -0.006 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBSD02


Tgt Ion:156 Resp: 54030
Ion Ratio Lower Upper
156 100
77 230.7 117.7 353.1
158 95.1 48.1 144.4



#78
n-propylbenzene
Concen: 19.168 ug/l
RT: 13.035 min Scan# 1884
Delta R.T. -0.000 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01

Tgt Ion: 91 Resp: 257426
Ion Ratio Lower Upper
91 100
120 23.2 11.4 34.2









#79
2-Chlorotoluene
Concen: 19.571 ug/l
RT: 13.123 min Scan# 1899
Delta R.T. -0.000 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBSD02


Tgt Ion: 91 Resp: 157627
Ion Ratio Lower Upper
91 100
126 33.1 16.2 48.5



#80
1,3,5-Trimethylbenzene
Concen: 19.228 ug/l
RT: 13.170 min Scan# 1907
Delta R.T. -0.000 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01

Tgt Ion: 105 Resp: 175428
Ion Ratio Lower Upper
105 100
120 50.7 24.9 74.6









#81
 trans-1,4-Dichloro-2-butene
 Concen: 18.963 ug/l
 RT: 12.735 min Scan# 1833
 Delta R.T. -0.000 min
 Lab File: VN086945.D
 Acq: 11 Jun 2025 14:01




Instrument :
 MSVOA_N
 ClientSampleId :
 VN0611WBSD02


Tgt Ion: 75 Resp: 29703
 Ion Ratio Lower Upper
 75 100
 53 89.6 66.6 99.8
 89 47.5 36.7 55.1



#82
 4-Chlorotoluene
 Concen: 19.931 ug/l
 RT: 13.223 min Scan# 1916
 Delta R.T. -0.000 min
 Lab File: VN086945.D
 Acq: 11 Jun 2025 14:01

Tgt Ion: 91 Resp: 162421
 Ion Ratio Lower Upper
 91 100
 126 33.2 16.7 50.1









#83
tert-Butylbenzene
Concen: 18.749 ug/l
RT: 13.435 min Scan# 1952
Delta R.T. -0.000 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBSD02


Tgt Ion:119 Resp: 156588
Ion Ratio Lower Upper
119 100
91 63.5 32.1 96.5
134 24.9 13.4 40.1



#84
1,2,4-Trimethylbenzene
Concen: 19.373 ug/l
RT: 13.482 min Scan# 1960
Delta R.T. -0.000 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01

Tgt Ion:105 Resp: 177241
Ion Ratio Lower Upper
105 100
120 46.8 23.2 69.6









#85
 sec-Butylbenzene
 Concen: 18.109 ug/l
 RT: 13.611 min Scan# 1983
 Delta R.T. -0.006 min
 Lab File: VN086945.D
 Acq: 11 Jun 2025 14:01




Instrument :
 MSVOA_N
 ClientSampleId :
 VN0611WBSD02


Tgt Ion:105 Resp: 219639
 Ion Ratio Lower Upper
 105 100
 134 20.3 10.0 30.0



#86
 p-Isopropyltoluene
 Concen: 18.238 ug/l
 RT: 13.729 min Scan# 2002
 Delta R.T. -0.000 min
 Lab File: VN086945.D
 Acq: 11 Jun 2025 14:01

Tgt Ion:119 Resp: 182853
 Ion Ratio Lower Upper
 119 100
 134 26.9 13.5 40.4
 91 22.2 11.5 34.4









#87
1,3-Dichlorobenzene
Concen: 20.517 ug/l
RT: 13.729 min Scan# 2003
Delta R.T. -0.006 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01




Instrument :
MSVOA_N
ClientSampleId :
VN0611WBSD02


Tgt Ion:146 Resp: 102302
Ion Ratio Lower Upper
146 100
111 41.7 20.9 62.8
148 63.8 32.0 96.2



#88
1,4-Dichlorobenzene
Concen: 20.382 ug/l
RT: 13.811 min Scan# 2016
Delta R.T. -0.000 min
Lab File: VN086945.D
Acq: 11 Jun 2025 14:01

Tgt Ion:146 Resp: 103614
Ion Ratio Lower Upper
146 100
111 42.1 19.9 59.7
148 63.3 31.8 95.4





#89

n-Butylbenzene

Concen: 17.668 ug/l

RT: 14.053 min Scan# 2057

Delta R.T. -0.000 min

Lab File: VN086945.D

Acq: 11 Jun 2025 14:01

Instrument :

MSVOA_N

ClientSampleId :

VN0611WBSD02





Tgt Ion: 91 Resp: 171576

Ion Ratio Lower Upper

91 100

92 54.4 27.5 82.4

134 25.4 13.0 39.0



#90

Hexachloroethane

Concen: 18.810 ug/l

RT: 14.329 min Scan# 2104

Delta R.T. -0.006 min

Lab File: VN086945.D




Acq: 11 Jun 2025 14:01


Tgt Ion: 117 Resp: 31966

Ion Ratio Lower Upper

117 100

201 66.7 33.2 99.6





#91

1,2-Dichlorobenzene

Concen: 20.638 ug/l

RT: 14.106 min Scan# 2066

Delta R.T. -0.000 min

Lab File: VN086945.D

Acq: 11 Jun 2025 14:01

Instrument :

MSVOA_N

ClientSampleId :

VN0611WBSD02





Tgt Ion:146 Resp: 98864

Ion Ratio Lower Upper

146 100

111 44.0 21.6 64.6

148 62.9 31.7 95.1



#92

1,2-Dibromo-3-Chloropropane

Concen: 20.786 ug/l

RT: 14.723 min Scan# 2171

Delta R.T. 0.006 min

Lab File: VN086945.D

Acq: 11 Jun 2025 14:01




Tgt Ion: 75 Resp: 18610


Ion Ratio Lower Upper

75 100

155 81.3 41.1 123.4

157 105.2 52.7 158.1





#93

1,2,4-Trichlorobenzene

Concen: 18.660 ug/l

RT: 15.388 min Scan# 21

Delta R.T. -0.000 min

Lab File: VN086945.D

Acq: 11 Jun 2025 14:01

Instrument :

MSVOA_N

ClientSampleId :

VN0611WBSD02





Tgt Ion:180 Resp: 57083

Ion Ratio Lower Upper

180 100

182 94.8 46.9 140.7

145 32.7 16.0 47.9



#94

Hexachlorobutadiene

Concen: 15.506 ug/l

RT: 15.500 min Scan# 2303

Delta R.T. -0.000 min

Lab File: VN086945.D

Acq: 11 Jun 2025 14:01




Tgt Ion:225 Resp: 17672


Ion Ratio Lower Upper

225 100

223 68.8 33.5 100.4

227 61.4 31.1 93.5





#95

Naphthalene

Concen: 20.281 ug/l

RT: 15.635 min Scan# 21

Delta R.T. -0.006 min

Lab File: VN086945.D

Acq: 11 Jun 2025 14:01

Instrument :

MSVOA_N

ClientSampleId :

VN0611WBSD02





Tgt Ion:128 Resp: 230927

Ion Ratio Lower Upper

128 100

127 12.8 10.2 15.2

129 10.6 8.8 13.2



#96

1,2,3-Trichlorobenzene

Concen: 18.105 ug/l

RT: 15.835 min Scan# 2360

Delta R.T. -0.000 min

Lab File: VN086945.D

Acq: 11 Jun 2025 14:01




Tgt Ion:180 Resp: 55026

Ion Ratio Lower Upper

180 100

182 95.9 47.0 141.0

145 34.9 16.6 49.8



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061125\
 Data File : VN086965.D
 Acq On : 11 Jun 2025 21:23
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC

Quant Time: Jun 12 01:48:05 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	177574	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	326287	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	286630	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	142002	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	123795	52.069	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 104.140%	
35) Dibromofluoromethane	8.177	113	102484	53.000	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 106.000%	
50) Toluene-d8	10.571	98	382080	49.914	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 99.820%	
62) 4-Bromofluorobenzene	12.847	95	146464	51.499	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 103.000%	

Target Compounds					Qvalue	
2) Dichlorodifluoromethane	2.153	85	97946	55.320	ug/l	98
3) Chloromethane	2.395	50	103222	45.148	ug/l	99
4) Vinyl Chloride	2.553	62	126441	53.612	ug/l	100
5) Bromomethane	2.983	94	43618	33.049	ug/l	94
6) Chloroethane	3.147	64	84712	55.606	ug/l	97
7) Trichlorofluoromethane	3.524	101	172024	55.827	ug/l	99
8) Diethyl Ether	3.983	74	75627	56.330	ug/l	99
9) 1,1,2-Trichlorotrifluo...	4.400	101	104252	53.851	ug/l	99
10) Methyl Iodide	4.606	142	46802	18.649	ug/l	98
11) Tert butyl alcohol	5.547	59	180885	280.391	ug/l	99
12) 1,1-Dichloroethene	4.365	96	106380	53.824	ug/l	96
13) Acrolein	4.200	56	152464	746.633	ug/l	100
14) Allyl chloride	5.047	41	166911	50.921	ug/l	97
15) Acrylonitrile	5.741	53	422921	280.476	ug/l	100
16) Acetone	4.447	43	331168	262.661	ug/l	100
17) Carbon Disulfide	4.736	76	268580	49.127	ug/l	99
18) Methyl Acetate	5.047	43	214312	58.328	ug/l	99
19) Methyl tert-butyl Ether	5.818	73	406405	56.790	ug/l	99
20) Methylene Chloride	5.300	84	125496	53.166	ug/l	96
21) trans-1,2-Dichloroethene	5.806	96	115879	52.697	ug/l	95
22) Diisopropyl ether	6.688	45	382028	55.291	ug/l	99
23) Vinyl Acetate	6.618	43	1623707	278.142	ug/l	99
24) 1,1-Dichloroethane	6.588	63	222894	56.057	ug/l	99
25) 2-Butanone	7.494	43	557966	272.256	ug/l	97
26) 2,2-Dichloropropane	7.500	77	149440	48.315	ug/l	98
27) cis-1,2-Dichloroethene	7.500	96	146360	55.654	ug/l	97
28) Bromochloromethane	7.824	49	111849	57.211	ug/l	98
29) Tetrahydrofuran	7.853	42	361046	270.436	ug/l	97
30) Chloroform	7.977	83	219091	55.175	ug/l	97
31) Cyclohexane	8.265	56	181510	47.071	ug/l	98
32) 1,1,1-Trichloroethane	8.177	97	181857	53.846	ug/l	99
36) 1,1-Dichloropropene	8.377	75	153059	53.121	ug/l	99
37) Ethyl Acetate	7.571	43	196171	53.483	ug/l	99
38) Carbon Tetrachloride	8.371	117	150782	53.303	ug/l	96
39) Methylcyclohexane	9.606	83	176079	44.605	ug/l	97
40) Benzene	8.612	78	505922	53.695	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061125\
 Data File : VN086965.D
 Acq On : 11 Jun 2025 21:23
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC

Quant Time: Jun 12 01:48:05 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.788	41	112567	54.646	ug/l	97
42) 1,2-Dichloroethane	8.677	62	153959	53.873	ug/l	100
43) Isopropyl Acetate	8.694	43	317059	53.857	ug/l	99
44) Trichloroethene	9.359	130	119433	53.458	ug/l	98
45) 1,2-Dichloropropane	9.624	63	124231	54.196	ug/l	100
46) Dibromomethane	9.712	93	85919	56.431	ug/l	99
47) Bromodichloromethane	9.894	83	175093	55.894	ug/l	99
48) Methyl methacrylate	9.682	41	143734	53.046	ug/l	98
49) 1,4-Dioxane	9.706	88	58859	1194.044	ug/l #	96
51) 4-Methyl-2-Pentanone	10.447	43	995984	281.028	ug/l	99
52) Toluene	10.629	92	310797	53.976	ug/l	99
53) t-1,3-Dichloropropene	10.835	75	191765	54.745	ug/l	96
54) cis-1,3-Dichloropropene	10.318	75	203300	54.242	ug/l	97
55) 1,1,2-Trichloroethane	11.018	97	123723	55.842	ug/l	98
56) Ethyl methacrylate	10.882	69	209075	59.246	ug/l	96
57) 1,3-Dichloropropane	11.165	76	212617	55.319	ug/l	98
58) 2-Chloroethyl Vinyl ether	10.165	63	621634	295.338	ug/l	100
59) 2-Hexanone	11.200	43	661460	289.751	ug/l	94
60) Dibromochloromethane	11.359	129	132918	57.580	ug/l	100
61) 1,2-Dibromoethane	11.470	107	127444	56.119	ug/l	98
64) Tetrachloroethene	11.106	164	91667	50.534	ug/l	95
65) Chlorobenzene	11.894	112	336350	53.206	ug/l	100
66) 1,1,1,2-Tetrachloroethane	11.959	131	113463	55.835	ug/l	100
67) Ethyl Benzene	11.965	91	576593	52.952	ug/l	100
68) m/p-Xylenes	12.070	106	444183	106.561	ug/l	100
69) o-Xylene	12.400	106	219424	54.963	ug/l	98
70) Styrene	12.412	104	377725	55.293	ug/l	99
71) Bromoform	12.576	173	89676	59.559	ug/l #	100
73) Isopropylbenzene	12.694	105	534031	51.622	ug/l	100
74) N-amyl acetate	12.512	43	188759	52.216	ug/l	97
75) 1,1,2,2-Tetrachloroethane	12.935	83	200820	57.301	ug/l	99
76) 1,2,3-Trichloropropane	12.994	75	262042	77.632	ug/l	68
77) Bromobenzene	12.982	156	131869	55.583	ug/l	99
78) n-propylbenzene	13.035	91	643398	51.178	ug/l	99
79) 2-Chlorotoluene	13.123	91	394854	52.371	ug/l	98
80) 1,3,5-Trimethylbenzene	13.170	105	443398	51.914	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.735	75	72331	49.328	ug/l	97
82) 4-Chlorotoluene	13.223	91	406685	53.310	ug/l	100
83) tert-Butylbenzene	13.435	119	396615	50.730	ug/l	97
84) 1,2,4-Trimethylbenzene	13.482	105	447506	52.251	ug/l	99
85) sec-Butylbenzene	13.611	105	549887	48.432	ug/l	100
86) p-Isopropyltoluene	13.729	119	459716	48.982	ug/l	99
87) 1,3-Dichlorobenzene	13.735	146	253292	54.264	ug/l	99
88) 1,4-Dichlorobenzene	13.811	146	252880	53.138	ug/l	99
89) n-Butylbenzene	14.053	91	422802	46.509	ug/l	99
90) Hexachloroethane	14.329	117	83746	52.642	ug/l	98
91) 1,2-Dichlorobenzene	14.106	146	244439	54.507	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	14.717	75	45867	54.724	ug/l	97
93) 1,2,4-Trichlorobenzene	15.388	180	141753	49.501	ug/l	100
94) Hexachlorobutadiene	15.500	225	44722	41.918	ug/l	100
95) Naphthalene	15.641	128	576105	54.049	ug/l	100
96) 1,2,3-Trichlorobenzene	15.841	180	138678	48.742	ug/l	99


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061125\
Data File : VN086965.D
Acq On : 11 Jun 2025 21:23
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Quant Time: Jun 12 01:48:05 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 02:12:50 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.235 min Scan# 1068
Delta R.T. -0.001 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:168 Resp: 177574
Ion Ratio Lower Upper
168 100
99 58.2 49.1 73.7



#2
Dichlorodifluoromethane
Concen: 55.320 ug/l
RT: 2.153 min Scan# 34
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23

Tgt Ion: 85 Resp: 97946
Ion Ratio Lower Upper
85 100
87 31.5 16.2 48.6









#3
Chloromethane
Concen: 45.148 ug/l
RT: 2.395 min Scan# 75
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 50 Resp: 103222
Ion Ratio Lower Upper
50 100
52 32.5 25.7 38.5



#4
Vinyl Chloride
Concen: 53.612 ug/l
RT: 2.553 min Scan# 102
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23

Tgt Ion: 62 Resp: 126441
Ion Ratio Lower Upper
62 100
64 32.3 26.0 39.0









#5
Bromomethane
Concen: 33.049 ug/l
RT: 2.983 min Scan# 11
Delta R.T. -0.012 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 94 Resp: 43618
Ion Ratio Lower Upper
94 100
96 91.4 77.9 116.9



#6
Chloroethane
Concen: 55.606 ug/l
RT: 3.147 min Scan# 203
Delta R.T. -0.006 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23

Tgt Ion: 64 Resp: 84712
Ion Ratio Lower Upper
64 100
66 32.4 24.6 36.8









#7
Trichlorofluoromethane
Concen: 55.827 ug/l
RT: 3.524 min Scan# 267
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:101 Resp: 172024
Ion Ratio Lower Upper
101 100
103 64.7 52.2 78.2



#8
Diethyl Ether
Concen: 56.330 ug/l
RT: 3.983 min Scan# 345
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23

Tgt Ion: 74 Resp: 75627
Ion Ratio Lower Upper
74 100
45 89.7 45.5 136.5









#9
1,1,2-Trichlorotrifluoroethane
Concen: 53.851 ug/l
RT: 4.400 min Scan# 415
Delta R.T. 0.006 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:101 Resp: 104252
Ion Ratio Lower Upper
101 100
85 41.8 33.8 50.8
151 72.4 58.7 88.1



#10
Methyl Iodide
Concen: 18.649 ug/l
RT: 4.606 min Scan# 451
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23

Tgt Ion:142 Resp: 46802
Ion Ratio Lower Upper
142 100
127 39.6 30.9 46.3
141 12.5 10.4 15.6









#11
Tert butyl alcohol
Concen: 280.391 ug/l
RT: 5.547 min Scan# 611
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 59 Resp: 180885
Ion Ratio Lower Upper
59 100
57 9.7 8.1 12.1



#12
1,1-Dichloroethene
Concen: 53.824 ug/l
RT: 4.365 min Scan# 410
Delta R.T. 0.006 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23

Tgt Ion: 96 Resp: 106380
Ion Ratio Lower Upper
96 100
61 152.7 125.8 188.6
98 62.5 52.5 78.7









#13
Acrolein
Concen: 746.633 ug/l
RT: 4.200 min Scan# 382
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 56 Resp: 152464
Ion Ratio Lower Upper
56 100
55 69.4 55.6 83.4



#14
Allyl chloride
Concen: 50.921 ug/l
RT: 5.047 min Scan# 526
Delta R.T. 0.006 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23

Tgt Ion: 41 Resp: 166911
Ion Ratio Lower Upper
41 100
39 61.7 51.3 76.9
76 40.8 31.4 47.0









#15
Acrylonitrile
Concen: 280.476 ug/l
RT: 5.741 min Scan# 644
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 53 Resp: 422921
Ion Ratio Lower Upper
53 100
52 81.6 65.0 97.6
51 34.0 27.2 40.8



#16
Acetone
Concen: 262.661 ug/l
RT: 4.447 min Scan# 424
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23

Tgt Ion: 43 Resp: 331168
Ion Ratio Lower Upper
43 100
58 35.0 28.0 42.0









#17
Carbon Disulfide
Concen: 49.127 ug/l
RT: 4.736 min Scan# 41
Delta R.T. 0.006 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 76 Resp: 268580
Ion Ratio Lower Upper
76 100
78 9.4 7.3 10.9



#18
Methyl Acetate
Concen: 58.328 ug/l
RT: 5.047 min Scan# 526
Delta R.T. 0.006 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23

Tgt Ion: 43 Resp: 214312
Ion Ratio Lower Upper
43 100
74 26.6 20.7 31.1









#19
Methyl tert-butyl Ether
Concen: 56.790 ug/l
RT: 5.818 min Scan# 61
Delta R.T. 0.006 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 73 Resp: 406405
Ion Ratio Lower Upper
73 100
57 21.3 17.4 26.2



#20
Methylene Chloride
Concen: 53.166 ug/l
RT: 5.300 min Scan# 569
Delta R.T. 0.006 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23

Tgt Ion: 84 Resp: 125496
Ion Ratio Lower Upper
84 100
49 107.5 90.5 135.7
51 34.3 28.5 42.7
86 62.4 48.1 72.1









#21
trans-1,2-Dichloroethene
Concen: 52.697 ug/l
RT: 5.806 min Scan# 61
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 96 Resp: 115879
Ion Ratio Lower Upper
96 100
61 128.3 109.2 163.8
98 60.9 49.8 74.8



#22
Diisopropyl ether
Concen: 55.291 ug/l
RT: 6.688 min Scan# 805
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23

Tgt Ion: 45 Resp: 382028
Ion Ratio Lower Upper
45 100
43 52.3 41.9 62.9
87 29.5 22.4 33.6
59 12.9 9.8 14.6









#23
 Vinyl Acetate
 Concen: 278.142 ug/l
 RT: 6.618 min Scan# 793
 Delta R.T. -0.000 min
 Lab File: VN086965.D
 Acq: 11 Jun 2025 21:23




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC


Tgt Ion: 43 Resp: 1623707
 Ion Ratio Lower Upper
 43 100
 86 12.1 9.5 14.3



#24
 1,1-Dichloroethane
 Concen: 56.057 ug/l
 RT: 6.588 min Scan# 788
 Delta R.T. 0.006 min
 Lab File: VN086965.D
 Acq: 11 Jun 2025 21:23

Tgt Ion: 63 Resp: 222894
 Ion Ratio Lower Upper
 63 100
 98 7.0 3.6 10.8
 100 4.3 2.6 7.8









#25
2-Butanone
Concen: 272.256 ug/l
RT: 7.494 min Scan# 942
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 43 Resp: 557966
Ion Ratio Lower Upper
43 100
72 30.6 23.0 34.6



#26
2,2-Dichloropropane
Concen: 48.315 ug/l
RT: 7.500 min Scan# 943
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23

Tgt Ion: 77 Resp: 149440
Ion Ratio Lower Upper
77 100
97 24.1 11.6 34.7









#27
 cis-1,2-Dichloroethene
 Concen: 55.654 ug/l
 RT: 7.500 min Scan# 943
 Delta R.T. -0.000 min
 Lab File: VN086965.D
 Acq: 11 Jun 2025 21:23




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC


Tgt Ion: 96 Resp: 146360
 Ion Ratio Lower Upper
 96 100
 61 132.9 0.0 278.0
 98 64.2 0.0 128.2



#28
 Bromochloromethane
 Concen: 57.211 ug/l
 RT: 7.824 min Scan# 998
 Delta R.T. -0.000 min
 Lab File: VN086965.D
 Acq: 11 Jun 2025 21:23

Tgt Ion: 49 Resp: 111849
 Ion Ratio Lower Upper
 49 100
 129 2.0 0.0 3.8
 130 79.9 62.2 93.2





#29

Tetrahydrofuran

Concen: 270.436 ug/l

RT: 7.853 min Scan# 1003

Delta R.T. -0.000 min

Lab File: VN086965.D

Acq: 11 Jun 2025 21:23

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC





Tgt Ion: 42 Resp: 361046

Ion Ratio Lower Upper

42 100

72 49.5 37.8 56.6

71 46.0 35.6 53.4



#30

Chloroform

Concen: 55.175 ug/l

RT: 7.977 min Scan# 1024

Delta R.T. -0.000 min

Lab File: VN086965.D




Acq: 11 Jun 2025 21:23


Tgt Ion: 83 Resp: 219091

Ion Ratio Lower Upper

83 100

85 63.6 52.9 79.3









#31
Cyclohexane
Concen: 47.071 ug/l
RT: 8.265 min Scan# 1073
Delta R.T. -0.006 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 56 Resp: 181510
Ion Ratio Lower Upper
56 100
69 31.3 25.4 38.0
84 88.0 68.8 103.2



#32
1,1,1-Trichloroethane
Concen: 53.846 ug/l
RT: 8.177 min Scan# 1058
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23

Tgt Ion: 97 Resp: 181857
Ion Ratio Lower Upper
97 100
99 65.4 52.0 78.0
61 44.6 35.9 53.9





#33

1,2-Dichloroethane-d4

Concen: 52.069 ug/l

RT: 8.588 min Scan# 1128

Delta R.T. -0.000 min

Lab File: VN086965.D

Acq: 11 Jun 2025 21:23

Instrument :

MSVOA_N

ClientSampleId :





VSTDCCC050EC

Tgt Ion: 65 Resp: 123795

Ion Ratio Lower Upper

65 100

67 54.0 0.0 105.6



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.106 min Scan# 1216

Delta R.T. -0.000 min

Lab File: VN086965.D

Acq: 11 Jun 2025 21:23




Tgt Ion: 114 Resp: 326287


Ion Ratio Lower Upper

114 100

63 19.3 0.0 39.6

88 15.8 0.0 30.2









#35
Dibromofluoromethane
Concen: 53.000 ug/l
RT: 8.177 min Scan# 1057
Delta R.T. 0.006 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:113 Resp: 102484
Ion Ratio Lower Upper
113 100
111 103.4 84.2 126.2
192 17.7 14.2 21.4



#36
1,1-Dichloropropene
Concen: 53.121 ug/l
RT: 8.377 min Scan# 1092
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23

Tgt Ion: 75 Resp: 153059
Ion Ratio Lower Upper
75 100
110 35.4 17.6 52.9
77 31.0 25.3 37.9









#37
Ethyl Acetate
Concen: 53.483 ug/l
RT: 7.571 min Scan# 91
Delta R.T. -0.006 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 43 Resp: 196171
Ion Ratio Lower Upper
43 100
61 15.7 12.1 18.1
70 12.7 9.8 14.6



#38
Carbon Tetrachloride
Concen: 53.303 ug/l
RT: 8.371 min Scan# 1091
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23

Tgt Ion: 117 Resp: 150782
Ion Ratio Lower Upper
117 100
119 97.9 75.0 112.6
121 32.3 24.9 37.3





#39

Methylcyclohexane

Concen: 44.605 ug/l

RT: 9.606 min Scan# 1130

Delta R.T. -0.000 min

Lab File: VN086965.D

Acq: 11 Jun 2025 21:23

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC





Tgt Ion: 83 Resp: 176079

Ion Ratio Lower Upper

83 100

55 71.5 61.0 91.6

98 48.7 38.9 58.3



#40

Benzene

Concen: 53.695 ug/l

RT: 8.612 min Scan# 1132

Delta R.T. -0.000 min

Lab File: VN086965.D




Acq: 11 Jun 2025 21:23


Tgt Ion: 78 Resp: 505922

Ion Ratio Lower Upper

78 100

77 23.4 18.7 28.1









#41
Methacrylonitrile
Concen: 54.646 ug/l
RT: 7.788 min Scan# 992
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 41 Resp: 112567




Ion	Ratio	Lower	Upper
41	100		
39	51.6	41.2	61.8
67	80.7	61.4	92.2
52	30.1	22.7	34.1




#42
1,2-Dichloroethane
Concen: 53.873 ug/l
RT: 8.677 min Scan# 1143
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23

Tgt Ion: 62 Resp: 153959

Ion	Ratio	Lower	Upper
62	100		
98	10.5	0.0	21.2





#43

Isopropyl Acetate

Concen: 53.857 ug/l

RT: 8.694 min Scan# 1146

Delta R.T. -0.000 min

Lab File: VN086965.D

Acq: 11 Jun 2025 21:23

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC





Tgt Ion: 43 Resp: 317059

Ion Ratio Lower Upper

43 100

61 28.1 22.1 33.1

87 15.2 11.8 17.6



#44

Trichloroethene

Concen: 53.458 ug/l

RT: 9.359 min Scan# 1259

Delta R.T. -0.000 min

Lab File: VN086965.D




Acq: 11 Jun 2025 21:23


Tgt Ion:130 Resp: 119433

Ion Ratio Lower Upper

130 100

95 96.2 0.0 196.2





#45

1,2-Dichloropropane

Concen: 54.196 ug/l

RT: 9.624 min Scan# 1304

Delta R.T. -0.006 min

Lab File: VN086965.D

Acq: 11 Jun 2025 21:23

Instrument :

MSVOA_N

ClientSampleId :





VSTDCCC050EC

Tgt Ion: 63 Resp: 124231

Ion Ratio Lower Upper

63 100

65 31.6 25.5 38.3



#46

Dibromomethane

Concen: 56.431 ug/l

RT: 9.712 min Scan# 1319

Delta R.T. -0.006 min

Lab File: VN086965.D

Acq: 11 Jun 2025 21:23




Tgt Ion: 93 Resp: 85919


Ion Ratio Lower Upper

93 100

95 85.6 67.1 100.7

174 93.3 74.2 111.4









#47
Bromodichloromethane
Concen: 55.894 ug/l
RT: 9.894 min Scan# 1350
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 83 Resp: 175093
Ion Ratio Lower Upper
83 100
85 62.6 50.9 76.3
127 7.9 6.6 10.0



#48
Methyl methacrylate
Concen: 53.046 ug/l
RT: 9.682 min Scan# 1314
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23

Tgt Ion: 41 Resp: 143734
Ion Ratio Lower Upper
41 100
69 96.2 74.6 112.0
39 49.3 40.2 60.4









#49
1,4-Dioxane
Concen: 1194.044 ug/l
RT: 9.706 min Scan# 11
Delta R.T. 0.006 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 88 Resp: 58859
Ion Ratio Lower Upper
88 100
43 0.0 0.0 0.0
58 69.1 57.7 86.5



#50
Toluene-d8
Concen: 49.914 ug/l
RT: 10.571 min Scan# 1465
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23

Tgt Ion: 98 Resp: 382080
Ion Ratio Lower Upper
98 100
100 68.3 53.4 80.0









#51
4-Methyl-2-Pentanone
Concen: 281.028 ug/l
RT: 10.447 min Scan# 1444
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 43 Resp: 995984
Ion Ratio Lower Upper
43 100
58 43.6 34.2 51.4



#52
Toluene
Concen: 53.976 ug/l
RT: 10.629 min Scan# 1475
Delta R.T. -0.006 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23

Tgt Ion: 92 Resp: 310797
Ion Ratio Lower Upper
92 100
91 169.1 136.5 204.7









#53
 t-1,3-Dichloropropene
 Concen: 54.745 ug/l
 RT: 10.835 min Scan# 1510
 Delta R.T. -0.006 min
 Lab File: VN086965.D
 Acq: 11 Jun 2025 21:23




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC


Tgt Ion: 75 Resp: 191765
 Ion Ratio Lower Upper
 75 100
 77 30.9 26.4 39.6



#54
 cis-1,3-Dichloropropene
 Concen: 54.242 ug/l
 RT: 10.318 min Scan# 1422
 Delta R.T. -0.000 min
 Lab File: VN086965.D
 Acq: 11 Jun 2025 21:23

Tgt Ion: 75 Resp: 203300
 Ion Ratio Lower Upper
 75 100
 77 31.0 26.2 39.4
 39 39.0 32.3 48.5









#55
1,1,2-Trichloroethane
Concen: 55.842 ug/l
RT: 11.018 min Scan# 1541
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:	97	Resp:	123723
Ion	Ratio	Lower	Upper
97	100		
83	89.6	69.7	104.5
85	57.7	45.1	67.7
99	65.2	50.6	75.8



#56
Ethyl methacrylate
Concen: 59.246 ug/l
RT: 10.882 min Scan# 1518
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23

Tgt Ion:	69	Resp:	209075
Ion	Ratio	Lower	Upper
69	100		
41	55.6	47.2	70.8
39	26.5	23.0	34.6









#57
1,3-Dichloropropane
Concen: 55.319 ug/l
RT: 11.165 min Scan# 1566
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 76 Resp: 212617
Ion Ratio Lower Upper
76 100
78 32.9 25.3 37.9



#58
2-Chloroethyl Vinyl ether
Concen: 295.338 ug/l
RT: 10.165 min Scan# 1396
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23

Tgt Ion: 63 Resp: 621634
Ion Ratio Lower Upper
63 100
106 29.1 23.0 34.6









#59
2-Hexanone
Concen: 289.751 ug/l
RT: 11.200 min Scan# 1572
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 43 Resp: 661460
Ion Ratio Lower Upper
43 100
58 65.1 30.4 91.3



#60
Dibromochloromethane
Concen: 57.580 ug/l
RT: 11.359 min Scan# 1599
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23

Tgt Ion: 129 Resp: 132918
Ion Ratio Lower Upper
129 100
127 77.1 38.6 115.8









#61
1,2-Dibromoethane
Concen: 56.119 ug/l
RT: 11.470 min Scan# 1618
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:107 Resp: 127444
Ion Ratio Lower Upper
107 100
109 93.8 76.3 114.5



#62
4-Bromofluorobenzene
Concen: 51.499 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23

Tgt Ion: 95 Resp: 146464
Ion Ratio Lower Upper
95 100
174 72.0 0.0 141.8
176 69.3 0.0 132.6





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.865 min Scan# 117

Delta R.T. -0.000 min

Lab File: VN086965.D

Acq: 11 Jun 2025 21:23

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC





Tgt Ion:117 Resp: 286630

Ion Ratio Lower Upper

117 100

82 55.2 44.6 67.0

119 32.2 25.5 38.3



#64

Tetrachloroethene

Concen: 50.534 ug/l

RT: 11.106 min Scan# 1556

Delta R.T. -0.000 min

Lab File: VN086965.D

Acq: 11 Jun 2025 21:23

Tgt Ion:164 Resp: 91667




Ion Ratio Lower Upper


164 100

166 121.1 101.8 152.8

129 98.1 83.0 124.4

131 95.3 78.2 117.2









#65
Chlorobenzene
Concen: 53.206 ug/l
RT: 11.894 min Scan# 11
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:112 Resp: 336350
Ion Ratio Lower Upper
112 100
114 32.8 26.2 39.2



#66
1,1,1,2-Tetrachloroethane
Concen: 55.835 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. -0.006 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23

Tgt Ion:131 Resp: 113463
Ion Ratio Lower Upper
131 100
133 95.7 48.3 144.8
119 67.8 34.0 101.8









#67
Ethyl Benzene
Concen: 52.952 ug/l
RT: 11.965 min Scan# 1702
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 91 Resp: 576593
Ion Ratio Lower Upper
91 100
106 31.5 25.4 38.2



#68
m/p-Xylenes
Concen: 106.561 ug/l
RT: 12.070 min Scan# 1720
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23

Tgt Ion:106 Resp: 444183
Ion Ratio Lower Upper
106 100
91 199.8 159.4 239.0









#69
o-Xylene
Concen: 54.963 ug/l
RT: 12.400 min Scan# 1776
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:106 Resp: 219424
Ion Ratio Lower Upper
106 100
91 208.2 106.1 318.1



#70
Styrene
Concen: 55.293 ug/l
RT: 12.412 min Scan# 1778
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23

Tgt Ion:104 Resp: 377725
Ion Ratio Lower Upper
104 100
78 47.3 38.5 57.7
103 53.9 43.1 64.7









#71
Bromoform
Concen: 59.559 ug/l
RT: 12.576 min Scan# 1806
Delta R.T. -0.006 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:173 Resp: 89676
Ion Ratio Lower Upper
173 100
175 48.5 24.1 72.4
254 0.0 0.0 0.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.794 min Scan# 2013
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23

Tgt Ion:152 Resp: 142002
Ion Ratio Lower Upper
152 100
115 60.8 30.1 90.5
150 176.0 0.0 345.0





#73

Isopropylbenzene

Concen: 51.622 ug/l

RT: 12.694 min Scan# 1826

Delta R.T. -0.000 min

Lab File: VN086965.D

Acq: 11 Jun 2025 21:23

Instrument :

MSVOA_N

ClientSampleId :





VSTDCCC050EC

Tgt Ion:105 Resp: 534031

Ion Ratio Lower Upper

105 100

120 27.0 13.5 40.4



#74

N-ethyl acetate

Concen: 52.216 ug/l

RT: 12.512 min Scan# 1795

Delta R.T. -0.000 min

Lab File: VN086965.D

Acq: 11 Jun 2025 21:23

Tgt Ion: 43 Resp: 188759




Ion Ratio Lower Upper


43 100

70 49.1 37.6 56.4

55 26.3 19.0 28.6

61 27.2 20.8 31.2









#75
1,1,2,2-Tetrachloroethane
Concen: 57.301 ug/l
RT: 12.935 min Scan# 1867
Delta R.T. -0.006 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 83 Resp: 200820
Ion Ratio Lower Upper
83 100
131 9.6 4.9 14.7
85 64.7 32.0 96.0



#76
1,2,3-Trichloropropane
Concen: 77.632 ug/l
RT: 12.994 min Scan# 1877
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23

Tgt Ion: 75 Resp: 262042
Ion Ratio Lower Upper
75 100
77 116.7 79.3 237.9









#77
Bromobenzene
Concen: 55.583 ug/l
RT: 12.982 min Scan# 1875
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:156 Resp: 131869
Ion Ratio Lower Upper
156 100
77 232.0 117.7 353.1
158 96.4 48.1 144.4



#78
n-propylbenzene
Concen: 51.178 ug/l
RT: 13.035 min Scan# 1884
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23

Tgt Ion: 91 Resp: 643398
Ion Ratio Lower Upper
91 100
120 23.2 11.4 34.2









#79
2-Chlorotoluene
Concen: 52.371 ug/l
RT: 13.123 min Scan# 1899
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 91 Resp: 394854
Ion Ratio Lower Upper
91 100
126 33.6 16.2 48.5



#80
1,3,5-Trimethylbenzene
Concen: 51.914 ug/l
RT: 13.170 min Scan# 1907
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23

Tgt Ion: 105 Resp: 443398
Ion Ratio Lower Upper
105 100
120 50.5 24.9 74.6









#81
 trans-1,4-Dichloro-2-butene
 Concen: 49.328 ug/l
 RT: 12.735 min Scan# 1833
 Delta R.T. -0.000 min
 Lab File: VN086965.D
 Acq: 11 Jun 2025 21:23




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC


Tgt Ion: 75 Resp: 72331
 Ion Ratio Lower Upper
 75 100
 53 81.4 66.6 99.8
 89 43.7 36.7 55.1



#82
 4-Chlorotoluene
 Concen: 53.310 ug/l
 RT: 13.223 min Scan# 1916
 Delta R.T. -0.000 min
 Lab File: VN086965.D
 Acq: 11 Jun 2025 21:23

Tgt Ion: 91 Resp: 406685
 Ion Ratio Lower Upper
 91 100
 126 33.7 16.7 50.1









#83
 tert-Butylbenzene
 Concen: 50.730 ug/l
 RT: 13.435 min Scan# 1952
 Delta R.T. -0.000 min
 Lab File: VN086965.D
 Acq: 11 Jun 2025 21:23




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC


Tgt Ion:119 Resp: 396615
 Ion Ratio Lower Upper
 119 100
 91 62.4 32.1 96.5
 134 25.1 13.4 40.1



#84
 1,2,4-Trimethylbenzene
 Concen: 52.251 ug/l
 RT: 13.482 min Scan# 1960
 Delta R.T. -0.000 min
 Lab File: VN086965.D
 Acq: 11 Jun 2025 21:23

Tgt Ion:105 Resp: 447506
 Ion Ratio Lower Upper
 105 100
 120 47.0 23.2 69.6





#85

sec-Butylbenzene

Concen: 48.432 ug/l

RT: 13.611 min Scan# 1983

Delta R.T. -0.006 min

Lab File: VN086965.D

Acq: 11 Jun 2025 21:23

Instrument :

MSVOA_N

ClientSampleId :





VSTDCCC050EC

Tgt Ion:105 Resp: 549887

Ion Ratio Lower Upper

105 100

134 20.1 10.0 30.0



#86

p-Isopropyltoluene

Concen: 48.982 ug/l

RT: 13.729 min Scan# 2002

Delta R.T. -0.000 min

Lab File: VN086965.D

Acq: 11 Jun 2025 21:23




Tgt Ion:119 Resp: 459716


Ion Ratio Lower Upper

119 100

134 26.2 13.5 40.4

91 22.3 11.5 34.4









#87
 1,3-Dichlorobenzene
 Concen: 54.264 ug/l
 RT: 13.735 min Scan# 2003
 Delta R.T. -0.000 min
 Lab File: VN086965.D
 Acq: 11 Jun 2025 21:23




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC


Tgt Ion:146 Resp: 253292
 Ion Ratio Lower Upper
 146 100
 111 41.0 20.9 62.8
 148 62.9 32.0 96.2



#88
 1,4-Dichlorobenzene
 Concen: 53.138 ug/l
 RT: 13.811 min Scan# 2016
 Delta R.T. -0.000 min
 Lab File: VN086965.D
 Acq: 11 Jun 2025 21:23

Tgt Ion:146 Resp: 252880
 Ion Ratio Lower Upper
 146 100
 111 40.2 19.9 59.7
 148 64.2 31.8 95.4





#89

n-Butylbenzene

Concen: 46.509 ug/l

RT: 14.053 min Scan# 2057

Delta R.T. -0.000 min

Lab File: VN086965.D

Acq: 11 Jun 2025 21:23

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC





Tgt Ion: 91 Resp: 422802

Ion Ratio Lower Upper

91 100

92 53.9 27.5 82.4

134 26.1 13.0 39.0



#90

Hexachloroethane

Concen: 52.642 ug/l

RT: 14.329 min Scan# 2104

Delta R.T. -0.006 min

Lab File: VN086965.D




Acq: 11 Jun 2025 21:23


Tgt Ion: 117 Resp: 83746

Ion Ratio Lower Upper

117 100

201 64.9 33.2 99.6





#91

1,2-Dichlorobenzene

Concen: 54.507 ug/l

RT: 14.106 min Scan# 2066

Delta R.T. -0.000 min

Lab File: VN086965.D

Acq: 11 Jun 2025 21:23

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC





Tgt Ion:146 Resp: 244439

Ion Ratio Lower Upper

146 100

111 43.5 21.6 64.6

148 63.6 31.7 95.1



#92

1,2-Dibromo-3-Chloropropane

Concen: 54.724 ug/l

RT: 14.717 min Scan# 2170

Delta R.T. -0.000 min

Lab File: VN086965.D

Acq: 11 Jun 2025 21:23




Tgt Ion: 75 Resp: 45867


Ion Ratio Lower Upper

75 100

155 84.4 41.1 123.4

157 108.4 52.7 158.1









#93
1,2,4-Trichlorobenzene
Concen: 49.501 ug/l
RT: 15.388 min Scan# 21
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:180 Resp: 141753
Ion Ratio Lower Upper
180 100
182 93.7 46.9 140.7
145 32.3 16.0 47.9



#94
Hexachlorobutadiene
Concen: 41.918 ug/l
RT: 15.500 min Scan# 2303
Delta R.T. -0.000 min
Lab File: VN086965.D
Acq: 11 Jun 2025 21:23

Tgt Ion:225 Resp: 44722
Ion Ratio Lower Upper
225 100
223 66.3 33.5 100.4
227 62.3 31.1 93.5





#95

Naphthalene

Concen: 54.049 ug/l

RT: 15.641 min Scan# 2327

Delta R.T. -0.000 min

Lab File: VN086965.D

Acq: 11 Jun 2025 21:23

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC





Tgt Ion:128 Resp: 576105

Ion Ratio Lower Upper

128 100

127 12.6 10.2 15.2

129 10.8 8.8 13.2



#96

1,2,3-Trichlorobenzene

Concen: 48.742 ug/l

RT: 15.841 min Scan# 2361

Delta R.T. 0.006 min

Lab File: VN086965.D

Acq: 11 Jun 2025 21:23

Tgt Ion:180 Resp: 138678

Ion Ratio Lower Upper

180 100

182 95.2 47.0 141.0

145 33.4 16.6 49.8

