



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	<u>Ardmore Chemical</u>				SDG No.:	<u>Q2264</u>			
Contract:	<u>ARDM01</u>	Lab Code:	<u>CHEM</u>		Case No.:	<u>Q2264</u>	SAS No.:	<u>Q2264</u>	
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB32	Mercury	0.027	+/-0.2	U	0.20	CV	06/11/2025	09:48	LB136099

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Ardmore Chemical

SDG No.: Q2264

Contract: ARDM01

Lab Code: CHEM

Case No.: Q2264

SAS No.: Q2264

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB23	Mercury	0.027	+/-0.2	U	0.20	CV	06/11/2025	09:52	LB136099
CCB24	Mercury	0.027	+/-0.2	U	0.20	CV	06/11/2025	10:28	LB136099
CCB25	Mercury	0.027	+/-0.2	U	0.20	CV	06/11/2025	10:37	LB136099

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Ardmore Chemical</u>		SDG No.: <u>Q2264</u>							
Contract: <u>ARDM01</u>		Lab Code: <u>CHEM</u>		Case No.: <u>Q2264</u>		SAS No.: <u>Q2264</u>			
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Lead	2.30	+/-6	U	12.0	P	06/17/2025	12:46	LB136187
	Zinc	3.50	+/-20	U	40.0	P	06/17/2025	12:46	LB136187

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Ardmore Chemical</u>		SDG No.: <u>Q2264</u>							
Contract: <u>ARDM01</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q2264</u>	SAS No.: <u>Q2264</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Lead	2.30	+/-6	U	12.0	P	06/17/2025	13:48	LB136187
	Zinc	3.50	+/-20	U	40.0	P	06/17/2025	13:48	LB136187
CCB02	Lead	2.30	+/-6	U	12.0	P	06/17/2025	14:50	LB136187
	Zinc	3.50	+/-20	U	40.0	P	06/17/2025	14:50	LB136187
CCB03	Lead	2.30	+/-6	U	12.0	P	06/17/2025	15:42	LB136187
	Zinc	3.50	+/-20	U	40.0	P	06/17/2025	15:42	LB136187
CCB04	Lead	2.30	+/-6	U	12.0	P	06/17/2025	16:34	LB136187
	Zinc	3.50	+/-20	U	40.0	P	06/17/2025	16:34	LB136187
CCB05	Lead	2.94	+/-6	J	12.0	P	06/17/2025	17:41	LB136187
	Zinc	3.50	+/-20	U	40.0	P	06/17/2025	17:41	LB136187
CCB06	Lead	2.30	+/-6	U	12.0	P	06/17/2025	18:42	LB136187
	Zinc	3.50	+/-20	U	40.0	P	06/17/2025	18:42	LB136187
CCB07	Lead	2.30	+/-6	U	12.0	P	06/17/2025	19:30	LB136187
	Zinc	3.50	+/-20	U	40.0	P	06/17/2025	19:30	LB136187
CCB08	Lead	4.53	+/-6	J	12.0	P	06/17/2025	20:22	LB136187
	Zinc	4.70	+/-20	J	40.0	P	06/17/2025	20:22	LB136187

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Ardmore Chemical

SDG No.: Q2264

Contract: ARDM01

Lab Code: CHEM

Case No.: Q2264

SAS No.: Q2264

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
-----------	---------	----------------	---------------------	--------------	------	---	------------------	------------------	---------------