

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Ardmore Chemical **SDG No.:** Q2264
Contract: ARDM01 **Lab Code:** CHEM **Case No.:** Q2264 **SAS No.:** Q2264
ICS Source: EPA **Instrument ID:** P5

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Lead	5.50			-12	12	06/17/2025	13:11	LB136187
	Zinc	10.3			-40	40	06/17/2025	13:11	LB136187
ICSAB01	Lead	53.1	49.0	108	37	61	06/17/2025	13:22	LB136187
	Zinc	1050	950	110	809	1095	06/17/2025	13:22	LB136187
ICSA	Lead	10.3			-12	12	06/17/2025	13:28	LB136187
	Zinc	24.2			-40	40	06/17/2025	13:28	LB136187
ICSAB	Lead	65.5	49.0	134	37	61	06/17/2025	13:35	LB136187
	Zinc	1040	950	110	809	1095	06/17/2025	13:35	LB136187