



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061125\
Data File : VX046627.D
Acq On : 11 Jun 2025 10:45
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT3.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X061025W.M
Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
Last Update : Wed Jun 11 06:20:28 2025



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1633

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.8	8713	PASS
75	95	30	60	53.4	22344	PASS
95	95	100	100	100.0	41819	PASS
96	95	5	9	6.6	2746	PASS
173	174	0.00	2	0.9	259	PASS
174	95	50	100	67.4	28179	PASS
175	174	5	9	7.7	2168	PASS
176	174	95	101	96.9	27299	PASS
177	176	5	9	6.2	1688	PASS