

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: ARDM01
 Lab Code: CHEM Case No.: Q2264 SAS No.: Q2264 SDG NO.: Q2264
 Lab File ID: VX046628.D Date Analyzed: 06/11/2025
 Instrument ID: MSVOA_X Time Analyzed: 11:12
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	20635	4.91	110776	6.76	99656	10.06
UPPER LIMIT	41270	5.41	221552	7.263	199312	10.555
LOWER LIMIT	10317.5	4.41	55388	6.263	49828	9.555
EPA SAMPLE NO.						
EFF-WW	16594	4.92	91514	6.77	85872	10.06
VX0611WBL01	19239	4.92	103534	6.77	93773	10.06
VX0611WBS01	20422	4.91	109005	6.76	100408	10.06
VX0611WBSD01	19624	4.91	102297	6.76	94515	10.06

IS1 = Bromochloromethane
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.