



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

PARSONS Engineering of New York, Inc.

Project Name: Con Ed Non MGP – Atlantic Ave 453957.600024.05

Project # N/A

Order ID # Q2267

Test Name: Flash Point,pH,Reactive Cyanide,Reactive Sulfide

A. Number of Samples and Date of Receipt:

1 Water sample was received on 06/06/2025.

B. Parameters:

According to the Chain of Custody document, the following analyses were requested: Flash Point, PCB, pH, RCRA CHARACTERISTICS, Reactive Cyanide, Reactive Sulfide, TCLP BNA, TCLP Extraction, TCLP ICP Metals, TCLP Mercury, TCLP METALS, TCLP VOA and TCLP ZHE Extraction. This data package contains results for Flash Point,pH,Reactive Cyanide,Reactive Sulfide.

C. Analytical Techniques:

The analysis of Flash Point was based on method 1010B, The analysis of Reactive Cyanide was based on method 9012B, The analysis of Reactive Sulfide was based on method 9034 and The analysis of pH was based on method 9040C.

D. QA/ QC Samples:

The Holding Times were met for all samples except for WC-20250605 of pH as sample was receive out of holding time.

The Blank Spike met requirements for all samples.

The Duplicate analysis met criteria for all samples.

The Blank analysis did not indicate the presence of lab contamination.

The Calibration met the requirements.

E. Additional Comments:

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____