

CASE NARRATIVE

PARSONS Engineering of New York, Inc. Project Name: Con Ed Non MGP – Atlantic Ave 453957.600024.05 Project # N/A Order ID # Q2267 Test Name: TCLP VOA

A. Number of Samples and Date of Receipt:

1 Water sample was received on 06/06/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: Flash Point, PCB, pH, RCRA CHARACTERISTICS, Reactive Cyanide, Reactive Sulfide, TCLP BNA, TCLP Extraction, TCLP ICP Metals, TCLP Mercury, TCLP METALS, TCLP VOA and TCLP ZHE Extraction. This data package contains results for TCLP VOA.

C. Analytical Techniques:

The analysis performed on instrument MSVOA_X were done using GC column DB-624UI 20m 0.18mm 1.0 um. Cat#121-1324UIThe analysis of TCLP VOA was based on method 8260D and TCLP extraction method was 1311.

D. QA/ QC Samples:

The Holding Times were met for all analysis. The Surrogate recoveries met the acceptable criteria. The Internal Standards Areas met the acceptable requirements. The Retention Times were acceptable for all samples. The RPD met criteria . The Blank Spike met requirements for all samples . The Blank Spike Duplicate met requirements for all samples . The Blank analysis did not indicate the presence of lab contamination. The Initial Calibration met the requirements . The Continuous Calibration File ID VX046657.D met the requirements except for 2-Butanone,Vinyl Chloride and 1,2-Dichloroethane-d4 . But associated samples have not positive hit therefore no corrective action was taken. The Tuning criteria met requirements.

E. Additional Comments:

Samples for MS/MSD for VOC analysis were not provided with this set of samples. The Blank Spike Duplicate is reported with the data.

Trip Blank was not provided with this set of samples.



Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <20% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 20% for the Initial Calibration curve for SW-846 analysis.

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____