

Analytical Summary Report

Analysis Method: 9040C
Parameter: pH
Run Number: LB136081

Analyst By : jignesh
Supervisor Review By : Iwona
Slope : 98.6
pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.05].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	06/10/2025	08:00
2	CAL2	1	Water	NA	NA	20.2	7.00	06/10/2025	08:02
3	CAL3	1	Water	NA	NA	20.2	10.02	06/10/2025	08:03
4	ICV	1	Water	NA	NA	20.2	7.01	06/10/2025	08:05
5	CCV1	1	Water	NA	NA	20.1	2.02	06/10/2025	09:10
6	Q2267-01	1	Water	NA	NA	20.2	7.11	06/10/2025	09:20
7	Q2267-01DUP	1	Water	NA	NA	20.3	7.12	06/10/2025	09:22
8	CCV2	1	Water	NA	NA	20.3	12.02	06/10/2025	09:25

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph w q276 WorkList ID : 190054 Department : Wet-Chemistry Date : 06-10-2025 07:35:24

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2267-01	WC-20250605	Water	pH	Cool 4 deg C	PARS02	D41	06/06/2025	9040C

Date/Time 06.10.25 07:50
Raw Sample Received by: SA wecc
Raw Sample Relinquished by: JF SN

Date/Time 06.10.25 13:30
Raw Sample Received by: JF SN
Raw Sample Relinquished by: JF wecc