



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Verification

Client: PARSONS Engineering of New York, Inc.

SDG No.: Q2267

Project: Con Ed Non MGP – Atlantic Ave 453957.600024.05

RunNo.: LB136068

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: Flash Point	ICV ° F	80.6	81	100	78-84	06/09/2025

Initial and Continuing Calibration Verification

Client: PARSONS Engineering of New York, Inc.

SDG No.: Q2267

Project: Con Ed Non MGP – Atlantic Ave 453957.600024.05

RunNo.: LB136081

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV						
pH		pH	7.01	7	100	90-110	06/10/2025
Sample ID:	CCV1						
pH		pH	2.02	2.00	101	90-110	06/10/2025
Sample ID:	CCV2						
pH		pH	12.02	12.00	100	90-110	06/10/2025

Initial and Continuing Calibration Verification

Client: PARSONS Engineering of New York, Inc.

SDG No.: Q2267

Project: Con Ed Non MGP – Atlantic Ave 453957.600024.05

RunNo.: LB136129

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Reactive Cyanide	mg/L	0.095	0.099	96	85-115	06/12/2025
Sample ID: CCV1 Reactive Cyanide	mg/L	0.24	0.25	96	90-110	06/12/2025
Sample ID: CCV2 Reactive Cyanide	mg/L	0.23	0.25	92	90-110	06/12/2025
Sample ID: CCV3 Reactive Cyanide	mg/L	0.24	0.25	96	90-110	06/12/2025
Sample ID: CCV4 Reactive Cyanide	mg/L	0.25	0.25	100	90-110	06/12/2025