

Initial and Continuing Calibration Blank Summary

Client: PARSONS Engineering of New York, Inc.

SDG No.: Q2267

Project: Con Ed Non MGP – Atlantic Ave 453957.600024.05

RunNo.: LB136129

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 Reactive Cyanide	mg/L	0.00098	0.0025	J	0.00096	0.005	06/12/2025
Sample ID: CCB1 Reactive Cyanide	mg/L	< 0.0025	0.0025	U	0.00096	0.005	06/12/2025
Sample ID: CCB2 Reactive Cyanide	mg/L	0.00096	0.0025	J	0.00096	0.005	06/12/2025
Sample ID: CCB3 Reactive Cyanide	mg/L	< 0.0025	0.0025	U	0.00096	0.005	06/12/2025
Sample ID: CCB4 Reactive Cyanide	mg/L	0.0013	0.0025	J	0.00096	0.005	06/12/2025