



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: PARSONS Engineering of New York, Inc.

SDG No.: Q2267

Contract: PARS02

Lab Code: CHEM

Case No.: Q2267

SAS No.: Q2267

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Arsenic	18.7	20.0	94	65 - 135	P	06/11/2025	14:07	LB136119
	Barium	89.9	100	90	65 - 135	P	06/11/2025	14:07	LB136119
	Cadmium	4.63	6.0	77	65 - 135	P	06/11/2025	14:07	LB136119
	Chromium	9.42	10.0	94	65 - 135	P	06/11/2025	14:07	LB136119
	Lead	9.92	12.0	83	65 - 135	P	06/11/2025	14:07	LB136119
	Selenium	20.1	20.0	101	65 - 135	P	06/11/2025	14:07	LB136119
	Silver	9.93	10.0	99	65 - 135	P	06/11/2025	14:07	LB136119
CRA	Mercury	0.19	0.2	93	70 - 130	CV	06/12/2025	10:21	LB136124