

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: PARSONS Engineering of New York, Inc.

SDG No.: Q2267

Instrument: CV1

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB168390TB		WATER		Batch Number:	PB168425		Prep Date:	06/11/2025	
	Mercury	0.76	<2	U	2.00	CV	06/12/2025	11:47	LB136124
Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB168425BL		WATER		Batch Number:	PB168425		Prep Date:	06/11/2025	
	Mercury	0.076	<0.2	U	0.20	CV	06/12/2025	10:54	LB136124

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: PARSONS Engineering of New York, Inc.

SDG No.: Q2267

Instrument: P4

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB168390TB		WATER		Batch Number:	PB168423		Prep Date:	06/11/2025	
	Arsenic	25.6	<50	U	100	P	06/11/2025	18:42	LB136119
	Barium	72.8	<250	U	500	P	06/11/2025	18:42	LB136119
	Cadmium	2.50	<15	U	30.0	P	06/11/2025	18:42	LB136119
	Chromium	10.6	<25	U	50.0	P	06/11/2025	18:42	LB136119
	Lead	11.5	<30	U	60.0	P	06/11/2025	18:42	LB136119
	Selenium	48.2	<50	U	100	P	06/11/2025	18:42	LB136119
	Silver	8.10	<25	U	50.0	P	06/11/2025	18:42	LB136119
Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB168423BL		WATER		Batch Number:	PB168423		Prep Date:	06/11/2025	
	Arsenic	25.6	<50	U	100	P	06/11/2025	18:33	LB136119
	Barium	72.8	<250	U	500	P	06/11/2025	18:33	LB136119
	Cadmium	2.50	<15	U	30.0	P	06/11/2025	18:33	LB136119
	Chromium	10.6	<25	U	50.0	P	06/11/2025	18:33	LB136119
	Lead	11.5	<30	U	60.0	P	06/11/2025	18:33	LB136119
	Selenium	48.2	<50	U	100	P	06/11/2025	18:33	LB136119
	Silver	8.10	<25	U	50.0	P	06/11/2025	18:33	LB136119