

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: PARSONS Engineering of New York, Inc. **SDG No.:** Q2267
Contract: PARS02 **Lab Code:** CHEM **Case No.:** Q2267 **SAS No.:** Q2267
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Arsenic	5.34			-20	20	06/11/2025	14:14	LB136119
	Barium	-5.25	6.0	88	-94	106	06/11/2025	14:14	LB136119
	Cadmium	3.57	1.0	357	-5	7	06/11/2025	14:14	LB136119
	Chromium	44.3	52.0	85	42	62	06/11/2025	14:14	LB136119
	Lead	-2.25			-12	12	06/11/2025	14:14	LB136119
	Selenium	13.7			-20	20	06/11/2025	14:14	LB136119
	Silver	-1.50			-10	10	06/11/2025	14:14	LB136119
ICSAB01	Arsenic	96.7	104	93	88.4	120	06/11/2025	14:27	LB136119
	Barium	477	537	89	437	637	06/11/2025	14:27	LB136119
	Cadmium	923	972	95	826	1120	06/11/2025	14:27	LB136119
	Chromium	506	542	93	460	624	06/11/2025	14:27	LB136119
	Lead	42.5	49.0	87	37	61	06/11/2025	14:27	LB136119
	Selenium	59.1	46.0	128	26	66	06/11/2025	14:27	LB136119
	Silver	201	201	100	170	232	06/11/2025	14:27	LB136119