

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: PARSONS Engineering of New York, Inc. **Level:** LOW **SDG No.:** Q2267
Contract: PARS02 **Lab Code:** CHEM **Case No.:** Q2267 **SAS No.:** Q2267
Matrix: Water **Sample ID:** Q2267-01 **Client ID:** WC-20250605DUP
Percent Solids for Sample: NA **Duplicate ID** Q2267-01DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	100	U	100	U			P
Barium	ug/L	20	500	U	500	U			P
Cadmium	ug/L	20	30.0	U	30.0	U			P
Chromium	ug/L	20	50.0	U	50.0	U			P
Lead	ug/L	20	60.0	U	60.0	U			P
Selenium	ug/L	20	100	U	100	U			P
Silver	ug/L	20	50.0	U	50.0	U			P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: PARSONS Engineering of New York, Inc. **Level:** LOW **SDG No.:** Q2267
Contract: PARS02 **Lab Code:** CHEM **Case No.:** Q2267 **SAS No.:** Q2267
Matrix: Water **Sample ID:** Q2267-01MS **Client ID:** WC-20250605MSD
Percent Solids for Sample: NA **Duplicate ID** Q2267-01MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	3850		4100		6		P
Barium	ug/L	20	963		991		3		P
Cadmium	ug/L	20	927		988		6		P
Chromium	ug/L	20	1910		2060		8		P
Lead	ug/L	20	4490		4800		7		P
Selenium	ug/L	20	9380		9970		6		P
Silver	ug/L	20	353		379		7		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: PARSONS Engineering of New York, Inc. **Level:** LOW **SDG No.:** Q2267
Contract: PARS02 **Lab Code:** CHEM **Case No.:** Q2267 **SAS No.:** Q2267
Matrix: Water **Sample ID:** Q2285-05 **Client ID:** HAM-CONCRETEDUP
Percent Solids for Sample: NA **Duplicate ID** Q2285-05DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Mercury	ug/L	20	2.00	U	2.00	U			CV

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: PARSONS Engineering of New York, Inc. **Level:** LOW **SDG No.:** Q2267
Contract: PARS02 **Lab Code:** CHEM **Case No.:** Q2267 **SAS No.:** Q2267
Matrix: Water **Sample ID:** Q2285-05MS **Client ID:** HAM-CONCRETEMSD
Percent Solids for Sample: NA **Duplicate ID** Q2285-05MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Mercury	ug/L	20	29.6		28.5		4		CV