

CALIBRATION VERIFICATION SUMMARY

Contract: PARS02

Lab Code: CHEM Case No.: Q2267 SAS No.: Q2267 SDG NO.: Q2267

Continuing Calib Date: 06/10/2025 Initial Calibration Date(s): 05/14/2025 05/15/2025

Continuing Calib Time: 11:03 Initial Calibration Time(s): 17:19 01:31

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1248-1	(1)	4.77	4.77	4.67	4.87	0.00
Aroclor-1248-2	(2)	5.01	5.01	4.91	5.11	0.00
Aroclor-1248-3	(3)	5.22	5.23	5.13	5.33	0.01
Aroclor-1248-4	(4)	5.57	5.58	5.48	5.68	0.01
Aroclor-1248-5	(5)	5.62	5.62	5.52	5.72	0.00
Tetrachloro-m-xylene		3.68	3.68	3.58	3.78	0.00
Decachlorobiphenyl		8.72	8.72	8.62	8.82	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: PARS02

Lab Code: CHEM Case No.: Q2267 SAS No.: Q2267 SDG NO.: Q2267

Continuing Calib Date: 06/10/2025 Initial Calibration Date(s): 05/14/2025 05/15/2025

Continuing Calib Time: 11:03 Initial Calibration Time(s): 17:19 01:31

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1248-1 (1)	4.75	4.76	4.66	4.86	0.01
Aroclor-1248-2 (2)	4.99	5.00	4.90	5.10	0.01
Aroclor-1248-3 (3)	5.03	5.04	4.94	5.14	0.01
Aroclor-1248-4 (4)	5.20	5.21	5.11	5.31	0.01
Aroclor-1248-5 (5)	5.59	5.60	5.50	5.70	0.01
Tetrachloro-m-xylene	3.68	3.68	3.58	3.78	0.01
Decachlorobiphenyl	8.67	8.68	8.58	8.78	0.02