

CALIBRATION VERIFICATION SUMMARY

Contract: PARS02

Lab Code: CHEM Case No.: Q2267 SAS No.: Q2267 SDG NO.: Q2267

Continuing Calib Date: 06/10/2025 Initial Calibration Date(s): 05/14/2025 05/15/2025

Continuing Calib Time: 11:21 Initial Calibration Time(s): 17:19 01:31

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1254-1 (1)	5.57	5.58	5.48	5.68	0.01
Aroclor-1254-2 (2)	5.72	5.73	5.63	5.83	0.01
Aroclor-1254-3 (3)	6.13	6.13	6.03	6.23	0.00
Aroclor-1254-4 (4)	6.36	6.36	6.26	6.46	0.00
Aroclor-1254-5 (5)	6.78	6.78	6.68	6.88	0.00
Tetrachloro-m-xylene	3.68	3.68	3.58	3.78	0.00
Decachlorobiphenyl	8.72	8.72	8.62	8.82	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: PARS02

Lab Code: CHEM Case No.: Q2267 SAS No.: Q2267 SDG NO.: Q2267

Continuing Calib Date: 06/10/2025 Initial Calibration Date(s): 05/14/2025 05/15/2025

Continuing Calib Time: 11:21 Initial Calibration Time(s): 17:19 01:31

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1254-1	(1)	5.55	5.56	5.46	5.66	0.01
Aroclor-1254-2	(2)	5.70	5.71	5.61	5.81	0.01
Aroclor-1254-3	(3)	6.10	6.11	6.01	6.21	0.01
Aroclor-1254-4	(4)	6.33	6.34	6.24	6.44	0.01
Aroclor-1254-5	(5)	6.75	6.75	6.65	6.85	0.01
Tetrachloro-m-xylene		3.68	3.68	3.58	3.78	0.01
Decachlorobiphenyl		8.66	8.67	8.57	8.77	0.01