

CALIBRATION VERIFICATION SUMMARY

Contract: PARS02

Lab Code: CHEM Case No.: Q2267 SAS No.: Q2267 SDG NO.: Q2267

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/14/2025 05/14/2025

Client Sample No.: CCAL01 Date Analyzed: 06/10/2025

Lab Sample No.: AR1660CCC500 Data File : PO111569.D Time Analyzed: 10:26

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.773	4.673	4.873	543.730	500.000	8.7
Aroclor-1016-2	4.792	4.692	4.892	544.840	500.000	9.0
Aroclor-1016-3	4.849	4.749	4.949	556.420	500.000	11.3
Aroclor-1016-4	4.968	4.869	5.069	554.460	500.000	10.9
Aroclor-1016-5	5.227	5.126	5.326	561.910	500.000	12.4
Aroclor-1260-1	6.265	6.166	6.366	579.850	500.000	16.0
Aroclor-1260-2	6.455	6.355	6.555	612.830	500.000	22.6
Aroclor-1260-3	6.822	6.722	6.922	607.480	500.000	21.5
Aroclor-1260-4	7.082	6.983	7.183	553.840	500.000	10.8
Aroclor-1260-5	7.325	7.225	7.425	555.070	500.000	11.0
Decachlorobiphenyl	8.724	8.624	8.824	47.480	50.000	-5.0
Tetrachloro-m-xylene	3.684	3.582	3.782	51.920	50.000	3.8

CALIBRATION VERIFICATION SUMMARY

Contract: PARS02

Lab Code: CHEM Case No.: Q2267 SAS No.: Q2267 SDG NO.: Q2267

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/14/2025 05/14/2025

Client Sample No.: CCAL01 Date Analyzed: 06/10/2025

Lab Sample No.: AR1660CCC500 Data File : PO111569.D Time Analyzed: 10:26

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.754	4.658	4.858	486.180	500.000	-2.8
Aroclor-1016-2	4.772	4.677	4.877	497.080	500.000	-0.6
Aroclor-1016-3	4.948	4.852	5.052	496.220	500.000	-0.8
Aroclor-1016-4	4.990	4.894	5.094	492.820	500.000	-1.4
Aroclor-1016-5	5.202	5.107	5.307	489.370	500.000	-2.1
Aroclor-1260-1	6.233	6.138	6.338	492.100	500.000	-1.6
Aroclor-1260-2	6.421	6.326	6.526	493.500	500.000	-1.3
Aroclor-1260-3	6.573	6.479	6.679	456.330	500.000	-8.7
Aroclor-1260-4	7.044	6.949	7.149	473.200	500.000	-5.4
Aroclor-1260-5	7.285	7.191	7.391	468.180	500.000	-6.4
Decachlorobiphenyl	8.668	8.574	8.774	44.690	50.000	-10.6
Tetrachloro-m-xylene	3.675	3.579	3.779	50.650	50.000	1.3