

CALIBRATION VERIFICATION SUMMARY

Contract: PARS02

Lab Code: CHEM Case No.: Q2267 SAS No.: Q2267 SDG NO.: Q2267

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/14/2025 05/14/2025

Client Sample No.: CCAL02 Date Analyzed: 06/10/2025

Lab Sample No.: AR1660CCC500 Data File : PO111580.D Time Analyzed: 17:04

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.765	4.673	4.873	533.610	500.000	6.7
Aroclor-1016-2	4.784	4.692	4.892	537.930	500.000	7.6
Aroclor-1016-3	4.841	4.749	4.949	542.340	500.000	8.5
Aroclor-1016-4	4.960	4.869	5.069	545.430	500.000	9.1
Aroclor-1016-5	5.217	5.126	5.326	567.340	500.000	13.5
Aroclor-1260-1	6.256	6.166	6.366	588.500	500.000	17.7
Aroclor-1260-2	6.446	6.355	6.555	611.860	500.000	22.4
Aroclor-1260-3	6.813	6.722	6.922	603.330	500.000	20.7
Aroclor-1260-4	7.072	6.983	7.183	552.320	500.000	10.5
Aroclor-1260-5	7.315	7.225	7.425	559.270	500.000	11.9
Decachlorobiphenyl	8.712	8.624	8.824	47.210	50.000	-5.6
Tetrachloro-m-xylene	3.677	3.582	3.782	50.760	50.000	1.5

CALIBRATION VERIFICATION SUMMARY

Contract: PARS02

Lab Code: CHEM Case No.: Q2267 SAS No.: Q2267 SDG NO.: Q2267

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/14/2025 05/14/2025

Client Sample No.: CCAL02 Date Analyzed: 06/10/2025

Lab Sample No.: AR1660CCC500 Data File : PO111580.D Time Analyzed: 17:04

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.753	4.658	4.858	499.890	500.000	0.0
Aroclor-1016-2	4.771	4.677	4.877	501.590	500.000	0.3
Aroclor-1016-3	4.946	4.852	5.052	495.710	500.000	-0.9
Aroclor-1016-4	4.989	4.894	5.094	459.850	500.000	-8.0
Aroclor-1016-5	5.201	5.107	5.307	501.140	500.000	0.2
Aroclor-1260-1	6.230	6.138	6.338	491.800	500.000	-1.6
Aroclor-1260-2	6.419	6.326	6.526	491.220	500.000	-1.8
Aroclor-1260-3	6.570	6.479	6.679	452.850	500.000	-9.4
Aroclor-1260-4	7.041	6.949	7.149	459.980	500.000	-8.0
Aroclor-1260-5	7.282	7.191	7.391	449.980	500.000	-10.0
Decachlorobiphenyl	8.663	8.574	8.774	44.710	50.000	-10.6
Tetrachloro-m-xylene	3.675	3.579	3.779	51.630	50.000	3.3