

Surrogate Summary

SDG No.: Q2267

Client: PARSONS Engineering of New York, Inc.

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PO111057.D	PIBLK-PO111057.D	Tetrachloro-m-xylene	1	20	16.4	82		60	140
		Decachlorobiphenyl	1	20	17.5	87		60	140
		Tetrachloro-m-xylene	2	20	15.4	77		60	140
		Decachlorobiphenyl	2	20	17.9	90		60	140
I.BLK-PO111573.D	PIBLK-PO111573.D	Tetrachloro-m-xylene	1	20	21.8	109		60	140
		Decachlorobiphenyl	1	20	20.7	104		60	140
		Tetrachloro-m-xylene	2	20	20.0	100		60	140
		Decachlorobiphenyl	2	20	19.4	97		60	140
PB168377BL	PB168377BL	Tetrachloro-m-xylene	1	20	23.5	118		30	173
		Decachlorobiphenyl	1	20	22.1	111		10	173
		Tetrachloro-m-xylene	2	20	21.4	107		30	173
		Decachlorobiphenyl	2	20	20.6	103		10	173
PB168377BS	PB168377BS	Tetrachloro-m-xylene	1	20	21.8	109		30	173
		Decachlorobiphenyl	1	20	22.1	111		10	173
		Tetrachloro-m-xylene	2	20	20.0	100		30	173
		Decachlorobiphenyl	2	20	20.6	103		10	173
PB168377BSD	PB168377BSD	Tetrachloro-m-xylene	1	20	21.2	106		30	173
		Decachlorobiphenyl	1	20	22.0	110		10	173
		Tetrachloro-m-xylene	2	20	20.0	100		30	173
		Decachlorobiphenyl	2	20	20.7	104		10	173
Q2267-01	WC-20250605	Tetrachloro-m-xylene	1	20	20.3	101		30	173
		Decachlorobiphenyl	1	20	4.19	21		10	173
		Tetrachloro-m-xylene	2	20	18.4	92		30	173
		Decachlorobiphenyl	2	20	3.98	20		10	173
I.BLK-PO111584.D	PIBLK-PO111584.D	Tetrachloro-m-xylene	1	20	21.6	108		60	140
		Decachlorobiphenyl	1	20	20.3	101		60	140
		Tetrachloro-m-xylene	2	20	19.9	100		60	140
		Decachlorobiphenyl	2	20	19.2	96		60	140