

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q2267

Client: PARSONS Engineering of New York, Inc.

Analytical Method: SW8270E

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID:	Q2262-02MSD	Client Sample ID:	ARS20-0032MSD					DataFile:	BM050282.D		
Pyridine	500	0	380	ug/L	76	5			10	109	20
1,4-Dichlorobenzene	500	0	440	ug/L	88	0			55	125	20
2-Methylphenol	500	0	470	ug/L	94	2			60	131	20
3+4-Methylphenols	500	0	460	ug/L	92	4			54	136	20
Hexachloroethane	500	0	450	ug/L	90	2			19	146	20
Nitrobenzene	500	0	500	ug/L	100	2			62	112	20
Hexachlorobutadiene	500	0	490	ug/L	98	4			52	125	20
2,4,6-Trichlorophenol	500	0	530	ug/L	106	4			78	112	20
2,4,5-Trichlorophenol	500	0	530	ug/L	106	2			71	111	20
2,4-Dinitrotoluene	500	0	570	ug/L	114	0			74	137	20
Hexachlorobenzene	500	0	540	ug/L	108	4			72	115	20
Pentachlorophenol	1000	0	1100	ug/L	110	0			52	162	20