

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2267

Client: PARSONS Engineering of New York, Inc.

Analytical Method: 8270E DataFile: BM050279.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB168422BS	Pyridine	50	39.9	ug/L	80				29	97	
	1,4-Dichlorobenzene	50	42.9	ug/L	86				76	103	
	2-Methylphenol	50	44.1	ug/L	88				69	109	
	3+4-Methylphenols	50	43.8	ug/L	88				67	106	
	Hexachloroethane	50	44.9	ug/L	90				76	118	
	Nitrobenzene	50	46.8	ug/L	94				58	106	
	Hexachlorobutadiene	50	43.7	ug/L	87				69	101	
	2,4,6-Trichlorophenol	50	46.1	ug/L	92				61	110	
	2,4,5-Trichlorophenol	50	47.0	ug/L	94				70	106	
	2,4-Dinitrotoluene	50	51.7	ug/L	103				60	115	
	Hexachlorobenzene	50	47.0	ug/L	94				73	106	
	Pentachlorophenol	100	100	ug/L	100				47	114	