

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2267 SAS No.: Q2267 SDG NO.: Q2267

EPA Sample No.: SSTDCCC040 Date Analyzed: 06/12/2025

Lab File ID: BM050294.D Time Analyzed: 04:31

Instrument ID: BNA_M GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	421256	7.763	1719770	10.55	980072	14.39
UPPER LIMIT	842512	8.263	3439540	11.051	1960140	14.892
LOWER LIMIT	210628	7.263	859885	10.051	490036	13.892
EPA SAMPLE NO.						
01 PB168390TB	380994	7.76	1407710	10.55	768840	14.39

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.