

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: PARS02

Lab Code: CHEM Case No.: Q2267 SAS No.: Q2267 SDG No.: Q2267

Instrument ID: MSVOA_X Calibration Date/Time: 06/12/2025 10:11

Lab File ID: VX046657.D Init. Calib. Date(s): 06/06/2025 06/06/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 09:42 12:57

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Vinyl Chloride	0.630	0.498		-20.95	20
1,1-Dichloroethene	0.594	0.579		-2.53	20
2-Butanone	0.495	0.394		-20.4	20
Carbon Tetrachloride	0.588	0.606		3.06	20
Chloroform	1.342	1.269		-5.44	20
Benzene	1.479	1.498		1.28	20
1,2-Dichloroethane	0.615	0.566		-7.97	20
Trichloroethene	0.376	0.394		4.79	20
Tetrachloroethene	0.334	0.342		2.39	20
Chlorobenzene	1.176	1.222	0.3	3.91	20
1,2-Dichloroethane-d4	0.890	0.711		-20.11	20
Dibromofluoromethane	0.368	0.434		17.93	20
Toluene-d8	1.212	1.368		12.87	20
4-Bromofluorobenzene	0.502	0.561		11.75	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.