



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061225\
Data File : VX046656.D
Acq On : 12 Jun 2025 09:44
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060625W.M
Title : SW846 8260
Last Update : Fri Jun 06 16:56:12 2025



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.2	7240	PASS
75	95	30	60	51.0	19246	PASS
95	95	100	100	100.0	37749	PASS
96	95	5	9	6.6	2477	PASS
173	174	0.00	2	0.8	230	PASS
174	95	50	100	73.3	27683	PASS
175	174	5	9	8.1	2237	PASS
176	174	95	101	98.6	27288	PASS
177	176	5	9	6.3	1728	PASS