

Surrogate Summary

SDG No.: Q2267

Client: PARSONS Engineering of New York, Inc.

Analytical Method: SW8260D

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q2267-01	WC-20250605	1,2-Dichloroethane-d4	50	48.7	97	74	125
		Dibromofluoromethane	50	53.9	108	75	124
		Toluene-d8	50	50.1	100	86	113
		4-Bromofluorobenzene	50	53.5	107	77	121

Surrogate Summary

SDG No.: Q2267

Client: PARSONS Engineering of New York, Inc.

Analytical Method: SW8260-Low

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
VX0612WBL01	VX0612WBL01	1,2-Dichloroethane-d4	50	46.2	92	74	125
		Dibromofluoromethane	50	48.5	97	75	124
		Toluene-d8	50	52.4	105	86	113
		4-Bromofluorobenzene	50	54.6	109	77	121
VX0612WBS01	VX0612WBS01	1,2-Dichloroethane-d4	50	41.5	83	74	125
		Dibromofluoromethane	50	48.0	96	75	124
		Toluene-d8	50	47.8	96	86	113
		4-Bromofluorobenzene	50	48.1	96	77	121
VX0612WBSD0	VX0612WBSD01	1,2-Dichloroethane-d4	50	44.8	90	74	125
		Dibromofluoromethane	50	49.7	99	75	124
		Toluene-d8	50	48.3	97	86	113
		4-Bromofluorobenzene	50	50.5	101	77	121