

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2267

Client: PARSONS Engineering of New York, Inc.

Analytical Method: SW8260-Low

Datafile : VX046660.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0612WBS01	Vinyl chloride	20	15.1	ug/L	76			65	117	
	1,1-Dichloroethene	20	17.3	ug/L	86			74	110	
	2-Butanone	100	88.0	ug/L	88			65	122	
	Carbon Tetrachloride	20	17.4	ug/L	87			77	113	
	Chloroform	20	18.3	ug/L	92			79	113	
	Benzene	20	18.7	ug/L	94			82	109	
	1,2-Dichloroethane	20	17.6	ug/L	88			80	115	
	Trichloroethene	20	18.5	ug/L	93			77	113	
	Tetrachloroethene	20	18.9	ug/L	95			67	123	
	Chlorobenzene	20	19.3	ug/L	97			82	109	