

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2267

Client: PARSONS Engineering of New York, Inc.

Analytical Method: SW8260-Low

Datafile : VX046662.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0612WBSD01	Vinyl chloride	20	15.3	ug/L	77	1		65	117	19
	1,1-Dichloroethene	20	17.2	ug/L	86	0		74	110	20
	2-Butanone	100	110	ug/L	110	22		65	122	26
	Carbon Tetrachloride	20	17.2	ug/L	86	1		77	113	15
	Chloroform	20	19.3	ug/L	97	5		79	113	20
	Benzene	20	19.0	ug/L	95	1		82	109	15
	1,2-Dichloroethane	20	19.1	ug/L	96	9		80	115	20
	Trichloroethene	20	18.9	ug/L	95	2		77	113	15
	Tetrachloroethene	20	18.5	ug/L	93	2		67	123	15
	Chlorobenzene	20	19.0	ug/L	95	2		82	109	15