



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

PARSONS Engineering of New York, Inc.

Project Name: Con Ed Non MGP – Atlantic Ave 453957.600024.05

Project # N/A

Order ID # Q2268

Test Name: SVOCMS Group1

A. Number of Samples and Date of Receipt:

10 Water samples were received on 06/06/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: Sulfate, SVOCMS Group1, TDS and VOCMS Group1. This data package contains results for SVOCMS Group1.

C. Analytical Techniques:

The samples were analyzed on instrument BNA_F using GC Column DB-UI 8270D which is 20 meters, 0.18 mm ID, 0.36 um df. The analysis of SVOCMS Group1 was based on method 8270E and extraction was done based on method 3510.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria.

The Internal Standards Areas met the acceptable requirements except for MW-6-20250605, MW-3-20250605. Due to high concentration of compounds, these samples required dilution. Therefore, samples were reanalyzed with dilution and reported.

The Retention Times were acceptable for all samples.

The MS {Q2268-04MS} with File ID: BF142735.D recoveries met the requirements for all compounds except for 2,3,4,6-Tetrachlorophenol[89%] and Hexachloroethane[211%] due to matrix interference.

The MSD {Q2268-05MSD} with File ID: BF142736.D recoveries met the acceptable requirements except for 2,3,4,6-Tetrachlorophenol[85%] and Hexachloroethane[192%] due to matrix interference.

The RPD for {Q2268-05MSD} with File ID: BF142736.D met criteria except for Naphthalene[42%] due to difference in MS and MSD concentrations.

The Blank Spike met requirements for all samples .

The Blank analysis did not indicate the presence of lab contamination.



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

The Initial Calibration met the requirements.

The Continuous Calibration File ID BF142748.D met the requirements except for 2,4-Dinitrophenol. But associated samples have not positive hit for this compound therefore no corrective action was taken.

The Tuning criteria met requirements.

Samples MW-2-20250605, MW-2-20250605-A, MW-6-20250605 and MW-3-20250605 were diluted due to high concentrations.

E. Additional Comments:

Alliance has analyzed samples for SVOCMS Group1 by Method 8270 E for Project "Con Ed Non MGP – Atlantic Ave". Alliance certification was in applied status for compound "2,4-Dimethylphenol" with NJDEP for Method 8270E for SVOC group 1 at the time when samples for Project "Con Ed Non MGP – Atlantic Ave" were analyzed.

The Form 6 is not included in the data package because the Initial Calibration was performed using 8 points.

The soil samples results are based on a dry weight basis.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <20% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 20% for the Initial Calibration curve for SW-846 analysis.

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____