

LAB CHRONICLE

OrderID:	Q2268	OrderDate:	6/6/2025 2:21:00 PM
Client:	PARSONS Engineering of New York, Inc.	Project:	Con Ed Non MGP – Atlantic Ave 453957.600024.05
Contact:	Stephen Liberatore	Location:	D41,VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2268-01	MW-4-20250605	WATER	Sulfate	300.0	06/05/25 09:20		06/09/25 10:51	06/06/25
			TDS	SM2540 C			06/09/25 12:30	
Q2268-01DL	MW-4-20250605DL	WATER	Sulfate	300.0	06/05/25 09:20		06/09/25 14:48	06/06/25
Q2268-02	MW-5-20250605	WATER	Sulfate	300.0	06/05/25 10:40		06/09/25 11:12	06/06/25
			TDS	SM2540 C			06/09/25 12:30	
Q2268-02DL	MW-5-20250605DL	WATER	Sulfate	300.0	06/05/25 10:40		06/09/25 15:09	06/06/25
Q2268-03	MW-2-20250605	WATER	Sulfate	300.0	06/05/25 12:15		06/09/25 11:34	06/06/25
			TDS	SM2540 C			06/09/25 12:30	
Q2268-06	MW-2-20250605-A	WATER			06/05/25 12:15			06/06/25

LAB CHRONICLE

Sulfate	300.0	06/09/25 12:39
TDS	SM2540 C	06/09/25 12:30

Q2268-07	MW-6-20250605	WATER	06/05/25 13:50	06/06/25
-----------------	----------------------	--------------	---------------------------	-----------------

Sulfate	300.0	06/09/25 13:00
TDS	SM2540 C	06/09/25 12:30

Q2268-08	MW-3-20250605	WATER	06/05/25 14:45	06/06/25
-----------------	----------------------	--------------	---------------------------	-----------------

Sulfate	300.0	06/09/25 13:22
TDS	SM2540 C	06/09/25 12:30

Q2268-10	FB-20250605	WATER	06/05/25 15:00	06/06/25
-----------------	--------------------	--------------	---------------------------	-----------------

Sulfate	300.0	06/09/25 14:26
TDS	SM2540 C	06/09/25 12:30



SAMPLE DATA

A

B

C

D

Report of Analysis

Client:	PARSONS Engineering of New York, Inc.	Date Collected:	06/05/25 09:20
Project:	Con Ed Non MGP – Atlantic Ave 453957.600024.05	Date Received:	06/06/25
Client Sample ID:	MW-4-20250605	SDG No.:	Q2268
Lab Sample ID:	Q2268-01	Matrix:	WATER
		% Solid:	0

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
Sulfate	71.5	OR	1	0.46	3.00	mg/L		06/09/25 10:51	300.0
TDS	2140		1	1.00	10.0	mg/L		06/09/25 12:30	SM 2540 C-15

Comments:

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 D = Dilution
 Q = indicates LCS control criteria did not meet requirements
 H = Sample Analysis Out Of Hold Time

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 * = indicates the duplicate analysis is not within control limits.
 E = Indicates the reported value is estimated because of the presence of interference.
 OR = Over Range
 N = Spiked sample recovery not within control limits

Report of Analysis

Client:	PARSONS Engineering of New York, Inc.	Date Collected:	06/05/25 09:20
Project:	Con Ed Non MGP – Atlantic Ave 453957.600024.05	Date Received:	06/06/25
Client Sample ID:	MW-4-20250605DL	SDG No.:	Q2268
Lab Sample ID:	Q2268-01DL	Matrix:	WATER
		% Solid:	0

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
Sulfate	66.3	D	5	2.30	15.0	mg/L		06/09/25 14:48	300.0

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits

Report of Analysis

Client:	PARSONS Engineering of New York, Inc.	Date Collected:	06/05/25 10:40
Project:	Con Ed Non MGP – Atlantic Ave 453957.600024.05	Date Received:	06/06/25
Client Sample ID:	MW-5-20250605	SDG No.:	Q2268
Lab Sample ID:	Q2268-02	Matrix:	WATER
		% Solid:	0

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
Sulfate	101	OR	1	0.46	3.00	mg/L		06/09/25 11:12	300.0
TDS	607		1	1.00	10.0	mg/L		06/09/25 12:30	SM 2540 C-15

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits

Report of Analysis

Client:	PARSONS Engineering of New York, Inc.	Date Collected:	06/05/25 10:40
Project:	Con Ed Non MGP – Atlantic Ave 453957.600024.05	Date Received:	06/06/25
Client Sample ID:	MW-5-20250605DL	SDG No.:	Q2268
Lab Sample ID:	Q2268-02DL	Matrix:	WATER
		% Solid:	0

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
Sulfate	91.6	D	10	4.60	30.0	mg/L		06/09/25 15:09	300.0

Comments: _____

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits

Report of Analysis

Client:	PARSONS Engineering of New York, Inc.	Date Collected:	06/05/25 12:15
Project:	Con Ed Non MGP – Atlantic Ave 453957.600024.05	Date Received:	06/06/25
Client Sample ID:	MW-2-20250605	SDG No.:	Q2268
Lab Sample ID:	Q2268-03	Matrix:	WATER
		% Solid:	0

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
Sulfate	12.4		1	0.46	3.00	mg/L		06/09/25 11:34	300.0
TDS	795		1	1.00	10.0	mg/L		06/09/25 12:30	SM 2540 C-15

Comments:

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 D = Dilution
 Q = indicates LCS control criteria did not meet requirements
 H = Sample Analysis Out Of Hold Time

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 * = indicates the duplicate analysis is not within control limits.
 E = Indicates the reported value is estimated because of the presence of interference.
 OR = Over Range
 N =Spiked sample recovery not within control limits

Report of Analysis

Client:	PARSONS Engineering of New York, Inc.	Date Collected:	06/05/25 12:15
Project:	Con Ed Non MGP – Atlantic Ave 453957.600024.05	Date Received:	06/06/25
Client Sample ID:	MW-2-20250605-A	SDG No.:	Q2268
Lab Sample ID:	Q2268-06	Matrix:	WATER
		% Solid:	0

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
Sulfate	12.3		1	0.46	3.00	mg/L		06/09/25 12:39	300.0
TDS	816		1	1.00	10.0	mg/L		06/09/25 12:30	SM 2540 C-15

Comments:

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 D = Dilution
 Q = indicates LCS control criteria did not meet requirements
 H = Sample Analysis Out Of Hold Time

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 * = indicates the duplicate analysis is not within control limits.
 E = Indicates the reported value is estimated because of the presence of interference.
 OR = Over Range
 N =Spiked sample recovery not within control limits

Report of Analysis

Client:	PARSONS Engineering of New York, Inc.	Date Collected:	06/05/25 13:50
Project:	Con Ed Non MGP – Atlantic Ave 453957.600024.05	Date Received:	06/06/25
Client Sample ID:	MW-6-20250605	SDG No.:	Q2268
Lab Sample ID:	Q2268-07	Matrix:	WATER
		% Solid:	0

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
Sulfate	20.6		1	0.46	3.00	mg/L		06/09/25 13:00	300.0
TDS	1440		1	1.00	10.0	mg/L		06/09/25 12:30	SM 2540 C-15

Comments: _____

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits

Report of Analysis

Client:	PARSONS Engineering of New York, Inc.	Date Collected:	06/05/25 14:45
Project:	Con Ed Non MGP – Atlantic Ave 453957.600024.05	Date Received:	06/06/25
Client Sample ID:	MW-3-20250605	SDG No.:	Q2268
Lab Sample ID:	Q2268-08	Matrix:	WATER
		% Solid:	0

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
Sulfate	14.1		1	0.46	3.00	mg/L		06/09/25 13:22	300.0
TDS	538		1	1.00	10.0	mg/L		06/09/25 12:30	SM 2540 C-15

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits

Report of Analysis

Client:	PARSONS Engineering of New York, Inc.	Date Collected:	06/05/25 15:00
Project:	Con Ed Non MGP – Atlantic Ave 453957.600024.05	Date Received:	06/06/25
Client Sample ID:	FB-20250605	SDG No.:	Q2268
Lab Sample ID:	Q2268-10	Matrix:	WATER
		% Solid:	0

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
Sulfate	0.46	U	1	0.46	3.00	mg/L		06/09/25 14:26	300.0
TDS	1.00	U	1	1.00	10.0	mg/L		06/09/25 12:30	SM 2540 C-15

Comments: _____

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits

QC RESULT SUMMARY

Initial and Continuing Calibration Verification

Client: PARSONS Engineering of New York, Inc.

SDG No.: Q2268

Project: Con Ed Non MGP – Atlantic Ave 453957.600024.05

RunNo.: LB136070

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1						
Bromide	mg/L	10.2	10	102	90-110	05/22/2025
Chloride	mg/L	3.1	3	103	90-110	05/22/2025
Fluoride	mg/L	2.1	2	105	90-110	05/22/2025
Nitrite	mg/L	3.1	3	103	90-110	05/22/2025
Nitrate	mg/L	2.6	2.5	104	90-110	05/22/2025
Sulfate	mg/L	15.1	15	101	90-110	05/22/2025
Orthophosphate as P	mg/L	5.3	5	106	90-110	05/22/2025
Sample ID: CCV1						
Bromide	mg/L	10.4	10	104	90-110	06/09/2025
Chloride	mg/L	3.1	3	103	90-110	06/09/2025
Fluoride	mg/L	2.1	2	105	90-110	06/09/2025
Nitrite	mg/L	3.1	3	103	90-110	06/09/2025
Nitrate	mg/L	2.6	2.5	104	90-110	06/09/2025
Sulfate	mg/L	15.4	15	103	90-110	06/09/2025
Orthophosphate as P	mg/L	5.1	5	102	90-110	06/09/2025
Sample ID: CCV2						
Bromide	mg/L	10.3	10	103	90-110	06/09/2025
Chloride	mg/L	3.1	3	103	90-110	06/09/2025
Fluoride	mg/L	2.1	2	105	90-110	06/09/2025
Nitrite	mg/L	3.1	3	103	90-110	06/09/2025
Nitrate	mg/L	2.6	2.5	104	90-110	06/09/2025
Sulfate	mg/L	15.2	15	101	90-110	06/09/2025
Orthophosphate as P	mg/L	5.3	5	106	90-110	06/09/2025
Sample ID: CCV3						
Bromide	mg/L	10.3	10	103	90-110	06/09/2025
Chloride	mg/L	3.1	3	103	90-110	06/09/2025
Fluoride	mg/L	2.1	2	105	90-110	06/09/2025
Nitrite	mg/L	3.1	3	103	90-110	06/09/2025
Nitrate	mg/L	2.6	2.5	104	90-110	06/09/2025
Sulfate	mg/L	15.3	15	102	90-110	06/09/2025
Orthophosphate as P	mg/L	5.3	5	106	90-110	06/09/2025

Initial and Continuing Calibration Blank Summary

Client: PARSONS Engineering of New York, Inc.

SDG No.: Q2268

Project: Con Ed Non MGP – Atlantic Ave 453957.600024.05

RunNo.: LB136070

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1							
Bromide	mg/L	< 1.0000	1.0000	U	0.37	2	05/22/2025
Chloride	mg/L	< 0.3000	0.3000	U	0.19	0.6	05/22/2025
Fluoride	mg/L	< 0.2000	0.2000	U	0.11	0.4	05/22/2025
Nitrite	mg/L	< 0.3000	0.3000	U	0.074	0.6	05/22/2025
Nitrate	mg/L	< 0.2500	0.2500	U	0.095	0.5	05/22/2025
Sulfate	mg/L	< 1.5000	1.5000	U	0.46	3	05/22/2025
Orthophosphate as P	mg/L	< 0.5000	0.5000	U	0.34	1	05/22/2025
Sample ID: CCB1							
Bromide	mg/L	< 1.0000	1.0000	U	0.37	2	06/09/2025
Chloride	mg/L	< 0.3000	0.3000	U	0.19	0.6	06/09/2025
Fluoride	mg/L	< 0.2000	0.2000	U	0.11	0.4	06/09/2025
Nitrite	mg/L	< 0.3000	0.3000	U	0.074	0.6	06/09/2025
Nitrate	mg/L	< 0.2500	0.2500	U	0.095	0.5	06/09/2025
Sulfate	mg/L	< 1.5000	1.5000	U	0.46	3	06/09/2025
Orthophosphate as P	mg/L	< 0.5000	0.5000	U	0.34	1	06/09/2025
Sample ID: CCB2							
Bromide	mg/L	< 1.0000	1.0000	U	0.37	2	06/09/2025
Chloride	mg/L	< 0.3000	0.3000	U	0.19	0.6	06/09/2025
Fluoride	mg/L	< 0.2000	0.2000	U	0.11	0.4	06/09/2025
Nitrite	mg/L	< 0.3000	0.3000	U	0.074	0.6	06/09/2025
Nitrate	mg/L	< 0.2500	0.2500	U	0.095	0.5	06/09/2025
Sulfate	mg/L	< 1.5000	1.5000	U	0.46	3	06/09/2025
Orthophosphate as P	mg/L	< 0.5000	0.5000	U	0.34	1	06/09/2025
Sample ID: CCB3							
Bromide	mg/L	< 1.0000	1.0000	U	0.37	2	06/09/2025
Chloride	mg/L	< 0.3000	0.3000	U	0.19	0.6	06/09/2025
Fluoride	mg/L	< 0.2000	0.2000	U	0.11	0.4	06/09/2025
Nitrite	mg/L	< 0.3000	0.3000	U	0.074	0.6	06/09/2025
Nitrate	mg/L	< 0.2500	0.2500	U	0.095	0.5	06/09/2025
Sulfate	mg/L	< 1.5000	1.5000	U	0.46	3	06/09/2025
Orthophosphate as P	mg/L	< 0.5000	0.5000	U	0.34	1	06/09/2025

Preparation Blank Summary

Client: PARSONS Engineering of New York, Inc.

SDG No.: Q2268

Project: Con Ed Non MGP – Atlantic Ave 453957.600024.05

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: LB136070BLW							
Bromide	mg/L	< 1.0000	1.0000	U	0.37	2	06/09/2025
Chloride	mg/L	< 0.3000	0.3000	U	0.19	0.6	06/09/2025
Fluoride	mg/L	< 0.2000	0.2000	U	0.11	0.4	06/09/2025
Nitrite	mg/L	< 0.3000	0.3000	U	0.074	0.6	06/09/2025
Nitrate	mg/L	< 0.2500	0.2500	U	0.095	0.5	06/09/2025
Sulfate	mg/L	< 1.5000	1.5000	U	0.46	3	06/09/2025
Orthophosphate as P	mg/L	< 0.5000	0.5000	U	0.34	1	06/09/2025
Sample ID: LB136082BL							
TDS	mg/L	< 5.0000	5.0000	U	1.0	10	06/09/2025

Matrix Spike Summary

Client:	PARSONS Engineering of New York, Inc.	SDG No.:	Q2268
Project:	Con Ed Non MGP – Atlantic Ave 453957.600024.05	Sample ID:	Q2268-03
Client ID:	MW-2-20250605MS	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Bromide	mg/L	80-120	10.6	OR	0.64	J	10	1	100		06/09/2025
Chloride	mg/L	80-120	409		426	OR	3	1	-567	*	06/09/2025
Fluoride	mg/L	80-120	2.20		0.16	J	2	1	102		06/09/2025
Nitrite	mg/L	80-120	3.00		0.074	U	3	1	100		06/09/2025
Nitrate	mg/L	80-120	2.60		0.17	J	2.5	1	97		06/09/2025
Sulfate	mg/L	80-120	27.0		12.4		15	1	97		06/09/2025
Orthophosphate as P	mg/L	80-120	5.00		0.34	U	5	1	100		06/09/2025

Matrix Spike Summary

Client:	PARSONS Engineering of New York, Inc.	SDG No.:	Q2268
Project:	Con Ed Non MGP – Atlantic Ave 453957.600024.05	Sample ID:	Q2268-03
Client ID:	MW-2-20250605MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Bromide	mg/L	80-120	10.9	OR	0.64	J	10	1	103		06/09/2025
Chloride	mg/L	80-120	408		426	OR	3	1	-600	*	06/09/2025
Fluoride	mg/L	80-120	2.20		0.16	J	2	1	102		06/09/2025
Nitrite	mg/L	80-120	3.10		0.074	U	3	1	103		06/09/2025
Nitrate	mg/L	80-120	2.70		0.17	J	2.5	1	101		06/09/2025
Sulfate	mg/L	80-120	27.4		12.4		15	1	100		06/09/2025
Orthophosphate as P	mg/L	80-120	5.40		0.34	U	5	1	108		06/09/2025

Duplicate Sample Summary

Client:	PARSONS Engineering of New York, Inc.	SDG No.:	Q2268
Project:	Con Ed Non MGP – Atlantic Ave 453957.600024.05	Sample ID:	Q2268-03
Client ID:	MW-2-20250605DUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
TDS	mg/L	+/-5	795		805		1	1.25		06/09/2025

Duplicate Sample Summary

Client:	PARSONS Engineering of New York, Inc.	SDG No.:	Q2268
Project:	Con Ed Non MGP – Atlantic Ave 453957.600024.05	Sample ID:	Q2268-03
Client ID:	MW-2-20250605MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
Fluoride	mg/L	+/-20	2.20	OR	2.20	OR	1	0		06/09/2025
Chloride	mg/L	+/-20	409		408		1	0		06/09/2025
Sulfate	mg/L	+/-20	27.0		27.4		1	1		06/09/2025
Bromide	mg/L	+/-20	10.6		10.9		1	3		06/09/2025
Nitrite	mg/L	+/-20	3.00		3.10		1	3		06/09/2025
Nitrate	mg/L	+/-20	2.60		2.70		1	4		06/09/2025
Orthophosphate as P	mg/L	+/-20	5.00		5.40		1	8		06/09/2025

Laboratory Control Sample Summary

Client:	PARSONS Engineering of New York, Inc.	SDG No.:	Q2268
Project:	Con Ed Non MGP – Atlantic Ave 453957.600024.05	Run No.:	LB136070

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136070BSW							
Bromide	mg/L	10	10.4		104	1	90-110	06/09/2025
Chloride	mg/L	3	3.10		103	1	90-110	06/09/2025
Fluoride	mg/L	2	2.10		105	1	90-110	06/09/2025
Nitrite	mg/L	3	3.10		103	1	90-110	06/09/2025
Nitrate	mg/L	2.5	2.60		104	1	90-110	06/09/2025
Sulfate	mg/L	15	15.4		103	1	90-110	06/09/2025
Orthophosphate as P	mg/L	5	5.30		106	1	90-110	06/09/2025

Laboratory Control Sample Summary

Client:	PARSONS Engineering of New York, Inc.	SDG No.:	Q2268
Project:	Con Ed Non MGP – Atlantic Ave 453957.600024.05	Run No.:	LB136082

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136082BS							
TDS	mg/L	100	95.0		95	1	90-110	06/09/2025