

Laboratory Control Sample Summary

Client:	PARSONS Engineering of New York, Inc.	SDG No.:	Q2268
Project:	Con Ed Non MGP – Atlantic Ave 453957.600024.05	Run No.:	LB136070

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136070BSW							
Bromide	mg/L	10	10.4		104	1	90-110	06/09/2025
Chloride	mg/L	3	3.10		103	1	90-110	06/09/2025
Fluoride	mg/L	2	2.10		105	1	90-110	06/09/2025
Nitrite	mg/L	3	3.10		103	1	90-110	06/09/2025
Nitrate	mg/L	2.5	2.60		104	1	90-110	06/09/2025
Sulfate	mg/L	15	15.4		103	1	90-110	06/09/2025
Orthophosphate as P	mg/L	5	5.30		106	1	90-110	06/09/2025

Laboratory Control Sample Summary

Client:	PARSONS Engineering of New York, Inc.	SDG No.:	Q2268
Project:	Con Ed Non MGP – Atlantic Ave 453957.600024.05	Run No.:	LB136082

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136082BS							
TDS	mg/L	100	95.0		95	1	90-110	06/09/2025