

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: PARS02

Lab Code: CHEM Case No.: Q2268 SAS No.: Q2268 SDG No.: Q2268

Instrument ID: BNA_F Calibration Date/Time: 06/11/2025 09:24

Lab File ID: BF142723.D Init. Calib. Date(s): 06/10/2025 06/10/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 16:54 20:19

GC Column: DB-UI ID: 0.18 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
2-Fluorophenol	1.174	1.141		-2.8	
Benzaldehyde	0.856	0.818		-4.4	
Phenol-d6	1.382	1.367		-1.1	
Phenol	1.536	1.528		-0.5	20.0
bis(2-Chloroethyl)ether	1.147	1.136		-1.0	
2-Chlorophenol	1.283	1.259		-1.9	
2-Methylphenol	0.984	0.985		0.1	
2,2-oxybis(1-Chloropropane)	1.806	1.799		-0.4	
Acetophenone	0.445	0.431		-3.1	
3+4-Methylphenols	1.240	1.242		0.2	
n-Nitroso-di-n-propylamine	0.880	0.889	0.050	1.0	
Nitrobenzene-d5	0.365	0.358		-1.9	
Hexachloroethane	0.530	0.508		-4.2	
Nitrobenzene	0.324	0.320		-1.2	
Isophorone	0.619	0.614		-0.8	
2-Nitrophenol	0.181	0.182		0.6	20.0
2,4-Dimethylphenol	0.308	0.305		-1.0	
bis(2-Chloroethoxy)methane	0.384	0.380		-1.0	
2,4-Dichlorophenol	0.284	0.284		0.0	20.0
Naphthalene	0.989	0.976		-1.3	
4-Chloroaniline	0.397	0.399		0.5	
Hexachlorobutadiene	0.200	0.197		-1.5	20.0
Caprolactam	0.077	0.080		3.9	
4-Chloro-3-methylphenol	0.296	0.295		-0.3	20.0
2-Methylnaphthalene	0.626	0.619		-1.1	
Hexachlorocyclopentadiene	0.372	0.369	0.050	-0.8	
2,4,6-Trichlorophenol	0.375	0.381		1.6	20.0
2-Fluorobiphenyl	1.505	1.452		-3.5	
2,4,5-Trichlorophenol	0.405	0.395		-2.5	
1,1-Biphenyl	1.553	1.516		-2.4	
2-Chloronaphthalene	1.144	1.134		-0.9	
2-Nitroaniline	0.325	0.326		0.3	
Dimethylphthalate	1.336	1.325		-0.8	
Acenaphthylene	1.929	1.891		-2.0	
2,6-Dinitrotoluene	0.289	0.285		-1.4	
3-Nitroaniline	0.313	0.309		-1.3	
Acenaphthene	1.196	1.172		-2.0	20.0
2,4-Dinitrophenol	0.158	0.159	0.050	0.6	
4-Nitrophenol	0.212	0.215	0.050	1.4	

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: PARS02

Lab Code: CHEM Case No.: Q2268 SAS No.: Q2268 SDG No.: Q2268

Instrument ID: BNA_F Calibration Date/Time: 06/11/2025 09:24

Lab File ID: BF142723.D Init. Calib. Date(s): 06/10/2025 06/10/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 16:54 20:19

GC Column: DB-UI ID: 0.18 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
Dibenzofuran	1.703	1.673		-1.8	
2,4-Dinitrotoluene	0.382	0.392		2.6	
Diethylphthalate	1.324	1.318		-0.5	
4-Chlorophenyl-phenylether	0.657	0.651		-0.9	
Fluorene	1.345	1.301		-3.3	
4-Nitroaniline	0.280	0.290		3.6	
4,6-Dinitro-2-methylphenol	0.125	0.125		0.0	
n-Nitrosodiphenylamine	0.687	0.667		-2.9	20.0
2,4,6-Tribromophenol	0.219	0.218		-0.5	
4-Bromophenyl-phenylether	0.236	0.233		-1.3	
Hexachlorobenzene	0.262	0.255		-2.7	
Atrazine	0.183	0.192		4.9	
Pentachlorophenol	0.137	0.138		0.7	20.0
Phenanthrene	1.071	1.043		-2.6	
Anthracene	1.108	1.070		-3.4	
Carbazole	0.928	0.921		-0.8	
Di-n-butylphthalate	1.036	1.101		6.3	
Fluoranthene	1.019	1.017		-0.2	20.0
Pyrene	1.859	1.904		2.4	
Terphenyl-d14	1.456	1.502		3.2	
Butylbenzylphthalate	0.550	0.589		7.1	
3,3-Dichlorobenzidine	0.427	0.407		-4.7	
Benzo (a) anthracene	1.325	1.330		0.4	
Chrysene	1.224	1.168		-4.6	
Bis(2-ethylhexyl)phthalate	0.825	0.821		-0.5	
Di-n-octyl phthalate	1.582	1.500		-5.2	20.0
Benzo (b) fluoranthene	1.178	1.192		1.2	
Benzo (k) fluoranthene	1.143	1.132		-1.0	
Benzo (a) pyrene	1.120	1.112		-0.7	20.0
Indeno (1,2,3-cd) pyrene	1.482	1.501		1.3	
Dibenzo (a,h) anthracene	1.210	1.240		2.5	
Benzo (g,h,i) perylene	1.203	1.196		-0.6	
1,2,4,5-Tetrachlorobenzene	0.579	0.567		-2.1	
1,4-Dioxane	0.465	0.450		-3.2	20.0
2,3,4,6-Tetrachlorophenol	0.340	0.335		-1.5	

All other compounds must meet a minimum RRF of 0.010.