

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2268 SAS No.: Q2268 SDG NO.: Q2268

EPA Sample No.: SSTDCCC040 Date Analyzed: 06/12/2025

Lab File ID: BF142748.D Time Analyzed: 09:50

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	201441	11.422	120471	14.068	146263	15.562
UPPER LIMIT	402882	11.922	240942	14.568	292526	16.062
LOWER LIMIT	100721	10.922	60235.5	13.568	73131.5	15.062
EPA SAMPLE NO.						
01 MW-2-20250605-ADL	217678	11.42	118180	14.06	146576	15.56
02 MW-6-20250605DL	172881	11.42	110385	14.06	149783	15.56
03 MW-3-20250605DL	175442	11.42	109504	14.06	154313	15.56

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.