

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: PARS02

Lab Code: CHEM Case No.: Q2268 SAS No.: Q2268 SDG No.: Q2268

Instrument ID: MSVOA_N Calibration Date/Time: 06/16/2025 09:48

Lab File ID: VN087017.D Init. Calib. Date(s): 06/06/2025 06/06/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 12:44 14:49

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.499	0.569		14.03	20
Chloromethane	0.644	0.574	0.1	-10.87	20
Vinyl Chloride	0.664	0.777		17.02	20
Bromomethane	0.372	0.322		-13.44	20
Chloroethane	0.429	0.487		13.52	20
Trichlorofluoromethane	0.868	0.959		10.48	20
1,1,2-Trichlorotrifluoroethane	0.545	0.599		9.91	20
1,1-Dichloroethene	0.557	0.625		12.21	20
Acetone	0.355	0.340		-4.22	20
Carbon Disulfide	1.539	1.877		21.96	20
Methyl tert-butyl Ether	2.015	2.103		4.37	20
Methyl Acetate	1.035	0.839		-18.94	20
Methylene Chloride	0.665	0.691		3.91	20
trans-1,2-Dichloroethene	0.619	0.673		8.72	20
1,1-Dichloroethane	1.120	1.186	0.1	5.89	20
Cyclohexane	1.086	1.066		-1.84	20
2-Butanone	0.577	0.514		-10.92	20
Carbon Tetrachloride	0.433	0.476		9.93	20
cis-1,2-Dichloroethene	0.740	0.796		7.57	20
Bromochloromethane	0.550	0.487		-11.45	20
Chloroform	1.118	1.158		3.58	20
1,1,1-Trichloroethane	0.951	0.991		4.21	20
Methylcyclohexane	0.605	0.618		2.15	20
Benzene	1.444	1.587		9.9	20
1,2-Dichloroethane	0.438	0.466		6.39	20
Trichloroethene	0.342	0.388		13.45	20
1,2-Dichloropropane	0.351	0.382		8.83	20
Bromodichloromethane	0.480	0.521		8.54	20
4-Methyl-2-Pentanone	0.543	0.536		-1.29	20
Toluene	0.882	0.980		11.11	20
t-1,3-Dichloropropene	0.537	0.599		11.55	20
cis-1,3-Dichloropropene	0.574	0.646		12.54	20
1,1,2-Trichloroethane	0.340	0.364		7.06	20
2-Hexanone	0.350	0.360		2.86	20
Dibromochloromethane	0.354	0.402		13.56	20
1,2-Dibromoethane	0.348	0.384		10.35	20
Tetrachloroethene	0.316	0.360		13.92	20
Chlorobenzene	1.103	1.247	0.3	13.06	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: PARS02

Lab Code: CHEM Case No.: Q2268 SAS No.: Q2268 SDG No.: Q2268

Instrument ID: MSVOA_N Calibration Date/Time: 06/16/2025 09:48

Lab File ID: VN087017.D Init. Calib. Date(s): 06/06/2025 06/06/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 12:44 14:49

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.900	2.085		9.74	20
m/p-Xylenes	0.727	0.842		15.82	20
o-Xylene	0.696	0.801		15.09	20
Styrene	1.192	1.369		14.85	20
Bromoform	0.263	0.310	0.1	17.87	20
Isopropylbenzene	3.643	3.951		8.45	20
1,1,2,2-Tetrachloroethane	1.234	1.321	0.3	7.05	20
1,3-Dichlorobenzene	1.644	1.844		12.16	20
1,4-Dichlorobenzene	1.676	1.861		11.04	20
1,2-Dichlorobenzene	1.579	1.743		10.39	20
1,2-Dibromo-3-Chloropropane	0.295	0.291		-1.36	20
1,2,4-Trichlorobenzene	1.008	1.007		-0.1	20
1,2,3-Trichlorobenzene	1.002	0.957		-4.49	20
1,2-Dichloroethane-d4	0.669	0.560		-16.29	20
Dibromofluoromethane	0.296	0.282		-4.73	20
Toluene-d8	1.173	1.051		-10.4	20
4-Bromofluorobenzene	0.436	0.397		-8.94	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.