

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PARS02
 Lab Code: CHEM Case No.: Q2268 SAS No.: Q2268 SDG NO.: Q2268
 Lab File ID: VN086967.D Date Analyzed: 06/12/2025
 Instrument ID: MSVOA_N Time Analyzed: 10:58
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	185417	8.23	334881	9.11	291913	11.87
UPPER LIMIT	370834	8.729	669762	9.606	583826	12.365
LOWER LIMIT	92708.5	7.729	167441	8.606	145957	11.365
EPA SAMPLE NO.						
MW-4-20250605	290913	8.24	573001	9.11	515947	11.87
MW-5-20250605	244434	8.24	483373	9.11	434597	11.87
MW-2-20250605	259006	8.23	488303	9.11	432945	11.87
MW-2-20250605MS	232985	8.24	427205	9.11	372041	11.87
MW-2-20250605MSD	173076	8.24	307412	9.11	267750	11.87
MW-2-20250605-A	287973	8.24	564200	9.11	499730	11.87
MW-6-20250605	198700	8.23	383576	9.11	340087	11.87
MW-3-20250605	240809	8.24	472342	9.11	416914	11.87
TB-20250605	283433	8.24	552519	9.11	494879	11.87
FB-20250605	274527	8.24	543078	9.11	482706	11.87
VN0612WBL01	293212	8.23	559485	9.11	496764	11.87
VN0612WBS01	173273	8.24	314156	9.11	274906	11.87

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.