

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PARS02
 Lab Code: CHEM Case No.: Q2268 SAS No.: Q2268 SDG NO.: Q2268
 Lab File ID: VN087017.D Date Analyzed: 06/16/2025
 Instrument ID: MSVOA_N Time Analyzed: 09:48
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	190027	8.23	334516	9.11	287112	11.87
UPPER LIMIT	380054	8.73	669032	9.606	574224	12.365
LOWER LIMIT	95013.5	7.73	167258	8.606	143556	11.365
EPA SAMPLE NO.						
MW-6-20250605DL	253695	8.23	507247	9.11	456316	11.87
MW-3-20250605DL	263151	8.23	520533	9.11	469764	11.87
VN0616WBL01	283331	8.23	556820	9.11	493060	11.87
VN0616WBS01	198479	8.23	351201	9.11	304875	11.87

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.