

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PARS02
 Lab Code: CHEM Case No.: Q2268 SAS No.: Q2268 SDG NO.: Q2268
 Lab File ID: VN086967.D Date Analyzed: 06/12/2025
 Instrument ID: MSVOA_N Time Analyzed: 10:58
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	144672	13.794				
UPPER LIMIT	289344	14.294				
LOWER LIMIT	72336	13.294				
EPA SAMPLE NO.						
MW-4-20250605	245935	13.79				
MW-5-20250605	206986	13.79				
MW-2-20250605	195427	13.79				
MW-2-20250605MS	171326	13.79				
MW-2-20250605MSD	123846	13.79				
MW-2-20250605-A	234750	13.79				
MW-6-20250605	149213	13.79				
MW-3-20250605	170552	13.79				
TB-20250605	234248	13.79				
FB-20250605	228705	13.79				
VN0612WBL01	239856	13.79				
VN0612WBS01	136822	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.