

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PARS02
 Lab Code: CHEM Case No.: Q2268 SAS No.: Q2268 SDG NO.: Q2268
 Lab File ID: VN087017.D Date Analyzed: 06/16/2025
 Instrument ID: MSVOA_N Time Analyzed: 09:48
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	143033	13.788				
UPPER LIMIT	286066	14.288				
LOWER LIMIT	71516.5	13.288				
EPA SAMPLE NO.						
MW-6-20250605DL	217980	13.79				
MW-3-20250605DL	224894	13.79				
VN0616WBL01	235789	13.79				
VN0616WBS01	154394	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.