

Surrogate Summary

SW-846

SDG No.: Q2275

Client: JACOBS Engineering Group, Inc.

Analytical Method: 8270-Modified

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB168391BL	PB168391BL	2-Methylnaphthalene-d10	0.4	0.32	81		30 (20)	150 (139)
		Fluoranthene-d10	0.4	0.41	103		30 (54)	150 (157)
		Nitrobenzene-d5	0.4	0.26	64		30 (27)	130 (154)
		2-Fluorobiphenyl	0.4	0.28	69		30 (30)	130 (155)
		Terphenyl-d14	0.4	0.35	88		30 (54)	130 (175)
PB168391BS	PB168391BS	2-Methylnaphthalene-d10	0.4	0.39	97		30 (20)	150 (139)
		Fluoranthene-d10	0.4	0.34	85		30 (54)	150 (157)
		Nitrobenzene-d5	0.4	0.36	90		30 (27)	130 (154)
		2-Fluorobiphenyl	0.4	0.37	92		30 (30)	130 (155)
		Terphenyl-d14	0.4	0.37	93		30 (54)	130 (175)
PB168391BSD	PB168391BSD	2-Methylnaphthalene-d10	0.4	0.37	91		30 (20)	150 (139)
		Fluoranthene-d10	0.4	0.34	85		30 (54)	150 (157)
		Nitrobenzene-d5	0.4	0.35	86		30 (27)	130 (154)
		2-Fluorobiphenyl	0.4	0.38	95		30 (30)	130 (155)
		Terphenyl-d14	0.4	0.37	92		30 (54)	130 (175)
Q2275-01	OW-08B-72.5-060925	2-Methylnaphthalene-d10	0.4	0.33	83		30 (20)	150 (139)
		Fluoranthene-d10	0.4	0.39	97		30 (54)	150 (157)
		Nitrobenzene-d5	0.4	0.26	66		30 (27)	130 (154)
		2-Fluorobiphenyl	0.4	0.37	92		30 (30)	130 (155)
		Terphenyl-d14	0.4	0.46	115		30 (54)	130 (175)
Q2275-03	EB01-060925	2-Methylnaphthalene-d10	0.4	0.36	90		30 (20)	150 (139)
		Fluoranthene-d10	0.4	0.40	99		30 (54)	150 (157)
		Nitrobenzene-d5	0.4	0.35	87		30 (27)	130 (154)
		2-Fluorobiphenyl	0.4	0.42	105		30 (30)	130 (155)
		Terphenyl-d14	0.4	0.54	134	*	30 (54)	130 (175)