

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: JAC005

Lab Code: CHEM Case No.: Q2275 SAS No.: Q2275 SDG No.: Q2275

Instrument ID: MSVOA_V Calibration Date/Time: 06/18/2025 10:19

Lab File ID: VV038811.D Init. Calib. Date(s): 06/16/2025 06/16/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 09:28 11:36

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRFCAL	MIN RRF	%D	MAX%D
Vinyl chloride	0.900	0.928	0.01	3.1	30
Vinyl Chloride-d3	0.475	0.512	0.01	7.8	30
1,2-Dichloroethane-d4	0.233	0.243	0.01	4.1	25
1,2-Dichloropropane-d6	0.283	0.325	0.1	14.7	20
Toluene-d8	0.572	0.630	0.2	10.2	20
1,1,2,2-Tetrachloroethane-d2	0.199	0.216	0.01	8.7	25

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.