

Surrogate Summary

SDG No.: Q2275

Client: JACOBS Engineering Group, Inc.

Analytical Method: SFAM_VOCSIM

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q2275-02	OW-08B-72.5-060925-SIM	Vinyl chloride-d3	0.5	0.49	98	40	130
		1,2-Dichloroethane-d4	0.5	0.50	100	70	130
		1,2-Dichloropropane-d6	0.5	0.53	107	60	140
		Toluene-d8	0.5	0.50	100	70	130
		1,1,2,2-Tetrachloroethane-d2	0.5	0.51	102	65	120
Q2275-04	EB01-060925-SIM	Vinyl chloride-d3	0.5	0.52	105	40	130
		1,2-Dichloroethane-d4	0.5	0.52	105	70	130
		1,2-Dichloropropane-d6	0.5	0.54	108	60	140
		Toluene-d8	0.5	0.52	104	70	130
		1,1,2,2-Tetrachloroethane-d2	0.5	0.50	101	65	120
VV0618WBL01	VV0618WBL01	Vinyl chloride-d3	0.5	0.43	87	40	130
		1,2-Dichloroethane-d4	0.5	0.49	99	70	130
		1,2-Dichloropropane-d6	0.5	0.53	106	60	140
		Toluene-d8	0.5	0.54	107	70	130
		1,1,2,2-Tetrachloroethane-d2	0.5	0.47	94	65	120