

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: JACO05
Lab Code: CHEM Case No.: Q2275 SAS No.: Q2275 SDG NO.: Q2275
Lab File ID: VV038811.D Date Analyzed: 06/18/2025
Instrument ID: MSVOA_V Time Analyzed: 10:19
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 (DFB) AREA #	RT #	IS2 (CBZ) AREA #	RT #	IS3 (DCB) AREA #	RT #
12 HOUR STD	6326	5.57	5881	8.79	2552	11.19
UPPER LIMIT	12652	5.736	11762	8.955	5104	11.361
LOWER LIMIT	3163	5.396	2940.5	8.615	1276	11.021
EPA SAMPLE NO.						
OW-08B-72.5-060925-SIM	6485	5.57	5976	8.79	2595	11.19
EB01-060925-SIM	6087	5.57	5929	8.80	2589	11.19
VBLK228	7230	5.57	6612	8.79	2640	11.19

IS1 (DFB) = 1,4-Difluorobenzene

IS2 (CBZ) = Chlorobenzene-d5

IS3 (DCB) = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.