

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: JACO05
 Lab Code: CHEM Case No.: Q2275 SAS No.: Q2275 SDG NO.: Q2275
 Lab File ID: VX046657.D Date Analyzed: 06/12/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:11
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	60636	12.018				
UPPER LIMIT	121272	12.518				
LOWER LIMIT	30318	11.518				
EPA SAMPLE NO.						
OW-08B-72.5-060925	61889	12.02				
EB01-060925	59204	12.02				
TB01-060925	70754	12.02				
VX0612WBL01	95358	12.02				
VX0612WBS01	69016	12.02				
VX0612WBSD01	65393	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.