

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: JACO05
Lab Code: CHEM Case No.: Q2275 SAS No.: Q2275 SDG NO.: Q2275
Lab File ID: VX046516.D BFB Injection Date: 06/06/2025
Instrument ID: MSVOA_X BFB Injection Time: 08:47
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	21
75	30.0 - 60.0% of mass 95	56.5
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.5
173	Less than 2.0% of mass 174	0.6 (1) 1
174	50.0 - 100.0% of mass 95	65.2
175	5.0 - 9.0% of mass 174	5 (7.7) 1
176	95.0 - 101.0% of mass 174	62.7 (96.1) 1
177	5.0 - 9.0% of mass 176	4.2 (6.6) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDIC005	VSTDIC005	VX046518.D	06/06/2025	09:42
VSTDIC020	VSTDIC020	VX046519.D	06/06/2025	10:18
VSTDIC050	VSTDIC050	VX046520.D	06/06/2025	10:40
VSTDIC100	VSTDIC100	VX046521.D	06/06/2025	11:02
VSTDIC150	VSTDIC150	VX046522.D	06/06/2025	11:25
VSTDIC001	VSTDIC001	VX046524.D	06/06/2025	12:57