

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: JACO05
Lab Code: CHEM Case No.: Q2275 SAS No.: Q2275 SDG NO.: Q2275
Lab File ID: VX046656.D BFB Injection Date: 06/12/2025
Instrument ID: MSVOA_X BFB Injection Time: 09:44
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	19.2
75	30.0 - 60.0% of mass 95	51
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.6
173	Less than 2.0% of mass 174	0.6 (0.8) 1
174	50.0 - 100.0% of mass 95	73.3
175	5.0 - 9.0% of mass 174	5.9 (8.1) 1
176	95.0 - 101.0% of mass 174	72.3 (98.6) 1
177	5.0 - 9.0% of mass 176	4.6 (6.3) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VX046657.D	06/12/2025	10:11
VX0612WBL01	VX0612WBL01	VX046659.D	06/12/2025	13:10
VX0612WBS01	VX0612WBS01	VX046660.D	06/12/2025	13:38
VX0612WBSD01	VX0612WBSD01	VX046662.D	06/12/2025	14:26
OW-08B-72.5-060925	Q2275-01	VX046665.D	06/12/2025	15:46
EB01-060925	Q2275-03	VX046667.D	06/12/2025	16:32
TB01-060925	Q2275-05	VX046668.D	06/12/2025	16:55