



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Verification

Client: Europastry

SDG No.: Q2276

Project: MCUA Permit No 14241 - 571 Jersey Ave NB NJ

RunNo.: LB136102

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV						
pH		pH	7.00	7	100	90-110	06/11/2025
Sample ID:	CCV1						
pH		pH	2.01	2.00	101	90-110	06/11/2025
Sample ID:	CCV2						
pH		pH	12.02	12.00	100	90-110	06/11/2025

Initial and Continuing Calibration Verification

Client: Europastry

SDG No.: Q2276

Project: MCUA Permit No 14241 - 571 Jersey Ave NB NJ

RunNo.: LB136134

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: COD	ICV	mg/L	50.962	50	102	95-105	05/28/2025
Sample ID: COD	CCV1	mg/L	49.946	50	100	95-105	06/12/2025
Sample ID: COD	CCV2	mg/L	48.931	50	98	95-105	06/12/2025
Sample ID: COD	CCV3	mg/L	50.962	50	102	95-105	06/12/2025