

Matrix Spike Summary

Client:	Europastry	SDG No.:	Q2276
Project:	MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Sample ID:	Q2276-01
Client ID:	MH-6-10-2025MS	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Oil and Grease	mg/L	78-114	31.8		803		20.0	1	-3850	*	06/13/2025

Matrix Spike Summary

Client:	Europastry	SDG No.:	Q2276
Project:	MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Sample ID:	Q2276-01
Client ID:	MH-6-10-2025MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Oil and Grease	mg/L	78-114	32.6		803		20.0	1	-3850	*	06/13/2025

Matrix Spike Summary

Client:	Europastry	SDG No.:	Q2276
Project:	MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Sample ID:	Q2294-01
Client ID:	Outfall 1MS	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
COD	mg/L	75-125	64.2		18.5		50.0	1	91		06/12/2025

Matrix Spike Summary

Client:	Europastry	SDG No.:	Q2276
Project:	MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Sample ID:	Q2294-01
Client ID:	Outfall 1MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
COD	mg/L	75-125	64.2		18.5		50.0	1	91		06/12/2025

Matrix Spike Summary

Client:	Europastry	SDG No.:	Q2276
Project:	MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Sample ID:	Q2309-01
Client ID:	EFFLUENTMS	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Oil and Grease	mg/L	78-114	41.9		22.1		20.0	1	99		06/13/2025

Matrix Spike Summary

Client:	Europastry	SDG No.:	Q2276
Project:	MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Sample ID:	Q2309-01
Client ID:	EFFLUENTMSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Oil and Grease	mg/L	78-114	42.3		22.1		20.0	1	101		06/13/2025