

Laboratory Control Sample Summary

Client:	Europastry	SDG No.:	Q2276
Project:	MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Run No.:	LB136118

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136118BS							
BOD5	mg/L	198	188		95	1	84.6-115.4	06/11/2025

Laboratory Control Sample Summary

Client:	Europastry	SDG No.:	Q2276
Project:	MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Run No.:	LB136134

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136134BS							
COD	mg/L	50	47.9		96	1	90-110	06/12/2025

Laboratory Control Sample Summary

Client:	Europastry	SDG No.:	Q2276
Project:	MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Run No.:	LB136143

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136143BS							
Oil and Grease	mg/L	20.0	16.8		84	1	78-114	06/13/2025

Laboratory Control Sample Summary

Client: Europastry

SDG No.: Q2276

Project: MCUA Permit No 14241 - 571 Jersey Ave NB NJ

Run No.: LB136144

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136144BS							
TPH	mg/L	20.0	17.0		85	1	78-114	06/13/2025

Laboratory Control Sample Summary

Client:	Europastry	SDG No.:	Q2276
Project:	MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Run No.:	LB136144

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136144BSD							
TPH	mg/L	20.0	17.0		85	1	78-114	06/13/2025

Laboratory Control Sample Summary

Client:	Europastry	SDG No.:	Q2276
Project:	MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Run No.:	LB136145

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136145BS							
TSS	mg/L	550	532		97	1	90-110	06/13/2025