

**Metals**

- 2a -

**INITIAL AND CONTINUING CALIBRATION VERIFICATION**

**Client:** Europastry **SDG No.:** Q2276  
**Contract:** EURO03 **Lab Code:** CHEM **Case No.:** Q2276 **SAS No.:** Q2276  
**Initial Calibration Source:** EPA  
**Continuing Calibration Source:** Inorganic Ventures

---

| Sample ID | Analyte | Result<br>ug/L | True Value | %<br>Recovery | Acceptance<br>Window (%R) | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|---------|----------------|------------|---------------|---------------------------|---|------------------|------------------|---------------|
| ICV01     | Arsenic | 3730           | 4000       | 93            | 90 - 110                  | P | 06/11/2025       | 13:53            | LB136119      |
|           | Copper  | 972            | 1000       | 97            | 90 - 110                  | P | 06/11/2025       | 13:53            | LB136119      |
|           | Zinc    | 1910           | 2000       | 96            | 90 - 110                  | P | 06/11/2025       | 13:53            | LB136119      |

---

**Metals**

- 2a -

**INITIAL AND CONTINUING CALIBRATION VERIFICATION**

**Client:** Europastry **SDG No.:** Q2276  
**Contract:** EURO03 **Lab Code:** CHEM **Case No.:** Q2276 **SAS No.:** Q2276  
**Initial Calibration Source:** EPA  
**Continuing Calibration Source:** Inorganic Ventures

---

| Sample ID | Analyte | Result<br>ug/L | True Value | %<br>Recovery | Acceptance<br>Window (%R) | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|---------|----------------|------------|---------------|---------------------------|---|------------------|------------------|---------------|
| LLICV01   | Arsenic | 20.1           | 20.0       | 100           | 80 - 120                  | P | 06/11/2025       | 13:57            | LB136119      |
|           | Copper  | 21.7           | 20.0       | 109           | 80 - 120                  | P | 06/11/2025       | 13:57            | LB136119      |
|           | Zinc    | 41.1           | 40.0       | 103           | 80 - 120                  | P | 06/11/2025       | 13:57            | LB136119      |

---

**Metals**

- 2a -

**INITIAL AND CONTINUING CALIBRATION VERIFICATION**

Client: Europastry SDG No.: Q2276  
Contract: EURO03 Lab Code: CHEM Case No.: Q2276 SAS No.: Q2276  
Initial Calibration Source: EPA  
Continuing Calibration Source: Inorganic Ventures

| Sample ID | Analyte | Result<br>ug/L | True Value | %<br>Recovery | Acceptance<br>Window (%R) | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|---------|----------------|------------|---------------|---------------------------|---|------------------|------------------|---------------|
| CCV01     | Arsenic | 4970           | 5000       | 99            | 90 - 110                  | P | 06/11/2025       | 14:42            | LB136119      |
|           | Copper  | 1240           | 1250       | 99            | 90 - 110                  | P | 06/11/2025       | 14:42            | LB136119      |
|           | Zinc    | 2490           | 2500       | 100           | 90 - 110                  | P | 06/11/2025       | 14:42            | LB136119      |
| CCV02     | Arsenic | 4770           | 5000       | 95            | 90 - 110                  | P | 06/11/2025       | 16:33            | LB136119      |
|           | Copper  | 1210           | 1250       | 97            | 90 - 110                  | P | 06/11/2025       | 16:33            | LB136119      |
|           | Zinc    | 2430           | 2500       | 97            | 90 - 110                  | P | 06/11/2025       | 16:33            | LB136119      |
| CCV03     | Arsenic | 4820           | 5000       | 96            | 90 - 110                  | P | 06/11/2025       | 17:27            | LB136119      |
|           | Copper  | 1220           | 1250       | 98            | 90 - 110                  | P | 06/11/2025       | 17:27            | LB136119      |
|           | Zinc    | 2400           | 2500       | 96            | 90 - 110                  | P | 06/11/2025       | 17:27            | LB136119      |
| CCV04     | Arsenic | 4860           | 5000       | 97            | 90 - 110                  | P | 06/11/2025       | 18:21            | LB136119      |
|           | Copper  | 1230           | 1250       | 98            | 90 - 110                  | P | 06/11/2025       | 18:21            | LB136119      |
|           | Zinc    | 2480           | 2500       | 99            | 90 - 110                  | P | 06/11/2025       | 18:21            | LB136119      |
| CCV05     | Arsenic | 4810           | 5000       | 96            | 90 - 110                  | P | 06/11/2025       | 19:09            | LB136119      |
|           | Copper  | 1220           | 1250       | 97            | 90 - 110                  | P | 06/11/2025       | 19:09            | LB136119      |
|           | Zinc    | 2460           | 2500       | 98            | 90 - 110                  | P | 06/11/2025       | 19:09            | LB136119      |
| CCV06     | Arsenic | 4810           | 5000       | 96            | 90 - 110                  | P | 06/11/2025       | 20:06            | LB136119      |
|           | Copper  | 1190           | 1250       | 96            | 90 - 110                  | P | 06/11/2025       | 20:06            | LB136119      |
|           | Zinc    | 2430           | 2500       | 97            | 90 - 110                  | P | 06/11/2025       | 20:06            | LB136119      |

**Metals**

- 2a -

**INITIAL AND CONTINUING CALIBRATION VERIFICATION**

**Client:** Europastry **SDG No.:** Q2276  
**Contract:** EURO03 **Lab Code:** CHEM **Case No.:** Q2276 **SAS No.:** Q2276  
**Initial Calibration Source:** EPA  
**Continuing Calibration Source:** Inorganic Ventures

---

| Sample ID | Analyte | Result<br>ug/L | True Value | %<br>Recovery | Acceptance<br>Window (%R) | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|---------|----------------|------------|---------------|---------------------------|---|------------------|------------------|---------------|
| ICV01     | Arsenic | 3860           | 4000       | 96            | 90 - 110                  | P | 06/12/2025       | 17:58            | LB136140      |
|           | Copper  | 1000           | 1000       | 100           | 90 - 110                  | P | 06/12/2025       | 17:58            | LB136140      |
|           | Zinc    | 2040           | 2000       | 102           | 90 - 110                  | P | 06/12/2025       | 17:58            | LB136140      |

---

**Metals**

- 2a -

**INITIAL AND CONTINUING CALIBRATION VERIFICATION**

**Client:** Europastry **SDG No.:** Q2276  
**Contract:** EURO03 **Lab Code:** CHEM **Case No.:** Q2276 **SAS No.:** Q2276  
**Initial Calibration Source:** EPA  
**Continuing Calibration Source:** Inorganic Ventures

---

| Sample ID | Analyte | Result<br>ug/L | True Value | %<br>Recovery | Acceptance<br>Window (%R) | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|---------|----------------|------------|---------------|---------------------------|---|------------------|------------------|---------------|
| LLICV01   | Arsenic | 22.0           | 20.0       | 110           | 80 - 120                  | P | 06/12/2025       | 18:02            | LB136140      |
|           | Copper  | 21.2           | 20.0       | 106           | 80 - 120                  | P | 06/12/2025       | 18:02            | LB136140      |
|           | Zinc    | 42.5           | 40.0       | 106           | 80 - 120                  | P | 06/12/2025       | 18:02            | LB136140      |

---

**Metals**

- 2a -

**INITIAL AND CONTINUING CALIBRATION VERIFICATION**

Client: Europastry SDG No.: Q2276  
Contract: EURO03 Lab Code: CHEM Case No.: Q2276 SAS No.: Q2276  
Initial Calibration Source: EPA  
Continuing Calibration Source: Inorganic Ventures

| Sample ID | Analyte | Result<br>ug/L | True Value | %<br>Recovery | Acceptance<br>Window (%R) | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|---------|----------------|------------|---------------|---------------------------|---|------------------|------------------|---------------|
| CCV01     | Arsenic | 4970           | 5000       | 99            | 90 - 110                  | P | 06/12/2025       | 18:50            | LB136140      |
|           | Copper  | 1240           | 1250       | 99            | 90 - 110                  | P | 06/12/2025       | 18:50            | LB136140      |
|           | Zinc    | 2500           | 2500       | 100           | 90 - 110                  | P | 06/12/2025       | 18:50            | LB136140      |
| CCV02     | Arsenic | 5010           | 5000       | 100           | 90 - 110                  | P | 06/12/2025       | 19:42            | LB136140      |
|           | Copper  | 1260           | 1250       | 101           | 90 - 110                  | P | 06/12/2025       | 19:42            | LB136140      |
|           | Zinc    | 2540           | 2500       | 101           | 90 - 110                  | P | 06/12/2025       | 19:42            | LB136140      |
| CCV03     | Arsenic | 4770           | 5000       | 95            | 90 - 110                  | P | 06/13/2025       | 01:46            | LB136140      |
|           | Copper  | 1200           | 1250       | 96            | 90 - 110                  | P | 06/13/2025       | 01:46            | LB136140      |
|           | Zinc    | 2480           | 2500       | 99            | 90 - 110                  | P | 06/13/2025       | 01:46            | LB136140      |
| CCV04     | Arsenic | 4750           | 5000       | 95            | 90 - 110                  | P | 06/13/2025       | 01:59            | LB136140      |
|           | Copper  | 1190           | 1250       | 95            | 90 - 110                  | P | 06/13/2025       | 01:59            | LB136140      |
|           | Zinc    | 2410           | 2500       | 97            | 90 - 110                  | P | 06/13/2025       | 01:59            | LB136140      |