



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

**Metals**

**- 2b -**

**CRDL STANDARD FOR AA & ICP**

**Client:** Europastry **SDG No.:** Q2276  
**Contract:** EURO03 **Lab Code:** CHEM **Case No.:** Q2276 **SAS No.:** Q2276  
**Initial Calibration Source:** \_\_\_\_\_  
**Continuing Calibration Source:** \_\_\_\_\_

| Sample ID | Analyte | Result<br>ug/L | True Value<br>ug/L | %<br>Recovery | Acceptance<br>Window (%R) | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|---------|----------------|--------------------|---------------|---------------------------|---|------------------|------------------|---------------|
| CRI01     | Arsenic | 18.7           | 20.0               | 94            | 65 - 135                  | P | 06/11/2025       | 14:07            | LB136119      |
|           | Copper  | 21.3           | 20.0               | 106           | 65 - 135                  | P | 06/11/2025       | 14:07            | LB136119      |
|           | Zinc    | 41.3           | 40.0               | 103           | 65 - 135                  | P | 06/11/2025       | 14:07            | LB136119      |
| CRI01     | Arsenic | 20.7           | 20.0               | 104           | 65 - 135                  | P | 06/12/2025       | 18:12            | LB136140      |
|           | Copper  | 21.8           | 20.0               | 109           | 65 - 135                  | P | 06/12/2025       | 18:12            | LB136140      |
|           | Zinc    | 43.4           | 40.0               | 108           | 65 - 135                  | P | 06/12/2025       | 18:12            | LB136140      |