



- 3a -

Client:	Europastry			SDG No.:	Q2276
Contract:	EURO03	Lab Code:	CHEM	Case No.:	Q2276
				SAS No.:	Q2276

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Arsenic	5.12	+/-10	U	20.0	P	06/11/2025	14:03	LB136119
	Copper	4.60	+/-10	U	20.0	P	06/11/2025	14:03	LB136119
	Zinc	16.7	+/-20	U	40.0	P	06/11/2025	14:03	LB136119

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Europastry</u>		SDG No.: <u>Q2276</u>							
Contract: <u>EURO03</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q2276</u>	SAS No.: <u>Q2276</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Arsenic	5.12	+/-10	U	20.0	P	06/11/2025	14:46	LB136119
	Copper	4.60	+/-10	U	20.0	P	06/11/2025	14:46	LB136119
	Zinc	16.7	+/-20	U	40.0	P	06/11/2025	14:46	LB136119
CCB02	Arsenic	5.12	+/-10	U	20.0	P	06/11/2025	16:44	LB136119
	Copper	4.60	+/-10	U	20.0	P	06/11/2025	16:44	LB136119
	Zinc	16.7	+/-20	U	40.0	P	06/11/2025	16:44	LB136119
CCB03	Arsenic	5.12	+/-10	U	20.0	P	06/11/2025	17:34	LB136119
	Copper	4.60	+/-10	U	20.0	P	06/11/2025	17:34	LB136119
	Zinc	16.7	+/-20	U	40.0	P	06/11/2025	17:34	LB136119
CCB04	Arsenic	5.12	+/-10	U	20.0	P	06/11/2025	18:25	LB136119
	Copper	4.60	+/-10	U	20.0	P	06/11/2025	18:25	LB136119
	Zinc	16.7	+/-20	U	40.0	P	06/11/2025	18:25	LB136119
CCB05	Arsenic	5.12	+/-10	U	20.0	P	06/11/2025	19:14	LB136119
	Copper	4.60	+/-10	U	20.0	P	06/11/2025	19:14	LB136119
	Zinc	16.7	+/-20	U	40.0	P	06/11/2025	19:14	LB136119
CCB06	Arsenic	5.12	+/-10	U	20.0	P	06/11/2025	20:11	LB136119
	Copper	4.60	+/-10	U	20.0	P	06/11/2025	20:11	LB136119
	Zinc	16.7	+/-20	U	40.0	P	06/11/2025	20:11	LB136119

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Europastry

SDG No.: Q2276

Contract: EURO03

Lab Code: CHEM

Case No.: Q2276

SAS No.: Q2276

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Arsenic	5.12	+/-10	U	20.0	P	06/12/2025	18:06	LB136140
	Copper	4.60	+/-10	U	20.0	P	06/12/2025	18:06	LB136140
	Zinc	16.7	+/-20	U	40.0	P	06/12/2025	18:06	LB136140

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Europastry **SDG No.:** Q2276
Contract: EURO03 **Lab Code:** CHEM **Case No.:** Q2276 **SAS No.:** Q2276

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Arsenic	5.12	+/-10	U	20.0	P	06/12/2025	18:55	LB136140
	Copper	4.60	+/-10	U	20.0	P	06/12/2025	18:55	LB136140
	Zinc	16.7	+/-20	U	40.0	P	06/12/2025	18:55	LB136140
CCB02	Arsenic	5.12	+/-10	U	20.0	P	06/12/2025	19:47	LB136140
	Copper	4.60	+/-10	U	20.0	P	06/12/2025	19:47	LB136140
	Zinc	16.7	+/-20	U	40.0	P	06/12/2025	19:47	LB136140
CCB03	Arsenic	5.12	+/-10	U	20.0	P	06/13/2025	01:50	LB136140
	Copper	4.60	+/-10	U	20.0	P	06/13/2025	01:50	LB136140
	Zinc	16.7	+/-20	U	40.0	P	06/13/2025	01:50	LB136140
CCB04	Arsenic	5.12	+/-10	U	20.0	P	06/13/2025	02:03	LB136140
	Copper	4.60	+/-10	U	20.0	P	06/13/2025	02:03	LB136140
	Zinc	16.7	+/-20	U	40.0	P	06/13/2025	02:03	LB136140